

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/29/2024

Lab File ID: BF140100.D Time Analyzed: 09:37

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	138305	6.893	518016	8.18	286088	9.93
UPPER LIMIT	276610	7.393	1036030	8.675	572176	10.428
LOWER LIMIT	69152.5	6.393	259008	7.675	143044	9.428
EPA SAMPLE NO.						
01 PT-TRIAZINE-SOIL	114288	6.89	452815	8.17	254415	9.92

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.