

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4495 SAS No.: P4495 SDG NO.: P4495

EPA Sample No.: SSTDCCC040 Date Analyzed: 10/29/2024

Lab File ID: BF140100.D Time Analyzed: 09:37

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	513809	11.416	264357	14.051	264628	15.527
UPPER LIMIT	1027620	11.916	528714	14.551	529256	16.027
LOWER LIMIT	256905	10.916	132179	13.551	132314	15.027
EPA SAMPLE NO.						
01 PT-TRIAZINE-SOIL	445374	11.41	216948	14.05	229298	15.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.