

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CHEM02

Lab Code: CHEM

SAS No.: P4495 SDG NO.: P4495

Lab File ID: BF140049.D

DFTPP Injection Date: 10/26/2024

Instrument ID: BNA_F

DFTPP Injection Time: 10:10

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	42.1
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	38.1
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	47.5
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
275	10.0 - 60.0% of mass 198	28.3
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	14.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.1 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140050.D	10/26/2024	10:38
PB164401BL	PB164401BL	BF140051.D	10/26/2024	11:06
PB164401BS	PB164401BS	BF140052.D	10/26/2024	11:34
BP-F-20MS	P4547-05MS	BF140057.D	10/26/2024	13:58
BP-F-20MSD	P4547-05MSD	BF140058.D	10/26/2024	14:27