

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51698 MA No. : SDG No.: MBHCH8
Initial Calibration Verification Source : WP110398
Continuing Calibration Verification Source : WP110399
Run Batch: LB133132 Analytical Method: Spectrophotometry
Concentration Units: µg/L

| | Initial Calibration Verification | | | | Continuing Calibration Verification | | | | | | |
|---------|----------------------------------|-------|----|------|-------------------------------------|-------|----|------|------------|----|------|
| | ID: ICV001 | | | | ID: CCV001 | | | | ID: CCV002 | | |
| Analyte | True | Found | %R | %RSD | True | Found | %R | %RSD | Found | %R | %RSD |
| Cyanide | 99.0 | 97 | 98 | | 250 | 250 | 99 | | 240 | 95 | |

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51698 MA No. : SDG No.: MBHCH8
Initial Calibration Verification Source : WP110398
Continuing Calibration Verification Source : WP110399
Run Batch: LB133132 Analytical Method: Spectrophotometry
Concentration Units: µg/L

| | Initial Calibration Verification | | | | Continuing Calibration Verification | | | | | | |
|---------|----------------------------------|-------|----|------|-------------------------------------|-------|-----|------|-------|----|------|
| | ID: | | | | ID: CCV003 | | | | ID: | | |
| Analyte | True | Found | %R | %RSD | True | Found | %R | %RSD | Found | %R | %RSD |
| Cyanide | | | | | 250 | 260 | 102 | | | | |