EPA SAMPLE NO.

MBHCN0A

## FORM 5B - IN

## POST-DIGESTION/DISTILLATION SPIKE SAMPLE RECOVERY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51698 MA No.: SDG No.: MBHCJ7

Matrix : Soil Analytical Method: ICP-AES

% Solids: 85.6

Concentration Units ( $\mu g/L$ , mg/L or mg/kg dry weight): mg/kg

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	%R	Q
Cobalt	75 - 125	92		48		94.0	47	
Copper	75 - 125	47		27		54.0	37	
Thallium	75 - 125	2.2	Ū	2.2	Ū	4.3	0	