

## FORM 2 - IN

## INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51698 MA No. :                      SDG No.: MBHCJ7  
Initial Calibration Verification Source : WP110398  
Continuing Calibration Verification Source : WP110399  
Run Batch: LB133132 Analytical Method: Spectrophotometry  
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Cyanide	99.0	97	98		250	250	99		240	95	

## FORM 2 - IN

## INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
Lab Code: ACE Case No.: 51698 MA No. :                      SDG No.: MBHCJ7  
Initial Calibration Verification Source : WP110398  
Continuing Calibration Verification Source : WP110399  
Run Batch: LB133132 Analytical Method: Spectrophotometry  
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV003				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Cyanide					250	260	102				