

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133346

Review By	Sarabjit Jaswal	Review On	11/8/2024 8:02:58 AM
Supervise By	Mohan Bera	Supervise On	11/8/2024 8:03:53 AM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	
Chk Standard	MP83148,MP83149

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/07/24 16:24		Kareem	OK
2	S1	S01	CAL2	11/07/24 16:29		Kareem	OK
3	S2	S02	CAL3	11/07/24 16:33		Kareem	OK
4	S3	S03	CAL4	11/07/24 16:38		Kareem	OK
5	S4	S04	CAL5	11/07/24 16:42		Kareem	OK
6	S5	S05	CAL6	11/07/24 16:46		Kareem	OK
7	S6	S06	CAL7	11/07/24 16:51		Kareem	OK
8	ICV005	ICV005	ICV	11/07/24 17:13	Inst. paused for calibration review	Kareem	OK
9	ICB005	ICB005	ICB	11/07/24 17:19	Inst. paused for review	Kareem	OK
10	ICSA005	ICSA005	ICSA	11/07/24 17:24		Kareem	OK
11	ICSAB005	ICSAB005	ICSAB	11/07/24 17:28		Kareem	OK
12	CCV029	CCV029	CCV	11/07/24 17:32		Kareem	OK
13	CCB029	CCB029	CCB	11/07/24 17:37		Kareem	OK
14	P4480-05A	MYE494A	PS	11/07/24 18:17	NOT USE	Kareem	Not Ok
15	P4481-12A	MYE4C1A	PS	11/07/24 18:21	NOT USE	Kareem	Not Ok
16	P4482-14A	MYE4E3A	PS	11/07/24 18:26	NOT USE	Kareem	Not Ok
17	P4483-20A	MYE4G9A	PS	11/07/24 18:30	NOT USE	Kareem	Not Ok
18	P4484-03ADL	MYE4H2A	PS	11/07/24 18:35	PS for Cr-0.031ml of M5658 - 10ml sample	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133346

Review By	Sarabjit Jaswal	Review On	11/8/2024 8:02:58 AM
Supervise By	Mohan Bera	Supervise On	11/8/2024 8:03:53 AM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	
Chk Standard	MP83148,MP83149

19	P4496-07A	MBHZC8A	PS	11/07/24 18:39	PS For Pb-0.018ml of MP83156, Ni-0.035ml of M5961, Se-0.07ml of MP83152, TI-0.05ml of MP83153	Kareem	OK
20	P4497-20A	MBHDD0A	PS	11/07/24 18:43	NOT USE	Kareem	Not Ok
21	P4498-10A	MBHDF0A	PS	11/07/24 18:48	PS For As-0.02ml MP83154, Pb-0.025ml of MP83156, Se-0.07ml of MP83152, TI-0.05ml-MP83153	Kareem	OK
22	P4499-20A	MBHCN0A	PS	11/07/24 18:52	NOT USE	Kareem	Not Ok
23	P4500-10A	MBHCP0A	PS	11/07/24 18:57	PS For As-0.02ml MP83154, Pb-0.025ml of MP83156, TI-0.05ml of MP83153	Kareem	OK
24	P4501-13A	MBHCW0A	PS	11/07/24 19:01	NOT USE	Kareem	Not Ok
25	P4519-03ADL	MYE4K2A	PS	11/07/24 19:05	PS for Pb-0.021ml of MP83156	Kareem	OK
26	CCV030	CCV030	CCV	11/07/24 19:10		Kareem	OK
27	CCB030	CCB030	CCB	11/07/24 19:14		Kareem	OK
28	P4480-05ADL	MYE494A	PS	11/07/24 19:25	PS for Pb-0.0028ml of M6025,Se-0.07ml of MP83152 - 10 ml sample	Kareem	OK
29	P4481-12ADL	MYE4C1A	PS	11/07/24 19:29	PS for As-0.0036ml of M5801,Cu-0.012ml of M5751,Zn-0.011ml of M5819 - 10ml sample	Kareem	OK
30	P4482-14ADL	MYE4E3A	PS	11/07/24 19:34	NOT USE	Kareem	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133346

Review By	Sarabjit Jaswal	Review On	11/8/2024 8:02:58 AM
Supervise By	Mohan Bera	Supervise On	11/8/2024 8:03:53 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

31	P4483-20ADL	MYE4G9A	PS	11/07/24 19:38	NOT USE	Kareem	Not Ok
32	P4496-07ADL	MBHZC8A	PS	11/07/24 19:43	NOT USE	Kareem	Not Ok
33	P4501-13ADL	MBHCW0A	PS	11/07/24 19:47	NOT USE	Kareem	Not Ok
34	CCV031	CCV031	CCV	11/07/24 19:51		Kareem	OK
35	CCB031	CCB031	CCB	11/07/24 19:56		Kareem	OK
36	P4497-11DL	MBHDB6	SAM	11/07/24 21:23	5X for Ca	Kareem	Confirms
37	P4497-13DL	MBHDB8	SAM	11/07/24 21:27	5X for Ni	Kareem	Confirms
38	P4499-19DL	MBHCM9	SAM	11/07/24 21:32	5X for Ca	Kareem	Confirms
39	CCV032	CCV032	CCV	11/07/24 21:36		Kareem	OK
40	CCB032	CCB032	CCB	11/07/24 21:40		Kareem	OK
41	P4499-20DL	MBHCN0	SAM	11/07/24 21:45	3X for Ca,Mg,Ni	Kareem	Confirms
42	P4499-21DL	MBHCN0D	DUP	11/07/24 21:49	3X for Ca,Mg,Ni	Kareem	Confirms
43	P4499-20LDL	MBHCN0L	SD	11/07/24 21:54	15X for Ca,Mg,Ni	Kareem	Confirms
44	P4499-22DL	MBHCN0S	MS	11/07/24 21:58	3X for Ca,Mg,Ni	Kareem	Confirms
45	P4496-04DL	MBHDO5	SAM	11/07/24 22:03	5X for Ca	Kareem	Confirms
46	P4496-05DL	MBHDO6	SAM	11/07/24 22:07	4X for Ca	Kareem	Confirms
47	P4496-06DL	MBHZC7	SAM	11/07/24 22:12	2X for Ca	Kareem	Confirms
48	P4496-07DL	MBHZC8	SAM	11/07/24 22:16	2X for Ca	Kareem	Confirms
49	P4496-08DL	MBHZC8D	DUP	11/07/24 22:20	2X for Ca	Kareem	Confirms
50	P4496-07LDL	MBHZC8L	SD	11/07/24 22:25	10X for Ca	Kareem	Confirms

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133346

Review By	Sarabjit Jaswal	Review On	11/8/2024 8:02:58 AM
Supervise By	Mohan Bera	Supervise On	11/8/2024 8:03:53 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

51	P4496-09DL	MBHZC8S	MS	11/07/24 22:29	2X for Ca	Kareem	Confirms
52	P4498-02DL	MBHDD2	SAM	11/07/24 22:39	5X for Ca	Kareem	Confirms
53	P4498-03DL	MBHDD8	SAM	11/07/24 22:44	2X For Ca,Mg,Ni High, Still Mg,Ni High	Kareem	Dilution
54	P4498-03DL2	MBHDD8	SAM	11/07/24 22:48	5X for Mg,Ni	Kareem	Confirms
55	P4498-04DL	MBHDD9	SAM	11/07/24 22:53	2X for Ca,Ni	Kareem	Confirms
56	P4498-05DL	MBHDE0	SAM	11/07/24 22:57	2X for Ca	Kareem	Confirms
57	P4498-06DL	MBHDE1	SAM	11/07/24 23:01	2X for Ni	Kareem	Confirms
58	P4498-07DL	MBHDE2	SAM	11/07/24 23:06	2X for Ni	Kareem	Confirms
59	P4498-08DL	MBHDE8	SAM	11/07/24 23:10	5X for Ca	Kareem	Confirms
60	P4498-09DL	MBHDE9	SAM	11/07/24 23:15	5X for Ca	Kareem	Confirms
61	CCV033	CCV033	CCV	11/07/24 23:19		Kareem	OK
62	CCB033	CCB033	CCB	11/07/24 23:23		Kareem	OK
63	P4498-10DL	MBHDF0	SAM	11/07/24 23:28	2x for Ni	Kareem	Confirms
64	P4498-11DL	MBHDF0D	DUP	11/07/24 23:32	2x for Ni	Kareem	Confirms
65	P4498-10LDL	MBHDF0L	SD	11/07/24 23:37	10x for Ni	Kareem	Confirms
66	P4498-12DL	MBHDF0S	MS	11/07/24 23:41	2x for Ni	Kareem	Confirms
67	P4498-13DL	MBHDF1	SAM	11/07/24 23:45	5x for Ca	Kareem	Confirms
68	P4498-14DL	MBHDF2	SAM	11/07/24 23:50	10x for Ca	Kareem	Confirms
69	P4498-15DL	MBHCH8	SAM	11/07/24 23:54	2X For Ca,Ni	Kareem	Confirms

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133346

Review By	Sarabjit Jaswal	Review On	11/8/2024 8:02:58 AM
Supervise By	Mohan Bera	Supervise On	11/8/2024 8:03:53 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

70	P4498-16DL	MBHCH9	SAM	11/07/24 23:59	3X For Ni	Kareem	Confirms
71	P4498-20DL	MBHCJ3	SAM	11/08/24 00:03	3X For Ca,Ni	Kareem	Confirms
72	P4498-21DL	MBHCJ5	SAM	11/08/24 00:08	5X For Ca	Kareem	Confirms
73	P4498-22DL	MBHCJ6	SAM	11/08/24 00:12	2X For Ca	Kareem	Confirms
74	CCV034	CCV034	CCV	11/08/24 00:16		Kareem	OK
75	CCB034	CCB034	CCB	11/08/24 00:21		Kareem	OK