

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51698 MA No. : SDG No.: MBHCN1

Initial Calibration Verification Source : MP83113

Continuing Calibration Verification Source : MP83115

Run Batch: LB133324 Analytical Method: CVAA

Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV072				ID: CCV040				ID: CCV041		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.8	94		5.0	5.0	99		5.1	102	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51698 MA No. : SDG No.: MBHCN1
Initial Calibration Verification Source : MP83113
Continuing Calibration Verification Source : MP83115
Run Batch: LB133324 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV042				ID: CCV043		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	5.1	102		5.1	103	