

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51698 MA No.: _____ SDG No.: MBHCS8
Instrument ID: CV1 Start Date: 11/07/2024
Analytical Method: CVAA Run Batch: LB133324
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0	0	0.2	0.19	-3	2.5	2.5	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51698 MA No.: _____ SDG No.: MBHCS8
Instrument ID: CV1 Start Date: 11/07/2024
Analytical Method: CVAA Run Batch: LB133324
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	4.9	-1	7.5	7.7	2	10	9.9	-1

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51698 MA No.: SDG No.: MBHCS8
Instrument ID: CV1 Start Date: 11/08/2024
Analytical Method: CVAA Run Batch: LB133356
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0	0	0.2	0.22	12	2.5	2.4	-2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51698 MA No.: _____ SDG No.: MBHCS8
Instrument ID: CV1 Start Date: 11/08/2024
Analytical Method: CVAA Run Batch: LB133356
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.0	1	7.5	7.7	2	10	9.9	-1