

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51698 MA No.: SDG No.: MBHCS8
Instrument ID: CV1 Start Date : 11/07/2024
Analytical Method: CVAA Run Batch : LB133324

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Mercury	0.999723	0.000229	-0.195606	Lin. Reg	NONE

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51698 MA No.: SDG No.: MBHCS8
Instrument ID: CV1 Start Date : 11/08/2024
Analytical Method: CVAA Run Batch : LB133356

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Mercury	0.999741	0.000266	-0.227364	Lin. Reg	NONE

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51698 MA No.: SDG No.: MBHCS8
Instrument ID: CV1 Start Date : 11/08/2024
Analytical Method: CVAA Run Batch : LB133356

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------