

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51698 MA No. : SDG No.: MBHCX4
 Initial Calibration Verification Source : MP83182
 Continuing Calibration Verification Source : MP83184
 Run Batch: LB133377 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV079				ID: CCV061				ID: CCV062		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	4.2	106		5.0	5.3	106		4.6	92	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51698 MA No. : SDG No.: MBHCX4
Initial Calibration Verification Source : MP83182
Continuing Calibration Verification Source : MP83184
Run Batch: LB133377 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV063				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	5.3	106				