

Laboratory Control Sample Summary

Client: Tetra Tech, EMI

SDG No.: P4506

Project: R36720

Run No.: LB133104

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133104BS							
TSS	mg/L	550	542		98	1	90-110	10/24/2024