

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: ENTA05
 ProjectID: North Point
 Lab Code: CHEM Case No.: P4518 SAS No.: P4518 SDG No.: P4518

Calibration Sequence : FF102124		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	140350994	140351	FF014706.D
500	71252825	142506	FF014707.D
200	29731497	148657	FF014708.D
100	14736133	147361	FF014709.D
50	7756416	155128	FF014710.D
AVG RF : 146801		% RSD : 3.93	AVG RT : 14.9988