

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF102524\
 Data File : FF014799.D
 Signal(s) : FID2B.ch
 Acq On : 25 Oct 2024 20:41
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
 Quant Time: Oct 26 04:05:35 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF102124.M
 Quant Title :
 QLast Update : Tue Oct 22 08:35:55 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.020	6049826	45.971 ug/ml
Target Compounds			
1) N-OCTANE	2.049	6594862	47.400 ug/ml
2) N-DECANE	4.575	7169971	50.191 ug/ml
3) N-DODECANE	6.739	7196331	49.591 ug/ml
4) N-TETRADECANE	8.564	7152038	48.791 ug/ml
5) N-HEXADECANE	10.170	7106577	48.360 ug/ml
6) N-OCTADECANE	11.613	7150274	47.628 ug/ml
7) N-EICOSANE	12.923	7146548	47.109 ug/ml
8) N-DOCOSANE	14.121	6850161	46.552 ug/ml
10) N-TETRACOSANE	15.224	6773583	46.089 ug/ml
11) N-HEXACOSANE	16.244	6624023	45.351 ug/ml
12) N-OCTACOSANE	17.194	6566333	45.445 ug/ml
13) N-TRIACONTANE	18.081	6735616	46.011 ug/ml
14) N-DOTRIACONTANE	18.914	6744020	46.877 ug/ml
15) N-TETRATRIACONTANE	19.697	6384220	46.551 ug/ml
16) N-HEXATRIACONTANE	20.436	6041323	46.568 ug/ml
17) N-OCTATRIACONTANE	21.172	5764722	50.121 ug/ml
18) N-TETRACONTANE	22.083	5514975	52.526 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_F\Data\FF102524\
Data File : FF014799.D
Signal(s) : FID2B.ch
Acq On : 25 Oct 2024 20:41
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: Sample.e
Quant Time: Oct 26 04:05:35 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_F\Method\FF102124.M
Quant Title :
QLast Update : Tue Oct 22 08:35:55 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

