

SOIL DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RI

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4518 **SAS No :** P4518 **SDG No:** P4518
Matrix Spike - EPA Sample No : PB164381BS **Datafile:** FF014760.D

COMPOUND	SPIKE ADDED ug/kg	CONCENTRATION ug/kg	LCS/LCSD CONCENTRATION ug/kg	% REC	QC LIMITS
DRO	6664	0	6639	100	68-131