

Continuing Calibration Report for SequenceID : FC102524AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067525.D

Aliphatic C9-C12	8099461.000	60.000	3.173	6.453	134991.017	135469.641	0.353
Aliphatic C12-C16	5283502.000	40.000	6.454	9.843	132087.550	131681.730	-0.308
Aliphatic C16-C21	7645839.000	60.000	9.844	13.201	127430.650	127235.092	-0.154
Aliphatic C21-C28	9642398.000	80.000	13.202	16.856	120529.975	120090.398	-0.366
Aliphatic C28-C40	12499640.000	120.000	16.857	21.702	104163.667	101093.195	-3.037
Aliphatic EPH	43170840.000	360.000	3.173	21.702	119919.000	118799.912	-0.942