

Continuing Calibration Report for SequenceID : FC102824AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067534.D

Aliphatic C16-C21	7789407.000	60.000	9.843	13.200	129823.450	127235.092	-2.034
Aliphatic C21-C28	9829212.000	80.000	13.201	16.853	122865.150	120090.398	-2.311
Aliphatic C28-C40	12777442.000	120.000	16.854	21.695	106478.683	101093.195	-5.327
Aliphatic EPH	43937898.000	360.000	3.173	21.695	122049.717	118799.912	-2.736
Aliphatic C9-C12	8169484.000	60.000	3.173	6.453	136158.067	135469.641	-0.508
Aliphatic C12-C16	5372353.000	40.000	6.454	9.842	134308.825	131681.730	-1.995