

Continuing Calibration Report for SequenceID : FC102824AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067541.D

Aliphatic C16-C21	7857628.000	60.000	9.843	13.200	130960.467	127235.092	-2.928
Aliphatic C21-C28	9913509.000	80.000	13.201	16.853	123918.863	120090.398	-3.188
Aliphatic C28-C40	13005709.000	120.000	16.854	21.695	108380.908	101093.195	-7.209
Aliphatic EPH	44286148.000	360.000	3.173	21.695	123017.078	118799.912	-3.550
Aliphatic C9-C12	8119479.000	60.000	3.173	6.453	135324.650	135469.641	0.107
Aliphatic C12-C16	5389823.000	40.000	6.454	9.842	134745.575	131681.730	-2.327