

Continuing Calibration Report for SequenceID : FC102824AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067552.D

Aliphatic C16-C21	7778385.000	60.000	9.843	13.200	129639.750	127235.092	-1.890
Aliphatic C21-C28	9854845.000	80.000	13.201	16.853	123185.563	120090.398	-2.577
Aliphatic C28-C40	13169327.000	120.000	16.854	21.695	109744.392	101093.195	-8.558
Aliphatic EPH	44364168.000	360.000	3.173	21.695	123233.800	118799.912	-3.732
Aliphatic C9-C12	8190095.000	60.000	3.173	6.453	136501.583	135469.641	-0.762
Aliphatic C12-C16	5371516.000	40.000	6.454	9.842	134287.900	131681.730	-1.979