

Continuing Calibration Report for SequenceID : FC102824AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067561.D

Aliphatic C16-C21	7784943.000	60.000	9.843	13.200	129749.050	127235.092	-1.976
Aliphatic C21-C28	9891372.000	80.000	13.201	16.853	123642.150	120090.398	-2.958
Aliphatic C28-C40	13200046.000	120.000	16.854	21.695	110000.383	101093.195	-8.811
Aliphatic EPH	44290968.000	360.000	3.173	21.695	123030.467	118799.912	-3.561
Aliphatic C9-C12	8074954.000	60.000	3.173	6.453	134582.567	135469.641	0.655
Aliphatic C12-C16	5339653.000	40.000	6.454	9.842	133491.325	131681.730	-1.374