

Continuing Calibration Report for SequenceID : FD102524AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048587.D

Aromatic C10-C12	7311656.000	40.000	4.094	5.812	182791.400	176202.694	-3.739
Aromatic C12-C16	11049841.000	60.000	5.813	8.418	184164.017	177273.493	-3.887
Aromatic C16-C21	13535485.000	80.000	8.419	12.680	169193.563	163832.550	-3.272
Aromatic C21-C36	28201866.000	180.000	12.681	18.085	156677.033	152713.153	-2.596
Aromatic EPH	60098848.000	360.000	4.094	18.085	166941.244	161887.469	-3.122