

Continuing Calibration Report for SequenceID : FD102524AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048594.D

Aromatic C10-C12	7580409.000	40.000	4.094	5.812	189510.225	176202.694	-7.552
Aromatic C12-C16	11427061.000	60.000	5.813	8.418	190451.017	177273.493	-7.433
Aromatic C16-C21	13923239.000	80.000	8.419	12.680	174040.488	163832.550	-6.231
Aromatic C21-C36	28939237.000	180.000	12.681	18.085	160773.539	152713.153	-5.278
Aromatic EPH	61869946.000	360.000	4.094	18.085	171860.961	161887.469	-6.161