

Continuing Calibration Report for SequenceID : FD102824AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048597.D

Aromatic C10-C12	6478652.000	40.000	4.094	5.812	161966.300	176202.694	8.080
Aromatic C12-C16	9681948.000	60.000	5.813	8.417	161365.800	177273.493	8.974
Aromatic C16-C21	11930066.000	80.000	8.418	12.677	149125.825	163832.550	8.977
Aromatic C21-C36	25141839.000	180.000	12.678	18.081	139676.883	152713.153	8.536
Aromatic EPH	53232505.000	360.000	4.094	18.081	147868.069	161887.469	8.660