

Continuing Calibration Report for SequenceID : FD102824AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048602.D

Aromatic C10-C12	6734968.000	40.000	4.094	5.812	168374.200	176202.694	4.443
Aromatic C12-C16	10158777.000	60.000	5.813	8.417	169312.950	177273.493	4.491
Aromatic C16-C21	12462820.000	80.000	8.418	12.677	155785.250	163832.550	4.912
Aromatic C21-C36	25673678.000	180.000	12.678	18.081	142631.544	152713.153	6.602
Aromatic EPH	55030243.000	360.000	4.094	18.081	152861.786	161887.469	5.575