

Continuing Calibration Report for SequenceID : FD102824AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048614.D

Aromatic C10-C12	6535868.000	40.000	4.094	5.812	163396.700	176202.694	7.268
Aromatic C12-C16	9728504.000	60.000	5.813	8.417	162141.733	177273.493	8.536
Aromatic C16-C21	11847029.000	80.000	8.418	12.677	148087.863	163832.550	9.610
Aromatic C21-C36	25259345.000	180.000	12.678	18.081	140329.694	152713.153	8.109
Aromatic EPH	53370746.000	360.000	4.094	18.081	148252.072	161887.469	8.423