

Continuing Calibration Report for SequenceID : FD102824AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048619.D

Aromatic C10-C12	6474941.000	40.000	4.094	5.812	161873.525	176202.694	8.132
Aromatic C12-C16	9658423.000	60.000	5.813	8.417	160973.717	177273.493	9.195
Aromatic C16-C21	11797080.000	80.000	8.418	12.677	147463.500	163832.550	9.991
Aromatic C21-C36	25330843.000	180.000	12.678	18.081	140726.906	152713.153	7.849
Aromatic EPH	53261287.000	360.000	4.094	18.081	147948.019	161887.469	8.611