

SOLID EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4518 **SAS No :** P4518 **SDG No:** P4518
Sample No : P4518-05MSD **Datafile:** FC067547.D
Client ID : EX-5-TPH-5MSD

COMPOUND	SPIKE ADDED mg/kg	SAMPLE CONCENTRATION mg/kg	MS/MSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS QC Limit Of RPD
Aliphatic C9-C12	13.5	23.2	42.2	140	*	3.51 (40-140) 25
Aliphatic C12-C16	9.0	184	262	856	*	7.32 (40-140) 25
Aliphatic C16-C21	13.5	200	302	748	*	1.85 (40-140) 25
Aliphatic C21-C28	18.0	36.9	100	350	*	7.43 (40-140) 25
Aliphatic C28-C40	27.0	38.9	122	307	*	4.46 (40-140) 25