

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4518 **SAS No :** P4518 **SDG No:** P4518
Sample No : PB164396BS **Datafile:** FC067543.D
Client ID : PB164396BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	10	5.70	57		(40-140)
Aliphatic C12-C16	6.7	5.41	81		(40-140)
Aliphatic C16-C21	10	9.37	94		(40-140)
Aliphatic C21-C28	13.3	12.7	95		(40-140)
Aliphatic C28-C40	20.0	22.7	113		(40-140)