

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P4518 **SAS No :** P4518 **SDG No:** P4518
Sample No : PB164396BSD **Datafile:** FC067544.D
Client ID : PB164396BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	10	5.66	57		0.813 (40-140)	50
Aliphatic C12-C16	6.7	5.38	81		0.551 (40-140)	50
Aliphatic C16-C21	10	9.32	93		0.575 (40-140)	50
Aliphatic C21-C28	13.3	12.6	95		0.776 (40-140)	50
Aliphatic C28-C40	20.0	22.6	113		0.06 (40-140)	50