

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133243

Review By	Sarabjit Jaswal	Review On	11/7/2024 4:10:01 AM
Supervise By	Mohan Bera	Supervise On	11/7/2024 4:26:12 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442		
ICV Standard	MP82450		
CCV Standard	MP82453		
ICSA Standard	MP82451,MP82452		
CRI Standard			
LCS Standard	MP82711		
Chk Standard	MP82455,MP82456		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/30/24 20:47		Kareem	OK
2	S1	S01	CAL2	10/30/24 20:52		Kareem	OK
3	S2	S02	CAL3	10/30/24 20:57		Kareem	OK
4	S3	S03	CAL4	10/30/24 21:01		Kareem	OK
5	S4	S04	CAL5	10/30/24 21:05		Kareem	OK
6	S5	S05	CAL6	10/30/24 21:09		Kareem	OK
7	S6	S06	CAL7	10/30/24 21:14		Kareem	OK
8	ICV001	ICV001	ICV	10/30/24 22:24	Inst. paused for calibration review	Kareem	OK
9	ICB001	ICB001	ICB	10/30/24 22:28		Kareem	OK
10	ICSA001	ICSA001	ICSA	10/30/24 22:45	Inst. paused for review	Kareem	OK
11	ICSAB001	ICSAB001	ICSAB	10/30/24 22:49		Kareem	OK
12	CCV001	CCV001	CCV	10/30/24 22:54		Kareem	OK
13	CCB001	CCB001	CCB	10/30/24 22:58		Kareem	OK
14	P4481-01DL	MYE4B0	SAM	10/30/24 23:02		Kareem	OK
15	P4481-02DL	MYE4B1	SAM	10/30/24 23:07		Kareem	OK
16	P4481-03DL	MYE4B2	SAM	10/30/24 23:11		Kareem	OK
17	P4481-04DL	MYE4B3	SAM	10/30/24 23:16		Kareem	OK
18	P4481-05DL	MYE4B4	SAM	10/30/24 23:20		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133243

Review By	Sarabjit Jaswal	Review On	11/7/2024 4:10:01 AM
Supervise By	Mohan Bera	Supervise On	11/7/2024 4:26:12 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442
ICV Standard	MP82450
CCV Standard	MP82453
ICSA Standard	MP82451,MP82452
CRI Standard	
LCS Standard	MP82711
Chk Standard	MP82455,MP82456

19	P4481-06DL	MYE4B5	SAM	10/30/24 23:25		Kareem	OK
20	P4481-07DL	MYE4B6	SAM	10/30/24 23:29		Kareem	OK
21	P4481-08DL	MYE4B7	SAM	10/30/24 23:33		Kareem	OK
22	P4481-09DL	MYE4B8	SAM	10/30/24 23:38		Kareem	OK
23	P4481-10DL	MYE4B9	SAM	10/30/24 23:42		Kareem	OK
24	P4481-11DL	MYE4C0	SAM	10/30/24 23:47		Kareem	OK
25	P4481-12DL	MYE4C1	SAM	10/30/24 23:51		Kareem	OK
26	P4481-13DL	MYE4C1D	DUP	10/30/24 23:56		Kareem	OK
27	P4481-14DL	MYE4C1S	MS	10/31/24	MS fail for As (ppb), Cu (ppb), Zn (ppb)	Kareem	OK
28	P4481-12LDL	MYE4C1L	SD	10/31/24 00:04		Kareem	OK
29	P4481-15DL	MYE4C2	SAM	10/31/24 00:09		Kareem	OK
30	P4481-16DL	MYE4C3	SAM	10/31/24 00:13		Kareem	OK
31	P4481-17DL	MYE4C4	SAM	10/31/24 00:18		Kareem	OK
32	P4481-18DL	MYE4C5	SAM	10/31/24 00:22		Kareem	OK
33	P4481-19DL	MYE4C6	SAM	10/31/24 00:27		Kareem	OK
34	CCV002	CCV002	CCV	10/31/24 00:31		Kareem	OK
35	CCB002	CCB002	CCB	10/31/24 00:35		Kareem	OK
36	P4481-20DL	MYE4C7	SAM	10/31/24 00:40		Kareem	OK
37	P4481-21DL	MYE4C8	SAM	10/31/24 00:44		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133243

Review By	Sarabjit Jaswal	Review On	11/7/2024 4:10:01 AM
Supervise By	Mohan Bera	Supervise On	11/7/2024 4:26:12 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442		
ICV Standard	MP82450		
CCV Standard	MP82453		
ICSA Standard	MP82451,MP82452		
CRI Standard			
LCS Standard	MP82711		
Chk Standard	MP82455,MP82456		

38	P4481-22DL	MYE4C9	SAM	10/31/24 00:49		Kareem	OK
39	PB164431BL	PBS431	MB	10/31/24 00:53		Kareem	OK
40	PB164431BS	LCS431	LCS	10/31/24 00:58		Kareem	OK
41	PB164431BSD	LCSD431	LCSD	10/31/24 01:02		Kareem	OK
42	P4519-01DL	MYE4K0	SAM	10/31/24 01:07		Kareem	OK
43	P4519-02DL	MYE4K1	SAM	10/31/24 01:11		Kareem	OK
44	P4519-03DL	MYE4K2	SAM	10/31/24 01:15		Kareem	OK
45	P4519-04DL	MYE4K2D	DUP	10/31/24 01:20		Kareem	OK
46	P4519-05DL	MYE4K2S	MS	10/31/24 01:24	MS fail for Pb (ppb)	Kareem	OK
47	P4519-03LDL	MYE4K2L	SD	10/31/24 01:29		Kareem	OK
48	P4519-06DL	MYE4K3	SAM	10/31/24 01:33		Kareem	OK
49	P4519-07DL	MYE4K4	SAM	10/31/24 01:38		Kareem	OK
50	P4519-08DL	MYE4K5	SAM	10/31/24 01:42		Kareem	OK
51	P4519-09DL	MYE4K6	SAM	10/31/24 01:46		Kareem	OK
52	P4519-10DL	MYE4K7	SAM	10/31/24 01:51		Kareem	OK
53	P4519-11DL	MYE4K8	SAM	10/31/24 01:55		Kareem	OK
54	P4519-12DL	MYE4K9	SAM	10/31/24 02:00		Kareem	OK
55	P4519-13DL	MYE4L0	SAM	10/31/24 02:04		Kareem	OK
56	CCV003	CCV003	CCV	10/31/24 02:09		Kareem	OK
57	CCB003	CCB003	CCB	10/31/24 02:13		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133243

Review By	Sarabjit Jaswal	Review On	11/7/2024 4:10:01 AM
Supervise By	Mohan Bera	Supervise On	11/7/2024 4:26:12 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442		
ICV Standard	MP82450		
CCV Standard	MP82453		
ICSA Standard	MP82451,MP82452		
CRI Standard			
LCS Standard	MP82711		
Chk Standard	MP82455,MP82456		

58	P4519-14DL	MYE4L1	SAM	10/31/24 02:17		Kareem	OK
59	P4519-15DL	MYE4L2	SAM	10/31/24 02:22		Kareem	OK
60	P4519-16DL	MYE4P6	SAM	10/31/24 02:26		Kareem	OK
61	P4519-17DL	MYE4P7	SAM	10/31/24 02:31		Kareem	OK
62	P4519-18DL	MYE4P8	SAM	10/31/24 02:35		Kareem	OK
63	P4519-19DL	MYE4P9	SAM	10/31/24 02:39		Kareem	OK
64	P4519-20DL	MYE4Q0	SAM	10/31/24 02:44		Kareem	OK
65	P4519-21DL	MYE4Q1	SAM	10/31/24 02:48		Kareem	OK
66	PB164457BL	PBS457	MB	10/31/24 02:53		Kareem	OK
67	PB164457BS	LCS457	LCS	10/31/24 02:57		Kareem	OK
68	PB164457BSD	LCSD457	LCSD	10/31/24 03:02		Kareem	OK
69	P4484-01DL	MYE4H0	SAM	10/31/24 03:06		Kareem	OK
70	P4484-02DL	MYE4H1	SAM	10/31/24 03:11		Kareem	OK
71	P4484-03DL	MYE4H2	SAM	10/31/24 03:15		Kareem	OK
72	P4484-04DL	MYE4H2D	DUP	10/31/24 03:19		Kareem	OK
73	P4484-05DL	MYE4H2S	MS	10/31/24 03:24	MS fail for Cr (ppb)	Kareem	OK
74	P4484-03LDL	MYE4H2L	SD	10/31/24 03:28		Kareem	OK
75	P4484-06DL	MYE4H3	SAM	10/31/24 03:33		Kareem	OK
76	P4484-07DL	MYE4H4	SAM	10/31/24 03:37		Kareem	OK
77	P4484-08DL	MYE4H5	SAM	10/31/24 03:42		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133243

Review By	Sarabjit Jaswal	Review On	11/7/2024 4:10:01 AM
Supervise By	Mohan Bera	Supervise On	11/7/2024 4:26:12 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442		
ICV Standard	MP82450		
CCV Standard	MP82453		
ICSA Standard	MP82451,MP82452		
CRI Standard			
LCS Standard	MP82711		
Chk Standard	MP82455,MP82456		

78	CCV004	CCV004	CCV	10/31/24 03:46		Kareem	OK
79	CCB004	CCB004	CCB	10/31/24 03:50		Kareem	OK
80	P4484-09DL	MYE4H6	SAM	10/31/24 03:55		Kareem	OK
81	P4484-10DL	MYE4H8	SAM	10/31/24 03:59		Kareem	OK
82	P4484-11DL	MYE4H9	SAM	10/31/24 04:04		Kareem	OK
83	P4484-12DL	MYE4J0	SAM	10/31/24 04:08		Kareem	OK
84	P4484-13DL	MYE4J1	SAM	10/31/24 04:12		Kareem	OK
85	P4484-14DL	MYE4J2	SAM	10/31/24 04:17		Kareem	OK
86	P4484-15DL	MYE4J3	SAM	10/31/24 04:21		Kareem	OK
87	P4484-16DL	MYE4J4	SAM	10/31/24 04:26		Kareem	OK
88	P4484-17DL	MYE4J5	SAM	10/31/24 04:30		Kareem	OK
89	P4484-18DL	MYE4J6	SAM	10/31/24 04:34		Kareem	OK
90	P4484-19DL	MYE4J7	SAM	10/31/24 04:39		Kareem	OK
91	P4484-20DL	MYE4J8	SAM	10/31/24 04:43		Kareem	OK
92	P4484-21DL	MYE4J9	SAM	10/31/24 04:48		Kareem	OK
93	P4484-22DL	MYE4L3	SAM	10/31/24 04:52		Kareem	OK
94	PB164455BL	PBS455	MB	10/31/24 04:57		Kareem	OK
95	PB164455BS	LCS455	LCS	10/31/24 05:01		Kareem	OK
96	PB164455BSD	LCSD455	LCSD	10/31/24 05:06		Kareem	OK
97	P4525-01DL	MYE4H7	SAM	10/31/24 05:10		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133243

Review By	Sarabjit Jaswal	Review On	11/7/2024 4:10:01 AM
Supervise By	Mohan Bera	Supervise On	11/7/2024 4:26:12 AM

STD. NAME	STD REF.#
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442
ICV Standard	MP82450
CCV Standard	MP82453
ICSA Standard	MP82451,MP82452
CRI Standard	
LCS Standard	MP82711
Chk Standard	MP82455,MP82456

98	P4525-02DL	MYE4H7D	DUP	10/31/24 05:15		Kareem	OK
99	P4525-03DL	MYE4H7S	MS	10/31/24 05:19		Kareem	OK
100	CCV005	CCV005	CCV	10/31/24 05:23		Kareem	OK
101	CCB005	CCB005	CCB	10/31/24 05:28		Kareem	OK
102	P4525-01LDL	MYE4H7L	SD	10/31/24 05:32		Kareem	OK
103	P4526-01DL	MYE4M3	SAM	10/31/24 05:37		Kareem	OK
104	P4526-02DL	MYE4M3D	DUP	10/31/24 05:41		Kareem	OK
105	P4526-03DL	MYE4M3S	MS	10/31/24 05:45	MS fail for As (ppb), Cu (ppb), Zn (ppb)	Kareem	OK
106	P4526-01LDL	MYE4M3L	SD	10/31/24 05:50		Kareem	OK
107	P4527-01DL	MYE4R9	SAM	10/31/24 05:54		Kareem	OK
108	P4527-02DL	MYE4R9D	DUP	10/31/24 05:59		Kareem	OK
109	P4527-03DL	MYE4R9S	MS	10/31/24 06:03		Kareem	OK
110	P4527-01LDL	MYE4R9L	SD	10/31/24 06:07		Kareem	OK
111	P4544-01DL	MYE5G1	SAM	10/31/24 06:12		Kareem	OK
112	P4544-02DL	MYE5G1D	DUP	10/31/24 06:16		Kareem	OK
113	P4544-03DL	MYE5G1S	MS	10/31/24 06:21	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
114	P4544-01LDL	MYE5G1L	SD	10/31/24 06:25		Kareem	OK
115	PB164451BL	PBS451	MB	10/31/24 06:30		Kareem	OK
116	PB164451BS	LCS451	LCS	10/31/24 06:34		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133243

Review By	Sarabjit Jaswal	Review On	11/7/2024 4:10:01 AM
Supervise By	Mohan Bera	Supervise On	11/7/2024 4:26:12 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442		
ICV Standard	MP82450		
CCV Standard	MP82453		
ICSA Standard	MP82451,MP82452		
CRI Standard			
LCS Standard	MP82711		
Chk Standard	MP82455,MP82456		

117	PB164451BSD	LCSD451	LCSD	10/31/24 06:39		Kareem	OK
118	P4496-01	MBHCX5	SAM	10/31/24 06:43		Kareem	OK
119	P4496-02	MBHCY6	SAM	10/31/24 06:48		Kareem	OK
120	P4496-03	MBHCY7	SAM	10/31/24 06:52		Kareem	OK
121	CCV006	CCV006	CCV	10/31/24 06:57		Kareem	OK
122	CCB006	CCB006	CCB	10/31/24 07:01		Kareem	OK
123	P4584-01	MC0PI1	SAM	10/31/24 07:05		Kareem	OK
124	P4584-02	MC0PI5	SAM	10/31/24 07:10		Kareem	OK
125	PB164537BL	PBW537	MB	10/31/24 07:15	NOT USE	Kareem	Not Ok
126	PB164537BS	LCS537	LCS	10/31/24 07:19	NOT USE	Kareem	Not Ok
127	P4501-01	MBHCS8	SAM	10/31/24 07:24	Ca,Mg,Ni high	Kareem	Dilution
128	P4501-02	MBHCS9	SAM	10/31/24 07:28	Mg,Ni high	Kareem	Dilution
129	P4501-03	MBHCT0	SAM	10/31/24 07:32	Mg,Ni high	Kareem	Dilution
130	P4501-04	MBHCT1	SAM	10/31/24 07:37	Ca,Ni high	Kareem	Dilution
131	P4501-05	MBHCT2	SAM	10/31/24 07:41	Ca high	Kareem	Dilution
132	P4501-06	MBHCT3	SAM	10/31/24 07:46		Kareem	OK
133	P4501-07	MBHCT4	SAM	10/31/24 07:50	Ni high	Kareem	Dilution
134	P4501-08	MBHCT5	SAM	10/31/24 07:54	Ca high	Kareem	Dilution
135	P4501-09	MBHCT6	SAM	10/31/24 07:59	Ca high	Kareem	Dilution
136	P4501-10	MBHCT7	SAM	10/31/24 08:03	Ca high	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133243

Review By	Sarabjit Jaswal	Review On	11/7/2024 4:10:01 AM
Supervise By	Mohan Bera	Supervise On	11/7/2024 4:26:12 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82441,MP82632,MP82446,MP82445,MP82444,MP82443,MP82442		
ICV Standard	MP82450		
CCV Standard	MP82453		
ICSA Standard	MP82451,MP82452		
CRI Standard			
LCS Standard	MP82711		
Chk Standard	MP82455,MP82456		

137	P4501-11	MBHCT8	SAM	10/31/24 08:08	Ca high	Kareem	Dilution
138	P4501-12	MBHCT9	SAM	10/31/24 08:12	Ca high	Kareem	Dilution
139	P4501-13	MBHCW0	SAM	10/31/24 08:17	Ca high	Kareem	Dilution
140	P4501-14	MBHCW0D	DUP	10/31/24 08:21	Ca high	Kareem	Dilution
141	P4501-15	MBHCW0S	MS	10/31/24 08:26	Ca high, MS fail for Pb (ppb), Ni (ppb), Zn (ppb), Sb,As,Se,Ag,Tl (Below RL)	Kareem	Dilution
142	P4501-13L	MBHCW0L	SD	10/31/24 08:30	Ca high	Kareem	Dilution
143	CCV007	CCV007	CCV	10/31/24 08:34		Kareem	OK
144	CCB007	CCB007	CCB	10/31/24 08:39		Kareem	OK
145	P4501-16	MBHCW1	SAM	10/31/24 08:43	Ca high	Kareem	Dilution
146	P4501-17	MBHCW2	SAM	10/31/24 08:47	Ni high	Kareem	Dilution
147	P4501-18	MBHCW3	SAM	10/31/24 08:52	Ni high	Kareem	Dilution
148	P4501-19	MBHCW4	SAM	10/31/24 08:56	Ni high	Kareem	Dilution
149	P4501-20	MBHCW5	SAM	10/31/24 09:00	Ni high	Kareem	Dilution
150	P4501-21	MBHCW6	SAM	10/31/24 09:05	Ca high	Kareem	Dilution
151	P4501-22	MBHCW7	SAM	10/31/24 09:09	Ca high	Kareem	Dilution
152	PB164544BL	PBS544	MB	10/31/24 09:14		Kareem	OK
153	PB164544BS	LCS544	LCS	10/31/24 09:18		Kareem	OK
154	CCV008	CCV008	CCV	10/31/24 09:23		Kareem	OK
155	CCB008	CCB008	CCB	10/31/24 09:27		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133346

Review By	Sarabjit Jaswal	Review On	11/8/2024 8:02:58 AM
Supervise By	Mohan Bera	Supervise On	11/8/2024 8:03:53 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/07/24 16:24		Kareem	OK
2	S1	S01	CAL2	11/07/24 16:29		Kareem	OK
3	S2	S02	CAL3	11/07/24 16:33		Kareem	OK
4	S3	S03	CAL4	11/07/24 16:38		Kareem	OK
5	S4	S04	CAL5	11/07/24 16:42		Kareem	OK
6	S5	S05	CAL6	11/07/24 16:46		Kareem	OK
7	S6	S06	CAL7	11/07/24 16:51		Kareem	OK
8	ICV005	ICV005	ICV	11/07/24 17:13		Kareem	OK
9	ICB005	ICB005	ICB	11/07/24 17:19		Kareem	OK
10	ICSA005	ICSA005	ICSA	11/07/24 17:24		Kareem	OK
11	ICSAB005	ICSAB005	ICSAB	11/07/24 17:28		Kareem	OK
12	CCV029	CCV029	CCV	11/07/24 17:32		Kareem	OK
13	CCB029	CCB029	CCB	11/07/24 17:37		Kareem	OK
14	P4480-05A	MYE494A	PS	11/07/24 18:17	Not Use	Kareem	Not Ok
15	P4481-12A	MYE4C1A	PS	11/07/24 18:21	Not Use	Kareem	Not Ok
16	P4482-14A	MYE4E3A	PS	11/07/24 18:26	Not Use	Kareem	Not Ok
17	P4483-20A	MYE4G9A	PS	11/07/24 18:30	Not Use	Kareem	Not Ok
18	P4484-03A	MYE4H2A	PS	11/07/24 18:35	Not Use	Kareem	Not Ok

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133346

Review By	Sarabjit Jaswal	Review On	11/8/2024 8:02:58 AM
Supervise By	Mohan Bera	Supervise On	11/8/2024 8:03:53 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

19	P4496-07A	MBHZC8A	PS	11/07/24 18:39	PS For Pb,Ni,Se,Tl	Kareem	OK
20	P4497-20A	MBHDD0A	PS	11/07/24 18:43	PS For As,Ni,Se,Tl,Zn	Kareem	OK
21	P4498-10A	MBHDF0A	PS	11/07/24 18:48	PS For As,Pb,Se,Tl	Kareem	OK
22	P4499-20A	MBHCN0A	PS	11/07/24 18:52	PS For Co,Cu,Tl	Kareem	OK
23	P4500-10A	MBHCP0A	PS	11/07/24 18:57	PS For As,Pb,Tl	Kareem	OK
24	P4501-13A	MBHCW0A	PS	11/07/24 19:01	PS For As,Pb,Ni,Se,Tl	Kareem	Dilution
25	P4519-03ADL	MYE4K2A	PS	11/07/24 19:05	PS for Pb-0021ml of MP83156	Kareem	OK
26	CCV030	CCV030	CCV	11/07/24 19:10		Kareem	OK
27	CCB030	CCB030	CCB	11/07/24 19:14		Kareem	OK
28	P4480-05ADL	MYE494A	PS	11/07/24 19:25	PS FOR Pb-0.0028ml of M6025,Se-0.07ml of MP83152	Kareem	OK
29	P4481-12ADL	MYE4C1A	PS	11/07/24 19:29	PS for As-0.0036ml of M5801,Cu-0.012ml of M5751,Zn-0.011ml of M5819	Kareem	OK
30	P4482-14ADL	MYE4E3A	PS	11/07/24 19:34	PS For Pb,Mn	Kareem	OK
31	P4483-20ADL	MYE4G9A	PS	11/07/24 19:38	Not Use	Kareem	Not Ok
32	P4496-07ADL	MBHZC8A	PS	11/07/24 19:43	Not Required	Kareem	Not Ok
33	P4501-13ADL	MBHCW0A	PS	11/07/24 19:47	2X For Ni	Kareem	Confirms
34	CCV031	CCV031	CCV	11/07/24 19:51		Kareem	OK
35	CCB031	CCB031	CCB	11/07/24 19:56		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133346

Review By	Sarabjit Jaswal	Review On	11/8/2024 8:02:58 AM
Supervise By	Mohan Bera	Supervise On	11/8/2024 8:03:53 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

36	P4497-11DL	MBHDB6	SAM	11/07/24 21:23	5X for Ca	Kareem	Confirms
37	P4497-13DL	MBHDB8	SAM	11/07/24 21:27	5X for Ni	Kareem	Confirms
38	P4499-19DL	MBHCM9	SAM	11/07/24 21:32	5X for Ca	Kareem	Confirms
39	CCV032	CCV032	CCV	11/07/24 21:36		Kareem	OK
40	CCB032	CCB032	CCB	11/07/24 21:40		Kareem	OK
41	P4499-20DL	MBHCN0	SAM	11/07/24 21:45	3X for Ca,Mg,Ni	Kareem	Confirms
42	P4499-21DL	MBHCN0D	DUP	11/07/24 21:49	3X for Ca,Mg,Ni	Kareem	Confirms
43	P4499-20LDL	MBHCN0L	SD	11/07/24 21:54	15X for Ca,Mg,Ni	Kareem	Confirms
44	P4499-22DL	MBHCN0S	MS	11/07/24 21:58	3X for Ca,Mg,Ni	Kareem	Confirms
45	P4496-04DL	MBHDO5	SAM	11/07/24 22:03	5X for Ca	Kareem	Confirms
46	P4496-05DL	MBHDO6	SAM	11/07/24 22:07	4X for Ca	Kareem	Confirms
47	P4496-06DL	MBHJC7	SAM	11/07/24 22:12	2X for Ca	Kareem	Confirms
48	P4496-07DL	MBHJC8	SAM	11/07/24 22:16	2X for Ca	Kareem	Confirms
49	P4496-08DL	MBHJC8D	DUP	11/07/24 22:20	2X for Ca	Kareem	Confirms
50	P4496-07LDL	MBHJC8L	SD	11/07/24 22:25	10X for Ca	Kareem	Confirms
51	P4496-09DL	MBHJC8S	MS	11/07/24 22:29	2X for Ca	Kareem	Confirms
52	P4498-02DL	MBHDD2	SAM	11/07/24 22:39	5X for Ca	Kareem	Confirms
53	P4498-03DL	MBHDD8	SAM	11/07/24 22:44	2X For Ca,Mg,Ni High, Still Mg,Ni High	Kareem	Dilution
54	P4498-03DL2	MBHDD8	SAM	11/07/24 22:48	5X for Mg,Ni	Kareem	Confirms

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133346

Review By	Sarabjit Jaswal	Review On	11/8/2024 8:02:58 AM
Supervise By	Mohan Bera	Supervise On	11/8/2024 8:03:53 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

55	P4498-04DL	MBHDD9	SAM	11/07/24 22:53	2X for Ca,Ni	Kareem	Confirms
56	P4498-05DL	MBHDE0	SAM	11/07/24 22:57	2X for Ca	Kareem	Confirms
57	P4498-06DL	MBHDE1	SAM	11/07/24 23:01	2X for Ni	Kareem	Confirms
58	P4498-07DL	MBHDE2	SAM	11/07/24 23:06	2X for Ni	Kareem	Confirms
59	P4498-08DL	MBHDE8	SAM	11/07/24 23:10	5X for Ca	Kareem	Confirms
60	P4498-09DL	MBHDE9	SAM	11/07/24 23:15	5X for Ca	Kareem	Confirms
61	CCV033	CCV033	CCV	11/07/24 23:19		Kareem	OK
62	CCB033	CCB033	CCB	11/07/24 23:23		Kareem	OK
63	P4498-10DL	MBHDF0	SAM	11/07/24 23:28	2X for Ni	Kareem	Confirms
64	P4498-11DL	MBHDF0D	DUP	11/07/24 23:32	2X for Ni	Kareem	Confirms
65	P4498-10LDL	MBHDF0L	SD	11/07/24 23:37	10X for Ni	Kareem	Confirms
66	P4498-12DL	MBHDF0S	MS	11/07/24 23:41	2X for Ni	Kareem	Confirms
67	P4498-13DL	MBHDF1	SAM	11/07/24 23:45	5X For Ca	Kareem	Confirms
68	P4498-14DL	MBHDF2	SAM	11/07/24 23:50	10X For Ca	Kareem	Confirms
69	P4498-15DL	MBHCH8	SAM	11/07/24 23:54	2X For Ca,Ni	Kareem	Confirms
70	P4498-16DL	MBHCH9	SAM	11/07/24 23:59	3X For Ni	Kareem	Confirms
71	P4498-20DL	MBHCJ3	SAM	11/08/24 00:03	3X For Ca,Ni	Kareem	Confirms
72	P4498-21DL	MBHCJ5	SAM	11/08/24 00:08	5X For Ca	Kareem	Confirms
73	P4498-22DL	MBHCJ6	SAM	11/08/24 00:12	2X for Ca	Kareem	Confirms
74	CCV034	CCV034	CCV	11/08/24 00:16		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133346

Review By	Sarabjit Jaswal	Review On	11/8/2024 8:02:58 AM
Supervise By	Mohan Bera	Supervise On	11/8/2024 8:03:53 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

75	CCB034	CCB034	CCB	11/08/24 00:21		Kareem	OK
----	--------	--------	-----	----------------	--	--------	----