


Batch Folder: D:\Agilent\ICPMH\1\DATA\P7110224 MS.b\
Analysis File: P7110224 MS.batch.bin
DA Date-Time: 2024-11-03 14:10:40
Calibration Title:
Calibration Method: External Calibration
VIS Interpolation Fit:

Level	Standard Data File	Sample Name	Acq. Date-Time
1	004CALB.d	S00	2024-11-02 17:41:38
2	006CALS.d	S02	2024-11-02 17:48:10
3	007CALS.d	S03	2024-11-02 17:51:25
4	008CALS.d	S04	2024-11-02 17:54:26
5	009CALS.d	S05	2024-11-02 17:57:16
6	010CALS.d	S06	2024-11-02 18:00:04
7	011CALS.d	S07	2024-11-02 18:02:47
8	012CALS.d	S08	2024-11-02 18:05:33

9 Be [No Gas] ISTD:6 Li [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	81.48	0.0001	P	35.0
2	☐	1.000	0.969	1572.33	0.0011	P	2.8
3	☐	50.000	53.691	82849.24	0.0588	P	2.0
4	☐	125.000	132.563	205241.95	0.1451	P	0.7
5	☐	250.000	261.871	408213.21	0.2865	P	0.8
6	☐	500.000	510.086	799774.54	0.5580	P	1.7
7	☐	1000.000	990.859	1542042.90	1.0839	A	2.3
8	☐			220.38	0.0002	P	2.4

$y = 0.0011 * x + 5.7197E-005$

$R = 0.9998$


$DL = 0.0549$

$BEC = 0.05229$

Weight: <None>


Min Conc: 0

12 C [No Gas] No available calibration points




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

12 C [He] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

23 Na [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	26365.94	0.0663	P	8.7
2	☐	500.000	477.810	348067.29	0.8507	P	0.2
3	☐	5000.000	4994.381	3417814.98	8.2656	A	0.8
4	☐	12500.000	12390.637	8317753.00	20.4080	A	0.7
5	☐	25000.000	24888.738	16503516.00	40.9260	A	0.1
6	☐	50000.000	48894.137	32481911.18	80.3355	A	1.8
7	☐	100000.00	99687.082	65240103.47	163.7220	A	1.0
8	☐	500000.00	500181.54	335412937.2	821.2115	A	0.5

$$y = 0.0016 * x + 0.0663$$


R = 1.0000

DL = 10.56

BEC = 40.4

Weight: <None>

Min Conc: 0

24 Mg [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	2027.93	0.0051	P	4.4
2	☐	500.000	503.064	202290.95	0.4944	P	0.5
3	☐	5000.000	4944.283	1990889.21	4.8145	A	0.8
4	☐	12500.000	12364.293	4903979.30	12.0322	A	0.5
5	☐	25000.000	24721.757	9699447.56	24.0527	A	0.8
6	☐	50000.000	48137.372	18935034.02	46.8297	A	1.7
7	☐	100000.00	98012.686	37993877.77	95.3449	A	0.6
8	☐	500000.00	500601.58	198890201.4	486.9548	A	0.8

$$y = 9.7273E-004 * x + 0.0051$$


$$R = 1.0000$$

$$DL = 0.6892$$

$$BEC = 5.228$$

Weight: <None>

Min Conc: 0

27 Al [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	488.90	0.0012	P	4.9
2	☐	20.000	17.802	4751.89	0.0116	P	2.8
3	☐	1000.000	994.072	240372.36	0.5813	P	0.5
4	☐	2500.000	2470.975	588194.05	1.4432	P	0.9
5	☐	5000.000	4947.151	1164649.04	2.8882	P	0.3
6	☐	10000.000	9578.825	2260724.17	5.5910	A	1.7
7	☐	20000.000	19137.132	4450562.05	11.1688	A	1.0
8	☐	100000.00	100218.11	23887246.87	58.4842	A	0.4

$$y = 5.8356E-004 * x + 0.0012$$

$$R = 1.0000$$


$$DL = 0.3123$$

$$BEC = 2.103$$

Weight: <None>

Min Conc: 0

28 Si [No Gas] ISTD :45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	217138.83	0.0453	P	1.6
2	☐	10.000	1.369	239875.98	0.0504	P	1.5
3	☐	50.000	41.963	966535.56	0.2025	P	1.5
4	☐	125.000	102.320	2007452.54	0.4286	A	1.0
5	☐	250.000	203.100	3675725.64	0.8061	A	1.2
6	☐	500.000	400.150	6864064.55	1.5442	A	1.0
7	☐	1000.000	1064.973	17951426.82	4.0344	A	0.7
8	☐			280164.02	0.0629	P	1.8

$$y = 0.0037 * x + 0.0453$$

R = 0.9923


DL = 0.5908

BEC = 12.1

Weight: <None>

Min Conc: 0

31 P [He] ISTD :45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	-15.426	208.89	0.0005	P	12.4
2	☐	0.000	15.426	586.68	0.0014	P	10.8
3	☐	1000.000	1038.220	13122.39	0.0317	P	0.5
4	☐	2500.000	2564.640	31363.23	0.0770	P	1.8
5	☐	5000.000	5153.548	61958.06	0.1536	P	1.0
6	☐	10000.000	9995.115	120108.67	0.2971	P	2.2
7	☐	20000.000	19954.064	235943.43	0.5921	P	0.8
8	☐			234.45	0.0006	P	7.2

$$y = 2.9624E-005 * x + 9.7749E-004$$

R = 1.0000


DL = 11.08

BEC = 33

Weight: <None>

Min Conc: 0

34 S [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	9.689	201620.04	0.0421	P	1.6
2	☐	0.000	-9.689	199290.22	0.0419	P	2.2
3	☐	1000.000	872.291	236237.45	0.0495	P	0.7
4	☐	2500.000	2227.534	286342.50	0.0611	P	1.7
5	☐	5000.000	4846.578	381391.21	0.0836	P	1.5
6	☐	10000.000	10057.398	570817.01	0.1284	P	0.8
7	☐	20000.000	20050.100	953439.74	0.2143	P	0.6
8	☐			162896.22	0.0366	P	1.8

$$y = 8.5922E-006 * x + 0.0420$$

$$R = 0.9999$$


$$DL = 281.3$$

$$BEC = 4888$$


Weight: <None>

Min Conc: 0


35 Cl [No Gas] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

35 Cl [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

39 K [He] ISTD:45 Sc [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	51892.49	0.1304	P	7.0
2	☐	500.000	480.603	284704.17	0.6959	P	0.3
3	☐	2500.000	2509.461	1274763.26	3.0829	A	0.3
4	☐	6250.000	6190.181	3021510.41	7.4134	A	1.5
5	☐	12500.000	12390.240	5930977.69	14.7080	A	0.4
6	☐	25000.000	24059.230	11497131.35	28.4369	A	2.2
7	☐	50000.000	48953.583	23003678.82	57.7260	A	0.4
8	☐	250000.00	250310.28	120337483.7	294.6289	A	0.9

$$y = 0.0012 * x + 0.1304$$

$$R = 1.0000$$


$$DL = 23.33$$

$$BEC = 110.8$$

Weight: <None>

Min Conc: 0

43 Ca [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	427.79	0.0001	P	7.6
2	☐	500.000	485.227	14602.68	0.0031	P	1.5
3	☐	5000.000	5970.557	175515.94	0.0368	P	1.6
4	☐	12500.000	14826.679	427097.38	0.0912	P	1.4
5	☐	25000.000	29863.214	837076.73	0.1836	P	1.5
6	☐	50000.000	57230.163	1563442.91	0.3517	A	0.9
7	☐	100000.00	112801.98	3084349.12	0.6932	A	0.3
8	☐	500000.00	496405.56	13591056.19	3.0500	A	1.4

$$y = 6.1441E-006 * x + 8.9293E-005$$

$$R = 0.9996$$


$$DL = 3.305$$

$$BEC = 14.53$$

Weight: <None>

Min Conc: 0

44 Ca [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	14508.76	0.0030	P	1.3
2	☐	500.000	485.329	243316.51	0.0512	P	1.8
3	☐	5000.000	5344.254	2544812.62	0.5332	A	1.2
4	☐	12500.000	13154.335	6126089.49	1.3079	A	1.0
5	☐	25000.000	26616.766	12054218.99	2.6433	A	0.6
6	☐	50000.000	52912.494	23344669.37	5.2517	A	1.0
7	☐	100000.00	104913.18	46320434.31	10.4099	A	0.8
8	☐	500000.00	498625.48	220415023.3	49.4642	A	1.5

$$y = 9.9195E-005 * x + 0.0030$$

$$R = 0.9999$$


$$DL = 1.182$$

$$BEC = 30.53$$

Weight: <None>

Min Conc: 0

47 Ti [No Gas] ISTD :45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	213.34	0.0000	P	18.6
2	☐	5.000	4.279	6562.58	0.0014	P	2.3
3	☐	500.000	511.600	762253.45	0.1597	P	1.7
4	☐	1250.000	1254.476	1833938.98	0.3915	A	1.1
5	☐	2500.000	2532.643	3604301.34	0.7904	A	1.5
6	☐	5000.000	5017.276	6960327.47	1.5658	A	1.2
7	☐	10000.000	9982.062	13861493.96	3.1152	A	0.8
8	☐			1922.36	0.0004	P	7.5

$$y = 3.1208E-004 * x + 4.4580E-005$$

$$R = 1.0000$$


$$DL = 0.07953$$

$$BEC = 0.1429$$

Weight: <None>

Min Conc: 0

51 V [He] ISTD :45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	31.11	0.0001	P	41.5
2	☐	5.000	4.887	18281.17	0.0447	P	1.4
3	☐	50.000	51.882	195837.98	0.4736	P	1.3
4	☐	125.000	128.357	477536.59	1.1717	P	0.6
5	☐	250.000	257.331	947198.33	2.3489	P	0.4
6	☐	500.000	499.510	1843265.54	4.5594	A	2.7
7	☐	1000.000	997.899	3629549.43	9.1085	A	1.2
8	☐			827.81	0.0020	P	5.4

$$y = 0.0091 * x + 7.7698E-005$$

$$R = 1.0000$$


$$DL = 0.01059$$

$$BEC = 0.008512$$

Weight: <None>

Min Conc: 0

52 Cr [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	580.02	0.0015	P	3.5
2	☐	2.000	1.922	8714.80	0.0213	P	0.7
3	☐	50.000	52.202	223472.91	0.5404	P	1.1
4	☐	125.000	129.511	545604.09	1.3387	P	0.5
5	☐	250.000	256.913	1070275.45	2.6541	P	0.4
6	☐	500.000	499.668	2086512.69	5.1605	A	2.0
7	☐	1000.000	997.764	4105796.36	10.3034	A	0.5
8	☐			13394.23	0.0328	P	0.1

$$y = 0.0103 * x + 0.0015$$

$$R = 1.0000$$


$$DL = 0.01464$$

$$BEC = 0.1407$$

Weight: <None>

Min Conc: 0

54 Fe [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1973.48	0.0050	P	11.4
2	☐	50.000	53.303	12475.24	0.0305	P	0.6
3	☐	2500.000	2795.015	555573.46	1.3436	P	0.5
4	☐	6250.000	6887.947	1346526.53	3.3038	M	2.6
5	☐	12500.000	13508.267	2610809.37	6.4746	A	0.9
6	☐	25000.000	26277.616	5090447.22	12.5903	A	2.0
7	☐	50000.000	52611.807	10043115.33	25.2026	A	0.8
8	☐	250000.00	249280.56	48765413.72	119.3944	A	0.1

$$y = 4.7894E-004 * x + 0.0050$$


$$R = 0.9999$$

$$DL = 3.556$$

$$BEC = 10.36$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	434.45	0.0011	P	10.8
2	☐	1.000	0.975	3472.66	0.0085	P	3.4
3	☐	500.000	515.910	1618864.94	3.9155	A	3.0
4	☐	1250.000	1269.299	3925596.61	9.6317	A	0.5
5	☐	2500.000	2540.541	7773493.36	19.2771	A	1.0
6	☐	5000.000	5002.724	15346379.08	37.9586	A	2.4
7	☐	10000.000	9985.295	30190555.38	75.7631	A	0.8
8	☐			14671.01	0.0359	P	4.7

$$y = 0.0076 * x + 0.0011$$


$$R = 1.0000$$

$$DL = 0.04664$$

$$BEC = 0.1435$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	18230.01	0.0458	P	5.1
2	☐	50.000	51.154	203048.51	0.4963	P	1.0
3	☐	2500.000	2658.663	9701116.31	23.4618	A	0.9
4	☐	6250.000	6577.142	23628441.04	57.9737	A	0.3
5	☐	12500.000	13098.666	46539750.42	115.4118	A	0.2
6	☐	25000.000	25812.266	91934679.74	227.3863	A	2.1
7	☐	50000.000	51465.053	180643697.3	453.3223	A	0.8
8	☐	250000.00	249586.06	897853995.5	2,198.265	A	0.7

$$y = 0.0088 * x + 0.0458$$

$$R = 1.0000$$


$$DL = 0.8018$$

$$BEC = 5.196$$

Weight: <None>

Min Conc: 0

Calibration for 016ICSB.d



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	1466.75	0.0037	P	3.9
2	☐	50.000	51.424	6137.98	0.0150	P	1.0
3	☐	2500.000	2749.160	251823.36	0.6090	P	1.2
4	☐	6250.000	6809.364	612606.00	1.5031	P	0.6
5	☐	12500.000	13584.489	1207691.73	2.9949	P	0.5
6	☐	25000.000	25731.790	2292402.29	5.6697	A	1.9
7	☐	50000.000	51342.265	4506647.16	11.3090	A	0.5
8	☐	250000.00	249587.66	22448500.50	54.9618	A	0.7

$$y = 2.2020\text{E-}004 * x + 0.0037$$


R = 1.0000

DL = 1.967

BEC = 16.71

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	120.00	0.0003	P	3.7
2	☐	1.000	0.991	6560.38	0.0160	P	1.0
3	☐	50.000	51.180	336022.48	0.8126	P	1.0
4	☐	125.000	127.698	826153.95	2.0270	P	0.7
5	☐	250.000	260.298	1665983.15	4.1315	A	1.1
6	☐	500.000	501.001	3214706.96	7.9517	A	2.8
7	☐	1000.000	996.529	6302653.38	15.8163	A	0.6
8	☐			16941.89	0.0415	P	1.1

$$y = 0.0159 * x + 3.0094\text{E-}004$$

R = 0.9999


DL = 0.002124

BEC = 0.01896

Weight: <None>

Min Conc: 0

60 Ni [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	78.89	0.0002	P	9.6
2	☐	1.000	0.960	1677.89	0.0041	P	7.2
3	☐	50.000	51.064	85900.80	0.2077	P	0.3
4	☐	125.000	126.917	210320.34	0.5160	P	0.8
5	☐	250.000	251.600	412440.95	1.0228	P	0.4
6	☐	500.000	482.721	793354.34	1.9622	P	1.8
7	☐	1000.000	1007.947	1632571.61	4.0969	A	0.5
8	☐			5764.49	0.0141	P	2.0

$$y = 0.0041 * x + 1.9809E-004$$

$$R = 0.9998$$


$$DL = 0.01405$$

$$BEC = 0.04874$$

Weight: <None>

Min Conc: 0

63 Cu [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	1297.85	0.0033	P	6.2
2	☐	2.000	1.833	9976.73	0.0244	P	0.3
3	☐	500.000	532.975	2540860.57	6.1448	A	0.6
4	☐	1250.000	1310.739	6157153.73	15.1072	A	1.3
5	☐	2500.000	2593.260	12051686.91	29.8859	A	0.5
6	☐	5000.000	5003.125	23311418.26	57.6553	A	1.9
7	☐	10000.000	9965.881	45762969.32	114.8421	A	0.8
8	☐			9910.02	0.0243	P	1.2

$$y = 0.0115 * x + 0.0033$$

$$R = 0.9999$$


$$DL = 0.05272$$

$$BEC = 0.2827$$

Weight: <None>

Min Conc: 0

66 Zn [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	684.47	0.0017	P	12.5
2	☐	2.000	4.707	4196.19	0.0103	P	2.9
3	☐	500.000	509.795	383351.39	0.9271	P	0.6
4	☐	1250.000	1265.610	937010.40	2.2990	P	0.5
5	☐	2500.000	2592.413	1898236.34	4.7074	A	0.3
6	☐	5000.000	5015.359	3681506.79	9.1054	A	1.9
7	☐	10000.000	9966.775	7210035.66	18.0930	A	0.2
8	☐			3221.49	0.0079	P	3.2

$$y = 0.0018 * x + 0.0017$$

$$R = 1.0000$$


$$DL = 0.3535$$

$$BEC = 0.9443$$

Weight: <None>

Min Conc: 0

75 As [He] ISTD:89 Y [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	10.00	0.0000	P	63.8
2	☐	1.000	0.952	494.46	0.0002	P	14.3
3	☐	50.000	52.091	26832.06	0.0082	P	2.7
4	☐	125.000	129.098	65692.15	0.0203	P	0.6
5	☐	250.000	255.161	130692.34	0.0401	P	1.0
6	☐	500.000	496.069	253613.08	0.0780	P	1.6
7	☐	1000.000	1000.058	501251.36	0.1572	P	1.6
8	☐			296.67	0.0001	P	13.0

$$y = 1.5722E-004 * x + 3.1313E-006$$

$$R = 1.0000$$


$$DL = 0.03811$$

$$BEC = 0.01992$$

Weight: <None>

Min Conc: 0

77 Se [He] ISTD:89 Y [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	3.33	0.0000	P	100.
2	☐	5.000	4.824	73.33	0.0000	P	16.5
3	☐	50.000	51.549	761.14	0.0002	P	7.3
4	☐	125.000	133.347	1940.15	0.0006	P	4.8
5	☐	250.000	256.113	3749.40	0.0012	P	1.6
6	☐	500.000	498.697	7285.18	0.0022	P	1.0
7	☐	1000.000	998.003	14285.83	0.0045	P	2.8
8	☐			10.00	0.0000	P	0.7

$$y = 4.4897E-006 * x + 1.0040E-006$$

$$R = 1.0000$$


$$DL = 0.6744$$

$$BEC = 0.2236$$

Weight: <None>

Min Conc: 0

78 Se [He] ISTD:89 Y [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	255.56	0.0001	P	16.1
2	☐	5.000	5.614	527.80	0.0002	P	8.4
3	☐	50.000	51.773	2754.73	0.0008	P	6.0
4	☐	125.000	133.740	6621.54	0.0020	P	1.4
5	☐	250.000	263.428	12874.58	0.0040	P	1.7
6	☐	500.000	493.166	23830.34	0.0073	P	1.4
7	☐	1000.000	998.876	47059.96	0.0148	P	1.0
8	☐			242.23	0.0001	P	10.3

$$y = 1.4697E-005 * x + 8.0602E-005$$


$$R = 0.9998$$

$$DL = 2.649$$


$$BEC = 5.484$$

Weight: <None>


Min Conc: 0

81 Br [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			11954.92		P	0.6
2	☐			11395.60		P	0.1
3	☐			11183.20		P	1.3
4	☐			10706.18		P	3.4
5	☐			10572.76		P	3.9
6	☐			10730.63		P	2.1
7	☐			10668.38		P	3.2
8	☐			11545.71		P	0.7

81 Br [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			60.00		P	29.4
2	☐			65.56		P	22.9
3	☐			48.89		P	30.7
4	☐			38.89		P	9.9
5	☐			42.22		P	16.4
6	☐			64.44		P	33.3
7	☐			65.55		P	29.8
8	☐			65.56		P	12.8



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	476.48	0.0000	P	6.9
2	☐	5.000	4.720	1221.18	0.0001	P	0.6
3	☐	50.000	52.555	8978.70	0.0006	P	0.9
4	☐	125.000	129.624	20908.01	0.0015	P	0.7
5	☐	250.000	253.860	40788.77	0.0030	P	1.2
6	☐	500.000	504.042	79947.37	0.0059	P	2.0
7	☐	1000.000	996.309	154738.43	0.0116	P	0.9
8	☐			609.08	0.0000	P	3.5

$$y = 1.1573E-005 * x + 3.4171E-005$$


R = 1.0000

DL = 0.6132

BEC = 2.953


Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			195.56		P	13.2
2	☐			178.89		P	16.3
3	☐			201.11		P	9.4
4	☐			200.00		P	3.3
5	☐			198.89		P	8.3
6	☐			216.67		P	13.1
7	☐			186.67		P	12.5
8	☐			233.33		P	9.9

86 Sr [No Gas] ISTD :89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	846.70	0.0001	P	1.9
2	☐	1.000	0.911	3966.14	0.0003	P	2.6
3	☐	1250.000	1268.176	4430165.59	0.3169	A	0.2
4	☐	3125.000	3177.397	10820305.39	0.7940	A	0.7
5	☐	6250.000	6263.986	21482327.60	1.5653	A	1.2
6	☐	12500.000	12402.859	42231512.71	3.0992	A	1.4
7	☐	25000.000	25037.616	83716045.98	6.2563	A	0.8
8	☐			13923.34	0.0011	P	1.0

$$y = 2.4987E-004 * x + 6.0708E-005$$

$$R = 1.0000$$


$$DL = 0.01383$$

$$BEC = 0.243$$

Weight: <None>

Min Conc: 0

88 Sr [No Gas] ISTD :89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	5298.79	0.0004	P	7.2
2	☐	1.000	0.902	32456.92	0.0024	P	1.8
3	☐	1250.000	1254.684	38505516.65	2.7548	A	0.3
4	☐	3125.000	3163.051	94632879.70	6.9443	A	0.8
5	☐	6250.000	6243.637	188128201.6	13.7073	A	1.0
6	☐	12500.000	12391.236	370675461.1	27.2033	A	1.8
7	☐	25000.000	25050.982	735895517.9	54.9958	A	1.4
8	☐			121880.94	0.0094	P	1.4

$$y = 0.0022 * x + 3.7998E-004$$

$$R = 1.0000$$


$$DL = 0.03715$$

$$BEC = 0.1731$$

Weight: <None>

Min Conc: 0

90 Zr [No Gas] ISTD : 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	733.36	0.0001	P	4.5
2	☐	1.000	0.908	17341.48	0.0013	P	0.5
3	☐	50.000	49.046	913086.58	0.0653	P	0.7
4	☐	125.000	126.830	2300933.25	0.1688	A	1.6
5	☐	250.000	250.703	4579795.73	0.3337	A	1.6
6	☐	500.000	493.694	8953737.78	0.6571	A	1.6
7	☐	1000.000	1002.796	17858759.04	1.3346	A	1.4
8	☐			10119.17	0.0008	P	0.7

$$y = 0.0013 * x + 5.2580E-005$$

$$R = 1.0000$$


$$DL = 0.005293$$

$$BEC = 0.03951$$

Weight: <None>

Min Conc: 0

91 Zr [No Gas] ISTD : 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	148.89	0.0000	P	20.6
2	☐	1.000	0.900	3822.76	0.0003	P	4.4
3	☐	50.000	49.254	204655.19	0.0146	P	1.0
4	☐	125.000	124.631	504661.21	0.0370	P	1.3
5	☐	250.000	246.531	1005262.63	0.0732	P	0.9
6	☐	500.000	493.624	1998186.29	0.1466	A	2.2
7	☐	1000.000	1004.139	3991586.33	0.2983	A	1.0
8	☐			2337.98	0.0002	P	4.8

$$y = 2.9706E-004 * x + 1.0680E-005$$

$$R = 1.0000$$

$$DL = 0.02225$$


$$BEC = 0.03595$$

Weight: <None>

Min Conc: 0

Calibration for 016ICSB.d

94 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	538.90	0.0000	P	6.7
2	☐	5.000	4.942	20475.48	0.0015	P	1.2
3	☐	500.000	499.801	2050526.53	0.1467	A	1.1
4	☐	1250.000	1260.855	5042652.22	0.3700	A	0.5
5	☐	2500.000	2481.182	9993645.20	0.7281	A	0.8
6	☐	5000.000	4968.170	19866764.84	1.4580	A	1.5
7	☐	10000.000	10019.273	39342842.19	2.9402	A	1.2
8	☐			9237.43	0.0007	P	1.5

$$y = 2.9345E-004 * x + 3.8643E-005$$

R = 1.0000


DL = 0.0264

BEC = 0.1317

Weight: <None>

Min Conc: 0

95 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	481.13	0.0000	P	9.6
2	☐	5.000	4.066	24552.87	0.0018	P	1.1
3	☐	500.000	500.726	3014235.68	0.2156	A	0.2
4	☐	1250.000	1263.854	7416616.90	0.5443	A	0.9
5	☐	2500.000	2507.388	14818717.84	1.0797	A	1.0
6	☐	5000.000	4959.228	29099667.90	2.1355	A	1.5
7	☐	10000.000	10016.771	57715310.25	4.3133	A	1.0
8	☐			8644.84	0.0007	P	3.7

$$y = 4.3060E-004 * x + 3.4496E-005$$

R = 1.0000

DL = 0.02304


BEC = 0.08011

Weight: <None>

Min Conc: 0

Calibration for 016ICSB.d

96 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	617.80	0.0000	P	5.8
2	☐	5.000	4.089	27045.09	0.0020	P	1.3
3	☐	500.000	497.728	3271053.74	0.2340	A	0.8
4	☐	1250.000	1258.491	8062351.06	0.5916	A	1.3
5	☐	2500.000	2491.788	16077162.40	1.1714	A	0.6
6	☐	5000.000	4946.103	31682430.08	2.3251	A	1.8
7	☐	10000.000	10028.054	63078596.28	4.7140	A	1.0
8	☐			21409.10	0.0016	P	3.4

$$y = 4.7008\text{E-}004 * x + 4.4289\text{E-}005$$

$$R = 1.0000$$


$$DL = 0.01642$$

$$BEC = 0.09422$$

Weight: <None>

Min Conc: 0

97 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	304.45	0.0000	P	10.7
2	☐	5.000	4.088	15464.96	0.0011	P	0.9
3	☐	500.000	497.149	1874432.56	0.1341	A	1.2
4	☐	1250.000	1254.870	4612198.61	0.3385	A	1.3
5	☐	2500.000	2488.036	9209878.20	0.6710	A	0.6
6	☐	5000.000	4952.161	18198396.54	1.3356	A	2.4
7	☐	10000.000	10026.445	36184035.29	2.7041	A	0.7
8	☐			5478.84	0.0004	P	2.7

$$y = 2.6970\text{E-}004 * x + 2.1839\text{E-}005$$

$$R = 1.0000$$


$$DL = 0.02606$$

$$BEC = 0.08098$$

Weight: <None>

Min Conc: 0

98 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	822.26	0.0001	P	4.6
2	☐	5.000	4.068	39808.62	0.0029	P	2.2
3	☐	500.000	494.039	4815441.32	0.3445	A	0.4
4	☐	1250.000	1253.976	11914562.18	0.8743	A	1.5
5	☐	2500.000	2480.819	23739032.98	1.7297	A	1.3
6	☐	5000.000	4924.866	46787383.75	3.4337	A	2.1
7	☐	10000.000	10042.164	93687794.16	7.0015	A	0.7
8	☐			12813.49	0.0010	P	2.2

$$y = 6.9721E-004 * x + 5.8946E-005$$

$$R = 1.0000$$


$$DL = 0.01174$$

$$BEC = 0.08455$$

Weight: <None>

Min Conc: 0

106 Cd [No Gas] ISTD:159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	2997.00	0.0003	P	0.8
2	☐	1.000	1.001	3235.95	0.0003	P	2.5
3	☐	50.000	53.963	17082.33	0.0015	P	1.1
4	☐	125.000	132.782	37290.06	0.0034	P	0.7
5	☐	250.000	259.379	70820.37	0.0064	P	0.4
6	☐	500.000	508.666	137270.20	0.0122	P	0.8
7	☐	1000.000	992.152	265634.19	0.0236	P	0.8
8	☐			2906.99	0.0003	P	3.1

$$y = 2.3506E-005 * x + 2.7626E-004$$

$$R = 0.9999$$


$$DL = 0.2781$$

$$BEC = 11.75$$

Weight: <None>

Min Conc: 0

107 Ag [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	176.67	0.0000	P	9.6
2	☐	1.000	0.902	11483.52	0.0011	P	3.3
3	☐	50.000	51.880	666347.96	0.0603	P	2.3
4	☐	125.000	136.630	1741648.43	0.1587	A	1.5
5	☐	250.000	266.345	3437214.36	0.3093	A	0.7
6	☐	500.000	510.408	6650921.22	0.5927	A	1.0
7	☐	1000.000	989.162	12931168.70	1.1487	A	0.7
8	☐			1762.35	0.0002	P	14.6

$$y = 0.0012 * x + 1.6281E-005$$

$$R = 0.9997$$


$$DL = 0.004023$$

$$BEC = 0.01402$$

Weight: <None>

Min Conc: 0

108 Cd [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	4.45	0.0000	P	86.6
2	☐	1.000	1.116	204.45	0.0000	P	13.3
3	☐	50.000	53.260	9797.84	0.0009	P	3.0
4	☐	125.000	133.940	24447.38	0.0022	P	2.3
5	☐	250.000	262.019	48417.11	0.0044	P	1.1
6	☐	500.000	510.795	95310.31	0.0085	P	1.7
7	☐	1000.000	990.317	185356.27	0.0165	P	1.1
8	☐			77.78	0.0000	P	30.5

$$y = 1.6628E-005 * x + 4.1075E-007$$

$$R = 0.9998$$


$$DL = 0.06418$$

$$BEC = 0.0247$$

Weight: <None>

Min Conc: 0

109 Ag [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	116.67	0.0000	P	6.9
2	☐	1.000	0.917	10988.70	0.0010	P	3.6
3	☐	50.000	51.662	627665.99	0.0568	P	1.6
4	☐	125.000	136.238	1642767.79	0.1497	A	1.2
5	☐	250.000	265.646	3242870.40	0.2918	A	0.9
6	☐	500.000	510.706	6294429.63	0.5610	A	1.9
7	☐	1000.000	989.248	12232934.26	1.0867	A	1.1
8	☐			1644.55	0.0002	P	10.5

$$y = 0.0011 * x + 1.0751E-005$$

$$R = 0.9997$$


$$DL = 0.002032$$

$$BEC = 0.009787$$

Weight: <None>

Min Conc: 0

111 Cd [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	2131.46	0.0002	P	0.8
2	☐	1.000	0.992	4664.52	0.0004	P	4.8
3	☐	50.000	51.037	136270.37	0.0123	P	1.4
4	☐	125.000	126.675	332531.06	0.0303	P	0.4
5	☐	250.000	249.211	660191.91	0.0594	P	0.9
6	☐	500.000	482.545	1288703.05	0.1149	P	1.0
7	☐	1000.000	1008.663	2700172.34	0.2399	A	0.6
8	☐			2420.46	0.0002	P	3.6

$$y = 2.3761E-004 * x + 1.9648E-004$$

$$R = 0.9998$$

$$DL = 0.01975$$


$$BEC = 0.8269$$

Weight: <None>

Min Conc: 0

Calibration for 016ICSB.d

118 Sn [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	1697.89	0.0002	P	2.5
2	☐	5.000	4.397	40263.90	0.0037	P	2.8
3	☐	50.000	49.219	444121.43	0.0402	P	1.7
4	☐	125.000	122.692	1096323.70	0.0999	P	1.4
5	☐	250.000	257.125	2324187.54	0.2092	A	1.8
6	☐	500.000	503.219	4591458.72	0.4092	A	0.7
7	☐	1000.000	996.940	9123377.57	0.8105	A	1.2
8	☐			4890.94	0.0005	P	4.2

$$y = 8.1285E-004 * x + 1.5652E-004$$

$$R = 1.0000$$


$$DL = 0.01449$$

$$BEC = 0.1926$$

Weight: <None>

Min Conc: 0

121 Sb [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	180.00	0.0000	P	16.9
2	☐	2.000	1.836	18896.95	0.0018	P	4.1
3	☐	50.000	49.079	512799.86	0.0464	P	2.0
4	☐	125.000	122.482	1270006.65	0.1157	P	1.3
5	☐	250.000	257.416	2702100.69	0.2432	A	1.4
6	☐	500.000	502.052	5321227.38	0.4742	A	1.8
7	☐	1000.000	997.481	10606211.16	0.9422	A	1.2
8	☐			7041.80	0.0007	P	3.6

$$y = 9.4459E-004 * x + 1.6601E-005$$

$$R = 1.0000$$


$$DL = 0.008932$$

$$BEC = 0.01758$$

Weight: <None>

Min Conc: 0

135 Ba [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	160.01	0.0000	P	18.6
2	☐	10.000	8.682	23907.91	0.0022	P	1.9
3	☐	250.000	238.404	668122.65	0.0604	P	2.0
4	☐	625.000	645.710	1795894.80	0.1636	A	1.8
5	☐	1250.000	1288.174	3627237.58	0.3264	A	0.7
6	☐	2500.000	2507.382	7129043.85	0.6354	A	1.0
7	☐	5000.000	4984.759	14217546.59	1.2631	A	1.6
8	☐			5221.01	0.0005	P	3.3

$$y = 2.5339E-004 * x + 1.4740E-005$$

$$R = 1.0000$$


$$DL = 0.03249$$

$$BEC = 0.05817$$

Weight: <None>

Min Conc: 0

137 Ba [No Gas] ISTD :159 Tb [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	280.00	0.0000	P	11.5
2	☐	10.000	8.600	40947.67	0.0038	P	1.2
3	☐	250.000	239.799	1161822.70	0.1051	P	1.5
4	☐	625.000	651.444	3132391.24	0.2854	A	0.8
5	☐	1250.000	1280.189	6231734.21	0.5608	A	0.4
6	☐	2500.000	2501.257	12293933.71	1.0957	A	1.2
7	☐	5000.000	4989.031	24600942.69	2.1854	A	0.5
8	☐			9288.65	0.0009	P	2.1

$$y = 4.3804E-004 * x + 2.5820E-005$$


$$R = 1.0000$$

$$DL = 0.02026$$


$$BEC = 0.05895$$

Weight: <None>


Min Conc: 0

150 Nd [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			8.89		P	94.4
2	☐			14.45		P	48.0
3	☐			33.33		P	20.0
4	☐			70.00		P	12.6
5	☐			158.89		P	21.0
6	☐			278.89		P	10.8
7	☐			604.46		P	4.8
8	☐			217.78		P	4.4

150 Nd [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			7.78		P	107.
2	☐			2.22		P	86.6
3	☐			4.44		P	114.
4	☐			17.78		P	21.7
5	☐			43.33		P	15.4
6	☐			53.33		P	41.0
7	☐			137.78		P	12.4
8	☐			102.23		P	5.0

156 Dy [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			72.22		P	16.2
2	☐			64.44		P	13.0
3	☐			85.56		P	2.3
4	☐			97.78		P	8.6
5	☐			124.45		P	24.3
6	☐			152.22		P	8.8
7	☐			278.89		P	3.4
8	☐			431.12		P	10.8

156 Dy [He] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			40.00		P	30.0
2	☐			28.89		P	17.6
3	☐			64.44		P	33.3
4	☐			87.78		P	11.0
5	☐			146.67		P	24.6
6	☐			257.78		P	15.6
7	☐			587.80		P	10.5
8	☐			311.12		P	18.9

160 Gd [No Gas] Noavailable calibration poin...


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			63.33		P	37.9
2	☐			45.55		P	29.6
3	☐			38.89		P	21.6
4	☐			40.00		P	16.7
5	☐			78.89		P	35.2
6	☐			91.11		P	12.8
7	☐			144.45		P	17.6
8	☐			355.56		P	3.0

160 Gd [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			233.34		P	17.3
2	☐			248.89		P	8.9
3	☐			260.01		P	11.1
4	☐			264.45		P	13.9
5	☐			284.45		P	11.1
6	☐			286.67		P	12.8
7	☐			325.56		P	7.2
8	☐			480.01		P	3.5


164 Er [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			64.45		P	10.8
2	☐			76.67		P	26.1
3	☐			67.78		P	19.9
4	☐			86.67		P	24.0
5	☐			115.56		P	19.6
6	☐			126.67		P	9.1
7	☐			217.78		P	0.9
8	☐			463.35		P	15.8

164 Er [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			37.78		P	18.4
2	☐			45.56		P	27.7
3	☐			53.33		P	49.6
4	☐			58.89		P	31.2
5	☐			56.67		P	30.6
6	☐			76.67		P	15.7
7	☐			131.11		P	5.9
8	☐			256.67		P	5.2

203 TI [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	224.45	0.0000	P	2.6
2	☐	1.000	0.758	8055.77	0.0012	P	2.8
3	☐	50.000	42.142	448686.60	0.0671	P	1.4
4	☐	125.000	105.821	1125384.97	0.1685	P	0.4
5	☐	250.000	245.870	2656991.19	0.3915	A	0.3
6	☐	500.000	498.344	5418887.25	0.7934	A	1.9
7	☐	1000.000	1004.651	10770002.34	1.5994	A	1.6
8	☐			1214.51	0.0002	P	6.2

$$y = 0.0016 * x + 3.3344E-005$$

R = 0.9998


DL = 0.001653

BEC = 0.02094

Weight: <None>

Min Conc: 0

205 TI [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	550.02	0.0001	P	12.2
2	☐	1.000	0.776	19828.94	0.0031	P	0.5
3	☐	50.000	41.690	1066417.18	0.1595	P	0.7
4	☐	125.000	122.500	3129713.04	0.4686	A	1.4
5	☐	250.000	243.162	6312484.00	0.9301	A	1.7
6	☐	500.000	492.933	12876812.45	1.8854	A	1.5
7	☐	1000.000	1005.971	25908021.28	3.8476	A	2.1
8	☐			3080.37	0.0005	P	5.3

$$y = 0.0038 * x + 8.1768E-005$$

R = 0.9999


DL = 0.007838

BEC = 0.02138

Weight: <None>

Min Conc: 0

206 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1108.95	0.0002	P	6.0
2	☐	1.000	0.766	7413.21	0.0011	P	2.9
3	☐	250.000	242.590	2067404.78	0.3093	A	0.5
4	☐	625.000	616.131	5244489.71	0.7852	A	1.0
5	☐	1250.000	1233.270	10666180.74	1.5716	A	1.7
6	☐	2500.000	2472.262	21516107.88	3.1503	A	1.8
7	☐	5000.000	5019.531	43067997.97	6.3960	A	1.9
8	☐			7260.89	0.0012	P	1.5

$$y = 0.0013 * x + 1.6470E-004$$

$$R = 1.0000$$


$$DL = 0.02333$$

$$BEC = 0.1293$$

Weight: <None>

Min Conc: 0

207 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	947.82	0.0001	P	6.5
2	☐	1.000	0.769	6554.98	0.0010	P	3.3
3	☐	250.000	243.904	1840046.97	0.2753	A	0.6
4	☐	625.000	613.789	4625111.53	0.6925	A	0.1
5	☐	1250.000	1236.043	9463573.06	1.3944	A	1.6
6	☐	2500.000	2466.097	18999883.47	2.7819	A	1.6
7	☐	5000.000	5022.147	38145065.82	5.6651	A	2.1
8	☐			6427.13	0.0011	P	3.2

$$y = 0.0011 * x + 1.4097E-004$$

$$R = 1.0000$$


$$DL = 0.0244$$

$$BEC = 0.125$$

Weight: <None>

Min Conc: 0

208 Pb [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	4303.64	0.0006	P	2.0
2	☐	1.000	0.767	29696.29	0.0046	P	3.1
3	☐	250.000	243.552	8342806.09	1.2480	A	0.4
4	☐	625.000	613.556	20992659.41	3.1431	A	0.1
5	☐	1250.000	1240.561	43128208.18	6.3544	A	1.1
6	☐	2500.000	2466.112	86270647.33	12.6314	A	1.5
7	☐	5000.000	5021.056	173166563.5	25.7171	A	1.6
8	☐			29450.60	0.0050	P	1.5

$$y = 0.0051 * x + 6.3964E-004$$

$$R = 1.0000$$


$$DL = 0.007365$$

$$BEC = 0.1249$$

Weight: <None>

Min Conc: 0

238 U [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	284.45	0.0000	P	17.8
2	☐	1.000	0.738	24557.13	0.0038	P	1.0
3	☐	50.000	39.782	1346852.40	0.2015	P	1.1
4	☐	125.000	117.557	3975939.18	0.5953	A	1.3
5	☐	250.000	239.041	8215013.21	1.2104	A	2.6
6	☐	500.000	486.411	16822230.17	2.4630	A	2.3
7	☐	1000.000	1010.976	34472178.93	5.1192	A	2.7
8	☐			1917.94	0.0003	P	4.1

$$y = 0.0051 * x + 4.2199E-005$$


$$R = 0.9998$$

$$DL = 0.004459$$


$$BEC = 0.008334$$

Weight: <None>


Min Conc: 0

6 Li (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		1424107.08		A	0.9
2	☐	1.000		1408244.03		A	1.3
3	☐	1.000		1409666.61		A	1.9
4	☐	1.000		1415005.58		A	1.3
5	☐	1.000		1424983.66		A	1.9
6	☐	1.000		1433436.03		A	1.0
7	☐	1.000		1423053.73		A	1.8
8	☐	1.000		1362701.78		A	1.3

45 Sc (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4791933.09		A	1.2
2	☐	1.000		4755613.44		A	1.1
3	☐	1.000		4773584.06		A	1.2
4	☐	1.000		4684398.99		A	1.3
5	☐	1.000		4560543.30		A	1.3
6	☐	1.000		4445352.85		A	0.6
7	☐	1.000		4449753.16		A	0.6
8	☐	1.000		4455690.49		A	0.8

45 Sc (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		399230.62		P	7.2
2	☐	1.000		409138.62		P	0.7
3	☐	1.000		413503.26		P	0.9
4	☐	1.000		407572.45		P	0.2
5	☐	1.000		403250.94		P	0.4
6	☐	1.000		404434.30		P	2.2
7	☐	1.000		398497.58		P	0.6
8	☐	1.000		408439.91		P	0.2

89 Y (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		13946360.76		A	0.7
2	☐	1.000		13753676.74		A	1.3
3	☐	1.000		13977590.07		A	0.5
4	☐	1.000		13627787.16		A	0.8
5	☐	1.000		13725099.24		A	0.7
6	☐	1.000		13627694.80		A	1.0
7	☐	1.000		13381425.77		A	0.6
8	☐	1.000		13027237.58		A	0.4

89 Y(ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		3195447.35		A	7.2
2	☐	1.000		3236289.29		A	0.2
3	☐	1.000		3275680.44		A	1.7
4	☐	1.000		3236213.25		A	0.9
5	☐	1.000		3257769.92		A	0.9
6	☐	1.000		3252373.87		A	2.1
7	☐	1.000		3188475.51		A	1.5
8	☐	1.000		3168554.43		A	0.7

103 Rh(ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		7984811.27		A	1.6
2	☐	1.000		7927462.52		A	1.0
3	☐	1.000		7930784.47		A	1.4
4	☐	1.000		7790558.84		A	0.3
5	☐	1.000		7709348.84		A	0.5
6	☐	1.000		7531992.25		A	0.8
7	☐	1.000		7286093.36		A	1.0
8	☐	1.000		6636653.51		A	1.6

103 Rh (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4360797.96		A	7.2
2	☐	1.000		4490168.96		A	0.8
3	☐	1.000		4507696.43		A	1.3
4	☐	1.000		4432656.57		A	1.4
5	☐	1.000		4380793.51		A	0.5
6	☐	1.000		4364698.82		A	2.3
7	☐	1.000		4217671.67		A	0.4
8	☐	1.000		3900791.51		A	0.3

115 In (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		8454121.03		A	0.5
2	☐	1.000		8444231.05		A	1.1
3	☐	1.000		8524258.86		A	0.2
4	☐	1.000		8308893.17		A	0.8
5	☐	1.000		8305644.16		A	0.5
6	☐	1.000		8151128.66		A	1.4
7	☐	1.000		7954653.79		A	0.7
8	☐	1.000		7676561.83		A	0.3

115 In (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2551039.90		A	7.7
2	☐	1.000		2581109.67		A	0.6
3	☐	1.000		2634228.69		A	0.7
4	☐	1.000		2586339.48		A	1.3
5	☐	1.000		2556029.67		A	0.9
6	☐	1.000		2543696.07		A	1.7
7	☐	1.000		2428547.72		A	0.3
8	☐	1.000		2423650.37		A	0.7

159 Tb (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		10848837.34		A	0.7
2	☐	1.000		10797281.58		A	1.8
3	☐	1.000		11058807.89		A	1.0
4	☐	1.000		10976333.03		A	0.7
5	☐	1.000		11112241.08		A	0.2
6	☐	1.000		11221322.89		A	1.5
7	☐	1.000		11257124.28		A	1.1
8	☐	1.000		10781371.23		A	1.6

159 Tb (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		6280668.66		A	8.5
2	☐	1.000		6492679.69		A	0.2
3	☐	1.000		6536123.45		A	0.4
4	☐	1.000		6600697.54		A	0.8
5	☐	1.000		6656903.30		A	0.5
6	☐	1.000		6820150.59		A	2.2
7	☐	1.000		6692266.64		A	0.7
8	☐	1.000		6424312.54		A	0.5

165 Ho (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		10607123.31		A	1.9
2	☐	1.000		10449300.61		A	0.7
3	☐	1.000		10669084.36		A	0.6
4	☐	1.000		10575440.67		A	0.6
5	☐	1.000		10759963.87		A	0.4
6	☐	1.000		10953370.95		A	1.4
7	☐	1.000		10953771.23		A	0.7
8	☐	1.000		10282893.32		A	2.1

165 Ho (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		6311314.35		A	7.4
2	☐	1.000		6454941.43		A	0.6
3	☐	1.000		6649752.40		A	0.8
4	☐	1.000		6566244.00		A	1.6
5	☐	1.000		6678544.07		A	0.4
6	☐	1.000		6830215.66		A	1.7
7	☐	1.000		6725705.94		A	0.4
8	☐	1.000		6455119.35		A	0.9

209 Bi (ISTD) [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		6729365.74		A	2.2
2	☐	1.000		6498734.21		A	1.2
3	☐	1.000		6684744.90		A	0.3
4	☐	1.000		6678952.26		A	0.2
5	☐	1.000		6787507.75		A	0.9
6	☐	1.000		6830241.57		A	0.6
7	☐	1.000		6734152.61		A	1.1
8	☐	1.000		5898710.33		A	4.2



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4596871.25		A	6.9
2	☐	1.000		4672169.06		A	1.3
3	☐	1.000		4769388.82		A	0.3
4	☐	1.000		4765129.31		A	2.1
5	☐	1.000		4757359.48		A	0.7
6	☐	1.000		4790297.88		A	1.3
7	☐	1.000		4633056.39		A	0.3
8	☐	1.000		4121310.11		A	0.9

US EPA Tune Check Report

Reviewed By:mohan
On:11/7/2024 3:50:57
AM
Inst Id :P7
ClientSampleId
:TUNE001

Operator Name Jaswal
Acq/Data Batch D:\Agilent\ICPMH\1\DATA\IP7110224 MS.b
Acq. Date-Time 2024-11-02 17:07:32
Report Comment ---
Instrument Name G8403A JP14410463

[No Gas]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
9		10809	108088.38			1.258	5.000
24		326802	3268017.55			0.800	5.000
25		43495	434950.78			0.899	5.000
26		50916	509163.79			1.132	5.000
59		69375	693746.18			1.385	5.000
113		5454	54539.22			1.619	5.000
115		128421	1284210.23			2.268	5.000
206		26576	265764.47			2.046	5.000
207		21856	218559.05			2.227	5.000
208		53699	536991.36			2.291	5.000
220		0	4.60			54.021	

Mass	RSD% (Flag)
9	
24	
25	
26	
59	
113	
115	
206	
207	
208	
220	

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
9	11006	10864	10803	10716	10656
24	330641	328240	325927	324776	324425
25	44035	43697	43455	43269	43020
26	51614	51270	50962	50607	50129
59	70587	70074	69243	68739	68229
113	5560	5501	5481	5386	5342
115	132428	129963	128192	126562	124960
206	27398	26774	26480	26245	25986
207	22607	22009	21769	21544	21351
208	55462	54245	53588	52936	52265


US EPA Tune Check Report

Reviewed By:mohan
On:11/7/2024 3:50:57
AM
Inst Id :P7
ClientSampleId
:TUNE001

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
220	0	0	1	1	0

Integration Time [sec] 0.1

Resolution/Axis



Mass	Peak Height	Axis	Axis (Required)	Axis (Flag)
9	19019.73	8.95	8.90 - 9.10	
24	577292.17	23.95	23.90 - 24.10	
25	77199.70	24.95	24.90 - 25.10	
26	90956.28	25.95	25.90 - 26.10	
59	120071.88	58.95	58.90 - 59.10	
113	10642.36	112.95	112.90 - 113.10	
115	250606.85	114.95	114.90 - 115.10	
206	55325.40	206.00	205.90 - 206.10	
207	45941.92	207.00	206.90 - 207.10	
208	112655.31	208.00	207.90 - 208.10	
220			-	

Mass	W-50%	W-5%	W-5% (Required)	W-5% (Flag)
9	0.59	0.735	0.900	
24	0.61	0.819	0.900	
25	0.59	0.743	0.900	
26	0.60	0.773	0.900	
59	0.60	0.768	0.900	
113	0.52	0.735	0.900	
115	0.51	0.754	0.900	
206	0.49	0.739	0.900	
207	0.49	0.734	0.900	
208	0.49	0.755	0.900	
220				

Integration Time [sec] 0.1

Acquisition Time [sec] 256.770000000002

Y Axis Linear

Tune Parameters

Plasma Parameters

Plasma Mode --- Nebulizer Gas 0.82 L/min Dilution Gas 0.40 L/min

US EPA Tune Check Report

Reviewed By:mohan
On:11/7/2024 3:50:57
AM
Inst Id :P7
ClientSampleId
TUNE001

RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.9 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	0.0 V	Omega Lens	8.9 V	Deflect	15.6 V
Extract 2	-180.0 V	Cell Entrance	-30 V	Plate Bias	-35 V
Omega Bias	-85 V	Cell Exit	-50 V		

Cell Parameters

Use Gas	No	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	0.0 mL/min	OctP Bias	-8.0 V		
H2 Flow	---	OctP RF	190 V		

QP Parameters

Mass Gain	135	Axis Gain	0.9979	QP Bias	-3.0 V
Mass Offset	130	Axis Offset	0.08		

Hardware Settings

Torch

Torch H	-0.6 mm	Torch V	0.5 mm

EM

Discriminator	4.9 mV	Analog HV	2429 V	Pulse HV	2000 V

[He]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
59		23978	239779.74			2.631	
89		114912	1149116.81			3.482	
205		31596	315961.45			2.862	

Mass	RSD% (Flag)
59	
89	
205	

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
59	24450	23029	23625	24398	24389
89	120136	109088	113984	115044	116308
205	32198	30242	31079	32198	32264

Integration Time [sec] 0.1

Tune Parameters

Plasma Parameters

Plasma Mode	---	Nebulizer Gas	0.82 L/min	Dilution Gas	0.40 L/min
RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	0.0 V	Omega Lens	8.5 V	Deflect	4.0 V

US EPA Tune Check Report

Reviewed By:mohan
On:11/7/2024 3:50:57
AM
Inst Id :P7
ClientSampleId
TUNE001

Extract 2	-155.0 V	Cell Entrance	-40 V	Plate Bias	-60 V
Omega Bias	-80 V	Cell Exit	-60 V		
Cell Parameters					
Use Gas	Yes	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	3.9 mL/min	OctP Bias	-18.0 V		
H2 Flow	---	OctP RF	200 V		
QP Parameters					
Mass Gain	135	Axis Gain	0.9979	QP Bias	-13.0 V
Mass Offset	130	Axis Offset	0.08		
Hardware Settings					
Torch					
Torch H	-0.6 mm	Torch V	0.5 mm		
EM					
Discriminator	4.9 mV	Analog HV	2429 V	Pulse HV	2000 V

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:41:38 DataFile Name : 004CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-0.02	-0.09	0.11	0.00	N/A	ppb
Antimony	121-1	0.00	0.00	0.00	0.00	N/A	ppb
Arsenic	75-2	0.01	-0.01	0.00	0.00	N/A	ppb
Barium	135-1	0.01	-0.01	0.00	0.00	N/A	ppb
Barium	137-1	0.00	0.01	-0.01	0.00	N/A	ppb
Beryllium	9-1	-0.02	0.00	0.02	0.00	N/A	ppb
Bismuth	209-1				100		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.01	0.01	-0.02	0.00	N/A	ppb
Cadmium	106-1	-0.04	0.11	-0.06	0.00	N/A	ppb
Cadmium	111-1	-0.01	0.01	0.00	0.00	N/A	ppb
Calcium	43-1	0.76	-1.26	0.50	0.00	N/A	ppb
Calcium	44-1	-0.28	-0.17	0.45	0.00	N/A	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.01	0.00	0.00	0.00	N/A	ppb
Cobalt	59-2	0.00	0.00	0.00	0.00	N/A	ppb
Copper	63-2	0.00	-0.02	0.02	0.00	N/A	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				100		%
Indium	115-1				100		%
Indium	115-2				100		%
Iron	56-2	-0.04	-0.24	0.29	0.00	N/A	ppb
Iron	57-2	0.20	-0.73	0.53	0.00	N/A	ppb
Iron	54-2	0.30	-1.31	1.00	0.00	N/A	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:41:38 DataFile Name : 004CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.00	0.01	-0.01	0.00	N/A	ppb
Lead	207-1	-0.01	0.00	0.01	0.00	N/A	ppb
Lead	208-1	0.00	0.00	0.00	0.00	N/A	ppb
Lithium	6-1				100		%
Magnesium	24-2	-0.26	0.12	0.14	0.00	N/A	ppb
Manganese	55-2	-0.02	0.02	0.00	0.00	N/A	ppb
Molybdenum	94-1	0.01	-0.01	0.00	0.00	N/A	ppb
Molybdenum	95-1	0.01	0.00	-0.01	0.00	N/A	ppb
Molybdenum	96-1	0.00	0.01	0.00	0.00	N/A	ppb
Molybdenum	97-1	0.00	-0.01	0.01	0.00	N/A	ppb
Molybdenum	98-1	0.00	0.00	0.00	0.00	N/A	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.01	0.00	0.00	0.00	N/A	ppb
Phosphorus	31-2	-13.27	-15.38	-17.63	-15.43		ppb
Potassium	39-2	-4.22	-4.76	8.97	0.00	N/A	ppb
Rhodium	103-1				100		%
Rhodium	103-2				100		%
Scandium	45-1				100		%
Scandium	45-2				100		%
Selenium	82-1	0.19	-0.22	0.02	0.00	N/A	ppb
Selenium	77-2	0.00	0.23	-0.22	0.00	N/A	ppb
Selenium	78-2	-0.55	-0.47	1.02	0.00	N/A	ppb
Silicon	28-1	-0.14	-0.08	0.22	0.00	N/A	ppb
Silver	107-1	0.00	0.00	0.00	0.00	N/A	ppb
Silver	109-1	0.00	0.00	0.00	0.00	N/A	ppb
Sodium	23-2	-1.78	-2.27	4.05	0.00	N/A	ppb
Strontium	86-1	0.00	0.00	-0.01	0.00	N/A	ppb
Strontium	88-1	0.01	-0.01	0.00	0.00	N/A	ppb
Sulfur	34-1	-73.48	16.23	86.32	9.69	826.70	ppb
Terbium	159-1				100		%
Terbium	159-2				100		%
Thallium	203-1	0.00	0.00	0.00	0.00	N/A	ppb
Thallium	205-1	0.00	0.00	0.00	0.00	N/A	ppb
Tin	118-1	0.00	0.01	0.00	0.00	N/A	ppb
Titanium	47-1	-0.01	-0.02	0.03	0.00	N/A	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:41:38 DataFile Name : 004CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	N/A	ppb
Vanadium	51-2	0.00	0.00	0.00	0.00	N/A	ppb
Yttrium	89-1				100		%
Yttrium	89-2				100		%
Zinc	66-2	0.12	-0.12	0.00	0.00	N/A	ppb
Zirconium	90-1	0.00	0.00	0.00	0.00	N/A	ppb
Zirconium	91-1	-0.01	0.00	0.01	0.00	N/A	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S02 Instrumnet Name : P7
Client Sample ID : S02 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:48:10 DataFile Name : 006CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	17.16	18.17	18.08	17.80	3.14	ppb
Antimony	121-1	1.76	1.92	1.83	1.84	4.14	ppb
Arsenic	75-2	0.88	0.86	1.11	0.95	14.58	ppb
Barium	135-1	8.50	8.82	8.73	8.68	1.92	ppb
Barium	137-1	8.50	8.71	8.59	8.60	1.20	ppb
Beryllium	9-1	0.94	0.98	0.99	0.97	2.91	ppb
Bismuth	209-1				97		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.94	1.22	1.18	1.12	13.59	ppb
Cadmium	106-1	0.64	1.25	1.11	1.00	32.08	ppb
Cadmium	111-1	0.95	1.09	0.94	0.99	8.77	ppb
Calcium	43-1	491.13	476.49	488.06	485.23	1.59	ppb
Calcium	44-1	474.98	488.43	492.58	485.33	1.90	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1.93	1.94	1.91	1.92	0.80	ppb
Cobalt	59-2	0.99	0.98	1.00	0.99	1.02	ppb
Copper	63-2	1.84	1.83	1.83	1.83	0.38	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				99		%
Holmium	165-2				102		%
Indium	115-1				100		%
Indium	115-2				101		%
Iron	56-2	50.55	51.56	51.36	51.15	1.05	ppb
Iron	57-2	51.33	50.83	52.11	51.42	1.26	ppb
Iron	54-2	53.50	53.51	52.90	53.30	0.66	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S02 Instrumnet Name : P7
Client Sample ID : S02 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:48:10 DataFile Name : 006CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.74	0.79	0.76	0.77	3.45	ppb
Lead	207-1	0.74	0.78	0.79	0.77	3.78	ppb
Lead	208-1	0.74	0.80	0.77	0.77	3.58	ppb
Lithium	6-1				99		%
Magnesium	24-2	504.06	505.14	499.98	503.06	0.54	ppb
Manganese	55-2	0.93	0.99	1.00	0.98	3.95	ppb
Molybdenum	94-1	4.87	4.98	4.97	4.94	1.18	ppb
Molybdenum	95-1	4.02	4.08	4.10	4.07	1.07	ppb
Molybdenum	96-1	4.03	4.10	4.14	4.09	1.30	ppb
Molybdenum	97-1	4.06	4.13	4.07	4.09	0.95	ppb
Molybdenum	98-1	3.97	4.09	4.14	4.07	2.21	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.04	0.91	0.93	0.96	7.56	ppb
Phosphorus	31-2	10.37	15.14	20.77	15.43	33.74	ppb
Potassium	39-2	481.02	482.32	478.47	480.60	0.41	ppb
Rhodium	103-1				99		%
Rhodium	103-2				103		%
Scandium	45-1				99		%
Scandium	45-2				102		%
Selenium	82-1	4.72	4.77	4.67	4.72	1.03	ppb
Selenium	77-2	5.51	5.07	3.90	4.82	17.23	ppb
Selenium	78-2	4.67	6.53	5.64	5.61	16.61	ppb
Silicon	28-1	1.16	1.38	1.56	1.37	14.42	ppb
Silver	107-1	0.87	0.93	0.90	0.90	3.35	ppb
Silver	109-1	0.89	0.95	0.91	0.92	3.68	ppb
Sodium	23-2	477.80	479.01	476.63	477.81	0.25	ppb
Strontium	86-1	0.88	0.92	0.93	0.91	3.26	ppb
Strontium	88-1	0.88	0.91	0.91	0.90	2.16	ppb
Sulfur	34-1	-127.31	15.06	83.18	-9.69		ppb
Terbium	159-1				100		%
Terbium	159-2				103		%
Thallium	203-1	0.73	0.76	0.78	0.76	2.89	ppb
Thallium	205-1	0.78	0.77	0.78	0.78	0.50	ppb
Tin	118-1	4.26	4.51	4.42	4.40	2.94	ppb
Titanium	47-1	4.20	4.40	4.24	4.28	2.39	ppb

LB Number :LB133265

Operator :Jaswal

Lab Sample ID :S02

Instrumnet Name :P7

Client Sample ID :S02

Dilution Factor :1

Date & Time Acquired :2024-11-02 17:48:10

DataFile Name :006CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.73	0.74	0.74	0.74	0.97	ppb
Vanadium	51-2	4.82	4.95	4.89	4.89	1.37	ppb
Yttrium	89-1				99		%
Yttrium	89-2				101		%
Zinc	66-2	4.52	4.76	4.84	4.71	3.50	ppb
Zirconium	90-1	0.90	0.91	0.91	0.91	0.48	ppb
Zirconium	91-1	0.87	0.95	0.88	0.90	4.53	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S03 Instrumnet Name : P7
Client Sample ID : S03 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:51:25 DataFile Name : 007CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	991.88	1000.02	990.32	994.07	0.52	ppb
Antimony	121-1	47.96	49.37	49.90	49.08	2.05	ppb
Arsenic	75-2	51.49	51.09	53.69	52.09	2.68	ppb
Barium	135-1	233.57	238.38	243.27	238.40	2.03	ppb
Barium	137-1	236.03	240.42	242.94	239.80	1.46	ppb
Beryllium	9-1	53.07	53.09	54.91	53.69	1.97	ppb
Bismuth	209-1				99		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	51.77	54.94	53.07	53.26	2.99	ppb
Cadmium	106-1	53.25	54.67	53.97	53.96	1.32	ppb
Cadmium	111-1	50.39	50.87	51.85	51.04	1.46	ppb
Calcium	43-1	5889.84	5949.55	6072.28	5970.56	1.56	ppb
Calcium	44-1	5319.14	5298.13	5415.49	5344.25	1.17	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	52.77	52.23	51.60	52.20	1.12	ppb
Cobalt	59-2	51.73	50.66	51.15	51.18	1.05	ppb
Copper	63-2	534.62	535.01	529.29	532.98	0.60	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				105		%
Indium	115-1				101		%
Indium	115-2				103		%
Iron	56-2	2661.64	2681.70	2632.64	2658.66	0.93	ppb
Iron	57-2	2755.36	2777.98	2714.14	2749.16	1.18	ppb
Iron	54-2	2803.01	2802.44	2779.59	2795.01	0.48	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S03 Instrumnet Name : P7
Client Sample ID : S03 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:51:25 DataFile Name : 007CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	241.47	242.27	244.03	242.59	0.54	ppb
Lead	207-1	242.38	245.28	244.05	243.90	0.60	ppb
Lead	208-1	243.16	244.72	242.78	243.55	0.42	ppb
Lithium	6-1				99		%
Magnesium	24-2	4927.48	4913.55	4991.82	4944.28	0.84	ppb
Manganese	55-2	520.55	528.32	498.86	515.91	2.96	ppb
Molybdenum	94-1	494.64	499.47	505.29	499.80	1.07	ppb
Molybdenum	95-1	499.96	501.65	500.57	500.73	0.17	ppb
Molybdenum	96-1	495.92	502.18	495.08	497.73	0.78	ppb
Molybdenum	97-1	491.45	503.01	496.99	497.15	1.16	ppb
Molybdenum	98-1	496.15	494.15	491.82	494.04	0.44	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	51.20	51.12	50.87	51.06	0.34	ppb
Phosphorus	31-2	1043.66	1033.46	1037.55	1038.22	0.49	ppb
Potassium	39-2	2510.29	2517.04	2501.05	2509.46	0.32	ppb
Rhodium	103-1				99		%
Rhodium	103-2				103		%
Scandium	45-1				100		%
Scandium	45-2				104		%
Selenium	82-1	52.91	51.96	52.80	52.56	0.99	ppb
Selenium	77-2	48.52	55.76	50.37	51.55	7.29	ppb
Selenium	78-2	48.71	51.11	55.51	51.77	6.66	ppb
Silicon	28-1	41.13	42.01	42.74	41.96	1.92	ppb
Silver	107-1	50.73	51.78	53.13	51.88	2.32	ppb
Silver	109-1	50.79	51.74	52.45	51.66	1.61	ppb
Sodium	23-2	5029.98	4999.92	4953.24	4994.38	0.77	ppb
Strontium	86-1	1268.52	1271.11	1264.90	1268.18	0.25	ppb
Strontium	88-1	1252.30	1252.09	1259.66	1254.68	0.34	ppb
Sulfur	34-1	825.58	884.94	906.35	872.29	4.80	ppb
Terbium	159-1				102		%
Terbium	159-2				104		%
Thallium	203-1	41.55	42.71	42.18	42.14	1.38	ppb
Thallium	205-1	41.39	41.70	41.98	41.69	0.70	ppb
Tin	118-1	48.64	48.84	50.18	49.22	1.69	ppb
Titanium	47-1	502.81	511.76	520.23	511.60	1.70	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

S03

Instrumnet Name :

P7

Client Sample ID :

S03

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 17:51:25

DataFile Name :

007CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	39.30	39.85	40.19	39.78	1.13	ppb
Vanadium	51-2	52.00	52.47	51.17	51.88	1.26	ppb
Yttrium	89-1				100		%
Yttrium	89-2				103		%
Zinc	66-2	513.04	508.93	507.42	509.80	0.57	ppb
Zirconium	90-1	48.67	49.22	49.25	49.05	0.67	ppb
Zirconium	91-1	48.76	49.22	49.78	49.25	1.04	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P7
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:54:26 DataFile Name : 008CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	2460.48	2496.01	2456.44	2470.98	0.88	ppb
Antimony	121-1	120.61	123.40	123.44	122.48	1.33	ppb
Arsenic	75-2	128.27	129.18	129.84	129.10	0.61	ppb
Barium	135-1	633.14	647.45	656.54	645.71	1.83	ppb
Barium	137-1	646.05	656.14	652.14	651.44	0.78	ppb
Beryllium	9-1	133.58	132.05	132.06	132.56	0.67	ppb
Bismuth	209-1				99		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	130.93	133.89	137.00	133.94	2.27	ppb
Cadmium	106-1	131.77	132.82	133.75	132.78	0.75	ppb
Cadmium	111-1	126.20	126.69	127.14	126.67	0.37	ppb
Calcium	43-1	14724.32	14685.59	15070.12	14826.68	1.43	ppb
Calcium	44-1	13092.13	13070.30	13300.58	13154.34	0.97	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	130.23	129.12	129.18	129.51	0.48	ppb
Cobalt	59-2	128.73	127.23	127.13	127.70	0.70	ppb
Copper	63-2	1309.59	1328.93	1293.70	1310.74	1.35	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				104		%
Indium	115-1				98		%
Indium	115-2				101		%
Iron	56-2	6593.69	6581.99	6555.75	6577.14	0.30	ppb
Iron	57-2	6807.97	6851.27	6768.86	6809.36	0.61	ppb
Iron	54-2	6995.27	6986.04	6682.53	6887.95	2.58	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P7
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:54:26 DataFile Name : 008CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	622.86	612.10	613.43	616.13	0.95	ppb
Lead	207-1	613.28	613.79	614.30	613.79	0.08	ppb
Lead	208-1	612.65	614.29	613.73	613.56	0.14	ppb
Lithium	6-1				99		%
Magnesium	24-2	12420.14	12372.53	12300.21	12364.29	0.49	ppb
Manganese	55-2	1266.66	1276.78	1264.46	1269.30	0.52	ppb
Molybdenum	94-1	1254.95	1259.30	1268.31	1260.86	0.54	ppb
Molybdenum	95-1	1251.82	1263.99	1275.76	1263.85	0.95	ppb
Molybdenum	96-1	1239.57	1265.59	1270.31	1258.49	1.32	ppb
Molybdenum	97-1	1236.65	1259.76	1268.20	1254.87	1.30	ppb
Molybdenum	98-1	1235.41	1254.66	1271.86	1253.98	1.45	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	126.20	128.03	126.52	126.92	0.77	ppb
Phosphorus	31-2	2569.43	2608.03	2516.47	2564.64	1.79	ppb
Potassium	39-2	6298.37	6125.51	6146.67	6190.18	1.52	ppb
Rhodium	103-1				98		%
Rhodium	103-2				102		%
Scandium	45-1				98		%
Scandium	45-2				102		%
Selenium	82-1	128.62	129.84	130.41	129.62	0.71	ppb
Selenium	77-2	126.40	134.45	139.20	133.35	4.85	ppb
Selenium	78-2	134.19	131.56	135.47	133.74	1.49	ppb
Silicon	28-1	101.52	101.85	103.59	102.32	1.09	ppb
Silver	107-1	134.25	138.27	137.37	136.63	1.54	ppb
Silver	109-1	134.33	137.39	136.99	136.24	1.22	ppb
Sodium	23-2	12495.35	12326.64	12349.92	12390.64	0.74	ppb
Strontium	86-1	3152.32	3184.12	3195.76	3177.40	0.71	ppb
Strontium	88-1	3134.27	3168.88	3186.00	3163.05	0.83	ppb
Sulfur	34-1	2129.58	2192.56	2360.46	2227.53	5.36	ppb
Terbium	159-1				101		%
Terbium	159-2				105		%
Thallium	203-1	106.03	106.04	105.39	105.82	0.35	ppb
Thallium	205-1	124.24	120.78	122.48	122.50	1.41	ppb
Tin	118-1	120.80	123.21	124.07	122.69	1.38	ppb
Titanium	47-1	1247.22	1245.77	1270.43	1254.48	1.10	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

S04

Instrumnet Name :

P7

Client Sample ID :

S04

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 17:54:26

DataFile Name :

008CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	118.96	115.98	117.73	117.56	1.27	ppb
Vanadium	51-2	128.28	129.10	127.69	128.36	0.55	ppb
Yttrium	89-1				98		%
Yttrium	89-2				101		%
Zinc	66-2	1272.97	1262.81	1261.05	1265.61	0.51	ppb
Zirconium	90-1	124.56	128.44	127.50	126.83	1.60	ppb
Zirconium	91-1	122.85	124.87	126.18	124.63	1.35	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P7
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:57:16 DataFile Name : 009CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	4935.40	4943.65	4962.41	4947.15	0.28	ppb
Antimony	121-1	253.32	259.83	259.10	257.42	1.39	ppb
Arsenic	75-2	256.22	257.02	252.24	255.16	1.00	ppb
Barium	135-1	1282.56	1282.94	1299.02	1288.17	0.73	ppb
Barium	137-1	1274.88	1282.05	1283.64	1280.19	0.36	ppb
Beryllium	9-1	262.53	263.57	259.51	261.87	0.81	ppb
Bismuth	209-1				101		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	260.33	260.23	265.49	262.02	1.15	ppb
Cadmium	106-1	258.05	260.24	259.85	259.38	0.45	ppb
Cadmium	111-1	246.89	251.21	249.53	249.21	0.87	ppb
Calcium	43-1	29339.11	30085.65	30164.88	29863.21	1.53	ppb
Calcium	44-1	26448.49	26765.06	26636.74	26616.77	0.60	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	256.43	258.14	256.17	256.91	0.42	ppb
Cobalt	59-2	259.52	258.03	263.34	260.30	1.05	ppb
Copper	63-2	2602.20	2599.93	2577.65	2593.26	0.52	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				106		%
Indium	115-1				98		%
Indium	115-2				100		%
Iron	56-2	13116.18	13066.21	13113.61	13098.67	0.21	ppb
Iron	57-2	13596.18	13507.17	13650.12	13584.49	0.53	ppb
Iron	54-2	13433.61	13443.65	13647.54	13508.27	0.89	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P7
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:57:16 DataFile Name : 009CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1218.44	1256.57	1224.80	1233.27	1.66	ppb
Lead	207-1	1215.56	1254.17	1238.40	1236.04	1.57	ppb
Lead	208-1	1225.66	1252.19	1243.83	1240.56	1.09	ppb
Lithium	6-1				100		%
Magnesium	24-2	24731.40	24908.01	24525.86	24721.76	0.77	ppb
Manganese	55-2	2565.73	2517.37	2538.53	2540.54	0.95	ppb
Molybdenum	94-1	2469.73	2503.76	2470.06	2481.18	0.79	ppb
Molybdenum	95-1	2485.00	2535.05	2502.12	2507.39	1.01	ppb
Molybdenum	96-1	2481.92	2509.19	2484.26	2491.79	0.61	ppb
Molybdenum	97-1	2472.77	2487.41	2503.93	2488.04	0.63	ppb
Molybdenum	98-1	2443.34	2502.62	2496.49	2480.82	1.31	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	252.48	250.44	251.88	251.60	0.42	ppb
Phosphorus	31-2	5203.37	5098.19	5159.08	5153.55	1.02	ppb
Potassium	39-2	12438.96	12333.78	12397.98	12390.24	0.43	ppb
Rhodium	103-1				97		%
Rhodium	103-2				100		%
Scandium	45-1				95		%
Scandium	45-2				101		%
Selenium	82-1	250.26	256.10	255.21	253.86	1.24	ppb
Selenium	77-2	251.51	257.24	259.59	256.11	1.62	ppb
Selenium	78-2	258.69	267.99	263.60	263.43	1.76	ppb
Silicon	28-1	200.14	204.04	205.12	203.10	1.29	ppb
Silver	107-1	264.52	266.33	268.19	266.35	0.69	ppb
Silver	109-1	264.83	263.85	268.26	265.65	0.87	ppb
Sodium	23-2	24915.74	24903.11	24847.36	24888.74	0.15	ppb
Strontium	86-1	6176.30	6308.14	6307.52	6263.99	1.21	ppb
Strontium	88-1	6197.12	6311.81	6221.98	6243.64	0.97	ppb
Sulfur	34-1	4679.20	4915.08	4945.46	4846.58	3.01	ppb
Terbium	159-1				102		%
Terbium	159-2				106		%
Thallium	203-1	245.31	245.58	246.73	245.87	0.31	ppb
Thallium	205-1	238.41	245.49	245.59	243.16	1.69	ppb
Tin	118-1	252.86	256.57	261.94	257.13	1.78	ppb
Titanium	47-1	2489.42	2554.20	2554.31	2532.64	1.48	ppb

LB Number :LB133265

Operator :Jaswal

Lab Sample ID :S05

Instrumnet Name :P7

Client Sample ID :S05

Dilution Factor :1

Date & Time Acquired :2024-11-02 17:57:16

DataFile Name :009CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	232.03	241.92	243.17	239.04	2.55	ppb
Vanadium	51-2	256.54	258.42	257.04	257.33	0.38	ppb
Yttrium	89-1				98		%
Yttrium	89-2				102		%
Zinc	66-2	2586.63	2590.57	2600.04	2592.41	0.27	ppb
Zirconium	90-1	246.25	251.67	254.19	250.70	1.62	ppb
Zirconium	91-1	244.01	248.52	247.07	246.53	0.93	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P7
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:00:04 DataFile Name : 010CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	9414.28	9591.28	9730.92	9578.82	1.66	ppb
Antimony	121-1	500.17	494.34	511.64	502.05	1.75	ppb
Arsenic	75-2	487.51	503.01	497.68	496.07	1.59	ppb
Barium	135-1	2506.10	2482.01	2534.03	2507.38	1.04	ppb
Barium	137-1	2504.70	2469.81	2529.26	2501.26	1.19	ppb
Beryllium	9-1	503.78	506.35	520.14	510.09	1.72	ppb
Bismuth	209-1				101		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	504.40	507.30	520.68	510.79	1.70	ppb
Cadmium	106-1	505.40	507.25	513.35	508.67	0.82	ppb
Cadmium	111-1	484.83	477.06	485.74	482.54	0.99	ppb
Calcium	43-1	56626.53	57393.99	57669.96	57230.16	0.94	ppb
Calcium	44-1	52299.87	53292.98	53144.63	52912.49	1.01	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	488.61	507.50	502.90	499.67	1.97	ppb
Cobalt	59-2	486.13	514.60	502.27	501.00	2.85	ppb
Copper	63-2	4895.88	5070.62	5042.88	5003.13	1.88	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				108		%
Indium	115-1				96		%
Indium	115-2				100		%
Iron	56-2	25192.50	26239.39	26004.91	25812.27	2.13	ppb
Iron	57-2	25174.65	25999.05	26021.67	25731.79	1.88	ppb
Iron	54-2	25672.19	26635.36	26525.30	26277.62	2.01	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P7
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:00:04 DataFile Name : 010CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2430.76	2466.62	2519.40	2472.26	1.80	ppb
Lead	207-1	2430.74	2458.55	2509.01	2466.10	1.61	ppb
Lead	208-1	2435.16	2455.17	2508.01	2466.11	1.53	ppb
Lithium	6-1				101		%
Magnesium	24-2	47217.60	48474.26	48720.26	48137.37	1.67	ppb
Manganese	55-2	4863.87	5074.30	5070.00	5002.72	2.40	ppb
Molybdenum	94-1	4886.72	4986.54	5031.25	4968.17	1.49	ppb
Molybdenum	95-1	4873.32	5002.91	5001.45	4959.23	1.50	ppb
Molybdenum	96-1	4856.42	4950.70	5031.19	4946.10	1.77	ppb
Molybdenum	97-1	4847.08	4928.68	5080.72	4952.16	2.39	ppb
Molybdenum	98-1	4822.67	4923.66	5028.27	4924.87	2.09	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	472.58	488.21	487.37	482.72	1.82	ppb
Phosphorus	31-2	9739.43	10135.89	10110.03	9995.12	2.22	ppb
Potassium	39-2	23435.08	24361.45	24381.16	24059.23	2.25	ppb
Rhodium	103-1				94		%
Rhodium	103-2				100		%
Scandium	45-1				93		%
Scandium	45-2				101		%
Selenium	82-1	493.76	504.77	513.59	504.04	1.97	ppb
Selenium	77-2	498.80	503.38	493.90	498.70	0.95	ppb
Selenium	78-2	485.56	499.66	494.27	493.17	1.44	ppb
Silicon	28-1	396.28	399.87	404.30	400.15	1.00	ppb
Silver	107-1	511.31	504.85	515.06	510.41	1.01	ppb
Silver	109-1	517.60	499.50	515.02	510.71	1.92	ppb
Sodium	23-2	47867.04	49344.06	49471.31	48894.14	1.82	ppb
Strontium	86-1	12226.54	12415.74	12566.30	12402.86	1.37	ppb
Strontium	88-1	12197.33	12347.48	12628.90	12391.24	1.77	ppb
Sulfur	34-1	9920.53	10129.93	10121.73	10057.40	1.18	ppb
Terbium	159-1				103		%
Terbium	159-2				109		%
Thallium	203-1	487.54	503.78	503.71	498.34	1.88	ppb
Thallium	205-1	487.72	489.40	501.68	492.93	1.55	ppb
Tin	118-1	504.85	498.95	505.85	503.22	0.74	ppb
Titanium	47-1	4955.86	5073.77	5022.20	5017.28	1.18	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

S06

Instrumnet Name :

P7

Client Sample ID :

S06

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:00:04

DataFile Name :

010CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	473.56	490.34	495.34	486.41	2.35	ppb
Vanadium	51-2	484.26	509.94	504.33	499.51	2.70	ppb
Yttrium	89-1				98		%
Yttrium	89-2				102		%
Zinc	66-2	4903.68	5070.22	5072.18	5015.36	1.93	ppb
Zirconium	90-1	485.42	494.71	500.95	493.69	1.58	ppb
Zirconium	91-1	483.12	493.11	504.64	493.62	2.18	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P7
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:02:47 DataFile Name : 011CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	18988.66	19356.07	19066.66	19137.13	1.01	ppb
Antimony	121-1	992.43	988.59	1011.42	997.48	1.23	ppb
Arsenic	75-2	984.60	1016.47	999.11	1000.06	1.60	ppb
Barium	135-1	4970.64	4911.89	5071.74	4984.76	1.62	ppb
Barium	137-1	4985.66	4964.14	5017.29	4989.03	0.54	ppb
Beryllium	9-1	987.59	970.22	1014.76	990.86	2.27	ppb
Bismuth	209-1				100		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	996.27	977.20	997.48	990.32	1.15	ppb
Cadmium	106-1	988.97	986.16	1001.33	992.15	0.81	ppb
Cadmium	111-1	1004.40	1006.56	1015.03	1008.66	0.56	ppb
Calcium	43-1	112516.68	112785.20	113104.09	112801.99	0.26	ppb
Calcium	44-1	104323.81	104518.10	105897.64	104913.18	0.82	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	993.99	1003.76	995.54	997.76	0.53	ppb
Cobalt	59-2	989.68	998.52	1001.39	996.53	0.61	ppb
Copper	63-2	9883.94	10030.74	9982.97	9965.88	0.75	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				107		%
Indium	115-1				94		%
Indium	115-2				95		%
Iron	56-2	50985.05	51664.83	51745.28	51465.05	0.81	ppb
Iron	57-2	51193.55	51168.06	51665.19	51342.26	0.55	ppb
Iron	54-2	52213.10	52577.43	53044.89	52611.81	0.79	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P7
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:02:47 DataFile Name : 011CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	4980.62	4951.18	5126.79	5019.53	1.87	ppb
Lead	207-1	4997.80	4931.98	5136.67	5022.15	2.08	ppb
Lead	208-1	5003.43	4949.66	5110.08	5021.06	1.63	ppb
Lithium	6-1				100		%
Magnesium	24-2	97407.06	98492.85	98138.14	98012.69	0.56	ppb
Manganese	55-2	9890.65	10043.97	10021.26	9985.29	0.83	ppb
Molybdenum	94-1	9912.70	9997.53	10147.58	10019.27	1.19	ppb
Molybdenum	95-1	9969.68	9944.84	10135.79	10016.77	1.04	ppb
Molybdenum	96-1	10093.38	9918.40	10072.38	10028.05	0.95	ppb
Molybdenum	97-1	10029.76	9951.67	10097.91	10026.44	0.73	ppb
Molybdenum	98-1	10035.22	9974.15	10117.12	10042.16	0.71	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1008.08	1013.08	1002.68	1007.95	0.52	ppb
Phosphorus	31-2	19779.83	20099.22	19983.14	19954.06	0.81	ppb
Potassium	39-2	48820.31	48874.15	49166.29	48953.58	0.38	ppb
Rhodium	103-1				91		%
Rhodium	103-2				97		%
Scandium	45-1				93		%
Scandium	45-2				100		%
Selenium	82-1	991.36	990.46	1007.11	996.31	0.94	ppb
Selenium	77-2	974.86	1029.06	990.09	998.00	2.80	ppb
Selenium	78-2	987.06	1006.90	1002.67	998.88	1.05	ppb
Silicon	28-1	1068.48	1056.77	1069.67	1064.97	0.67	ppb
Silver	107-1	984.01	986.60	996.88	989.16	0.69	ppb
Silver	109-1	985.18	981.35	1001.22	989.25	1.07	ppb
Sodium	23-2	98777.62	100753.42	99530.20	99687.08	1.00	ppb
Strontium	86-1	24902.36	24936.15	25274.34	25037.62	0.82	ppb
Strontium	88-1	24743.67	24988.52	25420.75	25050.98	1.37	ppb
Sulfur	34-1	19943.41	19987.04	20219.86	20050.10	0.74	ppb
Terbium	159-1				104		%
Terbium	159-2				107		%
Thallium	203-1	996.77	994.46	1022.72	1004.65	1.56	ppb
Thallium	205-1	994.32	993.69	1029.91	1005.97	2.06	ppb
Tin	118-1	1001.68	983.41	1005.73	996.94	1.19	ppb
Titanium	47-1	9961.82	9914.34	10070.03	9982.06	0.80	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P7
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:02:47 DataFile Name : 011CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	986.79	1004.78	1041.35	1010.98	2.75	ppb
Vanadium	51-2	990.52	1011.61	991.57	997.90	1.19	ppb
Yttrium	89-1				96		%
Yttrium	89-2				100		%
Zinc	66-2	9953.74	9956.25	9990.33	9966.78	0.21	ppb
Zirconium	90-1	989.25	1002.29	1016.85	1002.80	1.38	ppb
Zirconium	91-1	996.10	1001.10	1015.22	1004.14	0.99	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S08 Instrumnet Name : P7
Client Sample ID : S08 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:05:33 DataFile Name : 012CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	99817.78	100679.02	100157.56	100218.12	0.43	ppb
Antimony	121-1	0.66	0.66	0.70	0.67	3.71	ppb
Arsenic	75-2	0.62	0.49	0.62	0.58	13.44	ppb
Barium	135-1	1.80	1.92	1.83	1.85	3.43	ppb
Barium	137-1	1.91	1.95	1.87	1.91	2.21	ppb
Beryllium	9-1	0.10	0.09	0.10	0.10	3.76	ppb
Bismuth	209-1				88		%
Bismuth	209-2				90		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.26	0.46	0.50	0.41	32.36	ppb
Cadmium	106-1	-0.17	0.01	-0.68	-0.28		ppb
Cadmium	111-1	0.15	0.12	0.09	0.12	29.14	ppb
Calcium	43-1	488230.75	500624.37	500361.58	496405.57	1.43	ppb
Calcium	44-1	490206.20	501457.76	504212.51	498625.49	1.49	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3.03	3.04	3.03	3.04	0.14	ppb
Cobalt	59-2	2.57	2.59	2.63	2.59	1.11	ppb
Copper	63-2	1.79	1.84	1.83	1.82	1.37	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				97		%
Holmium	165-2				102		%
Indium	115-1				91		%
Indium	115-2				95		%
Iron	56-2	247642.10	250866.43	250249.66	249586.06	0.69	ppb
Iron	57-2	247744.59	251027.32	249991.09	249587.67	0.67	ppb
Iron	54-2	249136.61	249393.81	249311.27	249280.56	0.05	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S08 Instrumnet Name : P7
Client Sample ID : S08 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:05:33 DataFile Name : 012CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.85	0.82	0.84	0.84	1.75	ppb
Lead	207-1	0.81	0.86	0.86	0.84	3.72	ppb
Lead	208-1	0.83	0.86	0.86	0.85	1.77	ppb
Lithium	6-1				96		%
Magnesium	24-2	496081.60	503156.81	502566.34	500601.58	0.78	ppb
Manganese	55-2	4.85	4.47	4.46	4.59	4.85	ppb
Molybdenum	94-1	2.32	2.28	2.25	2.28	1.62	ppb
Molybdenum	95-1	1.44	1.42	1.52	1.46	3.90	ppb
Molybdenum	96-1	3.29	3.39	3.53	3.40	3.49	ppb
Molybdenum	97-1	1.45	1.53	1.46	1.48	2.80	ppb
Molybdenum	98-1	1.36	1.33	1.29	1.33	2.32	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	3.35	3.43	3.49	3.42	2.00	ppb
Phosphorus	31-2	-14.02	-12.07	-14.77	-13.62		ppb
Potassium	39-2	247747.07	252430.51	250753.29	250310.29	0.95	ppb
Rhodium	103-1				83		%
Rhodium	103-2				89		%
Scandium	45-1				93		%
Scandium	45-2				102		%
Selenium	82-1	1.24	0.95	1.07	1.09	13.12	ppb
Selenium	77-2	0.48	0.48	0.48	0.48	1.01	ppb
Selenium	78-2	-0.36	0.29	-0.78	-0.28		ppb
Silicon	28-1	4.99	4.71	4.37	4.69	6.62	ppb
Silver	107-1	0.15	0.13	0.11	0.13	16.25	ppb
Silver	109-1	0.14	0.13	0.12	0.13	11.27	ppb
Sodium	23-2	499114.30	498184.43	503245.91	500181.55	0.54	ppb
Strontium	86-1	4.03	3.99	4.08	4.03	1.05	ppb
Strontium	88-1	4.12	4.13	4.02	4.09	1.46	ppb
Sulfur	34-1	-719.67	-604.12	-575.75	-633.18		ppb
Terbium	159-1				99		%
Terbium	159-2				102		%
Thallium	203-1	0.12	0.10	0.11	0.11	7.41	ppb
Thallium	205-1	0.11	0.12	0.11	0.12	6.26	ppb
Tin	118-1	0.35	0.35	0.39	0.37	6.45	ppb
Titanium	47-1	1.12	1.29	1.31	1.24	8.34	ppb

LB Number :LB133265

Operator :Jaswal

Lab Sample ID :S08

Instrumnet Name :P7

Client Sample ID :S08

Dilution Factor :1

Date & Time Acquired :2024-11-02 18:05:33

DataFile Name :012CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.06	0.05	0.06	0.06	4.76	ppb
Vanadium	51-2	0.22	0.22	0.20	0.21	5.60	ppb
Yttrium	89-1				93		%
Yttrium	89-2				99		%
Zinc	66-2	3.32	3.56	3.32	3.40	4.11	ppb
Zirconium	90-1	0.54	0.54	0.55	0.54	0.74	ppb
Zirconium	91-1	0.53	0.58	0.59	0.57	5.11	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICV001 Instrumnet Name : P7
Client Sample ID : ICV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:09:41 DataFile Name : 013ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	494.18	498.14	502.09	498.14	0.79	ppb
Antimony	121-1	203.30	204.88	212.47	206.88	2.37	ppb
Arsenic	75-2	207.59	208.95	205.98	207.51	0.72	ppb
Barium	135-1	92.28	92.58	95.47	93.44	1.88	ppb
Barium	137-1	92.96	93.87	95.64	94.16	1.45	ppb
Beryllium	9-1	103.07	102.17	103.03	102.76	0.49	ppb
Bismuth	209-1				102		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	92.42	92.78	96.35	93.85	2.32	ppb
Cadmium	106-1	88.22	90.54	91.53	90.09	1.88	ppb
Cadmium	111-1	99.51	101.29	103.39	101.40	1.92	ppb
Calcium	43-1	2113.01	2123.18	2130.42	2122.20	0.41	ppb
Calcium	44-1	2131.78	2129.11	2142.32	2134.41	0.33	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	103.57	103.44	102.61	103.21	0.50	ppb
Cobalt	59-2	101.17	101.51	101.33	101.34	0.17	ppb
Copper	63-2	94.88	94.97	95.02	94.95	0.08	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				106		%
Indium	115-1				100		%
Indium	115-2				104		%
Iron	56-2	2070.93	2083.72	2087.98	2080.88	0.43	ppb
Iron	57-2	2134.89	2136.22	2148.61	2139.91	0.35	ppb
Iron	54-2	2246.85	2233.91	2251.15	2243.97	0.40	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICV001 Instrumnet Name : P7
Client Sample ID : ICV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:09:41 DataFile Name : 013ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	216.00	214.94	212.68	214.54	0.79	ppb
Lead	207-1	174.72	177.82	178.05	176.86	1.05	ppb
Lead	208-1	203.85	204.81	204.31	204.32	0.23	ppb
Lithium	6-1				105		%
Magnesium	24-2	1269.40	1282.57	1280.67	1277.54	0.56	ppb
Manganese	55-2	100.57	102.33	100.66	101.19	0.98	ppb
Molybdenum	94-1	0.35	0.42	0.37	0.38	8.64	ppb
Molybdenum	95-1	0.30	0.29	0.28	0.29	3.41	ppb
Molybdenum	96-1	0.29	0.32	0.30	0.30	4.22	ppb
Molybdenum	97-1	0.29	0.28	0.28	0.28	1.95	ppb
Molybdenum	98-1	0.25	0.28	0.30	0.27	9.01	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	102.64	102.26	102.19	102.36	0.23	ppb
Phosphorus	31-2	-15.13	-13.88	-12.34	-13.78		ppb
Potassium	39-2	2063.66	2079.10	2072.22	2071.66	0.37	ppb
Rhodium	103-1				98		%
Rhodium	103-2				105		%
Scandium	45-1				97		%
Scandium	45-2				106		%
Selenium	82-1	213.29	218.18	217.84	216.44	1.26	ppb
Selenium	77-2	220.64	211.90	214.25	215.60	2.10	ppb
Selenium	78-2	213.97	221.21	218.99	218.05	1.70	ppb
Silicon	28-1	0.59	0.78	0.94	0.77	22.92	ppb
Silver	107-1	47.85	49.19	50.43	49.16	2.62	ppb
Silver	109-1	48.14	49.12	50.32	49.19	2.22	ppb
Sodium	23-2	2053.87	2056.83	2062.07	2057.59	0.20	ppb
Strontium	86-1	0.82	0.85	0.76	0.81	5.49	ppb
Strontium	88-1	0.80	0.82	0.80	0.81	1.33	ppb
Sulfur	34-1	-476.77	-515.05	-394.53	-462.12		ppb
Terbium	159-1				102		%
Terbium	159-2				107		%
Thallium	203-1	215.62	218.89	215.21	216.57	0.93	ppb
Thallium	205-1	211.83	215.19	214.35	213.79	0.82	ppb
Tin	118-1	0.07	0.07	0.12	0.09	35.06	ppb
Titanium	47-1	0.27	0.27	0.35	0.30	14.00	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICV001 Instrumnet Name : P7
Client Sample ID : ICV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:09:41 DataFile Name : 013ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.02	0.02	0.02	9.38	ppb
Vanadium	51-2	100.18	100.36	100.75	100.43	0.29	ppb
Yttrium	89-1				98		%
Yttrium	89-2				104		%
Zinc	66-2	196.09	196.23	195.77	196.03	0.12	ppb
Zirconium	90-1	0.08	0.08	0.09	0.08	3.56	ppb
Zirconium	91-1	0.13	0.11	0.15	0.13	15.36	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICB001 Instrumnet Name : P7
Client Sample ID : ICB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:12:57 DataFile Name : 014CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1.50	1.85	1.39	1.58	15.10	ppb
Antimony	121-1	0.13	0.15	0.16	0.15	11.09	ppb
Arsenic	75-2	0.04	0.03	0.03	0.03	26.73	ppb
Barium	135-1	0.07	0.07	0.11	0.08	28.64	ppb
Barium	137-1	0.07	0.07	0.08	0.07	7.37	ppb
Beryllium	9-1	0.05	0.01	0.04	0.03	60.83	ppb
Bismuth	209-1				100		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.05	0.01	0.07	0.04	66.03	ppb
Cadmium	106-1	-0.40	-0.19	-0.75	-0.45		ppb
Cadmium	111-1	0.01	0.01	-0.03	0.00		ppb
Calcium	43-1	6.70	5.09	8.57	6.78	25.69	ppb
Calcium	44-1	5.62	6.29	6.74	6.22	9.03	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.01	0.04	0.02	0.02	74.66	ppb
Cobalt	59-2	0.01	0.01	0.01	0.01	34.00	ppb
Copper	63-2	0.13	0.19	0.17	0.16	18.92	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				100		%
Indium	115-1				102		%
Indium	115-2				100		%
Iron	56-2	3.91	4.94	4.16	4.34	12.46	ppb
Iron	57-2	3.75	4.48	5.88	4.70	23.04	ppb
Iron	54-2	3.33	5.94	5.29	4.86	28.00	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICB001 Instrumnet Name : P7
Client Sample ID : ICB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:12:57 DataFile Name : 014CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.10	0.10	0.11	0.10	7.82	ppb
Lead	207-1	0.12	0.12	0.13	0.12	5.31	ppb
Lead	208-1	0.11	0.11	0.12	0.11	4.43	ppb
Lithium	6-1				102		%
Magnesium	24-2	8.37	9.76	8.78	8.97	7.93	ppb
Manganese	55-2	0.14	0.16	0.14	0.15	6.08	ppb
Molybdenum	94-1	0.14	0.14	0.18	0.15	15.72	ppb
Molybdenum	95-1	0.14	0.15	0.17	0.16	7.65	ppb
Molybdenum	96-1	0.14	0.16	0.16	0.15	8.28	ppb
Molybdenum	97-1	0.16	0.17	0.16	0.16	3.76	ppb
Molybdenum	98-1	0.14	0.14	0.15	0.15	3.98	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.00	0.03	-0.01	0.01	361.45	ppb
Phosphorus	31-2	-19.13	-12.29	-15.56	-15.66		ppb
Potassium	39-2	5.01	18.83	11.98	11.94	57.88	ppb
Rhodium	103-1				102		%
Rhodium	103-2				99		%
Scandium	45-1				102		%
Scandium	45-2				102		%
Selenium	82-1	0.10	0.13	-0.41	-0.06		ppb
Selenium	77-2	-0.22	-0.22	0.01	-0.14		ppb
Selenium	78-2	-0.12	0.20	0.58	0.22	161.33	ppb
Silicon	28-1	-0.01	0.29	0.81	0.36	114.03	ppb
Silver	107-1	0.04	0.04	0.03	0.04	11.13	ppb
Silver	109-1	0.04	0.04	0.04	0.04	2.49	ppb
Sodium	23-2	36.63	43.88	38.97	39.83	9.29	ppb
Strontium	86-1	0.36	0.37	0.33	0.35	6.70	ppb
Strontium	88-1	0.33	0.34	0.35	0.34	2.58	ppb
Sulfur	34-1	-592.32	-528.29	-498.02	-539.54		ppb
Terbium	159-1				101		%
Terbium	159-2				100		%
Thallium	203-1	0.04	0.05	0.04	0.04	13.05	ppb
Thallium	205-1	0.05	0.05	0.04	0.05	6.76	ppb
Tin	118-1	-0.01	0.01	0.00	0.00	1054.20	ppb
Titanium	47-1	0.09	0.12	0.16	0.12	24.98	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICB001 Instrumnet Name : P7
Client Sample ID : ICB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:12:57 DataFile Name : 014CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.01	0.01	3.45	ppb
Vanadium	51-2	0.02	0.03	0.02	0.02	18.96	ppb
Yttrium	89-1				101		%
Yttrium	89-2				99		%
Zinc	66-2	0.19	0.25	0.24	0.23	15.37	ppb
Zirconium	90-1	0.02	0.02	0.02	0.02	7.63	ppb
Zirconium	91-1	0.01	0.03	0.02	0.02	62.09	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICSA001 Instrumnet Name : P7
Client Sample ID : ICSA001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:16:13 DataFile Name : 015ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	89268.82	88728.41	88724.21	88907.15	0.35	ppb
Antimony	121-1	1.08	1.10	1.16	1.11	3.34	ppb
Arsenic	75-2	0.37	0.48	0.42	0.42	13.30	ppb
Barium	135-1	1.32	1.32	1.33	1.32	0.56	ppb
Barium	137-1	1.31	1.32	1.28	1.30	1.81	ppb
Beryllium	9-1	0.31	0.26	0.28	0.28	10.39	ppb
Bismuth	209-1				98		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	6.82	7.25	6.64	6.90	4.57	ppb
Cadmium	106-1	-1.32	-1.12	-0.95	-1.13		ppb
Cadmium	111-1	0.22	0.24	0.26	0.24	8.22	ppb
Calcium	43-1	93386.67	95452.84	97634.06	95491.19	2.22	ppb
Calcium	44-1	92005.63	94891.42	96299.97	94399.01	2.32	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	19.88	20.12	20.01	20.00	0.61	ppb
Cobalt	59-2	1.22	1.18	1.21	1.20	1.94	ppb
Copper	63-2	7.07	7.08	6.83	6.99	2.01	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				108		%
Indium	115-1				98		%
Indium	115-2				102		%
Iron	56-2	99527.44	100352.61	100854.78	100244.94	0.67	ppb
Iron	57-2	99799.80	99218.49	100123.74	99714.01	0.46	ppb
Iron	54-2	101153.42	101450.60	101128.61	101244.21	0.18	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICSA001 Instrumnet Name : P7
Client Sample ID : ICSA001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:16:13 DataFile Name : 015ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	3.87	4.04	4.08	3.99	2.78	ppb
Lead	207-1	3.52	3.67	3.67	3.62	2.41	ppb
Lead	208-1	3.68	3.82	3.80	3.77	1.91	ppb
Lithium	6-1				101		%
Magnesium	24-2	96685.30	95095.38	96360.84	96047.17	0.87	ppb
Manganese	55-2	7.84	7.67	7.82	7.78	1.19	ppb
Molybdenum	94-1	1410.78	1456.74	1445.98	1437.83	1.67	ppb
Molybdenum	95-1	1716.91	1750.62	1753.43	1740.32	1.17	ppb
Molybdenum	96-1	1671.94	1702.86	1717.02	1697.27	1.36	ppb
Molybdenum	97-1	1657.92	1719.47	1761.73	1713.04	3.05	ppb
Molybdenum	98-1	1668.44	1736.03	1733.47	1712.65	2.24	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	5.10	5.23	5.09	5.14	1.57	ppb
Phosphorus	31-2	97570.80	95266.11	96356.97	96397.96	1.20	ppb
Potassium	39-2	95846.88	95276.94	95442.41	95522.08	0.31	ppb
Rhodium	103-1				93		%
Rhodium	103-2				99		%
Scandium	45-1				98		%
Scandium	45-2				104		%
Selenium	82-1	-0.03	-0.41	-0.20	-0.21		ppb
Selenium	77-2	0.00	0.23	0.00	0.08	171.78	ppb
Selenium	78-2	0.02	-0.82	-0.22	-0.34		ppb
Silicon	28-1	5.30	5.89	6.24	5.81	8.15	ppb
Silver	107-1	0.10	0.09	0.09	0.09	7.83	ppb
Silver	109-1	0.11	0.09	0.07	0.09	20.28	ppb
Sodium	23-2	97260.15	97566.87	97623.69	97483.57	0.20	ppb
Strontium	86-1	32.66	33.33	33.42	33.13	1.25	ppb
Strontium	88-1	32.26	33.52	33.32	33.03	2.05	ppb
Sulfur	34-1	93105.21	96339.12	97234.35	95559.56	2.27	ppb
Terbium	159-1				104		%
Terbium	159-2				108		%
Thallium	203-1	0.11	0.11	0.10	0.11	4.43	ppb
Thallium	205-1	0.11	0.10	0.10	0.11	5.00	ppb
Tin	118-1	0.12	0.12	0.13	0.12	2.31	ppb
Titanium	47-1	1739.07	1775.53	1818.76	1777.79	2.24	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

ICSA001

Instrumnet Name :

P7

Client Sample ID :

ICSA001

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:16:13

DataFile Name :

015ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.02	0.02	0.02	17.25	ppb
Vanadium	51-2	0.19	0.20	0.21	0.20	5.31	ppb
Yttrium	89-1				99		%
Yttrium	89-2				104		%
Zinc	66-2	10.97	10.92	10.62	10.84	1.74	ppb
Zirconium	90-1	0.04	0.04	0.04	0.04	7.86	ppb
Zirconium	91-1	0.04	0.03	0.05	0.04	23.94	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICSAB001 Instrumnet Name : P7
Client Sample ID : ICSAB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:19:16 DataFile Name : 016ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	86269.03	87737.38	85551.16	86519.19	1.29	ppb
Antimony	121-1	19.20	19.66	21.42	20.09	5.83	ppb
Arsenic	75-2	20.31	19.50	19.61	19.81	2.22	ppb
Barium	135-1	18.93	19.25	20.34	19.51	3.80	ppb
Barium	137-1	18.88	19.37	20.68	19.64	4.74	ppb
Beryllium	9-1	20.18	20.77	22.58	21.18	5.91	ppb
Bismuth	209-1				98		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	23.35	24.52	26.18	24.68	5.77	ppb
Cadmium	106-1	14.61	15.29	16.97	15.62	7.76	ppb
Cadmium	111-1	18.12	18.28	20.09	18.83	5.80	ppb
Calcium	43-1	93080.68	92719.19	105248.41	97016.09	7.35	ppb
Calcium	44-1	93020.15	92728.40	105404.15	97050.90	7.46	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	38.97	39.06	39.02	39.02	0.11	ppb
Cobalt	59-2	19.99	20.00	20.13	20.04	0.40	ppb
Copper	63-2	24.28	24.13	24.10	24.17	0.39	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				109		%
Indium	115-1				97		%
Indium	115-2				103		%
Iron	56-2	97622.03	99283.49	99278.49	98728.01	0.97	ppb
Iron	57-2	97904.05	98401.87	96930.25	97745.39	0.77	ppb
Iron	54-2	98641.19	98197.37	98710.35	98516.30	0.28	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICSAB001 Instrumnet Name : P7
Client Sample ID : ICSAB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:19:16 DataFile Name : 016ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	23.19	22.98	25.85	24.01	6.68	ppb
Lead	207-1	21.88	21.52	24.74	22.72	7.76	ppb
Lead	208-1	22.27	22.02	25.08	23.12	7.34	ppb
Lithium	6-1				100		%
Magnesium	24-2	93942.33	95103.18	93119.48	94055.00	1.06	ppb
Manganese	55-2	26.39	26.57	26.59	26.52	0.41	ppb
Molybdenum	94-1	1428.51	1422.47	1587.24	1479.41	6.32	ppb
Molybdenum	95-1	1697.27	1710.15	1906.62	1771.35	6.62	ppb
Molybdenum	96-1	1665.09	1682.69	1877.72	1741.83	6.78	ppb
Molybdenum	97-1	1709.44	1699.34	1915.98	1774.92	6.89	ppb
Molybdenum	98-1	1691.73	1681.85	1911.29	1761.63	7.36	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	23.35	22.90	23.21	23.15	1.00	ppb
Phosphorus	31-2	94324.84	92778.66	93500.86	93534.79	0.83	ppb
Potassium	39-2	93033.11	92562.48	93015.52	92870.37	0.29	ppb
Rhodium	103-1				92		%
Rhodium	103-2				100		%
Scandium	45-1				96		%
Scandium	45-2				106		%
Selenium	82-1	19.24	19.43	21.97	20.21	7.55	ppb
Selenium	77-2	18.65	19.34	19.19	19.06	1.90	ppb
Selenium	78-2	18.26	17.87	19.12	18.42	3.48	ppb
Silicon	28-1	6.01	6.02	8.66	6.90	22.16	ppb
Silver	107-1	17.19	17.53	19.25	17.99	6.15	ppb
Silver	109-1	17.18	17.57	19.31	18.02	6.30	ppb
Sodium	23-2	95573.92	96351.77	93994.63	95306.77	1.26	ppb
Strontium	86-1	32.74	32.36	37.00	34.03	7.57	ppb
Strontium	88-1	32.53	32.53	36.71	33.92	7.12	ppb
Sulfur	34-1	95432.85	95052.30	107082.77	99189.31	6.89	ppb
Terbium	159-1				104		%
Terbium	159-2				109		%
Thallium	203-1	18.31	17.99	20.05	18.78	5.91	ppb
Thallium	205-1	18.12	17.95	20.19	18.75	6.65	ppb
Tin	118-1	0.08	0.11	0.15	0.11	29.52	ppb
Titanium	47-1	1781.70	1763.31	1993.37	1846.13	6.93	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICSAB001 Instrumnet Name : P7
Client Sample ID : ICSAB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:19:16 DataFile Name : 016ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.02	0.02	0.02	16.55	ppb
Vanadium	51-2	19.45	19.42	19.60	19.49	0.49	ppb
Yttrium	89-1				97		%
Yttrium	89-2				104		%
Zinc	66-2	29.72	29.14	28.67	29.17	1.81	ppb
Zirconium	90-1	0.03	0.02	0.04	0.03	23.86	ppb
Zirconium	91-1	0.03	0.03	0.03	0.03	8.46	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV001 Instrumnet Name : P7
Client Sample ID : CCV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:22:18 DataFile Name : 017CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	47626.80	47842.71	47374.03	47614.51	0.49	ppb
Antimony	121-1	488.24	485.02	494.08	489.11	0.94	ppb
Arsenic	75-2	512.34	511.40	511.82	511.85	0.09	ppb
Barium	135-1	2453.13	2480.27	2514.53	2482.64	1.24	ppb
Barium	137-1	2459.02	2455.48	2516.59	2477.03	1.38	ppb
Beryllium	9-1	522.51	523.36	501.58	515.82	2.39	ppb
Bismuth	209-1				98		%
Bismuth	209-2				98		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	507.27	498.74	506.94	504.32	0.96	ppb
Cadmium	106-1	507.15	504.26	513.21	508.21	0.90	ppb
Cadmium	111-1	473.69	469.63	482.93	475.42	1.43	ppb
Calcium	43-1	242516.73	245293.13	245288.23	244366.03	0.66	ppb
Calcium	44-1	239735.25	241605.54	241542.34	240961.05	0.44	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	491.98	492.21	483.16	489.12	1.05	ppb
Cobalt	59-2	498.98	495.60	495.65	496.74	0.39	ppb
Copper	63-2	4795.81	4828.29	4778.61	4800.90	0.53	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				106		%
Holmium	165-2				108		%
Indium	115-1				98		%
Indium	115-2				100		%
Iron	56-2	121622.52	123442.44	122165.12	122410.02	0.76	ppb
Iron	57-2	122525.76	122884.41	121095.56	122168.58	0.77	ppb
Iron	54-2	121990.03	123877.44	123792.22	123219.90	0.87	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV001 Instrumnet Name : P7
Client Sample ID : CCV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:22:18 DataFile Name : 017CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2516.70	2491.92	2459.29	2489.30	1.16	ppb
Lead	207-1	2503.90	2502.81	2457.49	2488.07	1.06	ppb
Lead	208-1	2500.62	2475.57	2460.47	2478.88	0.82	ppb
Lithium	6-1				99		%
Magnesium	24-2	250504.16	252960.34	251729.43	251731.31	0.49	ppb
Manganese	55-2	4745.78	4841.29	4801.82	4796.29	1.00	ppb
Molybdenum	94-1	4933.95	4884.09	4989.50	4935.85	1.07	ppb
Molybdenum	95-1	4896.72	4837.57	4981.03	4905.10	1.47	ppb
Molybdenum	96-1	4876.49	4854.12	5000.86	4910.49	1.61	ppb
Molybdenum	97-1	4894.92	4866.91	4913.19	4891.67	0.48	ppb
Molybdenum	98-1	4841.07	4838.60	4913.66	4864.44	0.88	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	463.83	466.42	462.19	464.15	0.46	ppb
Phosphorus	31-2	9851.66	9873.36	9790.28	9838.43	0.44	ppb
Potassium	39-2	119005.37	120613.89	120032.59	119883.95	0.68	ppb
Rhodium	103-1				93		%
Rhodium	103-2				98		%
Scandium	45-1				101		%
Scandium	45-2				108		%
Selenium	82-1	493.78	498.60	500.85	497.74	0.73	ppb
Selenium	77-2	491.95	500.39	487.05	493.13	1.37	ppb
Selenium	78-2	502.47	508.82	499.63	503.64	0.93	ppb
Silicon	28-1	2263.66	2267.85	2295.19	2275.57	0.75	ppb
Silver	107-1	505.84	498.99	505.11	503.31	0.75	ppb
Silver	109-1	500.09	493.28	510.37	501.25	1.72	ppb
Sodium	23-2	238052.76	242298.26	238725.34	239692.12	0.95	ppb
Strontium	86-1	12226.83	12131.12	12212.59	12190.18	0.42	ppb
Strontium	88-1	12187.45	12110.96	12305.45	12201.29	0.80	ppb
Sulfur	34-1	9148.68	9056.50	9227.13	9144.10	0.93	ppb
Terbium	159-1				106		%
Terbium	159-2				108		%
Thallium	203-1	505.09	503.09	494.38	500.85	1.14	ppb
Thallium	205-1	500.56	496.56	494.56	497.23	0.61	ppb
Tin	118-1	485.18	478.71	479.12	481.01	0.75	ppb
Titanium	47-1	4887.30	4930.80	4947.78	4921.96	0.63	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV001 Instrumnet Name : P7
Client Sample ID : CCV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:22:18 DataFile Name : 017CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	500.42	498.03	511.22	503.23	1.40	ppb
Vanadium	51-2	495.97	498.49	489.78	494.75	0.91	ppb
Yttrium	89-1				102		%
Yttrium	89-2				105		%
Zinc	66-2	4806.56	4876.81	4836.35	4839.91	0.73	ppb
Zirconium	90-1	501.94	499.86	503.72	501.84	0.38	ppb
Zirconium	91-1	503.22	500.15	508.74	504.04	0.86	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB001 Instrumnet Name : P7
Client Sample ID : CCB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:28:13 DataFile Name : 019CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1.41	1.62	1.32	1.45	10.63	ppb
Antimony	121-1	0.22	0.25	0.34	0.27	21.29	ppb
Arsenic	75-2	0.00	0.04	0.00	0.01	187.50	ppb
Barium	135-1	0.09	0.06	0.05	0.07	27.54	ppb
Barium	137-1	0.08	0.07	0.04	0.07	32.71	ppb
Beryllium	9-1	0.05	0.04	0.02	0.03	45.44	ppb
Bismuth	209-1				106		%
Bismuth	209-2				107		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.03	0.03	0.08	0.05	69.70	ppb
Cadmium	106-1	-0.53	-0.38	-0.84	-0.58		ppb
Cadmium	111-1	0.00	0.01	-0.04	-0.01		ppb
Calcium	43-1	5.56	92.29	3.56	33.80	149.85	ppb
Calcium	44-1	7.11	94.72	3.67	35.17	146.73	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.00	-0.01	0.01	0.00	380.48	ppb
Cobalt	59-2	0.01	0.00	0.01	0.01	42.76	ppb
Copper	63-2	0.14	0.17	0.15	0.15	8.84	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				105		%
Holmium	165-2				108		%
Indium	115-1				104		%
Indium	115-2				106		%
Iron	56-2	3.40	3.14	3.39	3.31	4.37	ppb
Iron	57-2	1.22	2.49	0.85	1.52	56.34	ppb
Iron	54-2	3.37	2.73	2.50	2.87	15.71	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB001 Instrumnet Name : P7
Client Sample ID : CCB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:28:13 DataFile Name : 019CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.14	0.11	0.12	0.12	11.61	ppb
Lead	207-1	0.15	0.12	0.10	0.12	23.46	ppb
Lead	208-1	0.15	0.11	0.12	0.13	15.09	ppb
Lithium	6-1				106		%
Magnesium	24-2	6.57	6.55	6.29	6.47	2.47	ppb
Manganese	55-2	0.07	0.11	0.10	0.09	21.94	ppb
Molybdenum	94-1	0.34	0.51	0.32	0.39	26.16	ppb
Molybdenum	95-1	0.31	0.31	0.26	0.29	11.16	ppb
Molybdenum	96-1	0.29	0.38	0.28	0.32	18.17	ppb
Molybdenum	97-1	0.28	0.32	0.25	0.28	13.14	ppb
Molybdenum	98-1	0.27	0.33	0.27	0.29	12.75	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.01	0.00	0.01	0.01	91.52	ppb
Phosphorus	31-2	-13.66	-12.81	-17.88	-14.78		ppb
Potassium	39-2	9.52	10.83	9.00	9.78	9.66	ppb
Rhodium	103-1				104		%
Rhodium	103-2				107		%
Scandium	45-1				103		%
Scandium	45-2				107		%
Selenium	82-1	-0.50	-0.34	-0.26	-0.37		ppb
Selenium	77-2	-0.22	-0.22	-0.22	-0.22		ppb
Selenium	78-2	-0.76	1.50	-1.36	-0.21		ppb
Silicon	28-1	0.00	0.18	0.03	0.07	129.50	ppb
Silver	107-1	0.05	0.04	0.04	0.04	16.40	ppb
Silver	109-1	0.05	0.04	0.03	0.04	23.03	ppb
Sodium	23-2	39.23	39.63	37.95	38.94	2.25	ppb
Strontium	86-1	0.43	0.60	0.29	0.44	34.32	ppb
Strontium	88-1	0.40	0.53	0.26	0.39	33.65	ppb
Sulfur	34-1	-619.55	-501.66	-534.13	-551.78		ppb
Terbium	159-1				105		%
Terbium	159-2				108		%
Thallium	203-1	0.05	0.05	0.05	0.05	3.41	ppb
Thallium	205-1	0.05	0.04	0.04	0.04	10.71	ppb
Tin	118-1	0.05	0.06	0.07	0.06	16.51	ppb
Titanium	47-1	0.25	1.64	0.22	0.70	115.41	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

CCB001

Instrumnet Name :

P7

Client Sample ID :

CCB001

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:28:13

DataFile Name :

019CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.01	0.01	30.22	ppb
Vanadium	51-2	0.01	0.01	0.01	0.01	36.31	ppb
Yttrium	89-1				103		%
Yttrium	89-2				105		%
Zinc	66-2	0.01	0.01	0.07	0.03	117.15	ppb
Zirconium	90-1	0.06	0.07	0.06	0.06	10.27	ppb
Zirconium	91-1	0.07	0.08	0.07	0.07	9.48	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BL Instrumnet Name : P7
Client Sample ID : PBS454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:31:30 DataFile Name : 020CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1.02	1.13	0.86	1.01	13.41	ppb
Antimony	121-1	0.06	0.08	0.10	0.08	22.79	ppb
Arsenic	75-2	0.02	0.01	0.01	0.01	74.84	ppb
Barium	135-1	0.04	0.02	0.01	0.02	49.13	ppb
Barium	137-1	0.02	0.03	0.03	0.03	10.13	ppb
Beryllium	9-1	0.01	0.01	-0.02	0.00	7930.83	ppb
Bismuth	209-1				102		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.03	0.03	0.07	0.04	50.43	ppb
Cadmium	106-1	-0.51	-0.46	-1.47	-0.81		ppb
Cadmium	111-1	-0.03	-0.03	-0.09	-0.05		ppb
Calcium	43-1	3.15	1.21	0.31	1.56	93.31	ppb
Calcium	44-1	0.62	0.69	0.56	0.62	11.14	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.00	0.01	0.00	0.00	190.17	ppb
Cobalt	59-2	0.00	0.00	0.00	0.00	1188.86	ppb
Copper	63-2	0.10	0.10	0.09	0.09	5.89	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				105		%
Indium	115-1				101		%
Indium	115-2				104		%
Iron	56-2	1.93	2.01	2.17	2.04	6.03	ppb
Iron	57-2	-0.08	0.87	0.70	0.50	101.11	ppb
Iron	54-2	0.73	1.25	2.17	1.38	52.61	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BL Instrumnet Name : P7
Client Sample ID : PBS454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:31:30 DataFile Name : 020CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.06	0.07	0.04	0.06	27.01	ppb
Lead	207-1	0.06	0.06	0.06	0.06	4.50	ppb
Lead	208-1	0.06	0.06	0.05	0.06	13.10	ppb
Lithium	6-1				102		%
Magnesium	24-2	3.59	3.65	4.65	3.97	15.03	ppb
Manganese	55-2	0.05	0.07	0.05	0.06	24.19	ppb
Molybdenum	94-1	0.12	0.11	0.14	0.12	11.61	ppb
Molybdenum	95-1	0.13	0.12	0.11	0.12	6.83	ppb
Molybdenum	96-1	0.10	0.11	0.10	0.10	8.96	ppb
Molybdenum	97-1	0.11	0.10	0.11	0.11	7.28	ppb
Molybdenum	98-1	0.11	0.10	0.09	0.10	10.83	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.01	-0.01	-0.01	0.00		ppb
Phosphorus	31-2	-17.08	-14.68	-15.12	-15.63		ppb
Potassium	39-2	3.95	4.38	5.91	4.75	21.67	ppb
Rhodium	103-1				100		%
Rhodium	103-2				104		%
Scandium	45-1				100		%
Scandium	45-2				106		%
Selenium	82-1	-0.07	-0.60	-0.52	-0.40		ppb
Selenium	77-2	-0.22	-0.22	-0.22	-0.22		ppb
Selenium	78-2	-0.48	0.45	-0.42	-0.15		ppb
Silicon	28-1	-0.16	-0.05	0.14	-0.02		ppb
Silver	107-1	0.02	0.01	0.01	0.01	22.15	ppb
Silver	109-1	0.02	0.01	0.01	0.02	28.24	ppb
Sodium	23-2	26.04	26.14	26.33	26.17	0.57	ppb
Strontium	86-1	0.17	0.13	0.14	0.15	15.28	ppb
Strontium	88-1	0.17	0.16	0.12	0.15	16.51	ppb
Sulfur	34-1	-463.63	-495.82	-416.95	-458.80		ppb
Terbium	159-1				103		%
Terbium	159-2				105		%
Thallium	203-1	0.03	0.03	0.02	0.03	23.98	ppb
Thallium	205-1	0.03	0.03	0.03	0.03	6.79	ppb
Tin	118-1	-0.01	0.01	0.00	0.00		ppb
Titanium	47-1	0.10	0.10	0.07	0.09	18.58	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

PB164454BL

Instrumnet Name :

P7

Client Sample ID :

PBS454

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:31:30

DataFile Name :

020CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	37.19	ppb
Vanadium	51-2	0.01	0.01	0.01	0.01	20.19	ppb
Yttrium	89-1				101		%
Yttrium	89-2				105		%
Zinc	66-2	0.00	0.01	0.08	0.03	162.89	ppb
Zirconium	90-1	0.02	0.02	0.03	0.03	25.80	ppb
Zirconium	91-1	0.02	0.03	0.04	0.03	19.55	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BS Instrumnet Name : P7
Client Sample ID : LCS454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:34:44 DataFile Name : 021LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	40.08	50.39	37.28	42.58	16.21	ppb
Antimony	121-1	3.83	3.84	3.91	3.86	1.11	ppb
Arsenic	75-2	2.07	1.88	2.15	2.03	6.93	ppb
Barium	135-1	17.73	17.83	18.28	17.95	1.63	ppb
Barium	137-1	17.74	17.99	18.08	17.94	0.99	ppb
Beryllium	9-1	1.99	2.09	2.19	2.09	4.67	ppb
Bismuth	209-1				101		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.86	1.67	2.01	1.85	9.11	ppb
Cadmium	106-1	1.73	1.12	2.04	1.63	28.87	ppb
Cadmium	111-1	1.81	1.77	1.91	1.83	3.84	ppb
Calcium	43-1	955.58	999.44	985.29	980.10	2.28	ppb
Calcium	44-1	973.58	976.61	972.33	974.17	0.23	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	4.05	3.99	4.10	4.05	1.43	ppb
Cobalt	59-2	2.01	2.01	2.00	2.01	0.33	ppb
Copper	63-2	4.05	4.12	3.96	4.04	2.04	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				106		%
Indium	115-1				102		%
Indium	115-2				105		%
Iron	56-2	407.47	405.97	407.23	406.89	0.20	ppb
Iron	57-2	412.31	424.36	418.15	418.27	1.44	ppb
Iron	54-2	427.23	440.96	422.19	430.13	2.26	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BS Instrumnet Name : P7
Client Sample ID : LCS454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:34:44 DataFile Name : 021LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1.77	1.65	1.72	1.71	3.55	ppb
Lead	207-1	1.70	1.73	1.73	1.72	1.09	ppb
Lead	208-1	1.72	1.68	1.70	1.70	1.08	ppb
Lithium	6-1				101		%
Magnesium	24-2	1058.59	1080.90	1049.76	1063.08	1.51	ppb
Manganese	55-2	2.00	2.17	2.03	2.07	4.61	ppb
Molybdenum	94-1	10.39	10.19	10.52	10.37	1.59	ppb
Molybdenum	95-1	8.55	8.55	8.68	8.59	0.83	ppb
Molybdenum	96-1	8.69	8.71	8.81	8.74	0.77	ppb
Molybdenum	97-1	8.57	8.56	8.74	8.62	1.16	ppb
Molybdenum	98-1	8.45	8.47	8.55	8.49	0.65	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.90	1.99	2.00	1.97	2.89	ppb
Phosphorus	31-2	-16.80	-10.52	-14.69	-14.00		ppb
Potassium	39-2	1017.63	1032.37	1014.06	1021.35	0.95	ppb
Rhodium	103-1				101		%
Rhodium	103-2				106		%
Scandium	45-1				101		%
Scandium	45-2				107		%
Selenium	82-1	10.73	10.32	10.31	10.46	2.31	ppb
Selenium	77-2	12.66	14.29	9.03	11.99	22.46	ppb
Selenium	78-2	11.41	9.29	9.54	10.08	11.49	ppb
Silicon	28-1	5.56	5.66	5.64	5.62	0.96	ppb
Silver	107-1	1.88	1.94	1.94	1.92	1.93	ppb
Silver	109-1	1.88	1.94	1.93	1.91	1.77	ppb
Sodium	23-2	1019.06	1030.91	1016.22	1022.06	0.76	ppb
Strontium	86-1	2.09	1.99	2.10	2.06	2.82	ppb
Strontium	88-1	1.97	2.00	2.03	2.00	1.42	ppb
Sulfur	34-1	-443.28	-361.29	-424.10	-409.56		ppb
Terbium	159-1				102		%
Terbium	159-2				107		%
Thallium	203-1	1.66	1.67	1.63	1.65	1.17	ppb
Thallium	205-1	1.62	1.63	1.62	1.62	0.41	ppb
Tin	118-1	0.00	0.00	0.02	0.01	108.87	ppb
Titanium	47-1	8.96	9.14	9.19	9.10	1.36	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BS Instrumnet Name : P7
Client Sample ID : LCS454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:34:44 DataFile Name : 021LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.57	1.55	1.56	1.56	0.80	ppb
Vanadium	51-2	10.01	9.79	9.76	9.85	1.35	ppb
Yttrium	89-1				102		%
Yttrium	89-2				104		%
Zinc	66-2	10.21	9.81	9.61	9.88	3.14	ppb
Zirconium	90-1	1.91	1.91	1.96	1.93	1.48	ppb
Zirconium	91-1	1.95	2.00	1.95	1.97	1.25	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BSD Instrumnet Name : P7
Client Sample ID : LCSD454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:38:00 DataFile Name : 022LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	19.79	19.79	20.16	19.91	1.06	ppb
Antimony	121-1	2.07	1.97	2.01	2.02	2.61	ppb
Arsenic	75-2	1.10	1.22	1.00	1.11	9.66	ppb
Barium	135-1	9.83	8.97	9.08	9.29	5.04	ppb
Barium	137-1	9.65	9.06	9.02	9.24	3.82	ppb
Beryllium	9-1	1.12	1.07	1.10	1.09	2.23	ppb
Bismuth	209-1				99		%
Bismuth	209-2				107		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.17	1.25	1.15	1.19	4.63	ppb
Cadmium	106-1	0.90	0.45	0.72	0.69	32.59	ppb
Cadmium	111-1	1.02	0.91	0.89	0.94	7.34	ppb
Calcium	43-1	519.30	506.58	478.25	501.38	4.19	ppb
Calcium	44-1	518.92	486.37	494.56	499.95	3.39	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2.00	2.01	2.09	2.03	2.37	ppb
Cobalt	59-2	1.02	1.00	1.02	1.01	1.32	ppb
Copper	63-2	2.00	2.06	2.07	2.04	1.88	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				106		%
Indium	115-1				100		%
Indium	115-2				105		%
Iron	56-2	210.55	212.63	213.20	212.13	0.66	ppb
Iron	57-2	211.35	206.01	214.97	210.78	2.14	ppb
Iron	54-2	216.83	218.60	217.46	217.63	0.41	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BSD Instrumnet Name : P7
Client Sample ID : LCSD454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:38:00 DataFile Name : 022LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.92	0.91	0.84	0.89	5.01	ppb
Lead	207-1	0.88	0.89	0.89	0.89	0.44	ppb
Lead	208-1	0.92	0.88	0.88	0.89	2.70	ppb
Lithium	6-1				99		%
Magnesium	24-2	530.86	536.61	540.33	535.93	0.89	ppb
Manganese	55-2	1.06	1.04	1.12	1.07	3.55	ppb
Molybdenum	94-1	5.79	5.26	5.38	5.48	5.07	ppb
Molybdenum	95-1	4.72	4.19	4.54	4.48	5.98	ppb
Molybdenum	96-1	4.80	4.38	4.54	4.57	4.58	ppb
Molybdenum	97-1	4.72	4.37	4.39	4.49	4.36	ppb
Molybdenum	98-1	4.75	4.25	4.43	4.47	5.61	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.04	1.07	1.04	1.05	1.60	ppb
Phosphorus	31-2	-17.78	-15.04	-16.57	-16.46		ppb
Potassium	39-2	506.19	513.67	514.73	511.53	0.91	ppb
Rhodium	103-1				99		%
Rhodium	103-2				106		%
Scandium	45-1				100		%
Scandium	45-2				107		%
Selenium	82-1	5.69	4.59	4.89	5.06	11.20	ppb
Selenium	77-2	4.60	2.65	3.05	3.44	30.07	ppb
Selenium	78-2	4.17	4.84	5.33	4.78	12.16	ppb
Silicon	28-1	3.81	2.74	2.84	3.13	18.80	ppb
Silver	107-1	1.02	0.95	0.98	0.99	3.49	ppb
Silver	109-1	1.05	0.99	1.00	1.01	3.19	ppb
Sodium	23-2	518.26	520.79	523.07	520.70	0.46	ppb
Strontium	86-1	1.25	1.11	1.14	1.17	6.31	ppb
Strontium	88-1	1.16	1.07	1.08	1.10	4.38	ppb
Sulfur	34-1	-160.85	-377.08	-353.16	-297.03		ppb
Terbium	159-1				100		%
Terbium	159-2				107		%
Thallium	203-1	0.89	0.80	0.83	0.84	5.49	ppb
Thallium	205-1	0.90	0.83	0.84	0.85	4.53	ppb
Tin	118-1	0.00	-0.01	-0.01	-0.01		ppb
Titanium	47-1	4.93	4.50	4.76	4.73	4.59	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BSD Instrumnet Name : P7
Client Sample ID : LCSD454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:38:00 DataFile Name : 022LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.83	0.78	0.79	0.80	3.40	ppb
Vanadium	51-2	4.97	4.96	5.05	4.99	0.97	ppb
Yttrium	89-1				100		%
Yttrium	89-2				106		%
Zinc	66-2	5.44	4.76	5.18	5.13	6.62	ppb
Zirconium	90-1	1.07	0.99	0.99	1.02	4.28	ppb
Zirconium	91-1	1.02	0.98	1.07	1.03	4.41	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-01 Instrumnet Name : P7
Client Sample ID : MYE4H0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:41:15 DataFile Name : 023SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	38923.04	39704.85	39250.13	39292.67	1.00	ppb
Antimony	121-1	1.37	1.43	1.50	1.43	4.66	ppb
Arsenic	75-2	79.83	80.88	80.97	80.56	0.79	ppb
Barium	135-1	388.45	399.41	414.45	400.77	3.26	ppb
Barium	137-1	417.87	437.89	448.53	434.76	3.58	ppb
Beryllium	9-1	1.31	1.42	1.49	1.41	6.43	ppb
Bismuth	209-1				98		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.43	2.86	2.87	2.72	9.23	ppb
Cadmium	106-1	3.29	2.45	2.70	2.81	15.42	ppb
Cadmium	111-1	1.60	1.51	1.58	1.57	2.99	ppb
Calcium	43-1	16119.84	17028.41	17399.77	16849.34	3.91	ppb
Calcium	44-1	15141.30	15966.64	16283.83	15797.26	3.73	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	100.84	101.34	101.89	101.36	0.52	ppb
Cobalt	59-2	43.00	42.99	43.18	43.06	0.25	ppb
Copper	63-2	217.85	218.98	219.57	218.80	0.40	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				108		%
Indium	115-1				100		%
Indium	115-2				105		%
Iron	56-2	82715.71	83265.68	84106.70	83362.70	0.84	ppb
Iron	57-2	81728.92	81873.76	82636.99	82079.89	0.59	ppb
Iron	54-2	83787.00	82933.89	83618.58	83446.49	0.54	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-01 Instrumnet Name : P7
Client Sample ID : MYE4H0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:41:15 DataFile Name : 023SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	48.30	51.72	52.94	50.98	4.72	ppb
Lead	207-1	47.47	50.91	51.86	50.08	4.61	ppb
Lead	208-1	47.86	50.96	51.96	50.26	4.25	ppb
Lithium	6-1				99		%
Magnesium	24-2	19663.11	19735.75	19772.67	19723.84	0.28	ppb
Manganese	55-2	2042.01	2049.44	2069.24	2053.56	0.69	ppb
Molybdenum	94-1	17.05	18.04	18.54	17.88	4.22	ppb
Molybdenum	95-1	1.50	1.58	1.67	1.58	5.20	ppb
Molybdenum	96-1	3.79	3.93	4.15	3.96	4.64	ppb
Molybdenum	97-1	1.49	1.57	1.58	1.55	3.32	ppb
Molybdenum	98-1	1.43	1.49	1.58	1.50	4.96	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	80.91	79.70	81.16	80.59	0.97	ppb
Phosphorus	31-2	2791.34	2868.74	2877.42	2845.83	1.67	ppb
Potassium	39-2	12474.45	12538.26	12658.51	12557.08	0.74	ppb
Rhodium	103-1				98		%
Rhodium	103-2				105		%
Scandium	45-1				109		%
Scandium	45-2				117		%
Selenium	82-1	0.90	0.87	0.87	0.88	1.91	ppb
Selenium	77-2	16.53	23.69	16.67	18.96	21.59	ppb
Selenium	78-2	3.52	3.19	3.66	3.45	6.86	ppb
Silicon	28-1	1546.34	1628.35	1673.14	1615.94	3.98	ppb
Silver	107-1	0.61	0.56	0.58	0.58	3.55	ppb
Silver	109-1	0.57	0.57	0.59	0.57	1.66	ppb
Sodium	23-2	1082.07	1094.89	1099.08	1092.01	0.81	ppb
Strontium	86-1	80.24	82.93	85.60	82.92	3.24	ppb
Strontium	88-1	82.83	86.96	89.83	86.54	4.07	ppb
Sulfur	34-1	-459.61	-449.31	-536.66	-481.86		ppb
Terbium	159-1				103		%
Terbium	159-2				109		%
Thallium	203-1	0.36	0.39	0.41	0.39	6.24	ppb
Thallium	205-1	0.37	0.40	0.41	0.39	5.71	ppb
Tin	118-1	2.42	2.58	2.71	2.57	5.67	ppb
Titanium	47-1	1065.41	1116.32	1125.06	1102.26	2.92	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-01

Instrumnet Name :

P7

Client Sample ID :

MYE4H0

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:41:15

DataFile Name :

023SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.89	0.95	0.95	0.93	3.91	ppb
Vanadium	51-2	161.24	162.89	163.38	162.50	0.69	ppb
Yttrium	89-1				109		%
Yttrium	89-2				114		%
Zinc	66-2	226.54	229.01	230.58	228.71	0.89	ppb
Zirconium	90-1	9.72	10.11	10.25	10.03	2.71	ppb
Zirconium	91-1	9.61	10.18	10.48	10.09	4.38	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-02 Instrumnet Name : P7
Client Sample ID : MYE4H1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:44:21 DataFile Name : 024SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	35730.66	34977.47	35231.97	35313.37	1.08	ppb
Antimony	121-1	3.58	3.67	3.79	3.68	2.80	ppb
Arsenic	75-2	142.02	142.37	144.82	143.07	1.07	ppb
Barium	135-1	432.05	445.32	462.95	446.78	3.47	ppb
Barium	137-1	465.56	474.11	495.18	478.28	3.19	ppb
Beryllium	9-1	1.51	1.55	1.66	1.57	4.93	ppb
Bismuth	209-1				102		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	3.20	3.11	3.87	3.40	12.22	ppb
Cadmium	106-1	3.62	4.98	4.70	4.43	16.28	ppb
Cadmium	111-1	2.34	2.46	2.41	2.40	2.45	ppb
Calcium	43-1	6759.34	6795.98	6774.30	6776.54	0.27	ppb
Calcium	44-1	6352.02	6311.00	6430.07	6364.36	0.95	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	91.98	91.19	91.26	91.48	0.48	ppb
Cobalt	59-2	48.93	48.73	48.28	48.64	0.69	ppb
Copper	63-2	256.83	254.73	254.68	255.41	0.48	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				109		%
Indium	115-1				103		%
Indium	115-2				105		%
Iron	56-2	106130.35	105161.62	106503.73	105931.90	0.65	ppb
Iron	57-2	103546.58	104988.37	103440.21	103991.72	0.83	ppb
Iron	54-2	106237.25	105850.03	105870.62	105985.97	0.21	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-02 Instrumnet Name : P7
Client Sample ID : MYE4H1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:44:21 DataFile Name : 024SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	86.18	89.80	89.04	88.34	2.16	ppb
Lead	207-1	87.16	88.30	88.29	87.92	0.75	ppb
Lead	208-1	93.69	96.04	95.41	95.05	1.28	ppb
Lithium	6-1				102		%
Magnesium	24-2	19054.71	18623.11	18959.53	18879.12	1.20	ppb
Manganese	55-2	2427.03	2389.15	2385.86	2400.68	0.95	ppb
Molybdenum	94-1	23.21	23.53	23.58	23.44	0.85	ppb
Molybdenum	95-1	1.47	1.49	1.49	1.48	0.76	ppb
Molybdenum	96-1	4.67	4.68	4.76	4.70	1.11	ppb
Molybdenum	97-1	1.49	1.53	1.46	1.49	2.32	ppb
Molybdenum	98-1	1.48	1.48	1.47	1.48	0.44	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	89.51	88.70	89.01	89.08	0.46	ppb
Phosphorus	31-2	2105.33	2063.49	2105.58	2091.47	1.16	ppb
Potassium	39-2	7711.88	7656.04	7648.39	7672.10	0.45	ppb
Rhodium	103-1				102		%
Rhodium	103-2				106		%
Scandium	45-1				111		%
Scandium	45-2				117		%
Selenium	82-1	0.79	1.01	0.57	0.79	27.67	ppb
Selenium	77-2	16.01	18.65	18.73	17.80	8.67	ppb
Selenium	78-2	3.69	4.29	4.39	4.13	9.18	ppb
Silicon	28-1	1264.38	1290.24	1276.37	1277.00	1.01	ppb
Silver	107-1	0.59	0.59	0.61	0.60	1.93	ppb
Silver	109-1	0.55	0.58	0.60	0.58	4.32	ppb
Sodium	23-2	167.28	165.91	164.52	165.90	0.83	ppb
Strontium	86-1	56.33	57.43	58.32	57.36	1.74	ppb
Strontium	88-1	57.17	58.70	57.99	57.96	1.32	ppb
Sulfur	34-1	-1384.39	-1475.19	-1478.40	-1445.99		ppb
Terbium	159-1				104		%
Terbium	159-2				110		%
Thallium	203-1	0.43	0.44	0.46	0.44	3.05	ppb
Thallium	205-1	0.42	0.43	0.45	0.44	3.17	ppb
Tin	118-1	2.95	3.06	3.20	3.07	3.97	ppb
Titanium	47-1	822.48	840.66	844.77	835.97	1.42	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-02 Instrumnet Name : P7
Client Sample ID : MYE4H1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:44:21 DataFile Name : 024SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.03	1.02	1.04	1.03	0.88	ppb
Vanadium	51-2	189.12	187.38	187.05	187.85	0.59	ppb
Yttrium	89-1				110		%
Yttrium	89-2				114		%
Zinc	66-2	331.01	331.88	330.15	331.01	0.26	ppb
Zirconium	90-1	13.31	13.36	13.54	13.40	0.92	ppb
Zirconium	91-1	13.11	13.61	13.66	13.46	2.26	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03 Instrumnet Name : P7
Client Sample ID : MYE4H2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:47:25 DataFile Name : 025SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	51365.59	50548.24	50775.56	50896.46	0.83	ppb
Antimony	121-1	3.73	3.72	3.74	3.73	0.36	ppb
Arsenic	75-2	412.66	413.82	410.68	412.39	0.38	ppb
Barium	135-1	621.98	618.99	631.35	624.11	1.03	ppb
Barium	137-1	623.61	614.83	627.78	622.07	1.06	ppb
Beryllium	9-1	1.74	1.76	1.71	1.74	1.46	ppb
Bismuth	209-1				98		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.82	2.00	2.13	1.98	7.96	ppb
Cadmium	106-1	3.19	2.53	3.22	2.98	13.14	ppb
Cadmium	111-1	0.52	0.46	0.54	0.51	8.32	ppb
Calcium	43-1	77911.00	80281.62	80229.43	79474.02	1.70	ppb
Calcium	44-1	77698.63	80281.23	79950.63	79310.16	1.77	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	599.16	603.36	602.87	601.80	0.38	ppb
Cobalt	59-2	107.44	107.26	106.66	107.12	0.38	ppb
Copper	63-2	211.75	211.74	212.39	211.96	0.18	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				107		%
Indium	115-1				101		%
Indium	115-2				102		%
Iron	56-2	159871.57	159437.54	161178.74	160162.62	0.57	ppb
Iron	57-2	159315.27	158321.97	158013.27	158550.17	0.43	ppb
Iron	54-2	162505.78	162266.10	160128.27	161633.38	0.81	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03 Instrumnet Name : P7
Client Sample ID : MYE4H2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:47:25 DataFile Name : 025SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	17.03	16.83	17.18	17.01	1.04	ppb
Lead	207-1	16.22	16.23	16.48	16.31	0.92	ppb
Lead	208-1	16.47	16.36	16.74	16.53	1.19	ppb
Lithium	6-1				101		%
Magnesium	24-2	42755.12	42073.25	42154.66	42327.68	0.88	ppb
Manganese	55-2	5337.44	5323.61	5291.69	5317.58	0.44	ppb
Molybdenum	94-1	22.20	22.93	22.75	22.63	1.68	ppb
Molybdenum	95-1	2.58	2.66	2.62	2.62	1.62	ppb
Molybdenum	96-1	6.05	5.99	6.03	6.03	0.52	ppb
Molybdenum	97-1	2.49	2.55	2.60	2.55	2.14	ppb
Molybdenum	98-1	2.48	2.56	2.51	2.52	1.45	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	372.74	374.35	373.53	373.54	0.22	ppb
Phosphorus	31-2	1611.13	1620.25	1580.07	1603.82	1.31	ppb
Potassium	39-2	3567.12	3621.72	3572.46	3587.10	0.84	ppb
Rhodium	103-1				98		%
Rhodium	103-2				103		%
Scandium	45-1				120		%
Scandium	45-2				124		%
Selenium	82-1	1.10	0.85	0.69	0.88	23.82	ppb
Selenium	77-2	14.63	17.43	15.49	15.85	9.05	ppb
Selenium	78-2	1.73	4.30	2.68	2.90	44.77	ppb
Silicon	28-1	1049.05	1076.07	1078.45	1067.86	1.53	ppb
Silver	107-1	0.12	0.13	0.13	0.13	4.51	ppb
Silver	109-1	0.12	0.12	0.12	0.12	3.02	ppb
Sodium	23-2	195.53	191.49	194.45	193.82	1.08	ppb
Strontium	86-1	120.79	120.74	121.42	120.98	0.31	ppb
Strontium	88-1	122.67	123.79	126.02	124.16	1.37	ppb
Sulfur	34-1	-1303.03	-1171.93	-1210.14	-1228.37		ppb
Terbium	159-1				106		%
Terbium	159-2				107		%
Thallium	203-1	0.32	0.32	0.34	0.32	4.01	ppb
Thallium	205-1	0.33	0.32	0.34	0.33	4.00	ppb
Tin	118-1	1.95	2.04	2.00	2.00	2.29	ppb
Titanium	47-1	1343.74	1381.12	1371.09	1365.32	1.42	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-03

Instrumnet Name :

P7

Client Sample ID :

MYE4H2

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:47:25

DataFile Name :

025SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.08	1.07	1.10	1.08	1.37	ppb
Vanadium	51-2	356.84	354.93	355.77	355.84	0.27	ppb
Yttrium	89-1				113		%
Yttrium	89-2				114		%
Zinc	66-2	195.95	195.76	195.29	195.66	0.17	ppb
Zirconium	90-1	12.06	12.23	12.42	12.24	1.46	ppb
Zirconium	91-1	12.27	12.57	12.61	12.49	1.49	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-04 Instrumnet Name : P7
Client Sample ID : MYE4H2D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:50:27 DataFile Name : 026SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	50447.30	50483.49	50228.40	50386.40	0.27	ppb
Antimony	121-1	3.54	3.47	3.43	3.48	1.66	ppb
Arsenic	75-2	394.73	393.47	399.60	395.93	0.82	ppb
Barium	135-1	633.04	589.92	596.19	606.38	3.84	ppb
Barium	137-1	625.70	600.27	590.37	605.44	3.01	ppb
Beryllium	9-1	1.75	1.64	1.52	1.64	7.05	ppb
Bismuth	209-1				99		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.36	1.88	2.19	2.14	11.16	ppb
Cadmium	106-1	2.97	3.22	2.76	2.98	7.71	ppb
Cadmium	111-1	0.54	0.55	0.50	0.53	4.79	ppb
Calcium	43-1	81730.65	80045.20	80002.64	80592.83	1.22	ppb
Calcium	44-1	81742.49	80439.29	79564.99	80582.26	1.36	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	591.98	595.63	593.77	593.79	0.31	ppb
Cobalt	59-2	104.44	104.98	104.68	104.70	0.26	ppb
Copper	63-2	208.33	210.47	209.86	209.56	0.53	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				105		%
Indium	115-1				100		%
Indium	115-2				101		%
Iron	56-2	157616.78	157797.63	156984.25	157466.22	0.27	ppb
Iron	57-2	155703.98	157349.23	155055.44	156036.22	0.76	ppb
Iron	54-2	156451.93	157932.09	156448.82	156944.28	0.55	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-04 Instrumnet Name : P7
Client Sample ID : MYE4H2D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:50:27 DataFile Name : 026SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	17.31	16.27	16.55	16.71	3.23	ppb
Lead	207-1	16.69	15.89	15.96	16.18	2.75	ppb
Lead	208-1	16.70	15.98	16.14	16.27	2.34	ppb
Lithium	6-1				98		%
Magnesium	24-2	41995.19	41994.19	41986.40	41991.93	0.01	ppb
Manganese	55-2	5130.71	5170.93	5150.09	5150.58	0.39	ppb
Molybdenum	94-1	23.03	22.45	22.59	22.69	1.35	ppb
Molybdenum	95-1	2.61	2.59	2.57	2.59	0.81	ppb
Molybdenum	96-1	6.15	5.93	5.94	6.01	2.10	ppb
Molybdenum	97-1	2.60	2.58	2.54	2.57	1.29	ppb
Molybdenum	98-1	2.51	2.42	2.43	2.45	1.97	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	370.16	368.99	366.56	368.57	0.50	ppb
Phosphorus	31-2	1553.56	1562.59	1583.39	1566.51	0.98	ppb
Potassium	39-2	3594.86	3657.19	3643.74	3631.93	0.90	ppb
Rhodium	103-1				97		%
Rhodium	103-2				100		%
Scandium	45-1				115		%
Scandium	45-2				121		%
Selenium	82-1	1.12	0.65	0.29	0.69	60.24	ppb
Selenium	77-2	18.47	18.81	19.80	19.03	3.64	ppb
Selenium	78-2	2.74	3.30	3.69	3.24	14.84	ppb
Silicon	28-1	1120.39	1106.80	1093.84	1107.01	1.20	ppb
Silver	107-1	0.11	0.12	0.12	0.12	4.10	ppb
Silver	109-1	0.10	0.10	0.10	0.10	1.94	ppb
Sodium	23-2	189.95	189.44	188.62	189.34	0.35	ppb
Strontium	86-1	123.41	118.64	119.37	120.47	2.13	ppb
Strontium	88-1	126.52	121.19	122.94	123.55	2.20	ppb
Sulfur	34-1	-1243.05	-1364.00	-1373.37	-1326.80		ppb
Terbium	159-1				105		%
Terbium	159-2				106		%
Thallium	203-1	0.34	0.31	0.34	0.33	5.06	ppb
Thallium	205-1	0.32	0.33	0.31	0.32	2.69	ppb
Tin	118-1	2.17	2.05	2.04	2.09	3.33	ppb
Titanium	47-1	1380.42	1344.35	1340.85	1355.21	1.62	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-04

Instrumnet Name :

P7

Client Sample ID :

MYE4H2D

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:50:27

DataFile Name :

026SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.12	1.05	1.08	1.09	3.17	ppb
Vanadium	51-2	345.31	350.40	349.83	348.51	0.80	ppb
Yttrium	89-1				109		%
Yttrium	89-2				112		%
Zinc	66-2	197.06	197.77	196.63	197.15	0.29	ppb
Zirconium	90-1	12.53	12.18	12.40	12.37	1.41	ppb
Zirconium	91-1	12.80	12.42	12.43	12.55	1.73	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03LX5 Instrumnet Name : P7
Client Sample ID : MYE4H2L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 18:53:27 DataFile Name : 027SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	11659.51	11745.22	11506.55	11637.09	1.04	ppb
Antimony	121-1	0.79	0.75	0.79	0.77	3.13	ppb
Arsenic	75-2	89.69	89.68	89.36	89.57	0.21	ppb
Barium	135-1	117.61	115.71	118.33	117.22	1.16	ppb
Barium	137-1	117.12	115.81	118.96	117.30	1.34	ppb
Beryllium	9-1	0.36	0.40	0.38	0.38	5.08	ppb
Bismuth	209-1				101		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.55	0.38	0.37	0.44	23.14	ppb
Cadmium	106-1	-0.03	-0.43	-0.11	-0.19		ppb
Cadmium	111-1	0.05	0.05	0.09	0.06	36.01	ppb
Calcium	43-1	19009.77	19005.62	19051.21	19022.20	0.13	ppb
Calcium	44-1	18010.87	18105.79	17961.20	18025.96	0.41	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	137.36	136.17	137.14	136.89	0.46	ppb
Cobalt	59-2	24.80	25.12	24.99	24.97	0.63	ppb
Copper	63-2	49.80	50.32	50.07	50.07	0.52	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				104		%
Indium	115-1				99		%
Indium	115-2				102		%
Iron	56-2	36821.43	37076.87	36870.79	36923.03	0.37	ppb
Iron	57-2	36221.99	37290.02	36439.08	36650.37	1.54	ppb
Iron	54-2	36972.51	37092.88	36901.52	36988.97	0.26	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03LX5 Instrumnet Name : P7
Client Sample ID : MYE4H2L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 18:53:27 DataFile Name : 027SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	3.32	3.42	3.39	3.37	1.46	ppb
Lead	207-1	3.20	3.25	3.30	3.25	1.62	ppb
Lead	208-1	3.26	3.30	3.34	3.30	1.11	ppb
Lithium	6-1				98		%
Magnesium	24-2	9683.79	9636.40	9693.01	9671.07	0.31	ppb
Manganese	55-2	1219.38	1232.76	1221.46	1224.53	0.59	ppb
Molybdenum	94-1	4.91	4.81	4.98	4.90	1.79	ppb
Molybdenum	95-1	0.68	0.66	0.66	0.67	1.99	ppb
Molybdenum	96-1	1.35	1.30	1.36	1.34	2.51	ppb
Molybdenum	97-1	0.65	0.66	0.67	0.66	1.53	ppb
Molybdenum	98-1	0.60	0.59	0.61	0.60	1.23	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	87.42	87.16	87.82	87.46	0.38	ppb
Phosphorus	31-2	352.63	345.93	349.99	349.52	0.97	ppb
Potassium	39-2	863.21	866.25	861.52	863.66	0.28	ppb
Rhodium	103-1				96		%
Rhodium	103-2				103		%
Scandium	45-1				101		%
Scandium	45-2				106		%
Selenium	82-1	-0.03	-0.26	-0.58	-0.29		ppb
Selenium	77-2	4.24	4.21	4.24	4.23	0.43	ppb
Selenium	78-2	-0.23	1.90	-0.37	0.43	293.66	ppb
Silicon	28-1	246.76	243.50	245.01	245.09	0.67	ppb
Silver	107-1	0.03	0.03	0.03	0.03	6.19	ppb
Silver	109-1	0.03	0.03	0.03	0.03	13.96	ppb
Sodium	23-2	55.39	56.79	55.61	55.93	1.35	ppb
Strontium	86-1	26.38	26.83	26.43	26.55	0.93	ppb
Strontium	88-1	26.19	26.25	26.69	26.38	1.03	ppb
Sulfur	34-1	-1182.98	-1261.90	-1234.01	-1226.29		ppb
Terbium	159-1				103		%
Terbium	159-2				104		%
Thallium	203-1	0.08	0.08	0.08	0.08	3.95	ppb
Thallium	205-1	0.07	0.08	0.08	0.08	6.73	ppb
Tin	118-1	0.44	0.42	0.43	0.43	3.20	ppb
Titanium	47-1	312.58	312.39	312.52	312.50	0.03	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-03LX5

Instrumnet Name :

P7

Client Sample ID :

MYE4H2L

Dilution Factor :

5

Date & Time Acquired :

2024-11-02 18:53:27

DataFile Name :

027SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.21	0.22	0.23	0.22	4.42	ppb
Vanadium	51-2	81.68	82.63	81.20	81.84	0.89	ppb
Yttrium	89-1				101		%
Yttrium	89-2				104		%
Zinc	66-2	47.59	47.21	45.95	46.92	1.83	ppb
Zirconium	90-1	2.65	2.61	2.63	2.63	0.76	ppb
Zirconium	91-1	2.65	2.66	2.65	2.65	0.25	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:02:43 DataFile Name : 030SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	11289.11	12914.32	11168.83	11790.75	8.27	ppb
Antimony	121-1	0.74	0.74	0.76	0.75	1.94	ppb
Arsenic	75-2	87.45	99.47	84.00	90.30	8.99	ppb
Barium	135-1	111.61	113.84	116.26	113.90	2.04	ppb
Barium	137-1	111.78	114.31	116.35	114.15	2.01	ppb
Beryllium	9-1	0.26	0.38	0.35	0.33	18.60	ppb
Bismuth	209-1				103		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.52	0.39	0.32	0.41	25.39	ppb
Cadmium	106-1	-0.60	-0.48	-0.15	-0.41		ppb
Cadmium	111-1	0.01	0.04	0.05	0.03	60.04	ppb
Calcium	43-1	18223.23	18798.39	18948.31	18656.64	2.05	ppb
Calcium	44-1	17151.53	17662.01	17869.65	17561.06	2.10	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	132.80	153.49	129.79	138.69	9.31	ppb
Cobalt	59-2	24.12	27.27	23.26	24.88	8.49	ppb
Copper	63-2	48.34	54.53	46.75	49.87	8.24	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				102		%
Indium	115-1				100		%
Indium	115-2				100		%
Iron	56-2	35587.73	40769.40	34701.42	37019.51	8.85	ppb
Iron	57-2	35817.00	40765.62	34279.21	36953.94	9.17	ppb
Iron	54-2	36090.19	41420.76	34767.79	37426.25	9.41	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:02:43 DataFile Name : 030SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	3.23	3.35	3.16	3.25	2.93	ppb
Lead	207-1	3.06	3.32	3.17	3.18	4.06	ppb
Lead	208-1	3.11	3.25	3.16	3.18	2.18	ppb
Lithium	6-1				98		%
Magnesium	24-2	9232.96	10426.69	9086.27	9581.97	7.67	ppb
Manganese	55-2	1180.24	1350.41	1140.82	1223.82	9.10	ppb
Molybdenum	94-1	4.74	4.91	4.84	4.83	1.81	ppb
Molybdenum	95-1	0.61	0.63	0.64	0.63	2.20	ppb
Molybdenum	96-1	1.30	1.33	1.31	1.31	1.47	ppb
Molybdenum	97-1	0.54	0.56	0.58	0.56	2.94	ppb
Molybdenum	98-1	0.58	0.58	0.58	0.58	0.49	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	85.79	97.35	83.68	88.94	8.27	ppb
Phosphorus	31-2	329.65	377.47	319.38	342.17	9.06	ppb
Potassium	39-2	839.14	981.65	810.48	877.09	10.45	ppb
Rhodium	103-1				98		%
Rhodium	103-2				101		%
Scandium	45-1				100		%
Scandium	45-2				103		%
Selenium	82-1	0.07	0.14	-0.15	0.02	786.82	ppb
Selenium	77-2	2.22	4.59	3.59	3.47	34.27	ppb
Selenium	78-2	0.70	1.10	0.40	0.73	47.85	ppb
Silicon	28-1	231.53	237.96	241.35	236.95	2.10	ppb
Silver	107-1	0.03	0.03	0.04	0.03	11.28	ppb
Silver	109-1	0.03	0.03	0.03	0.03	6.77	ppb
Sodium	23-2	47.86	63.00	46.99	52.62	17.11	ppb
Strontium	86-1	24.70	25.97	25.61	25.42	2.58	ppb
Strontium	88-1	24.90	25.82	25.58	25.44	1.88	ppb
Sulfur	34-1	-1219.78	-1173.81	-1093.42	-1162.34		ppb
Terbium	159-1				103		%
Terbium	159-2				104		%
Thallium	203-1	0.06	0.07	0.06	0.06	4.96	ppb
Thallium	205-1	0.06	0.06	0.06	0.06	2.04	ppb
Tin	118-1	0.37	0.40	0.40	0.39	5.15	ppb
Titanium	47-1	298.05	308.19	312.38	306.20	2.41	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-03DLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4H2

Dilution Factor :

4

Date & Time Acquired :

2024-11-02 19:02:43

DataFile Name :

030SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.21	0.22	0.20	0.21	3.23	ppb
Vanadium	51-2	79.43	90.27	76.70	82.13	8.74	ppb
Yttrium	89-1				101		%
Yttrium	89-2				102		%
Zinc	66-2	46.13	54.05	45.25	48.48	10.00	ppb
Zirconium	90-1	2.46	2.60	2.57	2.54	2.78	ppb
Zirconium	91-1	2.53	2.57	2.63	2.58	2.06	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-04DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2D Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:06:09 DataFile Name : 031SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	13778.20	14052.23	13964.44	13931.62	1.00	ppb
Antimony	121-1	0.84	0.89	0.88	0.87	3.14	ppb
Arsenic	75-2	108.22	108.17	106.73	107.71	0.79	ppb
Barium	135-1	138.14	138.83	140.46	139.14	0.85	ppb
Barium	137-1	138.56	139.38	141.93	139.96	1.26	ppb
Beryllium	9-1	0.44	0.47	0.38	0.43	9.94	ppb
Bismuth	209-1				102		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.42	0.47	0.55	0.48	14.08	ppb
Cadmium	106-1	-0.73	-0.82	0.38	-0.39		ppb
Cadmium	111-1	0.04	0.04	0.11	0.06	67.04	ppb
Calcium	43-1	23116.70	23508.99	24035.81	23553.84	1.96	ppb
Calcium	44-1	22043.90	22202.16	22695.88	22313.98	1.52	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	164.73	165.25	164.08	164.69	0.36	ppb
Cobalt	59-2	29.29	29.86	29.52	29.56	0.96	ppb
Copper	63-2	59.56	60.35	59.55	59.82	0.76	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				106		%
Indium	115-1				100		%
Indium	115-2				102		%
Iron	56-2	43192.05	44508.79	43681.44	43794.09	1.52	ppb
Iron	57-2	43538.67	44290.71	43969.71	43933.03	0.86	ppb
Iron	54-2	43590.78	44592.61	44341.52	44174.97	1.18	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-04DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2D Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:06:09 DataFile Name : 031SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	4.10	4.19	4.19	4.16	1.19	ppb
Lead	207-1	3.90	4.02	3.97	3.96	1.47	ppb
Lead	208-1	3.94	4.03	4.05	4.01	1.46	ppb
Lithium	6-1				99		%
Magnesium	24-2	11590.35	11641.82	11506.96	11579.71	0.59	ppb
Manganese	55-2	1405.09	1463.39	1453.46	1440.65	2.16	ppb
Molybdenum	94-1	6.02	6.02	6.10	6.05	0.83	ppb
Molybdenum	95-1	0.76	0.71	0.77	0.75	4.01	ppb
Molybdenum	96-1	1.62	1.65	1.69	1.65	2.15	ppb
Molybdenum	97-1	0.74	0.77	0.70	0.73	4.92	ppb
Molybdenum	98-1	0.69	0.74	0.72	0.72	3.36	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	104.85	105.48	105.26	105.20	0.30	ppb
Phosphorus	31-2	446.80	438.88	419.75	435.15	3.20	ppb
Potassium	39-2	1052.50	1065.92	1060.05	1059.49	0.64	ppb
Rhodium	103-1				98		%
Rhodium	103-2				103		%
Scandium	45-1				101		%
Scandium	45-2				107		%
Selenium	82-1	-0.03	0.18	0.46	0.20	122.11	ppb
Selenium	77-2	6.92	4.44	5.20	5.52	23.07	ppb
Selenium	78-2	1.47	-0.94	1.00	0.51	250.09	ppb
Silicon	28-1	304.77	312.37	314.70	310.62	1.67	ppb
Silver	107-1	0.03	0.02	0.03	0.03	9.28	ppb
Silver	109-1	0.03	0.03	0.03	0.03	10.57	ppb
Sodium	23-2	56.45	57.20	56.07	56.57	1.02	ppb
Strontium	86-1	31.18	31.60	32.49	31.76	2.10	ppb
Strontium	88-1	31.44	31.79	32.10	31.78	1.05	ppb
Sulfur	34-1	-1085.80	-1007.41	-967.17	-1020.12		ppb
Terbium	159-1				103		%
Terbium	159-2				106		%
Thallium	203-1	0.07	0.08	0.08	0.08	7.97	ppb
Thallium	205-1	0.08	0.07	0.07	0.07	8.96	ppb
Tin	118-1	0.49	0.50	0.50	0.50	0.26	ppb
Titanium	47-1	372.97	377.40	384.59	378.32	1.55	ppb

LB Number :
Lab Sample ID :
Client Sample ID :
Date & Time Acquired :

LB133265
P4484-04DLX4
MYE4H2D
2024-11-02 19:06:09

Operator :
Instrumnet Name :
Dilution Factor :
DataFile Name :

Jaswal
P7
4
031SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.25	0.26	0.26	0.26	2.06	ppb
Vanadium	51-2	97.05	98.02	97.35	97.47	0.51	ppb
Yttrium	89-1				101		%
Yttrium	89-2				104		%
Zinc	66-2	56.53	57.18	56.43	56.71	0.72	ppb
Zirconium	90-1	3.14	3.23	3.28	3.22	2.19	ppb
Zirconium	91-1	3.29	3.34	3.36	3.33	1.17	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4H2L Dilution Factor : 20
Date & Time Acquired : 2024-11-02 19:09:20 DataFile Name : 032SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	2984.15	2976.07	2960.10	2973.44	0.41	ppb
Antimony	121-1	0.18	0.19	0.19	0.19	3.70	ppb
Arsenic	75-2	22.99	21.87	23.02	22.63	2.89	ppb
Barium	135-1	28.11	28.25	28.76	28.37	1.20	ppb
Barium	137-1	28.26	28.60	28.98	28.61	1.26	ppb
Beryllium	9-1	0.10	0.07	0.07	0.08	18.33	ppb
Bismuth	209-1				103		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.21	0.06	0.08	0.12	65.15	ppb
Cadmium	106-1	-0.36	-0.78	-0.72	-0.62		ppb
Cadmium	111-1	0.01	-0.03	-0.03	-0.02		ppb
Calcium	43-1	4720.38	4824.23	4807.02	4783.88	1.16	ppb
Calcium	44-1	4475.38	4525.42	4530.44	4510.41	0.68	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	34.41	34.14	34.48	34.34	0.52	ppb
Cobalt	59-2	6.37	6.23	6.30	6.30	1.13	ppb
Copper	63-2	12.74	12.54	12.63	12.64	0.78	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				104		%
Indium	115-1				100		%
Indium	115-2				102		%
Iron	56-2	9238.26	9235.94	9124.98	9199.73	0.70	ppb
Iron	57-2	9412.25	9403.80	9434.97	9417.01	0.17	ppb
Iron	54-2	9377.27	9455.67	9409.62	9414.19	0.42	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4H2L Dilution Factor : 20
Date & Time Acquired : 2024-11-02 19:09:20 DataFile Name : 032SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.85	0.83	0.81	0.83	2.29	ppb
Lead	207-1	0.81	0.78	0.84	0.81	3.25	ppb
Lead	208-1	0.82	0.81	0.81	0.81	0.58	ppb
Lithium	6-1				100		%
Magnesium	24-2	2480.45	2482.95	2460.72	2474.70	0.49	ppb
Manganese	55-2	304.21	306.58	304.22	305.00	0.45	ppb
Molybdenum	94-1	1.36	1.29	1.31	1.32	2.81	ppb
Molybdenum	95-1	0.22	0.23	0.20	0.21	8.20	ppb
Molybdenum	96-1	0.43	0.40	0.38	0.40	5.99	ppb
Molybdenum	97-1	0.23	0.23	0.21	0.22	4.15	ppb
Molybdenum	98-1	0.21	0.21	0.22	0.22	1.90	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	21.48	21.92	22.12	21.84	1.50	ppb
Phosphorus	31-2	80.01	75.13	77.13	77.43	3.17	ppb
Potassium	39-2	214.79	212.09	208.25	211.71	1.55	ppb
Rhodium	103-1				98		%
Rhodium	103-2				104		%
Scandium	45-1				98		%
Scandium	45-2				103		%
Selenium	82-1	-0.69	-0.52	-0.60	-0.61		ppb
Selenium	77-2	0.24	0.91	0.45	0.53	64.51	ppb
Selenium	78-2	0.00	0.82	-0.39	0.14	435.09	ppb
Silicon	28-1	61.93	63.42	63.35	62.90	1.33	ppb
Silver	107-1	0.01	0.01	0.01	0.01	41.82	ppb
Silver	109-1	0.01	0.01	0.01	0.01	34.87	ppb
Sodium	23-2	17.53	17.22	16.33	17.03	3.68	ppb
Strontium	86-1	6.69	6.73	6.58	6.67	1.13	ppb
Strontium	88-1	6.66	6.62	6.69	6.66	0.52	ppb
Sulfur	34-1	-981.38	-908.54	-908.36	-932.76		ppb
Terbium	159-1				103		%
Terbium	159-2				105		%
Thallium	203-1	0.02	0.02	0.02	0.02	1.40	ppb
Thallium	205-1	0.02	0.02	0.02	0.02	9.42	ppb
Tin	118-1	0.14	0.09	0.09	0.10	27.90	ppb
Titanium	47-1	78.07	77.71	79.69	78.49	1.35	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-03LDLX20

Instrumnet Name :

P7

Client Sample ID :

MYE4H2L

Dilution Factor :

20

Date & Time Acquired :

2024-11-02 19:09:20

DataFile Name :

032SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.06	0.06	0.06	0.06	5.53	ppb
Vanadium	51-2	20.34	20.30	20.04	20.23	0.82	ppb
Yttrium	89-1				100		%
Yttrium	89-2				102		%
Zinc	66-2	11.58	11.98	12.15	11.90	2.47	ppb
Zirconium	90-1	0.66	0.68	0.68	0.68	1.71	ppb
Zirconium	91-1	0.65	0.71	0.67	0.68	4.72	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-05 Instrumnet Name : P7
Client Sample ID : MYE4H2S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:15:50 DataFile Name : 034SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	61752.66	60062.81	61732.92	61182.80	1.59	ppb
Antimony	121-1	69.69	74.04	73.10	72.27	3.16	ppb
Arsenic	75-2	497.27	513.21	522.19	510.89	2.47	ppb
Barium	135-1	2083.55	2184.38	2170.12	2146.02	2.54	ppb
Barium	137-1	2074.31	2198.53	2172.36	2148.40	3.05	ppb
Beryllium	9-1	38.22	39.38	40.39	39.33	2.76	ppb
Bismuth	209-1				97		%
Bismuth	209-2				96		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	33.44	38.45	36.48	36.12	6.98	ppb
Cadmium	106-1	33.35	35.04	33.28	33.89	2.93	ppb
Cadmium	111-1	33.79	35.16	35.01	34.65	2.16	ppb
Calcium	43-1	91128.92	94084.23	94527.58	93246.91	1.98	ppb
Calcium	44-1	90630.32	94395.08	93871.22	92965.54	2.19	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	821.95	827.33	844.95	831.41	1.45	ppb
Cobalt	59-2	449.15	444.83	447.22	447.07	0.48	ppb
Copper	63-2	415.45	412.09	415.90	414.48	0.50	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				102		%
Indium	115-1				97		%
Indium	115-2				95		%
Iron	56-2	191160.72	186901.82	188682.86	188915.14	1.13	ppb
Iron	57-2	187947.55	185595.96	188903.00	187482.17	0.91	ppb
Iron	54-2	191737.57	187990.08	190366.96	190031.53	1.00	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-05 Instrumnet Name : P7
Client Sample ID : MYE4H2S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:15:50 DataFile Name : 034SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	32.37	32.97	33.54	32.96	1.77	ppb
Lead	207-1	31.13	32.08	32.42	31.88	2.11	ppb
Lead	208-1	31.46	32.39	32.87	32.24	2.21	ppb
Lithium	6-1				100		%
Magnesium	24-2	49667.62	49392.47	49898.38	49652.82	0.51	ppb
Manganese	55-2	6630.35	6432.40	6626.01	6562.92	1.72	ppb
Molybdenum	94-1	26.03	27.45	27.26	26.91	2.87	ppb
Molybdenum	95-1	2.84	3.03	2.97	2.95	3.28	ppb
Molybdenum	96-1	6.73	7.14	7.02	6.96	3.04	ppb
Molybdenum	97-1	2.83	2.88	2.84	2.85	0.96	ppb
Molybdenum	98-1	2.72	2.86	2.85	2.81	2.68	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	746.63	735.05	743.20	741.63	0.80	ppb
Phosphorus	31-2	1864.00	1838.45	1835.58	1846.01	0.85	ppb
Potassium	39-2	4228.05	4166.93	4143.49	4179.49	1.04	ppb
Rhodium	103-1				93		%
Rhodium	103-2				96		%
Scandium	45-1				113		%
Scandium	45-2				117		%
Selenium	82-1	73.96	77.23	77.57	76.25	2.61	ppb
Selenium	77-2	94.05	93.17	89.98	92.40	2.32	ppb
Selenium	78-2	76.51	77.49	79.48	77.82	1.94	ppb
Silicon	28-1	1493.73	1592.28	1569.80	1551.94	3.33	ppb
Silver	107-1	28.88	30.43	30.02	29.77	2.70	ppb
Silver	109-1	28.89	30.30	30.21	29.80	2.66	ppb
Sodium	23-2	210.64	208.19	213.19	210.67	1.19	ppb
Strontium	86-1	138.18	144.48	145.40	142.69	2.75	ppb
Strontium	88-1	140.85	148.55	147.43	145.61	2.86	ppb
Sulfur	34-1	-1431.45	-1422.63	-1484.69	-1446.26		ppb
Terbium	159-1				103		%
Terbium	159-2				102		%
Thallium	203-1	30.86	31.96	32.44	31.75	2.56	ppb
Thallium	205-1	31.23	31.61	31.87	31.57	1.00	ppb
Tin	118-1	2.22	2.35	2.29	2.29	2.83	ppb
Titanium	47-1	1587.42	1656.78	1643.19	1629.13	2.26	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-05 Instrumnet Name : P7
Client Sample ID : MYE4H2S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:15:50 DataFile Name : 034SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.28	1.34	1.33	1.32	2.36	ppb
Vanadium	51-2	727.09	731.28	744.08	734.15	1.21	ppb
Yttrium	89-1				108		%
Yttrium	89-2				108		%
Zinc	66-2	534.70	525.73	525.39	528.61	1.00	ppb
Zirconium	90-1	14.03	14.92	14.92	14.62	3.52	ppb
Zirconium	91-1	14.09	14.94	15.36	14.80	4.36	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03MS Instrumnet Name : P7
Client Sample ID : MYE4H2SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:18:48 DataFile Name : 035SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	63791.08	63545.94	63903.16	63746.73	0.29	ppb
Antimony	121-1	353.03	359.87	359.09	357.33	1.05	ppb
Arsenic	75-2	606.05	604.85	607.88	606.26	0.25	ppb
Barium	135-1	7682.38	7776.14	7795.33	7751.28	0.78	ppb
Barium	137-1	7560.39	7764.82	7697.36	7674.19	1.36	ppb
Beryllium	9-1	179.21	186.36	186.79	184.12	2.31	ppb
Bismuth	209-1				97		%
Bismuth	209-2				97		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	155.99	159.06	157.05	157.37	0.99	ppb
Cadmium	106-1	152.70	156.62	155.87	155.06	1.34	ppb
Cadmium	111-1	164.71	168.14	167.96	166.93	1.16	ppb
Calcium	43-1	89412.02	92377.01	92270.13	91353.05	1.84	ppb
Calcium	44-1	89706.51	91822.47	91412.07	90980.35	1.23	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1313.23	1310.89	1315.29	1313.14	0.17	ppb
Cobalt	59-2	1691.76	1677.60	1698.62	1689.33	0.63	ppb
Copper	63-2	978.01	985.75	989.20	984.32	0.58	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				103		%
Indium	115-1				95		%
Indium	115-2				96		%
Iron	56-2	188061.57	186065.59	187627.75	187251.64	0.56	ppb
Iron	57-2	186007.51	184503.78	186000.29	185503.86	0.47	ppb
Iron	54-2	188815.56	187974.89	189122.78	188637.74	0.32	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03MS Instrumnet Name : P7
Client Sample ID : MYE4H2SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:18:48 DataFile Name : 035SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	80.16	83.99	83.30	82.48	2.48	ppb
Lead	207-1	78.66	81.36	81.17	80.40	1.87	ppb
Lead	208-1	79.20	82.26	81.95	81.14	2.08	ppb
Lithium	6-1				96		%
Magnesium	24-2	48401.35	48314.77	48945.20	48553.77	0.70	ppb
Manganese	55-2	7642.03	7577.11	7634.95	7618.03	0.47	ppb
Molybdenum	94-1	25.82	26.17	26.84	26.28	1.96	ppb
Molybdenum	95-1	2.78	2.84	3.00	2.87	4.02	ppb
Molybdenum	96-1	6.67	6.88	6.85	6.80	1.68	ppb
Molybdenum	97-1	2.63	2.80	2.79	2.74	3.57	ppb
Molybdenum	98-1	2.67	2.79	2.82	2.76	2.96	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	2028.10	2005.55	1999.92	2011.19	0.74	ppb
Phosphorus	31-2	1833.33	1790.98	1798.77	1807.70	1.25	ppb
Potassium	39-2	4135.59	4104.04	4051.16	4096.93	1.04	ppb
Rhodium	103-1				90		%
Rhodium	103-2				96		%
Scandium	45-1				109		%
Scandium	45-2				117		%
Selenium	82-1	348.27	355.16	363.23	355.55	2.11	ppb
Selenium	77-2	371.01	380.30	374.81	375.37	1.24	ppb
Selenium	78-2	353.99	358.52	372.52	361.68	2.67	ppb
Silicon	28-1	1472.23	1505.39	1502.80	1493.47	1.23	ppb
Silver	107-1	151.39	153.15	153.77	152.77	0.81	ppb
Silver	109-1	151.34	154.58	154.77	153.56	1.25	ppb
Sodium	23-2	207.59	208.12	206.44	207.38	0.41	ppb
Strontium	86-1	136.55	140.57	141.82	139.65	1.97	ppb
Strontium	88-1	140.25	143.47	144.73	142.82	1.62	ppb
Sulfur	34-1	-1704.01	-1702.63	-1725.56	-1710.74		ppb
Terbium	159-1				102		%
Terbium	159-2				103		%
Thallium	203-1	178.78	186.33	183.65	182.92	2.10	ppb
Thallium	205-1	177.33	183.58	179.45	180.12	1.77	ppb
Tin	118-1	2.21	2.20	2.23	2.21	0.62	ppb
Titanium	47-1	1568.77	1612.72	1619.96	1600.48	1.73	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-03MS

Instrumnet Name :

P7

Client Sample ID :

MYE4H2SRE

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 19:18:48

DataFile Name :

035SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.27	1.31	1.30	1.29	1.53	ppb
Vanadium	51-2	1973.50	1941.70	1968.84	1961.35	0.88	ppb
Yttrium	89-1				105		%
Yttrium	89-2				108		%
Zinc	66-2	1689.65	1677.64	1690.26	1685.85	0.42	ppb
Zirconium	90-1	13.96	14.45	14.69	14.36	2.59	ppb
Zirconium	91-1	14.07	14.70	14.97	14.58	3.18	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-05DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2S Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:21:42 DataFile Name : 036SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	16454.77	17046.16	17046.48	16849.14	2.03	ppb
Antimony	121-1	18.54	18.21	18.87	18.54	1.79	ppb
Arsenic	75-2	133.51	138.03	138.57	136.70	2.03	ppb
Barium	135-1	501.12	497.61	509.94	502.89	1.26	ppb
Barium	137-1	541.34	539.92	548.78	543.35	0.88	ppb
Beryllium	9-1	10.36	10.05	10.36	10.26	1.76	ppb
Bismuth	209-1				100		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	8.24	8.79	8.50	8.51	3.22	ppb
Cadmium	106-1	7.88	6.41	7.30	7.20	10.26	ppb
Cadmium	111-1	8.98	8.91	9.02	8.97	0.60	ppb
Calcium	43-1	27544.71	28116.69	28527.24	28062.88	1.76	ppb
Calcium	44-1	26015.98	26324.53	26884.73	26408.42	1.67	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	227.74	236.45	234.41	232.87	1.96	ppb
Cobalt	59-2	120.63	124.66	123.75	123.01	1.72	ppb
Copper	63-2	109.97	113.51	113.20	112.23	1.75	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				105		%
Indium	115-1				97		%
Indium	115-2				100		%
Iron	56-2	51775.11	53383.23	52963.12	52707.15	1.58	ppb
Iron	57-2	51528.79	53442.95	52918.57	52630.11	1.88	ppb
Iron	54-2	52297.41	53747.42	53395.89	53146.91	1.42	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-05DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2S Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:21:42 DataFile Name : 036SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	8.13	8.33	8.22	8.22	1.22	ppb
Lead	207-1	7.92	8.14	7.92	7.99	1.57	ppb
Lead	208-1	7.97	8.06	8.06	8.03	0.65	ppb
Lithium	6-1				97		%
Magnesium	24-2	13371.40	13807.83	13795.14	13658.12	1.82	ppb
Manganese	55-2	1790.03	1855.25	1831.00	1825.43	1.81	ppb
Molybdenum	94-1	7.16	7.22	7.38	7.25	1.56	ppb
Molybdenum	95-1	0.82	0.77	0.82	0.80	3.34	ppb
Molybdenum	96-1	1.86	1.83	1.88	1.86	1.26	ppb
Molybdenum	97-1	0.83	0.73	0.82	0.79	7.05	ppb
Molybdenum	98-1	0.74	0.77	0.77	0.76	1.85	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	208.55	216.45	215.55	213.52	2.03	ppb
Phosphorus	31-2	513.36	495.88	483.20	497.48	3.04	ppb
Potassium	39-2	1202.69	1257.78	1247.10	1235.86	2.36	ppb
Rhodium	103-1				94		%
Rhodium	103-2				101		%
Scandium	45-1				96		%
Scandium	45-2				105		%
Selenium	82-1	20.11	20.99	20.71	20.60	2.17	ppb
Selenium	77-2	24.27	25.96	28.24	26.16	7.63	ppb
Selenium	78-2	22.52	20.39	20.68	21.20	5.46	ppb
Silicon	28-1	429.87	435.27	445.07	436.74	1.76	ppb
Silver	107-1	7.65	7.58	7.76	7.66	1.20	ppb
Silver	109-1	7.60	7.52	7.75	7.62	1.54	ppb
Sodium	23-2	58.02	61.53	60.49	60.02	3.00	ppb
Strontium	86-1	38.11	38.30	38.38	38.26	0.37	ppb
Strontium	88-1	38.25	38.43	38.62	38.43	0.48	ppb
Sulfur	34-1	-1439.24	-1348.45	-1344.39	-1377.36		ppb
Terbium	159-1				103		%
Terbium	159-2				105		%
Thallium	203-1	7.91	8.04	7.86	7.94	1.15	ppb
Thallium	205-1	7.89	7.92	7.91	7.91	0.24	ppb
Tin	118-1	0.55	0.55	0.53	0.55	1.96	ppb
Titanium	47-1	460.24	469.25	476.81	468.77	1.77	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-05DLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4H2S

Dilution Factor :

4

Date & Time Acquired :

2024-11-02 19:21:42

DataFile Name :

036SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.33	0.32	0.32	0.32	1.62	ppb
Vanadium	51-2	199.67	208.32	206.02	204.67	2.19	ppb
Yttrium	89-1				98		%
Yttrium	89-2				102		%
Zinc	66-2	149.29	153.43	152.38	151.70	1.42	ppb
Zirconium	90-1	3.92	3.93	3.93	3.93	0.16	ppb
Zirconium	91-1	4.03	3.94	4.10	4.02	1.96	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:24:50 DataFile Name : 037SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	18313.58	17891.30	18382.61	18195.83	1.46	ppb
Antimony	121-1	86.07	86.81	87.33	86.73	0.73	ppb
Arsenic	75-2	160.53	165.45	164.09	163.36	1.56	ppb
Barium	135-1	1958.49	1955.80	1973.03	1962.44	0.47	ppb
Barium	137-1	1958.58	1952.43	1956.71	1955.91	0.16	ppb
Beryllium	9-1	46.95	47.77	46.60	47.11	1.28	ppb
Bismuth	209-1				103		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	40.05	40.26	41.82	40.71	2.38	ppb
Cadmium	106-1	39.72	39.66	38.96	39.45	1.08	ppb
Cadmium	111-1	43.69	43.75	44.12	43.85	0.53	ppb
Calcium	43-1	27663.32	27363.33	28156.68	27727.78	1.44	ppb
Calcium	44-1	25858.18	25677.45	26469.81	26001.81	1.60	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	383.64	367.16	378.06	376.29	2.23	ppb
Cobalt	59-2	491.05	483.94	498.46	491.15	1.48	ppb
Copper	63-2	279.43	274.41	278.88	277.57	0.99	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				104		%
Indium	115-1				98		%
Indium	115-2				99		%
Iron	56-2	54432.17	53172.54	54262.77	53955.83	1.27	ppb
Iron	57-2	52890.06	52378.05	53450.43	52906.18	1.01	ppb
Iron	54-2	54543.75	53649.63	55091.18	54428.19	1.34	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:24:50 DataFile Name : 037SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	19.97	20.24	20.49	20.23	1.29	ppb
Lead	207-1	19.55	19.49	20.03	19.69	1.52	ppb
Lead	208-1	19.79	19.87	20.33	19.99	1.46	ppb
Lithium	6-1				98		%
Magnesium	24-2	13964.93	13545.19	13896.52	13802.22	1.63	ppb
Manganese	55-2	2222.50	2155.09	2180.76	2186.12	1.56	ppb
Molybdenum	94-1	6.96	7.10	7.20	7.09	1.72	ppb
Molybdenum	95-1	0.75	0.79	0.78	0.78	2.59	ppb
Molybdenum	96-1	1.86	1.81	1.89	1.85	2.02	ppb
Molybdenum	97-1	0.75	0.76	0.78	0.76	2.27	ppb
Molybdenum	98-1	0.73	0.73	0.79	0.75	4.87	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	564.86	556.60	568.12	563.19	1.05	ppb
Phosphorus	31-2	496.53	484.39	499.40	493.44	1.61	ppb
Potassium	39-2	1243.88	1213.73	1244.06	1233.89	1.41	ppb
Rhodium	103-1				94		%
Rhodium	103-2				100		%
Scandium	45-1				96		%
Scandium	45-2				103		%
Selenium	82-1	97.70	101.11	101.21	100.01	2.00	ppb
Selenium	77-2	98.00	105.58	112.00	105.19	6.66	ppb
Selenium	78-2	100.71	99.66	98.90	99.76	0.91	ppb
Silicon	28-1	430.61	430.91	441.15	434.22	1.38	ppb
Silver	107-1	37.96	37.73	38.04	37.91	0.43	ppb
Silver	109-1	37.41	37.30	37.52	37.41	0.30	ppb
Sodium	23-2	58.30	58.46	60.41	59.06	1.98	ppb
Strontium	86-1	37.69	38.30	38.32	38.11	0.94	ppb
Strontium	88-1	37.68	37.97	38.57	38.07	1.19	ppb
Sulfur	34-1	-1378.72	-1399.22	-1276.04	-1351.32		ppb
Terbium	159-1				103		%
Terbium	159-2				105		%
Thallium	203-1	38.07	37.66	38.90	38.21	1.65	ppb
Thallium	205-1	37.91	37.70	38.39	38.00	0.94	ppb
Tin	118-1	0.52	0.54	0.53	0.53	1.60	ppb
Titanium	47-1	457.06	457.64	471.50	462.07	1.77	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-03MSDLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4H2SRE

Dilution Factor :

4

Date & Time Acquired :

2024-11-02 19:24:50

DataFile Name :

037SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.30	0.31	0.33	0.31	4.17	ppb
Vanadium	51-2	554.82	549.98	554.56	553.12	0.49	ppb
Yttrium	89-1				98		%
Yttrium	89-2				102		%
Zinc	66-2	500.81	494.25	503.16	499.41	0.92	ppb
Zirconium	90-1	3.78	3.83	3.88	3.83	1.33	ppb
Zirconium	91-1	3.76	3.98	3.93	3.89	2.90	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-06 Instrumnet Name : P7
Client Sample ID : MYE4H3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:27:53 DataFile Name : 038SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	51080.08	52253.17	51671.11	51668.12	1.14	ppb
Antimony	121-1	5.00	5.11	5.12	5.08	1.40	ppb
Arsenic	75-2	120.29	119.55	119.17	119.67	0.48	ppb
Barium	135-1	209.36	215.29	216.18	213.61	1.73	ppb
Barium	137-1	211.34	216.49	217.04	214.95	1.46	ppb
Beryllium	9-1	1.05	1.13	1.21	1.13	7.01	ppb
Bismuth	209-1				100		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.78	2.62	2.22	2.54	11.26	ppb
Cadmium	106-1	1.85	2.38	1.86	2.03	14.88	ppb
Cadmium	111-1	1.88	1.89	1.84	1.87	1.35	ppb
Calcium	43-1	16627.80	16830.34	17117.14	16858.43	1.46	ppb
Calcium	44-1	15747.08	15569.68	16016.93	15777.90	1.43	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	53.05	54.36	53.85	53.75	1.22	ppb
Cobalt	59-2	59.65	60.95	60.11	60.24	1.10	ppb
Copper	63-2	187.15	189.64	188.25	188.35	0.66	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				106		%
Indium	115-1				98		%
Indium	115-2				102		%
Iron	56-2	115373.86	117510.04	116970.27	116618.06	0.95	ppb
Iron	57-2	114805.70	116469.32	116948.45	116074.49	0.97	ppb
Iron	54-2	117029.13	119379.90	117586.00	117998.34	1.04	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-06 Instrumnet Name : P7
Client Sample ID : MYE4H3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:27:53 DataFile Name : 038SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	28.18	28.43	27.91	28.17	0.92	ppb
Lead	207-1	27.49	27.62	27.58	27.56	0.24	ppb
Lead	208-1	27.42	27.68	27.45	27.52	0.50	ppb
Lithium	6-1				97		%
Magnesium	24-2	20159.43	20344.89	20180.26	20228.19	0.50	ppb
Manganese	55-2	2679.80	2737.71	2687.70	2701.74	1.16	ppb
Molybdenum	94-1	10.21	10.63	10.48	10.44	2.01	ppb
Molybdenum	95-1	1.55	1.58	1.56	1.56	0.85	ppb
Molybdenum	96-1	3.33	3.31	3.42	3.36	1.74	ppb
Molybdenum	97-1	1.43	1.54	1.58	1.52	5.32	ppb
Molybdenum	98-1	1.50	1.52	1.49	1.50	0.88	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	41.73	42.18	42.22	42.04	0.65	ppb
Phosphorus	31-2	1306.74	1300.26	1294.46	1300.48	0.47	ppb
Potassium	39-2	2416.72	2500.65	2435.16	2450.84	1.80	ppb
Rhodium	103-1				95		%
Rhodium	103-2				104		%
Scandium	45-1				101		%
Scandium	45-2				113		%
Selenium	82-1	-0.22	0.55	0.24	0.19	206.82	ppb
Selenium	77-2	10.70	10.65	8.26	9.87	14.14	ppb
Selenium	78-2	0.73	3.18	1.77	1.89	64.76	ppb
Silicon	28-1	783.24	783.89	792.91	786.68	0.69	ppb
Silver	107-1	0.87	0.87	0.83	0.86	2.91	ppb
Silver	109-1	0.86	0.86	0.86	0.86	0.18	ppb
Sodium	23-2	561.63	570.90	569.67	567.40	0.89	ppb
Strontium	86-1	98.27	100.58	100.68	99.85	1.36	ppb
Strontium	88-1	100.36	101.48	102.95	101.60	1.28	ppb
Sulfur	34-1	-1507.37	-1526.45	-1484.40	-1506.07		ppb
Terbium	159-1				102		%
Terbium	159-2				105		%
Thallium	203-1	0.23	0.23	0.21	0.22	5.51	ppb
Thallium	205-1	0.22	0.22	0.22	0.22	0.91	ppb
Tin	118-1	1.84	1.91	1.95	1.90	2.91	ppb
Titanium	47-1	1979.32	2012.48	2037.35	2009.72	1.45	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-06 Instrumnet Name : P7
Client Sample ID : MYE4H3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:27:53 DataFile Name : 038SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.50	0.51	0.50	0.50	1.16	ppb
Vanadium	51-2	197.25	201.69	199.93	199.63	1.12	ppb
Yttrium	89-1				103		%
Yttrium	89-2				109		%
Zinc	66-2	460.62	469.63	467.40	465.88	1.01	ppb
Zirconium	90-1	5.38	5.48	5.53	5.46	1.42	ppb
Zirconium	91-1	5.39	5.56	5.57	5.51	1.84	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-07 Instrumnet Name : P7
Client Sample ID : MYE4H4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:30:59 DataFile Name : 039SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	43444.12	41198.86	42218.99	42287.32	2.66	ppb
Antimony	121-1	6.75	6.86	6.93	6.85	1.33	ppb
Arsenic	75-2	205.86	196.88	201.92	201.55	2.23	ppb
Barium	135-1	596.97	590.93	600.92	596.28	0.84	ppb
Barium	137-1	592.13	588.85	604.73	595.24	1.41	ppb
Beryllium	9-1	1.76	1.78	1.94	1.83	5.37	ppb
Bismuth	209-1				101		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	4.71	4.29	4.49	4.50	4.71	ppb
Cadmium	106-1	4.59	5.24	4.58	4.80	7.89	ppb
Cadmium	111-1	3.31	3.35	3.41	3.36	1.39	ppb
Calcium	43-1	7712.37	7790.34	7805.82	7769.51	0.64	ppb
Calcium	44-1	7259.66	7278.17	7317.60	7285.14	0.41	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	111.31	105.56	109.11	108.66	2.67	ppb
Cobalt	59-2	63.17	60.23	61.88	61.76	2.39	ppb
Copper	63-2	393.99	375.85	387.25	385.70	2.38	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				107		%
Indium	115-1				100		%
Indium	115-2				102		%
Iron	56-2	135424.03	127306.46	131426.29	131385.60	3.09	ppb
Iron	57-2	132762.61	127140.85	128884.62	129596.02	2.22	ppb
Iron	54-2	135916.21	128144.61	133086.13	132382.32	2.97	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-07 Instrumnet Name : P7
Client Sample ID : MYE4H4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:30:59 DataFile Name : 039SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	136.21	136.46	138.95	137.21	1.10	ppb
Lead	207-1	136.30	137.71	138.71	137.57	0.88	ppb
Lead	208-1	146.64	147.51	149.95	148.04	1.16	ppb
Lithium	6-1				97		%
Magnesium	24-2	22458.16	21233.17	22032.41	21907.91	2.84	ppb
Manganese	55-2	3609.81	3396.23	3491.43	3499.16	3.06	ppb
Molybdenum	94-1	29.18	29.69	30.36	29.74	1.99	ppb
Molybdenum	95-1	1.55	1.58	1.61	1.58	1.89	ppb
Molybdenum	96-1	5.59	5.59	5.61	5.60	0.21	ppb
Molybdenum	97-1	1.54	1.53	1.58	1.55	1.73	ppb
Molybdenum	98-1	1.45	1.49	1.50	1.48	1.61	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	106.51	99.90	104.16	103.52	3.24	ppb
Phosphorus	31-2	2134.62	2117.17	2141.38	2131.06	0.59	ppb
Potassium	39-2	7869.77	7350.21	7608.16	7609.38	3.41	ppb
Rhodium	103-1				97		%
Rhodium	103-2				103		%
Scandium	45-1				106		%
Scandium	45-2				112		%
Selenium	82-1	0.80	0.88	0.81	0.83	5.11	ppb
Selenium	77-2	18.60	18.15	17.65	18.13	2.60	ppb
Selenium	78-2	5.30	4.60	4.79	4.90	7.33	ppb
Silicon	28-1	1315.39	1331.52	1335.89	1327.60	0.81	ppb
Silver	107-1	0.80	0.78	0.82	0.80	2.52	ppb
Silver	109-1	0.78	0.80	0.81	0.80	2.08	ppb
Sodium	23-2	141.65	130.84	138.15	136.88	4.03	ppb
Strontium	86-1	80.13	81.10	81.79	81.01	1.03	ppb
Strontium	88-1	81.72	83.46	85.91	83.70	2.51	ppb
Sulfur	34-1	-1424.73	-1510.47	-1472.52	-1469.24		ppb
Terbium	159-1				104		%
Terbium	159-2				106		%
Thallium	203-1	0.48	0.50	0.48	0.48	2.17	ppb
Thallium	205-1	0.47	0.46	0.47	0.47	1.29	ppb
Tin	118-1	3.03	3.04	3.16	3.08	2.28	ppb
Titanium	47-1	1013.00	1012.41	1025.92	1017.11	0.75	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-07 Instrumnet Name : P7
Client Sample ID : MYE4H4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:30:59 DataFile Name : 039SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.16	1.13	1.15	1.15	1.24	ppb
Vanadium	51-2	264.06	249.63	258.82	257.50	2.84	ppb
Yttrium	89-1				106		%
Yttrium	89-2				110		%
Zinc	66-2	390.60	370.83	384.77	382.07	2.66	ppb
Zirconium	90-1	16.86	16.92	17.18	16.99	1.01	ppb
Zirconium	91-1	16.81	17.28	17.37	17.16	1.74	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-08 Instrumnet Name : P7
Client Sample ID : MYE4H5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:34:02 DataFile Name : 040SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	52430.58	58390.01	52767.65	54529.41	6.14	ppb
Antimony	121-1	2.28	2.33	2.35	2.32	1.68	ppb
Arsenic	75-2	138.83	155.01	138.66	144.17	6.51	ppb
Barium	135-1	741.69	756.43	779.56	759.23	2.51	ppb
Barium	137-1	733.59	747.44	767.55	749.53	2.28	ppb
Beryllium	9-1	2.20	1.99	2.21	2.13	5.92	ppb
Bismuth	209-1				102		%
Bismuth	209-2				98		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	3.07	2.23	2.69	2.66	15.80	ppb
Cadmium	106-1	2.93	2.95	3.56	3.15	11.32	ppb
Cadmium	111-1	0.93	0.95	0.97	0.95	2.33	ppb
Calcium	43-1	11616.93	11812.47	11890.84	11773.41	1.20	ppb
Calcium	44-1	11020.73	11043.01	11190.76	11084.83	0.83	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	118.87	132.66	120.92	124.15	5.99	ppb
Cobalt	59-2	53.52	59.57	54.92	56.00	5.65	ppb
Copper	63-2	202.70	224.53	206.94	211.39	5.47	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				102		%
Indium	115-1				99		%
Indium	115-2				97		%
Iron	56-2	112124.75	126431.64	115758.98	118105.12	6.30	ppb
Iron	57-2	111611.99	125399.07	114064.15	117025.07	6.28	ppb
Iron	54-2	113452.72	125849.96	115130.23	118144.31	5.69	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-08 Instrumnet Name : P7
Client Sample ID : MYE4H5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:34:02 DataFile Name : 040SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	39.15	40.96	40.74	40.29	2.45	ppb
Lead	207-1	38.07	39.98	39.43	39.16	2.51	ppb
Lead	208-1	38.57	40.05	40.11	39.58	2.20	ppb
Lithium	6-1				96		%
Magnesium	24-2	23334.47	26139.99	23765.99	24413.48	6.19	ppb
Manganese	55-2	2321.47	2629.34	2393.42	2448.08	6.58	ppb
Molybdenum	94-1	30.39	31.21	31.46	31.02	1.80	ppb
Molybdenum	95-1	0.98	1.02	1.01	1.00	1.86	ppb
Molybdenum	96-1	4.99	5.12	5.15	5.09	1.75	ppb
Molybdenum	97-1	0.99	1.02	1.01	1.01	1.08	ppb
Molybdenum	98-1	0.98	0.98	0.97	0.97	0.59	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	121.24	133.97	122.91	126.04	5.49	ppb
Phosphorus	31-2	1983.76	2183.47	2085.40	2084.21	4.79	ppb
Potassium	39-2	6793.75	7627.01	7023.77	7148.17	6.02	ppb
Rhodium	103-1				96		%
Rhodium	103-2				98		%
Scandium	45-1				106		%
Scandium	45-2				109		%
Selenium	82-1	0.75	0.86	0.62	0.74	16.60	ppb
Selenium	77-2	20.41	25.19	20.35	21.98	12.62	ppb
Selenium	78-2	4.55	5.91	4.33	4.93	17.37	ppb
Silicon	28-1	1391.34	1417.69	1434.72	1414.59	1.54	ppb
Silver	107-1	0.35	0.35	0.37	0.36	3.22	ppb
Silver	109-1	0.35	0.35	0.34	0.35	1.38	ppb
Sodium	23-2	129.45	149.19	136.23	138.29	7.25	ppb
Strontium	86-1	115.22	118.40	118.86	117.49	1.69	ppb
Strontium	88-1	119.37	121.11	121.68	120.72	1.00	ppb
Sulfur	34-1	-1682.25	-1739.14	-1679.77	-1700.39		ppb
Terbium	159-1				103		%
Terbium	159-2				104		%
Thallium	203-1	0.40	0.42	0.41	0.41	2.36	ppb
Thallium	205-1	0.40	0.41	0.41	0.41	1.38	ppb
Tin	118-1	2.58	2.67	2.74	2.67	2.91	ppb
Titanium	47-1	1135.83	1164.67	1164.40	1154.96	1.43	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-08

Instrumnet Name :

P7

Client Sample ID :

MYE4H5

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 19:34:02

DataFile Name :

040SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.57	1.62	1.63	1.60	2.04	ppb
Vanadium	51-2	223.78	249.25	229.00	234.01	5.75	ppb
Yttrium	89-1				107		%
Yttrium	89-2				107		%
Zinc	66-2	198.98	221.64	201.84	207.49	5.95	ppb
Zirconium	90-1	17.82	18.28	18.37	18.16	1.65	ppb
Zirconium	91-1	17.47	18.54	18.41	18.14	3.22	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-09 Instrumnet Name : P7
Client Sample ID : MYE4H6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:37:05 DataFile Name : 041SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	35715.75	35275.38	35405.08	35465.40	0.64	ppb
Antimony	121-1	3.85	4.03	4.02	3.97	2.58	ppb
Arsenic	75-2	177.31	177.90	178.93	178.05	0.46	ppb
Barium	135-1	437.55	449.77	460.43	449.25	2.55	ppb
Barium	137-1	473.92	483.22	497.20	484.78	2.42	ppb
Beryllium	9-1	1.48	1.40	1.44	1.44	2.86	ppb
Bismuth	209-1				102		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.72	3.82	4.11	3.55	20.59	ppb
Cadmium	106-1	4.03	3.96	4.27	4.08	4.00	ppb
Cadmium	111-1	2.32	2.39	2.38	2.37	1.52	ppb
Calcium	43-1	7846.28	8001.66	7931.74	7926.56	0.98	ppb
Calcium	44-1	7367.40	7512.33	7501.23	7460.32	1.08	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	92.12	91.18	91.83	91.71	0.52	ppb
Cobalt	59-2	44.84	44.88	45.39	45.04	0.68	ppb
Copper	63-2	258.37	261.03	258.75	259.38	0.56	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				105		%
Indium	115-1				99		%
Indium	115-2				101		%
Iron	56-2	105908.86	106095.15	106195.42	106066.47	0.14	ppb
Iron	57-2	104556.71	105028.90	104455.28	104680.30	0.29	ppb
Iron	54-2	106254.76	105370.10	107380.96	106335.27	0.95	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-09 Instrumnet Name : P7
Client Sample ID : MYE4H6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:37:05 DataFile Name : 041SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	83.47	83.19	84.71	83.79	0.96	ppb
Lead	207-1	83.51	83.27	84.59	83.79	0.84	ppb
Lead	208-1	83.20	82.88	91.59	85.89	5.75	ppb
Lithium	6-1				97		%
Magnesium	24-2	17029.13	17131.36	16880.22	17013.57	0.74	ppb
Manganese	55-2	1883.62	1893.23	1896.14	1891.00	0.35	ppb
Molybdenum	94-1	28.85	29.60	29.91	29.45	1.86	ppb
Molybdenum	95-1	1.45	1.45	1.46	1.46	0.51	ppb
Molybdenum	96-1	5.06	5.29	5.33	5.22	2.78	ppb
Molybdenum	97-1	1.43	1.38	1.42	1.41	1.77	ppb
Molybdenum	98-1	1.35	1.40	1.39	1.38	1.95	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	89.32	88.37	88.97	88.89	0.54	ppb
Phosphorus	31-2	1916.88	1926.08	1976.20	1939.72	1.65	ppb
Potassium	39-2	6979.11	7020.14	6969.25	6989.50	0.39	ppb
Rhodium	103-1				96		%
Rhodium	103-2				101		%
Scandium	45-1				101		%
Scandium	45-2				107		%
Selenium	82-1	1.09	0.88	1.18	1.05	14.96	ppb
Selenium	77-2	19.10	21.89	13.29	18.09	24.23	ppb
Selenium	78-2	4.81	4.42	4.44	4.56	4.88	ppb
Silicon	28-1	1209.37	1256.61	1240.70	1235.56	1.95	ppb
Silver	107-1	0.51	0.55	0.56	0.54	5.53	ppb
Silver	109-1	0.53	0.55	0.54	0.54	1.49	ppb
Sodium	23-2	130.31	130.29	129.88	130.16	0.19	ppb
Strontium	86-1	77.98	78.14	79.40	78.51	0.99	ppb
Strontium	88-1	79.15	80.90	80.16	80.07	1.10	ppb
Sulfur	34-1	-1542.18	-1491.36	-1438.31	-1490.62		ppb
Terbium	159-1				103		%
Terbium	159-2				105		%
Thallium	203-1	0.39	0.38	0.40	0.39	3.75	ppb
Thallium	205-1	0.38	0.39	0.38	0.39	1.87	ppb
Tin	118-1	2.81	2.84	3.04	2.90	4.37	ppb
Titanium	47-1	1010.29	1050.91	1057.22	1039.47	2.45	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-09

Instrumnet Name :

P7

Client Sample ID :

MYE4H6

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 19:37:05

DataFile Name :

041SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.03	1.01	1.05	1.03	1.99	ppb
Vanadium	51-2	199.89	198.82	201.08	199.93	0.56	ppb
Yttrium	89-1				103		%
Yttrium	89-2				108		%
Zinc	66-2	318.74	314.91	317.12	316.93	0.61	ppb
Zirconium	90-1	16.60	17.23	17.29	17.04	2.25	ppb
Zirconium	91-1	16.71	17.10	17.32	17.04	1.82	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-10 Instrumnet Name : P7
Client Sample ID : MYE4H8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:40:10 DataFile Name : 042SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	36627.48	36572.89	36590.85	36597.07	0.08	ppb
Antimony	121-1	3.27	3.32	3.31	3.30	0.88	ppb
Arsenic	75-2	123.30	125.61	123.96	124.29	0.96	ppb
Barium	135-1	462.34	464.75	473.32	466.81	1.24	ppb
Barium	137-1	500.94	505.76	502.28	502.99	0.49	ppb
Beryllium	9-1	1.46	1.30	1.49	1.42	7.25	ppb
Bismuth	209-1				101		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	3.84	3.37	4.06	3.76	9.45	ppb
Cadmium	106-1	3.50	3.67	2.79	3.32	13.98	ppb
Cadmium	111-1	2.59	2.69	2.51	2.60	3.42	ppb
Calcium	43-1	8670.29	8785.63	8763.64	8739.85	0.70	ppb
Calcium	44-1	8148.00	8194.85	8286.07	8209.64	0.86	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	93.27	92.49	93.23	93.00	0.47	ppb
Cobalt	59-2	46.71	46.17	46.34	46.41	0.60	ppb
Copper	63-2	372.00	371.40	363.26	368.88	1.32	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				104		%
Indium	115-1				98		%
Indium	115-2				99		%
Iron	56-2	95070.60	95037.96	95873.92	95327.49	0.50	ppb
Iron	57-2	94342.75	93269.08	94237.50	93949.78	0.63	ppb
Iron	54-2	95636.61	95626.27	95280.25	95514.38	0.21	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-10 Instrumnet Name : P7
Client Sample ID : MYE4H8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:40:10 DataFile Name : 042SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	84.24	84.59	85.29	84.71	0.63	ppb
Lead	207-1	83.99	84.78	84.29	84.35	0.47	ppb
Lead	208-1	83.52	84.27	84.38	84.05	0.56	ppb
Lithium	6-1				95		%
Magnesium	24-2	17472.71	17571.14	17660.24	17568.03	0.53	ppb
Manganese	55-2	2219.42	2219.72	2210.45	2216.53	0.24	ppb
Molybdenum	94-1	22.37	22.42	22.44	22.41	0.17	ppb
Molybdenum	95-1	1.38	1.36	1.44	1.39	2.88	ppb
Molybdenum	96-1	4.39	4.29	4.42	4.36	1.56	ppb
Molybdenum	97-1	1.30	1.39	1.38	1.36	3.70	ppb
Molybdenum	98-1	1.34	1.30	1.35	1.33	1.83	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	89.78	88.60	89.47	89.28	0.69	ppb
Phosphorus	31-2	1948.43	1937.61	1902.25	1929.43	1.25	ppb
Potassium	39-2	7599.07	7604.10	7576.68	7593.28	0.19	ppb
Rhodium	103-1				95		%
Rhodium	103-2				100		%
Scandium	45-1				99		%
Scandium	45-2				106		%
Selenium	82-1	1.04	1.02	0.83	0.96	11.86	ppb
Selenium	77-2	15.38	16.56	15.56	15.83	4.01	ppb
Selenium	78-2	2.13	4.24	3.76	3.38	32.73	ppb
Silicon	28-1	1331.91	1337.00	1350.46	1339.79	0.72	ppb
Silver	107-1	0.70	0.68	0.69	0.69	1.14	ppb
Silver	109-1	0.69	0.70	0.69	0.69	0.34	ppb
Sodium	23-2	142.55	140.91	142.15	141.87	0.61	ppb
Strontium	86-1	89.00	89.54	90.15	89.56	0.64	ppb
Strontium	88-1	91.27	92.49	91.41	91.72	0.73	ppb
Sulfur	34-1	-1536.87	-1532.69	-1589.10	-1552.88		ppb
Terbium	159-1				102		%
Terbium	159-2				104		%
Thallium	203-1	0.32	0.35	0.34	0.34	3.90	ppb
Thallium	205-1	0.32	0.33	0.33	0.33	1.47	ppb
Tin	118-1	2.97	3.00	3.01	2.99	0.77	ppb
Titanium	47-1	964.53	951.59	970.87	962.33	1.02	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-10

Instrumnet Name :

P7

Client Sample ID :

MYE4H8

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 19:40:10

DataFile Name :

042SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.83	0.86	0.85	0.85	2.08	ppb
Vanadium	51-2	178.40	178.66	179.02	178.69	0.17	ppb
Yttrium	89-1				102		%
Yttrium	89-2				107		%
Zinc	66-2	332.44	332.58	330.07	331.70	0.43	ppb
Zirconium	90-1	12.70	12.78	12.91	12.80	0.85	ppb
Zirconium	91-1	12.75	12.83	13.11	12.90	1.47	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-11 Instrumnet Name : P7
Client Sample ID : MYE4H9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:43:13 DataFile Name : 043SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	44007.74	44805.98	43940.17	44251.30	1.09	ppb
Antimony	121-1	4.17	4.33	4.45	4.32	3.27	ppb
Arsenic	75-2	138.72	134.25	136.61	136.53	1.64	ppb
Barium	135-1	689.65	717.81	744.65	717.37	3.83	ppb
Barium	137-1	680.76	697.88	738.84	705.82	4.23	ppb
Beryllium	9-1	1.74	1.86	1.90	1.84	4.49	ppb
Bismuth	209-1				99		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	3.30	2.79	3.22	3.10	8.81	ppb
Cadmium	106-1	3.23	3.51	2.64	3.13	14.15	ppb
Cadmium	111-1	1.37	1.53	1.44	1.45	5.29	ppb
Calcium	43-1	9312.14	9620.82	9987.22	9640.06	3.51	ppb
Calcium	44-1	8733.41	8940.08	9271.80	8981.76	3.02	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	122.37	123.20	121.20	122.26	0.82	ppb
Cobalt	59-2	57.03	57.06	56.81	56.96	0.24	ppb
Copper	63-2	235.74	236.83	233.53	235.37	0.71	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				105		%
Indium	115-1				96		%
Indium	115-2				100		%
Iron	56-2	116059.40	116360.30	115696.35	116038.68	0.29	ppb
Iron	57-2	113889.99	113768.65	113172.73	113610.46	0.34	ppb
Iron	54-2	115366.43	116199.92	114565.61	115377.32	0.71	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-11 Instrumnet Name : P7
Client Sample ID : MYE4H9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:43:13 DataFile Name : 043SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	55.31	55.40	59.63	56.78	4.34	ppb
Lead	207-1	54.95	54.21	59.45	56.20	5.04	ppb
Lead	208-1	54.90	54.53	59.05	56.16	4.47	ppb
Lithium	6-1				93		%
Magnesium	24-2	20189.66	20500.81	20034.87	20241.78	1.17	ppb
Manganese	55-2	3257.99	3264.07	3238.87	3253.64	0.40	ppb
Molybdenum	94-1	27.81	29.01	30.14	28.99	4.02	ppb
Molybdenum	95-1	1.70	1.72	1.82	1.75	3.50	ppb
Molybdenum	96-1	5.20	5.57	5.76	5.51	5.19	ppb
Molybdenum	97-1	1.60	1.73	1.82	1.72	6.27	ppb
Molybdenum	98-1	1.63	1.64	1.69	1.66	2.02	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	110.52	110.22	108.33	109.69	1.08	ppb
Phosphorus	31-2	1863.81	1881.75	1810.79	1852.12	1.99	ppb
Potassium	39-2	8002.96	7898.29	7954.27	7951.84	0.66	ppb
Rhodium	103-1				93		%
Rhodium	103-2				101		%
Scandium	45-1				100		%
Scandium	45-2				109		%
Selenium	82-1	1.55	0.80	1.14	1.16	32.00	ppb
Selenium	77-2	22.20	24.58	20.09	22.29	10.09	ppb
Selenium	78-2	4.92	4.62	6.40	5.32	17.93	ppb
Silicon	28-1	480.66	499.02	522.95	500.88	4.23	ppb
Silver	107-1	0.41	0.43	0.45	0.43	3.94	ppb
Silver	109-1	0.40	0.42	0.43	0.42	3.31	ppb
Sodium	23-2	128.26	130.31	128.01	128.86	0.98	ppb
Strontium	86-1	98.48	102.18	105.38	102.01	3.39	ppb
Strontium	88-1	100.30	103.52	108.95	104.26	4.19	ppb
Sulfur	34-1	-1631.82	-1570.07	-1335.87	-1512.59		ppb
Terbium	159-1				102		%
Terbium	159-2				104		%
Thallium	203-1	0.46	0.47	0.51	0.48	5.59	ppb
Thallium	205-1	0.46	0.43	0.50	0.47	6.90	ppb
Tin	118-1	2.83	2.60	2.75	2.73	4.28	ppb
Titanium	47-1	950.69	965.78	1035.98	984.15	4.63	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-11 Instrumnet Name : P7
Client Sample ID : MYE4H9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:43:13 DataFile Name : 043SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.33	1.33	1.45	1.37	4.99	ppb
Vanadium	51-2	219.85	222.58	217.42	219.95	1.17	ppb
Yttrium	89-1				103		%
Yttrium	89-2				109		%
Zinc	66-2	237.54	236.14	235.76	236.48	0.40	ppb
Zirconium	90-1	15.87	16.58	17.04	16.50	3.57	ppb
Zirconium	91-1	15.85	16.37	17.21	16.48	4.18	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-12 Instrumnet Name : P7
Client Sample ID : MYE4J0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:46:19 DataFile Name : 044SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	34792.54	34783.95	34894.70	34823.73	0.18	ppb
Antimony	121-1	3.55	3.22	3.15	3.31	6.48	ppb
Arsenic	75-2	129.73	129.19	133.09	130.67	1.62	ppb
Barium	135-1	472.87	423.63	428.13	441.54	6.17	ppb
Barium	137-1	514.98	450.87	467.64	477.83	6.96	ppb
Beryllium	9-1	1.59	1.49	1.58	1.55	3.69	ppb
Bismuth	209-1				96		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	4.15	3.28	3.35	3.59	13.53	ppb
Cadmium	106-1	5.08	4.07	2.69	3.95	30.49	ppb
Cadmium	111-1	2.89	2.60	2.51	2.67	7.47	ppb
Calcium	43-1	7711.93	6973.94	7159.19	7281.69	5.27	ppb
Calcium	44-1	7269.29	6485.89	6706.92	6820.70	5.92	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	86.55	88.28	86.79	87.21	1.08	ppb
Cobalt	59-2	44.43	44.33	44.33	44.37	0.13	ppb
Copper	63-2	294.56	293.01	293.59	293.72	0.27	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				98		%
Holmium	165-2				105		%
Indium	115-1				94		%
Indium	115-2				99		%
Iron	56-2	93954.45	94792.39	95924.38	94890.41	1.04	ppb
Iron	57-2	94038.62	94390.06	94529.14	94319.28	0.27	ppb
Iron	54-2	94488.98	95999.06	94520.37	95002.80	0.91	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-12 Instrumnet Name : P7
Client Sample ID : MYE4J0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:46:19 DataFile Name : 044SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	98.82	90.42	91.07	93.44	5.00	ppb
Lead	207-1	97.21	89.26	91.16	92.54	4.48	ppb
Lead	208-1	105.07	95.93	98.43	99.81	4.73	ppb
Lithium	6-1				89		%
Magnesium	24-2	17598.91	17481.36	17279.83	17453.37	0.92	ppb
Manganese	55-2	1907.04	1912.50	1916.82	1912.12	0.26	ppb
Molybdenum	94-1	20.77	18.99	19.50	19.75	4.64	ppb
Molybdenum	95-1	1.39	1.27	1.20	1.29	7.48	ppb
Molybdenum	96-1	4.22	3.87	3.96	4.01	4.59	ppb
Molybdenum	97-1	1.39	1.24	1.24	1.29	6.75	ppb
Molybdenum	98-1	1.30	1.23	1.19	1.24	4.38	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	81.93	83.10	83.13	82.72	0.83	ppb
Phosphorus	31-2	1819.73	1859.01	1798.44	1825.73	1.68	ppb
Potassium	39-2	6752.39	6848.04	6847.03	6815.82	0.81	ppb
Rhodium	103-1				91		%
Rhodium	103-2				101		%
Scandium	45-1				96		%
Scandium	45-2				107		%
Selenium	82-1	1.20	0.86	1.07	1.04	16.64	ppb
Selenium	77-2	18.61	14.88	17.92	17.14	11.59	ppb
Selenium	78-2	2.85	4.14	3.73	3.57	18.50	ppb
Silicon	28-1	1089.60	986.56	1006.65	1027.60	5.32	ppb
Silver	107-1	0.71	0.61	0.64	0.65	7.66	ppb
Silver	109-1	0.69	0.63	0.61	0.64	6.95	ppb
Sodium	23-2	123.71	123.78	124.04	123.85	0.14	ppb
Strontium	86-1	77.11	69.87	71.64	72.87	5.18	ppb
Strontium	88-1	79.84	71.99	72.43	74.75	5.90	ppb
Sulfur	34-1	-1170.11	-1565.63	-1507.30	-1414.35		ppb
Terbium	159-1				98		%
Terbium	159-2				105		%
Thallium	203-1	0.40	0.37	0.39	0.39	3.79	ppb
Thallium	205-1	0.39	0.37	0.38	0.38	3.23	ppb
Tin	118-1	3.23	3.05	2.91	3.06	5.37	ppb
Titanium	47-1	915.49	827.82	851.62	864.98	5.24	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-12

Instrumnet Name :

P7

Client Sample ID :

MYE4J0

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 19:46:19

DataFile Name :

044SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.09	1.00	0.99	1.03	5.45	ppb
Vanadium	51-2	185.39	185.23	184.87	185.17	0.15	ppb
Yttrium	89-1				99		%
Yttrium	89-2				107		%
Zinc	66-2	310.87	311.21	314.61	312.23	0.66	ppb
Zirconium	90-1	11.84	10.63	10.95	11.14	5.64	ppb
Zirconium	91-1	11.79	10.64	11.01	11.15	5.24	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-13 Instrumnet Name : P7
Client Sample ID : MYE4J1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:49:23 DataFile Name : 045SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	33979.75	33879.14	33631.29	33830.06	0.53	ppb
Antimony	121-1	3.03	3.07	3.39	3.16	6.37	ppb
Arsenic	75-2	132.37	131.40	133.01	132.26	0.61	ppb
Barium	135-1	391.20	397.91	433.65	407.59	5.60	ppb
Barium	137-1	423.29	428.22	479.19	443.57	6.98	ppb
Beryllium	9-1	1.24	1.38	1.55	1.39	11.28	ppb
Bismuth	209-1				99		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.65	3.62	3.07	3.12	15.57	ppb
Cadmium	106-1	2.77	2.54	2.60	2.64	4.55	ppb
Cadmium	111-1	2.15	2.13	2.25	2.18	3.10	ppb
Calcium	43-1	6851.03	6962.24	7618.16	7143.81	5.80	ppb
Calcium	44-1	6537.24	6521.16	7152.62	6737.01	5.34	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	81.84	82.06	81.95	81.95	0.13	ppb
Cobalt	59-2	44.46	45.00	44.83	44.76	0.61	ppb
Copper	63-2	261.13	261.71	264.21	262.35	0.62	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				107		%
Indium	115-1				96		%
Indium	115-2				101		%
Iron	56-2	90191.52	89209.51	90499.37	89966.80	0.75	ppb
Iron	57-2	88299.17	89829.19	89438.04	89188.80	0.89	ppb
Iron	54-2	89903.54	90611.24	91044.48	90519.75	0.64	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-13 Instrumnet Name : P7
Client Sample ID : MYE4J1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:49:23 DataFile Name : 045SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	73.08	74.44	81.19	76.24	5.70	ppb
Lead	207-1	72.87	74.75	79.36	75.66	4.41	ppb
Lead	208-1	72.50	73.60	79.80	75.30	5.23	ppb
Lithium	6-1				92		%
Magnesium	24-2	15967.71	16016.92	16099.18	16027.94	0.41	ppb
Manganese	55-2	2003.16	2026.17	2025.17	2018.17	0.64	ppb
Molybdenum	94-1	19.61	19.92	21.61	20.38	5.28	ppb
Molybdenum	95-1	1.33	1.34	1.47	1.38	5.59	ppb
Molybdenum	96-1	3.89	4.04	4.52	4.15	7.99	ppb
Molybdenum	97-1	1.30	1.30	1.44	1.35	5.84	ppb
Molybdenum	98-1	1.23	1.25	1.39	1.29	6.74	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	80.32	80.61	80.00	80.31	0.38	ppb
Phosphorus	31-2	1760.19	1792.57	1798.56	1783.77	1.16	ppb
Potassium	39-2	6989.81	7039.18	7031.33	7020.11	0.38	ppb
Rhodium	103-1				92		%
Rhodium	103-2				101		%
Scandium	45-1				96		%
Scandium	45-2				105		%
Selenium	82-1	0.65	1.01	1.15	0.94	27.20	ppb
Selenium	77-2	17.45	19.52	13.27	16.75	18.99	ppb
Selenium	78-2	2.52	2.93	2.83	2.76	7.87	ppb
Silicon	28-1	1178.72	1193.72	1316.57	1229.67	6.15	ppb
Silver	107-1	0.51	0.51	0.58	0.53	7.23	ppb
Silver	109-1	0.52	0.50	0.58	0.54	7.85	ppb
Sodium	23-2	125.03	126.56	124.53	125.38	0.85	ppb
Strontium	86-1	65.90	66.33	72.69	68.31	5.56	ppb
Strontium	88-1	68.55	67.22	74.88	70.22	5.83	ppb
Sulfur	34-1	-1434.83	-1467.59	-1204.03	-1368.82		ppb
Terbium	159-1				100		%
Terbium	159-2				106		%
Thallium	203-1	0.31	0.33	0.34	0.33	4.82	ppb
Thallium	205-1	0.30	0.31	0.33	0.32	4.95	ppb
Tin	118-1	2.82	2.86	3.17	2.95	6.48	ppb
Titanium	47-1	833.59	839.74	927.16	866.83	6.04	ppb

LB Number :LB133265

Operator :Jaswal

Lab Sample ID :P4484-13

Instrumnet Name :P7

Client Sample ID :MYE4J1

Dilution Factor :1

Date & Time Acquired :2024-11-02 19:49:23

DataFile Name :045SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.94	0.96	1.04	0.98	5.25	ppb
Vanadium	51-2	165.51	166.32	167.17	166.34	0.50	ppb
Yttrium	89-1				99		%
Yttrium	89-2				106		%
Zinc	66-2	285.93	288.66	289.41	288.00	0.64	ppb
Zirconium	90-1	11.16	11.24	12.35	11.58	5.74	ppb
Zirconium	91-1	11.29	11.34	12.37	11.67	5.25	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-14 Instrumnet Name : P7
Client Sample ID : MYE4J2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:52:30 DataFile Name : 046SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	40271.24	40818.36	41000.29	40696.63	0.93	ppb
Antimony	121-1	3.59	3.60	3.87	3.69	4.24	ppb
Arsenic	75-2	151.72	149.19	147.94	149.62	1.29	ppb
Barium	135-1	501.68	510.18	534.45	515.43	3.30	ppb
Barium	137-1	559.62	558.15	577.04	564.94	1.86	ppb
Beryllium	9-1	1.59	1.63	1.70	1.64	3.59	ppb
Bismuth	209-1				101		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	4.00	4.32	4.58	4.30	6.78	ppb
Cadmium	106-1	3.77	3.11	5.21	4.03	26.62	ppb
Cadmium	111-1	2.86	2.68	3.03	2.86	6.24	ppb
Calcium	43-1	8398.27	8320.04	8679.13	8465.82	2.23	ppb
Calcium	44-1	7878.00	7960.16	8229.09	8022.42	2.29	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	107.68	108.59	108.37	108.21	0.44	ppb
Cobalt	59-2	56.25	56.73	56.64	56.54	0.45	ppb
Copper	63-2	356.61	358.90	367.28	360.93	1.56	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				105		%
Indium	115-1				96		%
Indium	115-2				100		%
Iron	56-2	112159.86	113437.02	112121.89	112572.92	0.66	ppb
Iron	57-2	111300.08	111565.15	111622.80	111496.01	0.15	ppb
Iron	54-2	112982.50	113373.50	113557.88	113304.63	0.26	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-14 Instrumnet Name : P7
Client Sample ID : MYE4J2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:52:30 DataFile Name : 046SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	107.60	109.79	116.62	111.34	4.23	ppb
Lead	207-1	110.57	110.09	115.86	112.17	2.86	ppb
Lead	208-1	117.75	116.78	123.89	119.48	3.23	ppb
Lithium	6-1				94		%
Magnesium	24-2	20606.01	20904.05	20891.32	20800.46	0.81	ppb
Manganese	55-2	2470.86	2471.32	2468.17	2470.12	0.07	ppb
Molybdenum	94-1	27.93	27.76	29.73	28.47	3.84	ppb
Molybdenum	95-1	1.63	1.59	1.72	1.65	4.08	ppb
Molybdenum	96-1	5.29	5.30	5.47	5.35	1.85	ppb
Molybdenum	97-1	1.58	1.58	1.70	1.62	4.35	ppb
Molybdenum	98-1	1.54	1.52	1.62	1.56	3.51	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	106.10	106.97	105.46	106.18	0.71	ppb
Phosphorus	31-2	2080.87	2063.06	2067.23	2070.39	0.45	ppb
Potassium	39-2	8066.02	8147.37	8084.87	8099.42	0.53	ppb
Rhodium	103-1				93		%
Rhodium	103-2				100		%
Scandium	45-1				99		%
Scandium	45-2				107		%
Selenium	82-1	1.17	1.29	1.00	1.15	12.51	ppb
Selenium	77-2	17.78	18.76	18.00	18.18	2.84	ppb
Selenium	78-2	5.32	3.68	4.34	4.44	18.57	ppb
Silicon	28-1	1152.50	1144.49	1191.56	1162.85	2.17	ppb
Silver	107-1	0.69	0.73	0.76	0.73	4.30	ppb
Silver	109-1	0.72	0.69	0.74	0.72	3.43	ppb
Sodium	23-2	131.96	135.63	135.01	134.20	1.47	ppb
Strontium	86-1	88.72	90.12	94.25	91.03	3.16	ppb
Strontium	88-1	91.04	94.08	98.04	94.39	3.72	ppb
Sulfur	34-1	-1503.10	-1438.19	-1305.08	-1415.46		ppb
Terbium	159-1				103		%
Terbium	159-2				105		%
Thallium	203-1	0.46	0.47	0.47	0.46	1.35	ppb
Thallium	205-1	0.45	0.47	0.48	0.47	4.08	ppb
Tin	118-1	2.96	3.05	3.19	3.07	3.77	ppb
Titanium	47-1	1123.06	1116.32	1167.64	1135.67	2.46	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-14

Instrumnet Name :

P7

Client Sample ID :

MYE4J2

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 19:52:30

DataFile Name :

046SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.15	1.15	1.19	1.16	2.10	ppb
Vanadium	51-2	220.60	222.41	221.93	221.65	0.42	ppb
Yttrium	89-1				102		%
Yttrium	89-2				108		%
Zinc	66-2	340.37	342.75	343.37	342.16	0.46	ppb
Zirconium	90-1	15.75	15.98	16.71	16.15	3.12	ppb
Zirconium	91-1	15.83	16.11	17.00	16.31	3.74	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-15 Instrumnet Name : P7
Client Sample ID : MYE4J3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:55:32 DataFile Name : 047SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	31165.89	31098.42	31242.46	31168.92	0.23	ppb
Antimony	121-1	4.63	4.65	4.68	4.65	0.56	ppb
Arsenic	75-2	134.09	136.39	136.62	135.70	1.03	ppb
Barium	135-1	394.26	404.16	405.36	401.26	1.52	ppb
Barium	137-1	420.31	437.17	440.59	432.69	2.51	ppb
Beryllium	9-1	1.36	1.28	1.22	1.29	5.41	ppb
Bismuth	209-1				102		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	3.01	3.00	2.67	2.89	6.78	ppb
Cadmium	106-1	2.68	3.38	2.54	2.87	15.82	ppb
Cadmium	111-1	1.89	2.01	2.02	1.98	3.64	ppb
Calcium	43-1	6173.89	6259.89	6247.78	6227.19	0.75	ppb
Calcium	44-1	5868.96	5898.25	5956.27	5907.82	0.75	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	82.52	82.35	81.91	82.26	0.38	ppb
Cobalt	59-2	53.06	53.58	53.03	53.22	0.58	ppb
Copper	63-2	243.89	243.71	242.28	243.29	0.36	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				105		%
Indium	115-1				97		%
Indium	115-2				99		%
Iron	56-2	102881.51	103557.39	102860.14	103099.68	0.38	ppb
Iron	57-2	101890.91	103851.34	100630.49	102124.25	1.59	ppb
Iron	54-2	103193.89	104361.44	103242.22	103599.19	0.64	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-15 Instrumnet Name : P7
Client Sample ID : MYE4J3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:55:32 DataFile Name : 047SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	78.60	79.63	80.83	79.69	1.41	ppb
Lead	207-1	78.38	78.88	81.01	79.42	1.76	ppb
Lead	208-1	78.22	78.95	80.10	79.09	1.20	ppb
Lithium	6-1				95		%
Magnesium	24-2	16308.33	16498.66	16469.21	16425.40	0.62	ppb
Manganese	55-2	2258.65	2274.75	2258.71	2264.04	0.41	ppb
Molybdenum	94-1	22.21	22.43	22.89	22.51	1.55	ppb
Molybdenum	95-1	1.37	1.41	1.34	1.37	2.85	ppb
Molybdenum	96-1	4.37	4.45	4.54	4.45	1.89	ppb
Molybdenum	97-1	1.30	1.33	1.38	1.34	3.08	ppb
Molybdenum	98-1	1.28	1.28	1.36	1.31	3.32	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	80.16	79.95	80.15	80.09	0.15	ppb
Phosphorus	31-2	1618.42	1635.84	1637.72	1630.66	0.65	ppb
Potassium	39-2	6411.85	6442.80	6451.35	6435.33	0.32	ppb
Rhodium	103-1				95		%
Rhodium	103-2				101		%
Scandium	45-1				97		%
Scandium	45-2				104		%
Selenium	82-1	1.52	1.01	1.34	1.29	20.29	ppb
Selenium	77-2	17.25	15.83	14.03	15.71	10.28	ppb
Selenium	78-2	2.46	3.03	2.41	2.63	13.15	ppb
Silicon	28-1	861.22	883.36	879.70	874.76	1.36	ppb
Silver	107-1	0.43	0.42	0.44	0.43	1.90	ppb
Silver	109-1	0.40	0.42	0.43	0.42	3.00	ppb
Sodium	23-2	113.89	115.72	114.79	114.80	0.80	ppb
Strontium	86-1	58.15	59.29	59.60	59.02	1.30	ppb
Strontium	88-1	59.80	60.82	61.52	60.71	1.42	ppb
Sulfur	34-1	-1268.99	-1264.14	-1296.25	-1276.46		ppb
Terbium	159-1				103		%
Terbium	159-2				105		%
Thallium	203-1	0.35	0.32	0.33	0.33	5.29	ppb
Thallium	205-1	0.33	0.35	0.34	0.34	2.53	ppb
Tin	118-1	2.67	2.80	2.69	2.72	2.50	ppb
Titanium	47-1	840.01	855.40	859.43	851.61	1.20	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-15

Instrumnet Name :

P7

Client Sample ID :

MYE4J3

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 19:55:32

DataFile Name :

047SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.87	0.87	0.88	0.87	0.33	ppb
Vanadium	51-2	193.59	194.83	194.02	194.15	0.32	ppb
Yttrium	89-1				101		%
Yttrium	89-2				105		%
Zinc	66-2	298.93	299.87	297.39	298.73	0.42	ppb
Zirconium	90-1	12.58	12.87	13.02	12.82	1.73	ppb
Zirconium	91-1	12.77	12.93	13.17	12.95	1.55	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-16 Instrumnet Name : P7
Client Sample ID : MYE4J4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:58:40 DataFile Name : 048SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	39771.25	40403.71	40385.50	40186.82	0.90	ppb
Antimony	121-1	6.06	6.01	7.05	6.37	9.20	ppb
Arsenic	75-2	138.96	142.06	142.13	141.05	1.28	ppb
Barium	135-1	566.73	574.02	661.57	600.78	8.79	ppb
Barium	137-1	559.61	574.50	673.09	602.40	10.24	ppb
Beryllium	9-1	1.54	1.54	1.84	1.64	10.56	ppb
Bismuth	209-1				98		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	4.32	4.08	5.56	4.65	17.08	ppb
Cadmium	106-1	3.17	4.42	4.93	4.18	21.67	ppb
Cadmium	111-1	2.77	2.94	3.39	3.03	10.55	ppb
Calcium	43-1	10526.92	10634.34	12034.23	11065.16	7.60	ppb
Calcium	44-1	9922.26	10004.68	11280.20	10402.38	7.32	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	93.02	94.73	94.70	94.15	1.04	ppb
Cobalt	59-2	49.43	50.34	49.89	49.89	0.91	ppb
Copper	63-2	388.53	393.34	397.83	393.23	1.18	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				99		%
Holmium	165-2				105		%
Indium	115-1				93		%
Indium	115-2				99		%
Iron	56-2	111053.68	110335.19	111851.95	111080.27	0.68	ppb
Iron	57-2	109201.67	108443.08	110408.65	109351.13	0.91	ppb
Iron	54-2	110383.55	110051.94	110866.19	110433.89	0.37	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-16 Instrumnet Name : P7
Client Sample ID : MYE4J4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:58:40 DataFile Name : 048SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	115.64	116.27	132.08	121.33	7.68	ppb
Lead	207-1	115.68	115.90	132.50	121.36	7.95	ppb
Lead	208-1	125.83	125.46	142.12	131.14	7.26	ppb
Lithium	6-1				92		%
Magnesium	24-2	20483.32	20393.65	20526.67	20467.88	0.33	ppb
Manganese	55-2	2476.92	2462.44	2517.19	2485.52	1.14	ppb
Molybdenum	94-1	26.01	26.64	30.37	27.67	8.51	ppb
Molybdenum	95-1	1.60	1.60	1.88	1.70	9.49	ppb
Molybdenum	96-1	4.95	5.05	5.86	5.29	9.42	ppb
Molybdenum	97-1	1.57	1.63	1.89	1.70	10.10	ppb
Molybdenum	98-1	1.54	1.55	1.81	1.63	9.49	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	94.63	95.25	95.95	95.28	0.69	ppb
Phosphorus	31-2	2456.94	2425.02	2475.68	2452.55	1.04	ppb
Potassium	39-2	7573.00	7551.65	7559.33	7561.33	0.14	ppb
Rhodium	103-1				90		%
Rhodium	103-2				100		%
Scandium	45-1				94		%
Scandium	45-2				105		%
Selenium	82-1	0.79	1.08	1.34	1.07	26.14	ppb
Selenium	77-2	18.82	17.51	17.61	17.98	4.06	ppb
Selenium	78-2	4.50	4.08	4.59	4.39	6.22	ppb
Silicon	28-1	983.71	996.98	1132.19	1037.63	7.92	ppb
Silver	107-1	0.79	0.82	0.93	0.84	8.56	ppb
Silver	109-1	0.80	0.79	0.94	0.84	9.79	ppb
Sodium	23-2	163.18	163.52	164.48	163.73	0.41	ppb
Strontium	86-1	99.76	101.58	114.94	105.43	7.86	ppb
Strontium	88-1	102.48	104.66	117.43	108.19	7.47	ppb
Sulfur	34-1	-1150.24	-1087.91	-611.63	-949.93		ppb
Terbium	159-1				99		%
Terbium	159-2				105		%
Thallium	203-1	0.37	0.38	0.43	0.39	7.64	ppb
Thallium	205-1	0.37	0.37	0.43	0.39	8.24	ppb
Tin	118-1	3.19	3.21	3.78	3.39	9.89	ppb
Titanium	47-1	975.94	962.49	1114.18	1017.54	8.25	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-16

Instrumnet Name :

P7

Client Sample ID :

MYE4J4

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 19:58:40

DataFile Name :

048SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.95	0.94	1.11	1.00	9.99	ppb
Vanadium	51-2	213.31	214.07	214.95	214.11	0.38	ppb
Yttrium	89-1				98		%
Yttrium	89-2				106		%
Zinc	66-2	370.11	367.77	375.64	371.17	1.09	ppb
Zirconium	90-1	14.90	15.29	17.28	15.82	8.09	ppb
Zirconium	91-1	14.83	15.40	17.41	15.88	8.56	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV002 Instrumnet Name : P7
Client Sample ID : CCV002 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:01:42 DataFile Name : 049CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	47640.05	48122.32	47937.17	47899.84	0.51	ppb
Antimony	121-1	480.19	468.03	483.24	477.15	1.69	ppb
Arsenic	75-2	505.49	516.41	517.33	513.08	1.28	ppb
Barium	135-1	2434.54	2433.49	2487.36	2451.80	1.26	ppb
Barium	137-1	2442.42	2412.86	2486.00	2447.09	1.50	ppb
Beryllium	9-1	505.39	478.98	493.34	492.57	2.68	ppb
Bismuth	209-1				97		%
Bismuth	209-2				97		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	498.58	493.81	508.36	500.25	1.48	ppb
Cadmium	106-1	496.84	489.45	501.28	495.86	1.21	ppb
Cadmium	111-1	468.73	467.38	478.05	471.39	1.23	ppb
Calcium	43-1	245914.67	241957.89	244622.95	244165.17	0.83	ppb
Calcium	44-1	240514.55	240065.89	240218.12	240266.19	0.09	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	497.77	496.04	493.60	495.80	0.42	ppb
Cobalt	59-2	517.70	521.51	520.11	519.77	0.37	ppb
Copper	63-2	5053.99	4976.76	5026.19	5018.98	0.78	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				106		%
Indium	115-1				93		%
Indium	115-2				96		%
Iron	56-2	126531.39	125099.30	124272.29	125300.99	0.91	ppb
Iron	57-2	125927.16	124256.89	124592.51	124925.52	0.71	ppb
Iron	54-2	125912.27	127174.31	125256.85	126114.48	0.77	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV002 Instrumnet Name : P7
Client Sample ID : CCV002 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:01:42 DataFile Name : 049CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2425.77	2481.03	2508.94	2471.92	1.71	ppb
Lead	207-1	2421.23	2489.62	2486.44	2465.76	1.57	ppb
Lead	208-1	2428.90	2474.15	2491.87	2464.97	1.32	ppb
Lithium	6-1				96		%
Magnesium	24-2	247241.44	248725.11	247462.05	247809.54	0.32	ppb
Manganese	55-2	4936.07	4920.65	4896.88	4917.87	0.40	ppb
Molybdenum	94-1	4942.33	4921.32	4963.70	4942.45	0.43	ppb
Molybdenum	95-1	4911.09	4886.14	4879.25	4892.16	0.34	ppb
Molybdenum	96-1	4896.85	4918.95	4859.00	4891.60	0.62	ppb
Molybdenum	97-1	4909.24	4883.05	4862.86	4885.05	0.48	ppb
Molybdenum	98-1	4898.50	4886.14	4885.57	4890.07	0.15	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	487.74	484.47	487.56	486.59	0.38	ppb
Phosphorus	31-2	9752.49	9790.36	9827.95	9790.27	0.39	ppb
Potassium	39-2	119223.13	120392.25	119081.41	119565.60	0.60	ppb
Rhodium	103-1				87		%
Rhodium	103-2				93		%
Scandium	45-1				90		%
Scandium	45-2				99		%
Selenium	82-1	499.98	504.37	502.56	502.30	0.44	ppb
Selenium	77-2	495.35	483.71	521.25	500.10	3.84	ppb
Selenium	78-2	503.95	505.97	515.62	508.51	1.23	ppb
Silicon	28-1	2559.22	2509.76	2571.53	2546.84	1.28	ppb
Silver	107-1	496.13	488.04	496.32	493.50	0.96	ppb
Silver	109-1	496.92	485.81	494.31	492.34	1.18	ppb
Sodium	23-2	241959.62	242528.05	242061.50	242183.05	0.13	ppb
Strontium	86-1	12147.12	12205.46	12240.00	12197.53	0.38	ppb
Strontium	88-1	12159.54	12280.25	12316.05	12251.94	0.67	ppb
Sulfur	34-1	9666.65	9623.24	9624.58	9638.16	0.26	ppb
Terbium	159-1				102		%
Terbium	159-2				105		%
Thallium	203-1	484.98	495.23	502.86	494.36	1.82	ppb
Thallium	205-1	484.92	488.15	503.06	492.04	1.97	ppb
Tin	118-1	470.44	461.90	475.56	469.30	1.47	ppb
Titanium	47-1	4917.75	4854.12	4889.73	4887.20	0.65	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV002 Instrumnet Name : P7
Client Sample ID : CCV002 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:01:42 DataFile Name : 049CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	498.40	508.37	507.06	504.61	1.07	ppb
Vanadium	51-2	501.13	502.44	491.92	498.49	1.15	ppb
Yttrium	89-1				94		%
Yttrium	89-2				99		%
Zinc	66-2	5009.36	4969.76	5041.18	5006.77	0.71	ppb
Zirconium	90-1	500.67	498.20	500.08	499.65	0.26	ppb
Zirconium	91-1	501.59	496.24	509.76	502.53	1.35	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB002 Instrumnet Name : P7
Client Sample ID : CCB002 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:04:26 DataFile Name : 050CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	11.03	10.78	10.99	10.94	1.23	ppb
Antimony	121-1	0.26	0.28	0.25	0.26	6.68	ppb
Arsenic	75-2	0.03	0.04	0.06	0.04	43.07	ppb
Barium	135-1	0.21	0.24	0.24	0.23	9.19	ppb
Barium	137-1	0.21	0.24	0.25	0.24	7.78	ppb
Beryllium	9-1	0.03	0.00	0.01	0.01	144.01	ppb
Bismuth	209-1				104		%
Bismuth	209-2				108		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.06	0.07	0.10	0.08	23.89	ppb
Cadmium	106-1	-0.48	-1.57	-1.40	-1.15		ppb
Cadmium	111-1	-0.02	-0.08	-0.06	-0.05		ppb
Calcium	43-1	11.87	9.36	11.00	10.74	11.85	ppb
Calcium	44-1	6.46	10.20	9.06	8.57	22.37	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.06	0.04	0.05	0.05	19.05	ppb
Cobalt	59-2	0.05	0.05	0.05	0.05	0.70	ppb
Copper	63-2	0.26	0.22	0.27	0.25	9.49	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				107		%
Indium	115-1				100		%
Indium	115-2				104		%
Iron	56-2	26.80	26.47	25.54	26.27	2.50	ppb
Iron	57-2	23.73	24.01	24.89	24.21	2.51	ppb
Iron	54-2	25.97	25.10	24.59	25.22	2.78	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB002 Instrumnet Name : P7
Client Sample ID : CCB002 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:04:26 DataFile Name : 050CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.10	0.11	0.09	0.10	6.56	ppb
Lead	207-1	0.09	0.11	0.09	0.10	9.61	ppb
Lead	208-1	0.11	0.11	0.10	0.10	7.03	ppb
Lithium	6-1				99		%
Magnesium	24-2	11.57	11.91	11.26	11.58	2.79	ppb
Manganese	55-2	0.73	0.72	0.69	0.71	2.68	ppb
Molybdenum	94-1	0.21	0.29	0.23	0.24	16.42	ppb
Molybdenum	95-1	0.26	0.25	0.25	0.25	2.10	ppb
Molybdenum	96-1	0.22	0.28	0.20	0.24	16.51	ppb
Molybdenum	97-1	0.24	0.30	0.24	0.26	13.13	ppb
Molybdenum	98-1	0.24	0.26	0.24	0.25	3.08	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.05	0.05	0.08	0.06	32.27	ppb
Phosphorus	31-2	-20.10	-18.71	-20.97	-19.92		ppb
Potassium	39-2	0.95	2.83	1.79	1.85	50.65	ppb
Rhodium	103-1				97		%
Rhodium	103-2				104		%
Scandium	45-1				94		%
Scandium	45-2				101		%
Selenium	82-1	-0.66	-0.25	-0.34	-0.42		ppb
Selenium	77-2	-0.22	-0.22	-0.22	-0.22		ppb
Selenium	78-2	-0.70	-1.27	-0.79	-0.92		ppb
Silicon	28-1	1.56	1.77	1.01	1.45	27.10	ppb
Silver	107-1	0.06	0.06	0.04	0.05	22.21	ppb
Silver	109-1	0.05	0.05	0.05	0.05	5.67	ppb
Sodium	23-2	16.52	17.03	15.23	16.26	5.72	ppb
Strontium	86-1	0.37	0.42	0.37	0.39	8.74	ppb
Strontium	88-1	0.35	0.39	0.33	0.36	9.16	ppb
Sulfur	34-1	-475.11	-387.82	-569.71	-477.55		ppb
Terbium	159-1				103		%
Terbium	159-2				106		%
Thallium	203-1	0.05	0.05	0.05	0.05	6.37	ppb
Thallium	205-1	0.05	0.05	0.04	0.05	11.88	ppb
Tin	118-1	0.05	0.06	0.04	0.05	16.27	ppb
Titanium	47-1	0.43	0.37	0.47	0.42	12.10	ppb

LB Number :LB133265

Operator :Jaswal

Lab Sample ID :CCB002

Instrumnet Name :P7

Client Sample ID :CCB002

Dilution Factor :1

Date & Time Acquired :2024-11-02 20:04:26

DataFile Name :050CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.01	0.01	13.18	ppb
Vanadium	51-2	0.10	0.07	0.07	0.08	18.49	ppb
Yttrium	89-1				97		%
Yttrium	89-2				103		%
Zinc	66-2	0.16	0.30	0.12	0.19	50.11	ppb
Zirconium	90-1	0.04	0.05	0.04	0.04	4.15	ppb
Zirconium	91-1	0.04	0.05	0.05	0.05	15.30	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-17 Instrumnet Name : P7
Client Sample ID : MYE4J5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:07:41 DataFile Name : 051SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	39787.77	40431.88	40080.50	40100.05	0.80	ppb
Antimony	121-1	1.46	1.35	1.40	1.40	3.83	ppb
Arsenic	75-2	71.08	75.10	73.91	73.36	2.81	ppb
Barium	135-1	427.64	397.48	405.60	410.24	3.80	ppb
Barium	137-1	461.76	425.83	445.17	444.25	4.05	ppb
Beryllium	9-1	1.66	1.33	1.56	1.52	11.23	ppb
Bismuth	209-1				100		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.03	1.92	2.17	2.04	6.10	ppb
Cadmium	106-1	2.90	2.07	2.22	2.40	18.30	ppb
Cadmium	111-1	1.35	1.19	1.24	1.26	6.43	ppb
Calcium	43-1	29763.42	27927.98	28749.01	28813.47	3.19	ppb
Calcium	44-1	28259.83	26401.13	26875.79	27178.92	3.55	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	88.70	88.49	88.01	88.40	0.40	ppb
Cobalt	59-2	41.23	41.64	41.39	41.42	0.50	ppb
Copper	63-2	203.63	202.98	201.12	202.58	0.64	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				107		%
Indium	115-1				101		%
Indium	115-2				105		%
Iron	56-2	83925.42	86290.75	84979.70	85065.29	1.39	ppb
Iron	57-2	83374.48	84215.65	82890.23	83493.45	0.80	ppb
Iron	54-2	84919.52	86756.82	84542.55	85406.29	1.39	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-17 Instrumnet Name : P7
Client Sample ID : MYE4J5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:07:41 DataFile Name : 051SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	47.38	43.63	46.06	45.69	4.16	ppb
Lead	207-1	46.09	41.80	44.06	43.98	4.87	ppb
Lead	208-1	46.22	42.50	44.91	44.55	4.24	ppb
Lithium	6-1				98		%
Magnesium	24-2	18959.95	18852.25	18875.80	18896.00	0.30	ppb
Manganese	55-2	2159.90	2199.55	2156.28	2171.91	1.11	ppb
Molybdenum	94-1	16.29	15.24	15.83	15.78	3.32	ppb
Molybdenum	95-1	1.95	1.76	1.85	1.85	5.09	ppb
Molybdenum	96-1	4.13	3.78	3.99	3.97	4.42	ppb
Molybdenum	97-1	1.93	1.70	1.78	1.80	6.54	ppb
Molybdenum	98-1	1.82	1.67	1.73	1.74	4.55	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	69.68	69.65	68.81	69.38	0.72	ppb
Phosphorus	31-2	2998.54	2979.17	2912.70	2963.47	1.52	ppb
Potassium	39-2	12584.17	12604.02	12687.54	12625.24	0.43	ppb
Rhodium	103-1				98		%
Rhodium	103-2				105		%
Scandium	45-1				106		%
Scandium	45-2				114		%
Selenium	82-1	0.72	1.07	1.48	1.09	35.16	ppb
Selenium	77-2	20.67	19.58	20.36	20.20	2.77	ppb
Selenium	78-2	3.91	5.03	3.22	4.05	22.54	ppb
Silicon	28-1	1457.54	1357.67	1413.82	1409.67	3.55	ppb
Silver	107-1	0.60	0.53	0.52	0.55	8.00	ppb
Silver	109-1	0.61	0.50	0.55	0.55	9.67	ppb
Sodium	23-2	397.09	394.92	399.26	397.09	0.55	ppb
Strontium	86-1	142.68	131.39	137.73	137.27	4.12	ppb
Strontium	88-1	148.83	136.62	141.23	142.23	4.34	ppb
Sulfur	34-1	-340.90	-800.83	-825.14	-655.62		ppb
Terbium	159-1				104		%
Terbium	159-2				108		%
Thallium	203-1	0.40	0.34	0.36	0.37	7.75	ppb
Thallium	205-1	0.38	0.33	0.36	0.36	7.21	ppb
Tin	118-1	2.52	2.35	2.50	2.46	3.73	ppb
Titanium	47-1	1094.08	1027.49	1063.33	1061.63	3.14	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-17

Instrumnet Name :

P7

Client Sample ID :

MYE4J5

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 20:07:41

DataFile Name :

051SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.18	1.09	1.13	1.13	4.16	ppb
Vanadium	51-2	162.66	163.38	162.67	162.90	0.26	ppb
Yttrium	89-1				109		%
Yttrium	89-2				113		%
Zinc	66-2	229.07	231.68	230.74	230.49	0.57	ppb
Zirconium	90-1	8.92	8.29	8.62	8.61	3.70	ppb
Zirconium	91-1	9.23	8.20	8.70	8.71	5.94	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-18 Instrumnet Name : P7
Client Sample ID : MYE4J6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:10:47 DataFile Name : 052SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	39011.54	38426.09	38563.13	38666.92	0.79	ppb
Antimony	121-1	1.90	2.11	1.95	1.99	5.67	ppb
Arsenic	75-2	83.98	86.41	83.74	84.71	1.74	ppb
Barium	135-1	358.91	402.31	379.49	380.23	5.71	ppb
Barium	137-1	393.71	438.15	403.94	411.93	5.65	ppb
Beryllium	9-1	1.24	1.37	1.31	1.30	4.96	ppb
Bismuth	209-1				98		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.33	2.25	2.65	2.41	8.66	ppb
Cadmium	106-1	2.45	2.94	1.09	2.16	44.48	ppb
Cadmium	111-1	1.73	2.07	1.71	1.84	11.21	ppb
Calcium	43-1	15669.44	17182.44	16278.18	16376.69	4.65	ppb
Calcium	44-1	14864.99	16135.73	15261.15	15420.62	4.22	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	92.55	91.55	91.64	91.91	0.60	ppb
Cobalt	59-2	43.31	43.10	42.59	43.00	0.86	ppb
Copper	63-2	282.25	294.35	280.03	285.54	2.70	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				107		%
Indium	115-1				96		%
Indium	115-2				102		%
Iron	56-2	86948.44	85901.42	86051.93	86300.59	0.66	ppb
Iron	57-2	85915.83	85038.42	84280.66	85078.31	0.96	ppb
Iron	54-2	86882.38	85065.75	85677.40	85875.18	1.08	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-18 Instrumnet Name : P7
Client Sample ID : MYE4J6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:10:47 DataFile Name : 052SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	64.14	71.55	67.57	67.75	5.47	ppb
Lead	207-1	63.79	69.90	65.95	66.55	4.66	ppb
Lead	208-1	63.43	70.10	66.15	66.56	5.04	ppb
Lithium	6-1				93		%
Magnesium	24-2	19402.65	19128.60	19113.23	19214.82	0.85	ppb
Manganese	55-2	1935.58	1904.36	1905.82	1915.25	0.92	ppb
Molybdenum	94-1	13.87	15.37	14.09	14.44	5.63	ppb
Molybdenum	95-1	1.97	2.20	2.07	2.08	5.63	ppb
Molybdenum	96-1	3.84	4.18	3.98	4.00	4.23	ppb
Molybdenum	97-1	1.93	2.17	1.97	2.02	6.24	ppb
Molybdenum	98-1	1.91	2.15	1.94	2.00	6.31	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	72.60	72.57	72.49	72.55	0.08	ppb
Phosphorus	31-2	2634.13	2667.96	2674.54	2658.88	0.82	ppb
Potassium	39-2	14226.12	13927.82	14087.87	14080.60	1.06	ppb
Rhodium	103-1				95		%
Rhodium	103-2				102		%
Scandium	45-1				102		%
Scandium	45-2				111		%
Selenium	82-1	1.74	0.94	1.16	1.28	32.17	ppb
Selenium	77-2	20.02	15.90	13.47	16.46	20.11	ppb
Selenium	78-2	3.92	4.56	4.02	4.17	8.29	ppb
Silicon	28-1	1187.99	1305.18	1245.38	1246.19	4.70	ppb
Silver	107-1	0.68	0.76	0.68	0.71	6.54	ppb
Silver	109-1	0.66	0.77	0.70	0.71	7.71	ppb
Sodium	23-2	1042.31	1031.10	1025.53	1032.98	0.83	ppb
Strontium	86-1	87.70	96.53	91.73	91.99	4.80	ppb
Strontium	88-1	89.96	99.20	94.45	94.54	4.89	ppb
Sulfur	34-1	-1362.32	-1158.53	-1409.22	-1310.02		ppb
Terbium	159-1				101		%
Terbium	159-2				106		%
Thallium	203-1	0.29	0.32	0.31	0.31	4.46	ppb
Thallium	205-1	0.28	0.33	0.31	0.30	7.63	ppb
Tin	118-1	2.38	2.63	2.44	2.48	5.19	ppb
Titanium	47-1	1024.67	1131.14	1067.25	1074.35	4.99	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-18

Instrumnet Name :

P7

Client Sample ID :

MYE4J6

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 20:10:47

DataFile Name :

052SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.01	1.14	1.08	1.08	5.72	ppb
Vanadium	51-2	174.08	172.13	172.13	172.78	0.65	ppb
Yttrium	89-1				105		%
Yttrium	89-2				112		%
Zinc	66-2	272.78	268.03	268.85	269.88	0.94	ppb
Zirconium	90-1	7.30	7.99	7.54	7.61	4.63	ppb
Zirconium	91-1	7.25	8.10	7.70	7.68	5.52	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-19 Instrumnet Name : P7
Client Sample ID : MYE4J7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:13:51 DataFile Name : 053SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	24296.40	24440.33	24215.20	24317.31	0.47	ppb
Antimony	121-1	1.02	1.02	1.04	1.02	0.95	ppb
Arsenic	75-2	39.26	39.28	39.36	39.30	0.14	ppb
Barium	135-1	231.74	235.20	233.84	233.59	0.75	ppb
Barium	137-1	231.89	236.51	234.68	234.36	0.99	ppb
Beryllium	9-1	0.90	1.04	0.88	0.94	9.03	ppb
Bismuth	209-1				101		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.85	1.05	0.82	0.91	13.80	ppb
Cadmium	106-1	0.94	0.58	-0.05	0.49	101.76	ppb
Cadmium	111-1	0.39	0.35	0.29	0.35	14.49	ppb
Calcium	43-1	15623.18	15558.40	15965.83	15715.80	1.39	ppb
Calcium	44-1	14748.26	14631.43	14949.46	14776.39	1.09	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	71.66	71.70	71.67	71.68	0.03	ppb
Cobalt	59-2	31.37	31.33	31.58	31.43	0.43	ppb
Copper	63-2	76.26	75.98	75.92	76.05	0.24	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				105		%
Holmium	165-2				107		%
Indium	115-1				100		%
Indium	115-2				102		%
Iron	56-2	70547.34	69253.04	70187.80	69996.06	0.95	ppb
Iron	57-2	69404.52	69223.12	69285.80	69304.48	0.13	ppb
Iron	54-2	70807.02	69539.20	70888.30	70411.51	1.07	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-19 Instrumnet Name : P7
Client Sample ID : MYE4J7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:13:51 DataFile Name : 053SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	25.17	25.61	25.79	25.53	1.25	ppb
Lead	207-1	23.72	24.30	24.55	24.19	1.77	ppb
Lead	208-1	24.19	24.49	24.59	24.42	0.85	ppb
Lithium	6-1				96		%
Magnesium	24-2	12915.27	12651.05	12717.90	12761.41	1.08	ppb
Manganese	55-2	1222.37	1207.74	1242.99	1224.37	1.45	ppb
Molybdenum	94-1	12.19	12.33	12.31	12.28	0.58	ppb
Molybdenum	95-1	0.87	0.83	0.91	0.87	4.26	ppb
Molybdenum	96-1	2.52	2.67	2.61	2.60	2.87	ppb
Molybdenum	97-1	0.87	0.87	0.86	0.87	0.65	ppb
Molybdenum	98-1	0.81	0.80	0.82	0.81	0.73	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	54.55	53.37	54.02	53.98	1.10	ppb
Phosphorus	31-2	1917.72	1932.17	1939.33	1929.74	0.57	ppb
Potassium	39-2	6165.51	6041.93	6056.25	6087.89	1.11	ppb
Rhodium	103-1				98		%
Rhodium	103-2				102		%
Scandium	45-1				101		%
Scandium	45-2				107		%
Selenium	82-1	0.13	0.95	0.35	0.48	88.47	ppb
Selenium	77-2	14.08	10.46	12.27	12.27	14.75	ppb
Selenium	78-2	3.12	2.94	2.74	2.93	6.52	ppb
Silicon	28-1	963.63	967.76	981.22	970.87	0.95	ppb
Silver	107-1	0.13	0.14	0.14	0.14	3.44	ppb
Silver	109-1	0.13	0.12	0.13	0.13	1.33	ppb
Sodium	23-2	288.80	284.57	287.21	286.86	0.75	ppb
Strontium	86-1	61.56	62.24	63.16	62.32	1.29	ppb
Strontium	88-1	64.08	65.19	65.25	64.84	1.02	ppb
Sulfur	34-1	-1428.18	-1441.51	-1364.52	-1411.40		ppb
Terbium	159-1				104		%
Terbium	159-2				107		%
Thallium	203-1	0.19	0.18	0.20	0.19	3.88	ppb
Thallium	205-1	0.19	0.18	0.19	0.19	2.39	ppb
Tin	118-1	1.87	1.99	1.90	1.92	3.31	ppb
Titanium	47-1	1091.75	1098.07	1117.18	1102.33	1.20	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-19

Instrumnet Name :

P7

Client Sample ID :

MYE4J7

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 20:13:51

DataFile Name :

053SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.71	0.72	0.72	0.72	0.92	ppb
Vanadium	51-2	175.57	174.53	175.00	175.03	0.30	ppb
Yttrium	89-1				106		%
Yttrium	89-2				108		%
Zinc	66-2	131.05	129.81	128.52	129.79	0.98	ppb
Zirconium	90-1	6.87	6.96	6.99	6.94	0.91	ppb
Zirconium	91-1	6.97	7.02	7.06	7.02	0.60	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-20 Instrumnet Name : P7
Client Sample ID : MYE4J8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:16:59 DataFile Name : 054SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	26061.57	25526.41	25947.84	25845.27	1.09	ppb
Antimony	121-1	1.42	1.41	1.43	1.42	0.49	ppb
Arsenic	75-2	49.88	49.40	51.09	50.12	1.74	ppb
Barium	135-1	276.02	275.37	279.63	277.01	0.83	ppb
Barium	137-1	277.56	277.52	276.92	277.34	0.13	ppb
Beryllium	9-1	0.91	0.96	0.96	0.95	2.88	ppb
Bismuth	209-1				103		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.31	1.73	1.39	1.48	14.96	ppb
Cadmium	106-1	1.68	0.60	0.49	0.92	71.25	ppb
Cadmium	111-1	0.95	0.86	0.90	0.90	4.69	ppb
Calcium	43-1	22453.27	22972.87	23129.08	22851.74	1.55	ppb
Calcium	44-1	21346.13	21531.35	21614.64	21497.37	0.64	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	62.26	62.33	63.12	62.57	0.76	ppb
Cobalt	59-2	28.39	28.45	28.91	28.58	0.99	ppb
Copper	63-2	140.75	140.06	142.87	141.23	1.04	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				106		%
Indium	115-1				100		%
Indium	115-2				102		%
Iron	56-2	58452.49	59009.12	59484.33	58981.98	0.88	ppb
Iron	57-2	57902.29	58150.04	59115.17	58389.17	1.10	ppb
Iron	54-2	58789.87	58611.69	59318.31	58906.63	0.62	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-20 Instrumnet Name : P7
Client Sample ID : MYE4J8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:16:59 DataFile Name : 054SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	35.81	35.66	36.58	36.02	1.37	ppb
Lead	207-1	34.05	34.20	35.23	34.49	1.85	ppb
Lead	208-1	34.78	34.63	35.51	34.97	1.35	ppb
Lithium	6-1				95		%
Magnesium	24-2	14549.53	14433.05	14608.02	14530.20	0.61	ppb
Manganese	55-2	1424.28	1397.94	1430.46	1417.56	1.22	ppb
Molybdenum	94-1	10.83	10.89	11.23	10.98	1.99	ppb
Molybdenum	95-1	1.31	1.32	1.32	1.32	0.52	ppb
Molybdenum	96-1	2.80	2.79	2.87	2.82	1.58	ppb
Molybdenum	97-1	1.28	1.33	1.36	1.33	3.01	ppb
Molybdenum	98-1	1.26	1.27	1.29	1.27	1.09	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	55.15	55.30	55.55	55.33	0.37	ppb
Phosphorus	31-2	2411.91	2419.42	2448.54	2426.62	0.80	ppb
Potassium	39-2	7880.64	7812.91	7914.97	7869.51	0.66	ppb
Rhodium	103-1				97		%
Rhodium	103-2				104		%
Scandium	45-1				100		%
Scandium	45-2				107		%
Selenium	82-1	0.95	0.66	0.85	0.82	17.66	ppb
Selenium	77-2	9.60	13.74	11.27	11.53	18.06	ppb
Selenium	78-2	3.58	1.67	1.67	2.31	47.81	ppb
Silicon	28-1	1240.29	1252.22	1269.67	1254.06	1.18	ppb
Silver	107-1	0.33	0.34	0.35	0.34	3.83	ppb
Silver	109-1	0.32	0.34	0.34	0.33	2.84	ppb
Sodium	23-2	391.90	388.86	396.30	392.35	0.95	ppb
Strontium	86-1	130.00	128.59	132.85	130.48	1.66	ppb
Strontium	88-1	134.38	133.33	135.46	134.39	0.79	ppb
Sulfur	34-1	-599.55	-489.23	-479.32	-522.70		ppb
Terbium	159-1				103		%
Terbium	159-2				107		%
Thallium	203-1	0.21	0.19	0.19	0.20	7.23	ppb
Thallium	205-1	0.20	0.20	0.20	0.20	0.93	ppb
Tin	118-1	2.16	2.16	2.29	2.21	3.44	ppb
Titanium	47-1	812.60	829.66	840.04	827.44	1.67	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-20

Instrumnet Name :

P7

Client Sample ID :

MYE4J8

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 20:16:59

DataFile Name :

054SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.73	0.69	0.72	0.71	2.68	ppb
Vanadium	51-2	118.23	117.39	120.32	118.65	1.27	ppb
Yttrium	89-1				104		%
Yttrium	89-2				108		%
Zinc	66-2	205.11	203.96	208.29	205.78	1.09	ppb
Zirconium	90-1	5.88	5.93	6.01	5.94	1.11	ppb
Zirconium	91-1	5.87	5.81	6.15	5.95	3.06	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-21 Instrumnet Name : P7
Client Sample ID : MYE4J9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:20:06 DataFile Name : 055SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	43811.46	42858.63	42329.17	42999.75	1.75	ppb
Antimony	121-1	1.23	1.21	1.23	1.22	1.19	ppb
Arsenic	75-2	51.46	51.65	50.00	51.04	1.77	ppb
Barium	135-1	421.73	422.44	427.67	423.95	0.76	ppb
Barium	137-1	451.81	454.74	460.83	455.80	1.01	ppb
Beryllium	9-1	1.63	1.64	1.64	1.63	0.22	ppb
Bismuth	209-1				101		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.90	1.80	1.99	1.89	5.03	ppb
Cadmium	106-1	2.46	2.06	1.13	1.88	36.45	ppb
Cadmium	111-1	0.84	0.85	0.74	0.81	7.42	ppb
Calcium	43-1	22198.70	22114.27	22295.91	22202.96	0.41	ppb
Calcium	44-1	20861.27	20696.86	21121.49	20893.21	1.02	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	90.13	89.23	86.89	88.75	1.88	ppb
Cobalt	59-2	42.52	42.49	41.18	42.07	1.82	ppb
Copper	63-2	143.38	140.55	137.93	140.62	1.94	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				105		%
Holmium	165-2				107		%
Indium	115-1				99		%
Indium	115-2				103		%
Iron	56-2	82470.87	81640.14	80649.60	81586.87	1.12	ppb
Iron	57-2	81099.38	81661.85	79405.83	80722.35	1.45	ppb
Iron	54-2	83299.94	82246.05	79773.40	81773.13	2.21	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-21 Instrumnet Name : P7
Client Sample ID : MYE4J9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:20:06 DataFile Name : 055SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	36.66	37.07	37.45	37.06	1.07	ppb
Lead	207-1	35.09	35.23	35.25	35.19	0.25	ppb
Lead	208-1	35.56	35.74	36.03	35.77	0.66	ppb
Lithium	6-1				96		%
Magnesium	24-2	19390.95	19125.80	18800.12	19105.62	1.55	ppb
Manganese	55-2	2056.29	2056.73	2026.44	2046.48	0.85	ppb
Molybdenum	94-1	20.23	20.08	20.49	20.27	1.02	ppb
Molybdenum	95-1	1.03	1.06	1.03	1.04	1.51	ppb
Molybdenum	96-1	3.63	3.74	3.67	3.68	1.58	ppb
Molybdenum	97-1	1.01	1.00	1.02	1.01	0.71	ppb
Molybdenum	98-1	1.00	0.96	1.01	0.99	2.71	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	75.50	73.97	73.38	74.28	1.47	ppb
Phosphorus	31-2	2770.10	2744.90	2674.06	2729.69	1.82	ppb
Potassium	39-2	11693.26	11575.29	11193.83	11487.46	2.27	ppb
Rhodium	103-1				96		%
Rhodium	103-2				103		%
Scandium	45-1				101		%
Scandium	45-2				108		%
Selenium	82-1	0.47	1.03	0.90	0.80	36.41	ppb
Selenium	77-2	15.64	17.89	21.07	18.20	14.98	ppb
Selenium	78-2	3.89	4.42	2.22	3.51	32.70	ppb
Silicon	28-1	1058.50	1048.24	1057.23	1054.65	0.53	ppb
Silver	107-1	0.27	0.27	0.29	0.28	4.21	ppb
Silver	109-1	0.30	0.29	0.27	0.29	5.05	ppb
Sodium	23-2	288.95	283.52	279.08	283.85	1.74	ppb
Strontium	86-1	94.58	95.91	95.68	95.39	0.74	ppb
Strontium	88-1	98.19	98.70	97.50	98.13	0.61	ppb
Sulfur	34-1	-1247.55	-1307.91	-1297.33	-1284.26		ppb
Terbium	159-1				104		%
Terbium	159-2				108		%
Thallium	203-1	0.32	0.34	0.33	0.33	3.60	ppb
Thallium	205-1	0.33	0.32	0.33	0.33	2.35	ppb
Tin	118-1	2.42	2.47	2.49	2.46	1.50	ppb
Titanium	47-1	1182.08	1170.66	1179.92	1177.56	0.52	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-21 Instrumnet Name : P7
Client Sample ID : MYE4J9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:20:06 DataFile Name : 055SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.04	1.07	1.08	1.06	1.96	ppb
Vanadium	51-2	157.34	154.52	152.16	154.68	1.68	ppb
Yttrium	89-1				106		%
Yttrium	89-2				111		%
Zinc	66-2	206.55	203.02	198.38	202.65	2.02	ppb
Zirconium	90-1	11.38	11.59	11.68	11.55	1.37	ppb
Zirconium	91-1	11.58	11.53	11.66	11.59	0.54	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-22 Instrumnet Name : P7
Client Sample ID : MYE4L3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:23:13 DataFile Name : 056SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	33634.52	33368.80	32930.95	33311.42	1.07	ppb
Antimony	121-1	1.45	1.47	1.46	1.46	0.72	ppb
Arsenic	75-2	51.43	52.36	52.11	51.97	0.92	ppb
Barium	135-1	288.65	291.73	297.42	292.60	1.52	ppb
Barium	137-1	289.94	290.93	296.61	292.49	1.23	ppb
Beryllium	9-1	1.06	1.09	1.27	1.14	9.71	ppb
Bismuth	209-1				102		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.17	1.48	1.30	1.32	11.55	ppb
Cadmium	106-1	0.53	1.11	0.58	0.74	43.87	ppb
Cadmium	111-1	0.72	0.74	0.67	0.71	4.84	ppb
Calcium	43-1	17253.99	17229.99	17630.36	17371.45	1.29	ppb
Calcium	44-1	16277.04	16394.33	16590.96	16420.78	0.97	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	70.63	70.13	69.85	70.21	0.56	ppb
Cobalt	59-2	34.29	34.10	33.91	34.10	0.55	ppb
Copper	63-2	147.96	145.98	145.49	146.48	0.89	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				105		%
Indium	115-1				98		%
Indium	115-2				101		%
Iron	56-2	69123.25	68829.93	68564.14	68839.11	0.41	ppb
Iron	57-2	69043.81	68533.47	67890.02	68489.10	0.84	ppb
Iron	54-2	69355.35	68941.75	69547.60	69281.57	0.45	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-22 Instrumnet Name : P7
Client Sample ID : MYE4L3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:23:13 DataFile Name : 056SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	31.86	32.10	32.47	32.14	0.96	ppb
Lead	207-1	31.09	31.07	31.29	31.15	0.38	ppb
Lead	208-1	31.04	31.22	31.57	31.28	0.86	ppb
Lithium	6-1				96		%
Magnesium	24-2	15442.26	15308.81	15222.57	15324.55	0.72	ppb
Manganese	55-2	1562.69	1543.14	1549.78	1551.87	0.64	ppb
Molybdenum	94-1	10.70	10.84	10.75	10.77	0.67	ppb
Molybdenum	95-1	1.26	1.27	1.31	1.28	2.34	ppb
Molybdenum	96-1	2.74	2.86	2.85	2.82	2.39	ppb
Molybdenum	97-1	1.21	1.24	1.23	1.23	1.13	ppb
Molybdenum	98-1	1.18	1.18	1.22	1.19	1.75	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	54.47	53.38	53.22	53.69	1.26	ppb
Phosphorus	31-2	2236.03	2220.22	2178.28	2211.51	1.35	ppb
Potassium	39-2	11594.63	11585.74	11626.07	11602.15	0.18	ppb
Rhodium	103-1				96		%
Rhodium	103-2				101		%
Scandium	45-1				99		%
Scandium	45-2				105		%
Selenium	82-1	0.54	1.08	0.76	0.79	34.17	ppb
Selenium	77-2	12.67	13.03	13.58	13.09	3.51	ppb
Selenium	78-2	0.82	2.68	2.42	1.97	51.15	ppb
Silicon	28-1	767.56	766.78	784.60	772.98	1.30	ppb
Silver	107-1	0.30	0.28	0.30	0.29	3.34	ppb
Silver	109-1	0.30	0.30	0.32	0.31	3.67	ppb
Sodium	23-2	341.31	335.67	334.21	337.06	1.11	ppb
Strontium	86-1	85.82	88.11	89.48	87.81	2.10	ppb
Strontium	88-1	89.64	90.61	91.76	90.67	1.17	ppb
Sulfur	34-1	-1377.04	-1343.40	-1292.69	-1337.71		ppb
Terbium	159-1				104		%
Terbium	159-2				106		%
Thallium	203-1	0.24	0.24	0.26	0.25	4.59	ppb
Thallium	205-1	0.24	0.22	0.24	0.23	4.32	ppb
Tin	118-1	1.86	1.92	1.92	1.90	1.79	ppb
Titanium	47-1	892.18	877.69	891.48	887.12	0.92	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-22

Instrumnet Name :

P7

Client Sample ID :

MYE4L3

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 20:23:13

DataFile Name :

056SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.91	0.90	0.91	0.91	0.64	ppb
Vanadium	51-2	134.81	133.58	133.46	133.95	0.56	ppb
Yttrium	89-1				103		%
Yttrium	89-2				108		%
Zinc	66-2	177.90	175.94	175.17	176.34	0.80	ppb
Zirconium	90-1	5.76	5.94	5.94	5.88	1.77	ppb
Zirconium	91-1	5.74	5.94	5.90	5.86	1.75	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-01 Instrumnet Name : P7
Client Sample ID : MYE4L4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:36:08 DataFile Name : 060SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	28667.05	35571.09	29981.65	31406.60	11.67	ppb
Antimony	121-1	0.21	0.22	0.21	0.21	2.70	ppb
Arsenic	75-2	26.92	33.37	28.00	29.43	11.75	ppb
Barium	135-1	203.07	205.29	206.26	204.87	0.80	ppb
Barium	137-1	201.73	207.13	209.08	205.98	1.85	ppb
Beryllium	9-1	0.53	0.60	0.48	0.53	11.34	ppb
Bismuth	209-1				100		%
Bismuth	209-2				97		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.24	1.20	1.31	1.25	4.63	ppb
Cadmium	106-1	0.29	-0.08	-0.21	0.00	32207.36	ppb
Cadmium	111-1	0.49	0.44	0.47	0.47	6.12	ppb
Calcium	43-1	17162.87	17743.41	17608.05	17504.78	1.74	ppb
Calcium	44-1	16435.22	16681.96	16643.27	16586.82	0.80	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	108.25	132.54	111.09	117.29	11.32	ppb
Cobalt	59-2	46.78	57.16	48.85	50.93	10.78	ppb
Copper	63-2	140.43	172.01	146.79	153.08	10.91	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				99		%
Indium	115-1				98		%
Indium	115-2				94		%
Iron	56-2	56990.87	68640.56	58794.32	61475.25	10.20	ppb
Iron	57-2	56665.11	69112.33	59015.00	61597.48	10.74	ppb
Iron	54-2	56788.41	70196.97	58451.37	61812.25	11.82	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-01 Instrumnet Name : P7
Client Sample ID : MYE4L4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:36:08 DataFile Name : 060SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	17.04	16.94	17.11	17.03	0.52	ppb
Lead	207-1	16.30	16.77	16.52	16.53	1.42	ppb
Lead	208-1	16.54	16.76	16.68	16.66	0.65	ppb
Lithium	6-1				94		%
Magnesium	24-2	22677.70	28091.53	23680.69	24816.64	11.61	ppb
Manganese	55-2	1624.76	1965.36	1667.40	1752.51	10.59	ppb
Molybdenum	94-1	10.28	10.56	10.62	10.49	1.72	ppb
Molybdenum	95-1	0.59	0.55	0.59	0.57	3.32	ppb
Molybdenum	96-1	2.02	2.06	2.09	2.06	1.64	ppb
Molybdenum	97-1	0.56	0.58	0.59	0.57	2.89	ppb
Molybdenum	98-1	0.52	0.52	0.55	0.53	2.55	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	86.85	105.49	89.41	93.92	10.76	ppb
Phosphorus	31-2	1933.98	2448.80	2043.41	2142.06	12.66	ppb
Potassium	39-2	5754.92	6949.74	5916.49	6207.05	10.44	ppb
Rhodium	103-1				94		%
Rhodium	103-2				93		%
Scandium	45-1				95		%
Scandium	45-2				95		%
Selenium	82-1	-0.02	0.61	0.10	0.23	145.75	ppb
Selenium	77-2	6.71	10.54	8.18	8.48	22.84	ppb
Selenium	78-2	0.80	2.82	1.87	1.83	55.24	ppb
Silicon	28-1	1098.60	1130.28	1125.16	1118.02	1.52	ppb
Silver	107-1	0.14	0.15	0.15	0.15	6.61	ppb
Silver	109-1	0.13	0.14	0.15	0.14	5.41	ppb
Sodium	23-2	216.53	272.66	223.30	237.50	12.90	ppb
Strontium	86-1	53.58	54.13	54.86	54.19	1.18	ppb
Strontium	88-1	56.00	56.05	55.68	55.91	0.36	ppb
Sulfur	34-1	-1337.28	-1401.67	-1395.36	-1378.10		ppb
Terbium	159-1				102		%
Terbium	159-2				98		%
Thallium	203-1	0.13	0.13	0.14	0.13	4.87	ppb
Thallium	205-1	0.13	0.13	0.14	0.13	2.01	ppb
Tin	118-1	0.30	0.30	0.29	0.30	1.31	ppb
Titanium	47-1	601.93	616.72	611.17	609.94	1.23	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-01

Instrumnet Name :

P7

Client Sample ID :

MYE4L4

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 20:36:08

DataFile Name :

060SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.46	0.46	0.46	0.46	0.49	ppb
Vanadium	51-2	119.91	145.26	123.56	129.58	10.58	ppb
Yttrium	89-1				101		%
Yttrium	89-2				99		%
Zinc	66-2	125.81	153.51	129.95	136.43	10.95	ppb
Zirconium	90-1	5.92	5.91	6.08	5.97	1.63	ppb
Zirconium	91-1	5.97	6.01	6.06	6.01	0.68	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-02 Instrumnet Name : P7
Client Sample ID : MYE4L5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:39:15 DataFile Name : 061SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	27880.08	27456.73	27982.84	27773.22	1.00	ppb
Antimony	121-1	0.20	0.20	0.20	0.20	2.48	ppb
Arsenic	75-2	32.22	31.56	32.34	32.04	1.32	ppb
Barium	135-1	321.11	326.95	331.52	326.53	1.60	ppb
Barium	137-1	352.45	356.94	361.34	356.91	1.25	ppb
Beryllium	9-1	1.17	1.12	1.26	1.18	5.69	ppb
Bismuth	209-1				101		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.30	1.00	1.28	1.19	14.37	ppb
Cadmium	106-1	0.08	0.39	0.58	0.35	72.09	ppb
Cadmium	111-1	0.65	0.73	0.69	0.69	5.82	ppb
Calcium	43-1	20088.42	20455.40	20285.60	20276.47	0.91	ppb
Calcium	44-1	19113.73	19397.88	19342.10	19284.57	0.78	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	58.18	57.77	57.96	57.97	0.36	ppb
Cobalt	59-2	35.48	34.97	35.65	35.37	1.00	ppb
Copper	63-2	129.56	128.84	128.50	128.97	0.42	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				107		%
Indium	115-1				97		%
Indium	115-2				102		%
Iron	56-2	53627.43	53545.10	53976.27	53716.26	0.43	ppb
Iron	57-2	53529.67	53045.54	54290.67	53621.96	1.17	ppb
Iron	54-2	54415.24	53369.84	53690.10	53825.06	1.00	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-02 Instrumnet Name : P7
Client Sample ID : MYE4L5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:39:15 DataFile Name : 061SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	31.54	31.65	32.23	31.81	1.17	ppb
Lead	207-1	30.23	30.45	30.77	30.48	0.90	ppb
Lead	208-1	30.41	30.66	31.09	30.72	1.13	ppb
Lithium	6-1				96		%
Magnesium	24-2	15314.95	15268.36	15315.04	15299.45	0.18	ppb
Manganese	55-2	1735.20	1738.46	1739.55	1737.74	0.13	ppb
Molybdenum	94-1	14.20	14.78	14.60	14.53	2.02	ppb
Molybdenum	95-1	0.73	0.79	0.76	0.76	4.17	ppb
Molybdenum	96-1	2.56	2.71	2.61	2.63	2.88	ppb
Molybdenum	97-1	0.76	0.75	0.74	0.75	1.44	ppb
Molybdenum	98-1	0.71	0.72	0.71	0.71	0.96	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	53.69	53.09	53.36	53.38	0.56	ppb
Phosphorus	31-2	2148.00	2161.36	2235.40	2181.58	2.16	ppb
Potassium	39-2	8628.15	8495.64	8569.61	8564.47	0.78	ppb
Rhodium	103-1				93		%
Rhodium	103-2				101		%
Scandium	45-1				95		%
Scandium	45-2				103		%
Selenium	82-1	0.47	0.48	0.88	0.61	38.22	ppb
Selenium	77-2	11.21	14.17	15.48	13.62	16.05	ppb
Selenium	78-2	2.76	2.26	3.58	2.86	23.28	ppb
Silicon	28-1	1278.00	1312.06	1291.37	1293.81	1.33	ppb
Silver	107-1	0.22	0.22	0.22	0.22	1.95	ppb
Silver	109-1	0.22	0.21	0.22	0.22	1.25	ppb
Sodium	23-2	240.51	237.78	241.90	240.06	0.87	ppb
Strontium	86-1	74.59	76.60	77.18	76.12	1.79	ppb
Strontium	88-1	76.67	79.16	79.19	78.34	1.85	ppb
Sulfur	34-1	-1312.16	-1310.70	-1356.84	-1326.57		ppb
Terbium	159-1				102		%
Terbium	159-2				108		%
Thallium	203-1	0.24	0.24	0.24	0.24	1.12	ppb
Thallium	205-1	0.24	0.24	0.25	0.24	2.55	ppb
Tin	118-1	0.37	0.38	0.42	0.39	6.55	ppb
Titanium	47-1	723.49	733.98	733.06	730.18	0.80	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-02

Instrumnet Name :

P7

Client Sample ID :

MYE4L5

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 20:39:15

DataFile Name :

061SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.85	0.83	0.87	0.85	2.02	ppb
Vanadium	51-2	104.21	102.72	103.61	103.51	0.72	ppb
Yttrium	89-1				101		%
Yttrium	89-2				107		%
Zinc	66-2	165.30	164.10	164.66	164.68	0.36	ppb
Zirconium	90-1	8.26	8.38	8.37	8.34	0.79	ppb
Zirconium	91-1	8.19	8.31	8.38	8.30	1.16	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BL Instrumnet Name : P7
Client Sample ID : PBS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:48:20 DataFile Name : 063CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	6.10	4.33	4.23	4.89	21.57	ppb
Antimony	121-1	0.00	0.00	0.00	0.00		ppb
Arsenic	75-2	0.00	0.01	0.04	0.02	136.88	ppb
Barium	135-1	0.07	0.08	0.09	0.08	13.65	ppb
Barium	137-1	0.10	0.07	0.08	0.08	16.91	ppb
Beryllium	9-1	-0.03	-0.03	-0.04	-0.04		ppb
Bismuth	209-1				99		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.01	0.05	0.07	0.04	66.48	ppb
Cadmium	106-1	-2.08	-2.98	-2.29	-2.45		ppb
Cadmium	111-1	-0.15	-0.22	-0.17	-0.18		ppb
Calcium	43-1	-1.00	-3.80	1.19	-1.20		ppb
Calcium	44-1	-0.09	-1.35	0.26	-0.40		ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.02	0.03	0.04	0.03	22.41	ppb
Cobalt	59-2	0.01	0.01	0.01	0.01	26.53	ppb
Copper	63-2	-0.03	-0.02	-0.05	-0.03		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				99		%
Holmium	165-2				103		%
Indium	115-1				98		%
Indium	115-2				103		%
Iron	56-2	14.66	13.74	13.84	14.08	3.57	ppb
Iron	57-2	12.70	13.42	13.59	13.24	3.54	ppb
Iron	54-2	14.76	14.20	14.29	14.42	2.08	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BL Instrumnet Name : P7
Client Sample ID : PBS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:48:20 DataFile Name : 063CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	-0.04	-0.04	-0.04	-0.04		ppb
Lead	207-1	-0.03	-0.05	-0.03	-0.03		ppb
Lead	208-1	-0.03	-0.03	-0.04	-0.03		ppb
Lithium	6-1				93		%
Magnesium	24-2	2.33	1.46	1.36	1.72	31.26	ppb
Manganese	55-2	0.39	0.35	0.34	0.36	6.76	ppb
Molybdenum	94-1	-0.03	-0.03	-0.03	-0.03		ppb
Molybdenum	95-1	-0.02	-0.03	-0.02	-0.02		ppb
Molybdenum	96-1	-0.02	-0.03	-0.03	-0.03		ppb
Molybdenum	97-1	-0.03	-0.03	-0.02	-0.03		ppb
Molybdenum	98-1	-0.03	-0.03	-0.03	-0.03		ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.07	0.04	0.06	0.06	27.63	ppb
Phosphorus	31-2	-18.46	-18.28	-20.84	-19.19		ppb
Potassium	39-2	-11.88	-11.73	-12.78	-12.13		ppb
Rhodium	103-1				96		%
Rhodium	103-2				103		%
Scandium	45-1				92		%
Scandium	45-2				97		%
Selenium	82-1	-0.20	-0.01	-0.10	-0.10		ppb
Selenium	77-2	0.01	-0.22	-0.22	-0.15		ppb
Selenium	78-2	-0.68	-0.15	0.56	-0.09		ppb
Silicon	28-1	-0.11	-0.17	0.19	-0.03		ppb
Silver	107-1	0.00	0.00	0.00	0.00		ppb
Silver	109-1	0.00	0.00	0.00	0.00	143.69	ppb
Sodium	23-2	-14.39	-14.74	-14.40	-14.51		ppb
Strontium	86-1	-0.02	-0.01	-0.03	-0.02		ppb
Strontium	88-1	-0.04	-0.04	-0.04	-0.04		ppb
Sulfur	34-1	-1519.93	-1502.44	-1437.62	-1486.66		ppb
Terbium	159-1				100		%
Terbium	159-2				104		%
Thallium	203-1	0.00	0.00	0.00	0.00		ppb
Thallium	205-1	0.00	0.00	0.00	0.00		ppb
Tin	118-1	-0.02	-0.01	-0.01	-0.01		ppb
Titanium	47-1	0.13	0.09	0.08	0.10	24.94	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BL Instrumnet Name : P7
Client Sample ID : PBS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:48:20 DataFile Name : 063CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00		ppb
Vanadium	51-2	0.04	0.02	0.05	0.04	45.71	ppb
Yttrium	89-1				95		%
Yttrium	89-2				100		%
Zinc	66-2	-0.11	-0.19	-0.20	-0.17		ppb
Zirconium	90-1	-0.01	-0.01	-0.01	-0.01		ppb
Zirconium	91-1	-0.01	0.00	-0.02	-0.01		ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BS Instrumnet Name : P7
Client Sample ID : LCS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:52:15 DataFile Name : 064LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	40.86	40.69	42.36	41.30	2.23	ppb
Antimony	121-1	3.63	3.64	3.67	3.65	0.71	ppb
Arsenic	75-2	1.89	1.95	1.91	1.92	1.77	ppb
Barium	135-1	17.34	17.25	17.61	17.40	1.06	ppb
Barium	137-1	17.25	17.37	17.54	17.39	0.83	ppb
Beryllium	9-1	1.96	1.93	1.84	1.91	3.08	ppb
Bismuth	209-1				98		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.13	1.47	2.07	1.89	19.20	ppb
Cadmium	106-1	0.44	-0.51	-0.82	-0.30		ppb
Cadmium	111-1	1.64	1.59	1.54	1.59	3.22	ppb
Calcium	43-1	906.13	937.98	943.29	929.13	2.16	ppb
Calcium	44-1	932.39	942.69	949.71	941.60	0.92	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	4.11	4.11	4.16	4.13	0.63	ppb
Cobalt	59-2	1.99	2.10	2.02	2.03	2.64	ppb
Copper	63-2	3.95	3.93	3.87	3.92	1.13	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				98		%
Holmium	165-2				102		%
Indium	115-1				96		%
Indium	115-2				100		%
Iron	56-2	422.91	423.43	420.71	422.35	0.34	ppb
Iron	57-2	429.31	433.83	434.54	432.56	0.66	ppb
Iron	54-2	439.37	436.35	438.50	438.08	0.35	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BS Instrumnet Name : P7
Client Sample ID : LCS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:52:15 DataFile Name : 064LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1.59	1.61	1.59	1.60	0.85	ppb
Lead	207-1	1.56	1.57	1.62	1.58	1.94	ppb
Lead	208-1	1.57	1.59	1.61	1.59	1.14	ppb
Lithium	6-1				91		%
Magnesium	24-2	1012.61	1014.96	1012.24	1013.27	0.15	ppb
Manganese	55-2	2.21	2.30	2.22	2.25	2.21	ppb
Molybdenum	94-1	10.12	10.29	10.30	10.23	0.97	ppb
Molybdenum	95-1	8.32	8.43	8.35	8.36	0.67	ppb
Molybdenum	96-1	8.52	8.56	8.76	8.62	1.50	ppb
Molybdenum	97-1	8.32	8.35	8.43	8.36	0.68	ppb
Molybdenum	98-1	8.29	8.41	8.45	8.38	0.98	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	2.08	2.02	2.02	2.04	1.75	ppb
Phosphorus	31-2	-19.12	-17.02	-20.15	-18.76		ppb
Potassium	39-2	987.15	979.30	981.68	982.71	0.41	ppb
Rhodium	103-1				95		%
Rhodium	103-2				101		%
Scandium	45-1				91		%
Scandium	45-2				97		%
Selenium	82-1	9.41	9.26	10.40	9.69	6.38	ppb
Selenium	77-2	10.15	8.16	9.48	9.26	10.95	ppb
Selenium	78-2	9.52	9.88	9.40	9.60	2.59	ppb
Silicon	28-1	5.27	5.63	5.72	5.54	4.29	ppb
Silver	107-1	1.77	1.82	1.81	1.80	1.42	ppb
Silver	109-1	1.80	1.82	1.88	1.83	2.26	ppb
Sodium	23-2	938.91	940.66	939.61	939.73	0.09	ppb
Strontium	86-1	1.93	1.86	1.90	1.90	1.92	ppb
Strontium	88-1	1.82	1.79	1.86	1.82	1.97	ppb
Sulfur	34-1	-1285.14	-1195.37	-1084.96	-1188.49		ppb
Terbium	159-1				99		%
Terbium	159-2				103		%
Thallium	203-1	1.59	1.62	1.61	1.60	0.96	ppb
Thallium	205-1	1.58	1.55	1.59	1.57	1.15	ppb
Tin	118-1	0.00	0.00	0.02	0.01	191.53	ppb
Titanium	47-1	8.68	9.03	9.04	8.92	2.29	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BS Instrumnet Name : P7
Client Sample ID : LCS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:52:15 DataFile Name : 064LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.51	1.46	1.56	1.51	3.30	ppb
Vanadium	51-2	9.74	9.82	9.73	9.77	0.50	ppb
Yttrium	89-1				94		%
Yttrium	89-2				100		%
Zinc	66-2	10.28	9.89	9.70	9.96	2.97	ppb
Zirconium	90-1	1.89	1.87	1.90	1.89	0.95	ppb
Zirconium	91-1	1.88	1.92	1.91	1.91	1.25	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BSD Instrumnet Name : P7
Client Sample ID : LCSD572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:55:31 DataFile Name : 065LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	22.53	22.79	22.70	22.67	0.58	ppb
Antimony	121-1	1.83	1.84	1.85	1.84	0.58	ppb
Arsenic	75-2	1.16	1.04	1.03	1.08	6.42	ppb
Barium	135-1	8.89	8.74	8.69	8.77	1.19	ppb
Barium	137-1	8.52	8.74	8.77	8.68	1.56	ppb
Beryllium	9-1	0.93	0.85	0.91	0.90	4.22	ppb
Bismuth	209-1				98		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.78	1.01	0.97	0.92	13.41	ppb
Cadmium	106-1	-0.75	-1.00	-0.30	-0.68		ppb
Cadmium	111-1	0.76	0.73	0.84	0.78	7.10	ppb
Calcium	43-1	454.64	480.21	458.57	464.47	2.96	ppb
Calcium	44-1	461.23	470.52	464.17	465.31	1.02	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2.03	1.96	2.00	2.00	1.86	ppb
Cobalt	59-2	0.98	1.00	1.03	1.00	2.48	ppb
Copper	63-2	1.87	2.03	2.00	1.96	4.39	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				98		%
Holmium	165-2				103		%
Indium	115-1				97		%
Indium	115-2				101		%
Iron	56-2	217.46	218.48	216.80	217.58	0.39	ppb
Iron	57-2	218.40	221.17	220.60	220.06	0.66	ppb
Iron	54-2	223.69	220.56	221.54	221.93	0.72	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BSD Instrumnet Name : P7
Client Sample ID : LCSD572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:55:31 DataFile Name : 065LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.73	0.74	0.78	0.75	3.62	ppb
Lead	207-1	0.75	0.73	0.75	0.74	1.49	ppb
Lead	208-1	0.75	0.75	0.77	0.76	1.37	ppb
Lithium	6-1				90		%
Magnesium	24-2	492.19	495.05	496.91	494.71	0.48	ppb
Manganese	55-2	1.30	1.32	1.23	1.29	3.72	ppb
Molybdenum	94-1	4.85	4.96	4.92	4.91	1.11	ppb
Molybdenum	95-1	4.05	4.10	4.21	4.12	1.89	ppb
Molybdenum	96-1	4.11	4.21	4.15	4.16	1.19	ppb
Molybdenum	97-1	4.04	4.12	4.04	4.07	1.22	ppb
Molybdenum	98-1	4.08	4.08	4.04	4.07	0.63	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.10	1.05	1.14	1.10	3.84	ppb
Phosphorus	31-2	-19.83	-16.35	-20.65	-18.94		ppb
Potassium	39-2	477.34	478.90	473.05	476.43	0.64	ppb
Rhodium	103-1				96		%
Rhodium	103-2				101		%
Scandium	45-1				92		%
Scandium	45-2				98		%
Selenium	82-1	4.41	4.75	4.61	4.59	3.81	ppb
Selenium	77-2	6.23	4.15	2.79	4.39	39.46	ppb
Selenium	78-2	5.72	4.36	5.08	5.05	13.46	ppb
Silicon	28-1	2.90	3.09	2.86	2.95	4.08	ppb
Silver	107-1	0.88	0.92	0.88	0.89	2.34	ppb
Silver	109-1	0.88	0.93	0.92	0.91	3.14	ppb
Sodium	23-2	453.87	455.16	454.19	454.41	0.15	ppb
Strontium	86-1	0.86	0.88	0.91	0.88	2.41	ppb
Strontium	88-1	0.88	0.87	0.88	0.88	0.84	ppb
Sulfur	34-1	-1096.66	-1015.58	-1021.76	-1044.67		ppb
Terbium	159-1				98		%
Terbium	159-2				104		%
Thallium	203-1	0.75	0.75	0.75	0.75	0.18	ppb
Thallium	205-1	0.76	0.77	0.77	0.77	1.28	ppb
Tin	118-1	-0.01	0.00	0.35	0.11	182.14	ppb
Titanium	47-1	4.28	4.44	4.42	4.38	1.96	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

PB164572BSD

Instrumnet Name :

P7

Client Sample ID :

LCSD572

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 20:55:31

DataFile Name :

065LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.74	0.76	0.74	0.75	1.40	ppb
Vanadium	51-2	4.88	4.88	4.87	4.88	0.10	ppb
Yttrium	89-1				96		%
Yttrium	89-2				101		%
Zinc	66-2	4.55	4.70	4.83	4.69	2.93	ppb
Zirconium	90-1	0.91	0.89	0.92	0.91	1.72	ppb
Zirconium	91-1	0.92	0.89	0.94	0.92	2.53	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-03 Instrumnet Name : P7
Client Sample ID : MYE4L6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:58:46 DataFile Name : 066SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	19385.65	19498.85	19444.90	19443.13	0.29	ppb
Antimony	121-1	0.19	0.20	0.23	0.21	10.65	ppb
Arsenic	75-2	19.01	18.33	18.92	18.75	1.98	ppb
Barium	135-1	197.94	202.62	211.54	204.03	3.39	ppb
Barium	137-1	200.44	204.29	211.05	205.26	2.62	ppb
Beryllium	9-1	0.69	0.82	0.76	0.76	9.11	ppb
Bismuth	209-1				97		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.94	1.06	0.93	0.98	7.58	ppb
Cadmium	106-1	0.41	0.28	-0.69	0.00	48574.77	ppb
Cadmium	111-1	0.37	0.34	0.29	0.34	12.09	ppb
Calcium	43-1	19650.45	19756.24	20953.43	20120.04	3.60	ppb
Calcium	44-1	18681.53	18574.32	19666.31	18974.05	3.17	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	45.79	45.70	46.13	45.87	0.50	ppb
Cobalt	59-2	26.40	26.20	26.49	26.36	0.57	ppb
Copper	63-2	69.12	68.78	69.37	69.09	0.43	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				99		%
Holmium	165-2				106		%
Indium	115-1				95		%
Indium	115-2				101		%
Iron	56-2	43356.94	42689.42	42525.97	42857.44	1.03	ppb
Iron	57-2	42745.14	42218.55	42855.87	42606.52	0.80	ppb
Iron	54-2	43478.16	43090.62	42784.76	43117.85	0.81	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-03 Instrumnet Name : P7
Client Sample ID : MYE4L6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:58:46 DataFile Name : 066SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	19.95	19.85	21.40	20.40	4.25	ppb
Lead	207-1	18.85	18.75	20.04	19.21	3.74	ppb
Lead	208-1	19.05	19.12	20.40	19.52	3.91	ppb
Lithium	6-1				92		%
Magnesium	24-2	15806.94	15838.66	15878.05	15841.22	0.22	ppb
Manganese	55-2	1248.17	1233.35	1211.65	1231.05	1.49	ppb
Molybdenum	94-1	11.76	11.59	12.25	11.87	2.89	ppb
Molybdenum	95-1	0.39	0.42	0.48	0.43	9.91	ppb
Molybdenum	96-1	1.88	1.96	2.03	1.96	3.83	ppb
Molybdenum	97-1	0.38	0.40	0.41	0.40	3.74	ppb
Molybdenum	98-1	0.38	0.42	0.40	0.40	4.36	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	47.43	46.81	48.45	47.56	1.74	ppb
Phosphorus	31-2	2069.83	2056.42	1977.03	2034.43	2.47	ppb
Potassium	39-2	4361.45	4337.40	4282.80	4327.22	0.93	ppb
Rhodium	103-1				93		%
Rhodium	103-2				100		%
Scandium	45-1				93		%
Scandium	45-2				101		%
Selenium	82-1	0.62	0.35	0.19	0.39	55.13	ppb
Selenium	77-2	9.04	10.65	13.21	10.97	19.20	ppb
Selenium	78-2	1.45	1.96	1.08	1.50	29.26	ppb
Silicon	28-1	1277.56	1290.00	1354.86	1307.48	3.17	ppb
Silver	107-1	0.08	0.09	0.09	0.09	4.49	ppb
Silver	109-1	0.08	0.10	0.09	0.09	9.62	ppb
Sodium	23-2	264.94	261.38	262.30	262.87	0.70	ppb
Strontium	86-1	58.99	59.44	62.66	60.36	3.32	ppb
Strontium	88-1	62.22	60.86	64.46	62.51	2.91	ppb
Sulfur	34-1	-1045.22	-1191.21	-1045.88	-1094.10		ppb
Terbium	159-1				99		%
Terbium	159-2				106		%
Thallium	203-1	0.12	0.15	0.13	0.13	8.22	ppb
Thallium	205-1	0.13	0.13	0.13	0.13	2.53	ppb
Tin	118-1	0.30	1.20	0.35	0.62	82.01	ppb
Titanium	47-1	596.86	599.06	635.43	610.45	3.55	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-03

Instrumnet Name :

P7

Client Sample ID :

MYE4L6

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 20:58:46

DataFile Name :

066SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.55	0.55	0.58	0.56	3.36	ppb
Vanadium	51-2	86.49	85.38	85.91	85.93	0.64	ppb
Yttrium	89-1				98		%
Yttrium	89-2				105		%
Zinc	66-2	131.73	132.26	132.91	132.30	0.45	ppb
Zirconium	90-1	6.69	6.77	7.26	6.91	4.49	ppb
Zirconium	91-1	6.76	6.95	7.20	6.97	3.15	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-04 Instrumnet Name : P7
Client Sample ID : MYE4L7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:01:55 DataFile Name : 067SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	28428.57	28148.13	27878.64	28151.78	0.98	ppb
Antimony	121-1	0.18	0.19	0.18	0.19	3.17	ppb
Arsenic	75-2	29.05	31.50	30.89	30.48	4.20	ppb
Barium	135-1	360.66	372.46	374.66	369.26	2.04	ppb
Barium	137-1	393.10	404.41	402.14	399.88	1.50	ppb
Beryllium	9-1	1.02	1.22	1.27	1.17	11.14	ppb
Bismuth	209-1				101		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.96	1.16	1.29	1.14	14.59	ppb
Cadmium	106-1	0.30	0.59	0.54	0.48	32.37	ppb
Cadmium	111-1	0.67	0.71	0.64	0.67	5.46	ppb
Calcium	43-1	27294.64	27879.02	27939.19	27704.28	1.29	ppb
Calcium	44-1	25560.75	26345.13	26278.27	26061.39	1.67	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	59.95	59.34	59.90	59.73	0.57	ppb
Cobalt	59-2	37.07	36.77	36.65	36.83	0.59	ppb
Copper	63-2	137.33	136.73	136.46	136.84	0.33	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				109		%
Indium	115-1				97		%
Indium	115-2				102		%
Iron	56-2	55369.22	55199.33	54930.87	55166.47	0.40	ppb
Iron	57-2	54660.19	53655.81	54588.68	54301.56	1.03	ppb
Iron	54-2	55390.13	54694.74	55049.34	55044.74	0.63	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-04 Instrumnet Name : P7
Client Sample ID : MYE4L7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:01:55 DataFile Name : 067SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	32.58	33.98	33.93	33.50	2.37	ppb
Lead	207-1	30.79	32.83	32.15	31.92	3.26	ppb
Lead	208-1	31.44	32.96	32.69	32.36	2.50	ppb
Lithium	6-1				97		%
Magnesium	24-2	16977.89	16673.18	16694.99	16782.02	1.01	ppb
Manganese	55-2	2093.72	2073.20	2078.77	2081.89	0.51	ppb
Molybdenum	94-1	11.43	11.75	11.86	11.68	1.90	ppb
Molybdenum	95-1	0.85	0.86	0.87	0.86	1.35	ppb
Molybdenum	96-1	2.35	2.45	2.38	2.39	2.02	ppb
Molybdenum	97-1	0.83	0.84	0.86	0.84	1.60	ppb
Molybdenum	98-1	0.77	0.80	0.80	0.79	2.31	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	55.24	54.27	53.85	54.45	1.31	ppb
Phosphorus	31-2	2001.91	1981.10	2022.25	2001.75	1.03	ppb
Potassium	39-2	7817.98	7696.95	7744.79	7753.24	0.79	ppb
Rhodium	103-1				92		%
Rhodium	103-2				101		%
Scandium	45-1				94		%
Scandium	45-2				103		%
Selenium	82-1	0.51	0.85	0.62	0.66	26.93	ppb
Selenium	77-2	14.61	18.08	17.95	16.88	11.65	ppb
Selenium	78-2	0.99	3.43	2.98	2.47	52.68	ppb
Silicon	28-1	1595.19	1617.46	1624.66	1612.44	0.95	ppb
Silver	107-1	0.21	0.21	0.20	0.21	2.90	ppb
Silver	109-1	0.22	0.22	0.21	0.22	1.47	ppb
Sodium	23-2	291.27	288.73	287.49	289.16	0.67	ppb
Strontium	86-1	81.59	83.07	83.67	82.78	1.29	ppb
Strontium	88-1	84.53	85.37	87.54	85.81	1.81	ppb
Sulfur	34-1	-1165.69	-1231.12	-1277.91	-1224.91		ppb
Terbium	159-1				103		%
Terbium	159-2				107		%
Thallium	203-1	0.25	0.27	0.27	0.27	3.84	ppb
Thallium	205-1	0.25	0.28	0.27	0.27	4.77	ppb
Tin	118-1	0.24	0.24	0.24	0.24	0.95	ppb
Titanium	47-1	873.50	892.09	891.80	885.80	1.20	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-04

Instrumnet Name :

P7

Client Sample ID :

MYE4L7

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:01:55

DataFile Name :

067SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.89	0.93	0.92	0.91	2.30	ppb
Vanadium	51-2	116.07	113.31	114.25	114.54	1.23	ppb
Yttrium	89-1				101		%
Yttrium	89-2				108		%
Zinc	66-2	166.71	165.18	164.57	165.49	0.67	ppb
Zirconium	90-1	6.47	6.66	6.62	6.58	1.57	ppb
Zirconium	91-1	6.44	6.63	6.63	6.57	1.68	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-05 Instrumnet Name : P7
Client Sample ID : MYE4L8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:05:01 DataFile Name : 068SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	20457.66	20614.82	20466.89	20513.12	0.43	ppb
Antimony	121-1	0.18	0.19	0.20	0.19	5.45	ppb
Arsenic	75-2	24.56	25.47	24.02	24.68	2.97	ppb
Barium	135-1	255.20	260.55	261.19	258.98	1.27	ppb
Barium	137-1	255.54	263.53	262.71	260.59	1.69	ppb
Beryllium	9-1	0.77	0.76	0.90	0.81	9.52	ppb
Bismuth	209-1				103		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.74	0.75	0.82	0.77	5.76	ppb
Cadmium	106-1	-0.18	-0.63	-0.66	-0.49		ppb
Cadmium	111-1	0.45	0.49	0.42	0.45	6.90	ppb
Calcium	43-1	19103.94	19449.95	19574.94	19376.28	1.26	ppb
Calcium	44-1	18050.04	18325.12	18507.75	18294.30	1.26	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	43.69	43.42	43.74	43.61	0.40	ppb
Cobalt	59-2	28.94	28.60	28.92	28.82	0.66	ppb
Copper	63-2	92.04	90.95	92.22	91.74	0.75	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				107		%
Indium	115-1				98		%
Indium	115-2				101		%
Iron	56-2	45253.21	44710.71	45044.81	45002.91	0.61	ppb
Iron	57-2	44860.34	44228.53	44910.64	44666.50	0.85	ppb
Iron	54-2	45127.50	44995.67	45473.71	45198.96	0.55	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-05 Instrumnet Name : P7
Client Sample ID : MYE4L8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:05:01 DataFile Name : 068SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	25.87	26.35	26.92	26.38	1.98	ppb
Lead	207-1	24.37	25.49	25.67	25.18	2.80	ppb
Lead	208-1	24.92	25.60	25.91	25.48	1.99	ppb
Lithium	6-1				98		%
Magnesium	24-2	12815.08	12693.33	12725.49	12744.63	0.50	ppb
Manganese	55-2	1499.97	1529.35	1534.35	1521.23	1.22	ppb
Molybdenum	94-1	8.43	8.64	8.79	8.62	2.15	ppb
Molybdenum	95-1	0.68	0.73	0.71	0.71	3.46	ppb
Molybdenum	96-1	1.82	1.82	1.85	1.83	0.95	ppb
Molybdenum	97-1	0.65	0.62	0.74	0.67	9.04	ppb
Molybdenum	98-1	0.65	0.69	0.65	0.66	3.54	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	44.03	43.23	43.63	43.63	0.92	ppb
Phosphorus	31-2	1955.32	1884.30	1966.78	1935.47	2.31	ppb
Potassium	39-2	5674.36	5628.05	5761.93	5688.11	1.20	ppb
Rhodium	103-1				93		%
Rhodium	103-2				100		%
Scandium	45-1				93		%
Scandium	45-2				99		%
Selenium	82-1	0.24	0.18	-0.09	0.11	156.70	ppb
Selenium	77-2	10.34	13.87	13.30	12.50	15.15	ppb
Selenium	78-2	2.65	1.63	1.22	1.83	40.31	ppb
Silicon	28-1	2193.25	2215.40	2250.19	2219.61	1.29	ppb
Silver	107-1	0.12	0.13	0.13	0.13	4.84	ppb
Silver	109-1	0.12	0.14	0.14	0.13	6.20	ppb
Sodium	23-2	232.27	229.71	233.70	231.89	0.87	ppb
Strontium	86-1	55.69	56.90	57.50	56.70	1.63	ppb
Strontium	88-1	56.66	58.49	58.79	57.98	1.99	ppb
Sulfur	34-1	-1467.04	-1515.18	-1513.06	-1498.43		ppb
Terbium	159-1				103		%
Terbium	159-2				107		%
Thallium	203-1	0.17	0.17	0.17	0.17	2.32	ppb
Thallium	205-1	0.16	0.17	0.17	0.17	4.57	ppb
Tin	118-1	0.24	0.26	0.23	0.24	5.75	ppb
Titanium	47-1	750.62	765.70	773.71	763.34	1.54	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-05

Instrumnet Name :

P7

Client Sample ID :

MYE4L8

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:05:01

DataFile Name :

068SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.73	0.74	0.76	0.75	2.19	ppb
Vanadium	51-2	90.78	89.82	90.47	90.36	0.54	ppb
Yttrium	89-1				100		%
Yttrium	89-2				105		%
Zinc	66-2	126.86	125.42	129.03	127.10	1.43	ppb
Zirconium	90-1	4.67	4.79	4.83	4.76	1.73	ppb
Zirconium	91-1	4.71	4.83	4.89	4.81	1.88	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-06 Instrumnet Name : P7
Client Sample ID : MYE4L9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:08:10 DataFile Name : 069SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	38446.10	38466.60	38467.16	38459.95	0.03	ppb
Antimony	121-1	0.14	0.16	0.16	0.15	6.16	ppb
Arsenic	75-2	15.74	15.50	15.71	15.65	0.82	ppb
Barium	135-1	85.17	85.52	88.01	86.23	1.79	ppb
Barium	137-1	86.39	86.82	88.59	87.27	1.34	ppb
Beryllium	9-1	0.22	0.26	0.23	0.24	7.33	ppb
Bismuth	209-1				100		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.50	0.51	0.58	0.53	8.19	ppb
Cadmium	106-1	-1.04	-1.26	-1.01	-1.10		ppb
Cadmium	111-1	0.09	0.10	0.14	0.11	22.05	ppb
Calcium	43-1	13368.88	13337.43	13450.75	13385.69	0.44	ppb
Calcium	44-1	12594.24	12474.97	12638.04	12569.08	0.67	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	127.02	126.05	127.79	126.95	0.69	ppb
Cobalt	59-2	63.94	63.40	63.76	63.70	0.43	ppb
Copper	63-2	157.44	156.40	157.41	157.08	0.38	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				105		%
Indium	115-1				96		%
Indium	115-2				99		%
Iron	56-2	80186.70	79848.65	80620.30	80218.55	0.48	ppb
Iron	57-2	81101.82	79678.69	80133.06	80304.52	0.91	ppb
Iron	54-2	81208.28	79906.74	81040.05	80718.36	0.88	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-06 Instrumnet Name : P7
Client Sample ID : MYE4L9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:08:10 DataFile Name : 069SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	7.59	7.77	7.98	7.78	2.50	ppb
Lead	207-1	7.65	7.72	8.13	7.83	3.27	ppb
Lead	208-1	7.64	7.64	7.99	7.76	2.60	ppb
Lithium	6-1				98		%
Magnesium	24-2	32117.05	32202.43	32226.71	32182.06	0.18	ppb
Manganese	55-2	1487.69	1488.25	1494.40	1490.11	0.25	ppb
Molybdenum	94-1	3.60	3.78	3.77	3.71	2.76	ppb
Molybdenum	95-1	0.18	0.18	0.19	0.18	4.42	ppb
Molybdenum	96-1	1.21	1.20	1.28	1.23	3.71	ppb
Molybdenum	97-1	0.16	0.19	0.19	0.18	10.42	ppb
Molybdenum	98-1	0.14	0.14	0.14	0.14	2.35	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	112.57	113.21	113.62	113.14	0.47	ppb
Phosphorus	31-2	794.19	746.81	796.76	779.25	3.61	ppb
Potassium	39-2	2308.78	2437.38	2443.49	2396.55	3.17	ppb
Rhodium	103-1				92		%
Rhodium	103-2				99		%
Scandium	45-1				97		%
Scandium	45-2				104		%
Selenium	82-1	0.15	-0.32	0.25	0.02	1368.28	ppb
Selenium	77-2	3.61	6.01	4.08	4.57	27.95	ppb
Selenium	78-2	1.67	0.16	0.95	0.93	81.54	ppb
Silicon	28-1	1215.62	1221.52	1235.38	1224.17	0.83	ppb
Silver	107-1	0.06	0.06	0.07	0.06	2.88	ppb
Silver	109-1	0.06	0.07	0.07	0.07	5.75	ppb
Sodium	23-2	366.69	364.65	365.07	365.47	0.29	ppb
Strontium	86-1	45.00	45.49	46.48	45.66	1.65	ppb
Strontium	88-1	45.03	45.04	46.49	45.52	1.85	ppb
Sulfur	34-1	-1825.36	-1783.70	-1700.06	-1769.71		ppb
Terbium	159-1				102		%
Terbium	159-2				105		%
Thallium	203-1	0.05	0.03	0.05	0.04	20.67	ppb
Thallium	205-1	0.04	0.03	0.04	0.04	8.93	ppb
Tin	118-1	0.27	0.29	0.30	0.28	6.01	ppb
Titanium	47-1	593.99	598.79	600.64	597.81	0.57	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-06

Instrumnet Name :

P7

Client Sample ID :

MYE4L9

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:08:10

DataFile Name :

069SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.15	0.15	0.16	0.16	0.78	ppb
Vanadium	51-2	214.26	215.33	215.34	214.98	0.29	ppb
Yttrium	89-1				97		%
Yttrium	89-2				102		%
Zinc	66-2	101.84	101.32	102.80	101.99	0.73	ppb
Zirconium	90-1	2.12	2.13	2.18	2.14	1.58	ppb
Zirconium	91-1	2.19	2.14	2.20	2.18	1.31	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-07 Instrumnet Name : P7
Client Sample ID : MYE4M0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:11:17 DataFile Name : 070SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	17666.93	17753.87	17509.87	17643.56	0.70	ppb
Antimony	121-1	0.28	0.29	0.30	0.29	4.11	ppb
Arsenic	75-2	40.55	40.01	40.22	40.26	0.68	ppb
Barium	135-1	298.57	306.48	309.24	304.76	1.82	ppb
Barium	137-1	300.60	306.80	309.80	305.73	1.54	ppb
Beryllium	9-1	0.61	0.63	0.62	0.62	1.80	ppb
Bismuth	209-1				101		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.06	1.09	1.40	1.18	15.97	ppb
Cadmium	106-1	0.61	0.48	0.38	0.49	24.14	ppb
Cadmium	111-1	0.50	0.51	0.42	0.48	10.82	ppb
Calcium	43-1	26833.28	27197.32	28127.30	27385.97	2.44	ppb
Calcium	44-1	25364.30	25648.48	26280.68	25764.49	1.82	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	49.09	48.31	48.03	48.48	1.13	ppb
Cobalt	59-2	23.13	23.09	23.20	23.14	0.22	ppb
Copper	63-2	51.22	51.40	51.75	51.46	0.52	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				105		%
Indium	115-1				97		%
Indium	115-2				100		%
Iron	56-2	34968.56	34905.63	34552.65	34808.94	0.64	ppb
Iron	57-2	35192.03	34300.51	34537.39	34676.64	1.33	ppb
Iron	54-2	34977.69	35008.10	34780.77	34922.19	0.35	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-07 Instrumnet Name : P7
Client Sample ID : MYE4M0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:11:17 DataFile Name : 070SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	15.23	15.65	15.63	15.51	1.52	ppb
Lead	207-1	14.75	15.26	15.29	15.10	2.02	ppb
Lead	208-1	14.81	15.31	15.33	15.15	1.97	ppb
Lithium	6-1				97		%
Magnesium	24-2	13643.95	13712.19	13689.90	13682.01	0.25	ppb
Manganese	55-2	1288.67	1285.81	1271.37	1281.95	0.72	ppb
Molybdenum	94-1	10.44	10.32	10.69	10.48	1.81	ppb
Molybdenum	95-1	0.46	0.45	0.43	0.45	3.86	ppb
Molybdenum	96-1	1.69	1.82	1.76	1.76	3.79	ppb
Molybdenum	97-1	0.45	0.39	0.42	0.42	6.95	ppb
Molybdenum	98-1	0.40	0.43	0.42	0.42	3.61	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	48.53	48.46	48.50	48.50	0.07	ppb
Phosphorus	31-2	1554.34	1546.55	1546.23	1549.04	0.30	ppb
Potassium	39-2	3252.24	3223.28	3219.37	3231.63	0.56	ppb
Rhodium	103-1				93		%
Rhodium	103-2				100		%
Scandium	45-1				93		%
Scandium	45-2				101		%
Selenium	82-1	0.21	0.24	0.69	0.38	70.06	ppb
Selenium	77-2	6.09	6.07	5.37	5.84	6.95	ppb
Selenium	78-2	0.71	1.45	1.56	1.24	37.18	ppb
Silicon	28-1	1190.04	1208.11	1240.64	1212.93	2.11	ppb
Silver	107-1	0.15	0.14	0.15	0.15	2.99	ppb
Silver	109-1	0.14	0.15	0.14	0.15	2.98	ppb
Sodium	23-2	344.20	342.63	344.80	343.87	0.33	ppb
Strontium	86-1	153.13	154.69	157.65	155.16	1.48	ppb
Strontium	88-1	156.70	162.13	161.74	160.19	1.89	ppb
Sulfur	34-1	-313.72	-204.11	-36.73	-184.85		ppb
Terbium	159-1				102		%
Terbium	159-2				106		%
Thallium	203-1	0.12	0.12	0.11	0.11	4.40	ppb
Thallium	205-1	0.12	0.12	0.11	0.11	1.14	ppb
Tin	118-1	0.31	0.34	0.33	0.33	4.65	ppb
Titanium	47-1	301.80	304.97	316.04	307.60	2.43	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-07

Instrumnet Name :

P7

Client Sample ID :

MYE4M0

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:11:17

DataFile Name :

070SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.64	0.66	0.64	0.65	1.09	ppb
Vanadium	51-2	72.43	72.04	72.11	72.19	0.29	ppb
Yttrium	89-1				96		%
Yttrium	89-2				103		%
Zinc	66-2	226.11	226.04	225.65	225.93	0.11	ppb
Zirconium	90-1	5.95	6.09	6.21	6.08	2.13	ppb
Zirconium	91-1	6.05	6.21	6.37	6.21	2.54	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08 Instrumnet Name : P7
Client Sample ID : MYE4M1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:14:26 DataFile Name : 071SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	21870.89	21667.56	21907.99	21815.48	0.59	ppb
Antimony	121-1	0.40	0.37	0.37	0.38	3.74	ppb
Arsenic	75-2	50.99	51.32	51.14	51.15	0.33	ppb
Barium	135-1	348.18	347.95	357.01	351.04	1.47	ppb
Barium	137-1	386.39	377.82	393.60	385.93	2.05	ppb
Beryllium	9-1	0.83	0.90	0.80	0.84	5.96	ppb
Bismuth	209-1				102		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.45	2.05	2.68	2.39	13.37	ppb
Cadmium	106-1	1.80	1.37	1.38	1.52	15.92	ppb
Cadmium	111-1	1.55	1.59	1.58	1.58	1.32	ppb
Calcium	43-1	22291.31	22505.23	22375.51	22390.69	0.48	ppb
Calcium	44-1	21112.10	21216.05	21036.42	21121.52	0.43	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	46.72	46.84	47.89	47.15	1.36	ppb
Cobalt	59-2	29.54	29.40	29.79	29.58	0.67	ppb
Copper	63-2	194.23	194.19	195.46	194.63	0.37	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				106		%
Indium	115-1				97		%
Indium	115-2				100		%
Iron	56-2	42924.21	43081.57	43411.56	43139.11	0.58	ppb
Iron	57-2	42755.05	42105.42	43019.59	42626.69	1.10	ppb
Iron	54-2	43456.30	43135.17	43306.92	43299.46	0.37	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08 Instrumnet Name : P7
Client Sample ID : MYE4M1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:14:26 DataFile Name : 071SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	98.72	99.88	101.48	100.02	1.38	ppb
Lead	207-1	96.26	98.37	99.44	98.03	1.65	ppb
Lead	208-1	106.58	106.36	108.81	107.25	1.26	ppb
Lithium	6-1				98		%
Magnesium	24-2	12414.05	12411.56	12586.21	12470.61	0.80	ppb
Manganese	55-2	1575.10	1582.56	1593.69	1583.78	0.59	ppb
Molybdenum	94-1	17.67	17.55	17.88	17.70	0.94	ppb
Molybdenum	95-1	0.29	0.31	0.29	0.30	4.12	ppb
Molybdenum	96-1	2.43	2.45	2.50	2.46	1.32	ppb
Molybdenum	97-1	0.26	0.28	0.26	0.27	4.56	ppb
Molybdenum	98-1	0.26	0.26	0.26	0.26	1.18	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	51.91	51.57	52.17	51.88	0.57	ppb
Phosphorus	31-2	1990.80	2062.84	2013.47	2022.37	1.82	ppb
Potassium	39-2	6241.96	6293.76	6415.89	6317.20	1.41	ppb
Rhodium	103-1				93		%
Rhodium	103-2				100		%
Scandium	45-1				94		%
Scandium	45-2				101		%
Selenium	82-1	-0.15	0.61	0.05	0.17	232.54	ppb
Selenium	77-2	12.36	9.87	8.69	10.30	18.17	ppb
Selenium	78-2	3.03	0.95	1.53	1.84	58.27	ppb
Silicon	28-1	1476.71	1494.28	1480.99	1483.99	0.62	ppb
Silver	107-1	0.43	0.46	0.47	0.45	4.28	ppb
Silver	109-1	0.45	0.46	0.46	0.46	0.58	ppb
Sodium	23-2	153.68	155.77	154.97	154.81	0.68	ppb
Strontium	86-1	68.68	70.51	71.38	70.19	1.96	ppb
Strontium	88-1	70.83	72.85	73.23	72.30	1.79	ppb
Sulfur	34-1	-800.09	-867.53	-993.06	-886.89		ppb
Terbium	159-1				102		%
Terbium	159-2				106		%
Thallium	203-1	0.18	0.19	0.19	0.19	2.76	ppb
Thallium	205-1	0.18	0.17	0.18	0.18	3.21	ppb
Tin	118-1	0.44	0.44	0.45	0.44	1.51	ppb
Titanium	47-1	463.96	473.97	470.88	469.60	1.09	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-08

Instrumnet Name :

P7

Client Sample ID :

MYE4M1

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:14:26

DataFile Name :

071SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.55	0.55	0.55	0.55	0.69	ppb
Vanadium	51-2	77.19	77.08	78.48	77.59	1.00	ppb
Yttrium	89-1				99		%
Yttrium	89-2				104		%
Zinc	66-2	269.71	268.67	272.84	270.41	0.80	ppb
Zirconium	90-1	10.22	10.47	10.50	10.40	1.49	ppb
Zirconium	91-1	10.51	10.65	10.56	10.57	0.68	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-09 Instrumnet Name : P7
Client Sample ID : MYE4M1D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:17:31 DataFile Name : 072SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	21506.37	21542.49	21482.84	21510.57	0.14	ppb
Antimony	121-1	0.36	0.37	0.39	0.37	4.67	ppb
Arsenic	75-2	49.00	48.15	49.40	48.85	1.31	ppb
Barium	135-1	336.84	344.90	345.72	342.49	1.43	ppb
Barium	137-1	368.71	369.04	373.65	370.46	0.75	ppb
Beryllium	9-1	0.73	0.87	0.91	0.84	10.93	ppb
Bismuth	209-1				101		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.06	2.47	2.46	2.33	9.89	ppb
Cadmium	106-1	1.04	1.62	1.11	1.26	25.43	ppb
Cadmium	111-1	1.53	1.54	1.60	1.56	2.29	ppb
Calcium	43-1	21516.01	21950.24	22348.60	21938.28	1.90	ppb
Calcium	44-1	20369.22	20686.38	21081.33	20712.31	1.72	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	46.05	46.37	45.79	46.07	0.64	ppb
Cobalt	59-2	28.97	29.02	28.32	28.77	1.35	ppb
Copper	63-2	190.18	190.12	188.82	189.71	0.40	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				106		%
Indium	115-1				96		%
Indium	115-2				99		%
Iron	56-2	42113.12	42420.54	42004.39	42179.35	0.51	ppb
Iron	57-2	41836.95	42393.63	41527.30	41919.29	1.05	ppb
Iron	54-2	42008.11	42869.67	41859.08	42245.62	1.29	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-09 Instrumnet Name : P7
Client Sample ID : MYE4M1D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:17:31 DataFile Name : 072SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	96.60	98.01	99.25	97.95	1.35	ppb
Lead	207-1	95.25	95.70	97.97	96.30	1.52	ppb
Lead	208-1	103.72	104.36	105.81	104.63	1.02	ppb
Lithium	6-1				97		%
Magnesium	24-2	12197.61	12315.36	12133.93	12215.63	0.75	ppb
Manganese	55-2	1531.02	1539.89	1540.38	1537.10	0.34	ppb
Molybdenum	94-1	17.62	18.11	18.09	17.94	1.54	ppb
Molybdenum	95-1	0.25	0.29	0.30	0.28	8.86	ppb
Molybdenum	96-1	2.38	2.41	2.40	2.40	0.64	ppb
Molybdenum	97-1	0.24	0.27	0.26	0.26	6.93	ppb
Molybdenum	98-1	0.25	0.26	0.24	0.25	3.19	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	50.87	51.12	50.81	50.94	0.32	ppb
Phosphorus	31-2	1972.80	1989.20	1965.77	1975.93	0.61	ppb
Potassium	39-2	6115.47	6212.25	6243.83	6190.52	1.08	ppb
Rhodium	103-1				92		%
Rhodium	103-2				99		%
Scandium	45-1				93		%
Scandium	45-2				100		%
Selenium	82-1	-0.01	-0.10	0.25	0.05	391.74	ppb
Selenium	77-2	9.88	12.49	11.99	11.45	12.11	ppb
Selenium	78-2	1.99	2.21	1.29	1.83	26.38	ppb
Silicon	28-1	1471.49	1500.95	1521.83	1498.09	1.69	ppb
Silver	107-1	0.44	0.45	0.45	0.45	1.24	ppb
Silver	109-1	0.43	0.43	0.46	0.44	3.94	ppb
Sodium	23-2	152.91	153.37	151.96	152.75	0.47	ppb
Strontium	86-1	67.78	69.60	69.59	68.99	1.52	ppb
Strontium	88-1	69.28	71.41	71.39	70.69	1.73	ppb
Sulfur	34-1	-1096.46	-1086.22	-1042.18	-1074.95		ppb
Terbium	159-1				101		%
Terbium	159-2				105		%
Thallium	203-1	0.18	0.19	0.18	0.19	1.63	ppb
Thallium	205-1	0.18	0.18	0.18	0.18	0.49	ppb
Tin	118-1	0.42	0.41	0.42	0.42	1.46	ppb
Titanium	47-1	452.15	458.84	467.01	459.33	1.62	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-09

Instrumnet Name :

P7

Client Sample ID :

MYE4M1D

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:17:31

DataFile Name :

072SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.52	0.55	0.56	0.54	3.19	ppb
Vanadium	51-2	75.58	76.58	75.47	75.88	0.81	ppb
Yttrium	89-1				97		%
Yttrium	89-2				103		%
Zinc	66-2	262.79	264.62	263.65	263.69	0.35	ppb
Zirconium	90-1	10.46	10.73	10.78	10.66	1.64	ppb
Zirconium	91-1	10.54	10.98	10.79	10.77	2.01	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08LX5 Instrumnet Name : P7
Client Sample ID : MYE4M1L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 21:20:38 DataFile Name : 073SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	4267.62	4286.52	4251.84	4268.66	0.41	ppb
Antimony	121-1	0.07	0.05	0.05	0.06	10.64	ppb
Arsenic	75-2	10.57	10.19	10.21	10.32	2.11	ppb
Barium	135-1	67.47	66.77	65.79	66.67	1.26	ppb
Barium	137-1	67.36	67.54	66.68	67.19	0.67	ppb
Beryllium	9-1	0.09	0.17	0.13	0.13	26.99	ppb
Bismuth	209-1				101		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.48	0.38	0.32	0.39	19.91	ppb
Cadmium	106-1	-1.81	-1.29	-2.00	-1.70		ppb
Cadmium	111-1	0.15	0.22	0.15	0.17	23.75	ppb
Calcium	43-1	4303.82	4406.32	4339.97	4350.04	1.20	ppb
Calcium	44-1	4078.40	4106.03	4126.50	4103.64	0.59	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	9.31	9.33	8.99	9.21	2.07	ppb
Cobalt	59-2	5.64	5.84	5.71	5.73	1.74	ppb
Copper	63-2	38.12	37.99	37.69	37.93	0.58	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				104		%
Indium	115-1				96		%
Indium	115-2				101		%
Iron	56-2	8363.31	8432.55	8206.85	8334.24	1.39	ppb
Iron	57-2	8574.27	8541.38	8516.51	8544.05	0.34	ppb
Iron	54-2	8557.87	8455.57	8456.77	8490.07	0.69	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08LX5 Instrumnet Name : P7
Client Sample ID : MYE4M1L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 21:20:38 DataFile Name : 073SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	18.50	18.57	18.87	18.65	1.05	ppb
Lead	207-1	18.38	18.22	18.64	18.42	1.16	ppb
Lead	208-1	18.38	18.34	18.67	18.46	0.96	ppb
Lithium	6-1				96		%
Magnesium	24-2	2444.31	2447.80	2445.69	2445.93	0.07	ppb
Manganese	55-2	308.33	306.04	305.96	306.78	0.44	ppb
Molybdenum	94-1	3.33	3.36	3.33	3.34	0.53	ppb
Molybdenum	95-1	0.03	0.01	0.02	0.02	56.74	ppb
Molybdenum	96-1	0.39	0.45	0.43	0.42	6.36	ppb
Molybdenum	97-1	-0.01	0.01	0.01	0.01	167.95	ppb
Molybdenum	98-1	0.00	0.00	0.01	0.00	60.29	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	10.15	10.21	10.41	10.26	1.36	ppb
Phosphorus	31-2	348.16	365.84	354.00	356.00	2.53	ppb
Potassium	39-2	1275.45	1271.72	1272.61	1273.26	0.15	ppb
Rhodium	103-1				94		%
Rhodium	103-2				101		%
Scandium	45-1				91		%
Scandium	45-2				98		%
Selenium	82-1	-0.02	-0.25	-0.43	-0.23		ppb
Selenium	77-2	2.09	0.94	1.39	1.48	39.33	ppb
Selenium	78-2	-1.17	-0.79	0.30	-0.55		ppb
Silicon	28-1	283.45	286.57	304.91	291.64	3.98	ppb
Silver	107-1	0.08	0.09	0.08	0.08	6.71	ppb
Silver	109-1	0.08	0.08	0.09	0.08	2.83	ppb
Sodium	23-2	15.70	16.43	17.36	16.49	5.04	ppb
Strontium	86-1	13.32	13.83	13.76	13.64	2.03	ppb
Strontium	88-1	13.43	13.71	13.68	13.61	1.12	ppb
Sulfur	34-1	-1608.09	-1562.50	-1566.07	-1578.89		ppb
Terbium	159-1				101		%
Terbium	159-2				105		%
Thallium	203-1	0.03	0.02	0.03	0.03	22.06	ppb
Thallium	205-1	0.02	0.03	0.03	0.03	12.50	ppb
Tin	118-1	0.07	0.07	0.07	0.07	2.38	ppb
Titanium	47-1	91.65	91.83	91.86	91.78	0.12	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-08LX5

Instrumnet Name :

P7

Client Sample ID :

MYE4M1L

Dilution Factor :

5

Date & Time Acquired :

2024-11-02 21:20:38

DataFile Name :

073SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.10	0.10	0.09	0.10	2.60	ppb
Vanadium	51-2	15.06	14.81	14.99	14.95	0.87	ppb
Yttrium	89-1				94		%
Yttrium	89-2				100		%
Zinc	66-2	52.29	51.95	52.94	52.39	0.96	ppb
Zirconium	90-1	2.04	2.04	2.09	2.06	1.32	ppb
Zirconium	91-1	1.98	2.04	2.10	2.04	3.02	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-10 Instrumnet Name : P7
Client Sample ID : MYE4M1S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:23:49 DataFile Name : 074SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	22305.64	22369.82	22225.55	22300.33	0.32	ppb
Antimony	121-1	55.78	55.77	63.13	58.23	7.29	ppb
Arsenic	75-2	71.43	71.40	71.84	71.56	0.34	ppb
Barium	135-1	1498.54	1500.65	1692.52	1563.91	7.12	ppb
Barium	137-1	1503.71	1523.56	1708.98	1578.75	7.17	ppb
Beryllium	9-1	30.76	31.95	34.33	32.35	5.62	ppb
Bismuth	209-1				99		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	28.13	27.93	31.81	29.29	7.47	ppb
Cadmium	106-1	26.76	25.77	30.06	27.53	8.16	ppb
Cadmium	111-1	29.32	29.49	33.42	30.75	7.54	ppb
Calcium	43-1	21384.54	21815.42	24007.58	22402.51	6.28	ppb
Calcium	44-1	20182.79	20462.85	22554.01	21066.55	6.15	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	165.41	165.40	164.12	164.98	0.45	ppb
Cobalt	59-2	330.64	327.24	328.28	328.72	0.53	ppb
Copper	63-2	341.28	336.44	340.16	339.29	0.75	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				98		%
Holmium	165-2				105		%
Indium	115-1				92		%
Indium	115-2				100		%
Iron	56-2	42961.32	42862.16	42830.27	42884.58	0.16	ppb
Iron	57-2	42907.22	42794.79	42420.59	42707.53	0.60	ppb
Iron	54-2	43171.01	43251.57	42551.72	42991.43	0.89	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-10 Instrumnet Name : P7
Client Sample ID : MYE4M1S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:23:49 DataFile Name : 074SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	106.60	107.36	117.47	110.48	5.49	ppb
Lead	207-1	105.28	105.64	115.34	108.75	5.25	ppb
Lead	208-1	113.36	115.10	125.76	118.07	5.69	ppb
Lithium	6-1				92		%
Magnesium	24-2	12077.53	12063.50	12130.31	12090.45	0.29	ppb
Manganese	55-2	1832.26	1833.83	1830.08	1832.06	0.10	ppb
Molybdenum	94-1	17.97	18.37	20.22	18.86	6.36	ppb
Molybdenum	95-1	0.30	0.29	0.32	0.30	5.46	ppb
Molybdenum	96-1	2.51	2.45	2.75	2.57	6.01	ppb
Molybdenum	97-1	0.26	0.27	0.30	0.28	7.02	ppb
Molybdenum	98-1	0.26	0.27	0.27	0.27	1.77	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	338.44	337.75	338.66	338.28	0.14	ppb
Phosphorus	31-2	1970.58	1923.39	1948.93	1947.63	1.21	ppb
Potassium	39-2	6189.04	6163.67	6177.81	6176.84	0.21	ppb
Rhodium	103-1				89		%
Rhodium	103-2				100		%
Scandium	45-1				90		%
Scandium	45-2				100		%
Selenium	82-1	62.66	63.42	71.06	65.71	7.08	ppb
Selenium	77-2	67.78	76.86	74.62	73.09	6.47	ppb
Selenium	78-2	68.46	67.84	62.70	66.34	4.76	ppb
Silicon	28-1	1479.83	1505.98	1650.95	1545.58	5.96	ppb
Silver	107-1	24.30	24.29	27.53	25.37	7.35	ppb
Silver	109-1	24.34	24.43	27.36	25.38	6.76	ppb
Sodium	23-2	149.74	149.50	151.71	150.32	0.81	ppb
Strontium	86-1	67.95	68.36	76.27	70.86	6.62	ppb
Strontium	88-1	70.35	69.78	79.05	73.06	7.11	ppb
Sulfur	34-1	-1179.68	-1240.78	-884.93	-1101.79		ppb
Terbium	159-1				99		%
Terbium	159-2				105		%
Thallium	203-1	24.92	25.35	27.34	25.87	4.99	ppb
Thallium	205-1	24.78	25.32	27.43	25.84	5.42	ppb
Tin	118-1	0.36	0.39	0.45	0.40	12.54	ppb
Titanium	47-1	442.43	454.62	499.96	465.67	6.51	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-10 Instrumnet Name : P7
Client Sample ID : MYE4M1S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:23:49 DataFile Name : 074SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.52	0.54	0.59	0.55	6.19	ppb
Vanadium	51-2	369.51	366.41	369.99	368.64	0.53	ppb
Yttrium	89-1				95		%
Yttrium	89-2				104		%
Zinc	66-2	546.93	552.41	549.23	549.52	0.50	ppb
Zirconium	90-1	10.50	10.69	11.87	11.02	6.72	ppb
Zirconium	91-1	10.68	11.04	11.89	11.20	5.58	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08MS Instrumnet Name : P7
Client Sample ID : MYE4M1SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:26:51 DataFile Name : 075SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	25493.97	25582.22	25302.48	25459.56	0.56	ppb
Antimony	121-1	289.33	297.14	296.76	294.41	1.50	ppb
Arsenic	75-2	153.82	157.98	156.15	155.98	1.33	ppb
Barium	135-1	5872.90	6100.90	5993.89	5989.23	1.90	ppb
Barium	137-1	5900.69	6005.08	6004.07	5969.95	1.00	ppb
Beryllium	9-1	146.39	148.58	146.91	147.29	0.78	ppb
Bismuth	209-1				102		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	128.41	134.19	132.58	131.73	2.27	ppb
Cadmium	106-1	125.58	125.30	127.49	126.12	0.95	ppb
Cadmium	111-1	138.41	139.90	139.09	139.13	0.53	ppb
Calcium	43-1	20743.63	20770.12	21145.53	20886.43	1.08	ppb
Calcium	44-1	19374.45	19717.35	20041.76	19711.19	1.69	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	611.29	612.34	606.11	609.91	0.55	ppb
Cobalt	59-2	1454.82	1473.71	1461.58	1463.37	0.65	ppb
Copper	63-2	862.81	880.19	866.81	869.93	1.05	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				106		%
Indium	115-1				95		%
Indium	115-2				99		%
Iron	56-2	43146.44	43744.45	43281.15	43390.68	0.72	ppb
Iron	57-2	43057.12	42946.50	43257.90	43087.17	0.37	ppb
Iron	54-2	43983.58	43942.61	43741.24	43889.15	0.30	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08MS Instrumnet Name : P7
Client Sample ID : MYE4M1SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:26:51 DataFile Name : 075SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	140.92	144.75	145.25	143.64	1.65	ppb
Lead	207-1	138.30	140.98	143.66	140.98	1.90	ppb
Lead	208-1	150.79	153.82	156.12	153.57	1.74	ppb
Lithium	6-1				95		%
Magnesium	24-2	11564.05	11603.40	11593.60	11587.02	0.18	ppb
Manganese	55-2	2887.91	2874.25	2852.72	2871.63	0.62	ppb
Molybdenum	94-1	17.92	18.00	18.14	18.02	0.63	ppb
Molybdenum	95-1	0.29	0.32	0.31	0.31	4.18	ppb
Molybdenum	96-1	2.54	2.52	2.48	2.51	1.12	ppb
Molybdenum	97-1	0.28	0.26	0.27	0.27	4.46	ppb
Molybdenum	98-1	0.30	0.30	0.29	0.30	1.08	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1503.85	1519.83	1493.15	1505.61	0.89	ppb
Phosphorus	31-2	1852.25	1855.58	1871.16	1859.66	0.54	ppb
Potassium	39-2	5875.90	5968.81	5901.81	5915.51	0.81	ppb
Rhodium	103-1				92		%
Rhodium	103-2				99		%
Scandium	45-1				92		%
Scandium	45-2				99		%
Selenium	82-1	309.88	303.06	307.57	306.84	1.13	ppb
Selenium	77-2	324.74	310.92	317.92	317.86	2.17	ppb
Selenium	78-2	306.31	309.52	302.40	306.08	1.16	ppb
Silicon	28-1	1453.53	1481.03	1480.51	1471.69	1.07	ppb
Silver	107-1	116.30	125.11	116.57	119.33	4.20	ppb
Silver	109-1	116.38	119.18	117.85	117.80	1.19	ppb
Sodium	23-2	146.25	144.73	144.96	145.32	0.56	ppb
Strontium	86-1	66.23	65.94	67.61	66.59	1.34	ppb
Strontium	88-1	68.51	68.36	68.64	68.51	0.21	ppb
Sulfur	34-1	-1289.23	-1359.50	-1387.42	-1345.38		ppb
Terbium	159-1				101		%
Terbium	159-2				105		%
Thallium	203-1	120.58	123.48	124.25	122.77	1.58	ppb
Thallium	205-1	141.18	142.07	144.07	142.44	1.04	ppb
Tin	118-1	0.37	0.40	0.41	0.40	5.18	ppb
Titanium	47-1	429.25	430.42	434.35	431.34	0.62	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-08MS

Instrumnet Name :

P7

Client Sample ID :

MYE4M1SRE

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:26:51

DataFile Name :

075SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.50	0.52	0.53	0.52	3.11	ppb
Vanadium	51-2	1453.18	1465.65	1446.64	1455.16	0.66	ppb
Yttrium	89-1				96		%
Yttrium	89-2				102		%
Zinc	66-2	1635.25	1640.52	1633.47	1636.42	0.22	ppb
Zirconium	90-1	10.75	10.67	10.72	10.71	0.39	ppb
Zirconium	91-1	10.82	10.65	10.82	10.76	0.91	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-11 Instrumnet Name : P7
Client Sample ID : MYE4M2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:29:46 DataFile Name : 076SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	18427.15	18439.22	18473.15	18446.51	0.13	ppb
Antimony	121-1	0.44	0.42	0.45	0.44	3.67	ppb
Arsenic	75-2	25.46	25.69	25.81	25.66	0.70	ppb
Barium	135-1	255.27	257.48	260.56	257.77	1.03	ppb
Barium	137-1	254.32	257.40	260.69	257.47	1.24	ppb
Beryllium	9-1	0.66	0.67	0.73	0.69	5.64	ppb
Bismuth	209-1				104		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.29	0.93	1.43	1.22	21.09	ppb
Cadmium	106-1	0.01	-0.14	0.17	0.01	1378.88	ppb
Cadmium	111-1	0.74	0.68	0.74	0.72	4.69	ppb
Calcium	43-1	16149.53	16449.93	16723.15	16440.87	1.75	ppb
Calcium	44-1	15270.25	15539.43	15553.10	15454.26	1.03	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	44.53	45.68	44.75	44.99	1.36	ppb
Cobalt	59-2	19.59	19.58	19.39	19.52	0.58	ppb
Copper	63-2	64.83	65.58	65.11	65.17	0.58	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				108		%
Indium	115-1				98		%
Indium	115-2				101		%
Iron	56-2	29881.27	29835.62	29581.92	29766.27	0.54	ppb
Iron	57-2	29441.40	29815.59	29573.32	29610.11	0.64	ppb
Iron	54-2	29772.18	30037.15	29955.21	29921.51	0.45	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-11 Instrumnet Name : P7
Client Sample ID : MYE4M2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:29:46 DataFile Name : 076SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	16.90	17.17	17.16	17.08	0.88	ppb
Lead	207-1	16.22	16.74	16.56	16.51	1.58	ppb
Lead	208-1	16.44	16.85	16.76	16.69	1.30	ppb
Lithium	6-1				98		%
Magnesium	24-2	12252.62	12270.62	12194.94	12239.39	0.32	ppb
Manganese	55-2	1172.85	1175.80	1186.43	1178.36	0.61	ppb
Molybdenum	94-1	11.11	11.24	11.37	11.24	1.14	ppb
Molybdenum	95-1	0.51	0.58	0.54	0.54	5.88	ppb
Molybdenum	96-1	1.85	1.89	1.88	1.87	1.04	ppb
Molybdenum	97-1	0.47	0.54	0.54	0.52	7.74	ppb
Molybdenum	98-1	0.50	0.48	0.50	0.49	2.50	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	35.07	35.11	34.94	35.04	0.24	ppb
Phosphorus	31-2	1263.31	1240.95	1204.41	1236.22	2.40	ppb
Potassium	39-2	4226.42	4293.66	4259.83	4259.97	0.79	ppb
Rhodium	103-1				94		%
Rhodium	103-2				101		%
Scandium	45-1				93		%
Scandium	45-2				101		%
Selenium	82-1	0.28	0.39	0.50	0.39	28.88	ppb
Selenium	77-2	11.92	10.29	7.38	9.86	23.31	ppb
Selenium	78-2	1.26	2.64	1.69	1.87	37.91	ppb
Silicon	28-1	1498.63	1519.31	1525.79	1514.58	0.94	ppb
Silver	107-1	0.14	0.13	0.15	0.14	4.67	ppb
Silver	109-1	0.14	0.14	0.13	0.14	3.30	ppb
Sodium	23-2	537.84	539.08	535.11	537.34	0.38	ppb
Strontium	86-1	177.98	181.52	182.03	180.51	1.22	ppb
Strontium	88-1	185.00	186.55	189.20	186.92	1.13	ppb
Sulfur	34-1	-673.44	-648.84	-629.60	-650.63		ppb
Terbium	159-1				104		%
Terbium	159-2				106		%
Thallium	203-1	0.18	0.18	0.18	0.18	1.25	ppb
Thallium	205-1	0.18	0.18	0.18	0.18	2.27	ppb
Tin	118-1	0.16	0.16	0.17	0.16	2.79	ppb
Titanium	47-1	343.16	349.18	352.87	348.40	1.41	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-11

Instrumnet Name :

P7

Client Sample ID :

MYE4M2

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:29:46

DataFile Name :

076SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.65	1.72	1.68	1.68	2.12	ppb
Vanadium	51-2	69.10	69.76	69.73	69.53	0.54	ppb
Yttrium	89-1				98		%
Yttrium	89-2				104		%
Zinc	66-2	194.83	195.41	194.77	195.00	0.18	ppb
Zirconium	90-1	6.33	6.45	6.54	6.44	1.59	ppb
Zirconium	91-1	6.46	6.54	6.72	6.57	2.03	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-12 Instrumnet Name : P7
Client Sample ID : MYE4M4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:32:54 DataFile Name : 077SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	21154.71	21227.25	20972.37	21118.11	0.62	ppb
Antimony	121-1	0.29	0.30	0.31	0.30	3.06	ppb
Arsenic	75-2	47.52	47.66	48.55	47.91	1.16	ppb
Barium	135-1	332.83	338.39	343.19	338.14	1.53	ppb
Barium	137-1	358.24	364.94	364.25	362.48	1.02	ppb
Beryllium	9-1	0.73	0.78	0.79	0.77	4.03	ppb
Bismuth	209-1				103		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.53	1.27	1.26	1.36	11.45	ppb
Cadmium	106-1	0.00	0.04	0.41	0.15	147.25	ppb
Cadmium	111-1	0.47	0.44	0.52	0.48	7.72	ppb
Calcium	43-1	9539.41	9548.13	9653.60	9580.38	0.66	ppb
Calcium	44-1	8986.38	9085.95	9150.75	9074.36	0.91	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	52.77	52.90	52.91	52.86	0.14	ppb
Cobalt	59-2	30.83	30.98	30.65	30.82	0.54	ppb
Copper	63-2	112.76	112.92	113.29	112.99	0.24	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				107		%
Indium	115-1				97		%
Indium	115-2				101		%
Iron	56-2	46485.52	46058.19	45653.93	46065.88	0.90	ppb
Iron	57-2	45916.86	45870.41	45869.83	45885.70	0.06	ppb
Iron	54-2	46200.09	47046.00	46441.66	46562.59	0.94	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-12 Instrumnet Name : P7
Client Sample ID : MYE4M4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:32:54 DataFile Name : 077SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	21.75	21.50	21.80	21.68	0.74	ppb
Lead	207-1	21.00	20.81	21.26	21.02	1.09	ppb
Lead	208-1	21.08	21.20	21.40	21.22	0.76	ppb
Lithium	6-1				97		%
Magnesium	24-2	12691.90	12593.18	12665.45	12650.18	0.40	ppb
Manganese	55-2	2294.88	2308.25	2279.02	2294.05	0.64	ppb
Molybdenum	94-1	14.57	15.01	14.83	14.80	1.47	ppb
Molybdenum	95-1	0.73	0.76	0.75	0.75	2.00	ppb
Molybdenum	96-1	2.57	2.57	2.54	2.56	0.68	ppb
Molybdenum	97-1	0.70	0.71	0.69	0.70	1.61	ppb
Molybdenum	98-1	0.70	0.69	0.68	0.69	1.62	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	49.55	48.96	48.86	49.12	0.76	ppb
Phosphorus	31-2	1511.17	1538.09	1503.61	1517.62	1.19	ppb
Potassium	39-2	5366.56	5319.15	5316.16	5333.96	0.53	ppb
Rhodium	103-1				94		%
Rhodium	103-2				100		%
Scandium	45-1				94		%
Scandium	45-2				101		%
Selenium	82-1	0.04	-0.24	0.60	0.13	322.63	ppb
Selenium	77-2	12.98	11.90	9.62	11.50	14.91	ppb
Selenium	78-2	1.98	1.17	1.08	1.41	35.15	ppb
Silicon	28-1	826.66	839.47	837.39	834.51	0.82	ppb
Silver	107-1	0.24	0.27	0.26	0.26	5.95	ppb
Silver	109-1	0.25	0.26	0.26	0.26	1.96	ppb
Sodium	23-2	80.62	80.40	79.27	80.09	0.90	ppb
Strontium	86-1	48.44	49.76	49.79	49.33	1.57	ppb
Strontium	88-1	50.74	51.03	50.51	50.76	0.51	ppb
Sulfur	34-1	-1214.40	-1163.54	-1082.57	-1153.50		ppb
Terbium	159-1				103		%
Terbium	159-2				107		%
Thallium	203-1	0.15	0.14	0.14	0.15	4.31	ppb
Thallium	205-1	0.13	0.13	0.14	0.14	4.54	ppb
Tin	118-1	0.22	0.23	0.23	0.23	2.71	ppb
Titanium	47-1	363.05	368.81	370.29	367.39	1.04	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-12

Instrumnet Name :

P7

Client Sample ID :

MYE4M4

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:32:54

DataFile Name :

077SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.65	0.63	0.65	0.64	1.31	ppb
Vanadium	51-2	76.74	77.01	77.03	76.93	0.21	ppb
Yttrium	89-1				99		%
Yttrium	89-2				104		%
Zinc	66-2	142.79	142.26	142.75	142.60	0.21	ppb
Zirconium	90-1	8.42	8.44	8.51	8.45	0.53	ppb
Zirconium	91-1	8.35	8.50	8.61	8.49	1.57	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-13 Instrumnet Name : P7
Client Sample ID : MYE4M5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:36:00 DataFile Name : 078SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	20448.31	20204.39	20374.25	20342.31	0.61	ppb
Antimony	121-1	0.22	0.24	0.23	0.23	4.29	ppb
Arsenic	75-2	34.19	33.48	33.06	33.58	1.71	ppb
Barium	135-1	286.53	291.55	294.45	290.84	1.38	ppb
Barium	137-1	286.30	293.27	292.84	290.80	1.34	ppb
Beryllium	9-1	0.69	0.81	0.84	0.78	9.66	ppb
Bismuth	209-1				102		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.97	0.99	1.30	1.09	17.17	ppb
Cadmium	106-1	-0.18	-0.43	-1.00	-0.54		ppb
Cadmium	111-1	0.46	0.46	0.40	0.44	7.65	ppb
Calcium	43-1	12418.14	12694.39	12639.49	12584.00	1.16	ppb
Calcium	44-1	11833.61	11952.68	11850.61	11878.97	0.54	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	44.91	43.75	44.40	44.35	1.31	ppb
Cobalt	59-2	21.86	21.80	21.68	21.78	0.43	ppb
Copper	63-2	56.60	55.66	56.29	56.19	0.86	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				106		%
Indium	115-1				96		%
Indium	115-2				100		%
Iron	56-2	33744.44	33535.85	33691.22	33657.17	0.32	ppb
Iron	57-2	33358.05	33218.39	33580.06	33385.50	0.55	ppb
Iron	54-2	33735.05	33076.73	33742.90	33518.23	1.14	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-13 Instrumnet Name : P7
Client Sample ID : MYE4M5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:36:00 DataFile Name : 078SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	16.17	16.62	16.70	16.49	1.74	ppb
Lead	207-1	15.49	15.97	16.10	15.85	2.01	ppb
Lead	208-1	15.71	16.12	16.26	16.03	1.78	ppb
Lithium	6-1				96		%
Magnesium	24-2	12243.26	12188.63	12397.89	12276.59	0.88	ppb
Manganese	55-2	1405.60	1401.00	1428.54	1411.71	1.04	ppb
Molybdenum	94-1	9.55	9.79	10.05	9.80	2.53	ppb
Molybdenum	95-1	0.52	0.48	0.52	0.50	4.81	ppb
Molybdenum	96-1	1.68	1.77	1.65	1.70	3.63	ppb
Molybdenum	97-1	0.44	0.46	0.45	0.45	2.23	ppb
Molybdenum	98-1	0.42	0.44	0.44	0.43	3.08	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	37.76	36.90	37.67	37.44	1.25	ppb
Phosphorus	31-2	1392.52	1361.22	1370.39	1374.71	1.17	ppb
Potassium	39-2	4247.38	4185.00	4252.05	4228.14	0.89	ppb
Rhodium	103-1				93		%
Rhodium	103-2				100		%
Scandium	45-1				92		%
Scandium	45-2				100		%
Selenium	82-1	1.05	0.26	0.55	0.62	64.68	ppb
Selenium	77-2	9.45	9.75	9.57	9.59	1.61	ppb
Selenium	78-2	2.48	2.20	2.73	2.47	10.57	ppb
Silicon	28-1	1383.25	1410.93	1389.81	1394.66	1.04	ppb
Silver	107-1	0.31	0.34	0.34	0.33	5.43	ppb
Silver	109-1	0.32	0.34	0.34	0.33	3.56	ppb
Sodium	23-2	273.41	272.59	277.07	274.36	0.87	ppb
Strontium	86-1	136.17	139.64	139.67	138.50	1.45	ppb
Strontium	88-1	141.15	142.21	143.90	142.42	0.97	ppb
Sulfur	34-1	-770.31	-814.84	-867.60	-817.58		ppb
Terbium	159-1				102		%
Terbium	159-2				106		%
Thallium	203-1	0.69	0.71	0.75	0.72	4.03	ppb
Thallium	205-1	0.70	0.73	0.73	0.72	2.31	ppb
Tin	118-1	0.51	0.61	0.55	0.56	9.09	ppb
Titanium	47-1	367.73	377.15	373.34	372.74	1.27	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-13

Instrumnet Name :

P7

Client Sample ID :

MYE4M5

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:36:00

DataFile Name :

078SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.61	0.64	0.63	0.63	2.50	ppb
Vanadium	51-2	67.70	66.69	67.67	67.35	0.85	ppb
Yttrium	89-1				97		%
Yttrium	89-2				103		%
Zinc	66-2	189.53	189.99	189.16	189.56	0.22	ppb
Zirconium	90-1	5.58	5.70	5.68	5.65	1.10	ppb
Zirconium	91-1	5.65	5.78	5.76	5.73	1.25	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-14 Instrumnet Name : P7
Client Sample ID : MYE4M6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:39:09 DataFile Name : 079SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	30966.71	31113.79	30837.69	30972.73	0.45	ppb
Antimony	121-1	0.12	0.12	0.12	0.12	3.83	ppb
Arsenic	75-2	25.93	27.28	26.92	26.71	2.61	ppb
Barium	135-1	422.54	435.02	433.11	430.22	1.56	ppb
Barium	137-1	463.73	469.70	470.20	467.88	0.77	ppb
Beryllium	9-1	1.32	1.22	1.11	1.22	8.96	ppb
Bismuth	209-1				102		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.02	0.79	0.70	0.83	19.60	ppb
Cadmium	106-1	0.84	-0.81	-1.07	-0.35		ppb
Cadmium	111-1	0.25	0.19	0.14	0.20	27.72	ppb
Calcium	43-1	10148.12	10374.94	10373.37	10298.81	1.27	ppb
Calcium	44-1	9545.50	9733.27	9625.41	9634.72	0.98	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	51.01	51.60	51.39	51.33	0.58	ppb
Cobalt	59-2	29.99	30.01	29.69	29.90	0.61	ppb
Copper	63-2	59.26	60.49	59.16	59.64	1.25	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				107		%
Indium	115-1				97		%
Indium	115-2				100		%
Iron	56-2	45093.56	45895.32	45337.55	45442.15	0.90	ppb
Iron	57-2	45093.86	45731.79	44627.33	45150.99	1.23	ppb
Iron	54-2	45317.39	45668.52	45786.40	45590.77	0.54	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-14 Instrumnet Name : P7
Client Sample ID : MYE4M6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:39:09 DataFile Name : 079SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	19.23	19.44	19.77	19.48	1.41	ppb
Lead	207-1	18.31	18.49	18.96	18.58	1.81	ppb
Lead	208-1	18.63	18.76	19.23	18.87	1.67	ppb
Lithium	6-1				96		%
Magnesium	24-2	14252.49	14674.14	14333.08	14419.90	1.55	ppb
Manganese	55-2	1677.33	1676.15	1663.48	1672.32	0.46	ppb
Molybdenum	94-1	15.31	15.44	15.81	15.52	1.65	ppb
Molybdenum	95-1	0.28	0.25	0.25	0.26	5.57	ppb
Molybdenum	96-1	2.19	2.22	2.16	2.19	1.40	ppb
Molybdenum	97-1	0.23	0.24	0.21	0.23	6.26	ppb
Molybdenum	98-1	0.20	0.22	0.21	0.21	5.81	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	52.64	53.30	53.36	53.10	0.75	ppb
Phosphorus	31-2	1188.08	1241.20	1206.07	1211.78	2.23	ppb
Potassium	39-2	6100.27	6208.21	6140.74	6149.74	0.89	ppb
Rhodium	103-1				93		%
Rhodium	103-2				101		%
Scandium	45-1				94		%
Scandium	45-2				101		%
Selenium	82-1	0.61	0.91	-0.05	0.49	101.20	ppb
Selenium	77-2	12.02	17.23	14.36	14.54	17.93	ppb
Selenium	78-2	2.93	2.48	1.76	2.39	24.61	ppb
Silicon	28-1	1326.90	1365.31	1362.56	1351.59	1.59	ppb
Silver	107-1	0.07	0.08	0.09	0.08	6.89	ppb
Silver	109-1	0.08	0.09	0.09	0.08	6.59	ppb
Sodium	23-2	76.17	79.57	79.02	78.25	2.33	ppb
Strontium	86-1	69.56	70.70	70.86	70.38	1.01	ppb
Strontium	88-1	71.88	72.80	73.58	72.75	1.17	ppb
Sulfur	34-1	-1667.15	-1669.45	-1682.67	-1673.09		ppb
Terbium	159-1				102		%
Terbium	159-2				106		%
Thallium	203-1	0.18	0.19	0.18	0.18	1.44	ppb
Thallium	205-1	0.18	0.19	0.19	0.18	1.33	ppb
Tin	118-1	0.21	0.20	0.20	0.20	3.89	ppb
Titanium	47-1	522.02	534.73	535.06	530.61	1.40	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-14

Instrumnet Name :

P7

Client Sample ID :

MYE4M6

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:39:09

DataFile Name :

079SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.61	0.61	0.64	0.62	2.69	ppb
Vanadium	51-2	81.40	82.10	81.83	81.78	0.43	ppb
Yttrium	89-1				100		%
Yttrium	89-2				106		%
Zinc	66-2	117.92	118.68	118.83	118.48	0.41	ppb
Zirconium	90-1	9.02	9.22	9.27	9.17	1.43	ppb
Zirconium	91-1	9.14	9.13	9.25	9.17	0.74	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-15 Instrumnet Name : P7
Client Sample ID : MYE4M7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:42:15 DataFile Name : 080SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	23773.04	23970.89	23809.18	23851.03	0.44	ppb
Antimony	121-1	0.33	0.34	0.35	0.34	2.41	ppb
Arsenic	75-2	36.36	35.84	36.40	36.20	0.86	ppb
Barium	135-1	326.56	336.73	335.63	332.97	1.68	ppb
Barium	137-1	352.67	355.51	359.91	356.03	1.03	ppb
Beryllium	9-1	0.95	0.98	0.90	0.94	4.51	ppb
Bismuth	209-1				103		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.62	1.66	1.87	1.72	7.56	ppb
Cadmium	106-1	0.31	0.16	-0.62	-0.05		ppb
Cadmium	111-1	0.91	1.04	0.93	0.96	7.32	ppb
Calcium	43-1	9291.14	9401.38	9393.17	9361.90	0.66	ppb
Calcium	44-1	8770.60	8919.22	8854.04	8847.95	0.84	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	50.22	49.73	50.68	50.21	0.94	ppb
Cobalt	59-2	31.31	30.94	31.27	31.18	0.65	ppb
Copper	63-2	132.29	131.60	132.82	132.23	0.46	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				106		%
Indium	115-1				96		%
Indium	115-2				101		%
Iron	56-2	43264.19	42849.62	43297.24	43137.02	0.58	ppb
Iron	57-2	43031.01	43286.03	43136.17	43151.07	0.30	ppb
Iron	54-2	43412.33	43055.67	43086.22	43184.74	0.46	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-15 Instrumnet Name : P7
Client Sample ID : MYE4M7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:42:15 DataFile Name : 080SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	43.50	44.05	43.97	43.84	0.67	ppb
Lead	207-1	43.04	42.79	42.93	42.92	0.28	ppb
Lead	208-1	42.68	42.90	43.09	42.89	0.48	ppb
Lithium	6-1				95		%
Magnesium	24-2	12871.26	12811.39	12716.96	12799.87	0.61	ppb
Manganese	55-2	1720.11	1708.19	1725.84	1718.05	0.52	ppb
Molybdenum	94-1	15.85	15.96	15.80	15.87	0.52	ppb
Molybdenum	95-1	0.35	0.36	0.37	0.36	2.87	ppb
Molybdenum	96-1	2.28	2.24	2.32	2.28	1.77	ppb
Molybdenum	97-1	0.30	0.31	0.34	0.32	7.32	ppb
Molybdenum	98-1	0.32	0.31	0.31	0.31	2.57	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	54.14	53.28	54.29	53.91	1.01	ppb
Phosphorus	31-2	1700.23	1751.12	1659.50	1703.61	2.69	ppb
Potassium	39-2	6716.45	6622.90	6706.04	6681.79	0.77	ppb
Rhodium	103-1				92		%
Rhodium	103-2				100		%
Scandium	45-1				92		%
Scandium	45-2				100		%
Selenium	82-1	0.86	0.46	0.26	0.53	57.23	ppb
Selenium	77-2	11.80	11.34	12.25	11.80	3.84	ppb
Selenium	78-2	1.86	1.51	2.00	1.79	13.97	ppb
Silicon	28-1	1696.45	1705.94	1714.30	1705.56	0.52	ppb
Silver	107-1	0.25	0.28	0.27	0.27	6.00	ppb
Silver	109-1	0.28	0.28	0.28	0.28	1.82	ppb
Sodium	23-2	65.41	67.41	66.62	66.48	1.52	ppb
Strontium	86-1	51.57	52.25	52.58	52.14	0.99	ppb
Strontium	88-1	52.81	54.59	54.62	54.01	1.92	ppb
Sulfur	34-1	-1644.58	-1695.89	-1712.62	-1684.36		ppb
Terbium	159-1				102		%
Terbium	159-2				107		%
Thallium	203-1	0.26	0.25	0.26	0.26	1.52	ppb
Thallium	205-1	0.26	0.26	0.25	0.26	1.01	ppb
Tin	118-1	0.27	0.27	0.29	0.28	3.32	ppb
Titanium	47-1	482.69	488.93	491.20	487.61	0.90	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-15 Instrumnet Name : P7
Client Sample ID : MYE4M7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:42:15 DataFile Name : 080SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.62	0.61	0.65	0.63	3.43	ppb
Vanadium	51-2	81.64	80.75	81.39	81.26	0.57	ppb
Yttrium	89-1				98		%
Yttrium	89-2				104		%
Zinc	66-2	178.75	177.86	179.24	178.62	0.39	ppb
Zirconium	90-1	9.26	9.33	9.39	9.32	0.72	ppb
Zirconium	91-1	9.25	9.33	9.48	9.35	1.28	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV003 Instrumnet Name : P7
Client Sample ID : CCV003 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:45:23 DataFile Name : 081CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	48637.89	47488.14	48312.91	48146.31	1.23	ppb
Antimony	121-1	487.29	477.69	489.04	484.67	1.26	ppb
Arsenic	75-2	514.52	514.13	515.27	514.64	0.11	ppb
Barium	135-1	2512.69	2427.30	2487.33	2475.77	1.77	ppb
Barium	137-1	2500.95	2415.96	2464.33	2460.41	1.73	ppb
Beryllium	9-1	489.38	494.82	500.32	494.84	1.10	ppb
Bismuth	209-1				94		%
Bismuth	209-2				94		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	508.12	494.96	505.24	502.78	1.38	ppb
Cadmium	106-1	506.79	491.58	497.71	498.70	1.54	ppb
Cadmium	111-1	478.76	469.66	474.72	474.38	0.96	ppb
Calcium	43-1	243538.74	246315.72	249414.58	246423.01	1.19	ppb
Calcium	44-1	241029.24	243002.70	245533.50	243188.48	0.93	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	497.17	492.88	490.60	493.55	0.68	ppb
Cobalt	59-2	522.70	522.29	524.21	523.07	0.19	ppb
Copper	63-2	5092.30	4990.22	5015.68	5032.73	1.06	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				103		%
Indium	115-1				90		%
Indium	115-2				93		%
Iron	56-2	126127.87	124259.56	125031.57	125139.67	0.75	ppb
Iron	57-2	126408.43	122785.03	125585.83	124926.43	1.52	ppb
Iron	54-2	127440.64	125057.73	125336.85	125945.07	1.03	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV003 Instrumnet Name : P7
Client Sample ID : CCV003 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:45:23 DataFile Name : 081CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2530.53	2552.03	2551.84	2544.80	0.49	ppb
Lead	207-1	2514.91	2522.32	2549.13	2528.78	0.71	ppb
Lead	208-1	2503.38	2529.12	2544.56	2525.69	0.82	ppb
Lithium	6-1				94		%
Magnesium	24-2	250689.17	246264.25	248394.99	248449.47	0.89	ppb
Manganese	55-2	4956.06	4929.31	4896.35	4927.24	0.61	ppb
Molybdenum	94-1	4973.72	4962.38	5053.04	4996.38	0.99	ppb
Molybdenum	95-1	4950.46	4922.20	5053.05	4975.24	1.38	ppb
Molybdenum	96-1	4913.69	4926.11	5060.16	4966.65	1.64	ppb
Molybdenum	97-1	4934.62	4945.10	5106.63	4995.45	1.93	ppb
Molybdenum	98-1	4927.40	4898.54	5056.12	4960.69	1.69	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	492.40	482.56	484.83	486.60	1.06	ppb
Phosphorus	31-2	9907.23	9669.56	9857.80	9811.53	1.28	ppb
Potassium	39-2	122091.54	119876.04	120259.22	120742.27	0.98	ppb
Rhodium	103-1				85		%
Rhodium	103-2				90		%
Scandium	45-1				88		%
Scandium	45-2				95		%
Selenium	82-1	504.46	503.33	514.48	507.42	1.21	ppb
Selenium	77-2	521.95	499.52	489.49	503.65	3.30	ppb
Selenium	78-2	510.11	506.79	497.75	504.89	1.27	ppb
Silicon	28-1	2542.59	2572.31	2586.69	2567.19	0.88	ppb
Silver	107-1	508.88	491.05	498.46	499.46	1.79	ppb
Silver	109-1	500.79	494.72	489.20	494.90	1.17	ppb
Sodium	23-2	243900.24	235985.67	240700.15	240195.35	1.66	ppb
Strontium	86-1	12319.20	12223.16	12542.60	12361.65	1.33	ppb
Strontium	88-1	12252.77	12198.72	12536.48	12329.32	1.47	ppb
Sulfur	34-1	9397.13	9383.13	9423.28	9401.18	0.22	ppb
Terbium	159-1				100		%
Terbium	159-2				102		%
Thallium	203-1	508.52	509.69	511.68	509.96	0.31	ppb
Thallium	205-1	509.47	511.07	510.91	510.48	0.17	ppb
Tin	118-1	481.22	473.03	473.23	475.82	0.98	ppb
Titanium	47-1	4881.41	4934.60	4956.20	4924.07	0.78	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

CCV003

Instrumnet Name :

P7

Client Sample ID :

CCV003

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:45:23

DataFile Name :

081CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	517.37	513.71	508.91	513.33	0.83	ppb
Vanadium	51-2	509.85	498.44	498.94	502.41	1.28	ppb
Yttrium	89-1				92		%
Yttrium	89-2				96		%
Zinc	66-2	5042.78	4981.54	4994.72	5006.35	0.64	ppb
Zirconium	90-1	505.94	501.72	514.48	507.38	1.28	ppb
Zirconium	91-1	501.00	503.43	518.03	507.48	1.82	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB003 Instrumnet Name : P7
Client Sample ID : CCB003 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:48:04 DataFile Name : 082CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	2.10	2.65	2.06	2.27	14.56	ppb
Antimony	121-1	0.32	0.32	0.31	0.32	1.32	ppb
Arsenic	75-2	0.01	0.00	0.01	0.01	101.42	ppb
Barium	135-1	0.11	0.09	0.09	0.10	8.31	ppb
Barium	137-1	0.12	0.10	0.10	0.11	14.32	ppb
Beryllium	9-1	0.01	0.00	0.00	0.00	173.97	ppb
Bismuth	209-1				103		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.05	0.08	-0.01	0.04	108.13	ppb
Cadmium	106-1	-1.42	-1.42	-1.23	-1.36		ppb
Cadmium	111-1	-0.07	-0.09	-0.07	-0.07		ppb
Calcium	43-1	4.35	3.80	4.89	4.35	12.50	ppb
Calcium	44-1	2.73	1.06	0.35	1.38	88.33	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.01	0.03	0.01	0.02	72.76	ppb
Cobalt	59-2	0.02	0.01	0.01	0.01	18.60	ppb
Copper	63-2	0.03	0.06	0.00	0.03	86.38	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				106		%
Indium	115-1				97		%
Indium	115-2				102		%
Iron	56-2	7.17	7.28	7.13	7.19	1.15	ppb
Iron	57-2	5.58	6.53	5.50	5.87	9.83	ppb
Iron	54-2	6.35	7.46	7.63	7.15	9.72	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB003 Instrumnet Name : P7
Client Sample ID : CCB003 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:48:04 DataFile Name : 082CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.05	0.05	0.04	0.04	10.93	ppb
Lead	207-1	0.07	0.05	0.03	0.05	39.11	ppb
Lead	208-1	0.06	0.04	0.04	0.05	23.73	ppb
Lithium	6-1				96		%
Magnesium	24-2	3.24	2.59	2.36	2.73	16.66	ppb
Manganese	55-2	0.23	0.19	0.23	0.21	11.72	ppb
Molybdenum	94-1	0.29	0.23	0.22	0.25	15.69	ppb
Molybdenum	95-1	0.20	0.21	0.16	0.19	14.94	ppb
Molybdenum	96-1	0.20	0.19	0.18	0.19	3.52	ppb
Molybdenum	97-1	0.21	0.19	0.14	0.18	20.63	ppb
Molybdenum	98-1	0.20	0.16	0.16	0.17	10.75	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.05	0.02	0.05	0.04	50.16	ppb
Phosphorus	31-2	-19.00	-23.48	-20.70	-21.06		ppb
Potassium	39-2	-2.98	-3.44	-3.49	-3.30		ppb
Rhodium	103-1				96		%
Rhodium	103-2				103		%
Scandium	45-1				92		%
Scandium	45-2				98		%
Selenium	82-1	-0.74	-0.49	-0.06	-0.43		ppb
Selenium	77-2	0.00	-0.22	-0.22	-0.15		ppb
Selenium	78-2	0.58	-0.47	-0.71	-0.20		ppb
Silicon	28-1	0.93	0.54	0.69	0.72	27.10	ppb
Silver	107-1	0.05	0.03	0.04	0.04	14.10	ppb
Silver	109-1	0.04	0.04	0.03	0.04	19.00	ppb
Sodium	23-2	1.31	-0.67	0.17	0.27	374.37	ppb
Strontium	86-1	0.24	0.21	0.14	0.20	25.14	ppb
Strontium	88-1	0.22	0.16	0.13	0.17	26.77	ppb
Sulfur	34-1	-657.77	-663.46	-636.50	-652.58		ppb
Terbium	159-1				101		%
Terbium	159-2				105		%
Thallium	203-1	0.05	0.04	0.04	0.04	18.86	ppb
Thallium	205-1	0.05	0.04	0.03	0.04	19.93	ppb
Tin	118-1	0.06	0.10	0.11	0.09	27.54	ppb
Titanium	47-1	0.18	0.16	0.15	0.16	10.99	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

CCB003

Instrumnet Name :

P7

Client Sample ID :

CCB003

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:48:04

DataFile Name :

082CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	16.18	ppb
Vanadium	51-2	0.02	0.02	0.03	0.02	20.22	ppb
Yttrium	89-1				95		%
Yttrium	89-2				101		%
Zinc	66-2	0.00	0.00	0.01	0.00	208.13	ppb
Zirconium	90-1	0.06	0.05	0.05	0.05	10.84	ppb
Zirconium	91-1	0.06	0.08	0.07	0.07	11.00	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-16 Instrumnet Name : P7
Client Sample ID : MYE4M8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:51:21 DataFile Name : 083SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	30899.93	30776.57	30716.65	30797.71	0.30	ppb
Antimony	121-1	0.32	0.38	0.35	0.35	7.59	ppb
Arsenic	75-2	117.78	120.37	118.16	118.77	1.18	ppb
Barium	135-1	999.32	1026.49	1016.33	1014.05	1.35	ppb
Barium	137-1	996.72	1023.72	1015.68	1012.04	1.37	ppb
Beryllium	9-1	0.96	1.12	1.04	1.04	7.85	ppb
Bismuth	209-1				103		%
Bismuth	209-2				107		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.82	1.28	1.63	1.57	17.21	ppb
Cadmium	106-1	2.50	1.50	0.95	1.65	47.60	ppb
Cadmium	111-1	0.82	0.74	0.71	0.76	7.03	ppb
Calcium	43-1	16494.45	16339.40	16396.68	16410.18	0.48	ppb
Calcium	44-1	15439.72	15410.18	15689.51	15513.14	0.99	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	71.00	72.16	72.34	71.83	1.02	ppb
Cobalt	59-2	40.09	40.19	40.50	40.26	0.53	ppb
Copper	63-2	97.86	98.07	99.13	98.35	0.70	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				106		%
Holmium	165-2				110		%
Indium	115-1				102		%
Indium	115-2				107		%
Iron	56-2	60199.98	60327.51	60237.97	60255.16	0.11	ppb
Iron	57-2	59130.09	59869.06	59693.17	59564.11	0.65	ppb
Iron	54-2	59771.30	60779.12	60211.83	60254.08	0.84	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-16 Instrumnet Name : P7
Client Sample ID : MYE4M8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:51:21 DataFile Name : 083SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	20.15	20.14	20.09	20.13	0.16	ppb
Lead	207-1	19.19	19.26	19.18	19.21	0.23	ppb
Lead	208-1	19.38	19.59	19.55	19.51	0.57	ppb
Lithium	6-1				99		%
Magnesium	24-2	17470.49	17409.68	17408.75	17429.64	0.20	ppb
Manganese	55-2	7176.18	7200.24	7267.97	7214.80	0.66	ppb
Molybdenum	94-1	20.68	20.47	20.94	20.70	1.15	ppb
Molybdenum	95-1	0.65	0.64	0.63	0.64	1.58	ppb
Molybdenum	96-1	3.02	3.01	3.07	3.03	1.06	ppb
Molybdenum	97-1	0.51	0.48	0.49	0.50	2.81	ppb
Molybdenum	98-1	0.48	0.46	0.48	0.48	2.31	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	93.94	94.76	94.65	94.45	0.47	ppb
Phosphorus	31-2	1871.59	1950.44	1913.43	1911.82	2.06	ppb
Potassium	39-2	3502.59	3470.79	3510.50	3494.63	0.60	ppb
Rhodium	103-1				98		%
Rhodium	103-2				106		%
Scandium	45-1				103		%
Scandium	45-2				111		%
Selenium	82-1	0.01	0.00	0.73	0.25	167.65	ppb
Selenium	77-2	16.09	14.88	15.78	15.59	4.05	ppb
Selenium	78-2	3.35	4.13	2.69	3.39	21.26	ppb
Silicon	28-1	1578.89	1554.91	1557.85	1563.88	0.84	ppb
Silver	107-1	0.09	0.10	0.10	0.10	3.36	ppb
Silver	109-1	0.10	0.11	0.10	0.10	6.29	ppb
Sodium	23-2	114.29	114.76	113.99	114.35	0.34	ppb
Strontium	86-1	97.07	96.82	99.34	97.74	1.42	ppb
Strontium	88-1	100.65	102.40	102.30	101.78	0.96	ppb
Sulfur	34-1	-1047.51	-1195.46	-1385.08	-1209.35		ppb
Terbium	159-1				105		%
Terbium	159-2				110		%
Thallium	203-1	0.26	0.31	0.25	0.27	11.08	ppb
Thallium	205-1	0.26	0.28	0.26	0.27	5.28	ppb
Tin	118-1	0.22	0.22	0.25	0.23	6.93	ppb
Titanium	47-1	483.55	479.83	486.50	483.29	0.69	ppb

LB Number :LB133265Operator :Jaswal

Lab Sample ID :P4520-16Instrumnet Name :P7

Client Sample ID :MYE4M8Dilution Factor :1

Date & Time Acquired :2024-11-02 21:51:21DataFile Name :083SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.61	0.63	0.63	0.62	2.21	ppb
Vanadium	51-2	123.04	123.97	124.08	123.70	0.46	ppb
Yttrium	89-1				106		%
Yttrium	89-2				111		%
Zinc	66-2	133.65	133.40	133.58	133.54	0.10	ppb
Zirconium	90-1	11.97	12.05	12.19	12.07	0.90	ppb
Zirconium	91-1	12.06	12.15	12.31	12.17	1.04	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-17 Instrumnet Name : P7
Client Sample ID : MYE4M9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:54:27 DataFile Name : 084SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	22452.17	22205.46	22226.22	22294.62	0.61	ppb
Antimony	121-1	0.24	0.24	0.22	0.24	4.63	ppb
Arsenic	75-2	32.00	33.09	32.18	32.42	1.79	ppb
Barium	135-1	236.59	238.68	239.81	238.36	0.69	ppb
Barium	137-1	236.16	239.91	240.11	238.73	0.93	ppb
Beryllium	9-1	0.86	0.88	0.92	0.89	3.42	ppb
Bismuth	209-1				105		%
Bismuth	209-2				109		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.89	1.44	1.00	1.11	25.93	ppb
Cadmium	106-1	0.63	-0.69	0.05	0.00		ppb
Cadmium	111-1	0.45	0.36	0.39	0.40	11.63	ppb
Calcium	43-1	16369.54	16486.53	16575.05	16477.04	0.63	ppb
Calcium	44-1	15463.83	15546.11	15746.86	15585.60	0.93	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	46.62	46.68	46.27	46.52	0.47	ppb
Cobalt	59-2	20.53	20.36	20.69	20.53	0.79	ppb
Copper	63-2	57.67	57.80	57.81	57.76	0.14	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				105		%
Holmium	165-2				111		%
Indium	115-1				103		%
Indium	115-2				106		%
Iron	56-2	35474.00	35123.16	35758.86	35452.01	0.90	ppb
Iron	57-2	35261.75	34967.48	35366.17	35198.47	0.59	ppb
Iron	54-2	35996.70	35492.40	35454.30	35647.80	0.85	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-17 Instrumnet Name : P7
Client Sample ID : MYE4M9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:54:27 DataFile Name : 084SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	15.81	16.01	16.52	16.11	2.28	ppb
Lead	207-1	15.23	15.52	15.86	15.54	2.04	ppb
Lead	208-1	15.46	15.64	16.12	15.74	2.18	ppb
Lithium	6-1				99		%
Magnesium	24-2	12533.33	12410.84	12551.39	12498.52	0.61	ppb
Manganese	55-2	1140.64	1151.29	1145.22	1145.72	0.47	ppb
Molybdenum	94-1	12.47	12.81	12.77	12.68	1.46	ppb
Molybdenum	95-1	0.44	0.52	0.48	0.48	8.20	ppb
Molybdenum	96-1	2.07	2.12	2.00	2.06	2.75	ppb
Molybdenum	97-1	0.44	0.50	0.46	0.47	5.76	ppb
Molybdenum	98-1	0.44	0.44	0.43	0.44	0.94	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	40.62	40.78	41.25	40.88	0.81	ppb
Phosphorus	31-2	1241.88	1211.18	1228.91	1227.32	1.26	ppb
Potassium	39-2	3676.10	3662.64	3661.34	3666.70	0.22	ppb
Rhodium	103-1				99		%
Rhodium	103-2				107		%
Scandium	45-1				102		%
Scandium	45-2				109		%
Selenium	82-1	0.12	0.17	0.58	0.29	85.74	ppb
Selenium	77-2	8.84	7.82	10.81	9.16	16.57	ppb
Selenium	78-2	1.47	2.47	1.13	1.69	41.39	ppb
Silicon	28-1	1696.61	1704.51	1725.13	1708.75	0.86	ppb
Silver	107-1	0.14	0.15	0.15	0.15	3.23	ppb
Silver	109-1	0.14	0.16	0.15	0.15	8.07	ppb
Sodium	23-2	338.39	337.99	337.71	338.03	0.10	ppb
Strontium	86-1	108.41	108.70	110.31	109.14	0.94	ppb
Strontium	88-1	111.14	111.46	113.03	111.88	0.90	ppb
Sulfur	34-1	-1160.66	-1275.86	-1287.04	-1241.19		ppb
Terbium	159-1				107		%
Terbium	159-2				111		%
Thallium	203-1	0.12	0.14	0.12	0.13	8.08	ppb
Thallium	205-1	0.13	0.12	0.12	0.12	3.72	ppb
Tin	118-1	0.34	0.35	0.34	0.34	0.92	ppb
Titanium	47-1	326.75	328.25	330.10	328.37	0.51	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-17

Instrumnet Name :

P7

Client Sample ID :

MYE4M9

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:54:27

DataFile Name :

084SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.73	0.75	0.78	0.76	3.28	ppb
Vanadium	51-2	78.44	78.20	77.66	78.10	0.51	ppb
Yttrium	89-1				105		%
Yttrium	89-2				110		%
Zinc	66-2	151.78	149.99	149.71	150.49	0.75	ppb
Zirconium	90-1	7.29	7.39	7.37	7.35	0.70	ppb
Zirconium	91-1	7.30	7.25	7.49	7.35	1.73	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-18 Instrumnet Name : P7
Client Sample ID : MYE4N0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:57:36 DataFile Name : 085SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	21169.41	21234.73	21242.73	21215.63	0.19	ppb
Antimony	121-1	0.57	0.62	0.66	0.62	7.19	ppb
Arsenic	75-2	40.44	40.67	40.60	40.57	0.29	ppb
Barium	135-1	343.91	349.20	403.35	365.49	9.00	ppb
Barium	137-1	368.73	382.85	439.88	397.15	9.48	ppb
Beryllium	9-1	1.00	0.85	1.01	0.95	9.50	ppb
Bismuth	209-1				101		%
Bismuth	209-2				108		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.61	1.94	1.87	1.81	9.72	ppb
Cadmium	106-1	1.64	2.11	1.85	1.87	12.60	ppb
Cadmium	111-1	0.68	0.76	0.83	0.76	9.64	ppb
Calcium	43-1	39078.84	38929.75	45083.75	41030.78	8.56	ppb
Calcium	44-1	36909.16	36702.14	42647.60	38752.97	8.71	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	43.16	42.63	43.74	43.18	1.28	ppb
Cobalt	59-2	23.74	23.54	23.63	23.64	0.43	ppb
Copper	63-2	92.50	91.05	92.13	91.89	0.82	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				112		%
Indium	115-1				98		%
Indium	115-2				107		%
Iron	56-2	36595.84	36318.47	36572.04	36495.45	0.42	ppb
Iron	57-2	36482.51	35932.73	36880.90	36432.05	1.31	ppb
Iron	54-2	37215.84	36501.02	37112.98	36943.28	1.05	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-18 Instrumnet Name : P7
Client Sample ID : MYE4N0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:57:36 DataFile Name : 085SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	25.35	25.54	29.29	26.73	8.31	ppb
Lead	207-1	24.51	24.48	28.59	25.86	9.16	ppb
Lead	208-1	24.60	24.87	28.52	26.00	8.42	ppb
Lithium	6-1				96		%
Magnesium	24-2	13431.98	13302.18	13501.30	13411.82	0.75	ppb
Manganese	55-2	1614.89	1597.32	1627.18	1613.13	0.93	ppb
Molybdenum	94-1	21.40	21.43	25.01	22.61	9.18	ppb
Molybdenum	95-1	0.55	0.64	0.70	0.63	11.98	ppb
Molybdenum	96-1	3.01	3.06	3.56	3.21	9.56	ppb
Molybdenum	97-1	0.61	0.63	0.62	0.62	1.53	ppb
Molybdenum	98-1	0.53	0.57	0.65	0.58	10.94	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	45.93	45.94	45.83	45.90	0.13	ppb
Phosphorus	31-2	2045.43	2002.67	2029.58	2025.90	1.07	ppb
Potassium	39-2	5855.96	5804.61	5923.91	5861.49	1.02	ppb
Rhodium	103-1				95		%
Rhodium	103-2				105		%
Scandium	45-1				98		%
Scandium	45-2				108		%
Selenium	82-1	0.16	0.53	0.11	0.27	85.52	ppb
Selenium	77-2	10.29	8.58	6.35	8.41	23.48	ppb
Selenium	78-2	2.22	2.07	2.16	2.15	3.51	ppb
Silicon	28-1	2196.68	2180.17	2529.11	2301.99	8.55	ppb
Silver	107-1	0.15	0.16	0.19	0.17	13.40	ppb
Silver	109-1	0.16	0.17	0.19	0.17	9.21	ppb
Sodium	23-2	192.17	190.31	192.67	191.72	0.65	ppb
Strontium	86-1	225.86	229.29	261.17	238.77	8.16	ppb
Strontium	88-1	230.84	234.93	270.21	245.33	8.82	ppb
Sulfur	34-1	-145.52	-132.83	681.69	134.44	352.54	ppb
Terbium	159-1				102		%
Terbium	159-2				111		%
Thallium	203-1	0.32	0.32	0.39	0.34	12.20	ppb
Thallium	205-1	0.31	0.32	0.37	0.34	9.67	ppb
Tin	118-1	0.74	0.81	0.98	0.84	14.73	ppb
Titanium	47-1	433.57	433.15	500.20	455.64	8.47	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-18

Instrumnet Name :

P7

Client Sample ID :

MYE4N0

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 21:57:36

DataFile Name :

085SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.52	0.54	0.60	0.56	7.82	ppb
Vanadium	51-2	70.26	69.20	70.65	70.04	1.07	ppb
Yttrium	89-1				102		%
Yttrium	89-2				110		%
Zinc	66-2	185.19	184.34	187.34	185.62	0.83	ppb
Zirconium	90-1	12.48	12.52	14.69	13.23	9.55	ppb
Zirconium	91-1	12.55	12.69	14.52	13.25	8.30	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-19 Instrumnet Name : P7
Client Sample ID : MYE4N1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:00:42 DataFile Name : 086SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	21989.91	21459.06	21316.62	21588.53	1.64	ppb
Antimony	121-1	1.45	1.48	1.51	1.48	2.00	ppb
Arsenic	75-2	39.09	38.15	37.34	38.19	2.30	ppb
Barium	135-1	320.82	330.13	330.36	327.10	1.66	ppb
Barium	137-1	349.11	356.72	353.27	353.03	1.08	ppb
Beryllium	9-1	1.08	1.02	1.02	1.04	3.66	ppb
Bismuth	209-1				104		%
Bismuth	209-2				107		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.33	2.62	2.45	2.47	5.89	ppb
Cadmium	106-1	1.22	1.42	0.95	1.20	19.44	ppb
Cadmium	111-1	1.73	1.82	1.85	1.80	3.56	ppb
Calcium	43-1	8642.41	8601.97	8745.20	8663.19	0.85	ppb
Calcium	44-1	8261.91	8142.60	8150.27	8184.93	0.82	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	50.69	49.44	49.86	49.99	1.28	ppb
Cobalt	59-2	27.98	28.06	28.23	28.09	0.44	ppb
Copper	63-2	257.47	257.37	257.16	257.33	0.06	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				105		%
Holmium	165-2				109		%
Indium	115-1				101		%
Indium	115-2				103		%
Iron	56-2	41285.57	41457.91	41279.02	41340.83	0.25	ppb
Iron	57-2	41146.67	39689.34	40668.30	40501.43	1.83	ppb
Iron	54-2	41379.69	41308.09	41592.33	41426.70	0.36	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-19 Instrumnet Name : P7
Client Sample ID : MYE4N1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:00:42 DataFile Name : 086SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	80.85	83.46	82.10	82.14	1.59	ppb
Lead	207-1	80.70	81.58	82.66	81.65	1.20	ppb
Lead	208-1	80.16	81.94	81.52	81.21	1.15	ppb
Lithium	6-1				99		%
Magnesium	24-2	11858.23	11789.03	11838.87	11828.71	0.30	ppb
Manganese	55-2	1618.99	1627.08	1609.49	1618.52	0.54	ppb
Molybdenum	94-1	13.61	13.75	13.99	13.79	1.40	ppb
Molybdenum	95-1	0.36	0.38	0.38	0.37	3.31	ppb
Molybdenum	96-1	2.04	2.04	2.08	2.05	1.18	ppb
Molybdenum	97-1	0.35	0.32	0.38	0.35	9.29	ppb
Molybdenum	98-1	0.30	0.32	0.33	0.31	4.44	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	48.27	47.70	47.15	47.71	1.18	ppb
Phosphorus	31-2	1551.96	1534.92	1536.77	1541.22	0.61	ppb
Potassium	39-2	5932.02	5850.94	5880.88	5887.95	0.70	ppb
Rhodium	103-1				97		%
Rhodium	103-2				103		%
Scandium	45-1				100		%
Scandium	45-2				106		%
Selenium	82-1	0.05	0.72	0.15	0.31	117.27	ppb
Selenium	77-2	11.72	10.33	12.43	11.49	9.28	ppb
Selenium	78-2	3.20	2.02	1.72	2.31	33.96	ppb
Silicon	28-1	2280.67	2285.10	2300.97	2288.91	0.47	ppb
Silver	107-1	0.64	0.66	0.68	0.66	2.55	ppb
Silver	109-1	0.63	0.67	0.67	0.66	3.78	ppb
Sodium	23-2	126.31	122.52	122.47	123.77	1.78	ppb
Strontium	86-1	54.65	55.20	55.35	55.07	0.67	ppb
Strontium	88-1	54.84	56.88	57.59	56.44	2.53	ppb
Sulfur	34-1	-1475.62	-1654.06	-1694.87	-1608.18		ppb
Terbium	159-1				105		%
Terbium	159-2				109		%
Thallium	203-1	0.22	0.22	0.24	0.23	4.82	ppb
Thallium	205-1	0.23	0.22	0.24	0.23	3.91	ppb
Tin	118-1	0.48	0.48	0.51	0.49	3.29	ppb
Titanium	47-1	513.01	511.50	515.18	513.23	0.36	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-19

Instrumnet Name :

P7

Client Sample ID :

MYE4N1

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 22:00:42

DataFile Name :

086SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.65	0.66	0.65	0.65	0.65	ppb
Vanadium	51-2	76.09	75.48	75.89	75.82	0.41	ppb
Yttrium	89-1				103		%
Yttrium	89-2				108		%
Zinc	66-2	257.21	252.13	253.38	254.24	1.04	ppb
Zirconium	90-1	7.95	8.07	8.14	8.06	1.17	ppb
Zirconium	91-1	7.90	8.00	8.14	8.02	1.48	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-20 Instrumnet Name : P7
Client Sample ID : MYE4N2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:03:50 DataFile Name : 087SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	44668.35	45342.39	45630.10	45213.61	1.09	ppb
Antimony	121-1	0.21	0.21	0.21	0.21	2.13	ppb
Arsenic	75-2	33.89	33.81	33.13	33.61	1.24	ppb
Barium	135-1	274.52	276.34	279.60	276.82	0.93	ppb
Barium	137-1	275.41	277.60	281.33	278.12	1.08	ppb
Beryllium	9-1	0.86	0.99	0.89	0.92	7.59	ppb
Bismuth	209-1				102		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.33	2.53	2.99	2.62	12.90	ppb
Cadmium	106-1	1.20	0.24	0.82	0.76	63.99	ppb
Cadmium	111-1	2.16	2.19	2.21	2.19	1.04	ppb
Calcium	43-1	18693.82	18995.02	18866.47	18851.77	0.80	ppb
Calcium	44-1	17675.39	17814.43	17954.53	17814.79	0.78	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	48.92	48.19	48.22	48.44	0.86	ppb
Cobalt	59-2	76.44	77.14	77.02	76.87	0.49	ppb
Copper	63-2	135.51	136.49	135.28	135.76	0.47	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				109		%
Indium	115-1				100		%
Indium	115-2				103		%
Iron	56-2	93869.63	93788.54	94520.76	94059.64	0.43	ppb
Iron	57-2	93517.28	93639.30	93830.99	93662.53	0.17	ppb
Iron	54-2	94376.14	94620.17	95155.38	94717.23	0.42	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-20 Instrumnet Name : P7
Client Sample ID : MYE4N2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:03:50 DataFile Name : 087SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	26.91	27.81	27.58	27.43	1.69	ppb
Lead	207-1	26.05	27.07	26.77	26.63	1.97	ppb
Lead	208-1	26.27	27.06	26.94	26.76	1.60	ppb
Lithium	6-1				99		%
Magnesium	24-2	24663.60	24743.64	24526.54	24644.59	0.45	ppb
Manganese	55-2	3474.99	3462.10	3476.75	3471.28	0.23	ppb
Molybdenum	94-1	3.94	3.89	4.04	3.96	1.90	ppb
Molybdenum	95-1	0.82	0.79	0.74	0.78	4.95	ppb
Molybdenum	96-1	1.79	1.82	1.83	1.81	1.09	ppb
Molybdenum	97-1	0.75	0.77	0.78	0.77	1.92	ppb
Molybdenum	98-1	0.67	0.70	0.70	0.69	2.76	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	44.65	44.32	44.45	44.47	0.37	ppb
Phosphorus	31-2	1169.93	1167.16	1166.81	1167.97	0.15	ppb
Potassium	39-2	2573.95	2604.41	2578.50	2585.62	0.64	ppb
Rhodium	103-1				96		%
Rhodium	103-2				102		%
Scandium	45-1				101		%
Scandium	45-2				108		%
Selenium	82-1	0.18	-0.03	0.20	0.12	106.98	ppb
Selenium	77-2	7.17	7.81	9.58	8.19	15.29	ppb
Selenium	78-2	0.58	1.88	-0.01	0.81	118.82	ppb
Silicon	28-1	2364.65	2421.52	2378.78	2388.31	1.24	ppb
Silver	107-1	0.72	0.73	0.71	0.72	1.04	ppb
Silver	109-1	0.71	0.72	0.74	0.72	2.19	ppb
Sodium	23-2	628.12	631.01	629.62	629.58	0.23	ppb
Strontium	86-1	105.55	106.69	107.86	106.70	1.08	ppb
Strontium	88-1	108.66	109.35	109.98	109.33	0.60	ppb
Sulfur	34-1	-2071.72	-2079.40	-2137.72	-2096.28		ppb
Terbium	159-1				105		%
Terbium	159-2				107		%
Thallium	203-1	0.22	0.23	0.23	0.23	1.98	ppb
Thallium	205-1	0.22	0.24	0.23	0.23	3.76	ppb
Tin	118-1	0.20	0.20	0.22	0.21	5.52	ppb
Titanium	47-1	1003.79	1038.77	1002.26	1014.94	2.03	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-20

Instrumnet Name :

P7

Client Sample ID :

MYE4N2

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 22:03:50

DataFile Name :

087SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.41	0.42	0.41	0.41	1.44	ppb
Vanadium	51-2	166.29	165.51	165.23	165.68	0.33	ppb
Yttrium	89-1				104		%
Yttrium	89-2				110		%
Zinc	66-2	441.58	445.63	444.38	443.87	0.47	ppb
Zirconium	90-1	2.00	1.97	2.02	1.99	1.35	ppb
Zirconium	91-1	1.93	1.96	2.01	1.97	1.86	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-21 Instrumnet Name : P7
Client Sample ID : MYE4N3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:06:56 DataFile Name : 088SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	23438.62	25922.06	27705.67	25688.78	8.34	ppb
Antimony	121-1	0.33	0.32	0.34	0.33	3.56	ppb
Arsenic	75-2	46.30	51.78	52.51	50.19	6.76	ppb
Barium	135-1	348.99	353.96	358.78	353.91	1.38	ppb
Barium	137-1	385.45	383.31	389.52	386.09	0.82	ppb
Beryllium	9-1	1.10	0.86	1.14	1.03	14.58	ppb
Bismuth	209-1				103		%
Bismuth	209-2				98		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.91	1.80	1.92	1.88	3.58	ppb
Cadmium	106-1	1.53	-0.31	0.80	0.67	137.17	ppb
Cadmium	111-1	0.97	0.82	0.91	0.90	8.06	ppb
Calcium	43-1	17510.48	17676.22	17612.86	17599.85	0.48	ppb
Calcium	44-1	16405.70	16803.02	16605.02	16604.58	1.20	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	50.62	55.98	58.34	54.98	7.20	ppb
Cobalt	59-2	30.40	33.70	35.03	33.05	7.22	ppb
Copper	63-2	138.02	151.30	159.76	149.69	7.32	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				100		%
Indium	115-1				100		%
Indium	115-2				94		%
Iron	56-2	44806.65	49189.06	51256.01	48417.24	6.80	ppb
Iron	57-2	44199.16	48431.67	50393.11	47674.65	6.64	ppb
Iron	54-2	44947.42	49279.01	51240.17	48488.87	6.64	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-21 Instrumnet Name : P7
Client Sample ID : MYE4N3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:06:56 DataFile Name : 088SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	43.00	43.75	43.47	43.41	0.88	ppb
Lead	207-1	42.28	42.98	42.72	42.66	0.82	ppb
Lead	208-1	42.44	43.12	42.95	42.84	0.83	ppb
Lithium	6-1				98		%
Magnesium	24-2	13253.01	14531.06	15323.63	14369.24	7.27	ppb
Manganese	55-2	1821.46	2004.93	2073.07	1966.49	6.62	ppb
Molybdenum	94-1	16.48	16.81	17.15	16.81	1.99	ppb
Molybdenum	95-1	0.45	0.49	0.49	0.48	5.09	ppb
Molybdenum	96-1	2.54	2.55	2.47	2.52	1.81	ppb
Molybdenum	97-1	0.40	0.45	0.51	0.45	12.60	ppb
Molybdenum	98-1	0.39	0.41	0.44	0.41	5.42	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	56.79	62.74	65.22	61.59	7.03	ppb
Phosphorus	31-2	1728.79	1973.07	2059.58	1920.48	8.93	ppb
Potassium	39-2	6339.77	7025.65	7326.77	6897.39	7.33	ppb
Rhodium	103-1				96		%
Rhodium	103-2				94		%
Scandium	45-1				97		%
Scandium	45-2				97		%
Selenium	82-1	0.29	-0.08	-0.24	-0.01		ppb
Selenium	77-2	13.81	13.95	16.63	14.80	10.74	ppb
Selenium	78-2	2.03	2.88	2.20	2.37	18.98	ppb
Silicon	28-1	1905.14	1931.85	1948.71	1928.57	1.14	ppb
Silver	107-1	0.23	0.26	0.24	0.24	6.14	ppb
Silver	109-1	0.24	0.24	0.25	0.24	2.27	ppb
Sodium	23-2	163.59	183.63	195.82	181.01	8.99	ppb
Strontium	86-1	72.80	74.19	74.20	73.73	1.09	ppb
Strontium	88-1	75.29	75.57	76.43	75.76	0.78	ppb
Sulfur	34-1	-1735.34	-1763.02	-1737.39	-1745.25		ppb
Terbium	159-1				104		%
Terbium	159-2				100		%
Thallium	203-1	0.22	0.22	0.22	0.22	1.16	ppb
Thallium	205-1	0.20	0.23	0.22	0.22	5.63	ppb
Tin	118-1	0.36	0.38	0.39	0.37	4.50	ppb
Titanium	47-1	506.75	509.74	512.08	509.52	0.52	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-21

Instrumnet Name :

P7

Client Sample ID :

MYE4N3

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 22:06:56

DataFile Name :

088SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.63	0.64	0.63	0.63	1.37	ppb
Vanadium	51-2	82.62	91.71	95.84	90.06	7.51	ppb
Yttrium	89-1				102		%
Yttrium	89-2				100		%
Zinc	66-2	179.63	198.96	207.08	195.23	7.22	ppb
Zirconium	90-1	9.77	9.95	9.93	9.88	1.03	ppb
Zirconium	91-1	9.90	10.11	9.92	9.98	1.15	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-22 Instrumnet Name : P7
Client Sample ID : MYE4N4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:10:04 DataFile Name : 089SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	46298.64	46286.47	45868.72	46151.28	0.53	ppb
Antimony	121-1	0.30	0.29	0.30	0.30	2.46	ppb
Arsenic	75-2	44.29	44.87	44.22	44.46	0.80	ppb
Barium	135-1	239.98	240.43	250.11	243.51	2.35	ppb
Barium	137-1	240.70	242.52	247.55	243.59	1.46	ppb
Beryllium	9-1	0.89	0.73	0.78	0.80	9.97	ppb
Bismuth	209-1				101		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.48	2.11	2.32	2.30	8.12	ppb
Cadmium	106-1	0.97	0.51	1.27	0.92	41.53	ppb
Cadmium	111-1	1.89	1.87	1.99	1.92	3.50	ppb
Calcium	43-1	16336.88	16547.12	16552.73	16478.91	0.75	ppb
Calcium	44-1	15400.00	15441.84	15424.48	15422.11	0.14	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	43.65	43.56	43.67	43.63	0.14	ppb
Cobalt	59-2	71.85	71.03	71.11	71.33	0.64	ppb
Copper	63-2	153.65	151.22	151.58	152.15	0.86	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				106		%
Indium	115-1				99		%
Indium	115-2				101		%
Iron	56-2	98801.33	97540.82	98375.62	98239.26	0.65	ppb
Iron	57-2	96875.32	95983.66	97262.94	96707.31	0.68	ppb
Iron	54-2	98718.45	97739.55	97569.03	98009.01	0.63	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-22 Instrumnet Name : P7
Client Sample ID : MYE4N4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:10:04 DataFile Name : 089SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	26.79	27.58	27.44	27.27	1.53	ppb
Lead	207-1	25.95	26.63	26.65	26.41	1.52	ppb
Lead	208-1	25.97	26.84	26.68	26.49	1.74	ppb
Lithium	6-1				97		%
Magnesium	24-2	21782.62	21510.77	21446.69	21580.03	0.83	ppb
Manganese	55-2	3201.46	3144.93	3172.10	3172.83	0.89	ppb
Molybdenum	94-1	7.56	7.77	7.80	7.71	1.70	ppb
Molybdenum	95-1	0.75	0.73	0.72	0.73	2.55	ppb
Molybdenum	96-1	2.25	2.26	2.25	2.25	0.19	ppb
Molybdenum	97-1	0.68	0.74	0.71	0.71	3.93	ppb
Molybdenum	98-1	0.66	0.66	0.68	0.67	2.16	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	40.80	40.68	40.60	40.70	0.25	ppb
Phosphorus	31-2	1124.78	1151.31	1151.41	1142.50	1.34	ppb
Potassium	39-2	2573.63	2551.04	2560.65	2561.77	0.44	ppb
Rhodium	103-1				95		%
Rhodium	103-2				100		%
Scandium	45-1				99		%
Scandium	45-2				105		%
Selenium	82-1	0.10	0.33	0.00	0.15	117.13	ppb
Selenium	77-2	10.27	9.63	6.50	8.80	22.95	ppb
Selenium	78-2	0.93	0.53	1.60	1.02	52.87	ppb
Silicon	28-1	1752.27	1789.63	1793.04	1778.31	1.27	ppb
Silver	107-1	0.67	0.72	0.73	0.70	4.75	ppb
Silver	109-1	0.69	0.70	0.73	0.70	2.83	ppb
Sodium	23-2	491.62	487.08	489.23	489.31	0.46	ppb
Strontium	86-1	96.86	97.48	96.91	97.08	0.35	ppb
Strontium	88-1	99.79	102.82	100.54	101.05	1.56	ppb
Sulfur	34-1	-2012.70	-2029.77	-2032.20	-2024.89		ppb
Terbium	159-1				104		%
Terbium	159-2				107		%
Thallium	203-1	0.20	0.21	0.22	0.21	4.32	ppb
Thallium	205-1	0.20	0.21	0.20	0.20	3.99	ppb
Tin	118-1	0.19	0.21	0.23	0.21	10.60	ppb
Titanium	47-1	1406.06	1419.27	1458.93	1428.09	1.93	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-22

Instrumnet Name :

P7

Client Sample ID :

MYE4N4

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 22:10:04

DataFile Name :

089SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.41	0.42	0.43	0.42	1.87	ppb
Vanadium	51-2	164.69	163.92	165.17	164.59	0.38	ppb
Yttrium	89-1				102		%
Yttrium	89-2				107		%
Zinc	66-2	418.84	416.25	420.50	418.53	0.51	ppb
Zirconium	90-1	4.22	4.20	4.21	4.21	0.27	ppb
Zirconium	91-1	4.32	4.26	4.17	4.25	1.72	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:13:10 DataFile Name : 090SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	5521.64	5551.79	5518.79	5530.74	0.33	ppb
Antimony	121-1	0.09	0.10	0.09	0.09	1.21	ppb
Arsenic	75-2	13.15	13.07	12.92	13.05	0.87	ppb
Barium	135-1	86.64	88.28	89.26	88.06	1.50	ppb
Barium	137-1	87.14	88.43	88.69	88.09	0.94	ppb
Beryllium	9-1	0.19	0.19	0.24	0.21	13.98	ppb
Bismuth	209-1				101		%
Bismuth	209-2				107		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.64	0.59	0.47	0.57	15.92	ppb
Cadmium	106-1	-0.83	-0.88	-1.64	-1.12		ppb
Cadmium	111-1	0.31	0.30	0.25	0.29	12.70	ppb
Calcium	43-1	5635.47	5769.77	5752.11	5719.12	1.28	ppb
Calcium	44-1	5402.83	5470.10	5416.55	5429.82	0.65	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	12.08	11.89	11.93	11.97	0.81	ppb
Cobalt	59-2	7.59	7.68	7.51	7.60	1.13	ppb
Copper	63-2	50.07	50.48	49.54	50.03	0.94	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				106		%
Indium	115-1				98		%
Indium	115-2				104		%
Iron	56-2	11107.10	11023.33	10828.95	10986.46	1.30	ppb
Iron	57-2	11206.77	11229.92	11116.85	11184.51	0.53	ppb
Iron	54-2	11373.65	11099.49	11017.18	11163.44	1.67	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:13:10 DataFile Name : 090SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	24.60	25.50	25.12	25.07	1.79	ppb
Lead	207-1	24.12	24.86	24.71	24.56	1.59	ppb
Lead	208-1	24.25	24.88	24.75	24.63	1.36	ppb
Lithium	6-1				96		%
Magnesium	24-2	3156.18	3190.63	3132.34	3159.72	0.93	ppb
Manganese	55-2	407.01	404.27	401.13	404.14	0.73	ppb
Molybdenum	94-1	4.43	4.28	4.55	4.42	3.12	ppb
Molybdenum	95-1	0.05	0.05	0.06	0.05	10.50	ppb
Molybdenum	96-1	0.58	0.58	0.60	0.59	1.69	ppb
Molybdenum	97-1	0.04	0.03	0.04	0.04	13.57	ppb
Molybdenum	98-1	0.01	0.03	0.03	0.02	36.21	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	14.02	13.77	13.65	13.81	1.36	ppb
Phosphorus	31-2	517.01	503.28	502.67	507.65	1.60	ppb
Potassium	39-2	1691.18	1711.90	1686.38	1696.49	0.80	ppb
Rhodium	103-1				95		%
Rhodium	103-2				103		%
Scandium	45-1				93		%
Scandium	45-2				101		%
Selenium	82-1	-0.48	-0.48	0.15	-0.27		ppb
Selenium	77-2	3.33	2.24	2.02	2.53	27.73	ppb
Selenium	78-2	0.49	-0.02	0.49	0.32	91.61	ppb
Silicon	28-1	370.71	375.68	379.19	375.19	1.14	ppb
Silver	107-1	0.13	0.12	0.11	0.12	7.02	ppb
Silver	109-1	0.11	0.11	0.11	0.11	4.23	ppb
Sodium	23-2	31.50	30.20	30.06	30.58	2.59	ppb
Strontium	86-1	17.83	17.94	18.15	17.97	0.90	ppb
Strontium	88-1	17.65	17.85	18.29	17.93	1.82	ppb
Sulfur	34-1	-1912.41	-1848.07	-1784.42	-1848.30		ppb
Terbium	159-1				102		%
Terbium	159-2				106		%
Thallium	203-1	0.03	0.04	0.04	0.04	11.65	ppb
Thallium	205-1	0.04	0.04	0.04	0.04	8.60	ppb
Tin	118-1	0.12	0.12	0.11	0.12	7.92	ppb
Titanium	47-1	118.07	120.13	121.25	119.82	1.35	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-08DLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4M1

Dilution Factor :

4

Date & Time Acquired :

2024-11-02 22:13:10

DataFile Name :

090SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.12	0.13	0.13	0.13	3.77	ppb
Vanadium	51-2	19.80	19.70	19.44	19.64	0.95	ppb
Yttrium	89-1				97		%
Yttrium	89-2				104		%
Zinc	66-2	70.75	69.89	69.72	70.12	0.78	ppb
Zirconium	90-1	2.60	2.69	2.65	2.65	1.69	ppb
Zirconium	91-1	2.71	2.65	2.71	2.69	1.26	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-09DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1D Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:16:21 DataFile Name : 091SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	5632.44	5717.32	5619.48	5656.41	0.94	ppb
Antimony	121-1	0.08	0.10	0.10	0.09	10.71	ppb
Arsenic	75-2	12.57	11.96	12.78	12.44	3.39	ppb
Barium	135-1	86.68	88.84	89.28	88.27	1.58	ppb
Barium	137-1	86.41	88.67	89.66	88.25	1.89	ppb
Beryllium	9-1	0.16	0.21	0.22	0.20	15.98	ppb
Bismuth	209-1				101		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.87	0.55	0.58	0.67	26.16	ppb
Cadmium	106-1	-0.53	-0.76	-0.92	-0.74		ppb
Cadmium	111-1	0.31	0.29	0.35	0.32	9.12	ppb
Calcium	43-1	5594.74	5700.07	5718.62	5671.14	1.18	ppb
Calcium	44-1	5312.48	5412.46	5444.01	5389.65	1.27	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	11.89	11.84	11.77	11.83	0.51	ppb
Cobalt	59-2	7.43	7.41	7.40	7.41	0.16	ppb
Copper	63-2	49.11	49.29	49.08	49.16	0.23	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				106		%
Indium	115-1				98		%
Indium	115-2				103		%
Iron	56-2	10892.29	10974.33	10831.70	10899.44	0.66	ppb
Iron	57-2	11045.91	11165.48	11056.09	11089.16	0.60	ppb
Iron	54-2	10928.20	11281.97	11085.95	11098.71	1.60	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-09DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1D Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:16:21 DataFile Name : 091SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	24.07	24.49	25.06	24.54	2.03	ppb
Lead	207-1	23.65	24.08	24.38	24.04	1.51	ppb
Lead	208-1	23.75	24.10	24.57	24.14	1.71	ppb
Lithium	6-1				95		%
Magnesium	24-2	3100.08	3192.95	3145.65	3146.22	1.48	ppb
Manganese	55-2	396.34	400.26	396.39	397.67	0.57	ppb
Molybdenum	94-1	4.67	4.70	4.77	4.71	1.13	ppb
Molybdenum	95-1	0.03	0.04	0.03	0.03	11.73	ppb
Molybdenum	96-1	0.59	0.60	0.60	0.60	1.27	ppb
Molybdenum	97-1	0.02	0.02	0.03	0.02	13.81	ppb
Molybdenum	98-1	0.02	0.02	0.02	0.02	9.05	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	12.95	13.47	12.96	13.13	2.26	ppb
Phosphorus	31-2	478.79	492.61	472.62	481.34	2.13	ppb
Potassium	39-2	1686.15	1692.71	1685.52	1688.13	0.24	ppb
Rhodium	103-1				96		%
Rhodium	103-2				102		%
Scandium	45-1				93		%
Scandium	45-2				101		%
Selenium	82-1	-0.14	-0.19	-0.02	-0.12		ppb
Selenium	77-2	3.35	2.71	1.61	2.56	34.49	ppb
Selenium	78-2	-0.02	1.54	1.45	0.99	88.29	ppb
Silicon	28-1	383.56	384.12	383.62	383.76	0.08	ppb
Silver	107-1	0.11	0.11	0.10	0.11	5.01	ppb
Silver	109-1	0.11	0.12	0.12	0.12	7.43	ppb
Sodium	23-2	27.18	28.71	26.68	27.52	3.84	ppb
Strontium	86-1	17.47	18.08	17.54	17.70	1.90	ppb
Strontium	88-1	17.74	18.08	17.76	17.86	1.08	ppb
Sulfur	34-1	-1814.97	-1757.77	-1741.95	-1771.56		ppb
Terbium	159-1				101		%
Terbium	159-2				105		%
Thallium	203-1	0.04	0.04	0.04	0.04	5.09	ppb
Thallium	205-1	0.04	0.04	0.04	0.04	5.04	ppb
Tin	118-1	0.10	0.10	0.10	0.10	2.84	ppb
Titanium	47-1	116.84	119.41	119.22	118.49	1.21	ppb

LB Number :	LB133265	Operator :	Jaswal
Lab Sample ID :	P4520-09DLX4	Instrumnet Name :	P7
Client Sample ID :	MYE4M1D	Dilution Factor :	4
Date & Time Acquired :	2024-11-02 22:16:21	DataFile Name :	091SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.12	0.12	0.13	0.13	4.99	ppb
Vanadium	51-2	19.58	19.71	19.57	19.62	0.38	ppb
Yttrium	89-1				97		%
Yttrium	89-2				103		%
Zinc	66-2	68.48	69.34	67.96	68.59	1.01	ppb
Zirconium	90-1	2.77	2.81	2.77	2.78	0.96	ppb
Zirconium	91-1	2.86	2.84	2.79	2.83	1.41	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4M1L Dilution Factor : 20
Date & Time Acquired : 2024-11-02 22:19:30 DataFile Name : 092SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1098.85	1120.61	1216.66	1145.37	5.47	ppb
Antimony	121-1	0.01	0.01	0.01	0.01	10.57	ppb
Arsenic	75-2	2.42	2.36	2.74	2.51	8.17	ppb
Barium	135-1	17.38	17.52	17.79	17.56	1.20	ppb
Barium	137-1	17.30	17.23	17.83	17.45	1.88	ppb
Beryllium	9-1	-0.01	0.00	0.02	0.01	277.76	ppb
Bismuth	209-1				103		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.12	0.16	0.10	0.13	22.06	ppb
Cadmium	106-1	-1.86	-1.37	-1.58	-1.61		ppb
Cadmium	111-1	-0.07	-0.05	-0.04	-0.05		ppb
Calcium	43-1	1113.27	1148.58	1158.55	1140.13	2.09	ppb
Calcium	44-1	1137.31	1163.14	1161.52	1153.99	1.25	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2.53	2.61	2.79	2.64	4.91	ppb
Cobalt	59-2	1.51	1.54	1.68	1.58	5.84	ppb
Copper	63-2	9.88	9.97	10.81	10.22	5.00	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				104		%
Indium	115-1				100		%
Indium	115-2				101		%
Iron	56-2	2189.62	2234.35	2412.21	2278.73	5.17	ppb
Iron	57-2	2224.75	2258.32	2445.72	2309.60	5.16	ppb
Iron	54-2	2273.87	2277.08	2505.76	2352.23	5.65	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4M1L Dilution Factor : 20
Date & Time Acquired : 2024-11-02 22:19:30 DataFile Name : 092SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	4.81	4.87	4.86	4.85	0.68	ppb
Lead	207-1	4.74	4.65	4.70	4.70	0.94	ppb
Lead	208-1	4.75	4.78	4.76	4.76	0.39	ppb
Lithium	6-1				95		%
Magnesium	24-2	629.78	648.10	698.99	658.96	5.44	ppb
Manganese	55-2	79.75	81.02	89.02	83.26	6.03	ppb
Molybdenum	94-1	0.76	0.81	0.84	0.80	4.70	ppb
Molybdenum	95-1	-0.04	-0.03	-0.03	-0.03		ppb
Molybdenum	96-1	0.07	0.05	0.06	0.06	12.69	ppb
Molybdenum	97-1	-0.04	-0.02	-0.03	-0.03		ppb
Molybdenum	98-1	-0.03	-0.04	-0.04	-0.04		ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	2.72	2.88	2.99	2.87	4.80	ppb
Phosphorus	31-2	73.08	78.81	91.87	81.25	11.85	ppb
Potassium	39-2	320.46	324.85	366.73	337.35	7.57	ppb
Rhodium	103-1				98		%
Rhodium	103-2				101		%
Scandium	45-1				94		%
Scandium	45-2				98		%
Selenium	82-1	0.03	0.23	0.08	0.12	88.92	ppb
Selenium	77-2	0.46	0.00	0.51	0.32	86.68	ppb
Selenium	78-2	-1.06	-0.68	0.90	-0.28		ppb
Silicon	28-1	76.41	76.50	76.06	76.32	0.31	ppb
Silver	107-1	0.02	0.02	0.02	0.02	12.21	ppb
Silver	109-1	0.02	0.02	0.03	0.02	12.57	ppb
Sodium	23-2	-7.81	-6.15	-3.61	-5.86		ppb
Strontium	86-1	3.40	3.60	3.50	3.50	2.90	ppb
Strontium	88-1	3.41	3.45	3.49	3.45	1.28	ppb
Sulfur	34-1	-1931.59	-1832.95	-1804.06	-1856.20		ppb
Terbium	159-1				102		%
Terbium	159-2				104		%
Thallium	203-1	0.00	0.00	0.00	0.00		ppb
Thallium	205-1	0.00	0.00	0.00	0.00	2382.40	ppb
Tin	118-1	0.00	0.02	0.00	0.01	181.48	ppb
Titanium	47-1	23.41	23.90	24.09	23.80	1.46	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-08LDLX20

Instrumnet Name :

P7

Client Sample ID :

MYE4M1L

Dilution Factor :

20

Date & Time Acquired :

2024-11-02 22:19:30

DataFile Name :

092SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.02	0.02	0.02	3.38	ppb
Vanadium	51-2	3.87	4.00	4.34	4.07	6.01	ppb
Yttrium	89-1				98		%
Yttrium	89-2				101		%
Zinc	66-2	14.27	13.90	15.58	14.58	6.04	ppb
Zirconium	90-1	0.51	0.52	0.52	0.52	1.11	ppb
Zirconium	91-1	0.50	0.50	0.51	0.50	1.57	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-10DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1S Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:22:45 DataFile Name : 093SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	5696.97	5835.59	5693.91	5742.16	1.41	ppb
Antimony	121-1	13.98	14.19	15.02	14.40	3.84	ppb
Arsenic	75-2	17.95	17.99	17.73	17.89	0.77	ppb
Barium	135-1	350.37	353.82	373.98	359.39	3.55	ppb
Barium	137-1	380.29	384.26	406.34	390.29	3.60	ppb
Beryllium	9-1	7.41	7.51	8.26	7.73	6.05	ppb
Bismuth	209-1				101		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	7.78	7.29	6.91	7.32	5.95	ppb
Cadmium	106-1	5.75	5.45	5.02	5.41	6.79	ppb
Cadmium	111-1	7.46	7.50	7.97	7.64	3.71	ppb
Calcium	43-1	5505.04	5594.52	5957.04	5685.54	4.21	ppb
Calcium	44-1	5258.00	5299.52	5638.87	5398.80	3.87	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	41.84	42.45	42.02	42.10	0.75	ppb
Cobalt	59-2	81.56	81.94	81.88	81.79	0.25	ppb
Copper	63-2	83.03	83.91	83.75	83.56	0.56	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				107		%
Indium	115-1				97		%
Indium	115-2				104		%
Iron	56-2	10858.65	10967.35	10897.71	10907.90	0.50	ppb
Iron	57-2	11077.10	11100.27	11118.91	11098.76	0.19	ppb
Iron	54-2	10871.70	11214.04	11120.67	11068.80	1.60	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-10DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1S Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:22:45 DataFile Name : 093SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	26.46	26.20	28.66	27.11	4.99	ppb
Lead	207-1	26.16	25.93	28.12	26.74	4.51	ppb
Lead	208-1	26.16	26.12	28.22	26.83	4.48	ppb
Lithium	6-1				95		%
Magnesium	24-2	3065.47	3076.28	3066.80	3069.52	0.19	ppb
Manganese	55-2	463.62	465.36	466.04	465.00	0.27	ppb
Molybdenum	94-1	4.55	4.66	5.13	4.78	6.54	ppb
Molybdenum	95-1	0.04	0.03	0.06	0.04	38.34	ppb
Molybdenum	96-1	0.57	0.59	0.65	0.61	7.16	ppb
Molybdenum	97-1	0.04	0.04	0.04	0.04	5.54	ppb
Molybdenum	98-1	0.02	0.02	0.03	0.02	25.16	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	86.60	87.60	87.07	87.09	0.57	ppb
Phosphorus	31-2	477.25	485.12	473.36	478.58	1.25	ppb
Potassium	39-2	1634.61	1646.99	1645.32	1642.31	0.41	ppb
Rhodium	103-1				95		%
Rhodium	103-2				103		%
Scandium	45-1				92		%
Scandium	45-2				101		%
Selenium	82-1	15.69	15.67	17.74	16.36	7.26	ppb
Selenium	77-2	18.18	20.44	23.25	20.62	12.31	ppb
Selenium	78-2	15.85	14.89	19.26	16.67	13.78	ppb
Silicon	28-1	384.97	383.35	411.02	393.11	3.95	ppb
Silver	107-1	6.09	6.23	6.70	6.34	5.04	ppb
Silver	109-1	6.06	6.27	6.64	6.32	4.59	ppb
Sodium	23-2	26.23	27.49	25.61	26.44	3.62	ppb
Strontium	86-1	17.47	17.70	18.72	17.96	3.70	ppb
Strontium	88-1	17.33	17.65	18.78	17.92	4.25	ppb
Sulfur	34-1	-1641.74	-1655.65	-1498.70	-1598.70		ppb
Terbium	159-1				102		%
Terbium	159-2				106		%
Thallium	203-1	6.22	6.17	6.76	6.38	5.08	ppb
Thallium	205-1	6.14	6.15	6.67	6.32	4.79	ppb
Tin	118-1	0.09	0.08	0.10	0.09	9.67	ppb
Titanium	47-1	115.12	115.52	123.19	117.94	3.86	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-10DLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4M1S

Dilution Factor :

4

Date & Time Acquired :

2024-11-02 22:22:45

DataFile Name :

093SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.13	0.14	0.14	0.14	2.30	ppb
Vanadium	51-2	93.31	94.22	93.91	93.81	0.49	ppb
Yttrium	89-1				96		%
Yttrium	89-2				104		%
Zinc	66-2	140.25	141.30	140.18	140.58	0.45	ppb
Zirconium	90-1	2.68	2.69	2.99	2.79	6.34	ppb
Zirconium	91-1	2.76	2.82	2.99	2.86	4.13	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:25:53 DataFile Name : 094SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	6527.54	6514.57	6558.74	6533.62	0.35	ppb
Antimony	121-1	68.31	68.96	77.30	71.52	7.01	ppb
Arsenic	75-2	40.10	38.57	39.34	39.33	1.95	ppb
Barium	135-1	1500.77	1502.51	1659.24	1554.17	5.86	ppb
Barium	137-1	1486.17	1477.68	1653.20	1539.02	6.43	ppb
Beryllium	9-1	37.18	36.43	39.77	37.80	4.63	ppb
Bismuth	209-1				100		%
Bismuth	209-2				107		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	33.90	33.57	36.71	34.73	4.96	ppb
Cadmium	106-1	30.31	31.68	35.02	32.34	7.49	ppb
Cadmium	111-1	34.55	35.39	39.51	36.49	7.28	ppb
Calcium	43-1	5243.62	5365.04	5933.36	5514.00	6.68	ppb
Calcium	44-1	4959.41	5095.85	5613.45	5222.90	6.61	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	156.67	157.93	155.09	156.56	0.91	ppb
Cobalt	59-2	375.16	379.13	367.83	374.04	1.53	ppb
Copper	63-2	213.67	214.62	212.28	213.53	0.55	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				99		%
Holmium	165-2				107		%
Indium	115-1				95		%
Indium	115-2				103		%
Iron	56-2	11122.21	11071.41	11034.95	11076.19	0.40	ppb
Iron	57-2	11241.24	11275.93	11256.80	11257.99	0.15	ppb
Iron	54-2	11217.25	11318.18	11143.68	11226.37	0.78	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:25:53 DataFile Name : 094SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	35.30	35.19	39.89	36.79	7.30	ppb
Lead	207-1	34.45	34.74	38.85	36.01	6.84	ppb
Lead	208-1	34.89	34.73	39.19	36.27	6.97	ppb
Lithium	6-1				93		%
Magnesium	24-2	2919.96	2917.06	2914.84	2917.29	0.09	ppb
Manganese	55-2	733.12	729.82	728.37	730.44	0.33	ppb
Molybdenum	94-1	4.48	4.62	5.19	4.76	7.92	ppb
Molybdenum	95-1	0.05	0.05	0.06	0.05	13.04	ppb
Molybdenum	96-1	0.57	0.61	0.69	0.63	9.60	ppb
Molybdenum	97-1	0.03	0.04	0.03	0.03	14.71	ppb
Molybdenum	98-1	0.03	0.03	0.04	0.03	31.54	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	368.66	371.79	365.43	368.63	0.86	ppb
Phosphorus	31-2	448.49	469.23	446.12	454.61	2.80	ppb
Potassium	39-2	1576.34	1579.88	1557.95	1571.39	0.75	ppb
Rhodium	103-1				92		%
Rhodium	103-2				103		%
Scandium	45-1				91		%
Scandium	45-2				100		%
Selenium	82-1	78.17	80.13	86.45	81.58	5.31	ppb
Selenium	77-2	75.49	79.92	79.29	78.23	3.06	ppb
Selenium	78-2	77.26	82.90	76.89	79.02	4.26	ppb
Silicon	28-1	363.78	377.86	419.49	387.05	7.49	ppb
Silver	107-1	29.44	29.99	33.62	31.02	7.32	ppb
Silver	109-1	29.67	29.91	33.50	31.03	6.91	ppb
Sodium	23-2	25.80	26.41	24.78	25.66	3.23	ppb
Strontium	86-1	16.48	16.46	18.78	17.24	7.73	ppb
Strontium	88-1	16.75	16.88	18.67	17.43	6.14	ppb
Sulfur	34-1	-1798.34	-1678.17	-1412.23	-1629.58		ppb
Terbium	159-1				100		%
Terbium	159-2				108		%
Thallium	203-1	30.35	29.97	33.76	31.36	6.66	ppb
Thallium	205-1	29.87	29.68	33.63	31.06	7.16	ppb
Tin	118-1	0.09	0.10	0.14	0.11	23.88	ppb
Titanium	47-1	107.00	110.66	123.90	113.85	7.81	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-08MSDLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4M1SRE

Dilution Factor :

4

Date & Time Acquired :

2024-11-02 22:25:53

DataFile Name :

094SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.12	0.12	0.14	0.13	6.75	ppb
Vanadium	51-2	379.41	377.68	364.98	374.02	2.11	ppb
Yttrium	89-1				94		%
Yttrium	89-2				103		%
Zinc	66-2	416.54	416.14	417.41	416.70	0.16	ppb
Zirconium	90-1	2.67	2.71	2.98	2.79	6.01	ppb
Zirconium	91-1	2.62	2.74	3.07	2.81	8.25	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BL Instrumnet Name : P7
Client Sample ID : PBS456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:31:57 DataFile Name : 095CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	2.68	3.28	3.20	3.05	10.68	ppb
Antimony	121-1	0.01	0.01	0.01	0.01	28.74	ppb
Arsenic	75-2	0.03	-0.02	-0.01	0.00	1796.95	ppb
Barium	135-1	0.08	0.09	0.08	0.08	8.00	ppb
Barium	137-1	0.09	0.09	0.06	0.08	18.22	ppb
Beryllium	9-1	-0.02	-0.01	-0.03	-0.02		ppb
Bismuth	209-1				95		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.06	0.01	0.10	0.06	78.89	ppb
Cadmium	106-1	-2.21	-2.55	-3.35	-2.70		ppb
Cadmium	111-1	-0.17	-0.18	-0.24	-0.20		ppb
Calcium	43-1	-1.82	-2.74	-3.66	-2.74		ppb
Calcium	44-1	-1.09	-3.22	-3.37	-2.56		ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.01	0.02	0.01	0.01	55.26	ppb
Cobalt	59-2	0.01	0.01	0.02	0.01	31.98	ppb
Copper	63-2	-0.09	-0.08	-0.05	-0.07		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				96		%
Holmium	165-2				105		%
Indium	115-1				94		%
Indium	115-2				103		%
Iron	56-2	7.82	7.61	8.17	7.87	3.56	ppb
Iron	57-2	4.02	4.62	8.09	5.58	39.38	ppb
Iron	54-2	6.86	8.26	8.04	7.72	9.78	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BL Instrumnet Name : P7
Client Sample ID : PBS456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:31:57 DataFile Name : 095CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	-0.06	-0.06	-0.07	-0.06		ppb
Lead	207-1	-0.06	-0.05	-0.06	-0.06		ppb
Lead	208-1	-0.05	-0.06	-0.06	-0.06		ppb
Lithium	6-1				92		%
Magnesium	24-2	-0.36	-0.59	-0.47	-0.48		ppb
Manganese	55-2	0.30	0.29	0.32	0.31	4.84	ppb
Molybdenum	94-1	-0.06	-0.06	-0.05	-0.06		ppb
Molybdenum	95-1	-0.04	-0.05	-0.05	-0.05		ppb
Molybdenum	96-1	-0.05	-0.06	-0.06	-0.06		ppb
Molybdenum	97-1	-0.06	-0.05	-0.05	-0.05		ppb
Molybdenum	98-1	-0.05	-0.06	-0.06	-0.06		ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.05	0.05	0.09	0.06	31.67	ppb
Phosphorus	31-2	-23.07	-19.57	-21.32	-21.32		ppb
Potassium	39-2	-16.65	-12.49	-14.05	-14.40		ppb
Rhodium	103-1				92		%
Rhodium	103-2				103		%
Scandium	45-1				89		%
Scandium	45-2				99		%
Selenium	82-1	-0.51	0.17	-0.77	-0.37		ppb
Selenium	77-2	-0.22	0.00	-0.22	-0.15		ppb
Selenium	78-2	-0.91	-0.46	-0.15	-0.51		ppb
Silicon	28-1	1.48	-0.22	-0.10	0.39	246.70	ppb
Silver	107-1	0.00	0.00	0.00	0.00	396.47	ppb
Silver	109-1	0.00	0.00	0.00	0.00	44.18	ppb
Sodium	23-2	-15.96	-16.75	-16.41	-16.37		ppb
Strontium	86-1	-0.06	-0.05	-0.09	-0.07		ppb
Strontium	88-1	-0.08	-0.10	-0.09	-0.09		ppb
Sulfur	34-1	-1379.24	-1773.97	-1755.37	-1636.19		ppb
Terbium	159-1				97		%
Terbium	159-2				105		%
Thallium	203-1	0.00	-0.01	-0.01	-0.01		ppb
Thallium	205-1	0.00	-0.01	-0.01	-0.01		ppb
Tin	118-1	0.01	-0.02	-0.02	-0.01		ppb
Titanium	47-1	0.04	0.02	0.24	0.10	118.18	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BL Instrumnet Name : P7
Client Sample ID : PBS456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:31:57 DataFile Name : 095CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	-0.01	-0.01	-0.01		ppb
Vanadium	51-2	0.02	0.03	0.03	0.03	20.45	ppb
Yttrium	89-1				92		%
Yttrium	89-2				102		%
Zinc	66-2	-0.18	-0.11	-0.14	-0.14		ppb
Zirconium	90-1	-0.01	-0.01	-0.01	-0.01		ppb
Zirconium	91-1	-0.01	-0.01	0.00	-0.01		ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BS Instrumnet Name : P7
Client Sample ID : LCS456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:35:14 DataFile Name : 096LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	39.75	38.08	38.96	38.93	2.14	ppb
Antimony	121-1	3.53	3.69	3.72	3.65	2.75	ppb
Arsenic	75-2	2.08	1.81	2.12	2.00	8.47	ppb
Barium	135-1	16.82	17.79	17.61	17.41	2.95	ppb
Barium	137-1	17.18	17.53	17.76	17.49	1.65	ppb
Beryllium	9-1	1.93	2.17	1.92	2.01	6.94	ppb
Bismuth	209-1				101		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.24	1.97	2.01	2.07	7.11	ppb
Cadmium	106-1	0.30	1.02	-0.23	0.36	172.75	ppb
Cadmium	111-1	1.74	1.90	1.88	1.84	4.73	ppb
Calcium	43-1	911.00	948.45	948.49	935.98	2.31	ppb
Calcium	44-1	930.89	956.80	942.92	943.54	1.37	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	4.13	3.85	3.92	3.97	3.67	ppb
Cobalt	59-2	2.00	1.91	2.00	1.97	2.68	ppb
Copper	63-2	3.94	3.74	3.88	3.86	2.67	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				104		%
Indium	115-1				98		%
Indium	115-2				103		%
Iron	56-2	420.91	399.69	410.70	410.43	2.59	ppb
Iron	57-2	423.59	402.39	419.38	415.12	2.70	ppb
Iron	54-2	431.08	408.46	423.32	420.95	2.73	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BS Instrumnet Name : P7
Client Sample ID : LCS456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:35:14 DataFile Name : 096LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1.50	1.54	1.54	1.53	1.40	ppb
Lead	207-1	1.54	1.54	1.58	1.55	1.31	ppb
Lead	208-1	1.54	1.57	1.55	1.55	0.92	ppb
Lithium	6-1				95		%
Magnesium	24-2	1033.70	976.11	1006.47	1005.42	2.87	ppb
Manganese	55-2	2.17	2.14	2.23	2.18	1.98	ppb
Molybdenum	94-1	9.95	10.05	10.20	10.07	1.23	ppb
Molybdenum	95-1	8.43	8.50	8.44	8.46	0.47	ppb
Molybdenum	96-1	8.43	8.61	8.66	8.57	1.47	ppb
Molybdenum	97-1	8.22	8.41	8.54	8.39	1.96	ppb
Molybdenum	98-1	8.17	8.35	8.31	8.28	1.12	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.99	2.01	2.19	2.06	5.25	ppb
Phosphorus	31-2	-20.04	-22.50	-23.39	-21.98		ppb
Potassium	39-2	990.06	929.20	962.71	960.66	3.17	ppb
Rhodium	103-1				96		%
Rhodium	103-2				104		%
Scandium	45-1				93		%
Scandium	45-2				100		%
Selenium	82-1	9.16	10.22	9.61	9.66	5.50	ppb
Selenium	77-2	9.25	8.53	9.35	9.04	4.93	ppb
Selenium	78-2	10.32	9.33	10.33	9.99	5.74	ppb
Silicon	28-1	5.30	5.72	5.66	5.56	4.14	ppb
Silver	107-1	1.78	1.79	1.89	1.82	3.29	ppb
Silver	109-1	1.72	1.85	1.87	1.82	4.45	ppb
Sodium	23-2	963.28	906.99	943.68	937.98	3.05	ppb
Strontium	86-1	1.73	1.81	1.84	1.79	3.30	ppb
Strontium	88-1	1.73	1.81	1.78	1.77	2.24	ppb
Sulfur	34-1	-1458.18	-1402.32	-1339.84	-1400.12		ppb
Terbium	159-1				100		%
Terbium	159-2				105		%
Thallium	203-1	1.56	1.57	1.54	1.56	1.00	ppb
Thallium	205-1	1.51	1.57	1.54	1.54	1.66	ppb
Tin	118-1	0.00	0.00	0.01	0.00	166.43	ppb
Titanium	47-1	8.51	9.11	9.02	8.88	3.61	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

PB164456BS

Instrumnet Name :

P7

Client Sample ID :

LCS456

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 22:35:14

DataFile Name :

096LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.46	1.54	1.54	1.51	3.17	ppb
Vanadium	51-2	9.93	9.36	9.67	9.66	2.94	ppb
Yttrium	89-1				95		%
Yttrium	89-2				102		%
Zinc	66-2	9.86	9.09	10.09	9.68	5.39	ppb
Zirconium	90-1	1.87	1.90	1.92	1.90	1.30	ppb
Zirconium	91-1	1.85	1.90	1.90	1.88	1.43	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BSD Instrumnet Name : P7
Client Sample ID : LCSD456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:38:27 DataFile Name : 097LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	19.72	19.50	18.79	19.34	2.53	ppb
Antimony	121-1	1.74	1.82	1.76	1.77	2.38	ppb
Arsenic	75-2	1.08	0.91	0.92	0.97	10.05	ppb
Barium	135-1	8.32	8.80	8.32	8.48	3.25	ppb
Barium	137-1	8.29	8.44	8.47	8.40	1.15	ppb
Beryllium	9-1	0.94	0.95	0.84	0.91	6.56	ppb
Bismuth	209-1				100		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.08	0.81	0.79	0.89	18.26	ppb
Cadmium	106-1	-0.21	-0.69	-0.86	-0.59		ppb
Cadmium	111-1	0.76	0.74	0.78	0.76	2.47	ppb
Calcium	43-1	438.45	459.74	469.75	455.98	3.51	ppb
Calcium	44-1	453.97	452.67	462.00	456.21	1.11	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1.89	1.99	1.90	1.92	2.86	ppb
Cobalt	59-2	0.96	0.92	0.94	0.94	2.18	ppb
Copper	63-2	1.79	1.84	1.85	1.83	1.72	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				99		%
Holmium	165-2				103		%
Indium	115-1				98		%
Indium	115-2				102		%
Iron	56-2	204.99	209.09	203.48	205.86	1.41	ppb
Iron	57-2	200.65	205.02	208.32	204.66	1.88	ppb
Iron	54-2	209.36	212.47	207.11	209.65	1.28	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BSD Instrumnet Name : P7
Client Sample ID : LCSD456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:38:27 DataFile Name : 097LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.73	0.78	0.68	0.73	6.93	ppb
Lead	207-1	0.74	0.72	0.73	0.73	1.49	ppb
Lead	208-1	0.73	0.74	0.72	0.73	1.08	ppb
Lithium	6-1				96		%
Magnesium	24-2	487.04	491.44	487.18	488.55	0.51	ppb
Manganese	55-2	1.09	1.14	1.15	1.13	2.53	ppb
Molybdenum	94-1	4.80	4.86	4.86	4.84	0.67	ppb
Molybdenum	95-1	3.99	4.05	4.09	4.04	1.23	ppb
Molybdenum	96-1	4.10	4.12	4.17	4.13	0.89	ppb
Molybdenum	97-1	4.06	4.01	3.94	4.00	1.45	ppb
Molybdenum	98-1	3.98	4.00	4.01	4.00	0.42	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.05	1.02	1.02	1.03	1.57	ppb
Phosphorus	31-2	-23.19	-20.79	-22.65	-22.21		ppb
Potassium	39-2	462.83	469.00	462.28	464.70	0.80	ppb
Rhodium	103-1				97		%
Rhodium	103-2				103		%
Scandium	45-1				93		%
Scandium	45-2				100		%
Selenium	82-1	4.48	4.42	4.66	4.52	2.76	ppb
Selenium	77-2	6.72	2.99	3.40	4.37	46.76	ppb
Selenium	78-2	5.76	4.00	3.16	4.31	30.77	ppb
Silicon	28-1	2.50	2.51	2.84	2.62	7.35	ppb
Silver	107-1	0.86	0.88	0.88	0.87	1.59	ppb
Silver	109-1	0.89	0.89	0.89	0.89	0.17	ppb
Sodium	23-2	447.13	449.91	444.96	447.33	0.55	ppb
Strontium	86-1	0.80	0.81	0.79	0.80	1.47	ppb
Strontium	88-1	0.80	0.80	0.80	0.80	0.51	ppb
Sulfur	34-1	-1256.56	-1240.89	-1170.46	-1222.64		ppb
Terbium	159-1				100		%
Terbium	159-2				104		%
Thallium	203-1	0.75	0.76	0.76	0.76	1.32	ppb
Thallium	205-1	0.75	0.76	0.76	0.76	0.73	ppb
Tin	118-1	-0.01	0.01	0.00	0.00		ppb
Titanium	47-1	4.32	4.24	4.22	4.26	1.24	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BSD Instrumnet Name : P7
Client Sample ID : LCSD456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:38:27 DataFile Name : 097LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.74	0.74	0.74	0.74	0.57	ppb
Vanadium	51-2	4.75	4.60	4.75	4.70	1.91	ppb
Yttrium	89-1				96		%
Yttrium	89-2				101		%
Zinc	66-2	4.59	4.66	4.74	4.66	1.54	ppb
Zirconium	90-1	0.87	0.90	0.93	0.90	3.81	ppb
Zirconium	91-1	0.89	0.95	0.94	0.93	3.68	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-01 Instrumnet Name : P7
Client Sample ID : MYE4K0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:41:45 DataFile Name : 098SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	36762.22	37127.30	37796.30	37228.60	1.41	ppb
Antimony	121-1	0.84	0.83	0.81	0.83	1.42	ppb
Arsenic	75-2	67.50	69.32	68.06	68.29	1.36	ppb
Barium	135-1	389.47	395.04	392.80	392.44	0.71	ppb
Barium	137-1	427.30	429.30	433.86	430.16	0.78	ppb
Beryllium	9-1	1.49	1.32	1.38	1.40	6.13	ppb
Bismuth	209-1				97		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.86	1.69	1.96	1.83	7.43	ppb
Cadmium	106-1	0.72	1.23	0.54	0.83	43.01	ppb
Cadmium	111-1	1.06	1.11	1.11	1.10	2.55	ppb
Calcium	43-1	16519.23	16585.74	16647.20	16584.06	0.39	ppb
Calcium	44-1	15568.98	15567.27	15539.63	15558.63	0.11	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	84.80	85.03	85.54	85.13	0.44	ppb
Cobalt	59-2	42.77	42.68	42.95	42.80	0.32	ppb
Copper	63-2	207.11	208.25	209.49	208.28	0.57	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				102		%
Indium	115-1				96		%
Indium	115-2				97		%
Iron	56-2	79182.06	80462.10	80733.20	80125.79	1.03	ppb
Iron	57-2	78564.95	79300.00	79458.19	79107.72	0.60	ppb
Iron	54-2	79942.70	79877.96	80906.51	80242.39	0.72	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-01 Instrumnet Name : P7
Client Sample ID : MYE4K0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:41:45 DataFile Name : 098SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	45.66	45.18	45.13	45.33	0.65	ppb
Lead	207-1	44.29	44.00	44.08	44.13	0.34	ppb
Lead	208-1	44.35	44.19	44.09	44.21	0.30	ppb
Lithium	6-1				93		%
Magnesium	24-2	18895.81	19115.41	19315.00	19108.74	1.10	ppb
Manganese	55-2	2039.53	2025.00	2052.60	2039.04	0.68	ppb
Molybdenum	94-1	12.29	12.65	12.49	12.47	1.46	ppb
Molybdenum	95-1	1.59	1.62	1.57	1.59	1.68	ppb
Molybdenum	96-1	3.32	3.38	3.35	3.35	0.91	ppb
Molybdenum	97-1	1.55	1.61	1.56	1.57	2.20	ppb
Molybdenum	98-1	1.51	1.57	1.53	1.54	1.94	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	68.25	68.29	68.40	68.31	0.11	ppb
Phosphorus	31-2	3040.44	3038.26	3078.80	3052.50	0.75	ppb
Potassium	39-2	12831.59	12976.72	13038.17	12948.83	0.82	ppb
Rhodium	103-1				92		%
Rhodium	103-2				98		%
Scandium	45-1				97		%
Scandium	45-2				103		%
Selenium	82-1	0.74	0.90	1.13	0.92	21.40	ppb
Selenium	77-2	18.49	18.55	18.11	18.38	1.29	ppb
Selenium	78-2	3.25	5.18	4.98	4.47	23.72	ppb
Silicon	28-1	1751.81	1753.45	1738.23	1747.83	0.48	ppb
Silver	107-1	0.50	0.46	0.46	0.48	5.32	ppb
Silver	109-1	0.51	0.48	0.46	0.49	4.76	ppb
Sodium	23-2	332.32	341.16	346.05	339.84	2.05	ppb
Strontium	86-1	114.63	117.15	116.01	115.93	1.09	ppb
Strontium	88-1	119.43	119.75	118.90	119.36	0.36	ppb
Sulfur	34-1	-1498.20	-1537.66	-1527.72	-1521.19		ppb
Terbium	159-1				100		%
Terbium	159-2				103		%
Thallium	203-1	0.33	0.33	0.34	0.33	1.23	ppb
Thallium	205-1	0.33	0.33	0.33	0.33	0.52	ppb
Tin	118-1	2.36	2.39	2.34	2.36	1.08	ppb
Titanium	47-1	889.06	879.89	884.23	884.39	0.52	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-01

Instrumnet Name :

P7

Client Sample ID :

MYE4K0

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 22:41:45

DataFile Name :

098SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.00	1.01	0.99	1.00	1.04	ppb
Vanadium	51-2	150.37	151.88	153.08	151.78	0.90	ppb
Yttrium	89-1				101		%
Yttrium	89-2				107		%
Zinc	66-2	246.82	244.80	250.99	247.54	1.27	ppb
Zirconium	90-1	6.75	6.85	6.80	6.80	0.77	ppb
Zirconium	91-1	6.63	6.69	6.76	6.69	0.98	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-02 Instrumnet Name : P7
Client Sample ID : MYE4K1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:44:49 DataFile Name : 099SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	27687.94	28326.93	27984.45	27999.77	1.14	ppb
Antimony	121-1	0.99	1.15	1.01	1.05	8.20	ppb
Arsenic	75-2	45.75	45.17	44.81	45.24	1.05	ppb
Barium	135-1	301.75	350.39	301.62	317.92	8.84	ppb
Barium	137-1	331.37	390.68	332.81	351.62	9.62	ppb
Beryllium	9-1	1.23	1.34	1.16	1.24	7.58	ppb
Bismuth	209-1				96		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.14	1.37	1.20	1.23	9.81	ppb
Cadmium	106-1	0.60	-0.25	0.23	0.19	220.93	ppb
Cadmium	111-1	0.36	0.35	0.33	0.35	3.44	ppb
Calcium	43-1	20418.87	24191.36	20631.24	21747.15	9.75	ppb
Calcium	44-1	19290.10	22957.96	19358.47	20535.51	10.22	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	80.25	81.84	80.34	80.81	1.11	ppb
Cobalt	59-2	38.05	38.27	37.62	37.98	0.87	ppb
Copper	63-2	90.49	91.22	90.73	90.82	0.41	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				97		%
Holmium	165-2				105		%
Indium	115-1				92		%
Indium	115-2				100		%
Iron	56-2	79336.30	80367.83	79948.06	79884.06	0.65	ppb
Iron	57-2	78761.06	77703.81	78803.58	78422.82	0.79	ppb
Iron	54-2	79401.30	80757.32	79420.31	79859.64	0.97	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-02 Instrumnet Name : P7
Client Sample ID : MYE4K1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:44:49 DataFile Name : 099SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	26.21	30.69	26.51	27.80	9.01	ppb
Lead	207-1	24.95	29.06	24.73	26.24	9.30	ppb
Lead	208-1	25.32	29.57	25.30	26.73	9.21	ppb
Lithium	6-1				89		%
Magnesium	24-2	15005.63	15377.92	15191.50	15191.68	1.23	ppb
Manganese	55-2	1545.48	1561.67	1555.81	1554.32	0.53	ppb
Molybdenum	94-1	16.88	20.03	17.54	18.15	9.16	ppb
Molybdenum	95-1	0.88	1.12	0.91	0.97	13.67	ppb
Molybdenum	96-1	3.22	3.79	3.24	3.42	9.51	ppb
Molybdenum	97-1	0.87	1.07	0.91	0.95	11.00	ppb
Molybdenum	98-1	0.85	0.99	0.85	0.90	8.80	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	63.66	64.46	63.68	63.93	0.71	ppb
Phosphorus	31-2	2376.96	2460.19	2392.06	2409.74	1.84	ppb
Potassium	39-2	6624.37	6642.20	6601.15	6622.57	0.31	ppb
Rhodium	103-1				90		%
Rhodium	103-2				101		%
Scandium	45-1				92		%
Scandium	45-2				104		%
Selenium	82-1	0.50	0.21	0.79	0.50	58.05	ppb
Selenium	77-2	15.08	15.67	17.69	16.15	8.49	ppb
Selenium	78-2	3.41	4.09	3.46	3.65	10.46	ppb
Silicon	28-1	1471.03	1745.05	1481.75	1565.94	9.91	ppb
Silver	107-1	0.11	0.15	0.12	0.13	17.07	ppb
Silver	109-1	0.11	0.14	0.13	0.13	13.20	ppb
Sodium	23-2	255.35	257.67	256.55	256.52	0.45	ppb
Strontium	86-1	71.43	83.25	72.72	75.80	8.56	ppb
Strontium	88-1	73.34	86.20	74.99	78.18	8.95	ppb
Sulfur	34-1	-1643.97	-1085.66	-1627.39	-1452.34		ppb
Terbium	159-1				98		%
Terbium	159-2				105		%
Thallium	203-1	0.22	0.26	0.23	0.24	9.71	ppb
Thallium	205-1	0.23	0.27	0.23	0.24	7.92	ppb
Tin	118-1	1.95	2.46	2.12	2.18	11.84	ppb
Titanium	47-1	1249.54	1468.04	1275.06	1330.88	8.98	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-02

Instrumnet Name :

P7

Client Sample ID :

MYE4K1

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 22:44:49

DataFile Name :

099SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.91	1.07	0.90	0.96	10.00	ppb
Vanadium	51-2	189.73	191.80	190.92	190.81	0.54	ppb
Yttrium	89-1				97		%
Yttrium	89-2				107		%
Zinc	66-2	147.17	146.34	145.74	146.42	0.49	ppb
Zirconium	90-1	9.72	11.51	9.91	10.38	9.49	ppb
Zirconium	91-1	9.71	11.37	9.88	10.32	8.86	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03 Instrumnet Name : P7
Client Sample ID : MYE4K2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:47:57 DataFile Name : 100SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	28718.50	30130.32	29218.85	29355.89	2.44	ppb
Antimony	121-1	0.43	0.43	0.44	0.43	1.10	ppb
Arsenic	75-2	36.98	39.15	38.46	38.19	2.90	ppb
Barium	135-1	300.26	294.86	309.78	301.63	2.50	ppb
Barium	137-1	328.48	323.32	329.11	326.97	0.97	ppb
Beryllium	9-1	1.06	0.98	0.92	0.99	7.00	ppb
Bismuth	209-1				102		%
Bismuth	209-2				99		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.87	0.92	0.97	0.92	5.19	ppb
Cadmium	106-1	-0.63	-0.79	-0.07	-0.50		ppb
Cadmium	111-1	0.46	0.46	0.50	0.48	4.47	ppb
Calcium	43-1	24575.27	24865.72	24892.33	24777.77	0.71	ppb
Calcium	44-1	23127.57	23314.05	23373.94	23271.85	0.55	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	64.25	68.06	64.77	65.69	3.15	ppb
Cobalt	59-2	36.95	38.78	37.39	37.71	2.52	ppb
Copper	63-2	112.11	119.27	114.83	115.40	3.13	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				103		%
Indium	115-1				99		%
Indium	115-2				99		%
Iron	56-2	65134.67	69565.46	66306.55	67002.23	3.43	ppb
Iron	57-2	64192.98	68312.06	65857.06	66120.70	3.13	ppb
Iron	54-2	65938.91	68913.68	67089.79	67314.12	2.23	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03 Instrumnet Name : P7
Client Sample ID : MYE4K2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:47:57 DataFile Name : 100SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	29.56	29.82	30.38	29.92	1.41	ppb
Lead	207-1	27.87	28.99	28.30	28.39	1.99	ppb
Lead	208-1	28.60	29.04	29.15	28.93	1.00	ppb
Lithium	6-1				91		%
Magnesium	24-2	15653.87	16398.79	15980.82	16011.16	2.33	ppb
Manganese	55-2	1974.76	2076.66	2021.90	2024.44	2.52	ppb
Molybdenum	94-1	8.24	8.48	8.44	8.39	1.50	ppb
Molybdenum	95-1	0.93	0.92	0.86	0.90	4.36	ppb
Molybdenum	96-1	2.19	2.17	2.11	2.16	1.98	ppb
Molybdenum	97-1	0.90	0.88	0.87	0.89	1.50	ppb
Molybdenum	98-1	0.83	0.83	0.83	0.83	0.36	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	52.31	55.73	53.48	53.84	3.22	ppb
Phosphorus	31-2	2282.62	2398.33	2290.75	2323.90	2.78	ppb
Potassium	39-2	9850.57	10427.46	10048.49	10108.84	2.90	ppb
Rhodium	103-1				96		%
Rhodium	103-2				100		%
Scandium	45-1				97		%
Scandium	45-2				104		%
Selenium	82-1	0.59	0.51	0.22	0.44	44.64	ppb
Selenium	77-2	14.25	13.90	13.75	13.96	1.83	ppb
Selenium	78-2	3.94	5.16	3.71	4.27	18.30	ppb
Silicon	28-1	2034.76	2042.40	2064.64	2047.26	0.76	ppb
Silver	107-1	0.19	0.22	0.22	0.21	8.47	ppb
Silver	109-1	0.20	0.21	0.23	0.21	6.93	ppb
Sodium	23-2	393.50	419.25	403.28	405.34	3.21	ppb
Strontium	86-1	93.36	95.32	94.06	94.25	1.05	ppb
Strontium	88-1	96.38	97.49	97.15	97.01	0.59	ppb
Sulfur	34-1	-1399.33	-1405.61	-1404.27	-1403.07		ppb
Terbium	159-1				104		%
Terbium	159-2				103		%
Thallium	203-1	0.23	0.24	0.23	0.23	1.95	ppb
Thallium	205-1	0.22	0.22	0.24	0.23	3.91	ppb
Tin	118-1	1.83	1.80	1.84	1.82	0.98	ppb
Titanium	47-1	892.78	901.91	914.06	902.91	1.18	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-03

Instrumnet Name :

P7

Client Sample ID :

MYE4K2

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 22:47:57

DataFile Name :

100SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.93	0.92	0.95	0.93	2.01	ppb
Vanadium	51-2	126.83	135.32	129.63	130.59	3.32	ppb
Yttrium	89-1				104		%
Yttrium	89-2				106		%
Zinc	66-2	158.48	168.79	158.47	161.91	3.68	ppb
Zirconium	90-1	4.58	4.60	4.61	4.60	0.32	ppb
Zirconium	91-1	4.62	4.66	4.73	4.67	1.15	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-04 Instrumnet Name : P7
Client Sample ID : MYE4K2D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:51:02 DataFile Name : 101SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	29812.53	28620.41	29268.20	29233.71	2.04	ppb
Antimony	121-1	0.45	0.43	0.43	0.44	3.09	ppb
Arsenic	75-2	38.29	37.60	37.70	37.86	0.99	ppb
Barium	135-1	306.55	301.65	301.82	303.34	0.92	ppb
Barium	137-1	336.75	329.18	331.61	332.51	1.16	ppb
Beryllium	9-1	1.13	1.07	1.10	1.10	2.57	ppb
Bismuth	209-1				101		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.27	1.21	0.92	1.13	16.29	ppb
Cadmium	106-1	0.33	-0.98	-0.17	-0.27		ppb
Cadmium	111-1	0.53	0.45	0.48	0.49	8.85	ppb
Calcium	43-1	24281.99	24710.55	25227.26	24739.93	1.91	ppb
Calcium	44-1	23070.42	23305.99	23760.75	23379.05	1.50	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	67.71	64.63	65.55	65.96	2.40	ppb
Cobalt	59-2	38.31	36.75	37.35	37.47	2.11	ppb
Copper	63-2	118.32	112.79	114.72	115.28	2.44	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				104		%
Indium	115-1				98		%
Indium	115-2				101		%
Iron	56-2	68650.46	65090.17	65958.26	66566.30	2.79	ppb
Iron	57-2	67168.23	65218.70	65670.01	66018.98	1.55	ppb
Iron	54-2	68666.31	65612.12	66395.18	66891.20	2.37	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-04 Instrumnet Name : P7
Client Sample ID : MYE4K2D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:51:02 DataFile Name : 101SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	29.81	30.43	30.61	30.28	1.38	ppb
Lead	207-1	28.52	29.37	29.58	29.15	1.92	ppb
Lead	208-1	29.11	29.63	30.00	29.58	1.52	ppb
Lithium	6-1				90		%
Magnesium	24-2	16395.68	15538.96	15995.59	15976.75	2.68	ppb
Manganese	55-2	2055.09	1962.32	1985.51	2000.97	2.41	ppb
Molybdenum	94-1	8.50	8.56	8.68	8.58	1.02	ppb
Molybdenum	95-1	0.94	0.91	0.89	0.91	3.14	ppb
Molybdenum	96-1	2.27	2.22	2.30	2.26	1.90	ppb
Molybdenum	97-1	0.81	0.86	0.92	0.87	6.51	ppb
Molybdenum	98-1	0.82	0.84	0.84	0.83	1.67	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	54.59	52.85	53.87	53.77	1.62	ppb
Phosphorus	31-2	2403.36	2284.54	2314.30	2334.07	2.65	ppb
Potassium	39-2	10286.56	9782.65	9915.33	9994.85	2.61	ppb
Rhodium	103-1				96		%
Rhodium	103-2				102		%
Scandium	45-1				98		%
Scandium	45-2				105		%
Selenium	82-1	0.84	0.53	0.45	0.61	34.52	ppb
Selenium	77-2	11.99	11.61	13.88	12.49	9.73	ppb
Selenium	78-2	3.31	1.68	3.59	2.86	36.10	ppb
Silicon	28-1	2034.50	2046.05	2066.75	2049.10	0.80	ppb
Silver	107-1	0.20	0.21	0.21	0.21	4.09	ppb
Silver	109-1	0.20	0.21	0.23	0.22	7.05	ppb
Sodium	23-2	409.91	393.27	406.62	403.27	2.18	ppb
Strontium	86-1	95.34	94.35	97.00	95.56	1.40	ppb
Strontium	88-1	98.05	98.16	99.27	98.49	0.69	ppb
Sulfur	34-1	-1478.56	-1489.87	-1413.69	-1460.71		ppb
Terbium	159-1				104		%
Terbium	159-2				105		%
Thallium	203-1	0.23	0.24	0.25	0.24	2.50	ppb
Thallium	205-1	0.23	0.23	0.23	0.23	1.74	ppb
Tin	118-1	1.81	1.82	1.85	1.83	1.31	ppb
Titanium	47-1	905.09	904.77	928.64	912.83	1.50	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-04

Instrumnet Name :

P7

Client Sample ID :

MYE4K2D

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 22:51:02

DataFile Name :

101SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.96	0.97	0.99	0.98	1.53	ppb
Vanadium	51-2	132.60	127.40	129.25	129.75	2.03	ppb
Yttrium	89-1				104		%
Yttrium	89-2				108		%
Zinc	66-2	165.94	157.86	161.43	161.74	2.50	ppb
Zirconium	90-1	4.69	4.71	4.69	4.70	0.28	ppb
Zirconium	91-1	4.75	4.66	4.79	4.73	1.36	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03LX5 Instrumnet Name : P7
Client Sample ID : MYE4K2L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 22:54:10 DataFile Name : 102SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	5836.31	5812.88	5949.04	5866.08	1.24	ppb
Antimony	121-1	0.07	0.08	0.08	0.08	6.79	ppb
Arsenic	75-2	7.24	8.10	7.75	7.70	5.58	ppb
Barium	135-1	58.45	65.10	60.11	61.22	5.66	ppb
Barium	137-1	58.23	65.64	59.69	61.18	6.42	ppb
Beryllium	9-1	0.24	0.28	0.19	0.23	20.14	ppb
Bismuth	209-1				99		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.28	0.16	0.34	0.26	36.71	ppb
Cadmium	106-1	-2.01	-1.99	-2.10	-2.03		ppb
Cadmium	111-1	-0.04	-0.05	-0.07	-0.05		ppb
Calcium	43-1	4933.06	5481.30	5068.38	5160.91	5.53	ppb
Calcium	44-1	4635.37	5185.61	4747.58	4856.19	5.99	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	12.89	12.76	12.85	12.83	0.50	ppb
Cobalt	59-2	7.46	7.36	7.36	7.40	0.80	ppb
Copper	63-2	22.99	22.91	23.00	22.97	0.21	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				98		%
Holmium	165-2				106		%
Indium	115-1				95		%
Indium	115-2				103		%
Iron	56-2	13212.46	13043.79	12960.91	13072.39	0.98	ppb
Iron	57-2	13389.15	13290.05	13387.27	13355.49	0.42	ppb
Iron	54-2	13318.49	13387.19	13124.98	13276.89	1.02	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03LX5 Instrumnet Name : P7
Client Sample ID : MYE4K2L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 22:54:10 DataFile Name : 102SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	5.69	6.33	5.95	5.99	5.44	ppb
Lead	207-1	5.39	5.99	5.52	5.63	5.58	ppb
Lead	208-1	5.50	6.06	5.68	5.75	4.92	ppb
Lithium	6-1				87		%
Magnesium	24-2	3155.09	3145.15	3154.96	3151.73	0.18	ppb
Manganese	55-2	395.71	392.84	394.96	394.51	0.38	ppb
Molybdenum	94-1	1.62	1.79	1.63	1.68	5.82	ppb
Molybdenum	95-1	0.13	0.16	0.14	0.14	11.64	ppb
Molybdenum	96-1	0.38	0.44	0.42	0.41	7.97	ppb
Molybdenum	97-1	0.13	0.19	0.12	0.15	26.49	ppb
Molybdenum	98-1	0.12	0.14	0.11	0.12	12.95	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	10.72	10.39	10.73	10.61	1.79	ppb
Phosphorus	31-2	459.60	439.99	471.14	456.91	3.45	ppb
Potassium	39-2	2080.57	2074.88	2076.65	2077.37	0.14	ppb
Rhodium	103-1				92		%
Rhodium	103-2				103		%
Scandium	45-1				89		%
Scandium	45-2				99		%
Selenium	82-1	0.30	0.06	-0.38	-0.01		ppb
Selenium	77-2	4.07	2.54	2.25	2.95	33.06	ppb
Selenium	78-2	0.04	-0.20	-0.05	-0.07		ppb
Silicon	28-1	409.37	462.72	423.25	431.78	6.41	ppb
Silver	107-1	0.03	0.04	0.04	0.04	5.18	ppb
Silver	109-1	0.04	0.04	0.04	0.04	2.63	ppb
Sodium	23-2	70.23	70.97	71.40	70.87	0.83	ppb
Strontium	86-1	18.96	20.98	19.25	19.73	5.52	ppb
Strontium	88-1	18.76	21.04	19.26	19.69	6.07	ppb
Sulfur	34-1	-1788.24	-1445.85	-1655.03	-1629.71		ppb
Terbium	159-1				98		%
Terbium	159-2				105		%
Thallium	203-1	0.04	0.05	0.04	0.04	13.32	ppb
Thallium	205-1	0.04	0.04	0.04	0.04	3.07	ppb
Tin	118-1	0.46	0.38	0.35	0.40	13.88	ppb
Titanium	47-1	179.90	203.53	185.49	189.64	6.51	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-03LX5

Instrumnet Name :

P7

Client Sample ID :

MYE4K2L

Dilution Factor :

5

Date & Time Acquired :

2024-11-02 22:54:10

DataFile Name :

102SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.18	0.20	0.18	0.19	7.11	ppb
Vanadium	51-2	25.24	25.25	25.56	25.35	0.72	ppb
Yttrium	89-1				94		%
Yttrium	89-2				102		%
Zinc	66-2	32.42	31.88	32.50	32.27	1.06	ppb
Zirconium	90-1	0.89	1.00	0.93	0.94	5.77	ppb
Zirconium	91-1	0.90	1.08	0.93	0.97	9.95	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-05 Instrumnet Name : P7
Client Sample ID : MYE4K2S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:57:20 DataFile Name : 103SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	29909.58	29983.02	30002.69	29965.10	0.16	ppb
Antimony	121-1	55.87	55.42	56.52	55.94	0.99	ppb
Arsenic	75-2	59.17	58.43	59.77	59.12	1.14	ppb
Barium	135-1	1496.94	1456.79	1467.71	1473.82	1.41	ppb
Barium	137-1	1491.93	1450.75	1467.20	1469.96	1.41	ppb
Beryllium	9-1	30.25	31.39	32.22	31.28	3.16	ppb
Bismuth	209-1				101		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	27.15	26.90	26.66	26.90	0.91	ppb
Cadmium	106-1	25.53	24.61	25.75	25.30	2.40	ppb
Cadmium	111-1	28.85	28.00	28.63	28.49	1.56	ppb
Calcium	43-1	24305.63	24608.06	24915.24	24609.64	1.24	ppb
Calcium	44-1	22721.07	23119.54	23258.76	23033.12	1.21	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	181.06	182.08	181.63	181.59	0.28	ppb
Cobalt	59-2	332.41	332.59	332.27	332.42	0.05	ppb
Copper	63-2	246.92	247.42	247.10	247.15	0.10	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				105		%
Indium	115-1				97		%
Indium	115-2				100		%
Iron	56-2	65814.06	66330.72	66357.65	66167.48	0.46	ppb
Iron	57-2	66008.18	65430.88	65897.11	65778.72	0.47	ppb
Iron	54-2	66689.56	67183.69	66785.29	66886.18	0.39	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-05 Instrumnet Name : P7
Client Sample ID : MYE4K2S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:57:20 DataFile Name : 103SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	39.41	40.40	40.08	39.97	1.26	ppb
Lead	207-1	38.05	38.16	39.29	38.50	1.78	ppb
Lead	208-1	38.26	38.89	39.11	38.76	1.14	ppb
Lithium	6-1				91		%
Magnesium	24-2	15476.87	15632.25	15726.12	15611.75	0.81	ppb
Manganese	55-2	2247.31	2272.35	2247.03	2255.56	0.64	ppb
Molybdenum	94-1	8.42	8.41	8.55	8.46	0.95	ppb
Molybdenum	95-1	0.89	0.90	0.91	0.90	1.17	ppb
Molybdenum	96-1	2.19	2.21	2.24	2.21	1.23	ppb
Molybdenum	97-1	0.85	0.92	0.85	0.87	4.70	ppb
Molybdenum	98-1	0.85	0.80	0.82	0.82	2.95	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	335.53	335.48	336.19	335.73	0.12	ppb
Phosphorus	31-2	2273.75	2339.04	2334.09	2315.63	1.57	ppb
Potassium	39-2	9682.60	9963.04	9691.26	9778.97	1.63	ppb
Rhodium	103-1				94		%
Rhodium	103-2				101		%
Scandium	45-1				95		%
Scandium	45-2				102		%
Selenium	82-1	54.56	53.95	56.14	54.88	2.06	ppb
Selenium	77-2	64.74	66.65	70.26	67.22	4.17	ppb
Selenium	78-2	54.26	55.80	59.59	56.55	4.85	ppb
Silicon	28-1	2020.77	2036.41	2065.23	2040.80	1.10	ppb
Silver	107-1	24.18	24.12	24.67	24.32	1.24	ppb
Silver	109-1	24.18	24.26	24.71	24.38	1.17	ppb
Sodium	23-2	393.76	396.59	396.76	395.71	0.43	ppb
Strontium	86-1	92.89	93.89	94.26	93.68	0.76	ppb
Strontium	88-1	94.27	96.22	97.55	96.02	1.72	ppb
Sulfur	34-1	-1409.49	-1454.18	-1468.13	-1443.94		ppb
Terbium	159-1				103		%
Terbium	159-2				105		%
Thallium	203-1	24.88	25.13	25.64	25.22	1.55	ppb
Thallium	205-1	24.62	24.96	25.24	24.94	1.24	ppb
Tin	118-1	1.81	1.80	1.83	1.82	0.77	ppb
Titanium	47-1	896.90	897.93	900.13	898.32	0.18	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-05

Instrumnet Name :

P7

Client Sample ID :

MYE4K2S

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 22:57:20

DataFile Name :

103SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.92	0.95	0.96	0.94	1.97	ppb
Vanadium	51-2	411.12	413.50	417.28	413.97	0.75	ppb
Yttrium	89-1				101		%
Yttrium	89-2				106		%
Zinc	66-2	437.04	437.53	435.68	436.75	0.22	ppb
Zirconium	90-1	4.54	4.70	4.64	4.63	1.77	ppb
Zirconium	91-1	4.61	4.60	4.63	4.61	0.37	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03MS Instrumnet Name : P7
Client Sample ID : MYE4K2SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:00:23 DataFile Name : 104SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	33091.60	33359.27	33218.51	33223.13	0.40	ppb
Antimony	121-1	281.59	284.36	280.53	282.16	0.70	ppb
Arsenic	75-2	140.66	145.89	140.84	142.46	2.09	ppb
Barium	135-1	5845.21	5950.42	5817.61	5871.08	1.19	ppb
Barium	137-1	5835.16	6018.47	5834.39	5896.01	1.80	ppb
Beryllium	9-1	142.52	145.00	146.47	144.66	1.38	ppb
Bismuth	209-1				102		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	125.24	127.01	127.72	126.66	1.01	ppb
Cadmium	106-1	119.76	121.87	120.70	120.77	0.88	ppb
Cadmium	111-1	133.13	136.93	136.09	135.38	1.47	ppb
Calcium	43-1	23399.86	23992.52	24163.58	23851.99	1.68	ppb
Calcium	44-1	21992.27	22638.80	22338.44	22323.17	1.45	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	631.56	634.23	629.78	631.86	0.35	ppb
Cobalt	59-2	1474.19	1458.25	1461.79	1464.75	0.57	ppb
Copper	63-2	796.42	793.59	801.18	797.06	0.48	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				104		%
Indium	115-1				96		%
Indium	115-2				98		%
Iron	56-2	67213.42	67213.88	67334.01	67253.77	0.10	ppb
Iron	57-2	66471.37	65889.64	66518.93	66293.31	0.53	ppb
Iron	54-2	67449.76	67852.61	67179.85	67494.07	0.50	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03MS Instrumnet Name : P7
Client Sample ID : MYE4K2SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:00:23 DataFile Name : 104SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	75.57	77.51	78.13	77.07	1.73	ppb
Lead	207-1	74.10	75.15	75.63	74.96	1.04	ppb
Lead	208-1	74.70	76.06	76.79	75.85	1.40	ppb
Lithium	6-1				91		%
Magnesium	24-2	15336.47	15277.24	15312.01	15308.57	0.19	ppb
Manganese	55-2	3288.87	3290.71	3318.21	3299.26	0.50	ppb
Molybdenum	94-1	8.19	8.34	8.32	8.28	1.02	ppb
Molybdenum	95-1	0.92	0.89	0.94	0.92	2.84	ppb
Molybdenum	96-1	2.20	2.18	2.21	2.20	0.63	ppb
Molybdenum	97-1	0.84	0.87	0.86	0.85	1.96	ppb
Molybdenum	98-1	0.83	0.87	0.87	0.85	2.77	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1510.18	1514.88	1499.66	1508.24	0.52	ppb
Phosphorus	31-2	2195.74	2248.29	2208.36	2217.47	1.24	ppb
Potassium	39-2	9540.11	9534.70	9481.49	9518.77	0.34	ppb
Rhodium	103-1				93		%
Rhodium	103-2				99		%
Scandium	45-1				93		%
Scandium	45-2				100		%
Selenium	82-1	296.41	301.98	301.95	300.11	1.07	ppb
Selenium	77-2	304.97	323.13	312.47	313.52	2.91	ppb
Selenium	78-2	295.11	300.11	303.28	299.50	1.38	ppb
Silicon	28-1	1952.89	2014.57	2006.50	1991.32	1.68	ppb
Silver	107-1	114.58	123.71	124.36	120.88	4.53	ppb
Silver	109-1	114.73	117.71	115.71	116.05	1.31	ppb
Sodium	23-2	382.06	383.93	386.53	384.17	0.58	ppb
Strontium	86-1	89.28	92.20	92.13	91.20	1.83	ppb
Strontium	88-1	93.89	94.69	93.96	94.18	0.47	ppb
Sulfur	34-1	-1498.81	-1500.90	-1521.17	-1506.96		ppb
Terbium	159-1				102		%
Terbium	159-2				105		%
Thallium	203-1	117.74	119.69	121.51	119.65	1.58	ppb
Thallium	205-1	135.89	138.32	139.25	137.82	1.26	ppb
Tin	118-1	1.72	1.77	1.75	1.75	1.42	ppb
Titanium	47-1	861.98	885.31	888.98	878.75	1.67	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-03MS

Instrumnet Name :

P7

Client Sample ID :

MYE4K2SRE

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:00:23

DataFile Name :

104SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.90	0.94	0.95	0.93	2.83	ppb
Vanadium	51-2	1510.14	1500.50	1499.31	1503.32	0.39	ppb
Yttrium	89-1				100		%
Yttrium	89-2				104		%
Zinc	66-2	1509.14	1513.15	1505.44	1509.25	0.26	ppb
Zirconium	90-1	4.42	4.52	4.55	4.50	1.48	ppb
Zirconium	91-1	4.58	4.53	4.77	4.63	2.68	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-06 Instrumnet Name : P7
Client Sample ID : MYE4K3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:03:19 DataFile Name : 105SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	25587.60	26036.11	25740.60	25788.10	0.88	ppb
Antimony	121-1	0.94	0.92	0.91	0.92	1.48	ppb
Arsenic	75-2	50.07	49.04	48.99	49.37	1.24	ppb
Barium	135-1	290.59	288.96	294.82	291.46	1.04	ppb
Barium	137-1	290.91	290.14	294.20	291.75	0.74	ppb
Beryllium	9-1	1.00	1.02	1.12	1.05	6.08	ppb
Bismuth	209-1				101		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.53	1.33	1.42	1.43	7.30	ppb
Cadmium	106-1	0.35	-0.20	0.41	0.19	178.65	ppb
Cadmium	111-1	0.78	0.78	0.81	0.79	2.80	ppb
Calcium	43-1	20361.84	20484.37	21179.31	20675.17	2.13	ppb
Calcium	44-1	19257.13	19223.31	19724.46	19401.63	1.44	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	59.28	59.59	58.89	59.25	0.59	ppb
Cobalt	59-2	33.54	33.61	33.47	33.54	0.20	ppb
Copper	63-2	152.59	153.32	152.84	152.92	0.24	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				105		%
Indium	115-1				96		%
Indium	115-2				100		%
Iron	56-2	61064.10	62005.52	60888.86	61319.49	0.98	ppb
Iron	57-2	61079.17	60745.77	60755.17	60860.04	0.31	ppb
Iron	54-2	61794.39	62039.95	61305.64	61713.33	0.61	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-06 Instrumnet Name : P7
Client Sample ID : MYE4K3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:03:19 DataFile Name : 105SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	36.56	36.16	37.47	36.73	1.82	ppb
Lead	207-1	35.35	35.29	37.08	35.91	2.83	ppb
Lead	208-1	35.69	35.42	36.74	35.95	1.95	ppb
Lithium	6-1				90		%
Magnesium	24-2	15130.66	15150.19	15163.69	15148.18	0.11	ppb
Manganese	55-2	1541.98	1579.31	1546.96	1556.08	1.30	ppb
Molybdenum	94-1	9.72	9.88	9.95	9.85	1.16	ppb
Molybdenum	95-1	1.22	1.19	1.21	1.20	1.16	ppb
Molybdenum	96-1	2.52	2.62	2.60	2.58	2.08	ppb
Molybdenum	97-1	1.19	1.18	1.19	1.19	0.36	ppb
Molybdenum	98-1	1.13	1.10	1.12	1.12	1.34	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	57.79	58.59	57.76	58.05	0.81	ppb
Phosphorus	31-2	2801.37	2839.32	2820.68	2820.46	0.67	ppb
Potassium	39-2	9542.75	9711.44	9472.57	9575.59	1.28	ppb
Rhodium	103-1				92		%
Rhodium	103-2				101		%
Scandium	45-1				92		%
Scandium	45-2				102		%
Selenium	82-1	0.56	0.79	0.82	0.72	20.27	ppb
Selenium	77-2	15.04	12.53	12.95	13.51	9.95	ppb
Selenium	78-2	3.16	3.04	2.30	2.84	16.53	ppb
Silicon	28-1	1989.29	1961.08	2015.73	1988.70	1.37	ppb
Silver	107-1	0.47	0.42	0.39	0.43	9.32	ppb
Silver	109-1	0.45	0.38	0.38	0.40	9.46	ppb
Sodium	23-2	617.24	626.55	625.15	622.98	0.81	ppb
Strontium	86-1	105.36	106.13	107.71	106.40	1.13	ppb
Strontium	88-1	109.00	109.61	110.00	109.53	0.46	ppb
Sulfur	34-1	-1093.38	-1133.25	-1020.06	-1082.23		ppb
Terbium	159-1				102		%
Terbium	159-2				106		%
Thallium	203-1	0.22	0.22	0.24	0.23	3.29	ppb
Thallium	205-1	0.23	0.23	0.23	0.23	1.53	ppb
Tin	118-1	2.53	2.52	2.55	2.53	0.73	ppb
Titanium	47-1	769.27	769.46	795.63	778.12	1.95	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-06

Instrumnet Name :

P7

Client Sample ID :

MYE4K3

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:03:19

DataFile Name :

105SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.74	0.73	0.78	0.75	3.28	ppb
Vanadium	51-2	119.54	120.94	119.27	119.92	0.75	ppb
Yttrium	89-1				99		%
Yttrium	89-2				106		%
Zinc	66-2	181.28	184.08	183.45	182.94	0.80	ppb
Zirconium	90-1	5.25	5.34	5.37	5.32	1.18	ppb
Zirconium	91-1	5.38	5.27	5.45	5.36	1.69	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-07 Instrumnet Name : P7
Client Sample ID : MYE4K4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:06:28 DataFile Name : 106SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	31759.07	31560.14	31738.46	31685.89	0.35	ppb
Antimony	121-1	0.52	0.52	0.53	0.52	1.44	ppb
Arsenic	75-2	43.20	42.69	43.91	43.27	1.42	ppb
Barium	135-1	304.16	316.66	308.61	309.81	2.05	ppb
Barium	137-1	336.28	314.74	333.21	328.08	3.55	ppb
Beryllium	9-1	1.20	1.24	1.10	1.18	6.34	ppb
Bismuth	209-1				101		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.14	1.18	1.06	1.13	5.44	ppb
Cadmium	106-1	0.17	0.58	-0.40	0.12	413.79	ppb
Cadmium	111-1	0.70	0.78	0.68	0.72	7.49	ppb
Calcium	43-1	19700.75	20234.08	19736.04	19890.29	1.50	ppb
Calcium	44-1	18373.62	19019.37	18635.47	18676.15	1.74	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	68.78	68.70	68.87	68.79	0.12	ppb
Cobalt	59-2	36.63	36.48	36.90	36.67	0.58	ppb
Copper	63-2	147.62	149.32	148.77	148.57	0.59	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				104		%
Indium	115-1				95		%
Indium	115-2				98		%
Iron	56-2	66872.39	65729.92	66722.27	66441.53	0.93	ppb
Iron	57-2	65784.98	65741.83	65924.89	65817.23	0.15	ppb
Iron	54-2	66794.16	66168.90	67112.49	66691.85	0.72	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-07 Instrumnet Name : P7
Client Sample ID : MYE4K4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:06:28 DataFile Name : 106SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	32.41	33.62	33.04	33.02	1.83	ppb
Lead	207-1	30.98	32.77	31.70	31.81	2.83	ppb
Lead	208-1	31.57	32.92	32.16	32.21	2.10	ppb
Lithium	6-1				89		%
Magnesium	24-2	15955.85	15915.09	16143.03	16004.66	0.76	ppb
Manganese	55-2	1850.94	1837.63	1851.58	1846.72	0.43	ppb
Molybdenum	94-1	9.36	9.72	9.42	9.50	2.04	ppb
Molybdenum	95-1	1.05	1.05	1.05	1.05	0.19	ppb
Molybdenum	96-1	2.44	2.46	2.53	2.47	1.88	ppb
Molybdenum	97-1	1.08	1.02	1.00	1.03	3.69	ppb
Molybdenum	98-1	0.97	1.00	0.99	0.99	1.52	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	54.66	54.67	55.39	54.91	0.76	ppb
Phosphorus	31-2	2338.14	2267.52	2252.17	2285.94	2.01	ppb
Potassium	39-2	10476.86	10456.58	10353.16	10428.87	0.64	ppb
Rhodium	103-1				92		%
Rhodium	103-2				100		%
Scandium	45-1				93		%
Scandium	45-2				101		%
Selenium	82-1	0.70	0.57	0.47	0.58	19.41	ppb
Selenium	77-2	14.27	12.13	14.59	13.66	9.80	ppb
Selenium	78-2	3.10	1.86	3.57	2.84	31.05	ppb
Silicon	28-1	1989.21	2051.67	2005.10	2015.32	1.61	ppb
Silver	107-1	0.29	0.33	0.29	0.31	7.38	ppb
Silver	109-1	0.32	0.31	0.30	0.31	2.03	ppb
Sodium	23-2	316.29	314.74	317.33	316.12	0.41	ppb
Strontium	86-1	87.18	89.08	88.06	88.11	1.08	ppb
Strontium	88-1	90.51	92.87	90.20	91.19	1.60	ppb
Sulfur	34-1	-1411.26	-1394.11	-1503.99	-1436.45		ppb
Terbium	159-1				101		%
Terbium	159-2				105		%
Thallium	203-1	0.25	0.25	0.24	0.25	1.99	ppb
Thallium	205-1	0.24	0.26	0.25	0.25	2.89	ppb
Tin	118-1	1.96	2.01	2.03	2.00	1.82	ppb
Titanium	47-1	879.47	898.78	888.64	888.96	1.09	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-07

Instrumnet Name :

P7

Client Sample ID :

MYE4K4

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:06:28

DataFile Name :

106SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.84	0.88	0.88	0.87	2.56	ppb
Vanadium	51-2	130.25	129.62	129.37	129.75	0.35	ppb
Yttrium	89-1				100		%
Yttrium	89-2				105		%
Zinc	66-2	185.19	185.07	185.60	185.29	0.15	ppb
Zirconium	90-1	5.08	5.18	5.15	5.14	0.98	ppb
Zirconium	91-1	5.10	5.21	5.30	5.21	1.94	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-08 Instrumnet Name : P7
Client Sample ID : MYE4K5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:09:35 DataFile Name : 107SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	25953.29	26417.43	26189.10	26186.61	0.89	ppb
Antimony	121-1	0.60	0.61	0.63	0.61	2.30	ppb
Arsenic	75-2	48.42	48.28	48.94	48.54	0.72	ppb
Barium	135-1	268.54	264.77	273.38	268.90	1.60	ppb
Barium	137-1	269.27	266.78	274.88	270.31	1.54	ppb
Beryllium	9-1	1.10	1.08	0.80	1.00	16.84	ppb
Bismuth	209-1				101		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.06	1.22	1.13	1.14	6.68	ppb
Cadmium	106-1	0.37	0.08	-0.55	-0.04		ppb
Cadmium	111-1	0.77	0.69	0.71	0.72	5.56	ppb
Calcium	43-1	18035.42	18072.02	18575.70	18227.71	1.66	ppb
Calcium	44-1	17014.21	17000.98	17392.17	17135.79	1.30	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	63.37	63.16	63.58	63.37	0.33	ppb
Cobalt	59-2	32.17	32.06	32.14	32.12	0.18	ppb
Copper	63-2	131.58	131.26	131.27	131.37	0.14	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				103		%
Indium	115-1				95		%
Indium	115-2				98		%
Iron	56-2	56712.19	56532.98	57062.44	56769.20	0.47	ppb
Iron	57-2	56759.66	55985.29	56479.35	56408.10	0.70	ppb
Iron	54-2	56687.93	56440.89	56902.32	56677.05	0.41	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-08 Instrumnet Name : P7
Client Sample ID : MYE4K5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:09:35 DataFile Name : 107SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	33.75	34.32	35.06	34.38	1.91	ppb
Lead	207-1	32.88	33.23	33.40	33.17	0.80	ppb
Lead	208-1	33.00	33.32	33.94	33.42	1.43	ppb
Lithium	6-1				90		%
Magnesium	24-2	15002.27	15125.26	15144.42	15090.65	0.51	ppb
Manganese	55-2	1417.22	1412.45	1417.72	1415.80	0.21	ppb
Molybdenum	94-1	8.08	8.22	8.16	8.15	0.88	ppb
Molybdenum	95-1	1.43	1.44	1.40	1.43	1.36	ppb
Molybdenum	96-1	2.60	2.58	2.57	2.58	0.52	ppb
Molybdenum	97-1	1.34	1.38	1.42	1.38	2.99	ppb
Molybdenum	98-1	1.37	1.41	1.35	1.38	2.03	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	54.88	55.38	55.88	55.38	0.90	ppb
Phosphorus	31-2	2451.59	2485.97	2440.85	2459.47	0.96	ppb
Potassium	39-2	10477.42	10500.12	10434.62	10470.72	0.32	ppb
Rhodium	103-1				93		%
Rhodium	103-2				99		%
Scandium	45-1				93		%
Scandium	45-2				100		%
Selenium	82-1	1.27	0.67	0.56	0.83	46.08	ppb
Selenium	77-2	13.33	12.19	12.52	12.68	4.62	ppb
Selenium	78-2	2.46	2.30	2.03	2.26	9.59	ppb
Silicon	28-1	2072.73	2060.67	2092.47	2075.29	0.77	ppb
Silver	107-1	0.29	0.29	0.29	0.29	1.75	ppb
Silver	109-1	0.27	0.28	0.30	0.28	5.36	ppb
Sodium	23-2	634.38	637.45	638.56	636.80	0.34	ppb
Strontium	86-1	86.39	86.33	87.03	86.59	0.45	ppb
Strontium	88-1	90.15	89.59	90.90	90.21	0.73	ppb
Sulfur	34-1	-1051.75	-1122.01	-1087.87	-1087.21		ppb
Terbium	159-1				102		%
Terbium	159-2				104		%
Thallium	203-1	0.19	0.20	0.21	0.20	5.25	ppb
Thallium	205-1	0.19	0.20	0.21	0.20	3.69	ppb
Tin	118-1	1.91	1.88	2.11	1.96	6.37	ppb
Titanium	47-1	714.71	721.01	738.16	724.63	1.68	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-08

Instrumnet Name :

P7

Client Sample ID :

MYE4K5

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:09:35

DataFile Name :

107SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.68	0.68	0.72	0.69	3.09	ppb
Vanadium	51-2	110.56	109.39	110.41	110.12	0.57	ppb
Yttrium	89-1				99		%
Yttrium	89-2				104		%
Zinc	66-2	198.40	199.28	197.46	198.38	0.46	ppb
Zirconium	90-1	4.13	4.17	4.22	4.17	1.03	ppb
Zirconium	91-1	4.28	4.23	4.28	4.27	0.72	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-09 Instrumnet Name : P7
Client Sample ID : MYE4K6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:12:45 DataFile Name : 108SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	26456.66	26041.38	26510.43	26336.16	0.97	ppb
Antimony	121-1	0.47	0.49	0.50	0.49	3.52	ppb
Arsenic	75-2	37.99	36.30	37.37	37.22	2.30	ppb
Barium	135-1	274.24	273.74	282.58	276.85	1.79	ppb
Barium	137-1	275.56	276.49	280.68	277.58	0.98	ppb
Beryllium	9-1	0.89	0.93	1.00	0.94	5.79	ppb
Bismuth	209-1				100		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.12	0.94	0.92	0.99	11.30	ppb
Cadmium	106-1	-0.43	-1.04	-1.04	-0.84		ppb
Cadmium	111-1	0.48	0.44	0.45	0.46	4.30	ppb
Calcium	43-1	20996.02	21022.19	21200.99	21073.07	0.53	ppb
Calcium	44-1	19780.14	19755.58	19986.97	19840.89	0.64	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	55.17	55.78	56.06	55.67	0.81	ppb
Cobalt	59-2	31.00	31.43	31.60	31.34	0.97	ppb
Copper	63-2	111.37	112.11	112.49	111.99	0.51	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				103		%
Indium	115-1				95		%
Indium	115-2				98		%
Iron	56-2	55100.90	55724.19	55854.36	55559.82	0.72	ppb
Iron	57-2	54919.68	55486.55	55312.66	55239.63	0.53	ppb
Iron	54-2	55277.49	56319.60	56155.58	55917.56	1.00	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-09 Instrumnet Name : P7
Client Sample ID : MYE4K6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:12:45 DataFile Name : 108SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	25.98	26.29	26.70	26.32	1.36	ppb
Lead	207-1	24.82	25.00	25.33	25.05	1.03	ppb
Lead	208-1	25.15	25.29	25.72	25.39	1.17	ppb
Lithium	6-1				88		%
Magnesium	24-2	13646.03	13825.38	13957.91	13809.77	1.13	ppb
Manganese	55-2	1578.41	1596.89	1583.68	1586.33	0.60	ppb
Molybdenum	94-1	8.24	8.23	8.41	8.29	1.21	ppb
Molybdenum	95-1	0.83	0.88	0.86	0.85	3.11	ppb
Molybdenum	96-1	2.11	2.13	2.11	2.12	0.56	ppb
Molybdenum	97-1	0.85	0.83	0.87	0.85	1.96	ppb
Molybdenum	98-1	0.82	0.83	0.87	0.84	3.07	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	48.35	48.82	48.91	48.69	0.62	ppb
Phosphorus	31-2	2455.01	2418.47	2445.10	2439.53	0.77	ppb
Potassium	39-2	9668.73	9730.13	9719.13	9706.00	0.34	ppb
Rhodium	103-1				92		%
Rhodium	103-2				99		%
Scandium	45-1				92		%
Scandium	45-2				100		%
Selenium	82-1	0.58	0.38	0.43	0.46	22.12	ppb
Selenium	77-2	12.33	10.54	13.46	12.11	12.15	ppb
Selenium	78-2	1.36	3.15	3.42	2.65	42.32	ppb
Silicon	28-1	1676.39	1670.52	1695.35	1680.76	0.77	ppb
Silver	107-1	0.20	0.20	0.20	0.20	1.70	ppb
Silver	109-1	0.21	0.21	0.22	0.21	3.29	ppb
Sodium	23-2	271.01	276.53	275.54	274.36	1.07	ppb
Strontium	86-1	75.06	75.20	75.97	75.41	0.65	ppb
Strontium	88-1	76.63	77.96	78.53	77.70	1.25	ppb
Sulfur	34-1	-1217.05	-1212.57	-1179.63	-1203.08		ppb
Terbium	159-1				101		%
Terbium	159-2				103		%
Thallium	203-1	0.22	0.22	0.22	0.22	1.35	ppb
Thallium	205-1	0.22	0.23	0.23	0.23	2.37	ppb
Tin	118-1	1.88	1.84	1.88	1.87	1.25	ppb
Titanium	47-1	725.30	730.84	735.76	730.63	0.72	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-09

Instrumnet Name :

P7

Client Sample ID :

MYE4K6

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:12:45

DataFile Name :

108SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.74	0.72	0.74	0.73	2.06	ppb
Vanadium	51-2	105.30	106.36	106.51	106.06	0.62	ppb
Yttrium	89-1				98		%
Yttrium	89-2				104		%
Zinc	66-2	159.84	160.99	160.74	160.52	0.38	ppb
Zirconium	90-1	4.55	4.56	4.57	4.56	0.25	ppb
Zirconium	91-1	4.59	4.61	4.66	4.62	0.76	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV004 Instrumnet Name : P7
Client Sample ID : CCV004 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:15:52 DataFile Name : 109CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	47516.11	47632.43	47328.78	47492.44	0.32	ppb
Antimony	121-1	479.80	493.42	484.75	485.99	1.42	ppb
Arsenic	75-2	506.00	503.50	504.70	504.73	0.25	ppb
Barium	135-1	2440.62	2498.41	2429.09	2456.04	1.51	ppb
Barium	137-1	2404.49	2469.23	2426.43	2433.38	1.35	ppb
Beryllium	9-1	495.14	494.65	491.75	493.84	0.37	ppb
Bismuth	209-1				95		%
Bismuth	209-2				95		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	488.30	494.88	486.93	490.04	0.87	ppb
Cadmium	106-1	485.72	499.58	479.73	488.34	2.08	ppb
Cadmium	111-1	463.40	474.45	464.81	467.55	1.29	ppb
Calcium	43-1	244197.76	245546.51	244213.16	244652.48	0.32	ppb
Calcium	44-1	241130.52	244092.85	240820.87	242014.74	0.75	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	500.65	501.29	488.28	496.74	1.48	ppb
Cobalt	59-2	523.81	523.05	517.09	521.31	0.71	ppb
Copper	63-2	4990.83	5007.86	4953.17	4983.95	0.56	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				103		%
Indium	115-1				91		%
Indium	115-2				93		%
Iron	56-2	125325.41	126263.87	124310.71	125300.00	0.78	ppb
Iron	57-2	126129.70	126486.94	122963.94	125193.53	1.55	ppb
Iron	54-2	125653.03	127685.26	125510.91	126283.07	0.96	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV004 Instrumnet Name : P7
Client Sample ID : CCV004 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:15:52 DataFile Name : 109CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2513.91	2521.55	2483.15	2506.20	0.81	ppb
Lead	207-1	2459.91	2499.96	2478.85	2479.57	0.81	ppb
Lead	208-1	2479.41	2504.73	2481.94	2488.69	0.56	ppb
Lithium	6-1				88		%
Magnesium	24-2	239390.02	240851.08	237287.22	239176.11	0.75	ppb
Manganese	55-2	4886.69	4992.03	4889.52	4922.75	1.22	ppb
Molybdenum	94-1	4912.38	4954.41	4919.40	4928.73	0.46	ppb
Molybdenum	95-1	4927.54	4932.75	4957.27	4939.18	0.32	ppb
Molybdenum	96-1	4918.94	4896.58	4942.43	4919.32	0.47	ppb
Molybdenum	97-1	4925.24	4870.73	4932.58	4909.52	0.69	ppb
Molybdenum	98-1	4880.01	4860.38	4863.89	4868.09	0.22	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	488.97	488.32	478.67	485.32	1.19	ppb
Phosphorus	31-2	9570.15	9513.28	9430.18	9504.54	0.74	ppb
Potassium	39-2	119591.22	120077.48	118183.87	119284.19	0.82	ppb
Rhodium	103-1				85		%
Rhodium	103-2				91		%
Scandium	45-1				86		%
Scandium	45-2				95		%
Selenium	82-1	493.53	492.79	489.90	492.07	0.39	ppb
Selenium	77-2	474.83	475.58	499.07	483.16	2.85	ppb
Selenium	78-2	484.07	493.73	493.75	490.52	1.14	ppb
Silicon	28-1	2756.54	2778.30	2785.52	2773.45	0.54	ppb
Silver	107-1	489.89	496.41	484.96	490.42	1.17	ppb
Silver	109-1	495.44	498.62	494.74	496.27	0.42	ppb
Sodium	23-2	240304.81	242831.74	235269.57	239468.71	1.61	ppb
Strontium	86-1	12224.86	12202.22	12185.32	12204.13	0.16	ppb
Strontium	88-1	12229.97	12235.79	12131.04	12198.93	0.48	ppb
Sulfur	34-1	9139.58	9163.20	9065.65	9122.81	0.56	ppb
Terbium	159-1				101		%
Terbium	159-2				104		%
Thallium	203-1	497.80	507.54	501.32	502.22	0.98	ppb
Thallium	205-1	500.27	502.35	503.33	501.98	0.31	ppb
Tin	118-1	469.53	476.43	470.51	472.16	0.79	ppb
Titanium	47-1	4908.41	4946.91	4913.38	4922.90	0.43	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

CCV004

Instrumnet Name :

P7

Client Sample ID :

CCV004

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:15:52

DataFile Name :

109CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	506.95	519.14	517.54	514.54	1.29	ppb
Vanadium	51-2	506.05	511.84	494.46	504.12	1.76	ppb
Yttrium	89-1				92		%
Yttrium	89-2				97		%
Zinc	66-2	4987.90	5022.91	4913.88	4974.90	1.12	ppb
Zirconium	90-1	496.54	500.56	506.86	501.32	1.04	ppb
Zirconium	91-1	510.24	503.31	503.17	505.57	0.80	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB004 Instrumnet Name : P7
Client Sample ID : CCB004 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:18:36 DataFile Name : 110CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	4.45	4.70	4.70	4.62	3.14	ppb
Antimony	121-1	0.30	0.29	0.29	0.29	2.43	ppb
Arsenic	75-2	0.03	-0.01	0.04	0.02	122.45	ppb
Barium	135-1	0.14	0.18	0.17	0.16	14.40	ppb
Barium	137-1	0.17	0.16	0.16	0.16	1.96	ppb
Beryllium	9-1	0.04	0.00	0.00	0.01	122.50	ppb
Bismuth	209-1				107		%
Bismuth	209-2				109		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.01	0.06	0.13	0.07	89.40	ppb
Cadmium	106-1	-1.60	-1.55	-1.78	-1.64		ppb
Cadmium	111-1	-0.09	-0.09	-0.11	-0.10		ppb
Calcium	43-1	4.45	5.88	6.91	5.74	21.49	ppb
Calcium	44-1	4.46	5.36	5.55	5.12	11.35	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.02	0.05	0.03	0.03	42.47	ppb
Cobalt	59-2	0.03	0.02	0.04	0.03	34.12	ppb
Copper	63-2	0.14	0.16	0.09	0.13	26.37	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				107		%
Indium	115-1				100		%
Indium	115-2				104		%
Iron	56-2	12.83	12.70	13.55	13.03	3.50	ppb
Iron	57-2	10.74	10.15	10.26	10.38	3.03	ppb
Iron	54-2	12.77	11.75	13.58	12.70	7.21	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB004 Instrumnet Name : P7
Client Sample ID : CCB004 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:18:36 DataFile Name : 110CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.08	0.07	0.05	0.07	19.66	ppb
Lead	207-1	0.09	0.07	0.07	0.08	15.94	ppb
Lead	208-1	0.08	0.07	0.06	0.07	11.93	ppb
Lithium	6-1				94		%
Magnesium	24-2	5.38	4.66	5.27	5.10	7.58	ppb
Manganese	55-2	0.38	0.43	0.45	0.42	7.98	ppb
Molybdenum	94-1	0.26	0.25	0.28	0.26	5.53	ppb
Molybdenum	95-1	0.21	0.18	0.19	0.19	7.43	ppb
Molybdenum	96-1	0.21	0.19	0.19	0.20	7.01	ppb
Molybdenum	97-1	0.23	0.16	0.20	0.20	16.38	ppb
Molybdenum	98-1	0.20	0.19	0.17	0.19	7.26	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.08	0.07	0.09	0.08	10.28	ppb
Phosphorus	31-2	-17.21	-18.78	-22.95	-19.65		ppb
Potassium	39-2	-3.52	-3.13	-3.19	-3.28		ppb
Rhodium	103-1				97		%
Rhodium	103-2				104		%
Scandium	45-1				92		%
Scandium	45-2				98		%
Selenium	82-1	-0.61	-0.33	0.05	-0.30		ppb
Selenium	77-2	0.00	-0.22	-0.22	-0.15		ppb
Selenium	78-2	-0.43	0.49	-0.94	-0.29		ppb
Silicon	28-1	0.73	0.69	0.82	0.75	8.93	ppb
Silver	107-1	0.05	0.04	0.04	0.04	13.93	ppb
Silver	109-1	0.05	0.04	0.03	0.04	18.52	ppb
Sodium	23-2	8.51	8.05	7.35	7.97	7.35	ppb
Strontium	86-1	0.27	0.25	0.22	0.25	11.01	ppb
Strontium	88-1	0.25	0.23	0.23	0.24	3.54	ppb
Sulfur	34-1	-804.49	-731.87	-661.98	-732.78		ppb
Terbium	159-1				105		%
Terbium	159-2				108		%
Thallium	203-1	0.06	0.06	0.03	0.05	29.16	ppb
Thallium	205-1	0.05	0.05	0.04	0.05	10.93	ppb
Tin	118-1	0.05	0.06	0.07	0.06	17.44	ppb
Titanium	47-1	0.28	0.33	0.31	0.31	7.95	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB004 Instrumnet Name : P7
Client Sample ID : CCB004 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:18:36 DataFile Name : 110CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.01	0.01	18.42	ppb
Vanadium	51-2	0.05	0.04	0.05	0.05	10.81	ppb
Yttrium	89-1				96		%
Yttrium	89-2				102		%
Zinc	66-2	-0.16	-0.10	-0.08	-0.11		ppb
Zirconium	90-1	0.05	0.05	0.04	0.05	8.21	ppb
Zirconium	91-1	0.07	0.06	0.05	0.06	18.26	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-10 Instrumnet Name : P7
Client Sample ID : MYE4K7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:21:51 DataFile Name : 111SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	32157.23	31909.90	31864.36	31977.17	0.49	ppb
Antimony	121-1	1.01	1.07	1.07	1.05	3.56	ppb
Arsenic	75-2	72.44	75.29	73.42	73.72	1.96	ppb
Barium	135-1	337.06	337.23	343.39	339.23	1.06	ppb
Barium	137-1	358.65	364.56	369.67	364.29	1.51	ppb
Beryllium	9-1	1.15	1.21	1.20	1.19	2.88	ppb
Bismuth	209-1				105		%
Bismuth	209-2				107		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.08	2.50	2.25	2.27	9.38	ppb
Cadmium	106-1	2.13	2.08	1.56	1.92	16.44	ppb
Cadmium	111-1	1.68	1.76	1.62	1.69	4.26	ppb
Calcium	43-1	15672.29	15714.87	15813.31	15733.49	0.46	ppb
Calcium	44-1	14760.15	14890.44	14897.08	14849.22	0.52	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	77.21	76.76	77.43	77.13	0.45	ppb
Cobalt	59-2	37.61	37.71	37.56	37.63	0.20	ppb
Copper	63-2	258.87	258.13	258.88	258.62	0.17	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				108		%
Holmium	165-2				110		%
Indium	115-1				102		%
Indium	115-2				105		%
Iron	56-2	69006.51	68528.39	68702.43	68745.77	0.35	ppb
Iron	57-2	68202.59	67612.99	67664.34	67826.64	0.48	ppb
Iron	54-2	69132.70	68715.09	69468.76	69105.52	0.55	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-10 Instrumnet Name : P7
Client Sample ID : MYE4K7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:21:51 DataFile Name : 111SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	58.22	59.50	60.22	59.32	1.71	ppb
Lead	207-1	57.43	58.62	58.56	58.21	1.15	ppb
Lead	208-1	57.49	58.59	58.81	58.29	1.21	ppb
Lithium	6-1				95		%
Magnesium	24-2	16293.74	16178.76	16322.80	16265.10	0.47	ppb
Manganese	55-2	1904.24	1893.06	1882.20	1893.17	0.58	ppb
Molybdenum	94-1	14.08	14.21	14.21	14.17	0.52	ppb
Molybdenum	95-1	1.32	1.33	1.34	1.33	0.87	ppb
Molybdenum	96-1	3.19	3.19	3.20	3.19	0.16	ppb
Molybdenum	97-1	1.25	1.22	1.28	1.25	2.24	ppb
Molybdenum	98-1	1.23	1.27	1.28	1.26	2.30	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	67.93	67.34	67.65	67.64	0.44	ppb
Phosphorus	31-2	2400.98	2379.96	2412.76	2397.90	0.69	ppb
Potassium	39-2	11568.65	11416.44	11471.89	11485.66	0.67	ppb
Rhodium	103-1				99		%
Rhodium	103-2				106		%
Scandium	45-1				104		%
Scandium	45-2				112		%
Selenium	82-1	1.47	1.75	1.36	1.53	13.00	ppb
Selenium	77-2	18.80	17.14	16.62	17.52	6.49	ppb
Selenium	78-2	3.83	4.22	4.43	4.16	7.35	ppb
Silicon	28-1	1786.67	1778.15	1782.94	1782.59	0.24	ppb
Silver	107-1	0.68	0.68	0.66	0.67	1.70	ppb
Silver	109-1	0.70	0.64	0.65	0.66	4.28	ppb
Sodium	23-2	239.65	240.32	239.90	239.95	0.14	ppb
Strontium	86-1	72.64	73.67	74.03	73.45	0.98	ppb
Strontium	88-1	75.38	75.37	76.80	75.85	1.08	ppb
Sulfur	34-1	-808.49	-981.76	-1168.75	-986.33		ppb
Terbium	159-1				108		%
Terbium	159-2				110		%
Thallium	203-1	0.36	0.36	0.35	0.36	2.14	ppb
Thallium	205-1	0.35	0.34	0.32	0.34	4.59	ppb
Tin	118-1	2.31	2.41	2.47	2.40	3.35	ppb
Titanium	47-1	785.87	792.91	797.13	791.97	0.72	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-10

Instrumnet Name :

P7

Client Sample ID :

MYE4K7

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:21:51

DataFile Name :

111SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.86	0.90	0.90	0.89	2.51	ppb
Vanadium	51-2	129.27	128.92	129.91	129.36	0.39	ppb
Yttrium	89-1				107		%
Yttrium	89-2				112		%
Zinc	66-2	252.36	250.55	249.26	250.72	0.62	ppb
Zirconium	90-1	7.69	7.83	7.94	7.82	1.64	ppb
Zirconium	91-1	7.81	7.99	7.95	7.91	1.22	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-11 Instrumnet Name : P7
Client Sample ID : MYE4K8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:24:58 DataFile Name : 112SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	31698.53	31709.90	31883.32	31763.92	0.33	ppb
Antimony	121-1	0.91	1.14	0.94	0.99	12.77	ppb
Arsenic	75-2	64.01	64.23	63.24	63.83	0.82	ppb
Barium	135-1	363.41	439.72	366.58	389.90	11.07	ppb
Barium	137-1	394.04	473.36	393.05	420.15	10.97	ppb
Beryllium	9-1	1.27	1.54	1.23	1.35	12.63	ppb
Bismuth	209-1				99		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.84	2.28	1.88	2.00	12.29	ppb
Cadmium	106-1	1.18	1.85	0.73	1.25	44.89	ppb
Cadmium	111-1	1.37	1.62	1.34	1.44	10.61	ppb
Calcium	43-1	15919.30	19119.69	16617.35	17218.78	9.77	ppb
Calcium	44-1	14997.44	18143.99	15395.04	16178.82	10.59	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	75.63	75.00	75.72	75.45	0.52	ppb
Cobalt	59-2	38.70	38.19	38.32	38.40	0.69	ppb
Copper	63-2	218.90	217.90	219.87	218.89	0.45	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				107		%
Indium	115-1				96		%
Indium	115-2				102		%
Iron	56-2	64285.04	64084.72	64952.52	64440.76	0.71	ppb
Iron	57-2	64705.73	63617.64	64023.50	64115.62	0.86	ppb
Iron	54-2	64562.96	64658.82	65444.62	64888.80	0.75	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-11 Instrumnet Name : P7
Client Sample ID : MYE4K8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:24:58 DataFile Name : 112SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	47.56	57.41	49.01	51.33	10.36	ppb
Lead	207-1	47.14	56.44	47.78	50.45	10.29	ppb
Lead	208-1	46.91	56.67	47.91	50.50	10.64	ppb
Lithium	6-1				88		%
Magnesium	24-2	16542.99	16541.84	16553.88	16546.24	0.04	ppb
Manganese	55-2	2010.18	2014.07	2030.35	2018.20	0.53	ppb
Molybdenum	94-1	10.94	13.35	11.53	11.94	10.51	ppb
Molybdenum	95-1	1.53	1.94	1.63	1.70	12.62	ppb
Molybdenum	96-1	2.99	3.71	3.24	3.32	11.03	ppb
Molybdenum	97-1	1.50	1.91	1.63	1.68	12.35	ppb
Molybdenum	98-1	1.49	1.85	1.60	1.65	11.30	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	65.54	64.97	65.68	65.40	0.57	ppb
Phosphorus	31-2	2809.14	2759.50	2776.15	2781.60	0.91	ppb
Potassium	39-2	11006.74	10976.57	11076.00	11019.77	0.46	ppb
Rhodium	103-1				93		%
Rhodium	103-2				104		%
Scandium	45-1				98		%
Scandium	45-2				109		%
Selenium	82-1	0.57	1.30	1.00	0.95	38.18	ppb
Selenium	77-2	17.02	14.74	14.08	15.28	10.09	ppb
Selenium	78-2	3.57	2.31	2.02	2.63	31.47	ppb
Silicon	28-1	1909.31	2318.67	1986.27	2071.42	10.50	ppb
Silver	107-1	0.60	0.74	0.62	0.66	11.37	ppb
Silver	109-1	0.60	0.74	0.61	0.65	11.84	ppb
Sodium	23-2	339.40	338.44	338.28	338.71	0.18	ppb
Strontium	86-1	90.52	110.52	94.72	98.59	10.70	ppb
Strontium	88-1	93.18	113.64	98.14	101.65	10.50	ppb
Sulfur	34-1	-1115.38	-558.10	-1140.32	-937.93		ppb
Terbium	159-1				100		%
Terbium	159-2				108		%
Thallium	203-1	0.30	0.38	0.30	0.32	14.12	ppb
Thallium	205-1	0.30	0.36	0.29	0.31	11.59	ppb
Tin	118-1	2.32	2.86	2.29	2.49	12.85	ppb
Titanium	47-1	739.18	893.56	773.57	802.11	10.10	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-11

Instrumnet Name :

P7

Client Sample ID :

MYE4K8

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:24:58

DataFile Name :

112SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.84	1.02	0.86	0.91	10.42	ppb
Vanadium	51-2	122.32	120.45	121.79	121.52	0.79	ppb
Yttrium	89-1				102		%
Yttrium	89-2				110		%
Zinc	66-2	253.76	251.94	253.19	252.97	0.37	ppb
Zirconium	90-1	5.73	7.15	6.07	6.32	11.78	ppb
Zirconium	91-1	5.80	7.05	6.05	6.30	10.50	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-12 Instrumnet Name : P7
Client Sample ID : MYE4K9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:28:04 DataFile Name : 113SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	34792.95	34079.78	34056.47	34309.73	1.22	ppb
Antimony	121-1	0.95	0.99	1.02	0.98	3.58	ppb
Arsenic	75-2	69.84	68.47	69.38	69.23	1.01	ppb
Barium	135-1	379.28	387.04	384.55	383.62	1.03	ppb
Barium	137-1	415.99	415.30	422.04	417.78	0.89	ppb
Beryllium	9-1	0.98	0.94	1.13	1.02	9.50	ppb
Bismuth	209-1				104		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.81	1.71	1.93	1.82	5.91	ppb
Cadmium	106-1	0.89	1.84	0.02	0.92	99.44	ppb
Cadmium	111-1	0.89	1.03	0.88	0.93	8.93	ppb
Calcium	43-1	21046.76	21753.67	21691.52	21497.32	1.82	ppb
Calcium	44-1	20055.76	20706.63	20321.72	20361.37	1.61	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	103.24	101.13	101.01	101.80	1.23	ppb
Cobalt	59-2	47.23	46.41	46.78	46.81	0.88	ppb
Copper	63-2	197.59	194.11	193.80	195.16	1.08	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				105		%
Holmium	165-2				107		%
Indium	115-1				100		%
Indium	115-2				101		%
Iron	56-2	76607.24	75123.06	75344.12	75691.47	1.06	ppb
Iron	57-2	75580.57	75025.62	74597.33	75067.84	0.66	ppb
Iron	54-2	76819.44	76522.41	76252.28	76531.38	0.37	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-12 Instrumnet Name : P7
Client Sample ID : MYE4K9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:28:04 DataFile Name : 113SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	35.62	35.44	35.81	35.62	0.51	ppb
Lead	207-1	35.01	34.28	35.17	34.82	1.35	ppb
Lead	208-1	35.01	34.59	34.84	34.81	0.61	ppb
Lithium	6-1				92		%
Magnesium	24-2	22313.71	22006.13	21934.58	22084.81	0.91	ppb
Manganese	55-2	2494.49	2472.60	2481.66	2482.92	0.44	ppb
Molybdenum	94-1	10.92	11.30	11.16	11.12	1.73	ppb
Molybdenum	95-1	1.24	1.29	1.28	1.27	2.18	ppb
Molybdenum	96-1	2.80	2.87	2.91	2.86	1.92	ppb
Molybdenum	97-1	1.25	1.24	1.19	1.22	2.62	ppb
Molybdenum	98-1	1.19	1.18	1.17	1.18	0.67	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	88.52	87.84	88.43	88.27	0.42	ppb
Phosphorus	31-2	2482.29	2481.71	2439.77	2467.92	0.99	ppb
Potassium	39-2	8965.15	8767.59	8739.13	8823.96	1.40	ppb
Rhodium	103-1				98		%
Rhodium	103-2				103		%
Scandium	45-1				102		%
Scandium	45-2				107		%
Selenium	82-1	0.74	0.08	0.68	0.50	73.00	ppb
Selenium	77-2	11.66	12.54	10.99	11.73	6.62	ppb
Selenium	78-2	2.81	2.90	2.75	2.82	2.61	ppb
Silicon	28-1	1563.26	1601.99	1592.48	1585.91	1.27	ppb
Silver	107-1	0.37	0.36	0.38	0.37	2.74	ppb
Silver	109-1	0.36	0.37	0.40	0.38	5.01	ppb
Sodium	23-2	356.75	352.10	351.31	353.39	0.83	ppb
Strontium	86-1	105.57	107.87	107.83	107.09	1.23	ppb
Strontium	88-1	109.41	110.73	110.48	110.21	0.64	ppb
Sulfur	34-1	-1255.19	-1195.00	-1219.72	-1223.30		ppb
Terbium	159-1				105		%
Terbium	159-2				107		%
Thallium	203-1	0.25	0.23	0.24	0.24	4.59	ppb
Thallium	205-1	0.25	0.25	0.23	0.24	3.89	ppb
Tin	118-1	2.26	2.28	2.36	2.30	2.17	ppb
Titanium	47-1	734.39	761.49	754.71	750.20	1.88	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-12

Instrumnet Name :

P7

Client Sample ID :

MYE4K9

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:28:04

DataFile Name :

113SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.71	0.68	0.71	0.70	2.13	ppb
Vanadium	51-2	161.01	157.89	157.51	158.81	1.21	ppb
Yttrium	89-1				106		%
Yttrium	89-2				109		%
Zinc	66-2	191.26	187.41	188.09	188.92	1.09	ppb
Zirconium	90-1	5.98	6.02	6.07	6.02	0.75	ppb
Zirconium	91-1	5.97	6.18	6.22	6.12	2.23	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-13 Instrumnet Name : P7
Client Sample ID : MYE4L0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:31:10 DataFile Name : 114SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	30786.41	30709.02	30531.72	30675.71	0.43	ppb
Antimony	121-1	0.64	0.65	0.66	0.65	2.01	ppb
Arsenic	75-2	53.96	52.60	51.72	52.76	2.14	ppb
Barium	135-1	391.82	393.67	396.09	393.86	0.54	ppb
Barium	137-1	422.22	426.02	419.38	422.54	0.79	ppb
Beryllium	9-1	1.31	1.20	1.35	1.29	6.14	ppb
Bismuth	209-1				102		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.45	1.44	1.40	1.43	1.69	ppb
Cadmium	106-1	0.31	0.91	1.06	0.76	52.16	ppb
Cadmium	111-1	0.44	0.44	0.49	0.46	6.30	ppb
Calcium	43-1	28509.88	28865.10	29308.42	28894.46	1.38	ppb
Calcium	44-1	27016.61	27245.67	27377.90	27213.39	0.67	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	70.47	70.94	70.47	70.63	0.39	ppb
Cobalt	59-2	36.74	36.38	36.64	36.59	0.50	ppb
Copper	63-2	98.26	98.11	98.04	98.14	0.12	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				105		%
Indium	115-1				99		%
Indium	115-2				101		%
Iron	56-2	61870.31	61959.99	62202.45	62010.92	0.28	ppb
Iron	57-2	61579.92	61854.67	61535.20	61656.59	0.28	ppb
Iron	54-2	62902.32	62843.28	63033.82	62926.47	0.16	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-13 Instrumnet Name : P7
Client Sample ID : MYE4L0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:31:10 DataFile Name : 114SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	24.71	24.69	25.29	24.90	1.37	ppb
Lead	207-1	23.32	22.97	23.76	23.35	1.69	ppb
Lead	208-1	23.68	23.54	24.25	23.82	1.57	ppb
Lithium	6-1				92		%
Magnesium	24-2	16226.97	16210.48	16263.50	16233.65	0.17	ppb
Manganese	55-2	1848.17	1858.19	1855.79	1854.05	0.28	ppb
Molybdenum	94-1	17.31	17.48	17.79	17.53	1.37	ppb
Molybdenum	95-1	0.85	0.86	0.85	0.85	0.49	ppb
Molybdenum	96-1	3.10	3.03	3.20	3.11	2.65	ppb
Molybdenum	97-1	0.82	0.82	0.85	0.83	2.32	ppb
Molybdenum	98-1	0.79	0.79	0.80	0.79	1.04	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	71.21	70.63	70.58	70.81	0.49	ppb
Phosphorus	31-2	2108.54	2127.48	2172.19	2136.07	1.53	ppb
Potassium	39-2	7505.28	7562.71	7546.02	7538.00	0.39	ppb
Rhodium	103-1				95		%
Rhodium	103-2				102		%
Scandium	45-1				97		%
Scandium	45-2				104		%
Selenium	82-1	0.34	0.07	0.62	0.34	80.42	ppb
Selenium	77-2	18.02	15.31	12.15	15.16	19.41	ppb
Selenium	78-2	2.35	2.75	3.38	2.82	18.36	ppb
Silicon	28-1	1894.81	1911.60	1931.78	1912.73	0.97	ppb
Silver	107-1	0.14	0.13	0.16	0.15	9.42	ppb
Silver	109-1	0.14	0.15	0.15	0.15	1.71	ppb
Sodium	23-2	275.89	276.94	275.25	276.03	0.31	ppb
Strontium	86-1	92.78	91.59	94.15	92.84	1.38	ppb
Strontium	88-1	94.77	96.50	95.96	95.74	0.93	ppb
Sulfur	34-1	-1367.52	-1335.64	-1322.19	-1341.78		ppb
Terbium	159-1				103		%
Terbium	159-2				106		%
Thallium	203-1	0.25	0.25	0.27	0.26	4.50	ppb
Thallium	205-1	0.25	0.25	0.26	0.25	2.55	ppb
Tin	118-1	2.00	2.01	2.13	2.05	3.41	ppb
Titanium	47-1	923.10	931.76	921.32	925.39	0.60	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-13

Instrumnet Name :

P7

Client Sample ID :

MYE4L0

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:31:10

DataFile Name :

114SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.84	0.83	0.85	0.84	1.16	ppb
Vanadium	51-2	120.43	118.80	118.08	119.10	1.01	ppb
Yttrium	89-1				104		%
Yttrium	89-2				108		%
Zinc	66-2	151.23	150.96	152.60	151.60	0.58	ppb
Zirconium	90-1	10.02	10.05	10.27	10.11	1.36	ppb
Zirconium	91-1	10.03	10.23	10.37	10.21	1.64	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-14 Instrumnet Name : P7
Client Sample ID : MYE4L1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:34:15 DataFile Name : 115SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	34514.85	34636.51	34800.50	34650.62	0.41	ppb
Antimony	121-1	0.55	0.55	0.63	0.57	8.64	ppb
Arsenic	75-2	62.31	62.47	60.58	61.79	1.69	ppb
Barium	135-1	353.48	358.20	403.24	371.64	7.39	ppb
Barium	137-1	382.60	390.30	439.13	404.01	7.59	ppb
Beryllium	9-1	1.33	1.35	1.39	1.36	2.19	ppb
Bismuth	209-1				97		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.65	1.26	1.85	1.59	18.82	ppb
Cadmium	106-1	-0.22	-0.10	-0.73	-0.35		ppb
Cadmium	111-1	0.84	0.93	0.93	0.90	5.80	ppb
Calcium	43-1	14962.62	15053.39	17139.76	15718.59	7.84	ppb
Calcium	44-1	14053.59	14099.44	16028.33	14727.12	7.65	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	77.62	77.66	78.46	77.91	0.61	ppb
Cobalt	59-2	39.59	39.76	39.48	39.61	0.36	ppb
Copper	63-2	182.73	181.78	182.97	182.49	0.34	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				98		%
Holmium	165-2				106		%
Indium	115-1				93		%
Indium	115-2				99		%
Iron	56-2	76759.69	76849.90	77270.32	76959.97	0.35	ppb
Iron	57-2	75786.21	76145.80	76837.21	76256.41	0.70	ppb
Iron	54-2	77105.17	76440.23	76945.30	76830.23	0.45	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-14 Instrumnet Name : P7
Client Sample ID : MYE4L1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:34:15 DataFile Name : 115SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	39.85	40.40	46.78	42.35	9.10	ppb
Lead	207-1	38.83	39.19	45.35	41.12	8.90	ppb
Lead	208-1	38.97	39.31	45.36	41.21	8.73	ppb
Lithium	6-1				88		%
Magnesium	24-2	17075.30	16692.33	16722.96	16830.20	1.26	ppb
Manganese	55-2	1663.99	1672.13	1662.96	1666.36	0.30	ppb
Molybdenum	94-1	10.94	11.15	12.51	11.53	7.40	ppb
Molybdenum	95-1	1.10	1.19	1.36	1.22	10.81	ppb
Molybdenum	96-1	2.80	2.81	3.23	2.95	8.28	ppb
Molybdenum	97-1	1.19	1.12	1.32	1.21	8.19	ppb
Molybdenum	98-1	1.09	1.11	1.24	1.15	6.93	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	62.19	61.36	61.69	61.75	0.67	ppb
Phosphorus	31-2	2354.14	2346.60	2316.11	2338.95	0.86	ppb
Potassium	39-2	11080.27	11078.41	11130.73	11096.47	0.27	ppb
Rhodium	103-1				90		%
Rhodium	103-2				101		%
Scandium	45-1				92		%
Scandium	45-2				102		%
Selenium	82-1	0.49	1.05	0.96	0.83	36.21	ppb
Selenium	77-2	12.44	17.42	15.88	15.24	16.73	ppb
Selenium	78-2	2.98	3.88	3.35	3.41	13.21	ppb
Silicon	28-1	1860.50	1872.68	2140.49	1957.89	8.08	ppb
Silver	107-1	0.37	0.37	0.42	0.38	7.13	ppb
Silver	109-1	0.38	0.40	0.43	0.41	6.42	ppb
Sodium	23-2	313.21	316.65	315.92	315.26	0.57	ppb
Strontium	86-1	71.71	73.46	83.32	76.17	8.22	ppb
Strontium	88-1	74.60	75.45	85.96	78.67	8.04	ppb
Sulfur	34-1	-1650.25	-1758.17	-1417.41	-1608.61		ppb
Terbium	159-1				100		%
Terbium	159-2				107		%
Thallium	203-1	0.28	0.26	0.34	0.29	13.35	ppb
Thallium	205-1	0.29	0.28	0.31	0.29	6.06	ppb
Tin	118-1	2.15	2.20	2.46	2.27	7.45	ppb
Titanium	47-1	893.58	897.37	1034.96	941.97	8.55	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-14

Instrumnet Name :

P7

Client Sample ID :

MYE4L1

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:34:15

DataFile Name :

115SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.03	1.02	1.20	1.08	9.40	ppb
Vanadium	51-2	139.54	140.96	140.16	140.22	0.51	ppb
Yttrium	89-1				98		%
Yttrium	89-2				106		%
Zinc	66-2	206.35	206.40	204.70	205.81	0.47	ppb
Zirconium	90-1	5.85	6.04	6.92	6.27	9.12	ppb
Zirconium	91-1	5.94	6.17	7.03	6.38	9.00	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-15 Instrumnet Name : P7
Client Sample ID : MYE4L2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:37:22 DataFile Name : 116SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	38966.32	38362.74	38568.06	38632.37	0.79	ppb
Antimony	121-1	0.51	0.57	0.53	0.54	4.87	ppb
Arsenic	75-2	70.13	68.74	69.44	69.44	1.00	ppb
Barium	135-1	401.78	405.73	401.84	403.12	0.56	ppb
Barium	137-1	437.50	436.90	433.90	436.10	0.44	ppb
Beryllium	9-1	1.50	1.31	1.45	1.42	6.97	ppb
Bismuth	209-1				102		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.64	1.54	2.05	1.74	15.42	ppb
Cadmium	106-1	0.11	0.61	0.43	0.39	65.10	ppb
Cadmium	111-1	1.07	1.00	1.03	1.03	3.21	ppb
Calcium	43-1	16703.75	17178.81	16786.89	16889.82	1.50	ppb
Calcium	44-1	15716.50	16273.00	15866.14	15951.88	1.81	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	86.38	87.37	87.36	87.03	0.65	ppb
Cobalt	59-2	48.46	48.26	48.29	48.34	0.22	ppb
Copper	63-2	207.43	208.83	207.27	207.85	0.41	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				106		%
Indium	115-1				96		%
Indium	115-2				99		%
Iron	56-2	86269.09	86885.61	87367.56	86840.76	0.63	ppb
Iron	57-2	84993.36	86043.27	86300.45	85779.03	0.81	ppb
Iron	54-2	87067.86	87176.75	86335.57	86860.06	0.53	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-15 Instrumnet Name : P7
Client Sample ID : MYE4L2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:37:22 DataFile Name : 116SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	45.24	46.43	45.80	45.83	1.30	ppb
Lead	207-1	43.14	45.07	44.19	44.13	2.19	ppb
Lead	208-1	43.65	45.48	44.45	44.52	2.06	ppb
Lithium	6-1				91		%
Magnesium	24-2	19272.11	19290.88	19548.38	19370.46	0.80	ppb
Manganese	55-2	1950.26	1935.30	1937.68	1941.08	0.41	ppb
Molybdenum	94-1	9.09	9.21	9.14	9.15	0.64	ppb
Molybdenum	95-1	1.20	1.19	1.21	1.20	0.83	ppb
Molybdenum	96-1	2.71	2.76	2.80	2.75	1.72	ppb
Molybdenum	97-1	1.17	1.21	1.20	1.19	1.81	ppb
Molybdenum	98-1	1.08	1.11	1.15	1.11	3.25	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	70.42	69.38	71.15	70.31	1.27	ppb
Phosphorus	31-2	2643.70	2630.85	2664.13	2646.23	0.63	ppb
Potassium	39-2	12759.92	12785.34	12876.43	12807.23	0.48	ppb
Rhodium	103-1				92		%
Rhodium	103-2				100		%
Scandium	45-1				95		%
Scandium	45-2				103		%
Selenium	82-1	1.28	0.86	0.69	0.94	32.02	ppb
Selenium	77-2	19.82	18.44	17.26	18.50	6.93	ppb
Selenium	78-2	3.23	2.27	3.09	2.86	18.12	ppb
Silicon	28-1	1745.95	1786.96	1757.51	1763.47	1.20	ppb
Silver	107-1	0.64	0.68	0.67	0.66	3.41	ppb
Silver	109-1	0.67	0.67	0.64	0.66	2.43	ppb
Sodium	23-2	345.67	345.72	346.50	345.96	0.13	ppb
Strontium	86-1	80.55	82.58	81.74	81.62	1.25	ppb
Strontium	88-1	83.05	86.76	84.49	84.77	2.21	ppb
Sulfur	34-1	-1755.57	-1760.26	-1785.94	-1767.26		ppb
Terbium	159-1				103		%
Terbium	159-2				106		%
Thallium	203-1	0.31	0.30	0.32	0.31	2.75	ppb
Thallium	205-1	0.30	0.31	0.31	0.30	1.94	ppb
Tin	118-1	2.24	2.31	2.28	2.28	1.43	ppb
Titanium	47-1	899.01	917.03	899.74	905.26	1.13	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-15 Instrumnet Name : P7
Client Sample ID : MYE4L2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:37:22 DataFile Name : 116SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.14	1.17	1.16	1.16	1.22	ppb
Vanadium	51-2	155.03	154.82	155.90	155.25	0.37	ppb
Yttrium	89-1				102		%
Yttrium	89-2				107		%
Zinc	66-2	231.39	232.55	233.42	232.45	0.44	ppb
Zirconium	90-1	4.80	4.92	4.88	4.87	1.24	ppb
Zirconium	91-1	4.84	4.85	5.00	4.90	1.87	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-16 Instrumnet Name : P7
Client Sample ID : MYE4P6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:40:26 DataFile Name : 117SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	31239.09	30657.98	30942.07	30946.38	0.94	ppb
Antimony	121-1	1.25	1.29	1.30	1.28	2.31	ppb
Arsenic	75-2	75.87	77.16	74.73	75.92	1.60	ppb
Barium	135-1	387.98	389.93	394.98	390.96	0.92	ppb
Barium	137-1	420.11	429.93	435.24	428.43	1.79	ppb
Beryllium	9-1	1.25	1.19	1.25	1.23	2.64	ppb
Bismuth	209-1				102		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.25	1.14	1.37	1.25	9.18	ppb
Cadmium	106-1	0.32	0.21	-0.23	0.10	296.48	ppb
Cadmium	111-1	0.56	0.56	0.43	0.52	14.70	ppb
Calcium	43-1	7170.58	7192.04	7291.44	7218.02	0.89	ppb
Calcium	44-1	6770.30	6780.57	6916.14	6822.33	1.19	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	64.33	63.85	63.91	64.03	0.41	ppb
Cobalt	59-2	37.89	37.78	38.50	38.06	1.02	ppb
Copper	63-2	121.10	120.29	120.47	120.62	0.35	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				104		%
Indium	115-1				96		%
Indium	115-2				98		%
Iron	56-2	74147.77	73803.20	73598.13	73849.70	0.38	ppb
Iron	57-2	72881.16	73381.29	73396.82	73219.76	0.40	ppb
Iron	54-2	74859.83	74108.80	73977.57	74315.40	0.64	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-16 Instrumnet Name : P7
Client Sample ID : MYE4P6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:40:26 DataFile Name : 117SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	26.67	27.03	27.06	26.92	0.81	ppb
Lead	207-1	25.82	26.12	26.31	26.08	0.95	ppb
Lead	208-1	26.09	26.32	26.56	26.32	0.91	ppb
Lithium	6-1				91		%
Magnesium	24-2	14273.95	14128.36	14209.30	14203.87	0.51	ppb
Manganese	55-2	1867.30	1872.65	1863.44	1867.80	0.25	ppb
Molybdenum	94-1	15.57	15.10	15.64	15.43	1.89	ppb
Molybdenum	95-1	0.69	0.67	0.68	0.68	1.73	ppb
Molybdenum	96-1	2.90	2.82	2.82	2.85	1.74	ppb
Molybdenum	97-1	0.65	0.61	0.65	0.64	3.56	ppb
Molybdenum	98-1	0.61	0.59	0.63	0.61	3.17	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	69.94	68.45	69.24	69.21	1.08	ppb
Phosphorus	31-2	1512.34	1528.01	1518.07	1519.47	0.52	ppb
Potassium	39-2	6783.38	6728.72	6697.56	6736.56	0.64	ppb
Rhodium	103-1				93		%
Rhodium	103-2				100		%
Scandium	45-1				93		%
Scandium	45-2				99		%
Selenium	82-1	0.06	-0.12	0.11	0.01	810.95	ppb
Selenium	77-2	18.14	19.22	14.83	17.39	13.15	ppb
Selenium	78-2	1.77	2.57	3.31	2.55	30.25	ppb
Silicon	28-1	1743.24	1733.61	1746.20	1741.02	0.38	ppb
Silver	107-1	0.19	0.18	0.19	0.18	3.11	ppb
Silver	109-1	0.18	0.16	0.17	0.17	4.76	ppb
Sodium	23-2	90.15	90.60	88.66	89.80	1.13	ppb
Strontium	86-1	60.35	59.51	61.83	60.56	1.94	ppb
Strontium	88-1	62.15	61.38	63.03	62.19	1.33	ppb
Sulfur	34-1	-1950.40	-1983.99	-1973.62	-1969.34		ppb
Terbium	159-1				102		%
Terbium	159-2				105		%
Thallium	203-1	0.27	0.27	0.26	0.27	2.16	ppb
Thallium	205-1	0.26	0.27	0.26	0.27	2.97	ppb
Tin	118-1	1.78	1.82	1.83	1.81	1.49	ppb
Titanium	47-1	702.38	704.03	712.77	706.39	0.79	ppb

LB Number :LB133265

Operator :Jaswal

Lab Sample ID :P4519-16

Instrumnet Name :P7

Client Sample ID :MYE4P6

Dilution Factor :1

Date & Time Acquired :2024-11-02 23:40:26

DataFile Name :117SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.90	0.92	0.93	0.91	1.79	ppb
Vanadium	51-2	132.73	132.41	132.20	132.45	0.20	ppb
Yttrium	89-1				99		%
Yttrium	89-2				104		%
Zinc	66-2	152.77	153.19	152.83	152.93	0.15	ppb
Zirconium	90-1	8.83	8.87	9.04	8.91	1.27	ppb
Zirconium	91-1	8.95	8.85	9.17	8.99	1.80	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-17 Instrumnet Name : P7
Client Sample ID : MYE4P7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:43:34 DataFile Name : 118SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	53422.14	52988.35	53266.97	53225.82	0.41	ppb
Antimony	121-1	1.37	1.34	1.36	1.36	1.04	ppb
Arsenic	75-2	188.65	190.47	193.88	191.00	1.39	ppb
Barium	135-1	538.14	582.14	580.45	566.91	4.40	ppb
Barium	137-1	583.49	574.60	588.43	582.17	1.20	ppb
Beryllium	9-1	1.12	1.21	1.27	1.20	6.38	ppb
Bismuth	209-1				96		%
Bismuth	209-2				96		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.48	1.17	1.26	1.30	12.30	ppb
Cadmium	106-1	1.68	0.52	0.58	0.93	70.75	ppb
Cadmium	111-1	0.45	0.33	0.36	0.38	16.57	ppb
Calcium	43-1	143367.19	141757.92	143725.70	142950.27	0.73	ppb
Calcium	44-1	143933.03	143695.01	144543.49	144057.18	0.30	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	853.86	839.00	846.60	846.48	0.88	ppb
Cobalt	59-2	131.17	130.33	130.58	130.70	0.33	ppb
Copper	63-2	216.91	216.37	216.97	216.75	0.15	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				100		%
Indium	115-1				91		%
Indium	115-2				92		%
Iron	56-2	160363.34	156716.29	157320.69	158133.44	1.24	ppb
Iron	57-2	157179.83	157675.58	156226.74	157027.38	0.47	ppb
Iron	54-2	160130.58	158386.85	159748.08	159421.84	0.57	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-17 Instrumnet Name : P7
Client Sample ID : MYE4P7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:43:34 DataFile Name : 118SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	8.48	8.72	8.49	8.56	1.63	ppb
Lead	207-1	8.27	8.41	8.34	8.34	0.82	ppb
Lead	208-1	8.33	8.47	8.39	8.40	0.82	ppb
Lithium	6-1				85		%
Magnesium	24-2	56710.20	55930.94	56502.19	56381.11	0.72	ppb
Manganese	55-2	6603.80	6498.43	6573.57	6558.60	0.83	ppb
Molybdenum	94-1	10.85	11.04	11.09	10.99	1.13	ppb
Molybdenum	95-1	1.12	1.08	1.15	1.11	2.88	ppb
Molybdenum	96-1	3.46	3.45	3.42	3.44	0.64	ppb
Molybdenum	97-1	1.02	1.01	1.04	1.02	1.42	ppb
Molybdenum	98-1	0.97	0.95	0.96	0.96	1.00	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	525.47	520.35	523.13	522.98	0.49	ppb
Phosphorus	31-2	1185.26	1157.19	1179.05	1173.83	1.26	ppb
Potassium	39-2	995.51	989.82	993.23	992.85	0.29	ppb
Rhodium	103-1				86		%
Rhodium	103-2				93		%
Scandium	45-1				98		%
Scandium	45-2				105		%
Selenium	82-1	0.62	0.98	0.38	0.66	45.71	ppb
Selenium	77-2	15.57	14.32	18.00	15.97	11.72	ppb
Selenium	78-2	2.24	1.67	1.84	1.92	15.25	ppb
Silicon	28-1	1637.74	1645.68	1638.73	1640.72	0.26	ppb
Silver	107-1	0.06	0.05	0.06	0.06	11.66	ppb
Silver	109-1	0.05	0.06	0.06	0.05	14.19	ppb
Sodium	23-2	166.79	165.67	168.81	167.09	0.95	ppb
Strontium	86-1	143.14	145.34	146.23	144.90	1.10	ppb
Strontium	88-1	149.03	149.37	152.41	150.27	1.24	ppb
Sulfur	34-1	-1100.72	-1151.31	-1233.72	-1161.92		ppb
Terbium	159-1				100		%
Terbium	159-2				101		%
Thallium	203-1	0.20	0.23	0.23	0.22	5.89	ppb
Thallium	205-1	0.23	0.22	0.22	0.22	1.64	ppb
Tin	118-1	1.64	1.68	1.69	1.67	1.35	ppb
Titanium	47-1	1208.45	1206.31	1211.70	1208.82	0.22	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-17

Instrumnet Name :

P7

Client Sample ID :

MYE4P7

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:43:34

DataFile Name :

118SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.74	0.75	0.77	0.75	1.90	ppb
Vanadium	51-2	394.24	399.29	401.24	398.26	0.91	ppb
Yttrium	89-1				98		%
Yttrium	89-2				102		%
Zinc	66-2	193.61	192.21	195.32	193.71	0.81	ppb
Zirconium	90-1	5.88	5.98	6.01	5.96	1.13	ppb
Zirconium	91-1	6.05	6.12	6.16	6.11	0.85	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-18 Instrumnet Name : P7
Client Sample ID : MYE4P8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:46:35 DataFile Name : 119SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	43804.68	43626.36	43634.50	43688.51	0.23	ppb
Antimony	121-1	1.28	1.17	1.21	1.22	4.20	ppb
Arsenic	75-2	103.69	104.04	105.95	104.56	1.16	ppb
Barium	135-1	658.65	596.99	623.74	626.46	4.94	ppb
Barium	137-1	656.28	614.45	614.04	628.26	3.86	ppb
Beryllium	9-1	2.07	1.68	1.78	1.84	11.05	ppb
Bismuth	209-1				98		%
Bismuth	209-2				101		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	3.96	3.12	3.09	3.39	14.65	ppb
Cadmium	106-1	2.94	2.55	3.52	3.00	16.14	ppb
Cadmium	111-1	2.07	2.00	2.15	2.07	3.50	ppb
Calcium	43-1	12823.16	11746.65	12016.81	12195.54	4.59	ppb
Calcium	44-1	12080.03	11134.41	11224.07	11479.50	4.55	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	99.10	99.29	97.86	98.75	0.79	ppb
Cobalt	59-2	53.48	54.25	53.73	53.82	0.73	ppb
Copper	63-2	276.61	279.02	277.12	277.58	0.46	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				99		%
Holmium	165-2				105		%
Indium	115-1				93		%
Indium	115-2				98		%
Iron	56-2	94638.05	94092.65	93230.68	93987.13	0.75	ppb
Iron	57-2	93807.07	93582.37	91988.38	93125.94	1.06	ppb
Iron	54-2	94155.55	94826.24	93085.18	94022.32	0.93	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-18 Instrumnet Name : P7
Client Sample ID : MYE4P8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:46:35 DataFile Name : 119SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	94.04	84.56	86.05	88.22	5.78	ppb
Lead	207-1	90.86	84.46	84.93	86.75	4.11	ppb
Lead	208-1	91.49	91.19	84.92	89.20	4.16	ppb
Lithium	6-1				84		%
Magnesium	24-2	21795.36	21708.48	21862.87	21788.90	0.36	ppb
Manganese	55-2	3095.21	3102.21	3087.17	3094.87	0.24	ppb
Molybdenum	94-1	26.96	24.91	25.25	25.71	4.28	ppb
Molybdenum	95-1	1.15	0.97	1.04	1.05	8.53	ppb
Molybdenum	96-1	4.60	4.19	4.25	4.35	5.13	ppb
Molybdenum	97-1	1.02	0.97	1.00	1.00	2.77	ppb
Molybdenum	98-1	1.03	0.92	0.95	0.97	5.93	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	102.71	102.59	103.42	102.90	0.43	ppb
Phosphorus	31-2	2503.84	2504.68	2475.18	2494.57	0.67	ppb
Potassium	39-2	10932.12	11031.12	10902.35	10955.20	0.62	ppb
Rhodium	103-1				89		%
Rhodium	103-2				100		%
Scandium	45-1				93		%
Scandium	45-2				105		%
Selenium	82-1	1.00	0.77	0.59	0.79	25.54	ppb
Selenium	77-2	24.52	17.19	22.26	21.32	17.62	ppb
Selenium	78-2	3.83	4.10	2.84	3.59	18.54	ppb
Silicon	28-1	2171.81	2000.62	2055.18	2075.87	4.21	ppb
Silver	107-1	0.70	0.67	0.69	0.69	1.77	ppb
Silver	109-1	0.73	0.66	0.68	0.69	5.07	ppb
Sodium	23-2	153.63	153.37	152.43	153.14	0.41	ppb
Strontium	86-1	95.65	87.50	89.74	90.96	4.62	ppb
Strontium	88-1	98.35	91.19	92.23	93.92	4.12	ppb
Sulfur	34-1	-1066.29	-1486.91	-1484.79	-1346.00		ppb
Terbium	159-1				100		%
Terbium	159-2				105		%
Thallium	203-1	0.53	0.51	0.51	0.52	2.38	ppb
Thallium	205-1	0.55	0.51	0.50	0.52	5.04	ppb
Tin	118-1	3.33	3.13	3.35	3.27	3.79	ppb
Titanium	47-1	980.34	901.19	928.79	936.77	4.29	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-18

Instrumnet Name :

P7

Client Sample ID :

MYE4P8

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:46:35

DataFile Name :

119SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.41	1.29	1.30	1.33	5.34	ppb
Vanadium	51-2	173.44	175.44	172.44	173.77	0.88	ppb
Yttrium	89-1				98		%
Yttrium	89-2				108		%
Zinc	66-2	327.81	330.17	325.90	327.96	0.65	ppb
Zirconium	90-1	15.57	14.31	14.56	14.81	4.50	ppb
Zirconium	91-1	15.72	14.47	14.43	14.87	4.94	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-19 Instrumnet Name : P7
Client Sample ID : MYE4P9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:49:40 DataFile Name : 120SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	27236.26	31681.89	26611.85	28510.00	9.70	ppb
Antimony	121-1	1.65	1.70	1.66	1.67	1.54	ppb
Arsenic	75-2	82.61	95.72	82.03	86.79	8.92	ppb
Barium	135-1	358.18	362.59	357.34	359.37	0.79	ppb
Barium	137-1	389.35	394.69	385.97	390.00	1.13	ppb
Beryllium	9-1	1.03	1.13	1.12	1.09	4.73	ppb
Bismuth	209-1				103		%
Bismuth	209-2				99		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.37	1.95	2.62	2.31	14.67	ppb
Cadmium	106-1	1.68	0.70	0.73	1.04	53.86	ppb
Cadmium	111-1	1.61	1.59	1.64	1.61	1.58	ppb
Calcium	43-1	6449.95	6459.83	6582.31	6497.36	1.13	ppb
Calcium	44-1	6019.42	6106.40	6213.37	6113.06	1.59	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	65.44	77.26	64.13	68.94	10.49	ppb
Cobalt	59-2	37.57	43.67	36.47	39.23	9.88	ppb
Copper	63-2	198.17	232.53	196.24	208.98	9.77	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				100		%
Indium	115-1				97		%
Indium	115-2				95		%
Iron	56-2	69463.08	81306.79	68557.78	73109.22	9.73	ppb
Iron	57-2	69091.96	80301.10	67813.62	72402.22	9.49	ppb
Iron	54-2	69898.57	81666.45	68929.14	73498.05	9.65	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-19 Instrumnet Name : P7
Client Sample ID : MYE4P9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:49:40 DataFile Name : 120SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	65.52	65.11	65.94	65.52	0.64	ppb
Lead	207-1	65.10	64.99	65.59	65.23	0.49	ppb
Lead	208-1	64.86	64.67	65.40	64.98	0.58	ppb
Lithium	6-1				86		%
Magnesium	24-2	14186.39	16640.67	14042.36	14956.47	9.76	ppb
Manganese	55-2	1827.99	2156.35	1832.65	1939.00	9.71	ppb
Molybdenum	94-1	14.12	14.20	14.47	14.26	1.26	ppb
Molybdenum	95-1	0.95	0.88	0.89	0.91	3.85	ppb
Molybdenum	96-1	2.88	2.94	2.88	2.90	1.18	ppb
Molybdenum	97-1	0.92	0.89	0.88	0.90	2.04	ppb
Molybdenum	98-1	0.82	0.80	0.88	0.84	5.01	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	68.59	80.71	67.54	72.28	10.13	ppb
Phosphorus	31-2	1562.84	1862.59	1538.44	1654.62	10.91	ppb
Potassium	39-2	6817.67	8053.08	6766.62	7212.46	10.10	ppb
Rhodium	103-1				94		%
Rhodium	103-2				96		%
Scandium	45-1				93		%
Scandium	45-2				95		%
Selenium	82-1	0.95	0.97	0.78	0.90	11.65	ppb
Selenium	77-2	10.62	11.48	13.00	11.70	10.31	ppb
Selenium	78-2	2.73	5.08	1.91	3.24	50.76	ppb
Silicon	28-1	1934.39	1948.81	1978.93	1954.04	1.16	ppb
Silver	107-1	0.42	0.41	0.42	0.42	1.15	ppb
Silver	109-1	0.42	0.42	0.42	0.42	0.43	ppb
Sodium	23-2	103.84	128.52	101.13	111.16	13.57	ppb
Strontium	86-1	49.03	50.22	50.68	49.98	1.71	ppb
Strontium	88-1	50.66	51.54	51.95	51.38	1.29	ppb
Sulfur	34-1	-1846.91	-1919.38	-1883.50	-1883.26		ppb
Terbium	159-1				103		%
Terbium	159-2				100		%
Thallium	203-1	0.28	0.28	0.27	0.28	0.95	ppb
Thallium	205-1	0.27	0.26	0.28	0.27	2.99	ppb
Tin	118-1	2.66	2.73	2.71	2.70	1.21	ppb
Titanium	47-1	659.71	668.70	677.48	668.63	1.33	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-19

Instrumnet Name :

P7

Client Sample ID :

MYE4P9

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:49:40

DataFile Name :

120SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.75	0.75	0.76	0.75	0.94	ppb
Vanadium	51-2	126.35	149.10	123.75	133.07	10.48	ppb
Yttrium	89-1				100		%
Yttrium	89-2				99		%
Zinc	66-2	273.43	321.45	269.71	288.20	10.01	ppb
Zirconium	90-1	7.96	8.16	8.23	8.12	1.69	ppb
Zirconium	91-1	8.05	8.30	8.34	8.23	1.91	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-20 Instrumnet Name : P7
Client Sample ID : MYE4Q0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:52:46 DataFile Name : 121SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	57279.73	57993.35	57820.15	57697.75	0.65	ppb
Antimony	121-1	1.99	2.07	2.13	2.06	3.21	ppb
Arsenic	75-2	144.80	146.69	144.05	145.18	0.93	ppb
Barium	135-1	227.89	238.50	239.67	235.35	2.76	ppb
Barium	137-1	230.21	238.87	241.88	236.99	2.56	ppb
Beryllium	9-1	1.07	1.14	1.14	1.12	3.81	ppb
Bismuth	209-1				101		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.95	3.21	2.83	3.00	6.55	ppb
Cadmium	106-1	1.42	1.06	1.53	1.34	18.26	ppb
Cadmium	111-1	2.43	2.44	2.64	2.50	4.60	ppb
Calcium	43-1	18909.26	19635.07	19796.81	19447.05	2.43	ppb
Calcium	44-1	17881.18	18346.41	18592.86	18273.48	1.98	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	55.64	55.74	56.05	55.81	0.38	ppb
Cobalt	59-2	93.12	93.31	95.91	94.11	1.65	ppb
Copper	63-2	220.25	222.06	224.99	222.43	1.07	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				105		%
Indium	115-1				98		%
Indium	115-2				99		%
Iron	56-2	143437.62	144128.74	144695.27	144087.21	0.44	ppb
Iron	57-2	142295.71	141903.81	143889.53	142696.35	0.74	ppb
Iron	54-2	142292.90	142929.80	143331.49	142851.40	0.37	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-20 Instrumnet Name : P7
Client Sample ID : MYE4Q0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:52:46 DataFile Name : 121SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	31.76	33.25	33.04	32.68	2.47	ppb
Lead	207-1	31.07	32.28	32.01	31.79	2.00	ppb
Lead	208-1	30.87	32.37	32.34	31.86	2.69	ppb
Lithium	6-1				85		%
Magnesium	24-2	24454.05	24717.64	24756.33	24642.67	0.67	ppb
Manganese	55-2	3712.89	3784.08	3827.76	3774.91	1.54	ppb
Molybdenum	94-1	5.39	5.82	5.71	5.64	3.96	ppb
Molybdenum	95-1	1.23	1.40	1.42	1.35	7.51	ppb
Molybdenum	96-1	2.90	3.09	3.08	3.02	3.53	ppb
Molybdenum	97-1	1.27	1.35	1.39	1.34	4.67	ppb
Molybdenum	98-1	1.19	1.26	1.27	1.24	3.43	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	45.18	45.89	46.74	45.94	1.70	ppb
Phosphorus	31-2	1348.88	1442.89	1403.26	1398.34	3.38	ppb
Potassium	39-2	2374.40	2381.56	2426.30	2394.09	1.17	ppb
Rhodium	103-1				94		%
Rhodium	103-2				100		%
Scandium	45-1				97		%
Scandium	45-2				103		%
Selenium	82-1	0.13	0.53	0.15	0.27	84.22	ppb
Selenium	77-2	7.03	8.16	7.54	7.58	7.44	ppb
Selenium	78-2	1.84	2.20	1.77	1.94	11.94	ppb
Silicon	28-1	1796.71	1871.18	1881.89	1849.93	2.51	ppb
Silver	107-1	1.07	1.09	1.14	1.10	3.39	ppb
Silver	109-1	1.05	1.12	1.10	1.09	3.52	ppb
Sodium	23-2	619.72	620.54	626.19	622.15	0.57	ppb
Strontium	86-1	114.58	120.08	120.85	118.50	2.88	ppb
Strontium	88-1	118.55	122.76	125.69	122.33	2.93	ppb
Sulfur	34-1	-2161.96	-2025.59	-1997.74	-2061.77		ppb
Terbium	159-1				105		%
Terbium	159-2				104		%
Thallium	203-1	0.18	0.20	0.21	0.20	7.09	ppb
Thallium	205-1	0.19	0.21	0.19	0.20	3.67	ppb
Tin	118-1	1.64	1.77	1.76	1.72	4.23	ppb
Titanium	47-1	1867.63	1946.79	1971.29	1928.57	2.81	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-20 Instrumnet Name : P7
Client Sample ID : MYE4Q0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:52:46 DataFile Name : 121SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.47	0.48	0.48	0.48	1.10	ppb
Vanadium	51-2	236.81	238.30	238.65	237.92	0.41	ppb
Yttrium	89-1				102		%
Yttrium	89-2				105		%
Zinc	66-2	564.81	566.47	574.27	568.52	0.89	ppb
Zirconium	90-1	2.57	2.82	2.72	2.70	4.61	ppb
Zirconium	91-1	2.69	2.66	2.84	2.73	3.42	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-21 Instrumnet Name : P7
Client Sample ID : MYE4Q1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:55:54 DataFile Name : 122SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	33497.14	33801.01	33675.71	33657.95	0.45	ppb
Antimony	121-1	1.36	1.35	1.38	1.37	1.16	ppb
Arsenic	75-2	106.02	107.17	106.49	106.56	0.54	ppb
Barium	135-1	461.98	471.29	467.31	466.86	1.00	ppb
Barium	137-1	502.23	502.20	504.83	503.09	0.30	ppb
Beryllium	9-1	1.34	1.41	1.48	1.41	4.86	ppb
Bismuth	209-1				103		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.85	2.90	2.28	2.68	12.95	ppb
Cadmium	106-1	0.83	1.03	1.47	1.11	29.41	ppb
Cadmium	111-1	1.48	1.49	1.53	1.50	1.51	ppb
Calcium	43-1	7323.94	7531.03	7496.53	7450.50	1.49	ppb
Calcium	44-1	6936.79	6998.03	6993.62	6976.15	0.49	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	82.14	81.55	81.71	81.80	0.37	ppb
Cobalt	59-2	49.37	49.11	49.53	49.34	0.43	ppb
Copper	63-2	246.10	244.84	247.75	246.23	0.59	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				107		%
Indium	115-1				97		%
Indium	115-2				100		%
Iron	56-2	84144.78	83787.79	84530.15	84154.24	0.44	ppb
Iron	57-2	83735.09	83320.78	84655.69	83903.85	0.81	ppb
Iron	54-2	85015.11	84100.42	84567.76	84561.10	0.54	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-21 Instrumnet Name : P7
Client Sample ID : MYE4Q1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:55:54 DataFile Name : 122SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	66.89	67.57	68.61	67.69	1.28	ppb
Lead	207-1	66.21	66.94	68.19	67.11	1.49	ppb
Lead	208-1	66.23	67.00	68.11	67.11	1.41	ppb
Lithium	6-1				86		%
Magnesium	24-2	17998.23	18015.03	17947.91	17987.06	0.19	ppb
Manganese	55-2	2458.45	2458.79	2474.19	2463.81	0.36	ppb
Molybdenum	94-1	19.70	19.72	20.21	19.88	1.48	ppb
Molybdenum	95-1	0.94	0.88	0.98	0.93	5.79	ppb
Molybdenum	96-1	3.63	3.62	3.63	3.63	0.16	ppb
Molybdenum	97-1	0.92	0.92	0.92	0.92	0.18	ppb
Molybdenum	98-1	0.86	0.88	0.87	0.87	0.94	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	89.90	90.26	90.69	90.28	0.44	ppb
Phosphorus	31-2	1580.79	1614.27	1622.44	1605.83	1.37	ppb
Potassium	39-2	6898.53	6879.66	6835.76	6871.31	0.47	ppb
Rhodium	103-1				93		%
Rhodium	103-2				100		%
Scandium	45-1				94		%
Scandium	45-2				101		%
Selenium	82-1	0.32	0.44	0.26	0.34	26.94	ppb
Selenium	77-2	14.02	17.74	17.41	16.39	12.56	ppb
Selenium	78-2	2.15	2.65	2.60	2.46	11.15	ppb
Silicon	28-1	1994.44	2024.25	2007.67	2008.78	0.74	ppb
Silver	107-1	0.46	0.47	0.47	0.47	1.47	ppb
Silver	109-1	0.47	0.46	0.45	0.46	2.35	ppb
Sodium	23-2	103.79	102.80	103.49	103.36	0.49	ppb
Strontium	86-1	60.34	60.34	61.10	60.59	0.73	ppb
Strontium	88-1	61.74	63.08	62.87	62.56	1.15	ppb
Sulfur	34-1	-2010.31	-2112.08	-2115.20	-2079.20		ppb
Terbium	159-1				103		%
Terbium	159-2				106		%
Thallium	203-1	0.37	0.38	0.38	0.38	1.59	ppb
Thallium	205-1	0.37	0.38	0.38	0.37	0.84	ppb
Tin	118-1	2.37	2.44	2.37	2.39	1.63	ppb
Titanium	47-1	801.56	809.90	808.76	806.74	0.56	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-21 Instrumnet Name : P7
Client Sample ID : MYE4Q1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:55:54 DataFile Name : 122SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.03	1.04	1.04	1.03	0.23	ppb
Vanadium	51-2	165.79	164.55	166.32	165.55	0.55	ppb
Yttrium	89-1				100		%
Yttrium	89-2				105		%
Zinc	66-2	227.48	224.70	229.33	227.17	1.03	ppb
Zirconium	90-1	11.37	11.50	11.57	11.48	0.91	ppb
Zirconium	91-1	11.35	11.45	11.68	11.49	1.49	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 23:58:59 DataFile Name : 123SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	8312.11	7483.87	7593.97	7796.65	5.77	ppb
Antimony	121-1	0.09	0.10	0.10	0.10	4.83	ppb
Arsenic	75-2	10.45	9.84	9.69	9.99	4.04	ppb
Barium	135-1	74.82	76.37	76.67	75.95	1.31	ppb
Barium	137-1	74.49	75.93	76.92	75.78	1.61	ppb
Beryllium	9-1	0.28	0.24	0.22	0.25	12.48	ppb
Bismuth	209-1				105		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.26	0.28	0.24	0.26	6.70	ppb
Cadmium	106-1	-1.82	-2.21	-2.05	-2.03		ppb
Cadmium	111-1	-0.02	-0.03	-0.01	-0.02		ppb
Calcium	43-1	6319.27	6504.47	6503.41	6442.38	1.65	ppb
Calcium	44-1	5955.78	6119.86	6085.06	6053.57	1.43	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	18.51	16.85	17.00	17.46	5.26	ppb
Cobalt	59-2	10.71	9.64	9.81	10.05	5.73	ppb
Copper	63-2	33.35	30.07	30.35	31.26	5.82	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				104		%
Indium	115-1				98		%
Indium	115-2				97		%
Iron	56-2	19279.92	17131.49	17565.33	17992.25	6.31	ppb
Iron	57-2	19428.31	17150.24	17467.26	18015.27	6.85	ppb
Iron	54-2	19261.19	17026.31	17245.06	17844.19	6.90	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 23:58:59 DataFile Name : 123SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	7.56	7.42	7.72	7.56	1.97	ppb
Lead	207-1	7.17	7.16	7.15	7.16	0.18	ppb
Lead	208-1	7.27	7.27	7.36	7.30	0.72	ppb
Lithium	6-1				86		%
Magnesium	24-2	4482.47	4003.67	4073.58	4186.57	6.18	ppb
Manganese	55-2	580.79	516.19	525.00	540.66	6.48	ppb
Molybdenum	94-1	2.22	2.23	2.21	2.22	0.36	ppb
Molybdenum	95-1	0.22	0.21	0.20	0.21	3.05	ppb
Molybdenum	96-1	0.54	0.55	0.56	0.55	2.06	ppb
Molybdenum	97-1	0.16	0.19	0.23	0.19	16.15	ppb
Molybdenum	98-1	0.17	0.18	0.19	0.18	7.17	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	15.56	14.14	14.01	14.57	5.88	ppb
Phosphorus	31-2	636.45	553.63	553.53	581.20	8.23	ppb
Potassium	39-2	2945.89	2515.92	2519.78	2660.53	9.29	ppb
Rhodium	103-1				94		%
Rhodium	103-2				99		%
Scandium	45-1				88		%
Scandium	45-2				94		%
Selenium	82-1	0.24	-0.27	0.12	0.03	808.27	ppb
Selenium	77-2	4.00	4.97	6.05	5.01	20.53	ppb
Selenium	78-2	0.96	1.27	0.05	0.76	82.92	ppb
Silicon	28-1	525.62	540.30	539.07	535.00	1.52	ppb
Silver	107-1	0.05	0.05	0.05	0.05	8.54	ppb
Silver	109-1	0.04	0.05	0.05	0.05	10.05	ppb
Sodium	23-2	101.28	89.43	90.45	93.72	7.01	ppb
Strontium	86-1	24.54	25.15	25.14	24.94	1.41	ppb
Strontium	88-1	24.60	24.95	25.01	24.85	0.90	ppb
Sulfur	34-1	-1940.91	-1902.98	-1873.63	-1905.84		ppb
Terbium	159-1				104		%
Terbium	159-2				102		%
Thallium	203-1	0.04	0.04	0.04	0.04	12.04	ppb
Thallium	205-1	0.04	0.04	0.05	0.04	4.25	ppb
Tin	118-1	0.40	0.46	0.45	0.44	6.98	ppb
Titanium	47-1	233.24	239.73	239.73	237.57	1.58	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-03DLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4K2

Dilution Factor :

4

Date & Time Acquired :

2024-11-02 23:58:59

DataFile Name :

123SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.23	0.23	0.23	0.23	0.30	ppb
Vanadium	51-2	36.34	32.54	33.55	34.14	5.76	ppb
Yttrium	89-1				95		%
Yttrium	89-2				99		%
Zinc	66-2	46.72	42.33	43.14	44.06	5.29	ppb
Zirconium	90-1	1.19	1.23	1.22	1.21	1.77	ppb
Zirconium	91-1	1.21	1.18	1.20	1.20	1.65	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-04DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2D Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:02:10 DataFile Name : 124SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	7620.62	7433.59	7557.33	7537.18	1.26	ppb
Antimony	121-1	0.08	0.10	0.10	0.10	14.17	ppb
Arsenic	75-2	9.79	9.61	9.47	9.62	1.69	ppb
Barium	135-1	72.77	75.20	74.52	74.16	1.69	ppb
Barium	137-1	72.82	75.18	75.34	74.45	1.90	ppb
Beryllium	9-1	0.27	0.22	0.25	0.25	11.34	ppb
Bismuth	209-1				103		%
Bismuth	209-2				106		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.26	0.25	0.22	0.24	6.95	ppb
Cadmium	106-1	-2.29	-2.43	-2.68	-2.46		ppb
Cadmium	111-1	-0.02	-0.05	-0.08	-0.05		ppb
Calcium	43-1	6154.85	6341.14	6302.65	6266.21	1.57	ppb
Calcium	44-1	5865.23	6028.98	5980.25	5958.15	1.41	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	16.82	16.93	16.71	16.82	0.65	ppb
Cobalt	59-2	9.75	9.76	9.75	9.76	0.04	ppb
Copper	63-2	29.89	30.08	30.02	29.99	0.32	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				106		%
Indium	115-1				98		%
Indium	115-2				101		%
Iron	56-2	17276.46	17309.32	17195.17	17260.32	0.34	ppb
Iron	57-2	16965.85	17218.84	17323.31	17169.33	1.07	ppb
Iron	54-2	17271.96	17472.31	17276.28	17340.18	0.66	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-04DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2D Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:02:10 DataFile Name : 124SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	7.15	7.27	7.58	7.33	3.05	ppb
Lead	207-1	6.78	6.82	7.26	6.96	3.85	ppb
Lead	208-1	6.97	7.02	7.33	7.10	2.71	ppb
Lithium	6-1				86		%
Magnesium	24-2	4048.01	4038.75	3996.46	4027.74	0.68	ppb
Manganese	55-2	519.82	518.69	520.86	519.79	0.21	ppb
Molybdenum	94-1	2.08	2.13	2.22	2.15	3.39	ppb
Molybdenum	95-1	0.16	0.19	0.21	0.19	11.86	ppb
Molybdenum	96-1	0.51	0.53	0.55	0.53	2.99	ppb
Molybdenum	97-1	0.16	0.21	0.19	0.19	15.18	ppb
Molybdenum	98-1	0.18	0.16	0.19	0.18	7.13	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	14.34	14.27	14.13	14.25	0.73	ppb
Phosphorus	31-2	548.56	560.80	545.47	551.61	1.47	ppb
Potassium	39-2	2621.43	2706.92	2652.53	2660.30	1.63	ppb
Rhodium	103-1				94		%
Rhodium	103-2				102		%
Scandium	45-1				89		%
Scandium	45-2				95		%
Selenium	82-1	-0.43	0.50	-0.07	0.00	25765.75	ppb
Selenium	77-2	3.85	3.72	3.21	3.59	9.36	ppb
Selenium	78-2	0.53	1.31	-0.10	0.58	121.48	ppb
Silicon	28-1	507.89	524.49	527.95	520.11	2.06	ppb
Silver	107-1	0.04	0.04	0.05	0.05	3.70	ppb
Silver	109-1	0.05	0.05	0.06	0.05	13.50	ppb
Sodium	23-2	87.29	88.42	87.64	87.79	0.66	ppb
Strontium	86-1	24.00	24.85	24.58	24.48	1.79	ppb
Strontium	88-1	23.89	24.45	24.55	24.29	1.47	ppb
Sulfur	34-1	-1798.20	-1783.51	-1708.76	-1763.49		ppb
Terbium	159-1				104		%
Terbium	159-2				106		%
Thallium	203-1	0.04	0.05	0.05	0.05	8.53	ppb
Thallium	205-1	0.04	0.04	0.04	0.04	3.11	ppb
Tin	118-1	0.45	0.43	0.42	0.43	3.89	ppb
Titanium	47-1	228.08	233.92	233.55	231.85	1.41	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-04DLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4K2D

Dilution Factor :

4

Date & Time Acquired :

2024-11-03 00:02:10

DataFile Name :

124SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.22	0.23	0.23	0.23	3.11	ppb
Vanadium	51-2	33.23	33.21	33.43	33.29	0.36	ppb
Yttrium	89-1				96		%
Yttrium	89-2				101		%
Zinc	66-2	42.92	43.05	42.69	42.89	0.43	ppb
Zirconium	90-1	1.15	1.18	1.20	1.18	2.25	ppb
Zirconium	91-1	1.10	1.20	1.20	1.17	5.02	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4K2L Dilution Factor : 20
Date & Time Acquired : 2024-11-03 00:05:18 DataFile Name : 125SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1387.70	1408.72	1380.96	1392.46	1.04	ppb
Antimony	121-1	0.01	0.01	0.01	0.01	18.47	ppb
Arsenic	75-2	1.76	1.82	1.60	1.73	6.50	ppb
Barium	135-1	13.09	13.45	14.07	13.54	3.69	ppb
Barium	137-1	13.33	13.34	13.85	13.51	2.19	ppb
Beryllium	9-1	0.01	0.03	0.01	0.02	63.18	ppb
Bismuth	209-1				102		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.08	0.06	0.05	0.07	25.22	ppb
Cadmium	106-1	-2.63	-2.61	-2.38	-2.54		ppb
Cadmium	111-1	-0.16	-0.18	-0.14	-0.16		ppb
Calcium	43-1	1153.16	1134.66	1213.20	1167.01	3.52	ppb
Calcium	44-1	1154.26	1141.08	1181.75	1159.03	1.79	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3.16	3.12	3.15	3.14	0.65	ppb
Cobalt	59-2	1.83	1.83	1.79	1.82	1.31	ppb
Copper	63-2	5.40	5.66	5.41	5.49	2.73	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				104		%
Indium	115-1				97		%
Indium	115-2				101		%
Iron	56-2	3165.59	3209.76	3147.79	3174.38	1.01	ppb
Iron	57-2	3171.41	3242.35	3208.45	3207.41	1.11	ppb
Iron	54-2	3272.65	3299.60	3240.70	3270.98	0.90	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4K2L Dilution Factor : 20
Date & Time Acquired : 2024-11-03 00:05:18 DataFile Name : 125SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1.29	1.27	1.28	1.28	0.95	ppb
Lead	207-1	1.22	1.26	1.25	1.24	1.88	ppb
Lead	208-1	1.23	1.25	1.26	1.25	1.18	ppb
Lithium	6-1				87		%
Magnesium	24-2	759.13	758.66	747.82	755.20	0.85	ppb
Manganese	55-2	94.96	95.12	93.56	94.55	0.91	ppb
Molybdenum	94-1	0.36	0.34	0.34	0.35	2.96	ppb
Molybdenum	95-1	-0.01	-0.01	-0.02	-0.01		ppb
Molybdenum	96-1	0.04	0.07	0.05	0.05	32.22	ppb
Molybdenum	97-1	-0.02	-0.02	0.00	-0.01		ppb
Molybdenum	98-1	-0.02	-0.02	-0.02	-0.02		ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	2.86	2.61	2.58	2.68	5.57	ppb
Phosphorus	31-2	94.62	102.24	90.31	95.72	6.31	ppb
Potassium	39-2	479.25	485.66	473.28	479.40	1.29	ppb
Rhodium	103-1				96		%
Rhodium	103-2				102		%
Scandium	45-1				88		%
Scandium	45-2				93		%
Selenium	82-1	-0.31	-0.37	-0.17	-0.29		ppb
Selenium	77-2	0.01	0.49	0.72	0.40	89.34	ppb
Selenium	78-2	-1.29	-0.63	0.41	-0.50		ppb
Silicon	28-1	95.95	95.48	100.01	97.15	2.56	ppb
Silver	107-1	0.00	0.00	0.00	0.00	102.57	ppb
Silver	109-1	0.01	0.00	0.00	0.00	62.38	ppb
Sodium	23-2	5.00	4.13	4.89	4.67	10.10	ppb
Strontium	86-1	4.27	4.53	4.48	4.43	3.09	ppb
Strontium	88-1	4.34	4.35	4.44	4.37	1.25	ppb
Sulfur	34-1	-1860.05	-1895.93	-1755.28	-1837.09		ppb
Terbium	159-1				102		%
Terbium	159-2				104		%
Thallium	203-1	0.00	0.00	0.00	0.00		ppb
Thallium	205-1	0.00	0.00	0.00	0.00		ppb
Tin	118-1	0.05	0.05	0.07	0.06	20.98	ppb
Titanium	47-1	42.39	41.83	44.00	42.74	2.64	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-03LDLX20

Instrumnet Name :

P7

Client Sample ID :

MYE4K2L

Dilution Factor :

20

Date & Time Acquired :

2024-11-03 00:05:18

DataFile Name :

125SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.04	0.03	0.04	0.04	2.92	ppb
Vanadium	51-2	6.08	6.15	6.02	6.08	1.01	ppb
Yttrium	89-1				94		%
Yttrium	89-2				99		%
Zinc	66-2	7.95	8.13	7.33	7.81	5.43	ppb
Zirconium	90-1	0.20	0.19	0.21	0.20	4.48	ppb
Zirconium	91-1	0.22	0.21	0.21	0.21	1.79	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-05DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2S Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:08:32 DataFile Name : 126SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	7673.55	7810.91	7733.63	7739.36	0.89	ppb
Antimony	121-1	13.67	14.63	13.89	14.06	3.58	ppb
Arsenic	75-2	15.35	15.45	14.85	15.22	2.11	ppb
Barium	135-1	335.07	359.08	333.28	342.47	4.21	ppb
Barium	137-1	367.62	383.52	366.09	372.41	2.59	ppb
Beryllium	9-1	7.39	8.00	8.06	7.81	4.71	ppb
Bismuth	209-1				102		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	7.60	8.13	6.53	7.42	10.96	ppb
Cadmium	106-1	5.72	4.16	4.92	4.93	15.84	ppb
Cadmium	111-1	7.10	7.27	6.85	7.08	2.99	ppb
Calcium	43-1	6185.43	6635.71	6267.03	6362.72	3.77	ppb
Calcium	44-1	5862.81	6375.07	5873.24	6037.04	4.85	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	46.79	47.77	46.60	47.06	1.34	ppb
Cobalt	59-2	84.69	85.73	83.72	84.72	1.19	ppb
Copper	63-2	64.96	65.76	65.54	65.42	0.63	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				104		%
Indium	115-1				96		%
Indium	115-2				100		%
Iron	56-2	16992.46	17344.04	17265.73	17200.74	1.07	ppb
Iron	57-2	17013.57	17412.19	17525.35	17317.04	1.55	ppb
Iron	54-2	17447.92	17384.69	17380.46	17404.35	0.22	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-05DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2S Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:08:32 DataFile Name : 126SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	9.54	10.63	9.56	9.91	6.27	ppb
Lead	207-1	9.10	10.18	9.19	9.49	6.30	ppb
Lead	208-1	9.24	10.30	9.32	9.62	6.15	ppb
Lithium	6-1				85		%
Magnesium	24-2	4005.20	4059.13	4070.52	4044.95	0.86	ppb
Manganese	55-2	581.72	597.81	588.60	589.38	1.37	ppb
Molybdenum	94-1	2.05	2.28	2.09	2.14	5.72	ppb
Molybdenum	95-1	0.18	0.22	0.19	0.20	11.29	ppb
Molybdenum	96-1	0.51	0.54	0.56	0.54	4.42	ppb
Molybdenum	97-1	0.21	0.19	0.18	0.19	8.17	ppb
Molybdenum	98-1	0.17	0.19	0.17	0.17	7.30	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	89.09	90.44	87.26	88.93	1.79	ppb
Phosphorus	31-2	566.36	537.43	565.56	556.45	2.96	ppb
Potassium	39-2	2673.66	2682.33	2622.13	2659.38	1.22	ppb
Rhodium	103-1				92		%
Rhodium	103-2				100		%
Scandium	45-1				87		%
Scandium	45-2				94		%
Selenium	82-1	13.65	15.34	14.08	14.36	6.11	ppb
Selenium	77-2	14.94	15.99	15.87	15.60	3.68	ppb
Selenium	78-2	12.83	13.68	13.89	13.47	4.18	ppb
Silicon	28-1	513.79	553.24	516.21	527.75	4.19	ppb
Silver	107-1	5.83	6.46	5.94	6.08	5.57	ppb
Silver	109-1	6.00	6.43	6.00	6.14	4.02	ppb
Sodium	23-2	83.85	85.79	85.40	85.01	1.20	ppb
Strontium	86-1	23.79	25.76	23.76	24.44	4.69	ppb
Strontium	88-1	23.71	25.76	23.75	24.41	4.80	ppb
Sulfur	34-1	-1645.30	-1522.46	-1743.68	-1637.15		ppb
Terbium	159-1				100		%
Terbium	159-2				105		%
Thallium	203-1	6.03	6.55	6.05	6.21	4.75	ppb
Thallium	205-1	5.98	6.60	5.96	6.18	5.91	ppb
Tin	118-1	0.41	0.47	0.43	0.44	6.75	ppb
Titanium	47-1	226.76	247.16	229.77	234.56	4.69	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-05DLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4K2S

Dilution Factor :

4

Date & Time Acquired :

2024-11-03 00:08:32

DataFile Name :

126SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.22	0.25	0.22	0.23	5.74	ppb
Vanadium	51-2	107.43	108.11	105.87	107.14	1.07	ppb
Yttrium	89-1				93		%
Yttrium	89-2				99		%
Zinc	66-2	115.42	117.05	115.87	116.11	0.72	ppb
Zirconium	90-1	1.14	1.24	1.14	1.17	4.71	ppb
Zirconium	91-1	1.19	1.36	1.07	1.21	12.34	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:11:39 DataFile Name : 127SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	8434.12	8448.00	8424.75	8435.62	0.14	ppb
Antimony	121-1	65.28	66.32	66.63	66.08	1.07	ppb
Arsenic	75-2	36.70	36.39	36.06	36.38	0.87	ppb
Barium	135-1	1437.90	1468.97	1469.64	1458.84	1.24	ppb
Barium	137-1	1436.61	1466.35	1456.84	1453.27	1.04	ppb
Beryllium	9-1	35.49	36.21	36.24	35.98	1.17	ppb
Bismuth	209-1				102		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	31.36	31.02	31.91	31.43	1.43	ppb
Cadmium	106-1	29.00	28.03	29.89	28.98	3.22	ppb
Cadmium	111-1	33.19	33.76	33.95	33.63	1.18	ppb
Calcium	43-1	5978.06	6092.00	6208.47	6092.84	1.89	ppb
Calcium	44-1	5710.13	5772.41	5844.93	5775.82	1.17	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	162.24	162.28	161.85	162.12	0.15	ppb
Cobalt	59-2	384.91	378.55	376.89	380.12	1.11	ppb
Copper	63-2	197.68	198.79	196.62	197.70	0.55	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				103		%
Indium	115-1				96		%
Indium	115-2				97		%
Iron	56-2	17291.25	17261.88	17033.80	17195.65	0.82	ppb
Iron	57-2	17235.14	16975.41	16807.33	17005.96	1.27	ppb
Iron	54-2	17618.68	17323.71	17388.10	17443.50	0.89	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:11:39 DataFile Name : 127SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	18.90	19.05	19.49	19.15	1.58	ppb
Lead	207-1	18.33	18.50	18.73	18.52	1.08	ppb
Lead	208-1	18.55	18.78	18.91	18.75	0.96	ppb
Lithium	6-1				86		%
Magnesium	24-2	3832.56	3861.00	3912.58	3868.71	1.05	ppb
Manganese	55-2	862.00	855.10	838.51	851.87	1.42	ppb
Molybdenum	94-1	2.07	2.01	2.13	2.07	2.92	ppb
Molybdenum	95-1	0.19	0.19	0.20	0.19	4.32	ppb
Molybdenum	96-1	0.52	0.50	0.56	0.53	5.24	ppb
Molybdenum	97-1	0.17	0.17	0.17	0.17	1.81	ppb
Molybdenum	98-1	0.17	0.17	0.18	0.17	3.52	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	372.66	372.31	374.10	373.02	0.25	ppb
Phosphorus	31-2	538.38	562.11	554.44	551.64	2.19	ppb
Potassium	39-2	2583.22	2556.41	2552.82	2564.15	0.65	ppb
Rhodium	103-1				93		%
Rhodium	103-2				100		%
Scandium	45-1				87		%
Scandium	45-2				93		%
Selenium	82-1	75.53	77.62	79.32	77.49	2.45	ppb
Selenium	77-2	80.82	76.55	75.52	77.63	3.62	ppb
Selenium	78-2	82.04	79.17	78.91	80.04	2.17	ppb
Silicon	28-1	498.43	512.65	514.16	508.41	1.71	ppb
Silver	107-1	28.57	29.25	29.17	29.00	1.28	ppb
Silver	109-1	28.54	28.99	29.32	28.95	1.35	ppb
Sodium	23-2	82.40	84.27	83.58	83.42	1.13	ppb
Strontium	86-1	22.90	23.40	24.06	23.46	2.48	ppb
Strontium	88-1	22.96	23.32	23.66	23.31	1.50	ppb
Sulfur	34-1	-1685.20	-1662.84	-1697.71	-1681.92		ppb
Terbium	159-1				101		%
Terbium	159-2				103		%
Thallium	203-1	29.20	29.54	29.58	29.44	0.72	ppb
Thallium	205-1	29.05	29.33	29.53	29.30	0.81	ppb
Tin	118-1	0.41	0.46	0.42	0.43	6.18	ppb
Titanium	47-1	222.85	226.67	228.49	226.00	1.27	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-03MSDLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4K2SRE

Dilution Factor :

4

Date & Time Acquired :

2024-11-03 00:11:39

DataFile Name :

127SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.22	0.23	0.22	0.22	2.03	ppb
Vanadium	51-2	386.92	380.53	382.32	383.25	0.86	ppb
Yttrium	89-1				94		%
Yttrium	89-2				98		%
Zinc	66-2	394.31	392.69	388.87	391.96	0.71	ppb
Zirconium	90-1	1.10	1.13	1.15	1.12	2.38	ppb
Zirconium	91-1	1.17	1.17	1.15	1.16	0.98	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BL Instrumnet Name : P7
Client Sample ID : PBS452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:14:44 DataFile Name : 128CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	3.07	4.01	4.43	3.84	18.17	ppb
Antimony	121-1	0.03	0.05	0.03	0.04	24.79	ppb
Arsenic	75-2	0.01	-0.01	0.01	0.01	205.67	ppb
Barium	135-1	0.14	0.13	0.13	0.13	4.75	ppb
Barium	137-1	0.10	0.14	0.12	0.12	13.83	ppb
Beryllium	9-1	-0.02	-0.02	-0.02	-0.02		ppb
Bismuth	209-1				95		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.07	0.02	0.10	0.06	66.27	ppb
Cadmium	106-1	-2.14	-2.10	-2.89	-2.38		ppb
Cadmium	111-1	-0.16	-0.15	-0.21	-0.17		ppb
Calcium	43-1	-0.43	3.55	3.00	2.04	105.94	ppb
Calcium	44-1	-1.34	2.72	-2.12	-0.25		ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.04	0.04	0.05	0.04	11.99	ppb
Cobalt	59-2	0.02	0.02	0.01	0.02	25.45	ppb
Copper	63-2	-0.04	-0.02	-0.01	-0.02		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				94		%
Holmium	165-2				103		%
Indium	115-1				90		%
Indium	115-2				99		%
Iron	56-2	17.72	19.24	19.97	18.98	6.05	ppb
Iron	57-2	15.40	19.81	18.35	17.85	12.60	ppb
Iron	54-2	18.24	19.21	20.06	19.17	4.76	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BL Instrumnet Name : P7
Client Sample ID : PBS452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:14:44 DataFile Name : 128CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	-0.05	-0.04	-0.04	-0.04		ppb
Lead	207-1	-0.05	-0.03	-0.05	-0.04		ppb
Lead	208-1	-0.05	-0.04	-0.05	-0.04		ppb
Lithium	6-1				82		%
Magnesium	24-2	-0.59	-0.63	-0.47	-0.57		ppb
Manganese	55-2	0.28	0.39	0.40	0.36	19.02	ppb
Molybdenum	94-1	-0.07	-0.06	-0.04	-0.06		ppb
Molybdenum	95-1	-0.05	-0.05	-0.05	-0.05		ppb
Molybdenum	96-1	-0.06	-0.05	-0.06	-0.06		ppb
Molybdenum	97-1	-0.06	-0.04	-0.06	-0.05		ppb
Molybdenum	98-1	-0.06	-0.05	-0.06	-0.06		ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.07	0.09	0.07	0.08	14.29	ppb
Phosphorus	31-2	-24.18	-22.04	-21.39	-22.54		ppb
Potassium	39-2	-11.85	-11.59	-13.57	-12.33		ppb
Rhodium	103-1				88		%
Rhodium	103-2				100		%
Scandium	45-1				81		%
Scandium	45-2				93		%
Selenium	82-1	-0.38	-0.38	-0.50	-0.42		ppb
Selenium	77-2	-0.22	-0.22	-0.22	-0.22		ppb
Selenium	78-2	-1.24	-1.86	-1.05	-1.39		ppb
Silicon	28-1	-0.14	2.94	0.24	1.01	165.82	ppb
Silver	107-1	0.02	0.02	0.01	0.01	40.16	ppb
Silver	109-1	0.03	0.02	0.01	0.02	38.65	ppb
Sodium	23-2	-15.84	-15.41	-15.60	-15.62		ppb
Strontium	86-1	-0.09	-0.02	-0.08	-0.06		ppb
Strontium	88-1	-0.10	-0.08	-0.09	-0.09		ppb
Sulfur	34-1	-1851.88	-1128.81	-1822.78	-1601.16		ppb
Terbium	159-1				94		%
Terbium	159-2				103		%
Thallium	203-1	-0.01	0.00	0.00	0.00		ppb
Thallium	205-1	-0.01	0.00	0.00	0.00		ppb
Tin	118-1	-0.01	0.04	0.01	0.01	173.34	ppb
Titanium	47-1	0.11	0.13	0.09	0.11	21.39	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

PB164452BL

Instrumnet Name :

P7

Client Sample ID :

PBS452

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:14:44

DataFile Name :

128CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	-0.01	-0.01	-0.01	-0.01		ppb
Vanadium	51-2	0.05	0.03	0.04	0.04	21.99	ppb
Yttrium	89-1				87		%
Yttrium	89-2				98		%
Zinc	66-2	-0.11	0.01	0.04	-0.02		ppb
Zirconium	90-1	-0.01	-0.01	-0.01	-0.01		ppb
Zirconium	91-1	-0.01	-0.01	-0.01	-0.01		ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BS Instrumnet Name : P7
Client Sample ID : LCS452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:17:59 DataFile Name : 129LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	37.78	38.34	39.24	38.45	1.91	ppb
Antimony	121-1	3.60	3.66	3.46	3.58	2.95	ppb
Arsenic	75-2	1.77	2.15	2.11	2.01	10.62	ppb
Barium	135-1	17.08	17.31	16.35	16.91	2.97	ppb
Barium	137-1	17.21	17.49	16.57	17.09	2.75	ppb
Beryllium	9-1	1.70	1.84	1.83	1.79	4.24	ppb
Bismuth	209-1				102		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.70	2.19	1.72	1.87	14.77	ppb
Cadmium	106-1	-0.06	-0.14	-1.15	-0.45		ppb
Cadmium	111-1	1.55	1.51	1.54	1.54	1.51	ppb
Calcium	43-1	939.67	934.83	904.19	926.23	2.08	ppb
Calcium	44-1	938.41	949.84	905.91	931.39	2.45	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	4.04	4.00	4.18	4.08	2.32	ppb
Cobalt	59-2	2.05	2.05	1.99	2.03	1.67	ppb
Copper	63-2	3.86	4.03	3.92	3.94	2.22	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				101		%
Indium	115-1				96		%
Indium	115-2				98		%
Iron	56-2	423.45	424.89	421.40	423.25	0.41	ppb
Iron	57-2	428.34	432.62	431.73	430.90	0.52	ppb
Iron	54-2	444.36	434.76	444.96	441.36	1.30	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BS Instrumnet Name : P7
Client Sample ID : LCS452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:17:59 DataFile Name : 129LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1.52	1.63	1.50	1.55	4.67	ppb
Lead	207-1	1.57	1.60	1.45	1.54	4.95	ppb
Lead	208-1	1.55	1.60	1.48	1.54	3.76	ppb
Lithium	6-1				88		%
Magnesium	24-2	996.72	1004.77	1004.95	1002.15	0.47	ppb
Manganese	55-2	2.31	2.17	2.11	2.20	4.83	ppb
Molybdenum	94-1	10.15	10.47	9.69	10.10	3.87	ppb
Molybdenum	95-1	8.47	8.50	8.00	8.32	3.35	ppb
Molybdenum	96-1	8.40	8.65	8.16	8.40	2.91	ppb
Molybdenum	97-1	8.38	8.61	8.16	8.38	2.70	ppb
Molybdenum	98-1	8.25	8.42	8.03	8.23	2.36	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	2.09	2.01	2.11	2.07	2.39	ppb
Phosphorus	31-2	-21.54	-20.43	-23.48	-21.82		ppb
Potassium	39-2	967.78	969.97	970.02	969.26	0.13	ppb
Rhodium	103-1				94		%
Rhodium	103-2				100		%
Scandium	45-1				87		%
Scandium	45-2				92		%
Selenium	82-1	9.63	9.49	8.63	9.25	5.83	ppb
Selenium	77-2	10.24	9.48	13.03	10.91	17.14	ppb
Selenium	78-2	8.90	7.70	7.40	8.00	9.91	ppb
Silicon	28-1	5.50	5.81	5.11	5.47	6.40	ppb
Silver	107-1	1.74	1.81	1.71	1.75	2.98	ppb
Silver	109-1	1.72	1.78	1.73	1.74	2.04	ppb
Sodium	23-2	926.77	928.45	931.39	928.87	0.25	ppb
Strontium	86-1	1.70	1.91	1.72	1.78	6.55	ppb
Strontium	88-1	1.74	1.81	1.70	1.75	3.33	ppb
Sulfur	34-1	-1576.02	-1464.28	-1507.36	-1515.89		ppb
Terbium	159-1				100		%
Terbium	159-2				102		%
Thallium	203-1	1.53	1.60	1.49	1.54	3.65	ppb
Thallium	205-1	1.55	1.61	1.54	1.56	2.29	ppb
Tin	118-1	0.00	0.03	0.00	0.01	156.51	ppb
Titanium	47-1	8.86	8.74	8.50	8.70	2.12	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

PB164452BS

Instrumnet Name :

P7

Client Sample ID :

LCS452

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:17:59

DataFile Name :

129LCSE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.52	1.54	1.48	1.51	2.08	ppb
Vanadium	51-2	9.67	9.93	9.90	9.83	1.45	ppb
Yttrium	89-1				92		%
Yttrium	89-2				97		%
Zinc	66-2	9.68	9.60	10.14	9.81	2.95	ppb
Zirconium	90-1	1.86	1.91	1.80	1.86	2.78	ppb
Zirconium	91-1	1.91	1.97	1.77	1.88	5.30	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BSD Instrumnet Name : P7
Client Sample ID : LCSD452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:21:14 DataFile Name : 130LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	23.86	25.94	25.67	25.16	4.49	ppb
Antimony	121-1	1.76	1.77	1.80	1.78	1.13	ppb
Arsenic	75-2	1.03	0.90	1.16	1.03	12.65	ppb
Barium	135-1	8.43	8.59	8.55	8.52	0.97	ppb
Barium	137-1	8.48	8.64	8.70	8.60	1.33	ppb
Beryllium	9-1	0.92	0.90	0.98	0.93	4.53	ppb
Bismuth	209-1				100		%
Bismuth	209-2				99		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.08	1.02	0.77	0.95	17.26	ppb
Cadmium	106-1	-1.07	-1.16	-1.36	-1.20		ppb
Cadmium	111-1	0.67	0.72	0.77	0.72	6.53	ppb
Calcium	43-1	452.33	459.46	464.20	458.66	1.30	ppb
Calcium	44-1	455.67	462.95	463.86	460.83	0.97	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2.01	2.18	2.14	2.11	4.09	ppb
Cobalt	59-2	0.99	1.08	0.99	1.02	4.54	ppb
Copper	63-2	1.94	2.01	1.86	1.93	3.86	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				98		%
Holmium	165-2				98		%
Indium	115-1				95		%
Indium	115-2				97		%
Iron	56-2	223.70	237.70	229.38	230.26	3.06	ppb
Iron	57-2	227.77	240.20	224.25	230.74	3.63	ppb
Iron	54-2	229.24	238.18	237.61	235.01	2.13	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BSD Instrumnet Name : P7
Client Sample ID : LCSD452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:21:14 DataFile Name : 130LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.71	0.71	0.77	0.73	5.19	ppb
Lead	207-1	0.76	0.74	0.73	0.74	1.99	ppb
Lead	208-1	0.73	0.72	0.75	0.73	1.98	ppb
Lithium	6-1				86		%
Magnesium	24-2	496.64	520.39	494.17	503.73	2.87	ppb
Manganese	55-2	1.26	1.29	1.45	1.33	7.83	ppb
Molybdenum	94-1	4.87	5.02	4.96	4.95	1.51	ppb
Molybdenum	95-1	3.93	4.18	4.26	4.12	4.22	ppb
Molybdenum	96-1	4.17	4.26	4.29	4.24	1.48	ppb
Molybdenum	97-1	4.00	4.14	4.23	4.12	2.76	ppb
Molybdenum	98-1	4.07	4.14	4.15	4.12	1.11	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.01	1.13	1.13	1.09	6.33	ppb
Phosphorus	31-2	-23.68	-18.98	-19.68	-20.78		ppb
Potassium	39-2	474.19	507.83	474.54	485.52	3.98	ppb
Rhodium	103-1				93		%
Rhodium	103-2				97		%
Scandium	45-1				87		%
Scandium	45-2				92		%
Selenium	82-1	4.53	4.41	4.09	4.35	5.21	ppb
Selenium	77-2	5.31	5.58	5.80	5.56	4.46	ppb
Selenium	78-2	5.68	5.31	5.26	5.42	4.20	ppb
Silicon	28-1	2.74	2.76	2.95	2.82	4.06	ppb
Silver	107-1	0.89	0.92	0.90	0.90	1.51	ppb
Silver	109-1	0.87	0.87	0.93	0.89	3.45	ppb
Sodium	23-2	452.08	479.55	449.45	460.36	3.62	ppb
Strontium	86-1	0.81	0.81	0.82	0.81	0.46	ppb
Strontium	88-1	0.80	0.82	0.81	0.81	1.26	ppb
Sulfur	34-1	-1363.01	-1371.92	-1303.21	-1346.05		ppb
Terbium	159-1				99		%
Terbium	159-2				99		%
Thallium	203-1	0.77	0.76	0.78	0.77	1.39	ppb
Thallium	205-1	0.76	0.75	0.79	0.77	2.39	ppb
Tin	118-1	-0.01	0.00	0.01	0.00		ppb
Titanium	47-1	4.29	4.62	4.55	4.48	3.85	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

PB164452BSD

Instrumnet Name :

P7

Client Sample ID :

LCSD452

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:21:14

DataFile Name :

130LLCC.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.76	0.75	0.75	0.75	0.32	ppb
Vanadium	51-2	4.81	5.10	4.94	4.95	2.88	ppb
Yttrium	89-1				92		%
Yttrium	89-2				95		%
Zinc	66-2	4.64	5.32	4.82	4.93	7.12	ppb
Zirconium	90-1	0.91	0.90	0.93	0.91	1.37	ppb
Zirconium	91-1	0.91	0.94	0.92	0.92	1.19	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01 Instrumnet Name : P7
Client Sample ID : MYE4H7 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:24:29 DataFile Name : 131SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	52193.49	52660.13	52755.76	52536.46	0.57	ppb
Antimony	121-1	2.52	2.52	2.53	2.52	0.30	ppb
Arsenic	75-2	156.70	160.34	157.88	158.31	1.17	ppb
Barium	135-1	786.41	796.54	792.63	791.86	0.65	ppb
Barium	137-1	779.47	793.96	793.50	788.97	1.04	ppb
Beryllium	9-1	2.08	2.18	2.45	2.24	8.63	ppb
Bismuth	209-1				99		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	4.16	3.95	4.63	4.25	8.19	ppb
Cadmium	106-1	3.31	2.48	3.82	3.20	21.22	ppb
Cadmium	111-1	2.89	2.95	3.05	2.96	2.84	ppb
Calcium	43-1	10379.32	10417.26	10620.81	10472.46	1.24	ppb
Calcium	44-1	9818.29	9987.48	9912.01	9905.93	0.86	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	129.77	129.59	130.64	130.00	0.43	ppb
Cobalt	59-2	76.56	77.12	76.71	76.80	0.38	ppb
Copper	63-2	390.48	395.27	395.83	393.86	0.75	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				103		%
Indium	115-1				93		%
Indium	115-2				94		%
Iron	56-2	137790.43	137927.88	137686.03	137801.45	0.09	ppb
Iron	57-2	135531.54	137146.82	137147.18	136608.51	0.68	ppb
Iron	54-2	137912.22	138148.54	138848.11	138302.96	0.35	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01 Instrumnet Name : P7
Client Sample ID : MYE4H7 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:24:29 DataFile Name : 131SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	103.54	105.02	107.87	105.48	2.09	ppb
Lead	207-1	103.44	104.05	108.45	105.31	2.60	ppb
Lead	208-1	112.62	112.62	117.67	114.30	2.55	ppb
Lithium	6-1				86		%
Magnesium	24-2	27826.30	28375.07	27657.97	27953.11	1.34	ppb
Manganese	55-2	3733.94	3766.08	3739.21	3746.41	0.46	ppb
Molybdenum	94-1	27.14	27.89	27.77	27.60	1.47	ppb
Molybdenum	95-1	1.51	1.55	1.50	1.52	1.96	ppb
Molybdenum	96-1	5.20	5.31	5.29	5.27	1.07	ppb
Molybdenum	97-1	1.44	1.51	1.44	1.46	2.79	ppb
Molybdenum	98-1	1.36	1.36	1.37	1.36	0.19	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	135.01	136.71	135.75	135.82	0.63	ppb
Phosphorus	31-2	2435.24	2495.10	2484.54	2471.63	1.29	ppb
Potassium	39-2	11123.75	11066.82	10983.47	11058.01	0.64	ppb
Rhodium	103-1				88		%
Rhodium	103-2				95		%
Scandium	45-1				93		%
Scandium	45-2				100		%
Selenium	82-1	0.88	1.39	1.40	1.22	24.42	ppb
Selenium	77-2	22.64	26.09	24.66	24.46	7.09	ppb
Selenium	78-2	5.95	7.09	4.70	5.91	20.22	ppb
Silicon	28-1	2335.62	2344.85	2355.74	2345.41	0.43	ppb
Silver	107-1	0.76	0.71	0.71	0.73	3.65	ppb
Silver	109-1	0.79	0.72	0.72	0.75	5.71	ppb
Sodium	23-2	150.25	152.29	153.85	152.13	1.19	ppb
Strontium	86-1	91.42	92.96	92.52	92.30	0.86	ppb
Strontium	88-1	94.65	95.11	95.87	95.21	0.65	ppb
Sulfur	34-1	-1801.86	-1822.89	-1789.41	-1804.72		ppb
Terbium	159-1				99		%
Terbium	159-2				102		%
Thallium	203-1	5.82	6.04	6.18	6.01	2.99	ppb
Thallium	205-1	5.91	6.04	6.15	6.04	1.97	ppb
Tin	118-1	3.53	3.51	3.58	3.54	1.02	ppb
Titanium	47-1	1174.11	1294.40	1172.42	1213.64	5.76	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4525-01

Instrumnet Name :

P7

Client Sample ID :

MYE4H7

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:24:29

DataFile Name :

131SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.54	1.58	1.63	1.59	2.87	ppb
Vanadium	51-2	257.35	259.46	259.37	258.73	0.46	ppb
Yttrium	89-1				99		%
Yttrium	89-2				105		%
Zinc	66-2	374.44	372.94	375.46	374.28	0.34	ppb
Zirconium	90-1	15.57	15.78	15.79	15.71	0.80	ppb
Zirconium	91-1	15.62	16.09	15.85	15.85	1.51	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-02 Instrumnet Name : P7
Client Sample ID : MYE4H7D Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:27:33 DataFile Name : 132SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	50505.02	50795.02	50355.76	50551.93	0.44	ppb
Antimony	121-1	2.42	2.42	2.90	2.58	10.69	ppb
Arsenic	75-2	155.98	152.82	155.79	154.86	1.14	ppb
Barium	135-1	744.39	745.67	881.14	790.40	9.94	ppb
Barium	137-1	758.42	744.33	882.54	795.10	9.57	ppb
Beryllium	9-1	2.00	2.05	2.32	2.12	8.09	ppb
Bismuth	209-1				97		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	3.41	3.98	5.40	4.26	23.97	ppb
Cadmium	106-1	2.77	2.84	3.57	3.06	14.44	ppb
Cadmium	111-1	2.74	2.68	3.29	2.90	11.51	ppb
Calcium	43-1	10078.42	10230.09	11481.53	10596.68	7.27	ppb
Calcium	44-1	9487.10	9557.13	10904.42	9982.88	8.00	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	124.37	125.13	124.56	124.69	0.31	ppb
Cobalt	59-2	74.25	74.04	73.91	74.07	0.23	ppb
Copper	63-2	377.65	382.40	377.15	379.07	0.76	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				97		%
Holmium	165-2				105		%
Indium	115-1				89		%
Indium	115-2				96		%
Iron	56-2	131677.35	134053.42	132983.43	132904.73	0.90	ppb
Iron	57-2	131884.89	132641.98	130951.37	131826.08	0.64	ppb
Iron	54-2	133191.76	133464.98	133643.08	133433.27	0.17	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-02 Instrumnet Name : P7
Client Sample ID : MYE4H7D Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:27:33 DataFile Name : 132SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	100.88	103.11	118.04	107.35	8.69	ppb
Lead	207-1	100.11	103.05	117.55	106.91	8.73	ppb
Lead	208-1	109.35	110.89	127.56	115.93	8.71	ppb
Lithium	6-1				84		%
Magnesium	24-2	26930.69	27199.10	26841.82	26990.54	0.69	ppb
Manganese	55-2	3590.27	3660.89	3624.14	3625.10	0.97	ppb
Molybdenum	94-1	26.66	26.75	30.60	28.00	8.04	ppb
Molybdenum	95-1	1.32	1.43	1.65	1.46	11.43	ppb
Molybdenum	96-1	5.02	5.23	5.82	5.36	7.72	ppb
Molybdenum	97-1	1.41	1.38	1.57	1.46	7.20	ppb
Molybdenum	98-1	1.31	1.35	1.53	1.39	8.42	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	130.66	131.91	129.72	130.76	0.84	ppb
Phosphorus	31-2	2387.62	2418.38	2396.40	2400.80	0.66	ppb
Potassium	39-2	10428.81	10646.27	10549.43	10541.50	1.03	ppb
Rhodium	103-1				85		%
Rhodium	103-2				97		%
Scandium	45-1				88		%
Scandium	45-2				101		%
Selenium	82-1	0.97	0.70	1.88	1.19	52.38	ppb
Selenium	77-2	26.03	26.99	24.61	25.88	4.61	ppb
Selenium	78-2	5.07	5.40	5.93	5.47	7.94	ppb
Silicon	28-1	2322.52	2361.73	2646.30	2443.52	7.23	ppb
Silver	107-1	0.65	0.68	0.79	0.71	10.58	ppb
Silver	109-1	0.63	0.65	0.78	0.69	11.47	ppb
Sodium	23-2	146.81	151.40	148.03	148.74	1.60	ppb
Strontium	86-1	88.67	89.18	102.05	93.30	8.12	ppb
Strontium	88-1	92.04	92.61	105.46	96.70	7.84	ppb
Sulfur	34-1	-1812.62	-1781.58	-1376.36	-1656.85		ppb
Terbium	159-1				97		%
Terbium	159-2				105		%
Thallium	203-1	5.14	5.41	6.07	5.54	8.59	ppb
Thallium	205-1	5.13	5.29	5.96	5.46	8.07	ppb
Tin	118-1	3.29	3.37	3.98	3.55	10.68	ppb
Titanium	47-1	1132.88	1151.16	1290.63	1191.56	7.24	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4525-02

Instrumnet Name :

P7

Client Sample ID :

MYE4H7D

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:27:33

DataFile Name :

132SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.52	1.57	1.79	1.63	8.59	ppb
Vanadium	51-2	250.40	249.87	249.87	250.05	0.12	ppb
Yttrium	89-1				95		%
Yttrium	89-2				104		%
Zinc	66-2	356.98	362.87	358.51	359.45	0.85	ppb
Zirconium	90-1	15.12	15.38	18.04	16.18	9.99	ppb
Zirconium	91-1	15.14	15.28	17.76	16.06	9.19	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01LX5 Instrumnet Name : P7
Client Sample ID : MYE4H7L Dilution Factor : 5
Date & Time Acquired : 2024-11-03 00:30:35 DataFile Name : 133SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	11132.61	10943.93	11172.80	11083.11	1.10	ppb
Antimony	121-1	0.49	0.48	0.47	0.48	2.58	ppb
Arsenic	75-2	33.56	33.49	33.67	33.57	0.26	ppb
Barium	135-1	141.41	141.07	140.16	140.88	0.46	ppb
Barium	137-1	142.47	141.56	140.59	141.54	0.66	ppb
Beryllium	9-1	0.40	0.42	0.41	0.41	3.03	ppb
Bismuth	209-1				104		%
Bismuth	209-2				105		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.69	0.72	0.86	0.76	11.72	ppb
Cadmium	106-1	-0.86	-1.47	-1.14	-1.16		ppb
Cadmium	111-1	0.50	0.50	0.44	0.48	6.76	ppb
Calcium	43-1	2197.71	2163.42	2230.81	2197.31	1.53	ppb
Calcium	44-1	2200.52	2190.87	2221.50	2204.30	0.71	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	27.29	26.93	27.50	27.24	1.05	ppb
Cobalt	59-2	16.23	16.27	16.65	16.38	1.41	ppb
Copper	63-2	80.57	80.01	81.10	80.56	0.68	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				105		%
Indium	115-1				95		%
Indium	115-2				98		%
Iron	56-2	29191.50	29031.78	29192.31	29138.53	0.32	ppb
Iron	57-2	28871.42	29043.90	29022.90	28979.41	0.32	ppb
Iron	54-2	29383.97	28901.16	29319.31	29201.48	0.90	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01LX5 Instrumnet Name : P7
Client Sample ID : MYE4H7L Dilution Factor : 5
Date & Time Acquired : 2024-11-03 00:30:35 DataFile Name : 133SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	20.55	20.43	20.22	20.40	0.81	ppb
Lead	207-1	20.01	20.33	20.54	20.29	1.32	ppb
Lead	208-1	20.21	20.22	20.15	20.19	0.20	ppb
Lithium	6-1				89		%
Magnesium	24-2	5888.74	5873.49	5915.45	5892.56	0.36	ppb
Manganese	55-2	800.54	782.27	785.12	789.31	1.25	ppb
Molybdenum	94-1	5.72	5.70	5.82	5.75	1.16	ppb
Molybdenum	95-1	0.26	0.27	0.27	0.27	2.54	ppb
Molybdenum	96-1	1.04	1.03	1.07	1.05	1.95	ppb
Molybdenum	97-1	0.27	0.26	0.26	0.26	3.09	ppb
Molybdenum	98-1	0.24	0.24	0.25	0.24	3.31	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	29.81	28.73	29.43	29.32	1.86	ppb
Phosphorus	31-2	480.04	467.95	510.44	486.14	4.50	ppb
Potassium	39-2	2417.11	2372.87	2442.73	2410.90	1.47	ppb
Rhodium	103-1				92		%
Rhodium	103-2				99		%
Scandium	45-1				86		%
Scandium	45-2				93		%
Selenium	82-1	0.16	0.27	-0.09	0.11	163.55	ppb
Selenium	77-2	3.82	4.77	4.00	4.20	12.01	ppb
Selenium	78-2	0.63	0.69	0.25	0.52	45.46	ppb
Silicon	28-1	497.95	487.68	493.45	493.03	1.04	ppb
Silver	107-1	0.14	0.14	0.12	0.13	6.14	ppb
Silver	109-1	0.13	0.14	0.14	0.14	2.91	ppb
Sodium	23-2	21.20	20.67	21.40	21.09	1.79	ppb
Strontium	86-1	19.11	19.35	19.22	19.22	0.62	ppb
Strontium	88-1	18.94	19.24	19.39	19.19	1.21	ppb
Sulfur	34-1	-1828.70	-1883.51	-1867.17	-1859.79		ppb
Terbium	159-1				101		%
Terbium	159-2				105		%
Thallium	203-1	1.14	1.14	1.14	1.14	0.05	ppb
Thallium	205-1	1.15	1.15	1.13	1.14	1.09	ppb
Tin	118-1	0.65	0.64	0.68	0.66	3.17	ppb
Titanium	47-1	244.34	242.66	247.17	244.73	0.93	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4525-01LX5

Instrumnet Name :

P7

Client Sample ID :

MYE4H7L

Dilution Factor :

5

Date & Time Acquired :

2024-11-03 00:30:35

DataFile Name :

133SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.31	0.31	0.29	0.30	3.38	ppb
Vanadium	51-2	54.32	53.18	54.23	53.91	1.18	ppb
Yttrium	89-1				93		%
Yttrium	89-2				98		%
Zinc	66-2	82.08	81.61	83.00	82.23	0.86	ppb
Zirconium	90-1	3.31	3.29	3.30	3.30	0.44	ppb
Zirconium	91-1	3.27	3.31	3.28	3.29	0.73	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-03 Instrumnet Name : P7
Client Sample ID : MYE4H7S Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:33:45 DataFile Name : 134SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	39162.28	39154.79	39390.48	39235.85	0.34	ppb
Antimony	121-1	54.21	55.49	55.85	55.18	1.56	ppb
Arsenic	75-2	137.82	138.16	139.18	138.39	0.51	ppb
Barium	135-1	1645.20	1689.55	1684.51	1673.08	1.45	ppb
Barium	137-1	1653.80	1694.10	1690.25	1679.38	1.32	ppb
Beryllium	9-1	30.99	30.47	31.04	30.83	1.01	ppb
Bismuth	209-1				103		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	27.50	29.40	26.81	27.90	4.81	ppb
Cadmium	106-1	24.73	25.93	25.62	25.43	2.46	ppb
Cadmium	111-1	28.03	28.72	29.40	28.72	2.38	ppb
Calcium	43-1	7634.14	7663.75	7756.03	7684.64	0.83	ppb
Calcium	44-1	7189.63	7129.84	7223.81	7181.09	0.66	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	209.33	209.78	210.42	209.84	0.26	ppb
Cobalt	59-2	352.96	349.57	354.75	352.43	0.75	ppb
Copper	63-2	432.07	435.29	431.08	432.81	0.51	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				105		%
Indium	115-1				94		%
Indium	115-2				96		%
Iron	56-2	101250.60	101588.56	102314.75	101717.97	0.53	ppb
Iron	57-2	100640.28	100935.66	101127.11	100901.01	0.24	ppb
Iron	54-2	101975.64	101881.45	101630.74	101829.28	0.18	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-03 Instrumnet Name : P7
Client Sample ID : MYE4H7S Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:33:45 DataFile Name : 134SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	83.68	84.04	85.70	84.47	1.27	ppb
Lead	207-1	83.34	84.43	86.40	84.73	1.83	ppb
Lead	208-1	83.05	91.75	93.13	89.31	6.12	ppb
Lithium	6-1				88		%
Magnesium	24-2	20179.70	20259.66	20642.41	20360.59	1.21	ppb
Manganese	55-2	3006.65	3011.67	3016.62	3011.65	0.17	ppb
Molybdenum	94-1	19.69	20.38	20.11	20.06	1.74	ppb
Molybdenum	95-1	1.03	1.05	1.09	1.05	3.15	ppb
Molybdenum	96-1	3.74	3.85	3.82	3.81	1.45	ppb
Molybdenum	97-1	1.04	1.02	1.08	1.05	3.16	ppb
Molybdenum	98-1	0.97	1.00	1.02	1.00	2.55	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	377.39	378.82	379.72	378.64	0.31	ppb
Phosphorus	31-2	1784.27	1783.94	1783.80	1784.00	0.01	ppb
Potassium	39-2	7955.67	7913.79	8053.98	7974.48	0.90	ppb
Rhodium	103-1				90		%
Rhodium	103-2				98		%
Scandium	45-1				90		%
Scandium	45-2				97		%
Selenium	82-1	53.62	53.70	55.39	54.24	1.84	ppb
Selenium	77-2	82.04	69.51	66.23	72.59	11.50	ppb
Selenium	78-2	58.51	54.75	56.68	56.65	3.32	ppb
Silicon	28-1	1732.77	1724.73	1753.03	1736.84	0.84	ppb
Silver	107-1	22.86	23.54	23.58	23.33	1.72	ppb
Silver	109-1	23.12	23.52	23.72	23.45	1.29	ppb
Sodium	23-2	107.41	108.01	109.94	108.45	1.22	ppb
Strontium	86-1	65.67	67.17	67.68	66.84	1.57	ppb
Strontium	88-1	68.97	69.13	69.71	69.27	0.56	ppb
Sulfur	34-1	-1768.91	-1765.29	-1731.94	-1755.38		ppb
Terbium	159-1				103		%
Terbium	159-2				105		%
Thallium	203-1	28.39	28.39	28.87	28.55	0.97	ppb
Thallium	205-1	28.07	28.53	29.17	28.59	1.94	ppb
Tin	118-1	2.45	2.46	2.57	2.49	2.80	ppb
Titanium	47-1	853.60	849.57	865.73	856.30	0.98	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-03 Instrumnet Name : P7
Client Sample ID : MYE4H7S Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:33:45 DataFile Name : 134SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.12	1.11	1.14	1.13	1.35	ppb
Vanadium	51-2	462.32	464.14	472.26	466.24	1.13	ppb
Yttrium	89-1				98		%
Yttrium	89-2				102		%
Zinc	66-2	548.34	551.72	554.26	551.44	0.54	ppb
Zirconium	90-1	11.28	11.47	11.56	11.44	1.24	ppb
Zirconium	91-1	11.39	11.40	11.68	11.49	1.40	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01MS Instrumnet Name : P7
Client Sample ID : MYE4H7SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:36:46 DataFile Name : 135SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	42286.04	42541.08	42682.38	42503.17	0.47	ppb
Antimony	121-1	284.91	291.12	290.90	288.98	1.22	ppb
Arsenic	75-2	223.74	219.90	220.51	221.38	0.93	ppb
Barium	135-1	5892.17	6032.13	6036.46	5986.92	1.37	ppb
Barium	137-1	5988.33	6040.09	5983.93	6004.12	0.52	ppb
Beryllium	9-1	139.49	142.99	143.43	141.97	1.52	ppb
Bismuth	209-1				103		%
Bismuth	209-2				102		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	127.26	132.37	130.46	130.03	1.99	ppb
Cadmium	106-1	122.93	125.30	123.45	123.89	1.01	ppb
Cadmium	111-1	134.93	138.51	137.65	137.03	1.36	ppb
Calcium	43-1	7338.39	7532.88	7563.61	7478.30	1.63	ppb
Calcium	44-1	6951.24	7082.66	7064.72	7032.87	1.01	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	653.18	641.68	655.86	650.24	1.16	ppb
Cobalt	59-2	1485.08	1469.31	1501.58	1485.32	1.09	ppb
Copper	63-2	961.82	961.80	967.29	963.64	0.33	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				103		%
Indium	115-1				93		%
Indium	115-2				95		%
Iron	56-2	101789.71	101491.27	102624.14	101968.37	0.58	ppb
Iron	57-2	100955.34	100250.19	101975.61	101060.38	0.86	ppb
Iron	54-2	102306.91	101977.91	103351.70	102545.50	0.70	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01MS Instrumnet Name : P7
Client Sample ID : MYE4H7SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:36:46 DataFile Name : 135SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	123.23	123.34	123.52	123.36	0.12	ppb
Lead	207-1	121.46	121.64	122.93	122.01	0.66	ppb
Lead	208-1	131.28	132.85	134.33	132.82	1.15	ppb
Lithium	6-1				88		%
Magnesium	24-2	19789.44	19795.78	19959.25	19848.16	0.48	ppb
Manganese	55-2	4034.70	4053.58	4109.10	4065.79	0.95	ppb
Molybdenum	94-1	19.50	19.97	19.92	19.80	1.31	ppb
Molybdenum	95-1	1.07	1.05	1.09	1.07	1.74	ppb
Molybdenum	96-1	3.77	3.82	3.90	3.83	1.73	ppb
Molybdenum	97-1	1.04	1.04	1.04	1.04	0.24	ppb
Molybdenum	98-1	1.00	1.04	0.99	1.01	2.32	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1536.64	1532.31	1550.89	1539.94	0.63	ppb
Phosphorus	31-2	1758.25	1746.17	1764.65	1756.36	0.53	ppb
Potassium	39-2	7763.48	7759.46	7795.68	7772.87	0.26	ppb
Rhodium	103-1				88		%
Rhodium	103-2				96		%
Scandium	45-1				87		%
Scandium	45-2				94		%
Selenium	82-1	293.15	299.09	299.85	297.37	1.23	ppb
Selenium	77-2	321.85	299.83	312.67	311.45	3.55	ppb
Selenium	78-2	296.51	296.57	289.76	294.28	1.33	ppb
Silicon	28-1	1673.88	1708.04	1717.74	1699.89	1.36	ppb
Silver	107-1	109.34	111.48	111.97	110.93	1.26	ppb
Silver	109-1	109.67	111.80	111.84	111.11	1.12	ppb
Sodium	23-2	106.83	105.42	107.10	106.45	0.84	ppb
Strontium	86-1	65.50	65.77	66.26	65.84	0.58	ppb
Strontium	88-1	68.14	68.16	68.38	68.23	0.19	ppb
Sulfur	34-1	-1808.47	-1758.08	-1739.67	-1768.74		ppb
Terbium	159-1				102		%
Terbium	159-2				104		%
Thallium	203-1	127.92	127.54	129.11	128.19	0.64	ppb
Thallium	205-1	151.26	147.49	147.94	148.89	1.38	ppb
Tin	118-1	2.35	2.45	2.39	2.40	1.99	ppb
Titanium	47-1	824.15	839.32	841.83	835.10	1.15	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4525-01MS

Instrumnet Name :

P7

Client Sample ID :

MYE4H7SRE

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:36:46

DataFile Name :

135SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.10	1.09	1.11	1.10	0.83	ppb
Vanadium	51-2	1550.45	1529.29	1554.97	1544.90	0.89	ppb
Yttrium	89-1				95		%
Yttrium	89-2				99		%
Zinc	66-2	1613.91	1608.37	1631.23	1617.84	0.74	ppb
Zirconium	90-1	11.09	11.28	11.36	11.24	1.23	ppb
Zirconium	91-1	11.30	11.36	11.54	11.40	1.07	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4526-01 Instrumnet Name : P7
Client Sample ID : MYE4M3 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:39:44 DataFile Name : 136SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	40286.44	40083.51	39995.46	40121.80	0.37	ppb
Antimony	121-1	2.62	2.62	2.63	2.62	0.24	ppb
Arsenic	75-2	93.59	96.72	95.49	95.26	1.66	ppb
Barium	135-1	442.47	442.05	450.09	444.87	1.02	ppb
Barium	137-1	482.34	482.99	485.11	483.48	0.30	ppb
Beryllium	9-1	1.69	1.52	1.62	1.61	5.23	ppb
Bismuth	209-1				104		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.96	2.42	2.83	2.74	10.25	ppb
Cadmium	106-1	1.38	1.40	1.60	1.46	8.35	ppb
Cadmium	111-1	1.25	1.20	1.24	1.23	2.10	ppb
Calcium	43-1	9907.30	9749.26	9653.15	9769.90	1.31	ppb
Calcium	44-1	9333.75	9200.12	9179.78	9237.89	0.91	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	97.28	96.63	96.83	96.91	0.35	ppb
Cobalt	59-2	46.06	46.23	45.85	46.05	0.42	ppb
Copper	63-2	198.80	200.06	198.98	199.28	0.34	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				104		%
Indium	115-1				93		%
Indium	115-2				95		%
Iron	56-2	91873.60	91681.96	91583.81	91713.12	0.16	ppb
Iron	57-2	90877.23	90568.68	90345.00	90596.97	0.29	ppb
Iron	54-2	91767.02	91810.90	91755.39	91777.77	0.03	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4526-01 Instrumnet Name : P7
Client Sample ID : MYE4M3 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:39:44 DataFile Name : 136SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	52.90	52.53	52.70	52.71	0.36	ppb
Lead	207-1	52.08	51.43	51.89	51.80	0.64	ppb
Lead	208-1	51.94	51.34	52.01	51.77	0.71	ppb
Lithium	6-1				85		%
Magnesium	24-2	17733.21	17783.20	17823.62	17780.01	0.25	ppb
Manganese	55-2	2236.05	2223.75	2213.43	2224.41	0.51	ppb
Molybdenum	94-1	29.17	29.12	29.31	29.20	0.34	ppb
Molybdenum	95-1	1.14	1.17	1.15	1.16	1.24	ppb
Molybdenum	96-1	4.86	4.78	4.85	4.83	0.85	ppb
Molybdenum	97-1	1.11	1.19	1.13	1.15	3.82	ppb
Molybdenum	98-1	1.10	1.11	1.11	1.11	0.65	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	91.79	91.44	91.60	91.61	0.19	ppb
Phosphorus	31-2	2052.62	2082.55	2085.40	2073.52	0.88	ppb
Potassium	39-2	9875.86	9797.36	9792.54	9821.92	0.48	ppb
Rhodium	103-1				88		%
Rhodium	103-2				97		%
Scandium	45-1				87		%
Scandium	45-2				96		%
Selenium	82-1	0.61	0.71	0.21	0.51	51.97	ppb
Selenium	77-2	20.49	23.02	19.35	20.95	8.96	ppb
Selenium	78-2	3.90	4.15	2.32	3.46	28.61	ppb
Silicon	28-1	1167.02	1134.25	1143.87	1148.38	1.47	ppb
Silver	107-1	0.50	0.44	0.43	0.46	8.08	ppb
Silver	109-1	0.49	0.47	0.42	0.46	7.67	ppb
Sodium	23-2	153.59	154.93	152.97	153.83	0.65	ppb
Strontium	86-1	82.30	82.30	83.81	82.80	1.05	ppb
Strontium	88-1	85.79	86.68	86.02	86.16	0.54	ppb
Sulfur	34-1	-1418.64	-1478.51	-1447.97	-1448.37		ppb
Terbium	159-1				101		%
Terbium	159-2				103		%
Thallium	203-1	0.42	0.39	0.38	0.40	6.23	ppb
Thallium	205-1	0.39	0.39	0.39	0.39	1.30	ppb
Tin	118-1	2.91	3.01	3.04	2.98	2.28	ppb
Titanium	47-1	1106.98	1086.10	1094.05	1095.71	0.96	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4526-01 Instrumnet Name : P7
Client Sample ID : MYE4M3 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:39:44 DataFile Name : 136SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.28	1.28	1.27	1.27	0.62	ppb
Vanadium	51-2	164.80	165.05	165.04	164.96	0.09	ppb
Yttrium	89-1				95		%
Yttrium	89-2				101		%
Zinc	66-2	263.40	264.90	262.99	263.76	0.38	ppb
Zirconium	90-1	16.65	16.81	16.86	16.77	0.65	ppb
Zirconium	91-1	16.94	17.01	16.92	16.96	0.29	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4526-02 Instrumnet Name : P7
Client Sample ID : MYE4M3D Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:42:48 DataFile Name : 137SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	40433.32	40159.62	39681.30	40091.41	0.95	ppb
Antimony	121-1	2.50	2.58	2.53	2.54	1.59	ppb
Arsenic	75-2	96.25	98.01	97.16	97.14	0.90	ppb
Barium	135-1	445.49	454.94	454.90	451.78	1.20	ppb
Barium	137-1	480.26	494.89	495.90	490.35	1.79	ppb
Beryllium	9-1	1.50	1.53	1.77	1.60	9.02	ppb
Bismuth	209-1				104		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	2.69	3.13	3.22	3.01	9.38	ppb
Cadmium	106-1	2.55	2.33	1.34	2.07	31.09	ppb
Cadmium	111-1	1.35	1.29	1.33	1.32	2.12	ppb
Calcium	43-1	9695.38	9761.37	9860.60	9772.45	0.85	ppb
Calcium	44-1	9106.42	9276.14	9334.88	9239.15	1.28	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	98.42	98.48	97.28	98.06	0.68	ppb
Cobalt	59-2	46.35	46.38	45.77	46.17	0.75	ppb
Copper	63-2	202.83	200.33	198.72	200.63	1.03	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				105		%
Indium	115-1				94		%
Indium	115-2				95		%
Iron	56-2	93411.13	92756.92	91864.14	92677.40	0.84	ppb
Iron	57-2	92558.24	92178.12	91405.56	92047.31	0.64	ppb
Iron	54-2	93136.36	92783.27	91810.84	92576.82	0.74	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4526-02 Instrumnet Name : P7
Client Sample ID : MYE4M3D Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:42:48 DataFile Name : 137SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	51.41	53.31	53.99	52.90	2.53	ppb
Lead	207-1	50.46	51.73	52.96	51.72	2.42	ppb
Lead	208-1	51.08	52.35	53.01	52.14	1.88	ppb
Lithium	6-1				87		%
Magnesium	24-2	17947.46	17979.84	17839.73	17922.34	0.41	ppb
Manganese	55-2	2281.39	2291.32	2256.29	2276.34	0.79	ppb
Molybdenum	94-1	28.45	29.08	29.62	29.05	2.02	ppb
Molybdenum	95-1	1.14	1.17	1.18	1.16	1.87	ppb
Molybdenum	96-1	4.79	4.76	4.85	4.80	1.05	ppb
Molybdenum	97-1	1.17	1.18	1.22	1.19	2.36	ppb
Molybdenum	98-1	1.11	1.12	1.12	1.12	0.82	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	92.30	92.59	90.98	91.95	0.93	ppb
Phosphorus	31-2	2062.17	2128.60	2056.33	2082.37	1.93	ppb
Potassium	39-2	9762.25	9900.14	9776.13	9812.84	0.77	ppb
Rhodium	103-1				90		%
Rhodium	103-2				97		%
Scandium	45-1				89		%
Scandium	45-2				97		%
Selenium	82-1	0.69	0.47	1.09	0.75	41.93	ppb
Selenium	77-2	16.93	21.01	22.26	20.06	13.90	ppb
Selenium	78-2	2.93	2.20	3.88	3.00	28.08	ppb
Silicon	28-1	1113.42	1134.86	1152.31	1133.53	1.72	ppb
Silver	107-1	0.41	0.42	0.41	0.41	0.83	ppb
Silver	109-1	0.42	0.43	0.41	0.42	3.19	ppb
Sodium	23-2	153.68	152.70	152.34	152.90	0.46	ppb
Strontium	86-1	81.22	83.26	83.70	82.72	1.60	ppb
Strontium	88-1	85.24	86.48	86.14	85.95	0.75	ppb
Sulfur	34-1	-1432.61	-1419.05	-1473.10	-1441.59		ppb
Terbium	159-1				103		%
Terbium	159-2				104		%
Thallium	203-1	0.39	0.39	0.37	0.38	2.87	ppb
Thallium	205-1	0.40	0.39	0.40	0.40	1.78	ppb
Tin	118-1	2.86	2.98	2.99	2.94	2.39	ppb
Titanium	47-1	1082.76	1095.99	1102.03	1093.59	0.90	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4526-02

Instrumnet Name :

P7

Client Sample ID :

MYE4M3D

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:42:48

DataFile Name :

137SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	1.24	1.28	1.30	1.27	2.30	ppb
Vanadium	51-2	167.39	166.17	164.10	165.89	1.00	ppb
Yttrium	89-1				97		%
Yttrium	89-2				102		%
Zinc	66-2	265.37	266.88	261.51	264.58	1.05	ppb
Zirconium	90-1	16.57	16.82	17.02	16.80	1.32	ppb
Zirconium	91-1	16.57	16.84	17.37	16.93	2.40	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV005 Instrumnet Name : P7
Client Sample ID : CCV005 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:45:54 DataFile Name : 138CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	53334.84	47152.92	46696.08	49061.28	7.56	ppb
Antimony	121-1	478.90	483.99	480.29	481.06	0.55	ppb
Arsenic	75-2	565.55	492.33	504.67	520.85	7.53	ppb
Barium	135-1	2437.01	2459.43	2406.53	2434.32	1.09	ppb
Barium	137-1	2436.52	2454.99	2408.49	2433.34	0.96	ppb
Beryllium	9-1	492.51	497.87	493.34	494.58	0.58	ppb
Bismuth	209-1				96		%
Bismuth	209-2				94		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	486.06	493.34	485.23	488.21	0.91	ppb
Cadmium	106-1	483.46	490.19	482.28	485.31	0.88	ppb
Cadmium	111-1	464.12	464.58	458.65	462.45	0.71	ppb
Calcium	43-1	242381.53	248837.44	251714.74	247644.57	1.93	ppb
Calcium	44-1	239023.21	244547.97	245536.74	243035.97	1.44	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	560.85	496.30	500.00	519.05	6.98	ppb
Cobalt	59-2	599.95	526.36	521.73	549.35	7.99	ppb
Copper	63-2	5737.80	5051.90	5053.70	5281.13	7.49	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				101		%
Indium	115-1				90		%
Indium	115-2				89		%
Iron	56-2	143533.17	125699.72	125161.42	131464.77	7.95	ppb
Iron	57-2	144566.16	126135.87	126361.87	132354.63	7.99	ppb
Iron	54-2	144947.94	126491.58	127253.93	132897.82	7.86	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV005 Instrumnet Name : P7
Client Sample ID : CCV005 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:45:54 DataFile Name : 138CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2522.38	2516.41	2548.28	2529.02	0.67	ppb
Lead	207-1	2513.54	2534.31	2518.34	2522.06	0.43	ppb
Lead	208-1	2506.57	2520.19	2523.55	2516.77	0.36	ppb
Lithium	6-1				86		%
Magnesium	24-2	270238.82	239201.14	236237.45	248559.14	7.58	ppb
Manganese	55-2	5639.74	4892.15	4943.52	5158.47	8.10	ppb
Molybdenum	94-1	5014.58	5051.04	5074.55	5046.72	0.60	ppb
Molybdenum	95-1	5004.19	4997.74	5007.42	5003.11	0.10	ppb
Molybdenum	96-1	4990.98	4962.81	5040.81	4998.20	0.79	ppb
Molybdenum	97-1	5027.60	5026.47	5086.86	5046.98	0.68	ppb
Molybdenum	98-1	4993.55	5011.92	5053.35	5019.61	0.61	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	554.01	488.56	487.06	509.87	7.50	ppb
Phosphorus	31-2	10838.00	9601.54	9623.68	10021.07	7.06	ppb
Potassium	39-2	133960.55	117895.86	117686.47	123180.96	7.58	ppb
Rhodium	103-1				83		%
Rhodium	103-2				88		%
Scandium	45-1				82		%
Scandium	45-2				88		%
Selenium	82-1	497.11	497.25	505.90	500.09	1.01	ppb
Selenium	77-2	544.53	474.44	479.71	499.56	7.81	ppb
Selenium	78-2	563.09	477.94	493.43	511.49	8.87	ppb
Silicon	28-1	2754.92	2829.04	2836.92	2806.96	1.61	ppb
Silver	107-1	484.89	486.45	484.60	485.32	0.21	ppb
Silver	109-1	492.90	483.83	479.15	485.29	1.44	ppb
Sodium	23-2	268107.54	238074.51	236130.95	247437.67	7.25	ppb
Strontium	86-1	12170.04	12450.53	12354.74	12325.10	1.16	ppb
Strontium	88-1	12317.05	12358.81	12420.67	12365.51	0.42	ppb
Sulfur	34-1	9465.21	9727.27	9709.75	9634.08	1.52	ppb
Terbium	159-1				101		%
Terbium	159-2				101		%
Thallium	203-1	513.85	504.64	510.79	509.76	0.92	ppb
Thallium	205-1	504.26	507.30	505.95	505.83	0.30	ppb
Tin	118-1	466.58	471.04	473.03	470.21	0.70	ppb
Titanium	47-1	4860.72	4984.73	5003.72	4949.72	1.57	ppb

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

CCV005

Instrumnet Name :

P7

Client Sample ID :

CCV005

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:45:54

DataFile Name :

138CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	523.33	524.84	524.03	524.07	0.14	ppb
Vanadium	51-2	568.55	498.82	504.79	524.05	7.38	ppb
Yttrium	89-1				89		%
Yttrium	89-2				92		%
Zinc	66-2	5703.75	5105.84	5091.51	5300.37	6.59	ppb
Zirconium	90-1	506.55	508.47	508.26	507.76	0.21	ppb
Zirconium	91-1	516.86	505.75	523.21	515.28	1.72	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB005 Instrumnet Name : P7
Client Sample ID : CCB005 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:48:35 DataFile Name : 139CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	4.12	4.97	5.49	4.86	14.15	ppb
Antimony	121-1	0.29	0.29	0.29	0.29	0.43	ppb
Arsenic	75-2	0.05	0.01	0.03	0.03	59.34	ppb
Barium	135-1	0.20	0.19	0.16	0.18	10.51	ppb
Barium	137-1	0.22	0.18	0.16	0.19	16.34	ppb
Beryllium	9-1	0.03	0.02	0.01	0.02	32.61	ppb
Bismuth	209-1				107		%
Bismuth	209-2				108		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.03	0.05	0.05	0.04	25.29	ppb
Cadmium	106-1	-2.23	-2.26	-1.98	-2.16		ppb
Cadmium	111-1	-0.13	-0.13	-0.11	-0.12		ppb
Calcium	43-1	5.68	4.53	1.90	4.04	47.97	ppb
Calcium	44-1	5.12	2.17	1.49	2.93	66.01	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.05	0.08	0.04	0.05	39.58	ppb
Cobalt	59-2	0.03	0.03	0.03	0.03	1.68	ppb
Copper	63-2	0.10	0.13	0.12	0.12	9.94	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				107		%
Indium	115-1				98		%
Indium	115-2				101		%
Iron	56-2	13.68	16.34	17.24	15.76	11.74	ppb
Iron	57-2	10.63	13.63	13.81	12.69	14.11	ppb
Iron	54-2	14.05	14.39	17.95	15.46	13.95	ppb
Krypton	83-1						cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB005 Instrumnet Name : P7
Client Sample ID : CCB005 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:48:35 DataFile Name : 139CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.08	0.06	0.04	0.06	27.82	ppb
Lead	207-1	0.08	0.06	0.04	0.06	31.27	ppb
Lead	208-1	0.08	0.06	0.05	0.06	22.80	ppb
Lithium	6-1				92		%
Magnesium	24-2	4.79	5.98	7.22	6.00	20.29	ppb
Manganese	55-2	0.36	0.46	0.45	0.42	12.24	ppb
Molybdenum	94-1	0.28	0.25	0.29	0.28	7.82	ppb
Molybdenum	95-1	0.23	0.20	0.19	0.21	9.64	ppb
Molybdenum	96-1	0.20	0.20	0.20	0.20	1.66	ppb
Molybdenum	97-1	0.22	0.23	0.19	0.21	9.20	ppb
Molybdenum	98-1	0.21	0.21	0.19	0.21	5.23	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.08	0.07	0.07	0.07	12.25	ppb
Phosphorus	31-2	-21.33	-23.92	-21.52	-22.25		ppb
Potassium	39-2	-6.47	-4.29	-3.79	-4.85		ppb
Rhodium	103-1				94		%
Rhodium	103-2				102		%
Scandium	45-1				87		%
Scandium	45-2				94		%
Selenium	82-1	-0.21	-0.02	-0.35	-0.19		ppb
Selenium	77-2	0.01	-0.22	-0.22	-0.14		ppb
Selenium	78-2	-1.01	0.13	-0.47	-0.45		ppb
Silicon	28-1	0.94	0.96	0.86	0.92	5.55	ppb
Silver	107-1	0.05	0.04	0.04	0.04	19.37	ppb
Silver	109-1	0.05	0.04	0.04	0.04	18.86	ppb
Sodium	23-2	3.73	3.78	2.91	3.48	14.09	ppb
Strontium	86-1	0.32	0.26	0.28	0.29	10.25	ppb
Strontium	88-1	0.32	0.26	0.22	0.26	19.61	ppb
Sulfur	34-1	-580.71	-614.07	-556.00	-583.59		ppb
Terbium	159-1				104		%
Terbium	159-2				106		%
Thallium	203-1	0.06	0.06	0.04	0.05	22.67	ppb
Thallium	205-1	0.06	0.05	0.04	0.05	17.43	ppb
Tin	118-1	0.07	0.05	0.08	0.06	20.57	ppb
Titanium	47-1	0.31	0.28	0.30	0.30	4.17	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB005 Instrumnet Name : P7
Client Sample ID : CCB005 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:48:35 DataFile Name : 139CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.01	0.01	0.00	0.01	32.63	ppb
Vanadium	51-2	0.04	0.04	0.04	0.04	11.36	ppb
Yttrium	89-1				93		%
Yttrium	89-2				98		%
Zinc	66-2	0.01	0.00	0.18	0.06	156.98	ppb
Zirconium	90-1	0.07	0.06	0.05	0.06	20.58	ppb
Zirconium	91-1	0.07	0.06	0.04	0.06	27.45	ppb

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:41:38 DataFile Name : 004CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	507	487	473	489	3.43	cps
Antimony	121-1	190	203	147	180	16.46	cps
Arsenic	75-2	17	3	10	10	66.70	cps
Barium	135-1	193	133	153	160	19.09	cps
Barium	137-1	287	307	247	280	10.91	cps
Beryllium	9-1	50	89	106	81	34.99	cps
Bismuth	209-1	6886545	6703046	6598507	6729366	2.17	cps
Bismuth	209-2	4816958	4741223	4232432	4596871	6.92	cps
Bromine	81-1	11875	11998	11992	11955	0.58	cps
Bromine	81-2	67	40	73	60	29.40	cps
Cadmium	108-1	7	7	0	4	86.60	cps
Cadmium	106-1	2997	3000	2994	2997	0.11	cps
Cadmium	111-1	2123	2132	2139	2131	0.37	cps
Calcium	43-1	453	393	437	428	7.24	cps
Calcium	44-1	14471	14528	14528	14509	0.23	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	583	623	533	580	7.77	cps
Cobalt	59-2	120	127	113	120	5.56	cps
Copper	63-2	1377	1260	1257	1298	5.27	cps
Dysprosium	156-1	60	83	73	72	16.21	cps
Dysprosium	156-2	30	37	53	40	30.04	cps
Erbium	164-1	70	57	67	64	10.76	cps
Erbium	164-2	30	43	40	38	18.36	cps
Gadolinium	160-1	83	37	70	63	37.95	cps
Gadolinium	160-2	190	240	270	233	17.32	cps
Holmium	165-1	10674060	10765086	10382224	10607123	1.89	cps
Holmium	165-2	6602622	6557958	5773363	6311314	7.39	cps
Indium	115-1	8502518	8439644	8420202	8454121	0.51	cps
Indium	115-2	2692805	2632835	2327480	2551040	7.68	cps
Iron	56-2	18905	18104	17681	18230	3.41	cps
Iron	57-2	1550	1460	1390	1467	5.47	cps
Iron	54-2	2127	1800	1993	1973	8.32	cps
Krypton	83-1	167	203	217	196	13.24	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:41:38 DataFile Name : 004CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1157	1157	1013	1109	7.46	cps
Lead	207-1	917	930	997	948	4.52	cps
Lead	208-1	4320	4374	4217	4304	1.85	cps
Lithium	6-1	1417300	1438284	1416737	1424107	0.86	cps
Magnesium	24-2	2010	2160	1913	2028	6.13	cps
Manganese	55-2	403	500	400	434	13.07	cps
Molybdenum	94-1	580	517	520	539	6.61	cps
Molybdenum	95-1	533	467	443	481	9.71	cps
Molybdenum	96-1	593	663	597	618	6.39	cps
Molybdenum	97-1	287	287	340	304	10.12	cps
Molybdenum	98-1	827	863	777	822	5.29	cps
Neodymium	150-1	0	10	17	9	94.38	cps
Neodymium	150-2	7	17	0	8	107.84	cps
Nickel	60-2	73	87	77	79	8.80	cps
Phosphorus	31-2	243	217	167	209	18.63	cps
Potassium	39-2	52236	51815	51627	51892	0.60	cps
Rhodium	103-1	7996309	8107544	7850580	7984811	1.61	cps
Rhodium	103-2	4589860	4487169	4005366	4360798	7.16	cps
Scandium	45-1	4823563	4824207	4728029	4791933	1.15	cps
Scandium	45-2	416370	415110	366212	399231	7.16	cps
Selenium	82-1	507	445	477	476	6.57	cps
Selenium	77-2	3	7	0	3	100.05	cps
Selenium	78-2	243	243	280	256	8.28	cps
Silicon	28-1	216009	217162	218246	217139	0.52	cps
Silver	107-1	197	163	170	177	9.98	cps
Silver	109-1	123	107	120	117	7.56	cps
Sodium	23-2	26393	25982	26723	26366	1.41	cps
Strontium	86-1	860	857	823	847	2.39	cps
Strontium	88-1	5718	5014	5164	5299	6.99	cps
Sulfur	34-1	199525	203270	202065	201620	0.95	cps
Terbium	159-1	10888895	10763033	10894584	10848837	0.69	cps
Terbium	159-2	6661854	6512652	5667500	6280669	8.54	cps
Thallium	203-1	233	227	213	224	4.54	cps
Thallium	205-1	573	477	600	550	11.80	cps
Tin	118-1	1680	1733	1680	1698	1.81	cps
Titanium	47-1	207	180	253	213	17.40	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:41:38 DataFile Name : 004CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	343	233	277	284	19.48	cps
Vanadium	51-2	47	20	27	31	44.61	cps
Yttrium	89-1	13944666	14040814	13853602	13946361	0.67	cps
Yttrium	89-2	3352294	3304301	2929747	3195447	7.24	cps
Zinc	66-2	803	623	627	684	15.05	cps
Zirconium	90-1	767	737	697	733	4.79	cps
Zirconium	91-1	117	153	177	149	20.31	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S02 Instrumnet Name : P7
Client Sample ID : S02 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:48:10 DataFile Name : 006CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4637	4811	4807	4752	2.09	cps
Antimony	121-1	18485	19353	18852	18897	2.31	cps
Arsenic	75-2	460	447	577	494	14.46	cps
Barium	135-1	23817	23843	24064	23908	0.57	cps
Barium	137-1	41213	40715	40915	40948	0.61	cps
Beryllium	9-1	1545	1589	1583	1572	1.54	cps
Bismuth	209-1	6588392	6448702	6459108	6498734	1.20	cps
Bismuth	209-2	4671357	4613354	4731795	4672169	1.27	cps
Bromine	81-1	11391	11404	11391	11396	0.07	cps
Bromine	81-2	50	67	80	66	22.93	cps
Cadmium	108-1	177	220	217	204	11.79	cps
Cadmium	106-1	3200	3240	3267	3236	1.04	cps
Cadmium	111-1	4630	4835	4529	4665	3.34	cps
Calcium	43-1	14931	14354	14524	14603	2.03	cps
Calcium	44-1	240981	244917	244052	243317	0.85	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	8803	8719	8623	8715	1.03	cps
Cobalt	59-2	6618	6455	6608	6560	1.40	cps
Copper	63-2	10090	9923	9917	9977	0.99	cps
Dysprosium	156-1	57	63	73	64	13.01	cps
Dysprosium	156-2	30	23	33	29	17.63	cps
Erbium	164-1	57	77	97	77	26.09	cps
Erbium	164-2	40	37	60	46	27.70	cps
Gadolinium	160-1	60	33	43	46	29.58	cps
Gadolinium	160-2	273	230	243	249	8.92	cps
Holmium	165-1	10492065	10367489	10488348	10449301	0.68	cps
Holmium	165-2	6469391	6407703	6487730	6454941	0.65	cps
Indium	115-1	8555013	8386924	8390756	8444231	1.14	cps
Indium	115-2	2577872	2566654	2598804	2581110	0.63	cps
Iron	56-2	202559	203305	203281	203049	0.21	cps
Iron	57-2	6181	6048	6185	6138	1.27	cps
Iron	54-2	12619	12442	12365	12475	1.04	cps
Krypton	83-1	187	147	203	179	16.28	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S02 Instrumnet Name : P7
Client Sample ID : S02 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:48:10 DataFile Name : 006CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	7309	7586	7345	7413	2.03	cps
Lead	207-1	6408	6562	6695	6555	2.19	cps
Lead	208-1	29206	30400	29483	29696	2.11	cps
Lithium	6-1	1428436	1403411	1392885	1408244	1.30	cps
Magnesium	24-2	204393	201914	200566	202291	0.96	cps
Manganese	55-2	3364	3510	3544	3473	2.76	cps
Molybdenum	94-1	20501	20474	20451	20475	0.12	cps
Molybdenum	95-1	24621	24430	24607	24553	0.43	cps
Molybdenum	96-1	27068	26881	27185	27045	0.57	cps
Molybdenum	97-1	15595	15505	15295	15465	1.00	cps
Molybdenum	98-1	39427	39730	40269	39809	1.07	cps
Neodymium	150-1	7	20	17	14	48.02	cps
Neodymium	150-2	3	0	3	2	86.60	cps
Nickel	60-2	1830	1580	1623	1678	7.96	cps
Phosphorus	31-2	530	580	650	587	10.28	cps
Potassium	39-2	287298	283838	282976	284704	0.80	cps
Rhodium	103-1	8015346	7908637	7858405	7927463	1.01	cps
Rhodium	103-2	4527294	4455465	4487748	4490169	0.80	cps
Scandium	45-1	4805807	4757697	4703337	4755613	1.08	cps
Scandium	45-2	412578	406714	408124	409139	0.75	cps
Selenium	82-1	1239	1219	1205	1221	1.40	cps
Selenium	77-2	83	77	60	73	16.39	cps
Selenium	78-2	483	570	530	528	8.22	cps
Silicon	28-1	238759	240258	240611	239876	0.41	cps
Silver	107-1	11331	11671	11448	11484	1.51	cps
Silver	109-1	10834	11224	10908	10989	1.89	cps
Sodium	23-2	350985	346806	346411	348067	0.73	cps
Strontium	86-1	3907	3974	4017	3966	1.40	cps
Strontium	88-1	32245	32503	32623	32457	0.59	cps
Sulfur	34-1	196568	200420	200883	199290	1.19	cps
Terbium	159-1	10987361	10600949	10803534	10797282	1.79	cps
Terbium	159-2	6508371	6478328	6491340	6492680	0.23	cps
Thallium	203-1	7912	8062	8193	8056	1.74	cps
Thallium	205-1	20108	19577	19801	19829	1.34	cps
Tin	118-1	39751	40557	40483	40264	1.11	cps
Titanium	47-1	6518	6738	6431	6563	2.41	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

S02

Instrumnet Name :

P7

Client Sample ID :

S02

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 17:48:10

DataFile Name :

006CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	24636	24422	24613	24557	0.48	cps
Vanadium	51-2	18181	18418	18244	18281	0.67	cps
Yttrium	89-1	13953892	13642606	13664532	13753677	1.26	cps
Yttrium	89-2	3239158	3227454	3242256	3236289	0.24	cps
Zinc	66-2	4094	4211	4284	4196	2.28	cps
Zirconium	90-1	17514	17200	17310	17341	0.92	cps
Zirconium	91-1	3754	3981	3734	3823	3.58	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S03 Instrumnet Name : P7
Client Sample ID : S03 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:51:25 DataFile Name : 007CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	241011	239379	240727	240372	0.36	cps
Antimony	121-1	505016	517839	515545	512800	1.33	cps
Arsenic	75-2	27018	26220	27258	26832	2.03	cps
Barium	135-1	659671	670581	674117	668123	1.13	cps
Barium	137-1	1152436	1169213	1163819	1161823	0.74	cps
Beryllium	9-1	81870	83484	83194	82849	1.04	cps
Bismuth	209-1	6697697	6664627	6691911	6684745	0.26	cps
Bismuth	209-2	4767051	4756583	4784533	4769389	0.30	cps
Bromine	81-1	11098	11101	11351	11183	1.30	cps
Bromine	81-2	33	50	63	49	30.75	cps
Cadmium	108-1	9597	10144	9653	9798	3.07	cps
Cadmium	106-1	17027	17330	16890	17082	1.32	cps
Cadmium	111-1	135602	136353	136856	136270	0.46	cps
Calcium	43-1	174789	175741	176017	175516	0.37	cps
Calcium	44-1	2556821	2535019	2542597	2544813	0.43	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	226993	221362	222064	223473	1.37	cps
Cobalt	59-2	341260	329224	337583	336022	1.84	cps
Copper	63-2	2561101	2524947	2536534	2540861	0.73	cps
Dysprosium	156-1	87	83	87	86	2.25	cps
Dysprosium	156-2	80	73	40	64	33.25	cps
Erbium	164-1	70	53	80	68	19.88	cps
Erbium	164-2	83	43	33	53	49.61	cps
Gadolinium	160-1	40	30	47	39	21.57	cps
Gadolinium	160-2	243	293	243	260	11.10	cps
Holmium	165-1	10733822	10606805	10666625	10669084	0.60	cps
Holmium	165-2	6712501	6603442	6633314	6649752	0.85	cps
Indium	115-1	8516592	8515697	8540488	8524259	0.16	cps
Indium	115-2	2647441	2614466	2640780	2634229	0.66	cps
Iron	56-2	9759416	9686948	9656985	9701116	0.54	cps
Iron	57-2	253623	251899	249948	251823	0.73	cps
Iron	54-2	559862	551443	555415	555573	0.76	cps
Krypton	83-1	180	207	217	201	9.43	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S03 Instrumnet Name : P7
Client Sample ID : S03 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:51:25 DataFile Name : 007CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	2061875	2058459	2081881	2067405	0.61	cps
Lead	207-1	1832083	1844894	1843164	1840047	0.38	cps
Lead	208-1	8345487	8357577	8325354	8342806	0.20	cps
Lithium	6-1	1408945	1436250	1383805	1409667	1.86	cps
Magnesium	24-2	1993670	1958548	2020450	1990889	1.56	cps
Manganese	55-2	1641537	1641300	1573757	1618865	2.41	cps
Molybdenum	94-1	2033906	2056889	2060784	2050527	0.71	cps
Molybdenum	95-1	3016284	3031068	2995354	3014236	0.60	cps
Molybdenum	96-1	3266333	3312582	3234247	3271054	1.20	cps
Molybdenum	97-1	1857024	1903589	1862685	1874433	1.36	cps
Molybdenum	98-1	4846649	4834472	4765203	4815441	0.91	cps
Neodymium	150-1	33	40	27	33	20.00	cps
Neodymium	150-2	10	0	3	4	114.60	cps
Nickel	60-2	86555	85127	86021	85901	0.84	cps
Phosphorus	31-2	13253	12932	13182	13122	1.28	cps
Potassium	39-2	1281361	1265598	1277331	1274763	0.64	cps
Rhodium	103-1	8025118	7958816	7808420	7930784	1.40	cps
Rhodium	103-2	4557086	4445184	4520820	4507696	1.27	cps
Scandium	45-1	4818195	4795941	4706616	4773584	1.24	cps
Scandium	45-2	415505	409340	415664	413503	0.87	cps
Selenium	82-1	9056	8916	8964	8979	0.79	cps
Selenium	77-2	730	820	733	761	6.70	cps
Selenium	78-2	2657	2714	2894	2755	4.49	cps
Silicon	28-1	960693	972092	966822	966536	0.59	cps
Silver	107-1	656644	667581	674818	666348	1.37	cps
Silver	109-1	621899	631017	630082	627666	0.80	cps
Sodium	23-2	3458666	3387152	3407628	3417815	1.08	cps
Strontium	86-1	4441148	4456950	4392400	4430166	0.76	cps
Strontium	88-1	38518031	38570033	38428486	38505517	0.19	cps
Sulfur	34-1	236523	237877	234312	236237	0.76	cps
Terbium	159-1	11143549	11099329	10933547	11058808	1.00	cps
Terbium	159-2	6523712	6515713	6568946	6536123	0.44	cps
Thallium	203-1	443201	453322	449536	448687	1.14	cps
Thallium	205-1	1060879	1063479	1074893	1066417	0.70	cps
Tin	118-1	442351	442373	447640	444121	0.69	cps
Titanium	47-1	756259	766165	764336	762253	0.69	cps

LB Number :LB133265

Operator :Jaswal

Lab Sample ID :S03

Instrumnet Name :P7

Client Sample ID :S03

Dilution Factor :1

Date & Time Acquired :2024-11-02 17:51:25

DataFile Name :007CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	1333109	1345183	1362265	1346852	1.09	cps
Vanadium	51-2	197258	196070	194186	195838	0.79	cps
Yttrium	89-1	14008553	14029811	13894406	13977590	0.52	cps
Yttrium	89-2	3336009	3262764	3228269	3275680	1.68	cps
Zinc	66-2	387650	378847	383557	383351	1.15	cps
Zirconium	90-1	908039	919802	911418	913087	0.66	cps
Zirconium	91-1	203065	205271	205629	204655	0.68	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P7
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:54:26 DataFile Name : 008CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	586088	592825	585670	588194	0.68	cps
Antimony	121-1	1260919	1276324	1272778	1270007	0.64	cps
Arsenic	75-2	65878	65741	65457	65692	0.33	cps
Barium	135-1	1775568	1796205	1815912	1795895	1.12	cps
Barium	137-1	3132121	3146861	3118192	3132391	0.46	cps
Beryllium	9-1	203760	205316	206650	205242	0.70	cps
Bismuth	209-1	6663961	6689331	6683566	6678952	0.20	cps
Bismuth	209-2	4780558	4658558	4856272	4765129	2.09	cps
Bromine	81-1	10581	10424	11114	10706	3.38	cps
Bromine	81-2	37	43	37	39	9.89	cps
Cadmium	108-1	24097	24377	24868	24447	1.60	cps
Cadmium	106-1	37335	37204	37331	37290	0.20	cps
Cadmium	111-1	334012	331701	331881	332531	0.39	cps
Calcium	43-1	427736	425875	427682	427097	0.25	cps
Calcium	44-1	6148463	6127634	6102171	6126089	0.38	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	548996	542729	545086	545604	0.58	cps
Cobalt	59-2	833355	821311	823795	826154	0.77	cps
Copper	63-2	6155864	6228732	6086865	6157154	1.15	cps
Dysprosium	156-1	107	90	97	98	8.58	cps
Dysprosium	156-2	77	93	93	88	10.96	cps
Erbium	164-1	63	93	103	87	24.02	cps
Erbium	164-2	60	40	77	59	31.18	cps
Gadolinium	160-1	47	40	33	40	16.68	cps
Gadolinium	160-2	247	240	307	264	13.88	cps
Holmium	165-1	10531812	10548892	10645618	10575441	0.58	cps
Holmium	165-2	6566975	6460298	6671459	6566244	1.61	cps
Indium	115-1	8303290	8381518	8241872	8308893	0.84	cps
Indium	115-2	2604437	2546435	2608146	2586339	1.34	cps
Iron	56-2	23703342	23592953	23589028	23628441	0.27	cps
Iron	57-2	612885	614992	609941	612606	0.41	cps
Iron	54-2	1368400	1362648	1308531	1346527	2.45	cps
Krypton	83-1	207	193	200	200	3.33	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P7
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:54:26 DataFile Name : 008CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	5289930	5218386	5225153	5244490	0.75	cps
Lead	207-1	4610875	4632324	4632135	4625112	0.27	cps
Lead	208-1	20914426	21050535	21013017	20992659	0.33	cps
Lithium	6-1	1394000	1420936	1430081	1415006	1.33	cps
Magnesium	24-2	4929350	4896271	4886316	4903979	0.46	cps
Manganese	55-2	3920015	3939918	3916857	3925597	0.32	cps
Molybdenum	94-1	5043724	5056767	5027465	5042652	0.29	cps
Molybdenum	95-1	7382203	7447485	7420163	7416617	0.44	cps
Molybdenum	96-1	7980335	8140713	8066006	8062351	1.00	cps
Molybdenum	97-1	4567689	4648978	4619929	4612199	0.89	cps
Molybdenum	98-1	11796276	11969687	11977724	11914562	0.86	cps
Neodymium	150-1	73	77	60	70	12.60	cps
Neodymium	150-2	20	20	13	18	21.66	cps
Nickel	60-2	209278	211690	209993	210320	0.59	cps
Phosphorus	31-2	31442	31816	30831	31363	1.59	cps
Potassium	39-2	3075372	2983782	3005377	3021510	1.58	cps
Rhodium	103-1	7772066	7816877	7782734	7790559	0.30	cps
Rhodium	103-2	4473817	4359810	4464343	4432657	1.43	cps
Scandium	45-1	4723405	4715248	4614544	4684399	1.29	cps
Scandium	45-2	407839	406660	408218	407572	0.20	cps
Selenium	82-1	20851	21027	20846	20908	0.49	cps
Selenium	77-2	1857	1957	2007	1940	3.94	cps
Selenium	78-2	6705	6518	6642	6622	1.43	cps
Silicon	28-1	2010132	2012585	1999640	2007453	0.34	cps
Silver	107-1	1725553	1758052	1741340	1741648	0.93	cps
Silver	109-1	1633224	1652439	1642641	1642768	0.58	cps
Sodium	23-2	8393264	8256378	8303618	8317753	0.84	cps
Strontium	86-1	10787667	10887008	10786241	10820305	0.53	cps
Strontium	88-1	94233509	95191079	94474052	94632880	0.53	cps
Sulfur	34-1	284792	286852	287383	286343	0.48	cps
Terbium	159-1	11066600	10947742	10914657	10976333	0.73	cps
Terbium	159-2	6585413	6559031	6657648	6600698	0.77	cps
Thallium	203-1	1125102	1129450	1121603	1125385	0.35	cps
Thallium	205-1	3167013	3090653	3131473	3129713	1.22	cps
Tin	118-1	1088396	1098116	1102459	1096324	0.66	cps
Titanium	47-1	1838695	1833382	1829741	1833939	0.25	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P7
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:54:26 DataFile Name : 008CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	4014289	3928845	3984684	3975939	1.09	cps
Vanadium	51-2	477576	479228	475806	477537	0.36	cps
Yttrium	89-1	13694397	13682481	13506484	13627787	0.77	cps
Yttrium	89-2	3266220	3236355	3206065	3236213	0.93	cps
Zinc	66-2	943067	932846	935118	937010	0.57	cps
Zirconium	90-1	2270784	2339552	2292464	2300933	1.53	cps
Zirconium	91-1	499900	507677	506407	504661	0.83	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P7
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:57:16 DataFile Name : 009CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1163945	1167167	1162835	1164649	0.19	cps
Antimony	121-1	2665602	2726990	2713710	2702101	1.20	cps
Arsenic	75-2	131075	130538	130464	130692	0.26	cps
Barium	135-1	3620218	3611926	3649569	3627238	0.55	cps
Barium	137-1	6220980	6239737	6234486	6231734	0.16	cps
Beryllium	9-1	402020	410069	412551	408213	1.35	cps
Bismuth	209-1	6843069	6724024	6795430	6787508	0.88	cps
Bismuth	209-2	4728986	4749715	4793378	4757359	0.69	cps
Bromine	81-1	11044	10277	10397	10573	3.90	cps
Bromine	81-2	50	40	37	42	16.43	cps
Cadmium	108-1	48223	48079	48949	48417	0.96	cps
Cadmium	106-1	70644	71033	70785	70820	0.28	cps
Cadmium	111-1	655665	665366	659545	660192	0.74	cps
Calcium	43-1	834522	837091	839617	837077	0.30	cps
Calcium	44-1	12153769	12030954	11977935	12054219	0.75	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1070136	1078469	1062221	1070275	0.76	cps
Cobalt	59-2	1664002	1656270	1677677	1665983	0.65	cps
Copper	63-2	12114436	12117137	11923487	12051687	0.92	cps
Dysprosium	156-1	137	90	147	124	24.31	cps
Dysprosium	156-2	117	137	187	147	24.58	cps
Erbium	164-1	90	123	133	116	19.64	cps
Erbium	164-2	47	47	77	57	30.56	cps
Gadolinium	160-1	110	70	57	79	35.18	cps
Gadolinium	160-2	317	283	253	284	11.14	cps
Holmium	165-1	10768459	10717752	10793680	10759964	0.36	cps
Holmium	165-2	6707393	6679370	6648869	6678544	0.44	cps
Indium	115-1	8298160	8348495	8270277	8305644	0.48	cps
Indium	115-2	2548920	2580521	2538648	2556030	0.85	cps
Iron	56-2	46684484	46557624	46377143	46539750	0.33	cps
Iron	57-2	1210882	1204281	1207913	1207692	0.27	cps
Iron	54-2	2601043	2605835	2625551	2610809	0.50	cps
Krypton	83-1	207	180	210	199	8.27	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P7
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:57:16 DataFile Name : 009CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	10625190	10767047	10606306	10666181	0.82	cps
Lead	207-1	9383755	9513420	9493544	9463573	0.74	cps
Lead	208-1	42961737	43128039	43294849	43128208	0.39	cps
Lithium	6-1	1399714	1422116	1453121	1424984	1.88	cps
Magnesium	24-2	9720190	9800311	9577841	9699448	1.16	cps
Manganese	55-2	7864436	7724663	7731381	7773493	1.01	cps
Molybdenum	94-1	10023311	10044598	9913026	9993645	0.71	cps
Molybdenum	95-1	14798531	14923079	14734544	14818718	0.65	cps
Molybdenum	96-1	16135441	16125217	15970829	16077162	0.57	cps
Molybdenum	97-1	9223167	9171108	9235360	9209878	0.37	cps
Molybdenum	98-1	23559503	23853704	23803892	23739033	0.66	cps
Neodymium	150-1	193	127	157	159	21.01	cps
Neodymium	150-2	50	37	43	43	15.38	cps
Nickel	60-2	414616	411721	410986	412441	0.47	cps
Phosphorus	31-2	62664	61473	61737	61958	1.01	cps
Potassium	39-2	5964635	5921120	5907178	5930978	0.51	cps
Rhodium	103-1	7749732	7675060	7703255	7709349	0.49	cps
Rhodium	103-2	4408444	4368459	4365477	4380794	0.55	cps
Scandium	45-1	4627213	4526340	4528078	4560543	1.27	cps
Scandium	45-2	403964	404406	401383	403251	0.40	cps
Selenium	82-1	40526	40983	40858	40789	0.58	cps
Selenium	77-2	3677	3734	3837	3749	2.16	cps
Selenium	78-2	12632	12982	13009	12875	1.63	cps
Silicon	28-1	3678552	3664513	3684112	3675726	0.27	cps
Silver	107-1	3421965	3436447	3453232	3437214	0.46	cps
Silver	109-1	3240727	3220380	3267504	3242870	0.73	cps
Sodium	23-2	16550532	16560276	16399739	16503516	0.55	cps
Strontium	86-1	21343755	21548746	21554482	21482328	0.56	cps
Strontium	88-1	188151571	189430564	186802471	188128202	0.70	cps
Sulfur	34-1	380361	381242	382571	381391	0.29	cps
Terbium	159-1	11139155	11110332	11087236	11112241	0.23	cps
Terbium	159-2	6672538	6679993	6618179	6656903	0.51	cps
Thallium	203-1	2672604	2628986	2669384	2656991	0.91	cps
Thallium	205-1	6240294	6313783	6383375	6312484	1.13	cps
Tin	118-1	2291283	2318837	2362442	2324188	1.54	cps
Titanium	47-1	3595031	3608164	3609709	3604301	0.22	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P7
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 17:57:16 DataFile Name : 009CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8040296	8237025	8367719	8215013	2.01	cps
Vanadium	51-2	945937	953931	941728	947198	0.65	cps
Yttrium	89-1	13829426	13670449	13675422	13725099	0.66	cps
Yttrium	89-2	3253586	3230192	3289532	3257770	0.92	cps
Zinc	66-2	1897367	1902331	1895011	1898236	0.20	cps
Zirconium	90-1	4532954	4579496	4626937	4579796	1.03	cps
Zirconium	91-1	1002584	1009367	1003837	1005263	0.36	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P7
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:00:04 DataFile Name : 010CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	2278264	2232535	2271373	2260724	1.09	cps
Antimony	121-1	5233930	5324500	5405252	5321227	1.61	cps
Arsenic	75-2	255399	253422	252018	253613	0.67	cps
Barium	135-1	7034738	7171176	7181217	7129044	1.15	cps
Barium	137-1	12154511	12336182	12391107	12293934	1.01	cps
Beryllium	9-1	789566	801818	807939	799775	1.17	cps
Bismuth	209-1	6845608	6860946	6784171	6830242	0.59	cps
Bismuth	209-2	4839830	4721327	4809736	4790298	1.29	cps
Bromine	81-1	10977	10694	10520	10731	2.15	cps
Bromine	81-2	40	80	73	64	33.25	cps
Cadmium	108-1	92915	96186	96830	95310	2.20	cps
Cadmium	106-1	134663	139105	138042	137270	1.69	cps
Cadmium	111-1	1278348	1294745	1293016	1288703	0.70	cps
Calcium	43-1	1557504	1563531	1569294	1563443	0.38	cps
Calcium	44-1	23231909	23446713	23355386	23344669	0.46	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2092249	2090221	2077068	2086513	0.40	cps
Cobalt	59-2	3198997	3257136	3187988	3214707	1.16	cps
Copper	63-2	23391929	23302456	23239870	23311418	0.33	cps
Dysprosium	156-1	137	160	160	152	8.85	cps
Dysprosium	156-2	227	303	243	258	15.64	cps
Erbium	164-1	120	120	140	127	9.12	cps
Erbium	164-2	90	73	67	77	15.68	cps
Gadolinium	160-1	90	103	80	91	12.85	cps
Gadolinium	160-2	287	250	323	287	12.79	cps
Holmium	165-1	10812762	11113100	10934251	10953371	1.38	cps
Holmium	165-2	6879265	6700092	6911291	6830216	1.67	cps
Indium	115-1	8175202	8248283	8029902	8151129	1.36	cps
Indium	115-2	2593302	2513897	2523890	2543696	1.70	cps
Iron	56-2	92012855	92179469	91611715	91934680	0.32	cps
Iron	57-2	2299839	2284490	2292878	2292402	0.34	cps
Iron	54-2	5099793	5089190	5082359	5090447	0.17	cps
Krypton	83-1	247	213	190	217	13.15	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P7
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:00:04 DataFile Name : 010CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	21203796	21564821	21779706	21516108	1.35	cps
Lead	207-1	18770638	19027888	19201123	18999883	1.14	cps
Lead	208-1	85384047	86278766	87149129	86270647	1.02	cps
Lithium	6-1	1432743	1447591	1419974	1433436	0.96	cps
Magnesium	24-2	19044994	18805880	18954228	18935034	0.64	cps
Manganese	55-2	15301104	15354035	15383999	15346379	0.27	cps
Molybdenum	94-1	19701207	20001422	19897666	19866765	0.77	cps
Molybdenum	95-1	28829418	29445551	29024035	29099668	1.08	cps
Molybdenum	96-1	31363677	31809920	31873694	31682430	0.88	cps
Molybdenum	97-1	17959471	18168970	18466749	18198397	1.40	cps
Molybdenum	98-1	46194101	46921666	47246384	46787384	1.15	cps
Neodymium	150-1	307	247	283	279	10.85	cps
Neodymium	150-2	50	33	77	53	40.99	cps
Nickel	60-2	796434	791388	792241	793354	0.34	cps
Phosphorus	31-2	120029	120133	120163	120109	0.06	cps
Potassium	39-2	11485691	11482141	11523563	11497131	0.20	cps
Rhodium	103-1	7531840	7589476	7474661	7531992	0.76	cps
Rhodium	103-2	4475624	4287682	4330790	4364699	2.26	cps
Scandium	45-1	4475497	4432754	4427807	4445353	0.59	cps
Scandium	45-2	414607	398790	399906	404434	2.18	cps
Selenium	82-1	78971	80311	80560	79947	1.07	cps
Selenium	77-2	7465	7245	7145	7285	2.25	cps
Selenium	78-2	24046	23789	23656	23830	0.83	cps
Silicon	28-1	6845889	6840249	6906056	6864065	0.53	cps
Silver	107-1	6578001	6685069	6689694	6650921	0.95	cps
Silver	109-1	6298982	6256792	6327515	6294430	0.57	cps
Sodium	23-2	32608603	32331532	32505599	32481911	0.43	cps
Strontium	86-1	41972253	42405034	42317251	42231513	0.54	cps
Strontium	88-1	367875688	370510648	373640048	370675461	0.78	cps
Sulfur	34-1	569442	571980	571029	570817	0.22	cps
Terbium	159-1	11077848	11402277	11183844	11221323	1.47	cps
Terbium	159-2	6989043	6758274	6713134	6820151	2.17	cps
Thallium	203-1	5313509	5502728	5440425	5418887	1.78	cps
Thallium	205-1	12770040	12842816	13017581	12876812	0.99	cps
Tin	118-1	4547763	4626262	4600351	4591459	0.87	cps
Titanium	47-1	6922017	7019026	6939940	6960327	0.74	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P7
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:00:04 DataFile Name : 010CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	16415292	17035105	17016294	16822230	2.10	cps
Vanadium	51-2	1832658	1856218	1840921	1843266	0.65	cps
Yttrium	89-1	13738168	13668314	13476603	13627695	0.99	cps
Yttrium	89-2	3332015	3204370	3220737	3252374	2.14	cps
Zinc	66-2	3691118	3670852	3682550	3681507	0.28	cps
Zirconium	90-1	8875981	8999792	8985440	8953738	0.76	cps
Zirconium	91-1	1971791	2002338	2020430	1998186	1.23	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P7
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:02:47 DataFile Name : 011CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4439985	4470503	4441197	4450562	0.39	cps
Antimony	121-1	10475457	10643918	10699258	10606211	1.10	cps
Arsenic	75-2	500236	501353	502165	501251	0.19	cps
Barium	135-1	14074249	14186494	14391896	14217547	1.13	cps
Barium	137-1	24404317	24785696	24612815	24600943	0.78	cps
Beryllium	9-1	1523656	1541064	1561409	1542043	1.23	cps
Bismuth	209-1	6688831	6823256	6690371	6734153	1.15	cps
Bismuth	209-2	4645218	4639140	4614811	4633056	0.35	cps
Bromine	81-1	10314	11001	10691	10668	3.23	cps
Bromine	81-2	43	80	73	66	29.80	cps
Cadmium	108-1	185115	185209	185744	185356	0.18	cps
Cadmium	106-1	262854	267366	266683	265634	0.92	cps
Cadmium	111-1	2669030	2728349	2703138	2700172	1.10	cps
Calcium	43-1	3071706	3104530	3076812	3084349	0.57	cps
Calcium	44-1	45988604	46455729	46516969	46320434	0.62	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	4112360	4101962	4103067	4105796	0.14	cps
Cobalt	59-2	6293136	6271596	6343228	6302653	0.58	cps
Copper	63-2	45632223	45743279	45913406	45762969	0.31	cps
Dysprosium	156-1	273	290	273	279	3.45	cps
Dysprosium	156-2	517	623	623	588	10.48	cps
Erbium	164-1	217	217	220	218	0.88	cps
Erbium	164-2	140	127	127	131	5.87	cps
Gadolinium	160-1	167	117	150	144	17.62	cps
Gadolinium	160-2	323	303	350	326	7.19	cps
Holmium	165-1	10880684	11034636	10945994	10953771	0.71	cps
Holmium	165-2	6700892	6721366	6754860	6725706	0.41	cps
Indium	115-1	8013890	7915903	7934168	7954654	0.66	cps
Indium	115-2	2432706	2420534	2432403	2428548	0.29	cps
Iron	56-2	179926044	180093751	181911297	180643697	0.61	cps
Iron	57-2	4517749	4460247	4541946	4506647	0.93	cps
Iron	54-2	10020727	9967162	10141457	10043115	0.89	cps
Krypton	83-1	170	177	213	187	12.50	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P7
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:02:47 DataFile Name : 011CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	42450336	43047459	43706199	43067998	1.46	cps
Lead	207-1	37709093	37960324	38765781	38145066	1.45	cps
Lead	208-1	171413461	172979077	175107152	173166564	1.07	cps
Lithium	6-1	1410417	1452072	1406673	1423054	1.77	cps
Magnesium	24-2	37963113	37916552	38101969	37993878	0.25	cps
Manganese	55-2	30066175	30158622	30346869	30190555	0.47	cps
Molybdenum	94-1	38919664	39509940	39598923	39342842	0.94	cps
Molybdenum	95-1	57437557	57669889	58038484	57715310	0.53	cps
Molybdenum	96-1	63481969	62790181	62963639	63078596	0.57	cps
Molybdenum	97-1	36191612	36145134	36215359	36184035	0.10	cps
Molybdenum	98-1	93611579	93651595	93800209	93687794	0.11	cps
Neodymium	150-1	610	573	630	604	4.75	cps
Neodymium	150-2	133	157	123	138	12.41	cps
Nickel	60-2	1641595	1629550	1626570	1632572	0.49	cps
Phosphorus	31-2	235152	236019	236658	235943	0.32	cps
Potassium	39-2	23064604	22807446	23138986	23003679	0.76	cps
Rhodium	103-1	7320476	7333599	7204205	7286093	0.98	cps
Rhodium	103-2	4220064	4231889	4201062	4217672	0.37	cps
Scandium	45-1	4442732	4479519	4427008	4449753	0.61	cps
Scandium	45-2	400642	395739	399112	398498	0.63	cps
Selenium	82-1	153954	154822	155439	154738	0.48	cps
Selenium	77-2	14147	14497	14214	14286	1.30	cps
Selenium	78-2	47138	46677	47365	47060	0.75	cps
Silicon	28-1	17981868	17934366	17938046	17951427	0.15	cps
Silver	107-1	12769373	13059419	12964714	12931169	1.14	cps
Silver	109-1	12093626	12287851	12317326	12232934	0.99	cps
Sodium	23-2	64995694	65484021	65240596	65240103	0.37	cps
Strontium	86-1	83253487	83912755	83981895	83716046	0.48	cps
Strontium	88-1	726783403	738784869	742118282	735895518	1.10	cps
Sulfur	34-1	947877	957405	955038	953440	0.52	cps
Terbium	159-1	11174380	11398202	11198790	11257124	1.09	cps
Terbium	159-2	6646607	6744135	6686058	6692267	0.73	cps
Thallium	203-1	10614316	10802556	10893136	10770002	1.32	cps
Thallium	205-1	25437547	25932420	26354097	25908021	1.77	cps
Tin	118-1	9100145	9113109	9156879	9123378	0.33	cps
Titanium	47-1	13811958	13859945	13912579	13861494	0.36	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P7
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:02:47 DataFile Name : 011CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	33422435	34715742	35278360	34472179	2.76	cps
Vanadium	51-2	3622267	3654135	3612245	3629549	0.60	cps
Yttrium	89-1	13379394	13467066	13297818	13381426	0.63	cps
Yttrium	89-2	3231480	3137141	3196806	3188476	1.50	cps
Zinc	66-2	7239340	7152563	7238204	7210036	0.69	cps
Zirconium	90-1	17615407	17964517	17996353	17858759	1.18	cps
Zirconium	91-1	3959155	4005086	4010518	3991586	0.71	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S08 Instrumnet Name : P7
Client Sample ID : S08 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:05:33 DataFile Name : 012CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	23837315	24015232	23809193	23887247	0.47	cps
Antimony	121-1	6778	6882	7465	7042	5.26	cps
Arsenic	75-2	317	253	320	297	12.66	cps
Barium	135-1	5018	5381	5264	5221	3.55	cps
Barium	137-1	9186	9427	9253	9289	1.33	cps
Beryllium	9-1	222	217	222	220	1.46	cps
Bismuth	209-1	6182766	5781594	5731770	5898710	4.19	cps
Bismuth	209-2	4104022	4161817	4098091	4121310	0.85	cps
Bromine	81-1	11568	11608	11461	11546	0.66	cps
Bromine	81-2	67	73	57	66	12.79	cps
Cadmium	108-1	50	87	97	78	31.59	cps
Cadmium	106-1	2900	2964	2857	2907	1.85	cps
Cadmium	111-1	2482	2401	2379	2420	2.24	cps
Calcium	43-1	13244432	13731760	13796976	13591056	2.22	cps
Calcium	44-1	214701097	222072890	224471083	220415023	2.31	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	13403	13423	13356	13394	0.26	cps
Cobalt	59-2	16803	16926	17097	16942	0.87	cps
Copper	63-2	9793	9997	9940	9910	1.06	cps
Dysprosium	156-1	433	477	383	431	10.83	cps
Dysprosium	156-2	243	340	350	311	18.94	cps
Erbium	164-1	517	493	380	463	15.78	cps
Erbium	164-2	243	257	270	257	5.20	cps
Gadolinium	160-1	363	360	343	356	3.01	cps
Gadolinium	160-2	497	463	480	480	3.47	cps
Holmium	165-1	10475709	10320613	10052358	10282893	2.08	cps
Holmium	165-2	6440358	6519108	6405892	6455119	0.90	cps
Indium	115-1	7692857	7687930	7648899	7676562	0.31	cps
Indium	115-2	2404512	2438142	2428297	2423650	0.71	cps
Iron	56-2	892568867	903146413	897846707	897853996	0.59	cps
Iron	57-2	22325430	22595124	22424948	22448501	0.61	cps
Iron	54-2	48830413	48824406	48641423	48765414	0.22	cps
Krypton	83-1	220	220	260	233	9.90	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : S08 Instrumnet Name : P7
Client Sample ID : S08 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:05:33 DataFile Name : 012CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	7689	6992	7102	7261	5.16	cps
Lead	207-1	6492	6442	6348	6427	1.13	cps
Lead	208-1	30340	29216	28796	29451	2.71	cps
Lithium	6-1	1372788	1373502	1341815	1362702	1.33	cps
Magnesium	24-2	197472657	200057850	199140097	198890201	0.66	cps
Manganese	55-2	15496	14293	14223	14671	4.88	cps
Molybdenum	94-1	9390	9186	9136	9237	1.45	cps
Molybdenum	95-1	8543	8356	9036	8645	4.07	cps
Molybdenum	96-1	20745	21222	22260	21409	3.62	cps
Molybdenum	97-1	5378	5621	5438	5479	2.31	cps
Molybdenum	98-1	13096	12769	12575	12813	2.05	cps
Neodymium	150-1	223	207	223	218	4.42	cps
Neodymium	150-2	107	103	97	102	4.98	cps
Nickel	60-2	5658	5778	5858	5764	1.75	cps
Phosphorus	31-2	230	253	220	234	7.30	cps
Potassium	39-2	119334088	121448648	120229715	120337484	0.88	cps
Rhodium	103-1	6562444	6588604	6758913	6636654	1.61	cps
Rhodium	103-2	3901762	3912781	3887832	3900792	0.32	cps
Scandium	45-1	4415083	4464216	4487773	4455690	0.83	cps
Scandium	45-2	409220	408748	407351	408440	0.24	cps
Selenium	82-1	632	586	609	609	3.75	cps
Selenium	77-2	10	10	10	10	0.00	cps
Selenium	78-2	237	270	220	242	10.51	cps
Silicon	28-1	282624	281009	276859	280164	1.06	cps
Silver	107-1	2007	1737	1543	1762	13.21	cps
Silver	109-1	1807	1613	1513	1645	9.07	cps
Sodium	23-2	335339575	334328555	336570682	335412937	0.33	cps
Strontium	86-1	13923	13730	14117	13923	1.39	cps
Strontium	88-1	122777	122498	120368	121881	1.08	cps
Sulfur	34-1	158115	164306	166267	162896	2.61	cps
Terbium	159-1	10650539	10714550	10979024	10781371	1.62	cps
Terbium	159-2	6433075	6386809	6453054	6424313	0.53	cps
Thallium	203-1	1357	1123	1163	1215	10.28	cps
Thallium	205-1	3224	3180	2837	3080	6.88	cps
Tin	118-1	4701	4751	5221	4891	5.87	cps
Titanium	47-1	1740	1997	2030	1922	8.26	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

S08

Instrumnet Name :

P7

Client Sample ID :

S08

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:05:33

DataFile Name :

012CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	2047	1790	1917	1918	6.69	cps
Vanadium	51-2	840	867	777	828	5.58	cps
Yttrium	89-1	13034264	12972006	13075443	13027238	0.40	cps
Yttrium	89-2	3143417	3182248	3179998	3168554	0.69	cps
Zinc	66-2	3167	3344	3154	3221	3.29	cps
Zirconium	90-1	10080	10040	10237	10119	1.03	cps
Zirconium	91-1	2210	2380	2424	2338	4.82	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICV001 Instrumnet Name : P7
Client Sample ID : ICV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:09:41 DataFile Name : 013ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	122188	123694	125017	123633	1.14	cps
Antimony	121-1	2146959	2163244	2207306	2172503	1.44	cps
Arsenic	75-2	108093	108220	108654	108323	0.27	cps
Barium	135-1	261565	262361	266202	263376	0.94	cps
Barium	137-1	455496	459899	461025	458807	0.64	cps
Beryllium	9-1	167014	166778	168862	167551	0.68	cps
Bismuth	209-1	6955017	6849729	6812246	6872331	1.08	cps
Bismuth	209-2	4815314	4828571	4768862	4804249	0.65	cps
Bromine	81-1	10280	10387	10644	10437	1.79	cps
Bromine	81-2	77	57	77	70	16.49	cps
Cadmium	108-1	17184	17247	17624	17352	1.37	cps
Cadmium	106-1	26270	26875	26698	26614	1.17	cps
Cadmium	111-1	266524	271219	272341	270028	1.14	cps
Calcium	43-1	59774	61377	61450	60867	1.56	cps
Calcium	44-1	980805	1001082	1005004	995630	1.30	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	451779	453166	450781	451909	0.27	cps
Cobalt	59-2	677579	682783	683479	681280	0.47	cps
Copper	63-2	462633	465088	466618	464780	0.43	cps
Dysprosium	156-1	60	33	47	47	28.58	cps
Dysprosium	156-2	27	20	43	30	40.05	cps
Erbium	164-1	63	77	43	61	27.46	cps
Erbium	164-2	57	50	63	57	11.76	cps
Gadolinium	160-1	53	50	73	59	21.43	cps
Gadolinium	160-2	260	283	253	266	5.93	cps
Holmium	165-1	10958575	10835485	10792136	10862066	0.79	cps
Holmium	165-2	6680076	6677273	6630473	6662607	0.42	cps
Indium	115-1	8466945	8425594	8446118	8446219	0.24	cps
Indium	115-2	2657573	2665142	2650078	2657598	0.28	cps
Iron	56-2	7714704	7795744	7833313	7781254	0.78	cps
Iron	57-2	199887	200874	202590	201117	0.68	cps
Iron	54-2	456104	455447	460219	457257	0.57	cps
Krypton	83-1	190	220	193	201	8.18	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICV001 Instrumnet Name : P7
Client Sample ID : ICV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:09:41 DataFile Name : 013ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1915373	1877089	1847174	1879879	1.82	cps
Lead	207-1	1371664	1374918	1369123	1371902	0.21	cps
Lead	208-1	7265869	7189437	7132943	7196083	0.93	cps
Lithium	6-1	1480707	1491573	1497666	1489982	0.58	cps
Magnesium	24-2	523106	530792	531479	528459	0.88	cps
Manganese	55-2	322390	329463	324979	325610	1.10	cps
Molybdenum	94-1	1930	2184	2007	2040	6.37	cps
Molybdenum	95-1	2224	2164	2114	2167	2.54	cps
Molybdenum	96-1	2467	2620	2504	2530	3.16	cps
Molybdenum	97-1	1363	1330	1330	1341	1.44	cps
Molybdenum	98-1	3150	3434	3624	3403	7.00	cps
Neodymium	150-1	13	27	20	20	33.35	cps
Neodymium	150-2	7	10	0	6	91.64	cps
Nickel	60-2	176084	176194	176569	176282	0.14	cps
Phosphorus	31-2	223	240	260	241	7.61	cps
Potassium	39-2	1079395	1091750	1091338	1087494	0.65	cps
Rhodium	103-1	7752328	7824947	7815701	7797659	0.51	cps
Rhodium	103-2	4551559	4584070	4591632	4575753	0.47	cps
Scandium	45-1	4572726	4673038	4662822	4636195	1.19	cps
Scandium	45-2	421906	423726	424901	423511	0.36	cps
Selenium	82-1	34061	34831	34868	34586	1.32	cps
Selenium	77-2	3284	3137	3230	3217	2.31	cps
Selenium	78-2	10681	10974	11068	10907	1.85	cps
Silicon	28-1	217276	225432	227697	223469	2.45	cps
Silver	107-1	621405	638708	644208	634774	1.87	cps
Silver	109-1	591261	603276	608008	600848	1.44	cps
Sodium	23-2	1450572	1458891	1466592	1458685	0.55	cps
Strontium	86-1	3620	3724	3434	3593	4.09	cps
Strontium	88-1	29079	29683	29272	29345	1.05	cps
Sulfur	34-1	173304	175569	180014	176295	1.94	cps
Terbium	159-1	11178831	11177310	10997464	11117868	0.94	cps
Terbium	159-2	6638101	6771286	6748923	6719437	1.06	cps
Thallium	203-1	2387600	2387130	2334208	2369646	1.30	cps
Thallium	205-1	5635470	5638152	5585325	5619649	0.53	cps
Tin	118-1	2364	2394	2810	2522	9.90	cps
Titanium	47-1	593	607	710	637	10.03	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICV001 Instrumnet Name : P7
Client Sample ID : ICV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:09:41 DataFile Name : 013ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	973	857	860	897	7.41	cps
Vanadium	51-2	385831	388177	390770	388259	0.64	cps
Yttrium	89-1	13610686	13610800	13645849	13622445	0.15	cps
Yttrium	89-2	3311588	3293900	3354854	3320114	0.94	cps
Zinc	66-2	150894	151654	151722	151423	0.30	cps
Zirconium	90-1	2227	2214	2320	2254	2.58	cps
Zirconium	91-1	680	600	767	682	12.22	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICB001 Instrumnet Name : P7
Client Sample ID : ICB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:12:57 DataFile Name : 014CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	903	890	817	870	5.36	cps
Antimony	121-1	1527	1717	1823	1689	8.90	cps
Arsenic	75-2	33	23	23	27	21.65	cps
Barium	135-1	370	347	470	396	16.56	cps
Barium	137-1	643	613	667	641	4.17	cps
Beryllium	9-1	161	100	144	135	23.37	cps
Bismuth	209-1	6849812	6722817	6707937	6760188	1.15	cps
Bismuth	209-2	4907761	4295742	4558556	4587353	6.69	cps
Bromine	81-1	11141	11745	11288	11391	2.76	cps
Bromine	81-2	60	80	83	74	16.95	cps
Cadmium	108-1	13	7	17	12	41.65	cps
Cadmium	106-1	2977	2964	2830	2924	2.77	cps
Cadmium	111-1	2225	2173	2063	2154	3.84	cps
Calcium	43-1	640	587	690	639	8.09	cps
Calcium	44-1	17594	17774	17971	17780	1.06	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	667	730	653	683	5.99	cps
Cobalt	59-2	177	200	207	194	8.10	cps
Copper	63-2	2030	2087	2067	2061	1.39	cps
Dysprosium	156-1	33	37	40	37	9.10	cps
Dysprosium	156-2	27	23	20	23	14.29	cps
Erbium	164-1	73	53	67	64	15.80	cps
Erbium	164-2	37	37	30	34	11.18	cps
Gadolinium	160-1	53	73	57	61	17.53	cps
Gadolinium	160-2	307	257	250	271	11.42	cps
Holmium	165-1	10978172	10863660	10523329	10788387	2.19	cps
Holmium	165-2	6787438	5959608	6178680	6308576	6.80	cps
Indium	115-1	8572305	8786090	8567833	8642076	1.44	cps
Indium	115-2	2732914	2415685	2498419	2549006	6.46	cps
Iron	56-2	34480	34463	32973	33972	2.55	cps
Iron	57-2	1937	1800	1990	1909	5.13	cps
Iron	54-2	2820	3014	3000	2945	3.67	cps
Krypton	83-1	190	177	227	198	13.09	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICB001 Instrumnet Name : P7
Client Sample ID : ICB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:12:57 DataFile Name : 014CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1960	1984	2057	2000	2.52	cps
Lead	207-1	1913	1820	1913	1882	2.86	cps
Lead	208-1	8154	8181	8328	8221	1.14	cps
Lithium	6-1	1449528	1452883	1455459	1452624	0.20	cps
Magnesium	24-2	5688	5624	5454	5589	2.16	cps
Manganese	55-2	920	877	873	890	2.93	cps
Molybdenum	94-1	1107	1117	1273	1166	8.02	cps
Molybdenum	95-1	1363	1423	1500	1429	4.79	cps
Molybdenum	96-1	1547	1677	1700	1641	5.03	cps
Molybdenum	97-1	907	957	930	931	2.69	cps
Molybdenum	98-1	2240	2227	2317	2261	2.15	cps
Neodymium	150-1	10	3	7	7	50.03	cps
Neodymium	150-2	0	10	0	3	173.21	cps
Nickel	60-2	93	117	57	89	34.03	cps
Phosphorus	31-2	177	237	207	207	14.52	cps
Potassium	39-2	58616	58877	57826	58439	0.94	cps
Rhodium	103-1	8148301	8112913	8062805	8108006	0.53	cps
Rhodium	103-2	4720816	4002330	4266119	4329755	8.39	cps
Scandium	45-1	4906413	4867034	4861471	4878306	0.50	cps
Scandium	45-2	430008	385889	400152	405350	5.55	cps
Selenium	82-1	499	505	413	472	10.90	cps
Selenium	77-2	0	0	3	1	173.21	cps
Selenium	78-2	263	253	280	266	5.07	cps
Silicon	28-1	222218	225784	235034	227679	2.91	cps
Silver	107-1	740	683	617	680	9.08	cps
Silver	109-1	613	613	627	618	1.25	cps
Sodium	23-2	54372	53389	52138	53300	2.10	cps
Strontium	86-1	2127	2177	2000	2101	4.33	cps
Strontium	88-1	15588	16079	16022	15896	1.69	cps
Sulfur	34-1	181079	182303	183360	182248	0.63	cps
Terbium	159-1	11151773	10906452	10944737	11000987	1.20	cps
Terbium	159-2	6694226	5925420	6241343	6286996	6.15	cps
Thallium	203-1	673	733	617	674	8.65	cps
Thallium	205-1	1840	1723	1643	1736	5.70	cps
Tin	118-1	1647	1790	1753	1730	4.30	cps
Titanium	47-1	363	400	453	406	11.16	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

ICB001

Instrumnet Name :

P7

Client Sample ID :

ICB001

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:12:57

DataFile Name :

014CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	670	683	663	672	1.51	cps
Vanadium	51-2	113	127	103	114	10.23	cps
Yttrium	89-1	14112592	14161726	14054117	14109479	0.38	cps
Yttrium	89-2	3342709	3032922	3141249	3172293	4.96	cps
Zinc	66-2	883	840	857	860	2.54	cps
Zirconium	90-1	1087	1133	1137	1119	2.50	cps
Zirconium	91-1	177	277	243	232	21.93	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICSA001 Instrumnet Name : P7
Client Sample ID : ICSA001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:16:13 DataFile Name : 015ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	21610757	21654160	21494950	21586622	0.38	cps
Antimony	121-1	11755	11992	12539	12095	3.32	cps
Arsenic	75-2	203	260	227	230	12.38	cps
Barium	135-1	3934	3964	3981	3959	0.60	cps
Barium	137-1	6768	6852	6618	6746	1.75	cps
Beryllium	9-1	578	483	517	526	9.11	cps
Bismuth	209-1	6662025	6556821	6657904	6625583	0.90	cps
Bismuth	209-2	4649005	4613812	4635852	4632890	0.38	cps
Bromine	81-1	11438	11188	11598	11408	1.81	cps
Bromine	81-2	180	143	127	150	18.19	cps
Cadmium	108-1	1287	1370	1253	1303	4.61	cps
Cadmium	106-1	2770	2830	2874	2825	1.84	cps
Cadmium	111-1	2812	2882	2921	2872	1.92	cps
Calcium	43-1	2755362	2750661	2789983	2765335	0.78	cps
Calcium	44-1	43834678	44155204	44435743	44141875	0.68	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	85732	87471	86355	86519	1.02	cps
Cobalt	59-2	8179	7939	8076	8064	1.49	cps
Copper	63-2	35145	35472	34025	34881	2.17	cps
Dysprosium	156-1	47	57	60	54	12.74	cps
Dysprosium	156-2	60	40	47	49	20.83	cps
Erbium	164-1	127	123	117	122	4.17	cps
Erbium	164-2	60	83	57	67	21.79	cps
Gadolinium	160-1	70	53	87	70	23.81	cps
Gadolinium	160-2	320	293	330	314	6.03	cps
Holmium	165-1	11118018	11101814	10907067	11042300	1.06	cps
Holmium	165-2	6796264	6890512	6764887	6817221	0.96	cps
Indium	115-1	8393640	8249581	8309042	8317421	0.87	cps
Indium	115-2	2565970	2600422	2616893	2594428	1.00	cps
Iron	56-2	363658141	369647168	368782861	367362723	0.88	cps
Iron	57-2	9117765	9138190	9154183	9136713	0.20	cps
Iron	54-2	20099233	20321784	20109304	20176774	0.62	cps
Krypton	83-1	213	247	237	232	7.37	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICSA001 Instrumnet Name : P7
Client Sample ID : ICSA001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:16:13 DataFile Name : 015ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	33932	34828	35677	34812	2.51	cps
Lead	207-1	27388	28062	28517	27989	2.03	cps
Lead	208-1	129956	132393	133693	132014	1.44	cps
Lithium	6-1	1442819	1433988	1434388	1437065	0.35	cps
Magnesium	24-2	39016906	38686607	38915009	38872841	0.44	cps
Manganese	55-2	25141	24797	25074	25004	0.73	cps
Molybdenum	94-1	5808504	5869376	5831928	5836603	0.53	cps
Molybdenum	95-1	10372247	10349580	10376689	10366172	0.14	cps
Molybdenum	96-1	11026779	10990394	11092952	11036709	0.47	cps
Molybdenum	97-1	6273257	6366920	6529974	6390050	2.03	cps
Molybdenum	98-1	16320227	16617946	16610205	16516126	1.03	cps
Neodymium	150-1	50	50	53	51	3.76	cps
Neodymium	150-2	33	17	27	26	32.81	cps
Nickel	60-2	8686	8979	8666	8777	2.00	cps
Phosphorus	31-2	1199474	1180650	1185441	1188522	0.82	cps
Potassium	39-2	46834008	46933458	46671413	46812959	0.28	cps
Rhodium	103-1	7600020	7438501	7280017	7439513	2.15	cps
Rhodium	103-2	4309877	4298297	4308105	4305426	0.14	cps
Scandium	45-1	4801421	4689491	4650280	4713730	1.66	cps
Scandium	45-2	414837	418201	415146	416061	0.45	cps
Selenium	82-1	474	404	437	439	8.03	cps
Selenium	77-2	3	7	3	4	43.40	cps
Selenium	78-2	270	227	253	250	8.74	cps
Silicon	28-1	312939	315913	319420	316091	1.03	cps
Silver	107-1	1493	1370	1310	1391	6.72	cps
Silver	109-1	1490	1190	1047	1242	18.21	cps
Sodium	23-2	66265052	67013127	66562294	66613491	0.57	cps
Strontium	86-1	115327	115156	115583	115355	0.19	cps
Strontium	88-1	998857	1015355	1010524	1008245	0.84	cps
Sulfur	34-1	4042691	4078753	4080418	4067287	0.52	cps
Terbium	159-1	11301200	11323263	11314250	11312904	0.10	cps
Terbium	159-2	6772244	6745787	6847371	6788467	0.78	cps
Thallium	203-1	1390	1383	1310	1361	3.26	cps
Thallium	205-1	3394	3147	3134	3225	4.54	cps
Tin	118-1	2887	2884	2930	2900	0.90	cps
Titanium	47-1	2606048	2598652	2639668	2614789	0.84	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICSA001 Instrumnet Name : P7
Client Sample ID : ICSA001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:16:13 DataFile Name : 015ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	843	1033	877	918	11.05	cps
Vanadium	51-2	757	790	837	794	5.06	cps
Yttrium	89-1	14029142	13728909	13742781	13833611	1.23	cps
Yttrium	89-2	3338927	3309844	3276637	3308469	0.94	cps
Zinc	66-2	8969	9009	8716	8898	1.79	cps
Zirconium	90-1	1413	1417	1493	1441	3.14	cps
Zirconium	91-1	300	270	343	304	12.11	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICSAB001 Instrumnet Name : P7
Client Sample ID : ICSAB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:19:16 DataFile Name : 016ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	21267339	21631091	21080270	21326233	1.31	cps
Antimony	121-1	212700	214412	214733	213948	0.51	cps
Arsenic	75-2	10691	10254	10270	10405	2.38	cps
Barium	135-1	56381	56431	54818	55877	1.64	cps
Barium	137-1	97203	98172	96330	97235	0.95	cps
Beryllium	9-1	32957	33069	33024	33017	0.17	cps
Bismuth	209-1	6728644	6862798	6117479	6569641	6.05	cps
Bismuth	209-2	4755022	4714730	4668985	4712912	0.91	cps
Bromine	81-1	12085	11601	11648	11778	2.27	cps
Bromine	81-2	183	147	137	156	15.80	cps
Cadmium	108-1	4554	4707	4621	4627	1.66	cps
Cadmium	106-1	7262	7332	7158	7251	1.20	cps
Cadmium	111-1	52748	52375	52699	52607	0.39	cps
Calcium	43-1	2729494	2739719	2763359	2744191	0.63	cps
Calcium	44-1	44046101	44244291	44686704	44325699	0.74	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	170603	170981	170740	170775	0.11	cps
Cobalt	59-2	134157	134220	135058	134478	0.37	cps
Copper	63-2	119549	118864	118616	119010	0.41	cps
Dysprosium	156-1	73	57	73	68	14.19	cps
Dysprosium	156-2	67	53	53	58	13.33	cps
Erbium	164-1	97	140	107	114	19.82	cps
Erbium	164-2	67	77	70	71	7.16	cps
Gadolinium	160-1	70	73	50	64	19.58	cps
Gadolinium	160-2	243	260	323	276	15.32	cps
Holmium	165-1	11306456	11380839	10329248	11005514	5.33	cps
Holmium	165-2	6970348	6870439	6863976	6901587	0.86	cps
Indium	115-1	8514050	8479752	7545246	8179683	6.72	cps
Indium	115-2	2642996	2652751	2564968	2620238	1.84	cps
Iron	56-2	363233981	369446021	369220101	367300035	0.96	cps
Iron	57-2	9108534	9155594	9013632	9092587	0.80	cps
Iron	54-2	19959334	19871175	19963761	19931423	0.26	cps
Krypton	83-1	203	233	273	237	14.84	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : ICSAB001 Instrumnet Name : P7
Client Sample ID : ICSAB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:19:16 DataFile Name : 016ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	199898	202101	202544	201514	0.70	cps
Lead	207-1	167046	167596	171586	168743	1.47	cps
Lead	208-1	771911	778398	789705	780004	1.15	cps
Lithium	6-1	1489322	1452028	1334066	1425139	5.69	cps
Magnesium	24-2	38604874	39085104	38248420	38646132	1.09	cps
Manganese	55-2	85048	85628	85639	85438	0.40	cps
Molybdenum	94-1	5884835	5885324	5872020	5880726	0.13	cps
Molybdenum	95-1	10259386	10381999	10349760	10330382	0.62	cps
Molybdenum	96-1	10987768	11152003	11127533	11089101	0.80	cps
Molybdenum	97-1	6471826	6461437	6514185	6482483	0.43	cps
Molybdenum	98-1	16557386	16531881	16798983	16629417	0.89	cps
Neodymium	150-1	67	40	93	67	40.00	cps
Neodymium	150-2	40	17	50	36	48.10	cps
Nickel	60-2	40174	39402	39920	39832	0.99	cps
Phosphorus	31-2	1180836	1161582	1169964	1170794	0.82	cps
Potassium	39-2	46293898	46063776	46262989	46206888	0.27	cps
Rhodium	103-1	7605986	7560035	6789126	7318382	6.27	cps
Rhodium	103-2	4376039	4395426	4365406	4378957	0.35	cps
Scandium	45-1	4771977	4808526	4272728	4617744	6.48	cps
Scandium	45-2	422440	422474	422237	422384	0.03	cps
Selenium	82-1	3605	3651	3636	3631	0.65	cps
Selenium	77-2	283	293	290	289	1.76	cps
Selenium	78-2	1167	1147	1203	1172	2.45	cps
Silicon	28-1	323695	326371	332287	327451	1.34	cps
Silver	107-1	234078	235037	237265	235460	0.69	cps
Silver	109-1	221330	222718	225101	223050	0.86	cps
Sodium	23-2	66310139	66855089	65183634	66116287	1.29	cps
Strontium	86-1	115680	114846	117313	115946	1.08	cps
Strontium	88-1	1007633	1012271	1020809	1013571	0.66	cps
Sulfur	34-1	4113338	4129119	4110692	4117716	0.24	cps
Terbium	159-1	11717730	11535086	10604461	11285759	5.29	cps
Terbium	159-2	6909560	6857285	6833589	6866811	0.57	cps
Thallium	203-1	196392	196725	195489	196202	0.33	cps
Thallium	205-1	466969	471639	472937	470515	0.67	cps
Tin	118-1	2627	2820	2950	2799	5.81	cps
Titanium	47-1	2653553	2646285	2658186	2652675	0.23	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

ICSAB001

Instrumnet Name :

P7

Client Sample ID :

ICSAB001

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:19:16

DataFile Name :

016ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	890	953	1007	950	6.15	cps
Vanadium	51-2	75032	74928	75575	75178	0.46	cps
Yttrium	89-1	14037033	14097790	12605896	13580240	6.22	cps
Yttrium	89-2	3344046	3340410	3327499	3337318	0.26	cps
Zinc	66-2	23512	23071	22694	23092	1.77	cps
Zirconium	90-1	1210	1167	1253	1210	3.58	cps
Zirconium	91-1	267	287	257	270	5.66	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV001 Instrumnet Name : P7
Client Sample ID : CCV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:22:18 DataFile Name : 017CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	12094214	11998969	12003286	12032156	0.45	cps
Antimony	121-1	5291456	5338363	5331550	5320456	0.48	cps
Arsenic	75-2	270314	266950	269204	268822	0.64	cps
Barium	135-1	7131801	7322961	7278649	7244470	1.38	cps
Barium	137-1	12358661	12533035	12593230	12494975	0.98	cps
Beryllium	9-1	791834	806544	797686	798688	0.93	cps
Bismuth	209-1	6545811	6598424	6571472	6571902	0.40	cps
Bismuth	209-2	4499679	4462176	4525549	4495801	0.71	cps
Bromine	81-1	11515	11578	11781	11625	1.20	cps
Bromine	81-2	97	97	110	101	7.61	cps
Cadmium	108-1	96777	96632	96297	96569	0.25	cps
Cadmium	106-1	139943	141329	140963	140745	0.51	cps
Cadmium	111-1	1293594	1302515	1313098	1303069	0.75	cps
Calcium	43-1	7199764	7322147	7271140	7264350	0.85	cps
Calcium	44-1	114913455	116445365	115606265	115655028	0.66	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2211006	2184685	2166552	2187414	1.02	cps
Cobalt	59-2	3446084	3380523	3415476	3414028	0.96	cps
Copper	63-2	24048344	23912122	23908795	23956420	0.33	cps
Dysprosium	156-1	317	310	323	317	2.10	cps
Dysprosium	156-2	460	423	403	429	6.70	cps
Erbium	164-1	290	353	363	336	11.85	cps
Erbium	164-2	187	183	187	186	1.04	cps
Gadolinium	160-1	250	293	243	262	10.35	cps
Gadolinium	160-2	400	393	327	373	10.86	cps
Holmium	165-1	11180526	11407147	11245416	11277696	1.03	cps
Holmium	165-2	6833533	6735530	6898995	6822686	1.21	cps
Indium	115-1	8333075	8273373	8275592	8294014	0.41	cps
Indium	115-2	2537775	2546795	2582734	2555768	0.93	cps
Iron	56-2	466130473	467261193	467169113	466853593	0.13	cps
Iron	57-2	11741397	11630302	11578566	11650088	0.71	cps
Iron	54-2	25425109	25499527	25743333	25555990	0.65	cps
Krypton	83-1	280	253	163	232	26.32	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV001 Instrumnet Name : P7
Client Sample ID : CCV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:22:18 DataFile Name : 017CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	20991983	20952356	20593567	20845969	1.05	cps
Lead	207-1	18488811	18629284	18217269	18445122	1.14	cps
Lead	208-1	83839565	83666643	82816667	83440958	0.66	cps
Lithium	6-1	1385342	1408783	1453825	1415983	2.46	cps
Magnesium	24-2	106032765	105749782	106314515	106032354	0.27	cps
Manganese	55-2	15668791	15786679	15818562	15758011	0.50	cps
Molybdenum	94-1	20515301	20650480	20896172	20687318	0.93	cps
Molybdenum	95-1	29876102	30012956	30610028	30166362	1.29	cps
Molybdenum	96-1	32480759	32876993	33549771	32969174	1.64	cps
Molybdenum	97-1	18705429	18912023	18910858	18842770	0.63	cps
Molybdenum	98-1	47824321	48606018	48892069	48440803	1.14	cps
Neodymium	150-1	500	460	403	454	10.69	cps
Neodymium	150-2	147	123	163	144	13.91	cps
Nickel	60-2	820399	814784	815678	816954	0.37	cps
Phosphorus	31-2	127419	126121	126346	126628	0.55	cps
Potassium	39-2	60981687	61041722	61370839	61131416	0.34	cps
Rhodium	103-1	7382760	7534998	7448852	7455537	1.02	cps
Rhodium	103-2	4260996	4257207	4287980	4268728	0.39	cps
Scandium	45-1	4831635	4858149	4824403	4838062	0.37	cps
Scandium	45-2	435135	429760	434168	433021	0.66	cps
Selenium	82-1	81450	83628	83206	82761	1.40	cps
Selenium	77-2	7415	7462	7319	7399	0.99	cps
Selenium	78-2	25051	25095	24834	24993	0.56	cps
Silicon	28-1	41185962	41488365	41694251	41456193	0.62	cps
Silver	107-1	6739791	6752044	6700988	6730941	0.40	cps
Silver	109-1	6303084	6314006	6404817	6340636	0.88	cps
Sodium	23-2	170083964	170978354	170185254	170415857	0.29	cps
Strontium	86-1	43289344	43674786	43551301	43505144	0.45	cps
Strontium	88-1	379103381	383077034	385538928	382573114	0.85	cps
Sulfur	34-1	582713	582063	585093	583289	0.27	cps
Terbium	159-1	11473160	11651804	11423513	11516159	1.04	cps
Terbium	159-2	6786872	6783210	6817609	6795897	0.28	cps
Thallium	203-1	5263617	5284919	5172204	5240247	1.14	cps
Thallium	205-1	12532201	12532159	12430485	12498281	0.47	cps
Tin	118-1	4526617	4535820	4450711	4504383	1.04	cps
Titanium	47-1	7369471	7475858	7449492	7431607	0.75	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV001 Instrumnet Name : P7
Client Sample ID : CCV001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:22:18 DataFile Name : 017CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	16586981	16640413	17011380	16746258	1.38	cps
Vanadium	51-2	1969906	1955441	1940994	1955447	0.74	cps
Yttrium	89-1	14168932	14407870	14271292	14282698	0.84	cps
Yttrium	89-2	3355750	3320016	3345333	3340367	0.55	cps
Zinc	66-2	3797163	3805060	3812200	3804808	0.20	cps
Zirconium	90-1	9465835	9585452	9567886	9539724	0.68	cps
Zirconium	91-1	2118238	2140818	2156921	2138659	0.91	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB001 Instrumnet Name : P7
Client Sample ID : CCB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:28:13 DataFile Name : 019CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	883	920	847	883	4.15	cps
Antimony	121-1	2650	2934	3734	3106	18.09	cps
Arsenic	75-2	10	30	10	17	69.28	cps
Barium	135-1	427	340	317	361	16.05	cps
Barium	137-1	727	670	497	631	18.99	cps
Beryllium	9-1	172	144	117	144	19.23	cps
Bismuth	209-1	7177268	7143365	7073000	7131211	0.75	cps
Bismuth	209-2	4974484	4838852	4939499	4917612	1.43	cps
Bromine	81-1	11551	11678	10874	11368	3.80	cps
Bromine	81-2	70	47	67	61	20.65	cps
Cadmium	108-1	10	10	20	13	43.30	cps
Cadmium	106-1	3054	3077	2870	3000	3.77	cps
Cadmium	111-1	2277	2277	2095	2216	4.74	cps
Calcium	43-1	617	3224	547	1462	104.33	cps
Calcium	44-1	18645	61020	16680	32115	78.01	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	647	573	680	633	8.62	cps
Cobalt	59-2	210	160	180	183	13.73	cps
Copper	63-2	2094	2190	2124	2136	2.32	cps
Dysprosium	156-1	17	30	30	26	30.11	cps
Dysprosium	156-2	17	17	17	17	0.00	cps
Erbium	164-1	67	83	90	80	15.02	cps
Erbium	164-2	27	60	37	41	41.60	cps
Gadolinium	160-1	37	53	27	39	34.63	cps
Gadolinium	160-2	257	233	240	243	4.94	cps
Holmium	165-1	11231024	11059838	11105722	11132195	0.80	cps
Holmium	165-2	6817501	6724721	6815308	6785843	0.78	cps
Indium	115-1	9025927	8735382	8734122	8831810	1.90	cps
Indium	115-2	2686369	2715191	2690074	2697211	0.58	cps
Iron	56-2	32612	31119	32077	31936	2.37	cps
Iron	57-2	1700	1790	1640	1710	4.42	cps
Iron	54-2	2834	2657	2614	2701	4.31	cps
Krypton	83-1	217	213	207	212	2.40	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB001 Instrumnet Name : P7
Client Sample ID : CCB001 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:28:13 DataFile Name : 019CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	2440	2174	2244	2286	6.05	cps
Lead	207-1	2260	1943	1777	1994	12.32	cps
Lead	208-1	10015	8701	8755	9157	8.12	cps
Lithium	6-1	1534385	1508704	1501594	1514895	1.14	cps
Magnesium	24-2	4944	4854	4751	4850	2.00	cps
Manganese	55-2	697	807	790	764	7.76	cps
Molybdenum	94-1	2017	2711	1913	2214	19.59	cps
Molybdenum	95-1	2460	2450	2070	2327	9.56	cps
Molybdenum	96-1	2640	3238	2497	2792	14.08	cps
Molybdenum	97-1	1423	1560	1260	1415	10.62	cps
Molybdenum	98-1	3594	4215	3547	3785	9.85	cps
Neodymium	150-1	7	0	20	9	114.55	cps
Neodymium	150-2	10	0	0	3	173.21	cps
Nickel	60-2	103	83	100	96	11.21	cps
Phosphorus	31-2	247	253	190	230	15.13	cps
Potassium	39-2	60998	60644	59807	60483	1.01	cps
Rhodium	103-1	8307786	8320277	8231763	8286609	0.58	cps
Rhodium	103-2	4666479	4660408	4653180	4660022	0.14	cps
Scandium	45-1	4994085	4911641	4916938	4940888	0.93	cps
Scandium	45-2	430728	423609	424161	426166	0.93	cps
Selenium	82-1	414	436	445	432	3.69	cps
Selenium	77-2	0	0	0	0	0.00	cps
Selenium	78-2	233	343	203	260	28.35	cps
Silicon	28-1	226409	225932	223457	225266	0.70	cps
Silver	107-1	837	667	673	726	13.27	cps
Silver	109-1	780	637	520	646	20.17	cps
Sodium	23-2	56302	55653	54556	55504	1.59	cps
Strontium	86-1	2450	3025	1920	2465	22.42	cps
Strontium	88-1	18312	22117	13583	18004	23.75	cps
Sulfur	34-1	183146	185098	183926	184057	0.53	cps
Terbium	159-1	11571721	11510057	11191505	11424427	1.79	cps
Terbium	159-2	6738633	6817741	6783811	6780062	0.59	cps
Thallium	203-1	773	803	793	790	1.93	cps
Thallium	205-1	1940	1813	1653	1802	7.97	cps
Tin	118-1	2290	2364	2397	2350	2.32	cps
Titanium	47-1	607	2729	560	1298	95.41	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

CCB001

Instrumnet Name :

P7

Client Sample ID :

CCB001

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:28:13

DataFile Name :

019CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	790	697	553	680	17.53	cps
Vanadium	51-2	53	60	73	62	16.37	cps
Yttrium	89-1	14578487	14420179	14288426	14429030	1.01	cps
Yttrium	89-2	3362467	3343149	3351365	3352327	0.29	cps
Zinc	66-2	747	733	783	754	3.43	cps
Zirconium	90-1	1870	2100	1967	1979	5.84	cps
Zirconium	91-1	440	483	430	451	6.29	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BL Instrumnet Name : P7
Client Sample ID : PBS454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:31:30 DataFile Name : 020CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	773	800	730	768	4.60	cps
Antimony	121-1	840	993	1223	1019	18.94	cps
Arsenic	75-2	20	13	13	16	24.76	cps
Barium	135-1	263	223	203	230	13.28	cps
Barium	137-1	403	437	410	417	4.23	cps
Beryllium	9-1	100	94	56	83	29.06	cps
Bismuth	209-1	6943071	6789957	6815326	6849451	1.20	cps
Bismuth	209-2	4843722	4770298	4851007	4821676	0.93	cps
Bromine	81-1	11171	11371	11221	11254	0.93	cps
Bromine	81-2	43	33	67	48	35.81	cps
Cadmium	108-1	10	10	17	12	31.50	cps
Cadmium	106-1	2910	2990	2690	2864	5.43	cps
Cadmium	111-1	2090	2143	1945	2059	4.98	cps
Calcium	43-1	523	463	433	473	9.68	cps
Calcium	44-1	14881	14831	14654	14789	0.81	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	620	663	607	630	4.70	cps
Cobalt	59-2	127	133	123	128	3.98	cps
Copper	63-2	1857	1857	1800	1838	1.78	cps
Dysprosium	156-1	47	30	23	33	36.07	cps
Dysprosium	156-2	17	7	7	10	57.72	cps
Erbium	164-1	63	83	100	82	22.33	cps
Erbium	164-2	50	20	20	30	57.74	cps
Gadolinium	160-1	40	43	37	40	8.32	cps
Gadolinium	160-2	170	300	213	228	29.06	cps
Holmium	165-1	10896010	11008348	10642415	10848925	1.73	cps
Holmium	165-2	6633161	6561790	6675024	6623325	0.86	cps
Indium	115-1	8632717	8575673	8496170	8568186	0.80	cps
Indium	115-2	2682494	2679778	2599886	2654053	1.77	cps
Iron	56-2	26637	26901	27375	26971	1.39	cps
Iron	57-2	1553	1640	1617	1603	2.80	cps
Iron	54-2	2254	2357	2530	2380	5.87	cps
Krypton	83-1	173	190	207	190	8.77	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BL Instrumnet Name : P7
Client Sample ID : PBS454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:31:30 DataFile Name : 020CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1663	1713	1463	1613	8.20	cps
Lead	207-1	1483	1430	1413	1442	2.54	cps
Lead	208-1	6747	6567	6134	6483	4.86	cps
Lithium	6-1	1493025	1424702	1427430	1448386	2.67	cps
Magnesium	24-2	3640	3660	4054	3785	6.16	cps
Manganese	55-2	620	697	617	644	7.02	cps
Molybdenum	94-1	1040	997	1077	1038	3.86	cps
Molybdenum	95-1	1273	1210	1127	1203	6.11	cps
Molybdenum	96-1	1273	1383	1277	1311	4.77	cps
Molybdenum	97-1	747	680	690	706	5.10	cps
Molybdenum	98-1	1910	1850	1640	1800	7.88	cps
Neodymium	150-1	10	7	13	10	33.30	cps
Neodymium	150-2	0	0	0	0	0.00	cps
Nickel	60-2	107	63	67	79	30.57	cps
Phosphorus	31-2	200	230	223	218	7.23	cps
Potassium	39-2	57297	57454	57930	57560	0.57	cps
Rhodium	103-1	8039768	7974343	7875301	7963137	1.04	cps
Rhodium	103-2	4530101	4596106	4495380	4540529	1.13	cps
Scandium	45-1	4817284	4789046	4753016	4786449	0.67	cps
Scandium	45-2	424207	423808	421709	423241	0.32	cps
Selenium	82-1	476	385	386	416	12.56	cps
Selenium	77-2	0	0	0	0	0.00	cps
Selenium	78-2	247	293	247	262	10.28	cps
Silicon	28-1	215505	216132	217902	216513	0.57	cps
Silver	107-1	407	340	347	364	10.07	cps
Silver	109-1	367	277	283	309	16.23	cps
Sodium	23-2	46266	46296	46196	46253	0.11	cps
Strontium	86-1	1487	1313	1323	1375	7.08	cps
Strontium	88-1	10677	10440	8896	10005	9.67	cps
Sulfur	34-1	183117	180718	182580	182138	0.69	cps
Terbium	159-1	11009161	11263364	11126180	11132902	1.14	cps
Terbium	159-2	6676538	6616470	6541644	6611551	1.02	cps
Thallium	203-1	617	503	467	529	14.79	cps
Thallium	205-1	1293	1350	1267	1303	3.27	cps
Tin	118-1	1607	1833	1703	1715	6.63	cps
Titanium	47-1	363	363	317	348	7.75	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BL Instrumnet Name : P7
Client Sample ID : PBS454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:31:30 DataFile Name : 020CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	400	440	360	400	10.00	cps
Vanadium	51-2	53	60	63	59	8.65	cps
Yttrium	89-1	14285915	14127427	13715492	14042945	2.10	cps
Yttrium	89-2	3354300	3361228	3313919	3343149	0.76	cps
Zinc	66-2	723	733	787	748	4.55	cps
Zirconium	90-1	1160	1157	1327	1215	8.00	cps
Zirconium	91-1	253	273	290	272	6.74	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BS Instrumnet Name : P7
Client Sample ID : LCS454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:34:44 DataFile Name : 021LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	10434	13036	9887	11119	15.13	cps
Antimony	121-1	40601	40393	40560	40518	0.27	cps
Arsenic	75-2	1097	1007	1123	1076	5.68	cps
Barium	135-1	50352	50288	50834	50491	0.59	cps
Barium	137-1	87057	87748	86910	87238	0.51	cps
Beryllium	9-1	3206	3295	3606	3369	6.24	cps
Bismuth	209-1	6684773	6791184	6825066	6767008	1.08	cps
Bismuth	209-2	4886325	4881808	4812234	4860122	0.85	cps
Bromine	81-1	11084	10907	11178	11056	1.24	cps
Bromine	81-2	60	47	53	53	12.50	cps
Cadmium	108-1	350	313	370	344	8.34	cps
Cadmium	106-1	3540	3357	3547	3482	3.10	cps
Cadmium	111-1	7004	6858	7115	6992	1.84	cps
Calcium	43-1	28791	30070	29879	29580	2.33	cps
Calcium	44-1	481107	482197	483862	482389	0.29	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	18335	18131	18842	18436	1.99	cps
Cobalt	59-2	13673	13673	13786	13711	0.48	cps
Copper	63-2	21165	21606	21025	21265	1.43	cps
Dysprosium	156-1	47	37	30	38	22.21	cps
Dysprosium	156-2	20	13	17	17	20.01	cps
Erbium	164-1	60	107	67	78	32.45	cps
Erbium	164-2	27	47	33	36	28.64	cps
Gadolinium	160-1	37	73	27	46	53.93	cps
Gadolinium	160-2	273	267	247	262	5.29	cps
Holmium	165-1	10804337	10949310	10588524	10780724	1.68	cps
Holmium	165-2	6666827	6649670	6729944	6682147	0.63	cps
Indium	115-1	8612428	8641321	8660979	8638243	0.28	cps
Indium	115-2	2687731	2656032	2692037	2678600	0.73	cps
Iron	56-2	1540596	1541079	1562668	1548114	0.81	cps
Iron	57-2	40043	41330	41193	40855	1.73	cps
Iron	54-2	88837	91986	89121	89981	1.94	cps
Krypton	83-1	187	180	193	187	3.57	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BS Instrumnet Name : P7
Client Sample ID : LCS454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:34:44 DataFile Name : 021LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	16150	15362	16046	15853	2.70	cps
Lead	207-1	13740	14194	14291	14075	2.09	cps
Lead	208-1	63064	62800	63845	63236	0.86	cps
Lithium	6-1	1431661	1408242	1470242	1436715	2.18	cps
Magnesium	24-2	438637	449604	441476	443239	1.28	cps
Manganese	55-2	6878	7482	7085	7148	4.29	cps
Molybdenum	94-1	44066	43280	44186	43844	1.12	cps
Molybdenum	95-1	53086	53093	53300	53159	0.23	cps
Molybdenum	96-1	58955	59115	59199	59090	0.21	cps
Molybdenum	97-1	33301	33291	33615	33403	0.55	cps
Molybdenum	98-1	84915	85210	85109	85078	0.18	cps
Neodymium	150-1	13	13	0	9	86.60	cps
Neodymium	150-2	3	0	3	2	86.60	cps
Nickel	60-2	3357	3527	3590	3492	3.46	cps
Phosphorus	31-2	203	283	233	240	16.84	cps
Potassium	39-2	562785	572388	569363	568179	0.86	cps
Rhodium	103-1	8164373	8031031	8080946	8092116	0.83	cps
Rhodium	103-2	4609843	4597965	4627178	4611662	0.32	cps
Scandium	45-1	4830289	4826660	4863979	4840309	0.43	cps
Scandium	45-2	423881	425557	430197	426545	0.77	cps
Selenium	82-1	2262	2194	2170	2209	2.15	cps
Selenium	77-2	193	220	137	183	23.21	cps
Selenium	78-2	830	733	727	763	7.58	cps
Silicon	28-1	319522	321152	323190	321288	0.57	cps
Silver	107-1	24551	25209	24878	24879	1.32	cps
Silver	109-1	23132	23763	23269	23388	1.42	cps
Sodium	23-2	737254	748449	746235	743979	0.80	cps
Strontium	86-1	8322	7979	8262	8188	2.24	cps
Strontium	88-1	67219	68016	68308	67848	0.83	cps
Sulfur	34-1	184456	187717	186543	186239	0.89	cps
Terbium	159-1	11168484	11096888	10939104	11068159	1.06	cps
Terbium	159-2	6682584	6738314	6794720	6738539	0.83	cps
Thallium	203-1	17929	18249	17959	18045	0.98	cps
Thallium	205-1	42005	42904	42777	42562	1.14	cps
Tin	118-1	1780	1763	1887	1810	3.70	cps
Titanium	47-1	13716	13987	14167	13957	1.62	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BS Instrumnet Name : P7
Client Sample ID : LCS454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:34:44 DataFile Name : 021LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	53534	53557	54042	53711	0.53	cps
Vanadium	51-2	38746	38061	38369	38392	0.89	cps
Yttrium	89-1	14278547	14284904	14136198	14233216	0.59	cps
Yttrium	89-2	3342921	3377090	3291243	3337085	1.30	cps
Zinc	66-2	8586	8306	8239	8377	2.20	cps
Zirconium	90-1	37113	37013	37611	37246	0.86	cps
Zirconium	91-1	8443	8626	8356	8475	1.63	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BSD Instrumnet Name : P7
Client Sample ID : LCSD454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:38:00 DataFile Name : 022LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5481	5444	5518	5481	0.67	cps
Antimony	121-1	20288	20902	21199	20796	2.23	cps
Arsenic	75-2	593	653	547	598	8.95	cps
Barium	135-1	25750	25500	25650	25633	0.49	cps
Barium	137-1	43687	44486	44065	44079	0.91	cps
Beryllium	9-1	1739	1756	1817	1771	2.31	cps
Bismuth	209-1	6431907	6802902	6824067	6686292	3.30	cps
Bismuth	209-2	4892443	4902391	4913234	4902690	0.21	cps
Bromine	81-1	11011	11294	11021	11109	1.45	cps
Bromine	81-2	73	63	100	79	24.03	cps
Cadmium	108-1	203	237	217	219	7.66	cps
Cadmium	106-1	3057	3197	3250	3168	3.15	cps
Cadmium	111-1	4498	4593	4514	4535	1.12	cps
Calcium	43-1	15034	15735	14771	15180	3.28	cps
Calcium	44-1	249824	251982	254101	251969	0.85	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	9473	9446	9763	9561	1.84	cps
Cobalt	59-2	7048	6875	7028	6984	1.36	cps
Copper	63-2	11274	11478	11521	11424	1.15	cps
Dysprosium	156-1	20	47	37	34	39.11	cps
Dysprosium	156-2	40	17	13	23	62.27	cps
Erbium	164-1	70	53	60	61	13.73	cps
Erbium	164-2	40	47	40	42	9.12	cps
Gadolinium	160-1	63	37	67	56	29.59	cps
Gadolinium	160-2	277	240	293	270	10.11	cps
Holmium	165-1	10126418	10842253	10734757	10567809	3.65	cps
Holmium	165-2	6672012	6720165	6622372	6671516	0.73	cps
Indium	115-1	8171777	8699165	8566231	8479057	3.23	cps
Indium	115-2	2684070	2689958	2674309	2682779	0.29	cps
Iron	56-2	815169	817401	817055	816542	0.15	cps
Iron	57-2	21542	20895	21669	21369	1.94	cps
Iron	54-2	46679	46722	46348	46583	0.44	cps
Krypton	83-1	203	200	197	200	1.67	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BSD Instrumnet Name : P7
Client Sample ID : LCSD454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:38:00 DataFile Name : 022LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	8613	8993	8416	8674	3.38	cps
Lead	207-1	7302	7776	7799	7626	3.68	cps
Lead	208-1	34384	34921	35015	34773	0.98	cps
Lithium	6-1	1359715	1431845	1442721	1411427	3.20	cps
Magnesium	24-2	223703	224562	225416	224560	0.38	cps
Manganese	55-2	3917	3834	4057	3936	2.87	cps
Molybdenum	94-1	22871	22798	22798	22822	0.19	cps
Molybdenum	95-1	27205	26501	28034	27247	2.82	cps
Molybdenum	96-1	30268	30315	30732	30438	0.84	cps
Molybdenum	97-1	17040	17287	17010	17112	0.89	cps
Molybdenum	98-1	44334	43545	44350	44076	1.04	cps
Neodymium	150-1	10	10	10	10	0.00	cps
Neodymium	150-2	7	3	3	4	43.40	cps
Nickel	60-2	1890	1933	1887	1903	1.37	cps
Phosphorus	31-2	193	227	207	209	8.03	cps
Potassium	39-2	311433	313060	312630	312374	0.27	cps
Rhodium	103-1	7589827	8096105	8112545	7932826	3.75	cps
Rhodium	103-2	4662865	4596930	4666233	4642009	0.84	cps
Scandium	45-1	4583713	4914471	4878470	4792218	3.79	cps
Scandium	45-2	428989	426064	424764	426606	0.51	cps
Selenium	82-1	1317	1258	1280	1285	2.30	cps
Selenium	77-2	73	43	50	56	28.35	cps
Selenium	78-2	480	510	540	510	5.88	cps
Silicon	28-1	273083	273130	273058	273090	0.01	cps
Silver	107-1	12345	12499	12832	12559	1.98	cps
Silver	109-1	11948	12195	12325	12156	1.58	cps
Sodium	23-2	393443	392529	392920	392964	0.12	cps
Strontium	86-1	4911	4864	4888	4888	0.48	cps
Strontium	88-1	38450	39262	38814	38842	1.05	cps
Sulfur	34-1	186163	190465	190073	188900	1.26	cps
Terbium	159-1	10275826	11141990	11083519	10833778	4.47	cps
Terbium	159-2	6721133	6699314	6727705	6716051	0.22	cps
Thallium	203-1	9307	8866	9213	9129	2.54	cps
Thallium	205-1	22639	22048	22479	22389	1.36	cps
Tin	118-1	1607	1620	1667	1631	1.93	cps
Titanium	47-1	7252	7115	7465	7277	2.43	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164454BSD Instrumnet Name : P7
Client Sample ID : LCSD454 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:38:00 DataFile Name : 022LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	27295	27088	27565	27316	0.88	cps
Vanadium	51-2	19496	19322	19606	19475	0.73	cps
Yttrium	89-1	13165283	14407805	14103512	13892200	4.66	cps
Yttrium	89-2	3382795	3361701	3398087	3380861	0.54	cps
Zinc	66-2	4968	4414	4724	4702	5.90	cps
Zirconium	90-1	19410	19800	19370	19526	1.22	cps
Zirconium	91-1	4141	4357	4644	4381	5.77	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-01 Instrumnet Name : P7
Client Sample ID : MYE4H0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:41:15 DataFile Name : 023SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	10737643	10874627	10646076	10752782	1.07	cps
Antimony	121-1	14955	15288	15475	15239	1.73	cps
Arsenic	75-2	46188	46212	46452	46284	0.32	cps
Barium	135-1	1124998	1133433	1133020	1130483	0.42	cps
Barium	137-1	2092075	2148179	2119754	2120003	1.32	cps
Beryllium	9-1	2139	2239	2350	2243	4.71	cps
Bismuth	209-1	6852663	6502917	6426725	6594102	3.44	cps
Bismuth	209-2	4858270	4804617	4762410	4808432	1.00	cps
Bromine	81-1	18675	19143	19503	19107	2.17	cps
Bromine	81-2	1570	1513	1370	1485	6.94	cps
Cadmium	108-1	467	537	520	508	7.20	cps
Cadmium	106-1	4041	3737	3664	3814	5.24	cps
Cadmium	111-1	6602	6231	6178	6337	3.65	cps
Calcium	43-1	535655	547108	544566	542443	1.11	cps
Calcium	44-1	8132127	8290954	8236584	8219888	0.98	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	492884	491765	489638	491429	0.34	cps
Cobalt	59-2	322755	320354	318680	320596	0.64	cps
Copper	63-2	1188203	1185784	1177441	1183809	0.48	cps
Dysprosium	156-1	79912	81572	80436	80640	1.05	cps
Dysprosium	156-2	59512	60098	59073	59561	0.86	cps
Erbium	164-1	64491	64079	64334	64301	0.32	cps
Erbium	164-2	39797	39713	39158	39556	0.88	cps
Gadolinium	160-1	74051	75301	75901	75084	1.26	cps
Gadolinium	160-2	49664	50333	48982	49660	1.36	cps
Holmium	165-1	11385319	10795244	10633513	10938025	3.62	cps
Holmium	165-2	6899224	6764239	6741211	6801558	1.26	cps
Indium	115-1	8719205	8412177	8231485	8454289	2.92	cps
Indium	115-2	2666343	2700646	2656519	2674502	0.87	cps
Iron	56-2	344399035	344198328	344310675	344302679	0.03	cps
Iron	57-2	8508821	8462666	8458894	8476793	0.33	cps
Iron	54-2	18971584	18643546	18615570	18743567	1.06	cps
Krypton	83-1	270	317	310	299	8.44	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-01 Instrumnet Name : P7
Client Sample ID : MYE4H0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:41:15 DataFile Name : 023SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	422831	429588	434597	429005	1.38	cps
Lead	207-1	367913	374382	376874	373056	1.24	cps
Lead	208-1	1684042	1701280	1714289	1699870	0.89	cps
Lithium	6-1	1435124	1394357	1392796	1407426	1.71	cps
Magnesium	24-2	9043892	9012088	8941584	8999188	0.58	cps
Manganese	55-2	7324444	7298288	7297514	7306748	0.21	cps
Molybdenum	94-1	78586	80745	81086	80139	1.69	cps
Molybdenum	95-1	10631	10794	11144	10856	2.42	cps
Molybdenum	96-1	28454	28638	29563	28885	2.06	cps
Molybdenum	97-1	6585	6752	6618	6652	1.33	cps
Molybdenum	98-1	16469	16616	17170	16752	2.21	cps
Neodymium	150-1	141556	142542	142840	142312	0.47	cps
Neodymium	150-2	74113	74030	74435	74193	0.29	cps
Nickel	60-2	155545	152120	153398	153688	1.13	cps
Phosphorus	31-2	39551	40343	40072	39989	1.01	cps
Potassium	39-2	6999449	6984400	6982578	6988809	0.13	cps
Rhodium	103-1	8077104	7745205	7714519	7845609	2.56	cps
Rhodium	103-2	4588298	4525856	4582420	4565525	0.76	cps
Scandium	45-1	5403518	5224826	5089646	5239330	3.00	cps
Scandium	45-2	472711	469315	464775	468934	0.85	cps
Selenium	82-1	694	670	654	673	3.00	cps
Selenium	77-2	277	390	277	314	20.81	cps
Selenium	78-2	487	463	490	480	3.03	cps
Silicon	28-1	31542578	32104402	32127547	31924842	1.04	cps
Silver	107-1	8216	7512	7469	7732	5.42	cps
Silver	109-1	7245	7152	7055	7151	1.33	cps
Sodium	23-2	871084	874703	869440	871742	0.31	cps
Strontium	86-1	313356	314661	317516	315178	0.67	cps
Strontium	88-1	2839303	2896137	2924654	2886698	1.51	cps
Sulfur	34-1	205587	199251	190276	198371	3.88	cps
Terbium	159-1	11427776	11197683	10787424	11137627	2.91	cps
Terbium	159-2	6868927	6849343	6818924	6845731	0.37	cps
Thallium	203-1	4184	4254	4417	4285	2.80	cps
Thallium	205-1	10207	10481	10638	10442	2.09	cps
Tin	118-1	24291	25262	25483	25012	2.54	cps
Titanium	47-1	1796842	1820446	1787222	1801504	0.95	cps

LB Number :LB133265Operator :Jaswal

Lab Sample ID :P4484-01Instrumnet Name :P7

Client Sample ID :MYE4H0Dilution Factor :1

Date & Time Acquired :2024-11-02 18:41:15DataFile Name :023SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	31016	31584	31033	31211	1.04	cps
Vanadium	51-2	695736	697835	693135	695569	0.34	cps
Yttrium	89-1	15582218	15140906	14801906	15175010	2.58	cps
Yttrium	89-2	3679280	3633129	3648061	3653490	0.64	cps
Zinc	66-2	195189	195891	195326	195469	0.19	cps
Zirconium	90-1	202413	204496	202604	203171	0.57	cps
Zirconium	91-1	44628	45939	46219	45595	1.86	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-02 Instrumnet Name : P7
Client Sample ID : MYE4H1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:44:21 DataFile Name : 024SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	9698113	9559087	9572501	9609900	0.80	cps
Antimony	121-1	39457	39614	39544	39539	0.20	cps
Arsenic	75-2	81683	81036	82082	81600	0.65	cps
Barium	135-1	1271103	1283571	1291262	1281979	0.79	cps
Barium	137-1	2367815	2362380	2387622	2372606	0.56	cps
Beryllium	9-1	2445	2622	2684	2584	4.80	cps
Bismuth	209-1	6874516	6784691	6837741	6832316	0.66	cps
Bismuth	209-2	4901892	4832005	4884347	4872748	0.75	cps
Bromine	81-1	14207	13626	13526	13787	2.67	cps
Bromine	81-2	590	673	537	600	11.48	cps
Cadmium	108-1	623	593	713	643	9.71	cps
Cadmium	106-1	4194	4474	4257	4308	3.41	cps
Cadmium	111-1	8744	8885	8462	8697	2.47	cps
Calcium	43-1	221714	223154	221350	222073	0.43	cps
Calcium	44-1	3372745	3354685	3400869	3376099	0.69	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	442386	441602	439339	441109	0.36	cps
Cobalt	59-2	361294	362318	356846	360152	0.81	cps
Copper	63-2	1377969	1376117	1367820	1373968	0.39	cps
Dysprosium	156-1	83815	85636	84613	84688	1.08	cps
Dysprosium	156-2	65274	64999	65883	65386	0.69	cps
Erbium	164-1	58298	60986	59931	59738	2.27	cps
Erbium	164-2	36758	37433	37614	37269	1.21	cps
Gadolinium	160-1	75093	76525	75653	75757	0.95	cps
Gadolinium	160-2	50360	50002	50079	50147	0.38	cps
Holmium	165-1	11134469	11018719	11001334	11051507	0.65	cps
Holmium	165-2	6884316	6964171	6791156	6879881	1.26	cps
Indium	115-1	8904284	8569860	8580195	8684779	2.19	cps
Indium	115-2	2682910	2693677	2692610	2689733	0.22	cps
Iron	56-2	434759780	433758967	436733380	435084042	0.35	cps
Iron	57-2	10606022	10827751	10605914	10679896	1.20	cps
Iron	54-2	23666515	23742686	23608857	23672686	0.28	cps
Krypton	83-1	253	297	327	292	12.62	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-02 Instrumnet Name : P7
Client Sample ID : MYE4H1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:44:21 DataFile Name : 024SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	756003	777408	776889	770100	1.59	cps
Lead	207-1	676810	676758	681925	678498	0.44	cps
Lead	208-1	3303282	3341829	3345822	3330311	0.71	cps
Lithium	6-1	1429399	1499062	1432992	1453818	2.70	cps
Magnesium	24-2	8622856	8485647	8588528	8565677	0.83	cps
Manganese	55-2	8565034	8489428	8428341	8494268	0.81	cps
Molybdenum	94-1	105346	106675	106218	106079	0.64	cps
Molybdenum	95-1	10257	10380	10307	10315	0.60	cps
Molybdenum	96-1	34441	34444	34855	34580	0.69	cps
Molybdenum	97-1	6515	6658	6331	6502	2.52	cps
Molybdenum	98-1	16806	16756	16549	16704	0.82	cps
Neodymium	150-1	161436	160135	163138	161570	0.93	cps
Neodymium	150-2	84183	84236	85152	84524	0.64	cps
Nickel	60-2	169291	168924	168528	168914	0.23	cps
Phosphorus	31-2	29462	29084	29495	29347	0.78	cps
Potassium	39-2	4280560	4279278	4250138	4269992	0.40	cps
Rhodium	103-1	8087561	8137739	8088964	8104755	0.35	cps
Rhodium	103-2	4590561	4640565	4575508	4602211	0.74	cps
Scandium	45-1	5327212	5332958	5306737	5322302	0.26	cps
Scandium	45-2	465091	468294	465564	466316	0.37	cps
Selenium	82-1	666	705	623	665	6.16	cps
Selenium	77-2	267	307	307	293	7.87	cps
Selenium	78-2	493	520	523	512	3.21	cps
Silicon	28-1	25470904	26014810	25611265	25698993	1.10	cps
Silver	107-1	8086	7999	7949	8011	0.86	cps
Silver	109-1	7185	7325	7415	7309	1.59	cps
Sodium	23-2	158566	158610	156617	157931	0.72	cps
Strontium	86-1	217378	221383	223396	220719	1.39	cps
Strontium	88-1	1936030	1985786	1949335	1957050	1.32	cps
Sulfur	34-1	160354	156367	155451	157391	1.66	cps
Terbium	159-1	11609237	11373717	11006251	11329735	2.68	cps
Terbium	159-2	6881470	6880569	6907438	6889826	0.22	cps
Thallium	203-1	4928	5028	5194	5050	2.67	cps
Thallium	205-1	11708	11839	12366	11971	2.91	cps
Tin	118-1	29684	30108	30322	30038	1.08	cps
Titanium	47-1	1367610	1399344	1399268	1388741	1.32	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-02 Instrumnet Name : P7
Client Sample ID : MYE4H1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:44:21 DataFile Name : 024SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	36099	35357	36256	35904	1.34	cps
Vanadium	51-2	802869	800996	794886	799584	0.52	cps
Yttrium	89-1	15378345	15363249	15266056	15335883	0.40	cps
Yttrium	89-2	3657761	3619826	3604519	3627369	0.76	cps
Zinc	66-2	280245	282908	279801	280984	0.60	cps
Zirconium	90-1	273134	273997	275940	274357	0.52	cps
Zirconium	91-1	60042	62278	62094	61471	2.02	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03 Instrumnet Name : P7
Client Sample ID : MYE4H2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:47:25 DataFile Name : 025SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	14841986	14543397	14631043	14672142	1.05	cps
Antimony	121-1	40357	40828	40564	40583	0.58	cps
Arsenic	75-2	236675	235278	234904	235619	0.40	cps
Barium	135-1	1799256	1816224	1827442	1814307	0.78	cps
Barium	137-1	3118579	3118689	3141315	3126194	0.42	cps
Beryllium	9-1	2761	2828	2817	2802	1.27	cps
Bismuth	209-1	6517777	6614137	6566935	6566283	0.73	cps
Bismuth	209-2	4540055	4671108	4677351	4629504	1.67	cps
Bromine	81-1	15361	15001	15145	15169	1.20	cps
Bromine	81-2	807	790	807	801	1.20	cps
Cadmium	108-1	350	390	410	383	7.97	cps
Cadmium	106-1	4011	3887	4021	3973	1.87	cps
Cadmium	111-1	3666	3548	3721	3645	2.42	cps
Calcium	43-1	2773733	2820538	2813419	2802563	0.90	cps
Calcium	44-1	44668554	45545918	45273373	45162615	0.99	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3063769	3072048	3074243	3070020	0.18	cps
Cobalt	59-2	844463	839396	836005	839955	0.51	cps
Copper	63-2	1209752	1204533	1210070	1208119	0.26	cps
Dysprosium	156-1	91103	92405	93988	92499	1.56	cps
Dysprosium	156-2	65883	66918	64320	65707	1.99	cps
Erbium	164-1	85174	85939	85583	85565	0.45	cps
Erbium	164-2	52769	52511	53080	52787	0.54	cps
Gadolinium	160-1	91956	93654	92821	92810	0.91	cps
Gadolinium	160-2	60965	59546	59422	59978	1.43	cps
Holmium	165-1	11001772	11100576	10916789	11006379	0.84	cps
Holmium	165-2	6751714	6685775	6737684	6725058	0.52	cps
Indium	115-1	8705674	8551234	8482111	8579673	1.33	cps
Indium	115-2	2613215	2625313	2587753	2608760	0.73	cps
Iron	56-2	697196416	692331936	700958256	696828870	0.62	cps
Iron	57-2	17371244	17189168	17181777	17247397	0.62	cps
Iron	54-2	38538331	38317027	37869813	38241724	0.89	cps
Krypton	83-1	287	327	340	318	8.73	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03 Instrumnet Name : P7
Client Sample ID : MYE4H2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:47:25 DataFile Name : 025SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	142522	142923	144840	143428	0.86	cps
Lead	207-1	120153	122016	123017	121729	1.19	cps
Lead	208-1	554063	558534	567351	559983	1.21	cps
Lithium	6-1	1406023	1427083	1461563	1431556	1.96	cps
Magnesium	24-2	20594542	20179520	20249346	20341136	1.09	cps
Manganese	55-2	20051846	19914411	19825170	19930476	0.57	cps
Molybdenum	94-1	103042	106717	104593	104784	1.76	cps
Molybdenum	95-1	17998	18625	18098	18240	1.85	cps
Molybdenum	96-1	45437	45113	44849	45133	0.65	cps
Molybdenum	97-1	10904	11191	11258	11118	1.69	cps
Molybdenum	98-1	28167	29035	28154	28452	1.78	cps
Neodymium	150-1	138021	138863	139496	138793	0.53	cps
Neodymium	150-2	71118	71299	71523	71313	0.28	cps
Nickel	60-2	750199	750212	749724	750045	0.04	cps
Phosphorus	31-2	24116	24146	23595	23952	1.29	cps
Potassium	39-2	2142554	2165067	2139747	2149123	0.65	cps
Rhodium	103-1	7947569	7784559	7803349	7845159	1.14	cps
Rhodium	103-2	4514539	4466518	4452202	4477753	0.73	cps
Scandium	45-1	5793331	5717165	5706445	5738980	0.83	cps
Scandium	45-2	495130	493014	493764	493969	0.22	cps
Selenium	82-1	738	693	656	696	5.88	cps
Selenium	77-2	243	287	257	262	8.46	cps
Selenium	78-2	387	520	437	448	15.04	cps
Silicon	28-1	23026775	23302776	23309841	23213131	0.70	cps
Silver	107-1	1797	1977	1870	1881	4.81	cps
Silver	109-1	1687	1623	1673	1661	2.01	cps
Sodium	23-2	191771	187679	190369	189940	1.09	cps
Strontium	86-1	475548	476750	473475	475258	0.35	cps
Strontium	88-1	4240619	4291682	4314748	4282350	0.89	cps
Sulfur	34-1	178435	182529	180313	180426	1.14	cps
Terbium	159-1	11415383	11578705	11422143	11472077	0.81	cps
Terbium	159-2	6828669	6670567	6731855	6743697	1.18	cps
Thallium	203-1	3534	3540	3771	3615	3.73	cps
Thallium	205-1	8666	8603	9187	8818	3.63	cps
Tin	118-1	19914	21059	20351	20441	2.83	cps
Titanium	47-1	2429684	2464439	2441954	2445359	0.72	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-03

Instrumnet Name :

P7

Client Sample ID :

MYE4H2

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 18:47:25

DataFile Name :

025SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	35775	36276	36884	36312	1.53	cps
Vanadium	51-2	1612706	1597222	1603457	1604461	0.49	cps
Yttrium	89-1	15724714	15769854	15575159	15689909	0.65	cps
Yttrium	89-2	3647801	3616142	3637922	3633955	0.45	cps
Zinc	66-2	176955	176027	175875	176286	0.33	cps
Zirconium	90-1	253290	257578	258270	256379	1.05	cps
Zirconium	91-1	57495	59055	58526	58359	1.36	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-04 Instrumnet Name : P7
Client Sample ID : MYE4H2D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:50:27 DataFile Name : 026SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	14321356	14188131	14180096	14229861	0.56	cps
Antimony	121-1	37064	38003	37997	37688	1.43	cps
Arsenic	75-2	224350	221988	223611	223317	0.54	cps
Barium	135-1	1769162	1726318	1764464	1753315	1.34	cps
Barium	137-1	3022971	3036716	3020502	3026730	0.29	cps
Beryllium	9-1	2728	2667	2361	2585	7.60	cps
Bismuth	209-1	6446122	6813287	6698384	6652597	2.82	cps
Bismuth	209-2	4639349	4608505	4599348	4615734	0.45	cps
Bromine	81-1	14561	14747	14971	14760	1.39	cps
Bromine	81-2	817	777	797	797	2.51	cps
Cadmium	108-1	437	367	430	411	9.40	cps
Cadmium	106-1	3817	4064	3984	3955	3.18	cps
Cadmium	111-1	3590	3768	3684	3681	2.42	cps
Calcium	43-1	2728035	2724577	2739395	2730669	0.28	cps
Calcium	44-1	44058624	44213116	43993978	44088573	0.26	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2974027	2962418	2966474	2967640	0.20	cps
Cobalt	59-2	806463	802516	803889	804289	0.25	cps
Copper	63-2	1169380	1169578	1171449	1170136	0.10	cps
Dysprosium	156-1	91083	92606	93009	92233	1.10	cps
Dysprosium	156-2	64115	64326	64350	64264	0.20	cps
Erbium	164-1	85992	85989	85081	85687	0.61	cps
Erbium	164-2	51367	51364	51621	51451	0.29	cps
Gadolinium	160-1	90889	91063	91923	91292	0.61	cps
Gadolinium	160-2	57598	59382	59101	58693	1.63	cps
Holmium	165-1	10811199	11056540	11196843	11021527	1.77	cps
Holmium	165-2	6700721	6613622	6633829	6649390	0.69	cps
Indium	115-1	8283506	8512037	8471228	8422257	1.45	cps
Indium	115-2	2568590	2550874	2579644	2566369	0.57	cps
Iron	56-2	675324883	669330043	668882737	671179221	0.54	cps
Iron	57-2	16680172	16687624	16518536	16628777	0.57	cps
Iron	54-2	36452919	36429321	36249934	36377391	0.31	cps
Krypton	83-1	307	303	353	321	8.71	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-04 Instrumnet Name : P7
Client Sample ID : MYE4H2D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:50:27 DataFile Name : 026SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	143273	142361	142387	142674	0.36	cps
Lead	207-1	122301	123085	121534	122307	0.63	cps
Lead	208-1	555548	561989	557872	558470	0.58	cps
Lithium	6-1	1381651	1444994	1371262	1399302	2.85	cps
Magnesium	24-2	19874247	19674784	19759837	19769623	0.51	cps
Manganese	55-2	18937626	18894956	18903685	18912089	0.12	cps
Molybdenum	94-1	101169	102559	102579	102102	0.79	cps
Molybdenum	95-1	17234	17804	17541	17526	1.63	cps
Molybdenum	96-1	43705	43812	43665	43727	0.17	cps
Molybdenum	97-1	10774	11121	10871	10922	1.64	cps
Molybdenum	98-1	26905	26998	27028	26977	0.24	cps
Neodymium	150-1	135759	139394	137066	137406	1.34	cps
Neodymium	150-2	69751	70254	69564	69856	0.51	cps
Nickel	60-2	731963	722348	720819	725043	0.83	cps
Phosphorus	31-2	22864	22764	23164	22931	0.91	cps
Potassium	39-2	2120905	2134982	2136953	2130947	0.41	cps
Rhodium	103-1	7559351	7836755	7820894	7739000	2.01	cps
Rhodium	103-2	4380004	4332492	4382846	4365114	0.65	cps
Scandium	45-1	5431644	5538963	5572050	5514219	1.33	cps
Scandium	45-2	486458	481587	483759	483934	0.50	cps
Selenium	82-1	701	646	578	641	9.62	cps
Selenium	77-2	303	307	320	310	2.85	cps
Selenium	78-2	437	463	480	460	4.75	cps
Silicon	28-1	23040735	23214012	23082064	23112270	0.39	cps
Silver	107-1	1603	1793	1807	1735	6.56	cps
Silver	109-1	1380	1397	1450	1409	2.60	cps
Sodium	23-2	183955	181716	181880	182517	0.68	cps
Strontium	86-1	459820	459845	459895	459853	0.01	cps
Strontium	88-1	4139315	4124352	4158824	4140830	0.42	cps
Sulfur	34-1	170094	167699	168252	168682	0.74	cps
Terbium	159-1	11028352	11547874	11678824	11418350	3.01	cps
Terbium	159-2	6625371	6661385	6776075	6687610	1.18	cps
Thallium	203-1	3694	3581	3814	3696	3.16	cps
Thallium	205-1	8503	9103	8523	8710	3.92	cps
Tin	118-1	21169	21075	21239	21161	0.39	cps
Titanium	47-1	2340171	2324062	2331854	2332029	0.35	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-04 Instrumnet Name : P7
Client Sample ID : MYE4H2D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 18:50:27 DataFile Name : 026SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	36934	36684	36961	36860	0.42	cps
Vanadium	51-2	1533294	1540287	1544732	1539438	0.37	cps
Yttrium	89-1	14882421	15479857	15387246	15249842	2.11	cps
Yttrium	89-2	3614871	3588315	3559134	3587440	0.78	cps
Zinc	66-2	174834	173704	173489	174009	0.42	cps
Zirconium	90-1	248888	251775	254828	251830	1.18	cps
Zirconium	91-1	56756	57285	56973	57005	0.47	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03LX5 Instrumnet Name : P7
Client Sample ID : MYE4H2L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 18:53:27 DataFile Name : 027SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	2905496	2891047	2839843	2878795	1.20	cps
Antimony	121-1	8469	8172	8409	8350	1.88	cps
Arsenic	75-2	46904	47225	46710	46946	0.55	cps
Barium	135-1	331422	332441	332192	332019	0.16	cps
Barium	137-1	570582	575255	577290	574376	0.60	cps
Beryllium	9-1	650	678	644	657	2.72	cps
Bismuth	209-1	6867835	6803214	6772308	6814452	0.72	cps
Bismuth	209-2	4676893	4726530	4740724	4714716	0.71	cps
Bromine	81-1	11414	11141	11258	11271	1.22	cps
Bromine	81-2	193	257	190	213	17.61	cps
Cadmium	108-1	107	77	73	86	21.46	cps
Cadmium	106-1	3064	3017	3030	3037	0.79	cps
Cadmium	111-1	2327	2355	2413	2365	1.87	cps
Calcium	43-1	565408	563319	568535	565754	0.46	cps
Calcium	44-1	8656777	8672068	8661829	8663558	0.09	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	606133	593551	599382	599688	1.05	cps
Cobalt	59-2	168184	168231	167837	168084	0.13	cps
Copper	63-2	246407	245916	245368	245897	0.21	cps
Dysprosium	156-1	18052	18272	18753	18359	1.95	cps
Dysprosium	156-2	13246	12773	13220	13080	2.04	cps
Erbium	164-1	17234	17041	17758	17344	2.14	cps
Erbium	164-2	10851	10644	10534	10676	1.51	cps
Gadolinium	160-1	18689	18412	18542	18548	0.75	cps
Gadolinium	160-2	12209	12209	11902	12106	1.46	cps
Holmium	165-1	10896811	10928017	10755783	10860204	0.84	cps
Holmium	165-2	6530679	6612402	6560951	6568010	0.63	cps
Indium	115-1	8317591	8415800	8472636	8402009	0.93	cps
Indium	115-2	2584544	2588201	2598551	2590432	0.28	cps
Iron	56-2	138481411	137736125	137335065	137850867	0.42	cps
Iron	57-2	3406908	3464410	3394395	3421905	1.09	cps
Iron	54-2	7562352	7494152	7475333	7510612	0.61	cps
Krypton	83-1	190	210	260	220	16.39	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03LX5 Instrumnet Name : P7
Client Sample ID : MYE4H2L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 18:53:27 DataFile Name : 027SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	30184	30732	30344	30420	0.93	cps
Lead	207-1	25734	25908	26179	25940	0.86	cps
Lead	208-1	119161	119354	120050	119522	0.39	cps
Lithium	6-1	1438230	1373145	1378221	1396532	2.59	cps
Magnesium	24-2	4023932	3955262	3989070	3989421	0.86	cps
Manganese	55-2	3950582	3945056	3919287	3938308	0.42	cps
Molybdenum	94-1	20765	20388	20992	20715	1.47	cps
Molybdenum	95-1	4601	4461	4464	4509	1.77	cps
Molybdenum	96-1	9506	9203	9573	9428	2.09	cps
Molybdenum	97-1	2754	2807	2820	2794	1.26	cps
Molybdenum	98-1	6725	6648	6755	6709	0.82	cps
Neodymium	150-1	28272	28248	28369	28296	0.23	cps
Neodymium	150-2	13773	14584	14431	14263	3.02	cps
Nickel	60-2	151775	149480	151008	150755	0.78	cps
Phosphorus	31-2	4877	4734	4797	4803	1.50	cps
Potassium	39-2	489292	484813	483750	485952	0.61	cps
Rhodium	103-1	7769400	7704711	7635672	7703261	0.87	cps
Rhodium	103-2	4439464	4520417	4483063	4480981	0.90	cps
Scandium	45-1	4837226	4820410	4853412	4837016	0.34	cps
Scandium	45-2	426952	421729	422851	423844	0.65	cps
Selenium	82-1	474	438	384	432	10.45	cps
Selenium	77-2	67	67	67	67	0.00	cps
Selenium	78-2	257	363	250	290	21.93	cps
Silicon	28-1	4690198	4615045	4674012	4659752	0.85	cps
Silver	107-1	533	577	577	562	4.45	cps
Silver	109-1	477	440	530	482	9.39	cps
Sodium	23-2	67135	67286	66643	67022	0.50	cps
Strontium	86-1	93344	95138	93223	93901	1.14	cps
Strontium	88-1	811863	815850	824702	817472	0.80	cps
Sulfur	34-1	153977	150173	152364	152171	1.25	cps
Terbium	159-1	11115702	11333384	11073379	11174155	1.25	cps
Terbium	159-2	6612058	6503301	6487153	6534171	1.04	cps
Thallium	203-1	1050	1050	1097	1066	2.53	cps
Thallium	205-1	2420	2520	2650	2530	4.56	cps
Tin	118-1	5758	5618	5598	5658	1.54	cps
Titanium	47-1	472077	470157	473576	471937	0.36	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03LX5 Instrumnet Name : P7
Client Sample ID : MYE4H2L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 18:53:27 DataFile Name : 027SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	7572	7779	8126	7826	3.57	cps
Vanadium	51-2	318361	318111	313434	316635	0.88	cps
Yttrium	89-1	14029077	14062040	13985796	14025638	0.27	cps
Yttrium	89-2	3325706	3348626	3324059	3332797	0.41	cps
Zinc	66-2	37617	36862	35993	36824	2.21	cps
Zirconium	90-1	50142	49520	49680	49781	0.65	cps
Zirconium	91-1	11208	11258	11141	11202	0.52	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:02:43 DataFile Name : 030SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	2793120	2794032	2841831	2809661	0.99	cps
Antimony	121-1	8069	8082	8152	8101	0.55	cps
Arsenic	75-2	45887	45820	46345	46018	0.62	cps
Barium	135-1	320557	324722	325424	323567	0.81	cps
Barium	137-1	554997	563693	562993	560561	0.86	cps
Beryllium	9-1	478	650	617	582	15.71	cps
Bismuth	209-1	6990210	6843506	6934954	6922890	1.07	cps
Bismuth	209-2	4696186	4219695	4981935	4632605	8.31	cps
Bromine	81-1	11294	11124	11391	11270	1.20	cps
Bromine	81-2	210	183	170	188	10.85	cps
Cadmium	108-1	103	77	63	81	25.12	cps
Cadmium	106-1	2970	2980	3010	2987	0.70	cps
Cadmium	111-1	2254	2329	2290	2291	1.65	cps
Calcium	43-1	543905	554431	549605	549313	0.96	cps
Calcium	44-1	8272934	8418104	8376014	8355684	0.89	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	581846	588009	584812	584889	0.53	cps
Cobalt	59-2	162381	160528	161033	161314	0.59	cps
Copper	63-2	237506	234120	236284	235970	0.73	cps
Dysprosium	156-1	17681	17798	17798	17759	0.38	cps
Dysprosium	156-2	12866	12719	12596	12727	1.06	cps
Erbium	164-1	16707	17421	17101	17076	2.09	cps
Erbium	164-2	10357	10374	10431	10387	0.37	cps
Gadolinium	160-1	17641	18132	18279	18017	1.85	cps
Gadolinium	160-2	12032	12099	11875	12002	0.96	cps
Holmium	165-1	11150273	10805767	11026954	10994331	1.59	cps
Holmium	165-2	6608379	5828398	6845847	6427541	8.28	cps
Indium	115-1	8575214	8420367	8407892	8467825	1.10	cps
Indium	115-2	2630492	2278827	2714698	2541339	9.10	cps
Iron	56-2	132886395	133121045	133256698	133088046	0.14	cps
Iron	57-2	3344775	3328797	3292133	3321901	0.81	cps
Iron	54-2	7329212	7355470	7261226	7315303	0.66	cps
Krypton	83-1	177	223	203	201	11.64	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:02:43 DataFile Name : 030SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	29890	30337	29091	29773	2.12	cps
Lead	207-1	25113	26576	25781	25823	2.84	cps
Lead	208-1	115889	118264	116849	117001	1.02	cps
Lithium	6-1	1398783	1388410	1391640	1392945	0.38	cps
Magnesium	24-2	3809295	3761512	3855257	3808688	1.23	cps
Manganese	55-2	3796462	3798458	3773881	3789600	0.36	cps
Molybdenum	94-1	20331	20608	20478	20472	0.68	cps
Molybdenum	95-1	4234	4237	4344	4272	1.47	cps
Molybdenum	96-1	9293	9346	9293	9311	0.33	cps
Molybdenum	97-1	2397	2404	2490	2430	2.14	cps
Molybdenum	98-1	6582	6411	6522	6505	1.33	cps
Neodymium	150-1	27003	27717	27440	27387	1.32	cps
Neodymium	150-2	13597	13867	13950	13805	1.34	cps
Nickel	60-2	147884	146737	148359	147660	0.56	cps
Phosphorus	31-2	4554	4507	4551	4537	0.57	cps
Potassium	39-2	473792	476467	472547	474269	0.42	cps
Rhodium	103-1	7931516	7714729	7717502	7787915	1.60	cps
Rhodium	103-2	4507451	3961977	4702885	4390771	8.75	cps
Scandium	45-1	4853945	4796614	4717266	4789275	1.43	cps
Scandium	45-2	423903	370686	435939	410176	8.47	cps
Selenium	82-1	497	498	456	484	4.97	cps
Selenium	77-2	37	63	60	53	27.24	cps
Selenium	78-2	303	283	303	297	3.89	cps
Silicon	28-1	4429517	4492671	4478205	4466798	0.74	cps
Silver	107-1	643	563	643	617	7.49	cps
Silver	109-1	537	487	523	516	5.02	cps
Sodium	23-2	61418	62921	62543	62294	1.25	cps
Strontium	86-1	88668	91190	90721	90193	1.49	cps
Strontium	88-1	783322	794734	793944	790667	0.81	cps
Sulfur	34-1	152974	153062	153788	153275	0.29	cps
Terbium	159-1	11328953	11251478	11040801	11207077	1.33	cps
Terbium	159-2	6755458	5883510	6879273	6506080	8.34	cps
Thallium	203-1	940	937	883	920	3.46	cps
Thallium	205-1	2087	2004	2090	2060	2.38	cps
Tin	118-1	5144	5391	5351	5295	2.50	cps
Titanium	47-1	451706	461541	460072	457773	1.16	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:02:43 DataFile Name : 030SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	7619	7799	7449	7622	2.30	cps
Vanadium	51-2	307367	305459	305212	306013	0.39	cps
Yttrium	89-1	14229201	13924549	14043048	14065599	1.09	cps
Yttrium	89-2	3336791	2929437	3508614	3258281	9.13	cps
Zinc	66-2	36221	37006	36555	36594	1.08	cps
Zirconium	90-1	47393	48844	48787	48341	1.70	cps
Zirconium	91-1	10844	10777	11141	10921	1.77	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-04DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2D Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:06:09 DataFile Name : 031SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	3478296	3489612	3484655	3484188	0.16	cps
Antimony	121-1	9126	9643	9446	9405	2.77	cps
Arsenic	75-2	56581	56899	55152	56210	1.66	cps
Barium	135-1	395506	397540	394901	395982	0.35	cps
Barium	137-1	685765	689998	689854	688539	0.35	cps
Beryllium	9-1	744	817	678	746	9.31	cps
Bismuth	209-1	6951770	6898866	6832496	6894377	0.87	cps
Bismuth	209-2	4755355	4731400	4752004	4746253	0.27	cps
Bromine	81-1	11831	11832	11848	11837	0.08	cps
Bromine	81-2	230	197	267	231	15.15	cps
Cadmium	108-1	83	93	107	94	12.40	cps
Cadmium	106-1	2927	2904	3164	2998	4.80	cps
Cadmium	111-1	2314	2327	2467	2369	3.57	cps
Calcium	43-1	690156	700969	702395	697840	0.96	cps
Calcium	44-1	10633363	10695999	10715788	10681717	0.40	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	736315	726582	724945	729281	0.84	cps
Cobalt	59-2	201217	201751	200413	201127	0.33	cps
Copper	63-2	298282	297261	294794	296779	0.60	cps
Dysprosium	156-1	22268	23373	22622	22754	2.48	cps
Dysprosium	156-2	16106	16243	16036	16128	0.65	cps
Erbium	164-1	21146	21266	21063	21158	0.48	cps
Erbium	164-2	12746	12839	12479	12688	1.47	cps
Gadolinium	160-1	22087	22251	22161	22166	0.37	cps
Gadolinium	160-2	15102	14888	15072	15020	0.77	cps
Holmium	165-1	10933164	10979068	10781804	10898012	0.95	cps
Holmium	165-2	6730516	6657638	6718564	6702239	0.58	cps
Indium	115-1	8471006	8525616	8403704	8466775	0.72	cps
Indium	115-2	2608320	2609522	2604054	2607299	0.11	cps
Iron	56-2	164562868	166813511	164508174	165294851	0.80	cps
Iron	57-2	4148353	4151159	4141100	4146871	0.13	cps
Iron	54-2	9032338	9089189	9081917	9067815	0.34	cps
Krypton	83-1	247	207	197	217	12.21	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-04DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2D Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:06:09 DataFile Name : 031SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	37498	37980	37592	37690	0.68	cps
Lead	207-1	31597	32265	31526	31796	1.28	cps
Lead	208-1	144838	146902	146199	145980	0.72	cps
Lithium	6-1	1396002	1432375	1419735	1416037	1.30	cps
Magnesium	24-2	4878756	4820497	4787821	4829024	0.95	cps
Manganese	55-2	4611743	4724741	4715455	4683980	1.34	cps
Molybdenum	94-1	25532	25522	25399	25484	0.29	cps
Molybdenum	95-1	5111	4838	5091	5013	3.04	cps
Molybdenum	96-1	11415	11595	11655	11555	1.08	cps
Molybdenum	97-1	3120	3244	2914	3093	5.39	cps
Molybdenum	98-1	7679	8146	7769	7864	3.15	cps
Neodymium	150-1	33770	34014	34302	34029	0.78	cps
Neodymium	150-2	17838	17434	17715	17662	1.17	cps
Nickel	60-2	184416	182499	182992	183302	0.54	cps
Phosphorus	31-2	6148	5948	5734	5943	3.48	cps
Potassium	39-2	592025	589088	588994	590036	0.29	cps
Rhodium	103-1	7849913	7785366	7728520	7787933	0.78	cps
Rhodium	103-2	4604317	4431639	4482712	4506223	1.97	cps
Scandium	45-1	4856145	4849979	4753395	4819840	1.20	cps
Scandium	45-2	432539	425485	427551	428525	0.85	cps
Selenium	82-1	478	513	548	513	6.80	cps
Selenium	77-2	107	70	80	86	22.16	cps
Selenium	78-2	340	223	313	292	20.92	cps
Silicon	28-1	5763703	5894504	5818591	5825599	1.13	cps
Silver	107-1	543	490	537	523	5.55	cps
Silver	109-1	480	480	540	500	6.93	cps
Sodium	23-2	68766	68173	67711	68217	0.78	cps
Strontium	86-1	111149	112519	113513	112393	1.06	cps
Strontium	88-1	982295	992334	983557	986062	0.55	cps
Sulfur	34-1	158634	161699	160122	160152	0.96	cps
Terbium	159-1	11294121	11296449	11091179	11227250	1.05	cps
Terbium	159-2	6709811	6694676	6628638	6677708	0.65	cps
Thallium	203-1	1017	1057	1130	1068	5.38	cps
Thallium	205-1	2697	2404	2337	2479	7.73	cps
Tin	118-1	6311	6321	6221	6285	0.88	cps
Titanium	47-1	565447	571440	570719	569202	0.57	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-04DLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4H2D

Dilution Factor :

4

Date & Time Acquired :

2024-11-02 19:06:09

DataFile Name :

031SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	9263	9497	9443	9401	1.30	cps
Vanadium	51-2	383201	380723	379937	381287	0.45	cps
Yttrium	89-1	14154929	14140339	13880191	14058486	1.10	cps
Yttrium	89-2	3324842	3345024	3286222	3318696	0.90	cps
Zinc	66-2	45121	44890	44529	44847	0.67	cps
Zirconium	90-1	59881	61598	61260	60913	1.49	cps
Zirconium	91-1	13980	14197	14020	14066	0.82	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4H2L Dilution Factor : 20
Date & Time Acquired : 2024-11-02 19:09:20 DataFile Name : 032SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	720099	718058	713550	717235	0.47	cps
Antimony	121-1	2100	2227	2184	2170	2.97	cps
Arsenic	75-2	11678	11271	11932	11627	2.87	cps
Barium	135-1	80620	80590	81297	80835	0.49	cps
Barium	137-1	140121	141074	141640	140945	0.54	cps
Beryllium	9-1	233	200	189	207	11.15	cps
Bismuth	209-1	6892268	6989002	6892539	6924603	0.81	cps
Bismuth	209-2	4747862	4792608	4758206	4766225	0.49	cps
Bromine	81-1	11451	11324	11415	11397	0.57	cps
Bromine	81-2	80	100	107	96	14.52	cps
Cadmium	108-1	43	17	20	27	54.47	cps
Cadmium	106-1	3024	2897	2887	2936	2.60	cps
Cadmium	111-1	2253	2120	2117	2163	3.60	cps
Calcium	43-1	137243	139534	138448	138408	0.83	cps
Calcium	44-1	2108599	2121080	2114407	2114695	0.30	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	147392	146241	147561	147064	0.49	cps
Cobalt	59-2	41895	40956	41404	41418	1.13	cps
Copper	63-2	62004	61060	61402	61489	0.78	cps
Dysprosium	156-1	4594	4467	4367	4476	2.54	cps
Dysprosium	156-2	3107	3230	3314	3217	3.23	cps
Erbium	164-1	4151	4064	4197	4137	1.64	cps
Erbium	164-2	2540	2487	2510	2512	1.06	cps
Gadolinium	160-1	4441	4644	4417	4501	2.77	cps
Gadolinium	160-2	3057	3180	3217	3151	2.66	cps
Holmium	165-1	10993632	10786989	10803402	10861341	1.06	cps
Holmium	165-2	6578031	6540515	6543985	6554177	0.32	cps
Indium	115-1	8492131	8380027	8483319	8451825	0.74	cps
Indium	115-2	2610258	2622517	2575456	2602744	0.94	cps
Iron	56-2	33640909	33628106	33193752	33487589	0.76	cps
Iron	57-2	857936	857057	859101	858031	0.12	cps
Iron	54-2	1857872	1873146	1862320	1864446	0.42	cps
Krypton	83-1	223	210	227	220	4.01	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4H2L Dilution Factor : 20
Date & Time Acquired : 2024-11-02 19:09:20 DataFile Name : 032SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	8569	8556	8239	8455	2.21	cps
Lead	207-1	7252	7165	7475	7298	2.19	cps
Lead	208-1	33343	33493	33083	33306	0.62	cps
Lithium	6-1	1422454	1453111	1402750	1426105	1.78	cps
Magnesium	24-2	999123	999998	990151	996424	0.55	cps
Manganese	55-2	954224	961537	953248	956336	0.47	cps
Molybdenum	94-1	6075	5801	5865	5913	2.42	cps
Molybdenum	95-1	1770	1863	1647	1760	6.18	cps
Molybdenum	96-1	3427	3214	3120	3254	4.83	cps
Molybdenum	97-1	1147	1163	1093	1134	3.22	cps
Molybdenum	98-1	2890	2900	2957	2916	1.23	cps
Neodymium	150-1	6858	6962	7122	6981	1.90	cps
Neodymium	150-2	3584	3387	3490	3487	2.82	cps
Nickel	60-2	36164	36889	37199	36751	1.45	cps
Phosphorus	31-2	1383	1323	1347	1351	2.24	cps
Potassium	39-2	158316	156982	154973	156757	1.07	cps
Rhodium	103-1	7924024	7821140	7807437	7850867	0.81	cps
Rhodium	103-2	4525189	4561191	4499994	4528791	0.68	cps
Scandium	45-1	4717617	4693415	4673501	4694844	0.47	cps
Scandium	45-2	413221	413167	412787	413059	0.06	cps
Selenium	82-1	363	391	377	377	3.73	cps
Selenium	77-2	7	17	10	11	45.82	cps
Selenium	78-2	260	303	247	270	10.97	cps
Silicon	28-1	1308156	1327629	1320704	1318830	0.75	cps
Silver	107-1	283	367	270	307	17.08	cps
Silver	109-1	227	223	300	250	17.33	cps
Sodium	23-2	39297	39080	38438	38938	1.15	cps
Strontium	86-1	24013	24243	23632	23963	1.29	cps
Strontium	88-1	208151	207602	208773	208175	0.28	cps
Sulfur	34-1	158341	160466	159793	159533	0.68	cps
Terbium	159-1	11294891	11236814	11134930	11222212	0.72	cps
Terbium	159-2	6639190	6643659	6534715	6605855	0.93	cps
Thallium	203-1	460	460	457	459	0.42	cps
Thallium	205-1	1110	1033	1047	1063	3.85	cps
Tin	118-1	3011	2540	2524	2692	10.28	cps
Titanium	47-1	115153	114031	116440	115208	1.05	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4H2L Dilution Factor : 20
Date & Time Acquired : 2024-11-02 19:09:20 DataFile Name : 032SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	2264	2250	2430	2315	4.33	cps
Vanadium	51-2	76747	76597	75524	76289	0.87	cps
Yttrium	89-1	13868731	13918599	13858346	13881892	0.23	cps
Yttrium	89-2	3228138	3274844	3293901	3265627	1.04	cps
Zinc	66-2	9393	9690	9813	9632	2.24	cps
Zirconium	90-1	12982	13420	13319	13240	1.73	cps
Zirconium	91-1	2840	3104	2897	2947	4.70	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-05 Instrumnet Name : P7
Client Sample ID : MYE4H2S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:15:50 DataFile Name : 034SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	16845602	16416281	16800744	16687542	1.41	cps
Antimony	121-1	756629	767362	764037	762676	0.72	cps
Arsenic	75-2	277089	278399	277082	277523	0.27	cps
Barium	135-1	6066855	6071996	6083566	6074139	0.14	cps
Barium	137-1	10441521	10564913	10527787	10511407	0.60	cps
Beryllium	9-1	61153	61298	60713	61055	0.50	cps
Bismuth	209-1	6640068	6525020	6508994	6558028	1.09	cps
Bismuth	209-2	4469752	4433122	4366545	4423140	1.18	cps
Bromine	81-1	14377	15078	14130	14528	3.38	cps
Bromine	81-2	933	900	893	909	2.36	cps
Cadmium	108-1	6395	7018	6715	6709	4.65	cps
Cadmium	106-1	12182	12065	11711	11986	2.04	cps
Cadmium	111-1	94531	93799	94199	94176	0.39	cps
Calcium	43-1	3138111	3087956	3088104	3104724	0.93	cps
Calcium	44-1	50395958	50027644	49519179	49980927	0.88	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3967751	4001459	4069223	4012811	1.29	cps
Cobalt	59-2	3332323	3306669	3310240	3316411	0.42	cps
Copper	63-2	2239337	2225550	2236492	2233793	0.33	cps
Dysprosium	156-1	106816	108351	110091	108419	1.51	cps
Dysprosium	156-2	76859	75686	74690	75745	1.43	cps
Erbium	164-1	101223	102861	101848	101977	0.81	cps
Erbium	164-2	60108	60815	60661	60528	0.61	cps
Gadolinium	160-1	106891	107592	111019	108500	2.04	cps
Gadolinium	160-2	69260	69052	69313	69209	0.20	cps
Holmium	165-1	11159423	10834799	10782681	10925634	1.87	cps
Holmium	165-2	6459726	6450157	6332995	6414293	1.10	cps
Indium	115-1	8363806	8026857	8104953	8165205	2.16	cps
Indium	115-2	2481909	2409997	2414489	2435465	1.65	cps
Iron	56-2	787035615	770987669	775013322	777678868	1.07	cps
Iron	57-2	19347168	19142014	19399933	19296372	0.71	cps
Iron	54-2	42928058	42170360	42521307	42539908	0.89	cps
Krypton	83-1	340	330	333	334	1.52	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-05 Instrumnet Name : P7
Client Sample ID : MYE4H2S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:15:50 DataFile Name : 034SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	275005	275196	279242	276481	0.87	cps
Lead	207-1	234066	237058	238947	236690	1.04	cps
Lead	208-1	1074223	1086573	1099824	1086873	1.18	cps
Lithium	6-1	1460892	1421196	1372496	1418195	3.12	cps
Magnesium	24-2	22586246	22504514	22637988	22576249	0.30	cps
Manganese	55-2	23516357	22858336	23445875	23273523	1.55	cps
Molybdenum	94-1	119254	119818	118192	119088	0.69	cps
Molybdenum	95-1	19556	19837	19333	19575	1.29	cps
Molybdenum	96-1	49841	50360	49149	49783	1.22	cps
Molybdenum	97-1	12192	11822	11578	11864	2.60	cps
Molybdenum	98-1	30432	30375	30108	30305	0.57	cps
Neodymium	150-1	160212	162916	166132	163087	1.82	cps
Neodymium	150-2	82811	81501	82228	82180	0.80	cps
Nickel	60-2	1418598	1399307	1408772	1408892	0.68	cps
Phosphorus	31-2	26269	25965	25815	26017	0.89	cps
Potassium	39-2	2386256	2357184	2334266	2359235	1.10	cps
Rhodium	103-1	7665410	7300667	7350671	7438916	2.66	cps
Rhodium	103-2	4246364	4156235	4139848	4180816	1.37	cps
Scandium	45-1	5603848	5341100	5316309	5420419	2.94	cps
Scandium	45-2	467448	468351	466352	467383	0.21	cps
Selenium	82-1	13828	13735	13702	13755	0.48	cps
Selenium	77-2	1500	1447	1367	1438	4.67	cps
Selenium	78-2	4271	4207	4214	4231	0.82	cps
Silicon	28-1	31607690	32097280	31500505	31735158	1.00	cps
Silver	107-1	385550	387825	385839	386404	0.32	cps
Silver	109-1	364768	365218	367319	365768	0.37	cps
Sodium	23-2	192646	191137	194144	192642	0.78	cps
Strontium	86-1	537352	535285	535116	535918	0.23	cps
Strontium	88-1	4809577	4832961	4764804	4802448	0.72	cps
Sulfur	34-1	166415	159018	155445	160293	3.49	cps
Terbium	159-1	11491111	10970012	11063145	11174756	2.49	cps
Terbium	159-2	6430315	6482402	6366041	6426253	0.91	cps
Thallium	203-1	326418	332157	336409	331661	1.51	cps
Thallium	205-1	793776	789297	793800	792291	0.33	cps
Tin	118-1	22524	22655	22324	22501	0.74	cps
Titanium	47-1	2776375	2761811	2726447	2754878	0.93	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-05 Instrumnet Name : P7
Client Sample ID : MYE4H2S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:15:50 DataFile Name : 034SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	43377	44618	43952	43982	1.41	cps
Vanadium	51-2	3102283	3126209	3167352	3131948	1.05	cps
Yttrium	89-1	15535659	14802167	14704518	15014115	3.03	cps
Yttrium	89-2	3544038	3450237	3374840	3456372	2.45	cps
Zinc	66-2	454487	447745	445545	449259	1.04	cps
Zirconium	90-1	290859	294693	292689	292747	0.66	cps
Zirconium	91-1	65194	65867	67233	66098	1.57	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03MS Instrumnet Name : P7
Client Sample ID : MYE4H2SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:18:48 DataFile Name : 035SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	17278518	17368792	17302690	17316667	0.27	cps
Antimony	121-1	3701200	3750875	3750984	3734353	0.77	cps
Arsenic	75-2	329181	329603	327766	328850	0.29	cps
Barium	135-1	21605136	21741002	21842690	21729609	0.55	cps
Barium	137-1	36756660	37529940	37285918	37190839	1.06	cps
Beryllium	9-1	273256	278621	276905	276261	0.99	cps
Bismuth	209-1	6598632	6453385	6504416	6518811	1.13	cps
Bismuth	209-2	4454642	4473227	4481469	4469779	0.31	cps
Bromine	81-1	13403	14010	14020	13811	2.56	cps
Bromine	81-2	920	973	973	956	3.22	cps
Cadmium	108-1	28792	29186	28882	28953	0.71	cps
Cadmium	106-1	42903	43669	43572	43381	0.96	cps
Cadmium	111-1	436543	442989	443482	441005	0.88	cps
Calcium	43-1	2886646	2950701	2952079	2929809	1.28	cps
Calcium	44-1	46766181	47360844	47225884	47117637	0.66	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	6294048	6339986	6301661	6311899	0.39	cps
Cobalt	59-2	12462310	12470456	12508403	12480390	0.20	cps
Copper	63-2	5232270	5321690	5290258	5281406	0.86	cps
Dysprosium	156-1	105042	104002	104720	104588	0.51	cps
Dysprosium	156-2	74262	74962	74550	74591	0.47	cps
Erbium	164-1	97611	98706	99193	98503	0.82	cps
Erbium	164-2	60219	59101	59646	59656	0.94	cps
Gadolinium	160-1	103119	105644	106796	105187	1.79	cps
Gadolinium	160-2	67917	67542	67980	67813	0.35	cps
Holmium	165-1	10796673	10911413	10703036	10803707	0.97	cps
Holmium	165-2	6524960	6471774	6569929	6522221	0.75	cps
Indium	115-1	8147840	8118223	7860594	8042219	1.96	cps
Indium	115-2	2460003	2473725	2437493	2457074	0.74	cps
Iron	56-2	768797775	767561989	766750335	767703366	0.13	cps
Iron	57-2	19011998	19029969	19004468	19015478	0.07	cps
Iron	54-2	41974768	42168261	42028040	42057023	0.24	cps
Krypton	83-1	313	373	357	348	8.91	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03MS Instrumnet Name : P7
Client Sample ID : MYE4H2SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:18:48 DataFile Name : 035SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	675034	691672	691460	686055	1.39	cps
Lead	207-1	586438	593169	596433	592013	0.86	cps
Lead	208-1	2680845	2722982	2734353	2712727	1.04	cps
Lithium	6-1	1393610	1366490	1354931	1371677	1.45	cps
Magnesium	24-2	21854736	22014221	22092378	21987112	0.55	cps
Manganese	55-2	26912677	26926947	26878147	26905924	0.09	cps
Molybdenum	94-1	112743	112810	114692	113415	0.98	cps
Molybdenum	95-1	18218	18402	19239	18620	2.92	cps
Molybdenum	96-1	47055	47925	47303	47428	0.94	cps
Molybdenum	97-1	10804	11354	11221	11127	2.58	cps
Molybdenum	98-1	28384	29263	29360	29002	1.85	cps
Neodymium	150-1	157128	157953	159214	158098	0.66	cps
Neodymium	150-2	80552	80418	80203	80391	0.22	cps
Nickel	60-2	3825993	3817882	3771477	3805117	0.77	cps
Phosphorus	31-2	25662	25308	25177	25382	0.99	cps
Potassium	39-2	2318879	2322604	2271963	2304482	1.22	cps
Rhodium	103-1	7244541	7241804	7159799	7215381	0.67	cps
Rhodium	103-2	4161849	4165393	4224766	4184003	0.84	cps
Scandium	45-1	5253764	5198003	5206453	5219407	0.58	cps
Scandium	45-2	464140	468365	463974	465493	0.53	cps
Selenium	82-1	60162	60579	61417	60720	1.05	cps
Selenium	77-2	5758	5921	5775	5818	1.55	cps
Selenium	78-2	18251	18542	19052	18615	2.18	cps
Silicon	28-1	29209987	29545447	29543085	29432840	0.66	cps
Silver	107-1	1951478	1962579	1974834	1962964	0.60	cps
Silver	109-1	1845359	1873800	1880229	1866463	0.99	cps
Sodium	23-2	188958	191086	188014	189353	0.83	cps
Strontium	86-1	505937	514323	514479	511580	0.96	cps
Strontium	88-1	4563135	4609703	4610441	4594426	0.59	cps
Sulfur	34-1	143715	142252	141457	142475	0.80	cps
Terbium	159-1	11098728	11033872	11058188	11063596	0.30	cps
Terbium	159-2	6534302	6487266	6424532	6482033	0.85	cps
Thallium	203-1	1878231	1914553	1901910	1898231	0.97	cps
Thallium	205-1	4475967	4531728	4464735	4490810	0.80	cps
Tin	118-1	21640	21500	21767	21635	0.62	cps
Titanium	47-1	2572342	2616348	2632358	2607016	1.19	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03MS Instrumnet Name : P7
Client Sample ID : MYE4H2SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:18:48 DataFile Name : 035SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	42758	43053	43100	42970	0.43	cps
Vanadium	51-2	8360768	8300915	8338028	8333237	0.36	cps
Yttrium	89-1	14801841	14617514	14492901	14637419	1.06	cps
Yttrium	89-2	3454638	3465952	3429468	3450019	0.54	cps
Zinc	66-2	1424305	1427067	1424312	1425228	0.11	cps
Zirconium	90-1	275751	281794	284110	280552	1.54	cps
Zirconium	91-1	62010	63985	64601	63532	2.13	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-05DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2S Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:21:42 DataFile Name : 036SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4089090	4111342	4116953	4105795	0.36	cps
Antimony	121-1	195155	194206	196430	195264	0.57	cps
Arsenic	75-2	69371	70781	70466	70206	1.05	cps
Barium	135-1	1413690	1422596	1422737	1419674	0.37	cps
Barium	137-1	2640044	2668364	2646880	2651763	0.56	cps
Beryllium	9-1	15836	15491	15424	15584	1.42	cps
Bismuth	209-1	6742006	6714430	6807730	6754722	0.71	cps
Bismuth	209-2	4765802	4702013	4624364	4697393	1.51	cps
Bromine	81-1	10734	10641	10757	10711	0.58	cps
Bromine	81-2	247	310	297	284	11.74	cps
Cadmium	108-1	1530	1653	1560	1581	4.07	cps
Cadmium	106-1	5138	4818	4931	4962	3.27	cps
Cadmium	111-1	25938	26101	25749	25929	0.68	cps
Calcium	43-1	790996	795339	791997	792777	0.29	cps
Calcium	44-1	12069519	12029860	12057970	12052450	0.17	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1001833	1009504	1002139	1004492	0.43	cps
Cobalt	59-2	815311	817732	812880	815308	0.30	cps
Copper	63-2	540966	541906	541116	541329	0.09	cps
Dysprosium	156-1	26969	26779	26782	26843	0.41	cps
Dysprosium	156-2	18619	18329	19096	18681	2.07	cps
Erbium	164-1	24792	24632	25530	24985	1.92	cps
Erbium	164-2	15172	15489	15472	15377	1.16	cps
Gadolinium	160-1	27166	27490	26659	27105	1.55	cps
Gadolinium	160-2	17498	17574	17778	17617	0.82	cps
Holmium	165-1	11014792	10834513	10815431	10888245	1.01	cps
Holmium	165-2	6741001	6657962	6560531	6653165	1.36	cps
Indium	115-1	8199054	8210924	8167050	8192343	0.28	cps
Indium	115-2	2596493	2529081	2537231	2554268	1.44	cps
Iron	56-2	194183290	194320424	193050850	193851522	0.36	cps
Iron	57-2	4832771	4864696	4823463	4840310	0.45	cps
Iron	54-2	10666945	10639942	10584597	10630494	0.39	cps
Krypton	83-1	227	237	220	228	3.68	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-05DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2S Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:21:42 DataFile Name : 036SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	70930	72361	72415	71902	1.17	cps
Lead	207-1	61174	62580	61790	61848	1.14	cps
Lead	208-1	279426	281381	285298	282035	1.06	cps
Lithium	6-1	1390569	1402073	1353784	1382142	1.82	cps
Magnesium	24-2	5540324	5552673	5555041	5549346	0.14	cps
Manganese	55-2	5783383	5817634	5749350	5783456	0.59	cps
Molybdenum	94-1	29156	29503	29967	29542	1.38	cps
Molybdenum	95-1	5278	5014	5258	5183	2.83	cps
Molybdenum	96-1	12502	12372	12609	12494	0.95	cps
Molybdenum	97-1	3337	2980	3310	3209	6.19	cps
Molybdenum	98-1	7872	8136	8069	8026	1.71	cps
Neodymium	150-1	40588	40839	40839	40755	0.36	cps
Neodymium	150-2	21086	20351	20265	20567	2.19	cps
Nickel	60-2	360997	363641	362620	362419	0.37	cps
Phosphorus	31-2	6892	6475	6328	6565	4.45	cps
Potassium	39-2	658026	665445	661139	661536	0.56	cps
Rhodium	103-1	7601792	7490224	7351162	7481059	1.68	cps
Rhodium	103-2	4465413	4345581	4398198	4403064	1.36	cps
Scandium	45-1	4671441	4601581	4516334	4596452	1.69	cps
Scandium	45-2	425791	413257	413814	417621	1.70	cps
Selenium	82-1	3639	3787	3724	3717	2.00	cps
Selenium	77-2	363	383	413	387	6.51	cps
Selenium	78-2	1360	1240	1243	1281	5.33	cps
Silicon	28-1	7733413	7710882	7733733	7726009	0.17	cps
Silver	107-1	99088	99481	99424	99331	0.21	cps
Silver	109-1	93056	93264	93814	93378	0.42	cps
Sodium	23-2	68796	69151	68538	68828	0.45	cps
Strontium	86-1	130643	131615	131256	131171	0.37	cps
Strontium	88-1	1149916	1158351	1158053	1155440	0.41	cps
Sulfur	34-1	138413	139933	137499	138615	0.89	cps
Terbium	159-1	11132057	11281222	11009669	11140982	1.22	cps
Terbium	159-2	6676782	6594388	6560721	6610631	0.90	cps
Thallium	203-1	85104	86160	85463	85576	0.63	cps
Thallium	205-1	203903	204000	206568	204824	0.74	cps
Tin	118-1	6725	6845	6502	6691	2.61	cps
Titanium	47-1	671167	674065	672233	672488	0.22	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-05DLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2S Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:21:42 DataFile Name : 036SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	11478	11095	11322	11298	1.71	cps
Vanadium	51-2	776053	785831	778199	780028	0.66	cps
Yttrium	89-1	13632922	13667548	13599454	13633308	0.25	cps
Yttrium	89-2	3304367	3261134	3234072	3266524	1.09	cps
Zinc	66-2	116114	115798	115170	115694	0.42	cps
Zirconium	90-1	71858	72246	71860	71988	0.31	cps
Zirconium	91-1	16476	16139	16693	16436	1.70	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:24:50 DataFile Name : 037SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4404810	4360987	4368028	4377942	0.54	cps
Antimony	121-1	905607	916076	916458	912714	0.67	cps
Arsenic	75-2	83577	83801	84163	83847	0.35	cps
Barium	135-1	5527058	5535644	5553240	5538647	0.24	cps
Barium	137-1	9555334	9553252	9520713	9543100	0.20	cps
Beryllium	9-1	71586	72111	71670	71789	0.39	cps
Bismuth	209-1	6905005	6977806	6816825	6899879	1.17	cps
Bismuth	209-2	4712711	4619660	4725344	4685905	1.23	cps
Bromine	81-1	10504	11311	10831	10882	3.73	cps
Bromine	81-2	313	277	307	299	6.54	cps
Cadmium	108-1	7422	7482	7729	7544	2.16	cps
Cadmium	106-1	13476	13500	13239	13405	1.07	cps
Cadmium	111-1	117799	118304	118621	118241	0.35	cps
Calcium	43-1	783183	783648	784103	783645	0.06	cps
Calcium	44-1	11826997	11880164	11908369	11871843	0.35	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1633047	1583877	1589879	1602268	1.67	cps
Cobalt	59-2	3211964	3207900	3221048	3213637	0.21	cps
Copper	63-2	1328326	1321986	1309693	1320002	0.72	cps
Dysprosium	156-1	26235	26979	26862	26692	1.50	cps
Dysprosium	156-2	18839	18649	19333	18941	1.86	cps
Erbium	164-1	25053	25570	25557	25393	1.16	cps
Erbium	164-2	15442	15152	15105	15233	1.20	cps
Gadolinium	160-1	26776	27257	26675	26903	1.15	cps
Gadolinium	160-2	17775	17631	17304	17570	1.37	cps
Holmium	165-1	10816460	11097544	10674784	10862930	1.98	cps
Holmium	165-2	6643017	6518288	6595989	6585764	0.96	cps
Indium	115-1	8297831	8378774	8066291	8247632	1.97	cps
Indium	115-2	2547049	2519960	2481422	2516143	1.31	cps
Iron	56-2	197591810	195609404	194598757	195933324	0.78	cps
Iron	57-2	4801106	4818433	4793378	4804306	0.27	cps
Iron	54-2	10767776	10733395	10744516	10748562	0.16	cps
Krypton	83-1	177	253	260	230	20.13	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:24:50 DataFile Name : 037SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	176821	181134	179101	179018	1.21	cps
Lead	207-1	153260	154379	155016	154218	0.58	cps
Lead	208-1	704225	714411	714026	710887	0.81	cps
Lithium	6-1	1392521	1378524	1404562	1391869	0.94	cps
Magnesium	24-2	5600340	5504955	5505631	5536975	0.99	cps
Manganese	55-2	6949962	6829598	6737107	6838889	1.56	cps
Molybdenum	94-1	28545	28949	28995	28829	0.86	cps
Molybdenum	95-1	4934	5128	5018	5026	1.93	cps
Molybdenum	96-1	12585	12215	12549	12450	1.64	cps
Molybdenum	97-1	3077	3077	3140	3098	1.18	cps
Molybdenum	98-1	7812	7739	8259	7937	3.55	cps
Neodymium	150-1	40000	40682	40448	40377	0.86	cps
Neodymium	150-2	20846	21146	20969	20987	0.72	cps
Nickel	60-2	946215	944893	940198	943769	0.34	cps
Phosphorus	31-2	6465	6401	6421	6429	0.50	cps
Potassium	39-2	656867	650866	649022	652252	0.63	cps
Rhodium	103-1	7521208	7563649	7379532	7488130	1.29	cps
Rhodium	103-2	4388812	4360757	4375074	4374881	0.32	cps
Scandium	45-1	4605477	4658702	4530138	4598105	1.40	cps
Scandium	45-2	412118	417647	407142	412302	1.27	cps
Selenium	82-1	15988	16420	16253	16220	1.34	cps
Selenium	77-2	1460	1530	1643	1545	5.99	cps
Selenium	78-2	5168	4978	5004	5050	2.04	cps
Silicon	28-1	7636916	7730560	7690844	7686107	0.61	cps
Silver	107-1	491147	489531	490803	490494	0.17	cps
Silver	109-1	457835	457758	457938	457844	0.02	cps
Sodium	23-2	66777	67782	67377	67312	0.75	cps
Strontium	86-1	130106	131329	129931	130455	0.58	cps
Strontium	88-1	1140651	1141593	1146673	1142972	0.28	cps
Sulfur	34-1	138854	139638	140579	139690	0.62	cps
Terbium	159-1	11137202	11169836	11107478	11138172	0.28	cps
Terbium	159-2	6590483	6625427	6542958	6586289	0.63	cps
Thallium	203-1	418713	418609	422388	419903	0.51	cps
Thallium	205-1	1001793	1006569	1001543	1003302	0.28	cps
Tin	118-1	6495	6668	6562	6575	1.33	cps
Titanium	47-1	657113	665561	666780	663151	0.79	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-03MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4H2SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-02 19:24:50 DataFile Name : 037SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	10788	11151	11518	11153	3.28	cps
Vanadium	51-2	2087060	2096613	2060921	2081531	0.89	cps
Yttrium	89-1	13725012	13634764	13482634	13614137	0.90	cps
Yttrium	89-2	3311067	3221194	3261932	3264731	1.38	cps
Zinc	66-2	375343	375405	372549	374432	0.44	cps
Zirconium	90-1	69792	70193	70391	70125	0.44	cps
Zirconium	91-1	15482	16249	15875	15869	2.42	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-06 Instrumnet Name : P7
Client Sample ID : MYE4H3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:27:53 DataFile Name : 038SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	13524975	13614241	13593119	13577445	0.34	cps
Antimony	121-1	52826	53666	53696	53396	0.92	cps
Arsenic	75-2	65550	65487	65617	65552	0.10	cps
Barium	135-1	591944	604184	605879	600669	1.27	cps
Barium	137-1	1032969	1050280	1051583	1044944	0.99	cps
Beryllium	9-1	1711	1795	1872	1793	4.49	cps
Bismuth	209-1	6700860	6725439	6731641	6719314	0.24	cps
Bismuth	209-2	4686883	4630212	4638204	4651766	0.66	cps
Bromine	81-1	11968	12342	12439	12250	2.03	cps
Bromine	81-2	433	437	390	420	6.20	cps
Cadmium	108-1	520	487	413	473	11.53	cps
Cadmium	106-1	3567	3677	3537	3594	2.05	cps
Cadmium	111-1	7181	7141	7011	7111	1.25	cps
Calcium	43-1	488579	506723	512956	502753	2.52	cps
Calcium	44-1	7478164	7576462	7757469	7604031	1.86	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	249197	251225	251304	250575	0.48	cps
Cobalt	59-2	429666	432017	430206	430630	0.29	cps
Copper	63-2	979933	977056	979353	978781	0.16	cps
Dysprosium	156-1	41765	41909	42203	41959	0.53	cps
Dysprosium	156-2	29939	29625	30293	29952	1.12	cps
Erbium	164-1	43427	44878	44501	44269	1.70	cps
Erbium	164-2	27036	27507	27230	27258	0.87	cps
Gadolinium	160-1	41956	42962	43086	42668	1.45	cps
Gadolinium	160-2	28359	28088	28359	28269	0.55	cps
Holmium	165-1	10878034	10949729	10775037	10867600	0.81	cps
Holmium	165-2	6789821	6653025	6625765	6689537	1.31	cps
Indium	115-1	8309809	8403857	8243662	8319109	0.97	cps
Indium	115-2	2633569	2589083	2618902	2613851	0.87	cps
Iron	56-2	461064713	462088766	464425753	462526411	0.37	cps
Iron	57-2	11471488	11451523	11610107	11511039	0.75	cps
Iron	54-2	25432787	25528585	25388792	25450055	0.28	cps
Krypton	83-1	313	233	277	274	14.59	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-06 Instrumnet Name : P7
Client Sample ID : MYE4H3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:27:53 DataFile Name : 038SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	241748	244703	240497	242316	0.89	cps
Lead	207-1	208743	210483	210353	209860	0.46	cps
Lead	208-1	945425	957617	950786	951276	0.64	cps
Lithium	6-1	1417397	1391832	1356273	1388501	2.21	cps
Magnesium	24-2	8899525	8837701	8851195	8862807	0.37	cps
Manganese	55-2	9225716	9274314	9193168	9231066	0.44	cps
Molybdenum	94-1	43090	45203	44735	44342	2.50	cps
Molybdenum	95-1	9960	10210	10167	10112	1.32	cps
Molybdenum	96-1	22881	22931	23759	23191	2.13	cps
Molybdenum	97-1	5768	6271	6438	6159	5.67	cps
Molybdenum	98-1	15715	15972	15765	15818	0.86	cps
Neodymium	150-1	61199	63780	63613	62864	2.30	cps
Neodymium	150-2	32434	32912	32187	32511	1.13	cps
Nickel	60-2	77040	76628	77438	77035	0.53	cps
Phosphorus	31-2	18007	17634	17727	17789	1.09	cps
Potassium	39-2	1349247	1371748	1350315	1357104	0.94	cps
Rhodium	103-1	7596704	7612337	7609372	7606137	0.11	cps
Rhodium	103-2	4535400	4523961	4490380	4516580	0.52	cps
Scandium	45-1	4778199	4896057	4873312	4849190	1.29	cps
Scandium	45-2	453716	446458	450786	450320	0.81	cps
Selenium	82-1	448	580	531	520	12.81	cps
Selenium	77-2	170	170	133	158	13.41	cps
Selenium	78-2	317	443	373	378	16.80	cps
Silicon	28-1	14234562	14597740	14694584	14508962	1.67	cps
Silver	107-1	11491	11365	10817	11224	3.19	cps
Silver	109-1	10634	10527	10551	10571	0.53	cps
Sodium	23-2	448425	448044	451480	449316	0.42	cps
Strontium	86-1	349485	360760	362442	357562	1.97	cps
Strontium	88-1	3133268	3195805	3253511	3194195	1.88	cps
Sulfur	34-1	138780	141400	142504	140894	1.36	cps
Terbium	159-1	11155077	11072327	11058038	11095148	0.47	cps
Terbium	159-2	6611137	6590445	6664821	6622134	0.58	cps
Thallium	203-1	2627	2714	2460	2600	4.95	cps
Thallium	205-1	6161	6168	6095	6141	0.66	cps
Tin	118-1	18442	18886	19270	18866	2.20	cps
Titanium	47-1	2951704	3075164	3098712	3041860	2.60	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-06 Instrumnet Name : P7
Client Sample ID : MYE4H3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:27:53 DataFile Name : 038SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	17308	17568	17188	17355	1.12	cps
Vanadium	51-2	816921	821957	822687	820522	0.38	cps
Yttrium	89-1	14196987	14320100	14371924	14296337	0.63	cps
Yttrium	89-2	3465544	3483493	3501702	3483580	0.52	cps
Zinc	66-2	380128	381348	383223	381566	0.41	cps
Zirconium	90-1	102417	105173	106603	104731	2.03	cps
Zirconium	91-1	22878	23796	23936	23537	2.44	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-07 Instrumnet Name : P7
Client Sample ID : MYE4H4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:30:59 DataFile Name : 039SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	11090336	11055051	10972717	11039368	0.55	cps
Antimony	121-1	71661	74545	73413	73206	1.99	cps
Arsenic	75-2	112361	111360	110756	111493	0.73	cps
Barium	135-1	1696036	1718440	1703424	1705967	0.67	cps
Barium	137-1	2908257	2960289	2963440	2943996	1.05	cps
Beryllium	9-1	2673	2817	3034	2841	6.40	cps
Bismuth	209-1	6768104	6892581	6819801	6826828	0.92	cps
Bismuth	209-2	4614228	4902350	4800454	4772344	3.06	cps
Bromine	81-1	13196	13169	13186	13184	0.10	cps
Bromine	81-2	597	487	523	536	10.46	cps
Cadmium	108-1	883	823	840	849	3.65	cps
Cadmium	106-1	4307	4584	4294	4395	3.73	cps
Cadmium	111-1	11033	11389	11254	11225	1.60	cps
Calcium	43-1	240226	244082	243050	242453	0.82	cps
Calcium	44-1	3659196	3690142	3687049	3678796	0.46	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	503375	501805	502368	502516	0.16	cps
Cobalt	59-2	438716	439704	437516	438645	0.25	cps
Copper	63-2	1987381	1992885	1988775	1989680	0.14	cps
Dysprosium	156-1	89674	89936	89043	89551	0.51	cps
Dysprosium	156-2	68067	67300	67629	67665	0.57	cps
Erbium	164-1	66300	67824	67576	67233	1.22	cps
Erbium	164-2	40900	40161	40977	40679	1.11	cps
Gadolinium	160-1	80168	82039	81365	81190	1.17	cps
Gadolinium	160-2	53562	53144	53970	53559	0.77	cps
Holmium	165-1	10818309	10992832	10861163	10890768	0.84	cps
Holmium	165-2	6613012	6901816	6711091	6741973	2.18	cps
Indium	115-1	8558521	8498904	8315930	8457785	1.49	cps
Indium	115-2	2518960	2675432	2613078	2602490	3.03	cps
Iron	56-2	521762539	515572179	515527672	517620797	0.69	cps
Iron	57-2	12789369	12874242	12640613	12768075	0.93	cps
Iron	54-2	28476819	28221714	28388720	28362417	0.46	cps
Krypton	83-1	317	283	373	324	14.02	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-07 Instrumnet Name : P7
Client Sample ID : MYE4H4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:30:59 DataFile Name : 039SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1175795	1199634	1208536	1194655	1.42	cps
Lead	207-1	1041503	1071616	1067990	1060370	1.55	cps
Lead	208-1	5087655	5211836	5242117	5180536	1.58	cps
Lithium	6-1	1347442	1402311	1390969	1380241	2.10	cps
Magnesium	24-2	9558217	9499138	9546818	9534725	0.33	cps
Manganese	55-2	11981241	11848851	11798148	11876080	0.80	cps
Molybdenum	94-1	127910	129681	132200	129930	1.66	cps
Molybdenum	95-1	10457	10581	10767	10602	1.47	cps
Molybdenum	96-1	39764	39610	39644	39673	0.20	cps
Molybdenum	97-1	6492	6428	6612	6510	1.43	cps
Molybdenum	98-1	15932	16262	16286	16160	1.22	cps
Neodymium	150-1	170563	170767	173271	171533	0.88	cps
Neodymium	150-2	90455	89193	89365	89671	0.76	cps
Nickel	60-2	189455	186782	188624	188287	0.73	cps
Phosphorus	31-2	28089	29288	28687	28688	2.09	cps
Potassium	39-2	4107245	4036224	4044524	4062664	0.96	cps
Rhodium	103-1	7836071	7768399	7585317	7729929	1.68	cps
Rhodium	103-2	4385443	4596456	4482555	4488151	2.35	cps
Scandium	45-1	5060077	5089955	5058395	5069476	0.35	cps
Scandium	45-2	437431	459802	445350	447528	2.53	cps
Selenium	82-1	646	657	644	649	1.11	cps
Selenium	77-2	293	297	280	290	3.04	cps
Selenium	78-2	550	533	527	537	2.24	cps
Silicon	28-1	25160455	25616504	25540380	25439113	0.96	cps
Silver	107-1	10631	10557	10811	10666	1.22	cps
Silver	109-1	9727	10190	10110	10009	2.48	cps
Sodium	23-2	130735	129258	130538	130177	0.62	cps
Strontium	86-1	298682	301193	302850	300908	0.70	cps
Strontium	88-1	2673616	2720620	2792148	2728795	2.19	cps
Sulfur	34-1	150560	147699	148432	148897	1.00	cps
Terbium	159-1	11211240	11475432	11186041	11290904	1.42	cps
Terbium	159-2	6515487	6765447	6645180	6642038	1.88	cps
Thallium	203-1	5354	5671	5441	5489	2.98	cps
Thallium	205-1	12746	12669	12739	12718	0.33	cps
Tin	118-1	29380	30195	30472	30016	1.89	cps
Titanium	47-1	1599881	1608390	1619747	1609339	0.62	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-07 Instrumnet Name : P7
Client Sample ID : MYE4H4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:30:59 DataFile Name : 039SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	39940	39706	39956	39867	0.35	cps
Vanadium	51-2	1054368	1047725	1052115	1051403	0.32	cps
Yttrium	89-1	14871679	14817686	14775464	14821610	0.33	cps
Yttrium	89-2	3471299	3597313	3488465	3519026	1.94	cps
Zinc	66-2	310892	310287	311803	310994	0.25	cps
Zirconium	90-1	334498	334372	338637	335836	0.72	cps
Zirconium	91-1	74444	76233	76397	75692	1.43	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-08 Instrumnet Name : P7
Client Sample ID : MYE4H5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:34:02 DataFile Name : 040SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	13874073	13942761	13570228	13795687	1.44	cps
Antimony	121-1	24451	25049	24815	24771	1.22	cps
Arsenic	75-2	76978	77622	77099	77233	0.44	cps
Barium	135-1	2120458	2166981	2188702	2158714	1.62	cps
Barium	137-1	3625690	3701638	3725443	3684257	1.41	cps
Beryllium	9-1	3367	3034	3406	3269	6.26	cps
Bismuth	209-1	6960240	6749706	6830811	6846919	1.55	cps
Bismuth	209-2	4696915	4155969	4639286	4497390	6.61	cps
Bromine	81-1	16222	17110	16696	16676	2.66	cps
Bromine	81-2	1243	1297	1183	1241	4.57	cps
Cadmium	108-1	580	423	500	501	15.63	cps
Cadmium	106-1	3894	3907	3987	3929	1.28	cps
Cadmium	111-1	4707	4766	4738	4737	0.62	cps
Calcium	43-1	363961	369719	369565	367748	0.89	cps
Calcium	44-1	5582965	5588773	5623718	5598485	0.39	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	557197	561074	550841	556371	0.93	cps
Cobalt	59-2	385310	386962	384259	385510	0.35	cps
Copper	63-2	1060619	1059991	1052273	1057628	0.44	cps
Dysprosium	156-1	106218	107843	107850	107304	0.88	cps
Dysprosium	156-2	80660	81924	81846	81477	0.87	cps
Erbium	164-1	79695	82361	81325	81127	1.66	cps
Erbium	164-2	49324	49561	49962	49616	0.65	cps
Gadolinium	160-1	95996	98104	97962	97354	1.21	cps
Gadolinium	160-2	64176	64568	63138	63961	1.16	cps
Holmium	165-1	11074269	10971804	10747509	10931194	1.53	cps
Holmium	165-2	6680914	5933549	6690236	6434900	6.75	cps
Indium	115-1	8330035	8420112	8331609	8360585	0.62	cps
Indium	115-2	2585688	2275952	2529140	2463593	6.70	cps
Iron	56-2	447807500	455654967	449307953	450923473	0.92	cps
Iron	57-2	11145610	11299854	11069829	11171764	1.05	cps
Iron	54-2	24640584	24664820	24301048	24535484	0.83	cps
Krypton	83-1	270	300	320	297	8.48	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-08 Instrumnet Name : P7
Client Sample ID : MYE4H5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:34:02 DataFile Name : 040SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	348396	353427	355698	352507	1.06	cps
Lead	207-1	299847	305306	304765	303306	0.99	cps
Lead	208-1	1379586	1388956	1407772	1392105	1.03	cps
Lithium	6-1	1364520	1360256	1378379	1367718	0.69	cps
Magnesium	24-2	10294519	10406288	10189725	10296844	1.05	cps
Manganese	55-2	7987308	8163436	8003033	8051259	1.21	cps
Molybdenum	94-1	134840	137013	137885	136580	1.15	cps
Molybdenum	95-1	6878	7042	6955	6958	1.17	cps
Molybdenum	96-1	35954	36522	36703	36393	1.07	cps
Molybdenum	97-1	4364	4404	4371	4380	0.49	cps
Molybdenum	98-1	11168	11011	10917	11032	1.15	cps
Neodymium	150-1	207765	208493	211399	209219	0.92	cps
Neodymium	150-2	107429	107809	106797	107345	0.48	cps
Nickel	60-2	223529	222874	220235	222213	0.78	cps
Phosphorus	31-2	27091	26867	27655	27204	1.49	cps
Potassium	39-2	3683507	3725091	3699082	3702560	0.57	cps
Rhodium	103-1	7732274	7721133	7516138	7656515	1.59	cps
Rhodium	103-2	4397198	3949678	4413304	4253393	6.19	cps
Scandium	45-1	5092884	5087923	5052327	5077711	0.44	cps
Scandium	45-2	453439	409177	440676	434431	5.24	cps
Selenium	82-1	645	658	615	639	3.49	cps
Selenium	77-2	327	363	327	339	6.25	cps
Selenium	78-2	520	533	510	521	2.25	cps
Silicon	28-1	26772470	27248501	27380182	27133718	1.18	cps
Silver	107-1	4801	4758	4941	4833	1.98	cps
Silver	109-1	4467	4417	4271	4385	2.33	cps
Sodium	23-2	126434	127351	127781	127189	0.54	cps
Strontium	86-1	434328	441675	442722	439575	1.04	cps
Strontium	88-1	3950764	3966868	3979332	3965655	0.36	cps
Sulfur	34-1	140267	137643	139257	139056	0.95	cps
Terbium	159-1	11282010	11304993	11079519	11222174	1.11	cps
Terbium	159-2	6821177	6029689	6691841	6514236	6.52	cps
Thallium	203-1	4718	4784	4724	4742	0.77	cps
Thallium	205-1	11278	11201	11311	11264	0.50	cps
Tin	118-1	25469	26311	26408	26063	1.98	cps
Titanium	47-1	1805474	1849505	1836144	1830374	1.23	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-08 Instrumnet Name : P7
Client Sample ID : MYE4H5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:34:02 DataFile Name : 040SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	55512	55712	56496	55907	0.93	cps
Vanadium	51-2	926212	930945	921149	926102	0.53	cps
Yttrium	89-1	15054652	14898528	14875575	14942918	0.65	cps
Yttrium	89-2	3526139	3184694	3536018	3415617	5.86	cps
Zinc	66-2	164549	165321	162203	164024	0.99	cps
Zirconium	90-1	357741	363250	364535	361842	1.00	cps
Zirconium	91-1	78284	82215	81511	80670	2.60	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-09 Instrumnet Name : P7
Client Sample ID : MYE4H6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:37:05 DataFile Name : 041SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8927184	8838463	8836284	8867311	0.58	cps
Antimony	121-1	41811	43064	41767	42214	1.75	cps
Arsenic	75-2	96407	96749	95810	96322	0.49	cps
Barium	135-1	1269747	1283083	1278713	1277181	0.53	cps
Barium	137-1	2377498	2383089	2387098	2382562	0.20	cps
Beryllium	9-1	2295	2211	2272	2259	1.91	cps
Bismuth	209-1	6789728	6922924	6799135	6837262	1.09	cps
Bismuth	209-2	4772046	4756811	4712648	4747169	0.65	cps
Bromine	81-1	12976	13199	12746	12974	1.75	cps
Bromine	81-2	520	480	490	497	4.19	cps
Cadmium	108-1	523	720	753	666	18.67	cps
Cadmium	106-1	4247	4157	4127	4177	1.50	cps
Cadmium	111-1	8575	8611	8350	8512	1.66	cps
Calcium	43-1	235601	235834	233983	235139	0.43	cps
Calcium	44-1	3579755	3582673	3580552	3580993	0.04	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	407977	404807	406096	406293	0.39	cps
Cobalt	59-2	304923	305940	308215	306359	0.55	cps
Copper	63-2	1276525	1292817	1276505	1281949	0.73	cps
Dysprosium	156-1	76943	77804	78813	77853	1.20	cps
Dysprosium	156-2	60222	59074	59850	59715	0.98	cps
Erbium	164-1	56725	59171	58368	58088	2.15	cps
Erbium	164-2	35368	35298	34436	35034	1.48	cps
Gadolinium	160-1	70550	70898	72479	71309	1.44	cps
Gadolinium	160-2	45771	46554	46584	46303	0.99	cps
Holmium	165-1	11056807	10981603	10801515	10946642	1.20	cps
Holmium	165-2	6680590	6650356	6548331	6626426	1.05	cps
Indium	115-1	8524516	8372602	8249356	8382158	1.64	cps
Indium	115-2	2580719	2561050	2570948	2570906	0.38	cps
Iron	56-2	399531101	401202621	400012167	400248630	0.21	cps
Iron	57-2	9862261	9930774	9837942	9876992	0.49	cps
Iron	54-2	21797886	21668743	21995913	21820847	0.76	cps
Krypton	83-1	270	253	237	253	6.58	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-09 Instrumnet Name : P7
Client Sample ID : MYE4H6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:37:05 DataFile Name : 041SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	723215	734980	734967	731054	0.93	cps
Lead	207-1	640566	651194	649712	647158	0.89	cps
Lead	208-1	2897628	2943088	3193770	3011496	5.30	cps
Lithium	6-1	1368658	1393874	1390179	1384237	0.98	cps
Magnesium	24-2	7096822	7156694	7024248	7092588	0.94	cps
Manganese	55-2	6121571	6167708	6153036	6147439	0.38	cps
Molybdenum	94-1	123774	125902	125899	125192	0.98	cps
Molybdenum	95-1	9583	9527	9490	9533	0.49	cps
Molybdenum	96-1	35242	36509	36379	36043	1.93	cps
Molybdenum	97-1	5921	5691	5785	5799	2.00	cps
Molybdenum	98-1	14564	14931	14728	14741	1.25	cps
Neodymium	150-1	149490	153074	153020	151861	1.35	cps
Neodymium	150-2	77284	77578	77880	77581	0.38	cps
Nickel	60-2	155568	154294	154726	154863	0.42	cps
Phosphorus	31-2	24740	24917	25455	25037	1.49	cps
Potassium	39-2	3572683	3602052	3562371	3579035	0.58	cps
Rhodium	103-1	7762504	7599599	7606863	7656322	1.20	cps
Rhodium	103-2	4444955	4376916	4366073	4395982	0.97	cps
Scandium	45-1	4878133	4788303	4792522	4819653	1.05	cps
Scandium	45-2	428298	429335	427657	428430	0.20	cps
Selenium	82-1	681	639	683	668	3.69	cps
Selenium	77-2	300	343	207	283	24.65	cps
Selenium	78-2	523	503	497	508	2.73	cps
Silicon	28-1	22318487	22754763	22489217	22520822	0.98	cps
Silver	107-1	6908	7419	7309	7212	3.72	cps
Silver	109-1	6782	6865	6588	6745	2.11	cps
Sodium	23-2	120030	120309	119546	119962	0.32	cps
Strontium	86-1	284487	282627	284178	283764	0.35	cps
Strontium	88-1	2534585	2568504	2518213	2540434	1.01	cps
Sulfur	34-1	140223	139732	142039	140665	0.86	cps
Terbium	159-1	11451172	11257093	10958970	11222412	2.21	cps
Terbium	159-2	6663075	6535466	6566336	6588292	1.01	cps
Thallium	203-1	4421	4371	4607	4466	2.79	cps
Thallium	205-1	10434	10985	10481	10633	2.87	cps
Tin	118-1	27940	27767	28822	28176	2.01	cps
Titanium	47-1	1538227	1570598	1581420	1563415	1.44	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-09 Instrumnet Name : P7
Client Sample ID : MYE4H6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:37:05 DataFile Name : 041SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	35558	35715	36473	35915	1.36	cps
Vanadium	51-2	781474	779184	784945	781868	0.37	cps
Yttrium	89-1	14554050	14430514	14279019	14421195	0.96	cps
Yttrium	89-2	3457882	3458632	3405341	3440619	0.89	cps
Zinc	66-2	248533	246150	246906	247196	0.49	cps
Zirconium	90-1	322316	331701	329376	327798	1.49	cps
Zirconium	91-1	72394	73443	73627	73154	0.91	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-10 Instrumnet Name : P7
Client Sample ID : MYE4H8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:40:10 DataFile Name : 042SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8984993	9041575	8988711	9005093	0.35	cps
Antimony	121-1	34404	35026	34551	34661	0.94	cps
Arsenic	75-2	66417	67288	66940	66882	0.66	cps
Barium	135-1	1299892	1308737	1318194	1308941	0.70	cps
Barium	137-1	2434790	2462096	2418208	2438364	0.91	cps
Beryllium	9-1	2284	2006	2222	2171	6.73	cps
Bismuth	209-1	6830769	6770476	6802964	6801403	0.44	cps
Bismuth	209-2	4626508	4694667	4681056	4667410	0.77	cps
Bromine	81-1	12549	12976	12632	12719	1.78	cps
Bromine	81-2	577	390	583	517	21.24	cps
Cadmium	108-1	713	627	747	696	8.91	cps
Cadmium	106-1	3977	4027	3757	3921	3.67	cps
Cadmium	111-1	8996	9285	8722	9001	3.13	cps
Calcium	43-1	254063	256713	255638	255471	0.52	cps
Calcium	44-1	3862680	3873878	3910215	3882258	0.64	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	405421	405179	405821	405473	0.08	cps
Cobalt	59-2	311742	310544	309694	310660	0.33	cps
Copper	63-2	1803225	1814348	1763355	1793643	1.49	cps
Dysprosium	156-1	76493	77026	76011	76510	0.66	cps
Dysprosium	156-2	57691	58330	57775	57932	0.60	cps
Erbium	164-1	56922	57852	57611	57462	0.84	cps
Erbium	164-2	34335	35134	35264	34911	1.44	cps
Gadolinium	160-1	70315	70248	70473	70345	0.16	cps
Gadolinium	160-2	45601	44952	45142	45232	0.74	cps
Holmium	165-1	10998512	11005222	10854319	10952685	0.78	cps
Holmium	165-2	6550790	6525131	6527514	6534479	0.22	cps
Indium	115-1	8264707	8331215	8182550	8259491	0.90	cps
Indium	115-2	2531287	2508532	2535860	2525226	0.58	cps
Iron	56-2	351984501	354608035	355460955	354017830	0.51	cps
Iron	57-2	8733679	8701646	8736286	8723870	0.22	cps
Iron	54-2	19255352	19403436	19210763	19289850	0.52	cps
Krypton	83-1	280	303	270	284	6.01	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-10 Instrumnet Name : P7
Client Sample ID : MYE4H8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:40:10 DataFile Name : 042SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	734326	730834	740482	735214	0.66	cps
Lead	207-1	648093	648432	647756	648094	0.05	cps
Lead	208-1	2926191	2926374	2944345	2932303	0.36	cps
Lithium	6-1	1379162	1356977	1318310	1351483	2.28	cps
Magnesium	24-2	7146362	7242658	7233263	7207428	0.74	cps
Manganese	55-2	7078824	7135005	7060189	7091339	0.55	cps
Molybdenum	94-1	94311	94888	94227	94475	0.38	cps
Molybdenum	95-1	8993	8899	9303	9065	2.33	cps
Molybdenum	96-1	30111	29530	30151	29931	1.16	cps
Molybdenum	97-1	5321	5688	5621	5543	3.52	cps
Molybdenum	98-1	14180	13860	14204	14081	1.36	cps
Neodymium	150-1	144427	145724	145966	145373	0.57	cps
Neodymium	150-2	75420	75698	74784	75301	0.62	cps
Nickel	60-2	153463	152627	153156	153082	0.28	cps
Phosphorus	31-2	24673	24730	24132	24512	1.35	cps
Potassium	39-2	3812916	3845160	3807220	3821765	0.54	cps
Rhodium	103-1	7584230	7644637	7517839	7582235	0.84	cps
Rhodium	103-2	4353827	4360160	4330913	4348300	0.35	cps
Scandium	45-1	4761285	4747887	4739857	4749676	0.23	cps
Scandium	45-2	420342	423620	420937	421633	0.41	cps
Selenium	82-1	660	658	623	647	3.25	cps
Selenium	77-2	240	257	243	247	3.58	cps
Selenium	78-2	383	487	467	446	12.30	cps
Silicon	28-1	23969365	23992493	24190779	24050879	0.51	cps
Silver	107-1	9196	9010	9033	9080	1.12	cps
Silver	109-1	8566	8626	8479	8557	0.86	cps
Sodium	23-2	126249	126088	126151	126163	0.06	cps
Strontium	86-1	318532	321646	321277	320485	0.53	cps
Strontium	88-1	2867367	2916619	2859785	2881257	1.07	cps
Sulfur	34-1	137081	136866	134338	136095	1.12	cps
Terbium	159-1	11094428	11111947	10989690	11065355	0.60	cps
Terbium	159-2	6558017	6561641	6534152	6551270	0.23	cps
Thallium	203-1	3751	4001	3891	3881	3.23	cps
Thallium	205-1	8910	9040	9093	9014	1.05	cps
Tin	118-1	28502	28876	28615	28664	0.67	cps
Titanium	47-1	1433384	1410190	1436314	1426629	1.00	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-10 Instrumnet Name : P7
Client Sample ID : MYE4H8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:40:10 DataFile Name : 042SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	28868	29737	29696	29434	1.67	cps
Vanadium	51-2	684500	690837	687841	687726	0.46	cps
Yttrium	89-1	14283845	14337674	14223315	14281611	0.40	cps
Yttrium	89-2	3425670	3406678	3434147	3422165	0.41	cps
Zinc	66-2	254371	256462	252917	254583	0.70	cps
Zirconium	90-1	242110	244595	245163	243956	0.67	cps
Zirconium	91-1	54273	54795	55558	54875	1.18	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-11 Instrumnet Name : P7
Client Sample ID : MYE4H9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:43:13 DataFile Name : 043SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	11154844	11287647	11199334	11213942	0.60	cps
Antimony	121-1	44870	45602	45094	45189	0.83	cps
Arsenic	75-2	75095	73923	74599	74539	0.79	cps
Barium	135-1	1981791	2019608	2015057	2005485	1.03	cps
Barium	137-1	3381872	3394419	3456327	3410873	1.17	cps
Beryllium	9-1	2678	2739	2756	2724	1.50	cps
Bismuth	209-1	6764816	6905692	6353404	6674637	4.30	cps
Bismuth	209-2	4649609	4706821	4580222	4645551	1.36	cps
Bromine	81-1	14427	14781	15058	14755	2.14	cps
Bromine	81-2	953	877	973	934	5.46	cps
Cadmium	108-1	627	520	577	574	9.29	cps
Cadmium	106-1	3994	3984	3614	3864	5.61	cps
Cadmium	111-1	5928	6205	5746	5960	3.88	cps
Calcium	43-1	281038	287574	284874	284495	1.15	cps
Calcium	44-1	4263543	4322519	4277619	4287894	0.72	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	549395	549752	547177	548775	0.25	cps
Cobalt	59-2	393249	391042	393889	392727	0.38	cps
Copper	63-2	1181271	1179487	1176726	1179161	0.19	cps
Dysprosium	156-1	95458	96680	96965	96368	0.83	cps
Dysprosium	156-2	73608	73397	73487	73497	0.14	cps
Erbium	164-1	72614	72467	73760	72947	0.97	cps
Erbium	164-2	44223	44280	43574	44026	0.89	cps
Gadolinium	160-1	87753	89030	89369	88717	0.96	cps
Gadolinium	160-2	58063	57862	57056	57660	0.92	cps
Holmium	165-1	11021083	10813334	10391832	10742083	2.98	cps
Holmium	165-2	6539886	6606130	6655065	6600360	0.88	cps
Indium	115-1	8338690	8284681	7792612	8138661	3.70	cps
Indium	115-2	2562388	2557198	2566841	2562143	0.19	cps
Iron	56-2	443998260	442424460	445058060	443826927	0.30	cps
Iron	57-2	10894040	10815822	10885330	10865064	0.39	cps
Iron	54-2	24000849	24026256	23966119	23997741	0.13	cps
Krypton	83-1	237	320	277	278	15.00	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-11 Instrumnet Name : P7
Client Sample ID : MYE4H9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:43:13 DataFile Name : 043SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	477903	488639	483745	483429	1.11	cps
Lead	207-1	420260	423276	426941	423493	0.79	cps
Lead	208-1	1906617	1933140	1925572	1921776	0.71	cps
Lithium	6-1	1362788	1308472	1288202	1319821	2.92	cps
Magnesium	24-2	8532260	8610689	8513703	8552217	0.60	cps
Manganese	55-2	10737169	10691357	10733222	10720583	0.24	cps
Molybdenum	94-1	120611	123364	124002	122659	1.47	cps
Molybdenum	95-1	11288	11174	11391	11284	0.96	cps
Molybdenum	96-1	36586	38400	38404	37797	2.77	cps
Molybdenum	97-1	6678	7052	7142	6957	3.53	cps
Molybdenum	98-1	17591	17380	17307	17426	0.84	cps
Neodymium	150-1	184960	184865	187351	185725	0.76	cps
Neodymium	150-2	96186	97034	95568	96262	0.76	cps
Nickel	60-2	195186	193469	192378	193677	0.73	cps
Phosphorus	31-2	24406	24486	23855	24249	1.42	cps
Potassium	39-2	4146292	4067753	4144213	4119419	1.09	cps
Rhodium	103-1	7540630	7534384	7248313	7441109	2.24	cps
Rhodium	103-2	4442762	4381583	4397566	4407304	0.72	cps
Scandium	45-1	4904353	4857645	4635761	4799253	2.99	cps
Scandium	45-2	434341	431682	436744	434256	0.58	cps
Selenium	82-1	766	627	662	685	10.56	cps
Selenium	77-2	347	390	317	351	10.50	cps
Selenium	78-2	527	520	607	551	8.75	cps
Silicon	28-1	9052014	9299963	9290564	9214180	1.53	cps
Silver	107-1	5611	5778	5698	5696	1.46	cps
Silver	109-1	5144	5288	5154	5195	1.54	cps
Sodium	23-2	120262	120974	120748	120661	0.30	cps
Strontium	86-1	362846	369243	368395	366828	0.95	cps
Strontium	88-1	3244570	3284444	3343861	3290959	1.52	cps
Sulfur	34-1	137199	138470	141474	139048	1.58	cps
Terbium	159-1	11339899	11102877	10678649	11040475	3.03	cps
Terbium	159-2	6451721	6502794	6638815	6531110	1.48	cps
Thallium	203-1	5164	5411	5371	5316	2.49	cps
Thallium	205-1	12549	12025	12623	12399	2.63	cps
Tin	118-1	27904	25242	25583	26243	5.52	cps
Titanium	47-1	1455276	1464301	1498975	1472851	1.57	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-11 Instrumnet Name : P7
Client Sample ID : MYE4H9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:43:13 DataFile Name : 043SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	45829	46732	46822	46461	1.18	cps
Vanadium	51-2	871646	877035	866766	871816	0.59	cps
Yttrium	89-1	14709311	14427775	13957968	14365018	2.64	cps
Yttrium	89-2	3442803	3501718	3472880	3472467	0.85	cps
Zinc	66-2	188018	185771	187650	187146	0.64	cps
Zirconium	90-1	311421	319132	317199	315917	1.27	cps
Zirconium	91-1	69393	70334	71516	70414	1.51	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-12 Instrumnet Name : P7
Client Sample ID : MYE4J0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:46:19 DataFile Name : 044SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8655968	8628923	8679541	8654811	0.29	cps
Antimony	121-1	33018	33800	33095	33304	1.29	cps
Arsenic	75-2	69950	68909	71086	69982	1.56	cps
Barium	135-1	1172993	1186365	1198902	1186087	1.09	cps
Barium	137-1	2208392	2182828	2263824	2218348	1.87	cps
Beryllium	9-1	2111	2245	2350	2235	5.36	cps
Bismuth	209-1	6080038	6696302	6614178	6463506	5.18	cps
Bismuth	209-2	4684282	4671108	4646758	4667382	0.41	cps
Bromine	81-1	12345	12512	12976	12611	2.59	cps
Bromine	81-2	647	553	550	583	9.41	cps
Cadmium	108-1	680	607	620	636	6.15	cps
Cadmium	106-1	3874	4111	3751	3912	4.68	cps
Cadmium	111-1	8644	9004	8753	8800	2.10	cps
Calcium	43-1	202665	205768	208512	205648	1.42	cps
Calcium	44-1	3091306	3097695	3161655	3116885	1.25	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	381559	388082	382557	384066	0.91	cps
Cobalt	59-2	300758	299207	300009	299991	0.26	cps
Copper	63-2	1448402	1436637	1443308	1442782	0.41	cps
Dysprosium	156-1	73782	73712	74808	74101	0.83	cps
Dysprosium	156-2	55292	55750	56824	55955	1.41	cps
Erbium	164-1	54616	56410	56333	55786	1.82	cps
Erbium	164-2	33049	33981	33941	33657	1.57	cps
Gadolinium	160-1	67904	66869	68319	67697	1.10	cps
Gadolinium	160-2	43925	44701	44337	44321	0.88	cps
Holmium	165-1	9663645	10907086	10764247	10444993	6.51	cps
Holmium	165-2	6593434	6698814	6574199	6622149	1.01	cps
Indium	115-1	7338965	8196232	8236713	7923970	6.40	cps
Indium	115-2	2514496	2521906	2564925	2533775	1.07	cps
Iron	56-2	352786968	354908688	360106941	355934199	1.06	cps
Iron	57-2	8829026	8836509	8873188	8846241	0.27	cps
Iron	54-2	19294191	19546088	19296504	19378928	0.75	cps
Krypton	83-1	240	263	293	266	10.07	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-12 Instrumnet Name : P7
Client Sample ID : MYE4J0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:46:19 DataFile Name : 044SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	766564	772635	768644	769281	0.40	cps
Lead	207-1	667528	675162	681066	674585	1.01	cps
Lead	208-1	3275803	3294369	3338631	3302934	0.98	cps
Lithium	6-1	1174943	1331977	1312498	1273139	6.72	cps
Magnesium	24-2	7300072	7230452	7166233	7232252	0.93	cps
Manganese	55-2	6168838	6168699	6199157	6178898	0.28	cps
Molybdenum	94-1	78372	80533	81813	80239	2.17	cps
Molybdenum	95-1	8096	8366	7829	8097	3.32	cps
Molybdenum	96-1	25920	26724	27042	26562	2.18	cps
Molybdenum	97-1	5071	5098	5074	5081	0.29	cps
Molybdenum	98-1	12305	13136	12616	12686	3.31	cps
Neodymium	150-1	140458	142253	144939	142550	1.58	cps
Neodymium	150-2	73005	74834	73681	73840	1.25	cps
Nickel	60-2	142041	143657	144091	143263	0.75	cps
Phosphorus	31-2	23398	23825	23124	23449	1.51	cps
Potassium	39-2	3442342	3480269	3489069	3470560	0.72	cps
Rhodium	103-1	6697242	7541666	7523068	7253992	6.65	cps
Rhodium	103-2	4365828	4425989	4386969	4392929	0.69	cps
Scandium	45-1	4269141	4792255	4730748	4597381	6.22	cps
Scandium	45-2	426305	425078	426214	425866	0.16	cps
Selenium	82-1	614	633	661	636	3.69	cps
Selenium	77-2	290	230	277	266	11.87	cps
Selenium	78-2	420	480	460	453	6.74	cps
Silicon	28-1	17617038	17926194	18052008	17865080	1.25	cps
Silver	107-1	8219	8046	8352	8206	1.88	cps
Silver	109-1	7562	7729	7505	7599	1.53	cps
Sodium	23-2	114854	114572	115056	114827	0.21	cps
Strontium	86-1	246997	251479	255071	251182	1.61	cps
Strontium	88-1	2244696	2274086	2263585	2260789	0.66	cps
Sulfur	34-1	136365	136789	137404	136853	0.38	cps
Terbium	159-1	9788544	11050734	11050039	10629772	6.85	cps
Terbium	159-2	6513403	6590633	6609053	6571030	0.77	cps
Thallium	203-1	4087	4191	4324	4201	2.83	cps
Thallium	205-1	9607	9954	10131	9897	2.69	cps
Tin	118-1	27262	29113	27827	28068	3.38	cps
Titanium	47-1	1219894	1238261	1257510	1238555	1.52	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-12 Instrumnet Name : P7
Client Sample ID : MYE4J0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:46:19 DataFile Name : 044SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	33816	34221	33332	33789	1.32	cps
Vanadium	51-2	721430	718730	719231	719797	0.20	cps
Yttrium	89-1	12778779	14355215	14201764	13778586	6.31	cps
Yttrium	89-2	3429077	3392022	3396668	3405922	0.59	cps
Zinc	66-2	241289	240856	244125	242090	0.73	cps
Zirconium	90-1	202013	203778	207708	204500	1.43	cps
Zirconium	91-1	44889	45547	46620	45685	1.91	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-13 Instrumnet Name : P7
Client Sample ID : MYE4J1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:49:23 DataFile Name : 045SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8399187	8318555	8215018	8310920	1.11	cps
Antimony	121-1	32079	32507	32991	32526	1.40	cps
Arsenic	75-2	70811	70711	70175	70565	0.48	cps
Barium	135-1	1105396	1125527	1124909	1118610	1.02	cps
Barium	137-1	2067696	2093928	2148880	2103501	1.97	cps
Beryllium	9-1	1900	2095	2172	2056	6.82	cps
Bismuth	209-1	6799593	6756969	6334923	6630495	3.87	cps
Bismuth	209-2	4775417	4782452	4745344	4767738	0.41	cps
Bromine	81-1	12856	13083	13376	13105	1.99	cps
Bromine	81-2	553	463	577	531	11.27	cps
Cadmium	108-1	497	677	527	567	17.02	cps
Cadmium	106-1	3807	3751	3454	3671	5.17	cps
Cadmium	111-1	7888	7830	7490	7736	2.78	cps
Calcium	43-1	198963	203082	203364	201803	1.22	cps
Calcium	44-1	3072894	3078959	3089887	3080580	0.28	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	358498	357071	354752	356774	0.53	cps
Cobalt	59-2	299016	300588	297952	299185	0.44	cps
Copper	63-2	1275849	1270184	1275669	1273901	0.25	cps
Dysprosium	156-1	73246	74771	75796	74604	1.72	cps
Dysprosium	156-2	56118	56035	56978	56377	0.93	cps
Erbium	164-1	55088	54292	55891	55090	1.45	cps
Erbium	164-2	34463	33871	33995	34109	0.91	cps
Gadolinium	160-1	68215	68979	67890	68361	0.82	cps
Gadolinium	160-2	44955	44908	44975	44946	0.08	cps
Holmium	165-1	10919744	10834075	10046582	10600134	4.54	cps
Holmium	165-2	6838146	6598200	6747196	6727847	1.80	cps
Indium	115-1	8340298	8314874	7582061	8079078	5.33	cps
Indium	115-2	2582771	2562356	2552337	2565821	0.60	cps
Iron	56-2	336471915	330592395	333638462	333567591	0.88	cps
Iron	57-2	8236735	8323633	8244555	8268307	0.58	cps
Iron	54-2	18239443	18260589	18253044	18251025	0.06	cps
Krypton	83-1	290	200	233	241	18.87	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-13 Instrumnet Name : P7
Client Sample ID : MYE4J1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:49:23 DataFile Name : 045SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	634313	641991	656410	644238	1.74	cps
Lead	207-1	559848	570706	567953	566169	1.00	cps
Lead	208-1	2529251	2551290	2593230	2557924	1.27	cps
Lithium	6-1	1342682	1338195	1235995	1305624	4.62	cps
Magnesium	24-2	6580882	6557198	6556795	6564959	0.21	cps
Manganese	55-2	6437921	6468504	6431869	6446098	0.30	cps
Molybdenum	94-1	82293	84086	83345	83241	1.08	cps
Molybdenum	95-1	8653	8733	8723	8703	0.50	cps
Molybdenum	96-1	26584	27763	28351	27566	3.26	cps
Molybdenum	97-1	5304	5314	5348	5322	0.43	cps
Molybdenum	98-1	12969	13269	13386	13208	1.63	cps
Neodymium	150-1	141020	141831	142694	141849	0.59	cps
Neodymium	150-2	74331	75028	73999	74453	0.71	cps
Nickel	60-2	138358	137921	136178	137486	0.84	cps
Phosphorus	31-2	22500	22754	22710	22655	0.60	cps
Potassium	39-2	3538427	3539305	3517143	3531625	0.36	cps
Rhodium	103-1	7607652	7590476	6954452	7384193	5.04	cps
Rhodium	103-2	4422884	4388531	4437288	4416234	0.57	cps
Scandium	45-1	4716720	4737629	4336492	4596947	4.91	cps
Scandium	45-2	423553	420732	418557	420947	0.59	cps
Selenium	82-1	593	656	620	623	5.06	cps
Selenium	77-2	270	303	203	259	19.67	cps
Selenium	78-2	400	423	410	411	2.85	cps
Silicon	28-1	21038589	21397968	21581681	21339413	1.29	cps
Silver	107-1	6755	6818	7018	6864	2.00	cps
Silver	109-1	6538	6295	6685	6506	3.03	cps
Sodium	23-2	115029	115321	113326	114559	0.94	cps
Strontium	86-1	234769	237763	238095	236876	0.77	cps
Strontium	88-1	2143041	2114623	2152702	2136789	0.93	cps
Sulfur	34-1	139934	139221	137253	138802	1.00	cps
Terbium	159-1	11149839	11161462	10236150	10849150	4.89	cps
Terbium	159-2	6681344	6682997	6622610	6662317	0.52	cps
Thallium	203-1	3611	3804	3681	3698	2.65	cps
Thallium	205-1	8453	8613	8609	8558	1.07	cps
Tin	118-1	27272	27730	27967	27657	1.28	cps
Titanium	47-1	1227239	1241766	1254929	1241311	1.12	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-13 Instrumnet Name : P7
Client Sample ID : MYE4J1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:49:23 DataFile Name : 045SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	32720	33165	33599	33161	1.33	cps
Vanadium	51-2	639904	638753	638705	639121	0.11	cps
Yttrium	89-1	14204111	14292583	13064879	13853858	4.94	cps
Yttrium	89-2	3402146	3422295	3355343	3393261	1.01	cps
Zinc	66-2	220555	221170	220599	220775	0.16	cps
Zirconium	90-1	211754	214570	215447	213924	0.90	cps
Zirconium	91-1	47775	48312	48165	48084	0.58	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-14 Instrumnet Name : P7
Client Sample ID : MYE4J2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:52:30 DataFile Name : 046SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	10143231	10172414	10234744	10183463	0.46	cps
Antimony	121-1	38628	38652	39665	38982	1.52	cps
Arsenic	75-2	81660	80731	80574	80988	0.72	cps
Barium	135-1	1439156	1462171	1463480	1454936	0.94	cps
Barium	137-1	2775284	2765408	2731562	2757418	0.83	cps
Beryllium	9-1	2395	2478	2539	2471	2.94	cps
Bismuth	209-1	6884797	6890354	6600526	6791892	2.44	cps
Bismuth	209-2	4689984	4742014	4731129	4721043	0.58	cps
Bromine	81-1	12852	13319	12395	12856	3.59	cps
Bromine	81-2	600	467	540	536	12.47	cps
Cadmium	108-1	757	817	827	800	4.73	cps
Cadmium	106-1	4131	3951	4307	4130	4.32	cps
Cadmium	111-1	9927	9412	9908	9749	3.00	cps
Calcium	43-1	246938	247199	247247	247128	0.07	cps
Calcium	44-1	3747795	3826303	3792465	3788854	1.04	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	480456	479396	479247	479699	0.14	cps
Cobalt	59-2	385462	384647	384638	384916	0.12	cps
Copper	63-2	1774974	1767466	1811696	1784712	1.33	cps
Dysprosium	156-1	87816	88574	87786	88058	0.51	cps
Dysprosium	156-2	67136	66386	66148	66557	0.77	cps
Erbium	164-1	67539	67211	67589	67446	0.30	cps
Erbium	164-2	41177	40261	40639	40692	1.13	cps
Gadolinium	160-1	79973	80392	81686	80684	1.11	cps
Gadolinium	160-2	52033	52886	52461	52460	0.81	cps
Holmium	165-1	10987113	11016317	10694534	10899321	1.63	cps
Holmium	165-2	6541945	6626146	6620694	6596262	0.71	cps
Indium	115-1	8286722	8144571	8001153	8144148	1.75	cps
Indium	115-2	2525946	2570669	2530656	2542424	0.97	cps
Iron	56-2	426367474	426666647	422421674	425151931	0.56	cps
Iron	57-2	10579018	10492182	10515060	10528753	0.43	cps
Iron	54-2	23356305	23189526	23265865	23270565	0.36	cps
Krypton	83-1	263	277	273	271	2.56	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-14 Instrumnet Name : P7
Client Sample ID : MYE4J2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:52:30 DataFile Name : 046SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	945072	965010	981944	964009	1.91	cps
Lead	207-1	859631	856617	863564	859937	0.41	cps
Lead	208-1	4156575	4125756	4192503	4158278	0.80	cps
Lithium	6-1	1334477	1348752	1321621	1334950	1.02	cps
Magnesium	24-2	8653098	8685488	8694636	8677741	0.25	cps
Manganese	55-2	8091716	8007722	8010787	8036742	0.59	cps
Molybdenum	94-1	119267	118370	121202	119613	1.21	cps
Molybdenum	95-1	10644	10417	10734	10598	1.54	cps
Molybdenum	96-1	36679	36683	36168	36510	0.81	cps
Molybdenum	97-1	6471	6478	6642	6530	1.47	cps
Molybdenum	98-1	16423	16172	16469	16355	0.98	cps
Neodymium	150-1	165358	167870	167338	166855	0.79	cps
Neodymium	150-2	85111	86654	86617	86128	1.02	cps
Nickel	60-2	186205	185749	183434	185129	0.80	cps
Phosphorus	31-2	27027	26516	26613	26719	1.02	cps
Potassium	39-2	4152095	4149099	4124548	4141914	0.36	cps
Rhodium	103-1	7557754	7573457	7210995	7447402	2.75	cps
Rhodium	103-2	4351195	4392356	4398654	4380735	0.59	cps
Scandium	45-1	4777387	4827318	4628838	4744514	2.18	cps
Scandium	45-2	431594	427035	427745	428791	0.57	cps
Selenium	82-1	691	710	633	678	5.92	cps
Selenium	77-2	277	293	283	284	2.95	cps
Selenium	78-2	543	463	500	502	7.97	cps
Silicon	28-1	20839876	20912975	20869219	20874023	0.18	cps
Silver	107-1	9316	9767	9677	9587	2.49	cps
Silver	109-1	9070	8709	8903	8894	2.03	cps
Sodium	23-2	122119	123406	123171	122899	0.56	cps
Strontium	86-1	322016	326563	326573	325050	0.81	cps
Strontium	88-1	2900699	2992425	2982209	2958445	1.70	cps
Sulfur	34-1	138931	143076	142487	141498	1.58	cps
Terbium	159-1	11320089	11309499	10805600	11145063	2.64	cps
Terbium	159-2	6654419	6561867	6595552	6603946	0.71	cps
Thallium	203-1	5244	5374	5134	5251	2.29	cps
Thallium	205-1	12322	12829	12773	12642	2.20	cps
Tin	118-1	29053	29801	29741	29531	1.41	cps
Titanium	47-1	1674597	1681938	1686913	1681149	0.37	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-14 Instrumnet Name : P7
Client Sample ID : MYE4J2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:52:30 DataFile Name : 046SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	40528	40301	40130	40320	0.50	cps
Vanadium	51-2	869059	866954	866518	867510	0.16	cps
Yttrium	89-1	14485533	14462172	13830894	14259533	2.60	cps
Yttrium	89-2	3422866	3441466	3463653	3442662	0.59	cps
Zinc	66-2	267391	266414	267334	267046	0.21	cps
Zirconium	90-1	304332	308371	308380	307028	0.76	cps
Zirconium	91-1	68281	69376	69996	69218	1.25	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-15 Instrumnet Name : P7
Client Sample ID : MYE4J3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:55:32 DataFile Name : 047SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7534853	7490281	7541076	7522070	0.37	cps
Antimony	121-1	49829	49130	49602	49520	0.72	cps
Arsenic	75-2	71658	71404	71605	71556	0.19	cps
Barium	135-1	1134143	1141089	1148022	1141084	0.61	cps
Barium	137-1	2090138	2133757	2157109	2127001	1.60	cps
Beryllium	9-1	2078	1961	1889	1976	4.82	cps
Bismuth	209-1	6901967	6898033	6813474	6871158	0.73	cps
Bismuth	209-2	4804034	4743055	4783451	4776846	0.65	cps
Bromine	81-1	12482	12299	12185	12322	1.22	cps
Bromine	81-2	453	397	417	422	6.81	cps
Cadmium	108-1	573	560	500	544	7.18	cps
Cadmium	106-1	3851	3964	3754	3856	2.73	cps
Cadmium	111-1	7335	7514	7565	7471	1.62	cps
Calcium	43-1	177633	178253	178287	178058	0.21	cps
Calcium	44-1	2733952	2719322	2751769	2735014	0.59	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	353550	351497	350382	351810	0.46	cps
Cobalt	59-2	349018	351051	348204	349424	0.42	cps
Copper	63-2	1165600	1160363	1156059	1160674	0.41	cps
Dysprosium	156-1	67471	67655	68516	67881	0.82	cps
Dysprosium	156-2	51715	51557	52223	51832	0.67	cps
Erbium	164-1	50113	50748	51300	50721	1.17	cps
Erbium	164-2	30360	31058	30670	30696	1.14	cps
Gadolinium	160-1	61762	62586	62281	62209	0.67	cps
Gadolinium	160-2	40000	40040	40351	40131	0.48	cps
Holmium	165-1	11218118	10863050	10876890	10986020	1.83	cps
Holmium	165-2	6640577	6570158	6585199	6598645	0.56	cps
Indium	115-1	8292812	8249926	8073887	8205541	1.41	cps
Indium	115-2	2562136	2521936	2524803	2536292	0.88	cps
Iron	56-2	375398914	376444768	374710474	375518052	0.23	cps
Iron	57-2	9296053	9439273	9166144	9300490	1.47	cps
Iron	54-2	20476607	20630426	20452874	20519969	0.47	cps
Krypton	83-1	243	253	253	250	2.31	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-15 Instrumnet Name : P7
Client Sample ID : MYE4J3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:55:32 DataFile Name : 047SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	692342	701058	702901	698767	0.81	cps
Lead	207-1	611219	614707	623590	616505	1.03	cps
Lead	208-1	2769432	2793639	2799673	2787581	0.57	cps
Lithium	6-1	1346927	1343839	1358200	1349655	0.56	cps
Magnesium	24-2	6573907	6625608	6627937	6609151	0.46	cps
Manganese	55-2	7099872	7123599	7088497	7103989	0.25	cps
Molybdenum	94-1	92788	93254	94063	93368	0.69	cps
Molybdenum	95-1	8856	9066	8493	8805	3.30	cps
Molybdenum	96-1	29687	30094	30298	30026	1.04	cps
Molybdenum	97-1	5264	5361	5484	5370	2.05	cps
Molybdenum	98-1	13493	13400	13987	13626	2.32	cps
Neodymium	150-1	127544	130124	130642	129437	1.28	cps
Neodymium	150-2	67045	66787	66030	66621	0.79	cps
Nickel	60-2	135055	134191	134812	134686	0.33	cps
Phosphorus	31-2	20267	20404	20470	20380	0.51	cps
Potassium	39-2	3179178	3182255	3193228	3184887	0.23	cps
Rhodium	103-1	7562158	7585686	7555456	7567767	0.21	cps
Rhodium	103-2	4406813	4374407	4371003	4384074	0.45	cps
Scandium	45-1	4671833	4623877	4633701	4643137	0.55	cps
Scandium	45-2	414270	412712	413596	413526	0.19	cps
Selenium	82-1	733	645	691	690	6.36	cps
Selenium	77-2	267	240	213	240	11.11	cps
Selenium	78-2	397	417	387	400	3.82	cps
Silicon	28-1	15282285	15508969	15478358	15423204	0.80	cps
Silver	107-1	5878	5651	5881	5803	2.27	cps
Silver	109-1	5168	5308	5378	5284	2.02	cps
Sodium	23-2	104931	105777	105371	105359	0.40	cps
Strontium	86-1	206535	209540	208197	208090	0.72	cps
Strontium	88-1	1863731	1886165	1885451	1878449	0.68	cps
Sulfur	34-1	145259	143961	142988	144069	0.79	cps
Terbium	159-1	11350997	11140807	11175507	11222437	1.00	cps
Terbium	159-2	6613109	6478835	6692573	6594839	1.64	cps
Thallium	203-1	4111	3744	3767	3874	5.30	cps
Thallium	205-1	9267	9707	9313	9429	2.57	cps
Tin	118-1	26438	27085	26191	26571	1.74	cps
Titanium	47-1	1224913	1234544	1242993	1234150	0.73	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-15 Instrumnet Name : P7
Client Sample ID : MYE4J3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:55:32 DataFile Name : 047SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	30719	30759	30518	30665	0.42	cps
Vanadium	51-2	732060	733969	732467	732832	0.14	cps
Yttrium	89-1	14155010	14085531	13922331	14054291	0.85	cps
Yttrium	89-2	3398657	3329374	3333172	3353734	1.16	cps
Zinc	66-2	225493	225355	223975	224941	0.37	cps
Zirconium	90-1	237767	241914	241958	240546	1.00	cps
Zirconium	91-1	53842	54250	54604	54232	0.70	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-16 Instrumnet Name : P7
Client Sample ID : MYE4J4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:58:40 DataFile Name : 048SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	9825250	9875365	9838923	9846513	0.26	cps
Antimony	121-1	64454	63647	64959	64353	1.03	cps
Arsenic	75-2	74958	75742	74435	75045	0.88	cps
Barium	135-1	1611501	1627296	1631475	1623424	0.65	cps
Barium	137-1	2750871	2815524	2869491	2811962	2.11	cps
Beryllium	9-1	2350	2395	2461	2402	2.33	cps
Bismuth	209-1	6892039	6829354	6076792	6599395	6.87	cps
Bismuth	209-2	4761723	4748591	4695187	4735167	0.74	cps
Bromine	81-1	12325	12862	13116	12768	3.16	cps
Bromine	81-2	520	487	550	519	6.11	cps
Cadmium	108-1	810	763	903	826	8.64	cps
Cadmium	106-1	3937	4254	3817	4003	5.64	cps
Cadmium	111-1	9588	10026	9752	9788	2.26	cps
Calcium	43-1	304059	306970	310729	307252	1.09	cps
Calcium	44-1	4634838	4670377	4709361	4671526	0.80	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	407181	410255	408778	408738	0.38	cps
Cobalt	59-2	332213	334732	330667	332537	0.62	cps
Copper	63-2	1896618	1899679	1915105	1903801	0.52	cps
Dysprosium	156-1	81203	83040	83510	82584	1.48	cps
Dysprosium	156-2	61212	61105	61608	61308	0.43	cps
Erbium	164-1	61926	62251	63584	62587	1.40	cps
Erbium	164-2	37968	38129	38035	38044	0.21	cps
Gadolinium	160-1	73950	76025	75334	75103	1.41	cps
Gadolinium	160-2	50099	47989	49079	49056	2.15	cps
Holmium	165-1	10980822	10804565	9605426	10463604	7.15	cps
Holmium	165-2	6666217	6661184	6580548	6635983	0.72	cps
Indium	115-1	8332808	8119745	7055523	7836025	8.73	cps
Indium	115-2	2543608	2546345	2504540	2531498	0.92	cps
Iron	56-2	414068807	407016194	411273874	410786292	0.86	cps
Iron	57-2	10180595	10002399	10150663	10111219	0.94	cps
Iron	54-2	22381593	22077121	22168368	22209027	0.70	cps
Krypton	83-1	347	303	247	299	16.78	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-16 Instrumnet Name : P7
Client Sample ID : MYE4J4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:58:40 DataFile Name : 048SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1016665	1012868	1023730	1017754	0.54	cps
Lead	207-1	900290	893826	909071	901062	0.85	cps
Lead	208-1	4446083	4392611	4427221	4421972	0.61	cps
Lithium	6-1	1349661	1373070	1188518	1303750	7.71	cps
Magnesium	24-2	8436679	8310453	8337533	8361555	0.79	cps
Manganese	55-2	7956014	7825434	7973487	7918312	1.02	cps
Molybdenum	94-1	109284	111513	113139	111312	1.74	cps
Molybdenum	95-1	10337	10290	10687	10438	2.08	cps
Molybdenum	96-1	33812	34350	35403	34522	2.34	cps
Molybdenum	97-1	6331	6545	6718	6532	2.97	cps
Molybdenum	98-1	16089	16126	16680	16298	2.03	cps
Neodymium	150-1	153505	153300	155116	153974	0.65	cps
Neodymium	150-2	80230	77924	78524	78892	1.52	cps
Nickel	60-2	162897	162224	162887	162669	0.24	cps
Phosphorus	31-2	31225	30497	31025	30916	1.22	cps
Potassium	39-2	3826939	3775732	3767269	3789980	0.85	cps
Rhodium	103-1	7469117	7472696	6634343	7192052	6.72	cps
Rhodium	103-2	4362529	4389829	4345037	4365798	0.52	cps
Scandium	45-1	4694620	4691761	4197417	4527933	6.32	cps
Scandium	45-2	423320	418819	417462	419867	0.73	cps
Selenium	82-1	616	662	629	636	3.73	cps
Selenium	77-2	293	270	267	277	5.25	cps
Selenium	78-2	503	477	493	491	2.74	cps
Silicon	28-1	17510768	17733426	17990654	17744950	1.35	cps
Silver	107-1	10464	10807	10624	10632	1.62	cps
Silver	109-1	9957	9800	10100	9952	1.51	cps
Sodium	23-2	141476	140208	140413	140699	0.48	cps
Strontium	86-1	355990	361136	363835	360320	1.11	cps
Strontium	88-1	3210752	3266478	3263996	3247075	0.97	cps
Sulfur	34-1	150758	153179	154216	152718	1.16	cps
Terbium	159-1	11220816	11186924	9731481	10713073	7.94	cps
Terbium	159-2	6691503	6537175	6489932	6572870	1.60	cps
Thallium	203-1	4324	4377	4354	4352	0.61	cps
Thallium	205-1	10351	10327	10454	10377	0.65	cps
Tin	118-1	30840	30910	31411	31053	1.00	cps
Titanium	47-1	1430038	1409478	1459669	1433062	1.76	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-16 Instrumnet Name : P7
Client Sample ID : MYE4J4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 19:58:40 DataFile Name : 048SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	33281	32747	34558	33529	2.78	cps
Vanadium	51-2	824243	818391	819089	820574	0.39	cps
Yttrium	89-1	14246807	14193369	12641940	13694039	6.66	cps
Yttrium	89-2	3430495	3390783	3330629	3383969	1.49	cps
Zinc	66-2	285112	280307	285358	283592	1.00	cps
Zirconium	90-1	283181	289471	291395	288016	1.49	cps
Zirconium	91-1	62904	65066	65529	64500	2.17	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV002 Instrumnet Name : P7
Client Sample ID : CCV002 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:01:42 DataFile Name : 049CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	10925585	11073000	11028877	11009154	0.69	cps
Antimony	121-1	5028490	4957247	5019419	5001719	0.78	cps
Arsenic	75-2	254512	255920	255066	255166	0.28	cps
Barium	135-1	6838844	6914055	6930527	6894475	0.71	cps
Barium	137-1	11860891	11851291	11974573	11895585	0.58	cps
Beryllium	9-1	732278	733306	739854	735146	0.56	cps
Bismuth	209-1	6541045	6574261	6520670	6545325	0.41	cps
Bismuth	209-2	4473061	4398741	4545091	4472298	1.64	cps
Bromine	81-1	10971	10324	10537	10611	3.11	cps
Bromine	81-2	77	93	93	88	10.96	cps
Cadmium	108-1	91909	92070	92952	92310	0.61	cps
Cadmium	106-1	132531	132100	132605	132412	0.21	cps
Cadmium	111-1	1236882	1247406	1251177	1245155	0.59	cps
Calcium	43-1	6527484	6451340	6466806	6481877	0.62	cps
Calcium	44-1	103077848	103348182	102532355	102986128	0.40	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2020278	2020005	2009792	2016692	0.30	cps
Cobalt	59-2	3229046	3263620	3254421	3249029	0.55	cps
Copper	63-2	22887713	22613163	22834587	22778488	0.64	cps
Dysprosium	156-1	337	287	267	297	12.15	cps
Dysprosium	156-2	393	407	393	398	1.94	cps
Erbium	164-1	347	310	390	349	11.48	cps
Erbium	164-2	210	177	190	192	8.73	cps
Gadolinium	160-1	270	250	217	246	10.97	cps
Gadolinium	160-2	410	433	373	406	7.46	cps
Holmium	165-1	10924090	11074078	11031244	11009804	0.70	cps
Holmium	165-2	6740638	6561199	6701941	6667926	1.42	cps
Indium	115-1	7829444	7920929	7843213	7864529	0.63	cps
Indium	115-2	2453645	2443572	2432692	2443303	0.43	cps
Iron	56-2	437962900	434450767	431519300	434644322	0.74	cps
Iron	57-2	10898242	10789585	10817232	10835020	0.52	cps
Iron	54-2	23700180	24017570	23652214	23789988	0.83	cps
Krypton	83-1	217	213	180	203	9.97	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV002 Instrumnet Name : P7
Client Sample ID : CCV002 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:01:42 DataFile Name : 049CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	20218860	20784409	20846908	20616726	1.68	cps
Lead	207-1	17865363	18463269	18289342	18205991	1.69	cps
Lead	208-1	81375629	83312589	83225408	82637875	1.32	cps
Lithium	6-1	1324546	1399541	1370927	1365004	2.77	cps
Magnesium	24-2	94513269	95397619	94900099	94936995	0.47	cps
Manganese	55-2	14718209	14721196	14648067	14695824	0.28	cps
Molybdenum	94-1	19165436	19002050	19167722	19111736	0.50	cps
Molybdenum	95-1	27944721	27683414	27647357	27758497	0.58	cps
Molybdenum	96-1	30418541	30424678	30057085	30300101	0.69	cps
Molybdenum	97-1	17496013	17327970	17258186	17360723	0.70	cps
Molybdenum	98-1	45130839	44823786	44823276	44925967	0.39	cps
Neodymium	150-1	530	490	470	497	6.15	cps
Neodymium	150-2	167	167	137	157	11.06	cps
Nickel	60-2	779108	776458	781316	778961	0.31	cps
Phosphorus	31-2	113920	114742	115166	114609	0.55	cps
Potassium	39-2	55174561	55900983	55285263	55453602	0.71	cps
Rhodium	103-1	6974839	7005438	6911116	6963797	0.69	cps
Rhodium	103-2	4104512	4044859	4077072	4075481	0.73	cps
Scandium	45-1	4319956	4339379	4302396	4320577	0.43	cps
Scandium	45-2	392980	394291	394237	393836	0.19	cps
Selenium	82-1	76910	77248	76981	77046	0.23	cps
Selenium	77-2	7125	6848	7342	7105	3.48	cps
Selenium	78-2	23976	23692	24016	23895	0.74	cps
Silicon	28-1	41606841	40990086	41636038	41410988	0.88	cps
Silver	107-1	6387284	6355059	6337997	6360113	0.39	cps
Silver	109-1	6051738	5984027	5971036	6002267	0.72	cps
Sodium	23-2	156127028	157015834	156692221	156611694	0.29	cps
Strontium	86-1	40109179	40128800	40246646	40161542	0.19	cps
Strontium	88-1	352748448	354720755	355792781	354420661	0.44	cps
Sulfur	34-1	540228	541039	536477	539248	0.45	cps
Terbium	159-1	11085884	11212648	10995942	11098158	0.98	cps
Terbium	159-2	6581939	6606537	6557060	6581845	0.38	cps
Thallium	203-1	5050369	5183365	5220284	5151339	1.73	cps
Thallium	205-1	12131771	12274664	12546513	12317649	1.71	cps
Tin	118-1	4240976	4211632	4252297	4234969	0.50	cps
Titanium	47-1	6630071	6573725	6565508	6589768	0.53	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV002 Instrumnet Name : P7
Client Sample ID : CCV002 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:01:42 DataFile Name : 049CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	16508014	16923675	16742530	16724739	1.25	cps
Vanadium	51-2	1797555	1808277	1770159	1791997	1.10	cps
Yttrium	89-1	13214189	13157458	13158860	13176836	0.25	cps
Yttrium	89-2	3202365	3151992	3135885	3163414	1.10	cps
Zinc	66-2	3573959	3557546	3608160	3579889	0.72	cps
Zirconium	90-1	8805638	8724480	8758309	8762809	0.47	cps
Zirconium	91-1	1969099	1939751	1992774	1967208	1.35	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB002 Instrumnet Name : P7
Client Sample ID : CCB002 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:04:26 DataFile Name : 050CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	3077	3020	3074	3057	1.04	cps
Antimony	121-1	2964	3087	2887	2979	3.39	cps
Arsenic	75-2	23	33	43	33	30.00	cps
Barium	135-1	753	833	867	818	7.12	cps
Barium	137-1	1353	1440	1553	1449	6.92	cps
Beryllium	9-1	128	78	89	98	26.75	cps
Bismuth	209-1	7049690	6810956	7135519	6998722	2.40	cps
Bismuth	209-2	4957939	4918167	4953027	4943044	0.44	cps
Bromine	81-1	10627	10230	10537	10465	1.99	cps
Bromine	81-2	90	40	43	58	48.39	cps
Cadmium	108-1	17	17	23	19	20.36	cps
Cadmium	106-1	2990	2604	2800	2798	6.91	cps
Cadmium	111-1	2162	1934	2084	2060	5.61	cps
Calcium	43-1	730	647	720	699	6.51	cps
Calcium	44-1	16509	17794	18021	17442	4.67	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	833	753	803	797	5.07	cps
Cobalt	59-2	413	417	413	414	0.47	cps
Copper	63-2	2504	2344	2557	2468	4.50	cps
Dysprosium	156-1	43	20	20	28	48.49	cps
Dysprosium	156-2	20	7	17	14	48.02	cps
Erbium	164-1	90	83	77	83	8.00	cps
Erbium	164-2	47	60	57	54	12.74	cps
Gadolinium	160-1	57	37	73	56	33.04	cps
Gadolinium	160-2	247	310	283	280	11.36	cps
Holmium	165-1	11066167	10464252	11160376	10896932	3.47	cps
Holmium	165-2	6802345	6734996	6646848	6728063	1.16	cps
Indium	115-1	8535242	8133727	8619200	8429389	3.08	cps
Indium	115-2	2658579	2626155	2658779	2647838	0.71	cps
Iron	56-2	113108	112027	108857	111331	1.98	cps
Iron	57-2	3574	3600	3684	3619	1.59	cps
Iron	54-2	6985	6822	6732	6846	1.88	cps
Krypton	83-1	213	137	240	197	27.28	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB002 Instrumnet Name : P7
Client Sample ID : CCB002 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:04:26 DataFile Name : 050CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	2104	2037	2027	2056	2.03	cps
Lead	207-1	1723	1787	1740	1750	1.88	cps
Lead	208-1	8485	8134	8091	8237	2.62	cps
Lithium	6-1	1422376	1407809	1389454	1406546	1.17	cps
Magnesium	24-2	6558	6695	6451	6568	1.86	cps
Manganese	55-2	2654	2627	2547	2609	2.13	cps
Molybdenum	94-1	1370	1620	1437	1476	8.77	cps
Molybdenum	95-1	1957	1860	1930	1916	2.61	cps
Molybdenum	96-1	2027	2304	1913	2081	9.64	cps
Molybdenum	97-1	1170	1347	1183	1233	7.98	cps
Molybdenum	98-1	3097	3120	3114	3110	0.39	cps
Neodymium	150-1	43	47	33	41	16.89	cps
Neodymium	150-2	23	13	13	17	34.65	cps
Nickel	60-2	157	157	210	174	17.65	cps
Phosphorus	31-2	153	170	143	156	8.66	cps
Potassium	39-2	52795	53714	53296	53268	0.86	cps
Rhodium	103-1	7795594	7567036	7966325	7776318	2.58	cps
Rhodium	103-2	4552401	4542716	4556647	4550588	0.16	cps
Scandium	45-1	4500148	4404754	4589389	4498097	2.05	cps
Scandium	45-2	401371	401626	402171	401722	0.10	cps
Selenium	82-1	358	411	414	395	8.00	cps
Selenium	77-2	0	0	0	0	0.00	cps
Selenium	78-2	230	203	227	220	6.60	cps
Silicon	28-1	230261	228842	225374	228159	1.10	cps
Silver	107-1	963	890	700	851	15.97	cps
Silver	109-1	720	720	693	711	2.16	cps
Sodium	23-2	37502	37863	36724	37363	1.56	cps
Strontium	86-1	2060	2197	2087	2115	3.43	cps
Strontium	88-1	15568	16296	15034	15633	4.05	cps
Sulfur	34-1	170617	170304	170271	170397	0.11	cps
Terbium	159-1	11283700	10879976	11507804	11223826	2.83	cps
Terbium	159-2	6664276	6717997	6596097	6659457	0.92	cps
Thallium	203-1	813	750	753	772	4.62	cps
Thallium	205-1	1933	1870	1693	1832	6.79	cps
Tin	118-1	2180	2194	2180	2185	0.35	cps
Titanium	47-1	807	700	873	793	11.02	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB002 Instrumnet Name : P7
Client Sample ID : CCB002 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:04:26 DataFile Name : 050CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	677	653	603	644	5.81	cps
Vanadium	51-2	383	283	300	322	16.63	cps
Yttrium	89-1	13507902	13162626	13721735	13464088	2.10	cps
Yttrium	89-2	3271371	3283500	3287363	3280745	0.25	cps
Zinc	66-2	803	910	777	830	8.50	cps
Zirconium	90-1	1517	1490	1490	1499	1.03	cps
Zirconium	91-1	300	347	330	326	7.26	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-17 Instrumnet Name : P7
Client Sample ID : MYE4J5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:07:41 DataFile Name : 051SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	10572416	10680066	10660456	10637646	0.54	cps
Antimony	121-1	14951	15005	15308	15088	1.28	cps
Arsenic	75-2	41314	42056	41508	41626	0.92	cps
Barium	135-1	1164602	1171776	1173426	1169935	0.40	cps
Barium	137-1	2173872	2170152	2226398	2190140	1.44	cps
Beryllium	9-1	2556	2134	2467	2385	9.33	cps
Bismuth	209-1	6481044	7069233	6707167	6752481	4.39	cps
Bismuth	209-2	4765240	4752732	4774855	4764276	0.23	cps
Bromine	81-1	17737	18525	18669	18310	2.74	cps
Bromine	81-2	1503	1407	1307	1406	7.00	cps
Cadmium	108-1	367	377	417	387	6.84	cps
Cadmium	106-1	3701	3781	3751	3744	1.08	cps
Cadmium	111-1	5554	5576	5601	5577	0.42	cps
Calcium	43-1	877527	902199	911602	897110	1.96	cps
Calcium	44-1	13459763	13778268	13767354	13668461	1.32	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	417647	414216	414792	415552	0.44	cps
Cobalt	59-2	298059	299262	299542	298954	0.26	cps
Copper	63-2	1069897	1060177	1057717	1062597	0.61	cps
Dysprosium	156-1	84904	84211	85233	84783	0.62	cps
Dysprosium	156-2	63513	62836	63355	63235	0.56	cps
Erbium	164-1	70855	70678	72303	71278	1.25	cps
Erbium	164-2	43260	43293	43267	43273	0.04	cps
Gadolinium	160-1	79079	81234	81301	80538	1.57	cps
Gadolinium	160-2	52873	52371	52357	52534	0.56	cps
Holmium	165-1	10514388	11630927	10991554	11045623	5.07	cps
Holmium	165-2	6810466	6796607	6740810	6782628	0.54	cps
Indium	115-1	8052285	8912885	8631594	8532255	5.14	cps
Indium	115-2	2641307	2678548	2679106	2666320	0.81	cps
Iron	56-2	336581295	344021555	341137528	340580126	1.10	cps
Iron	57-2	8360802	8395227	8320259	8358763	0.45	cps
Iron	54-2	18520671	18809505	18456228	18595468	1.01	cps
Krypton	83-1	307	263	213	261	17.89	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-17 Instrumnet Name : P7
Client Sample ID : MYE4J5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:07:41 DataFile Name : 051SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	392306	394149	394707	393721	0.32	cps
Lead	207-1	337840	334323	334296	335486	0.61	cps
Lead	208-1	1538491	1543294	1547190	1542992	0.28	cps
Lithium	6-1	1361728	1409829	1394552	1388703	1.77	cps
Magnesium	24-2	8399774	8302711	8370560	8357682	0.60	cps
Manganese	55-2	7462297	7554355	7456958	7491203	0.73	cps
Molybdenum	94-1	69012	71912	72150	71024	2.46	cps
Molybdenum	95-1	12515	12632	12836	12661	1.28	cps
Molybdenum	96-1	28441	29042	29603	29029	2.00	cps
Molybdenum	97-1	7782	7669	7712	7721	0.74	cps
Molybdenum	98-1	19052	19463	19450	19322	1.21	cps
Neodymium	150-1	149712	149733	150062	149836	0.13	cps
Neodymium	150-2	76711	76519	77059	76763	0.36	cps
Nickel	60-2	129046	128226	127547	128273	0.59	cps
Phosphorus	31-2	40891	40390	39772	40351	1.39	cps
Potassium	39-2	6800747	6771121	6862725	6811531	0.69	cps
Rhodium	103-1	7345214	8183006	7908584	7812268	5.47	cps
Rhodium	103-2	4554085	4578526	4600106	4577572	0.50	cps
Scandium	45-1	4796334	5255096	5158301	5069910	4.77	cps
Scandium	45-2	455322	452631	455761	454571	0.37	cps
Selenium	82-1	608	742	790	713	13.22	cps
Selenium	77-2	347	317	330	331	4.54	cps
Selenium	78-2	510	550	457	506	9.26	cps
Silicon	28-1	26402707	26962250	27550669	26971875	2.13	cps
Silver	107-1	7679	7315	7115	7370	3.88	cps
Silver	109-1	7309	6548	7002	6953	5.50	cps
Sodium	23-2	327024	323474	328961	326486	0.85	cps
Strontium	86-1	511555	524329	531181	522355	1.91	cps
Strontium	88-1	4685660	4787451	4782636	4751916	1.21	cps
Sulfur	34-1	187378	184533	180057	183989	2.01	cps
Terbium	159-1	10746096	11632764	11415797	11264885	4.10	cps
Terbium	159-2	6837851	6742013	6793312	6791059	0.71	cps
Thallium	203-1	4314	4067	4117	4166	3.13	cps
Thallium	205-1	10050	9570	9684	9768	2.57	cps
Tin	118-1	23673	24043	24972	24229	2.76	cps
Titanium	47-1	1637848	1685312	1711962	1678374	2.24	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-17 Instrumnet Name : P7
Client Sample ID : MYE4J5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:07:41 DataFile Name : 051SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	39104	39281	38710	39032	0.75	cps
Vanadium	51-2	676032	675046	676746	675941	0.13	cps
Yttrium	89-1	14324013	15941497	15407053	15224188	5.41	cps
Yttrium	89-2	3695826	3561074	3571344	3609415	2.08	cps
Zinc	66-2	190101	191120	191665	190962	0.42	cps
Zirconium	90-1	170869	176644	177566	175026	2.07	cps
Zirconium	91-1	39444	39002	39971	39472	1.23	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-18 Instrumnet Name : P7
Client Sample ID : MYE4J6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:10:47 DataFile Name : 052SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	10078273	9999384	10006860	10028172	0.43	cps
Antimony	121-1	20645	20902	20658	20735	0.70	cps
Arsenic	75-2	47472	48171	47135	47593	1.11	cps
Barium	135-1	1038913	1059131	1068763	1055602	1.44	cps
Barium	137-1	1970151	1994057	1966661	1976956	0.75	cps
Beryllium	9-1	1928	1978	2011	1972	2.13	cps
Bismuth	209-1	6824379	6217127	6643149	6561552	4.75	cps
Bismuth	209-2	4772837	4869737	4760724	4801099	1.24	cps
Bromine	81-1	15185	14701	14861	14915	1.65	cps
Bromine	81-2	967	900	833	900	7.41	cps
Cadmium	108-1	447	393	493	444	11.26	cps
Cadmium	106-1	3814	3587	3354	3585	6.42	cps
Cadmium	111-1	6941	7160	6686	6929	3.42	cps
Calcium	43-1	486366	492085	494681	491044	0.87	cps
Calcium	44-1	7457537	7468452	7495832	7473940	0.26	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	423664	422138	421365	422389	0.28	cps
Cobalt	59-2	304426	305124	300717	303422	0.78	cps
Copper	63-2	1441193	1513903	1436287	1463795	2.97	cps
Dysprosium	156-1	79651	81086	81052	80596	1.02	cps
Dysprosium	156-2	60560	59348	60141	60016	1.03	cps
Erbium	164-1	65610	66768	67288	66555	1.29	cps
Erbium	164-2	40823	40646	40629	40699	0.26	cps
Gadolinium	160-1	76612	75726	77499	76612	1.16	cps
Gadolinium	160-2	49815	50471	49637	49974	0.88	cps
Holmium	165-1	11010979	10061489	10973654	10682041	5.03	cps
Holmium	165-2	6830463	6724130	6777468	6777354	0.78	cps
Indium	115-1	8514586	7647391	8301054	8154344	5.54	cps
Indium	115-2	2621037	2602695	2601480	2608404	0.42	cps
Iron	56-2	339019862	337378722	337020648	337806411	0.32	cps
Iron	57-2	8376320	8351218	8253572	8327037	0.78	cps
Iron	54-2	18422456	18168764	18247993	18279738	0.71	cps
Krypton	83-1	243	300	290	278	10.89	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-18 Instrumnet Name : P7
Client Sample ID : MYE4J6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:10:47 DataFile Name : 052SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	558856	567822	573063	566580	1.27	cps
Lead	207-1	492036	491097	495119	492751	0.43	cps
Lead	208-1	2221421	2236238	2254916	2237525	0.75	cps
Lithium	6-1	1368070	1275832	1349597	1331166	3.67	cps
Magnesium	24-2	8357118	8299158	8269202	8308492	0.54	cps
Manganese	55-2	6501593	6443369	6430198	6458387	0.59	cps
Molybdenum	94-1	61863	63446	61642	62317	1.58	cps
Molybdenum	95-1	13286	13707	13670	13554	1.72	cps
Molybdenum	96-1	27850	27997	28301	28049	0.82	cps
Molybdenum	97-1	8186	8459	8166	8270	1.98	cps
Molybdenum	98-1	20972	21679	20892	21181	2.05	cps
Neodymium	150-1	138236	140903	142077	140405	1.40	cps
Neodymium	150-2	73647	74361	72117	73375	1.56	cps
Nickel	60-2	130704	131608	131087	131133	0.35	cps
Phosphorus	31-2	34977	35678	35665	35440	1.13	cps
Potassium	39-2	7467040	7364962	7428006	7420003	0.69	cps
Rhodium	103-1	7840615	7123252	7754557	7572808	5.17	cps
Rhodium	103-2	4500749	4430955	4446850	4459518	0.82	cps
Scandium	45-1	5047198	4657272	4941687	4882052	4.13	cps
Scandium	45-2	442677	445903	444651	444410	0.37	cps
Selenium	82-1	817	628	703	716	13.28	cps
Selenium	77-2	327	257	220	268	20.24	cps
Selenium	78-2	497	523	500	507	2.87	cps
Silicon	28-1	22687917	22979502	23275911	22981110	1.28	cps
Silver	107-1	9143	9323	8996	9154	1.79	cps
Silver	109-1	8449	8916	8666	8677	2.69	cps
Sodium	23-2	786842	784371	778104	783106	0.58	cps
Strontium	86-1	330855	337144	339467	335822	1.33	cps
Strontium	88-1	2979324	3041928	3068567	3029940	1.51	cps
Sulfur	34-1	152883	149227	147695	149935	1.78	cps
Terbium	159-1	11421992	10388206	11113092	10974430	4.84	cps
Terbium	159-2	6709473	6702225	6627699	6679799	0.68	cps
Thallium	203-1	3424	3390	3464	3426	1.07	cps
Thallium	205-1	7859	8259	8303	8140	3.00	cps
Tin	118-1	23880	23810	23800	23830	0.18	cps
Titanium	47-1	1614194	1644236	1646110	1634847	1.10	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-18 Instrumnet Name : P7
Client Sample ID : MYE4J6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:10:47 DataFile Name : 052SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	35337	36065	36456	35953	1.58	cps
Vanadium	51-2	703405	700603	698622	700877	0.34	cps
Yttrium	89-1	15056071	13943117	14771291	14590160	3.96	cps
Yttrium	89-2	3594509	3544886	3579251	3572882	0.71	cps
Zinc	66-2	219943	217701	217752	218465	0.59	cps
Zirconium	90-1	146987	149020	149017	148341	0.79	cps
Zirconium	91-1	32579	33679	33943	33400	2.16	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-19 Instrumnet Name : P7
Client Sample ID : MYE4J7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:13:51 DataFile Name : 053SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	6052438	6135127	6050940	6079502	0.79	cps
Antimony	121-1	11028	10964	11275	11089	1.48	cps
Arsenic	75-2	21479	21472	21342	21431	0.36	cps
Barium	135-1	662345	666219	671441	666669	0.68	cps
Barium	137-1	1145749	1158117	1164961	1156276	0.84	cps
Beryllium	9-1	1456	1639	1372	1489	9.16	cps
Bismuth	209-1	6823152	6808084	6779863	6803699	0.32	cps
Bismuth	209-2	4756749	4847136	4764325	4789403	1.05	cps
Bromine	81-1	13743	13690	13386	13606	1.42	cps
Bromine	81-2	507	533	503	514	3.20	cps
Cadmium	108-1	163	200	160	174	12.72	cps
Cadmium	106-1	3364	3240	3117	3240	3.81	cps
Cadmium	111-1	3270	3132	3017	3140	4.05	cps
Calcium	43-1	466845	469527	469427	468600	0.32	cps
Calcium	44-1	7123147	7136992	7104384	7121508	0.23	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	316452	319047	317471	317657	0.41	cps
Cobalt	59-2	212642	214035	214752	213810	0.50	cps
Copper	63-2	376481	377993	375970	376814	0.28	cps
Dysprosium	156-1	58264	58424	58575	58421	0.27	cps
Dysprosium	156-2	43317	43644	43631	43531	0.43	cps
Erbium	164-1	49256	49333	49889	49493	0.70	cps
Erbium	164-2	30343	30590	30183	30372	0.68	cps
Gadolinium	160-1	54064	55158	54951	54724	1.06	cps
Gadolinium	160-2	35993	36945	35886	36275	1.61	cps
Holmium	165-1	11205269	11039898	11114834	11120000	0.74	cps
Holmium	165-2	6804595	6778059	6732060	6771571	0.54	cps
Indium	115-1	8613998	8468193	8382558	8488250	1.38	cps
Indium	115-2	2599016	2626354	2598123	2607831	0.62	cps
Iron	56-2	265235363	262372563	264702909	264103612	0.58	cps
Iron	57-2	6524831	6557846	6533894	6538857	0.26	cps
Iron	54-2	14477223	14327421	14538836	14447826	0.75	cps
Krypton	83-1	273	200	257	243	15.80	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-19 Instrumnet Name : P7
Client Sample ID : MYE4J7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:13:51 DataFile Name : 053SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	219987	223321	223947	222418	0.96	cps
Lead	207-1	183519	187564	188736	186606	1.47	cps
Lead	208-1	849619	858372	858063	855351	0.58	cps
Lithium	6-1	1393372	1371684	1339466	1368174	1.98	cps
Magnesium	24-2	5364622	5295341	5299076	5319680	0.73	cps
Manganese	55-2	3959249	3941961	4038537	3979916	1.29	cps
Molybdenum	94-1	53450	53802	53577	53610	0.33	cps
Molybdenum	95-1	6015	5788	6238	6014	3.74	cps
Molybdenum	96-1	18181	19136	18642	18653	2.56	cps
Molybdenum	97-1	3771	3787	3734	3764	0.73	cps
Molybdenum	98-1	9220	9123	9220	9187	0.61	cps
Neodymium	150-1	101348	103466	103338	102718	1.16	cps
Neodymium	150-2	52594	54050	53370	53338	1.37	cps
Nickel	60-2	94727	93378	94100	94068	0.72	cps
Phosphorus	31-2	24667	25041	25017	24908	0.84	cps
Potassium	39-2	3151954	3113663	3106699	3124106	0.78	cps
Rhodium	103-1	7898829	7853914	7823649	7858797	0.48	cps
Rhodium	103-2	4485098	4468325	4451026	4468150	0.38	cps
Scandium	45-1	4858962	4907198	4781072	4849077	1.31	cps
Scandium	45-2	426843	430126	428168	428379	0.39	cps
Selenium	82-1	528	664	561	584	12.18	cps
Selenium	77-2	223	167	193	194	14.58	cps
Selenium	78-2	440	430	417	429	2.73	cps
Silicon	28-1	17758317	18010612	17788687	17852539	0.77	cps
Silver	107-1	1897	1997	1994	1962	2.89	cps
Silver	109-1	1690	1647	1710	1682	1.92	cps
Sodium	23-2	230685	229471	230283	230147	0.27	cps
Strontium	86-1	228232	229789	232535	230185	0.95	cps
Strontium	88-1	2084511	2111976	2108049	2101512	0.71	cps
Sulfur	34-1	144431	145303	144731	144822	0.31	cps
Terbium	159-1	11276696	11175845	11329347	11260629	0.69	cps
Terbium	159-2	6722241	6739873	6682959	6715024	0.43	cps
Thallium	203-1	2330	2207	2350	2296	3.38	cps
Thallium	205-1	5424	5251	5451	5376	2.02	cps
Tin	118-1	18906	19851	19286	19348	2.46	cps
Titanium	47-1	1655715	1681824	1667106	1668215	0.78	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-19 Instrumnet Name : P7
Client Sample ID : MYE4J7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:13:51 DataFile Name : 053SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	24880	25274	25007	25054	0.80	cps
Vanadium	51-2	684074	685238	683947	684420	0.10	cps
Yttrium	89-1	14778073	14717136	14677146	14724118	0.35	cps
Yttrium	89-2	3478374	3474951	3447025	3466783	0.50	cps
Zinc	66-2	102267	102089	100616	101657	0.89	cps
Zirconium	90-1	135875	137187	137271	136777	0.57	cps
Zirconium	91-1	30769	30832	30926	30842	0.26	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-20 Instrumnet Name : P7
Client Sample ID : MYE4J8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:16:59 DataFile Name : 054SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	6524860	6379556	6382648	6429021	1.29	cps
Antimony	121-1	15128	15155	15218	15167	0.30	cps
Arsenic	75-2	27288	26858	27505	27217	1.21	cps
Barium	135-1	778604	782852	790684	784047	0.78	cps
Barium	137-1	1353523	1363953	1353664	1357047	0.44	cps
Beryllium	9-1	1422	1472	1533	1476	3.77	cps
Bismuth	209-1	6907253	6993977	6817116	6906115	1.28	cps
Bismuth	209-2	4656310	4787634	4710734	4718226	1.40	cps
Bromine	81-1	17157	17160	18355	17557	3.93	cps
Bromine	81-2	1140	1187	1240	1189	4.21	cps
Cadmium	108-1	247	327	263	279	15.13	cps
Cadmium	106-1	3514	3257	3210	3327	4.91	cps
Cadmium	111-1	4691	4502	4587	4593	2.06	cps
Calcium	43-1	668396	678206	678925	675176	0.87	cps
Calcium	44-1	10267053	10270497	10251437	10262996	0.10	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	276400	276198	275282	275960	0.22	cps
Cobalt	59-2	193450	193467	193491	193469	0.01	cps
Copper	63-2	697188	692560	695290	695013	0.33	cps
Dysprosium	156-1	53940	54314	54057	54104	0.35	cps
Dysprosium	156-2	40712	40288	41120	40707	1.02	cps
Erbium	164-1	45507	46330	46266	46034	0.99	cps
Erbium	164-2	27581	27698	28325	27868	1.44	cps
Gadolinium	160-1	51692	50952	52725	51790	1.72	cps
Gadolinium	160-2	33794	33497	33493	33595	0.51	cps
Holmium	165-1	10970013	10900814	10935319	10935382	0.32	cps
Holmium	165-2	6666178	6674928	6690827	6677311	0.19	cps
Indium	115-1	8526019	8416227	8381443	8441230	0.89	cps
Indium	115-2	2618769	2584564	2613373	2605569	0.71	cps
Iron	56-2	220874257	222582110	220837537	221431301	0.45	cps
Iron	57-2	5471202	5484865	5487977	5481348	0.16	cps
Iron	54-2	12081176	12023214	11976328	12026906	0.44	cps
Krypton	83-1	283	277	263	274	3.71	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-20 Instrumnet Name : P7
Client Sample ID : MYE4J8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:16:59 DataFile Name : 054SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	316274	318962	318843	318026	0.48	cps
Lead	207-1	266275	270825	271839	269646	1.10	cps
Lead	208-1	1234984	1244889	1244251	1241375	0.45	cps
Lithium	6-1	1346191	1326653	1386078	1352974	2.24	cps
Magnesium	24-2	6073648	6014361	5991289	6026433	0.70	cps
Manganese	55-2	4636414	4542623	4575017	4584685	1.04	cps
Molybdenum	94-1	46500	47330	47988	47273	1.58	cps
Molybdenum	95-1	8643	8796	8686	8708	0.91	cps
Molybdenum	96-1	19677	19834	20057	19856	0.96	cps
Molybdenum	97-1	5321	5564	5604	5497	2.79	cps
Molybdenum	98-1	13573	13813	13777	13721	0.94	cps
Neodymium	150-1	96511	96866	97252	96876	0.38	cps
Neodymium	150-2	50513	49286	50805	50201	1.61	cps
Nickel	60-2	96246	96334	95246	95942	0.63	cps
Phosphorus	31-2	31072	31112	30985	31056	0.21	cps
Potassium	39-2	4033524	3992251	3979928	4001901	0.70	cps
Rhodium	103-1	7758469	7775523	7743802	7759264	0.20	cps
Rhodium	103-2	4500275	4540576	4534821	4525224	0.48	cps
Scandium	45-1	4841907	4801925	4774570	4806134	0.70	cps
Scandium	45-2	428995	428235	421485	426238	0.97	cps
Selenium	82-1	653	612	633	633	3.18	cps
Selenium	77-2	153	217	177	182	17.58	cps
Selenium	78-2	463	363	360	396	14.84	cps
Silicon	28-1	22713527	22740539	22923092	22792386	0.50	cps
Silver	107-1	4401	4594	4747	4581	3.79	cps
Silver	109-1	4081	4334	4267	4227	3.11	cps
Sodium	23-2	304458	301777	302173	302803	0.48	cps
Strontium	86-1	470540	471156	478543	473413	0.94	cps
Strontium	88-1	4270948	4289552	4284603	4281701	0.23	cps
Sulfur	34-1	178398	181476	180849	180241	0.90	cps
Terbium	159-1	11130029	11217399	11156992	11168140	0.40	cps
Terbium	159-2	6732888	6721621	6670191	6708233	0.50	cps
Thallium	203-1	2587	2307	2344	2412	6.31	cps
Thallium	205-1	5865	5848	5778	5830	0.79	cps
Tin	118-1	21299	21486	22554	21780	3.11	cps
Titanium	47-1	1228087	1243521	1251901	1241169	0.97	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-20 Instrumnet Name : P7
Client Sample ID : MYE4J8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:16:59 DataFile Name : 054SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	25732	24736	25007	25158	2.05	cps
Vanadium	51-2	463005	458871	462916	461597	0.51	cps
Yttrium	89-1	14458781	14635772	14389726	14494760	0.88	cps
Yttrium	89-2	3478423	3456839	3423029	3452763	0.81	cps
Zinc	66-2	160455	159273	160074	159934	0.38	cps
Zirconium	90-1	113917	116212	115873	115334	1.07	cps
Zirconium	91-1	25385	25422	26454	25754	2.36	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-21 Instrumnet Name : P7
Client Sample ID : MYE4J9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:20:06 DataFile Name : 055SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	10877255	10748388	10787390	10804344	0.61	cps
Antimony	121-1	13276	13099	13290	13222	0.80	cps
Arsenic	75-2	28404	28644	28244	28431	0.71	cps
Barium	135-1	1204869	1214331	1221033	1213411	0.67	cps
Barium	137-1	2231481	2259756	2274529	2255255	0.97	cps
Beryllium	9-1	2534	2517	2534	2528	0.38	cps
Bismuth	209-1	6795035	6862591	6785565	6814397	0.62	cps
Bismuth	209-2	4738851	4778019	4819102	4778657	0.84	cps
Bromine	81-1	16005	16102	16079	16062	0.31	cps
Bromine	81-2	1050	1080	990	1040	4.41	cps
Cadmium	108-1	360	343	377	360	4.63	cps
Cadmium	106-1	3767	3684	3410	3620	5.15	cps
Cadmium	111-1	4455	4525	4197	4393	3.93	cps
Calcium	43-1	658663	662742	660708	660704	0.31	cps
Calcium	44-1	10001390	10022233	10113142	10045589	0.59	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	396504	396558	392427	395163	0.60	cps
Cobalt	59-2	287239	289949	285540	287576	0.77	cps
Copper	63-2	704289	697393	695511	699064	0.66	cps
Dysprosium	156-1	91184	92619	92935	92246	1.01	cps
Dysprosium	156-2	69457	68821	67990	68756	1.07	cps
Erbium	164-1	76850	77192	78170	77404	0.89	cps
Erbium	164-2	47046	46504	46922	46824	0.61	cps
Gadolinium	160-1	87159	88209	88550	87973	0.82	cps
Gadolinium	160-2	56487	56547	55392	56142	1.16	cps
Holmium	165-1	11022284	11308248	11147147	11159226	1.28	cps
Holmium	165-2	6788182	6742240	6806825	6779082	0.49	cps
Indium	115-1	8389112	8422708	8374863	8395561	0.29	cps
Indium	115-2	2627010	2588271	2653459	2622913	1.25	cps
Iron	56-2	309034202	309017409	310208122	309419911	0.22	cps
Iron	57-2	7598779	7728903	7637037	7654906	0.87	cps
Iron	54-2	16974771	16929650	16686455	16863625	0.92	cps
Krypton	83-1	363	260	283	302	17.93	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-21 Instrumnet Name : P7
Client Sample ID : MYE4J9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:20:06 DataFile Name : 055SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	318546	325312	324938	322932	1.18	cps
Lead	207-1	269911	273708	270770	271463	0.73	cps
Lead	208-1	1241797	1260618	1256380	1252932	0.79	cps
Lithium	6-1	1376860	1362096	1372491	1370482	0.55	cps
Magnesium	24-2	8026678	7997074	7988134	8003962	0.25	cps
Manganese	55-2	6637931	6706540	6714696	6686389	0.63	cps
Molybdenum	94-1	88519	87787	89689	88665	1.08	cps
Molybdenum	95-1	7098	7249	7075	7141	1.32	cps
Molybdenum	96-1	25923	26694	26234	26284	1.48	cps
Molybdenum	97-1	4354	4327	4391	4357	0.73	cps
Molybdenum	98-1	11168	10791	11351	11103	2.57	cps
Neodymium	150-1	161490	163277	165212	163326	1.14	cps
Neodymium	150-2	84502	82831	82989	83441	1.11	cps
Nickel	60-2	130630	129282	130328	130080	0.54	cps
Phosphorus	31-2	35328	35364	35020	35237	0.54	cps
Potassium	39-2	5908339	5908487	5808110	5874979	0.99	cps
Rhodium	103-1	7696781	7624285	7662130	7661065	0.47	cps
Rhodium	103-2	4434692	4473642	4511135	4473156	0.85	cps
Scandium	45-1	4826085	4874504	4819975	4840188	0.62	cps
Scandium	45-2	425430	429735	436689	430618	1.32	cps
Selenium	82-1	587	682	661	643	7.73	cps
Selenium	77-2	250	287	343	293	16.03	cps
Selenium	78-2	483	513	407	468	11.76	cps
Silicon	28-1	19353118	19359903	19305643	19339555	0.15	cps
Silver	107-1	3674	3771	3964	3803	3.88	cps
Silver	109-1	3784	3784	3460	3676	5.08	cps
Sodium	23-2	230024	228521	229035	229193	0.33	cps
Strontium	86-1	351029	355517	355209	353918	0.71	cps
Strontium	88-1	3199395	3211876	3177755	3196342	0.54	cps
Sulfur	34-1	150944	149931	148692	149856	0.75	cps
Terbium	159-1	11273561	11342867	11266200	11294209	0.37	cps
Terbium	159-2	6613372	6728625	6915456	6752485	2.26	cps
Thallium	203-1	3687	3974	3747	3803	3.98	cps
Thallium	205-1	9133	8883	9133	9050	1.60	cps
Tin	118-1	23927	24528	24571	24342	1.48	cps
Titanium	47-1	1780556	1781045	1775052	1778884	0.19	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4484-21

Instrumnet Name :

P7

Client Sample ID :

MYE4J9

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 20:20:06

DataFile Name :

055SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	36089	37596	37359	37015	2.19	cps
Vanadium	51-2	611019	606146	606547	607904	0.45	cps
Yttrium	89-1	14815665	14797765	14819920	14811117	0.08	cps
Yttrium	89-2	3509348	3525960	3591135	3542147	1.22	cps
Zinc	66-2	160229	159098	158000	159109	0.70	cps
Zirconium	90-1	225068	229075	231191	228445	1.36	cps
Zirconium	91-1	51132	50854	51473	51153	0.61	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-22 Instrumnet Name : P7
Client Sample ID : MYE4L3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:23:13 DataFile Name : 056SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8222017	8187032	8089813	8166288	0.84	cps
Antimony	121-1	15725	15922	15635	15761	0.93	cps
Arsenic	75-2	27953	28143	28210	28102	0.47	cps
Barium	135-1	832045	839451	845441	838979	0.80	cps
Barium	137-1	1444791	1447234	1457613	1449879	0.47	cps
Beryllium	9-1	1672	1706	1967	1782	9.05	cps
Bismuth	209-1	6882612	6872539	6803318	6852823	0.63	cps
Bismuth	209-2	4770797	4755417	4666924	4731046	1.18	cps
Bromine	81-1	15678	16166	16262	16035	1.95	cps
Bromine	81-2	1050	1153	1080	1094	4.86	cps
Cadmium	108-1	227	283	247	252	11.39	cps
Cadmium	106-1	3284	3434	3250	3323	2.94	cps
Cadmium	111-1	4170	4232	3999	4134	2.92	cps
Calcium	43-1	501845	505544	507600	504996	0.58	cps
Calcium	44-1	7651323	7773962	7719795	7715027	0.80	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	306071	305048	304217	305112	0.30	cps
Cobalt	59-2	228063	227651	226693	227469	0.31	cps
Copper	63-2	715534	708592	707067	710398	0.64	cps
Dysprosium	156-1	67401	68791	69424	68538	1.51	cps
Dysprosium	156-2	50420	49932	51547	50633	1.64	cps
Erbium	164-1	56882	58572	57906	57787	1.47	cps
Erbium	164-2	35107	35462	35144	35238	0.55	cps
Gadolinium	160-1	64936	66125	65512	65524	0.91	cps
Gadolinium	160-2	42852	42227	42026	42368	1.02	cps
Holmium	165-1	11048152	11239983	10947404	11078513	1.34	cps
Holmium	165-2	6696088	6582683	6634038	6637603	0.86	cps
Indium	115-1	8326353	8344959	8256083	8309132	0.56	cps
Indium	115-2	2565563	2599867	2589813	2585081	0.68	cps
Iron	56-2	255029650	254880430	254216810	254708963	0.17	cps
Iron	57-2	6369743	6345899	6294269	6336637	0.61	cps
Iron	54-2	13915711	13883551	14023221	13940828	0.52	cps
Krypton	83-1	297	227	267	263	13.34	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-22 Instrumnet Name : P7
Client Sample ID : MYE4L3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:23:13 DataFile Name : 056SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	280496	282236	282606	281779	0.40	cps
Lead	207-1	242352	241822	241049	241741	0.27	cps
Lead	208-1	1098595	1103368	1104445	1102136	0.28	cps
Lithium	6-1	1370267	1365029	1362864	1366053	0.28	cps
Magnesium	24-2	6294089	6262636	6235231	6263985	0.47	cps
Manganese	55-2	4966925	4922785	4950225	4946645	0.45	cps
Molybdenum	94-1	46333	46343	45651	46109	0.86	cps
Molybdenum	95-1	8399	8342	8579	8440	1.46	cps
Molybdenum	96-1	19413	19984	19793	19730	1.47	cps
Molybdenum	97-1	5081	5114	5071	5089	0.45	cps
Molybdenum	98-1	12882	12702	12989	12858	1.13	cps
Neodymium	150-1	120359	121068	121040	120822	0.33	cps
Neodymium	150-2	61513	62026	61534	61691	0.47	cps
Nickel	60-2	92808	91302	91131	91747	1.01	cps
Phosphorus	31-2	28156	28062	27575	27931	1.12	cps
Potassium	39-2	5768685	5785476	5812753	5788971	0.38	cps
Rhodium	103-1	7691289	7634198	7590300	7638596	0.66	cps
Rhodium	103-2	4401409	4399338	4409181	4403310	0.12	cps
Scandium	45-1	4729963	4771459	4682161	4727861	0.95	cps
Scandium	45-2	418874	420412	420943	420076	0.26	cps
Selenium	82-1	589	671	615	625	6.73	cps
Selenium	77-2	200	203	213	206	3.38	cps
Selenium	78-2	320	410	400	377	13.10	cps
Silicon	28-1	13813165	13920288	13972296	13901916	0.58	cps
Silver	107-1	4097	3921	4124	4047	2.73	cps
Silver	109-1	3901	3811	4034	3915	2.87	cps
Sodium	23-2	262481	259556	258875	260304	0.74	cps
Strontium	86-1	313425	317671	320508	317201	1.12	cps
Strontium	88-1	2873557	2867770	2885115	2875481	0.31	cps
Sulfur	34-1	142676	145306	144627	144203	0.95	cps
Terbium	159-1	11373586	11353907	11216253	11314582	0.76	cps
Terbium	159-2	6668633	6675336	6636336	6660102	0.31	cps
Thallium	203-1	2840	2854	3027	2907	3.58	cps
Thallium	205-1	6845	6395	6802	6681	3.72	cps
Tin	118-1	18973	19470	19260	19234	1.30	cps
Titanium	47-1	1317169	1307141	1302829	1309046	0.56	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4484-22 Instrumnet Name : P7
Client Sample ID : MYE4L3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:23:13 DataFile Name : 056SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	31888	31591	31671	31717	0.49	cps
Vanadium	51-2	515455	512633	512823	513637	0.31	cps
Yttrium	89-1	14573776	14389547	14295419	14419581	0.98	cps
Yttrium	89-2	3455404	3417437	3442004	3438282	0.56	cps
Zinc	66-2	135979	134983	134567	135176	0.54	cps
Zirconium	90-1	112490	114455	113819	113588	0.88	cps
Zirconium	91-1	25021	25539	25192	25251	1.04	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-01 Instrumnet Name : P7
Client Sample ID : MYE4L4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:36:08 DataFile Name : 060SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	6846272	6964681	6896809	6902587	0.86	cps
Antimony	121-1	2347	2440	2390	2392	1.95	cps
Arsenic	75-2	14521	14480	14397	14466	0.44	cps
Barium	135-1	571987	572308	572905	572400	0.08	cps
Barium	137-1	982294	998257	1003929	994827	1.13	cps
Beryllium	9-1	861	944	778	861	9.68	cps
Bismuth	209-1	6658445	6735596	6740133	6711391	0.68	cps
Bismuth	209-2	4785449	3907266	4623011	4438575	10.53	cps
Bromine	81-1	13526	13616	13937	13693	1.58	cps
Bromine	81-2	657	737	700	698	5.74	cps
Cadmium	108-1	233	223	243	233	4.29	cps
Cadmium	106-1	3147	3017	2974	3046	2.96	cps
Cadmium	111-1	3483	3299	3371	3384	2.74	cps
Calcium	43-1	490198	494902	491858	492319	0.48	cps
Calcium	44-1	7586282	7519701	7513420	7539801	0.54	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	457966	459603	452669	456746	0.79	cps
Cobalt	59-2	303963	304448	305698	304703	0.29	cps
Copper	63-2	663542	666082	668009	665877	0.34	cps
Dysprosium	156-1	44286	43343	43310	43646	1.27	cps
Dysprosium	156-2	31933	32284	31876	32031	0.69	cps
Erbium	164-1	40943	41746	41732	41474	1.11	cps
Erbium	164-2	25383	25907	25527	25606	1.06	cps
Gadolinium	160-1	43240	43537	44460	43746	1.45	cps
Gadolinium	160-2	29204	29294	29424	29307	0.38	cps
Holmium	165-1	10881179	10840022	10774197	10831800	0.50	cps
Holmium	165-2	6774189	5490125	6454846	6239720	10.71	cps
Indium	115-1	8370149	8220328	8212597	8267692	1.07	cps
Indium	115-2	2590649	2124838	2477143	2397543	10.13	cps
Iron	56-2	205424430	202842937	204128550	204131973	0.63	cps
Iron	57-2	5107510	5106989	5123580	5112693	0.18	cps
Iron	54-2	11131991	11281200	11036423	11149871	1.11	cps
Krypton	83-1	280	223	283	262	12.86	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-01 Instrumnet Name : P7
Client Sample ID : MYE4L4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:36:08 DataFile Name : 060SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	145691	146469	148069	146743	0.83	cps
Lead	207-1	123344	128335	126546	126075	2.01	cps
Lead	208-1	568429	582455	580114	576999	1.30	cps
Lithium	6-1	1353611	1330517	1346425	1343518	0.88	cps
Magnesium	24-2	9029161	9169454	9081577	9093398	0.78	cps
Manganese	55-2	5045161	5003343	4987076	5011860	0.60	cps
Molybdenum	94-1	43281	44404	44046	43910	1.31	cps
Molybdenum	95-1	4057	3857	4007	3974	2.62	cps
Molybdenum	96-1	14067	14340	14300	14236	1.04	cps
Molybdenum	97-1	2430	2524	2514	2489	2.06	cps
Molybdenum	98-1	6015	5961	6128	6035	1.41	cps
Neodymium	150-1	69895	68653	70391	69646	1.29	cps
Neodymium	150-2	36724	36551	35983	36419	1.07	cps
Nickel	60-2	144539	143914	143314	143922	0.43	cps
Phosphorus	31-2	23845	24667	24246	24253	1.69	cps
Potassium	39-2	2824139	2787023	2795175	2802112	0.70	cps
Rhodium	103-1	7594757	7436167	7430588	7487171	1.24	cps
Rhodium	103-2	4416199	3506860	4188834	4037297	11.72	cps
Scandium	45-1	4644702	4535967	4542694	4574455	1.33	cps
Scandium	45-2	409220	335502	394166	379629	10.26	cps
Selenium	82-1	481	584	493	519	10.88	cps
Selenium	77-2	107	133	123	121	11.12	cps
Selenium	78-2	317	337	353	336	5.47	cps
Silicon	28-1	19323494	19409298	19351024	19361272	0.23	cps
Silver	107-1	1930	2154	2060	2048	5.48	cps
Silver	109-1	1767	1827	1927	1840	4.39	cps
Sodium	23-2	172606	172427	170637	171890	0.63	cps
Strontium	86-1	190455	192278	192180	191638	0.54	cps
Strontium	88-1	1746466	1746826	1711517	1734936	1.17	cps
Sulfur	34-1	141690	135864	136311	137955	2.35	cps
Terbium	159-1	11112904	10998928	10958708	11023513	0.73	cps
Terbium	159-2	6669985	5367349	6451289	6162874	11.32	cps
Thallium	203-1	1563	1610	1713	1629	4.71	cps
Thallium	205-1	3917	3894	4034	3948	1.90	cps
Tin	118-1	4431	4381	4307	4373	1.42	cps
Titanium	47-1	872698	873206	866640	870848	0.42	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-01 Instrumnet Name : P7
Client Sample ID : MYE4L4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:36:08 DataFile Name : 060SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	15706	15876	16026	15870	1.01	cps
Vanadium	51-2	447904	444867	444571	445781	0.41	cps
Yttrium	89-1	14162232	14151505	13958587	14090775	0.81	cps
Yttrium	89-2	3428767	2758266	3268665	3151899	11.11	cps
Zinc	66-2	94154	94064	93654	93957	0.28	cps
Zirconium	90-1	112275	112007	113692	112658	0.80	cps
Zirconium	91-1	25282	25415	25258	25319	0.33	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-02 Instrumnet Name : P7
Client Sample ID : MYE4L5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:39:15 DataFile Name : 061SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	6646537	6581470	6672833	6633614	0.71	cps
Antimony	121-1	2247	2220	2284	2250	1.41	cps
Arsenic	75-2	17437	17160	17317	17305	0.80	cps
Barium	135-1	909021	911329	915591	911980	0.37	cps
Barium	137-1	1724837	1719925	1725174	1723312	0.17	cps
Beryllium	9-1	1872	1728	1928	1843	5.60	cps
Bismuth	209-1	6812038	6815888	6794681	6807536	0.17	cps
Bismuth	209-2	4783159	4829862	4837583	4816868	0.61	cps
Bromine	81-1	16483	18308	18879	17890	7.00	cps
Bromine	81-2	1640	1643	1707	1663	2.26	cps
Cadmium	108-1	247	187	237	223	14.39	cps
Cadmium	106-1	3107	3140	3160	3136	0.86	cps
Cadmium	111-1	3922	4071	3917	3970	2.20	cps
Calcium	43-1	560200	567756	565772	564576	0.69	cps
Calcium	44-1	8613011	8699936	8716942	8676630	0.64	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	245979	245568	245111	245553	0.18	cps
Cobalt	59-2	230168	228097	231339	229868	0.71	cps
Copper	63-2	611186	611123	606377	609562	0.45	cps
Dysprosium	156-1	68633	69240	68399	68757	0.63	cps
Dysprosium	156-2	51992	52668	52277	52313	0.65	cps
Erbium	164-1	59292	60962	60972	60409	1.60	cps
Erbium	164-2	37440	36795	37166	37134	0.87	cps
Gadolinium	160-1	66523	68087	68362	67658	1.47	cps
Gadolinium	160-2	45293	44952	44664	44970	0.70	cps
Holmium	165-1	10893494	10691197	10944717	10843136	1.24	cps
Holmium	165-2	6756003	6757528	6753068	6755533	0.03	cps
Indium	115-1	8236809	8143184	8147059	8175684	0.65	cps
Indium	115-2	2617167	2595409	2611275	2607950	0.43	cps
Iron	56-2	192959210	193718297	194266704	193648070	0.34	cps
Iron	57-2	4816425	4799021	4886194	4833880	0.95	cps
Iron	54-2	10647959	10500636	10508927	10552507	0.78	cps
Krypton	83-1	243	277	263	261	6.42	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-02 Instrumnet Name : P7
Client Sample ID : MYE4L5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:39:15 DataFile Name : 061SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	274894	276016	280174	277028	1.00	cps
Lead	207-1	233240	235050	236820	235037	0.76	cps
Lead	208-1	1065310	1074759	1086423	1075497	0.98	cps
Lithium	6-1	1396467	1344163	1346739	1362456	2.16	cps
Magnesium	24-2	6087533	6102264	6089212	6093003	0.13	cps
Manganese	55-2	5378505	5418126	5393440	5396690	0.37	cps
Molybdenum	94-1	59581	61264	60638	60494	1.41	cps
Molybdenum	95-1	4938	5264	5084	5095	3.21	cps
Molybdenum	96-1	17657	18438	17844	17980	2.27	cps
Molybdenum	97-1	3210	3120	3100	3144	1.86	cps
Molybdenum	98-1	7862	7856	7739	7819	0.89	cps
Neodymium	150-1	117367	119441	121602	119470	1.77	cps
Neodymium	150-2	63184	63666	62591	63147	0.85	cps
Nickel	60-2	89213	88714	88694	88874	0.33	cps
Phosphorus	31-2	26393	26700	27458	26850	2.04	cps
Potassium	39-2	4200023	4158993	4173016	4177344	0.50	cps
Rhodium	103-1	7430483	7431150	7402428	7421354	0.22	cps
Rhodium	103-2	4386741	4405742	4412357	4401614	0.30	cps
Scandium	45-1	4535519	4514288	4536136	4528647	0.27	cps
Scandium	45-2	408494	410731	408604	409277	0.31	cps
Selenium	82-1	562	555	622	580	6.33	cps
Selenium	77-2	177	223	240	213	15.39	cps
Selenium	78-2	417	393	453	421	7.18	cps
Silicon	28-1	21916941	22390307	22147145	22151464	1.07	cps
Silver	107-1	2980	2934	3000	2971	1.15	cps
Silver	109-1	2767	2690	2730	2729	1.41	cps
Sodium	23-2	188381	187574	189366	188440	0.48	cps
Strontium	86-1	264811	268843	271397	268351	1.24	cps
Strontium	88-1	2388995	2438531	2444084	2423870	1.25	cps
Sulfur	34-1	139338	138743	137616	138566	0.63	cps
Terbium	159-1	11170099	10998459	10897476	11022011	1.25	cps
Terbium	159-2	6798569	6848009	6773728	6806769	0.56	cps
Thallium	203-1	2807	2867	2834	2836	1.06	cps
Thallium	205-1	6718	6685	6955	6786	2.17	cps
Tin	118-1	5148	5141	5448	5245	3.34	cps
Titanium	47-1	1024251	1034236	1037929	1032139	0.69	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-02 Instrumnet Name : P7
Client Sample ID : MYE4L5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:39:15 DataFile Name : 061SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	29570	29028	30121	29573	1.85	cps
Vanadium	51-2	388576	385135	386462	386724	0.45	cps
Yttrium	89-1	14162297	14001461	14028833	14064197	0.61	cps
Yttrium	89-2	3440488	3456562	3403368	3433473	0.79	cps
Zinc	66-2	123264	123045	122827	123045	0.18	cps
Zirconium	90-1	156480	156925	156999	156801	0.18	cps
Zirconium	91-1	34624	34721	35092	34812	0.71	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BL Instrumnet Name : P7
Client Sample ID : PBS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:48:20 DataFile Name : 063CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1854	1463	1437	1585	14.73	cps
Antimony	121-1	163	143	140	149	8.48	cps
Arsenic	75-2	10	13	30	18	60.28	cps
Barium	135-1	350	393	400	381	7.12	cps
Barium	137-1	747	613	640	667	10.58	cps
Beryllium	9-1	28	33	11	24	48.04	cps
Bismuth	209-1	6712390	6780321	6566311	6686340	1.64	cps
Bismuth	209-2	4727050	4799871	4751608	4759510	0.78	cps
Bromine	81-1	11178	11531	11631	11447	2.08	cps
Bromine	81-2	83	87	77	82	6.19	cps
Cadmium	108-1	7	13	17	12	41.65	cps
Cadmium	106-1	2497	2237	2407	2380	5.55	cps
Cadmium	111-1	1776	1570	1686	1678	6.18	cps
Calcium	43-1	367	293	423	361	18.05	cps
Calcium	44-1	13316	12869	13383	13190	2.12	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	660	703	720	694	4.46	cps
Cobalt	59-2	177	197	163	179	9.38	cps
Copper	63-2	1143	1163	1030	1112	6.47	cps
Dysprosium	156-1	27	20	33	27	24.99	cps
Dysprosium	156-2	23	10	17	17	39.99	cps
Erbium	164-1	117	57	70	81	38.84	cps
Erbium	164-2	43	40	43	42	4.55	cps
Gadolinium	160-1	60	50	37	49	23.94	cps
Gadolinium	160-2	250	277	297	274	8.53	cps
Holmium	165-1	10449803	10657609	10437355	10514922	1.18	cps
Holmium	165-2	6514418	6570082	6510777	6531759	0.51	cps
Indium	115-1	8277497	8359429	8105639	8247522	1.57	cps
Indium	115-2	2594548	2642684	2616820	2618017	0.92	cps
Iron	56-2	67682	65033	65194	65970	2.25	cps
Iron	57-2	2507	2587	2594	2562	1.88	cps
Iron	54-2	4657	4587	4591	4612	0.86	cps
Krypton	83-1	183	137	187	169	16.55	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BL Instrumnet Name : P7
Client Sample ID : PBS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:48:20 DataFile Name : 063CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	803	747	780	777	3.67	cps
Lead	207-1	740	600	703	681	10.66	cps
Lead	208-1	3264	3144	2997	3135	4.26	cps
Lithium	6-1	1342399	1319703	1305958	1322687	1.39	cps
Magnesium	24-2	2847	2537	2490	2625	7.39	cps
Manganese	55-2	1560	1473	1427	1487	4.55	cps
Molybdenum	94-1	417	403	377	399	5.11	cps
Molybdenum	95-1	343	283	330	319	9.88	cps
Molybdenum	96-1	487	400	380	422	13.43	cps
Molybdenum	97-1	197	180	207	194	6.93	cps
Molybdenum	98-1	517	533	533	528	1.82	cps
Neodymium	150-1	33	40	23	32	26.04	cps
Neodymium	150-2	23	10	17	17	39.99	cps
Nickel	60-2	187	140	177	168	14.65	cps
Phosphorus	31-2	167	170	140	159	10.35	cps
Potassium	39-2	45073	45474	44863	45137	0.69	cps
Rhodium	103-1	7705255	7627934	7584405	7639198	0.80	cps
Rhodium	103-2	4416094	4478203	4521960	4472086	1.19	cps
Scandium	45-1	4410977	4446993	4382724	4413564	0.73	cps
Scandium	45-2	387102	389942	388828	388624	0.37	cps
Selenium	82-1	427	451	438	439	2.79	cps
Selenium	77-2	3	0	0	1	173.21	cps
Selenium	78-2	227	247	283	252	11.39	cps
Silicon	28-1	198047	198628	201724	199467	0.99	cps
Silver	107-1	153	163	163	160	3.61	cps
Silver	109-1	107	143	157	136	19.10	cps
Sodium	23-2	16529	16422	16596	16516	0.53	cps
Strontium	86-1	747	780	717	748	4.24	cps
Strontium	88-1	4031	3924	3897	3951	1.79	cps
Sulfur	34-1	127638	129348	129920	128969	0.92	cps
Terbium	159-1	10983212	10844900	10816247	10881453	0.82	cps
Terbium	159-2	6549042	6578541	6523637	6550406	0.42	cps
Thallium	203-1	207	223	220	217	4.07	cps
Thallium	205-1	473	583	560	539	10.76	cps
Tin	118-1	1540	1583	1623	1582	2.63	cps
Titanium	47-1	377	330	307	338	10.55	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BL Instrumnet Name : P7
Client Sample ID : PBS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:48:20 DataFile Name : 063CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	200	153	127	160	23.20	cps
Vanadium	51-2	157	97	210	154	36.71	cps
Yttrium	89-1	13403536	13241984	13251227	13298916	0.68	cps
Yttrium	89-2	3212928	3149057	3189992	3183992	1.02	cps
Zinc	66-2	587	533	527	549	5.99	cps
Zirconium	90-1	550	557	557	554	0.69	cps
Zirconium	91-1	103	147	80	110	30.76	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BS Instrumnet Name : P7
Client Sample ID : LCS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:52:15 DataFile Name : 064LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	9756	9673	10003	9811	1.75	cps
Antimony	121-1	36877	36974	37335	37062	0.65	cps
Arsenic	75-2	967	990	973	977	1.23	cps
Barium	135-1	47255	46977	47917	47383	1.02	cps
Barium	137-1	81244	81787	82507	81846	0.77	cps
Beryllium	9-1	2867	2828	2678	2791	3.57	cps
Bismuth	209-1	6552762	6619215	6519817	6563931	0.77	cps
Bismuth	209-2	4697601	4693273	4657080	4682651	0.48	cps
Bromine	81-1	10864	11011	10854	10910	0.80	cps
Bromine	81-2	127	73	67	89	37.00	cps
Cadmium	108-1	383	267	373	341	18.96	cps
Cadmium	106-1	3070	2830	2750	2884	5.78	cps
Cadmium	111-1	6285	6162	6018	6155	2.17	cps
Calcium	43-1	24814	25508	25682	25335	1.81	cps
Calcium	44-1	418997	420778	424330	421368	0.64	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	17096	17016	17116	17076	0.31	cps
Cobalt	59-2	12425	12999	12449	12624	2.57	cps
Copper	63-2	18985	18809	18431	18742	1.51	cps
Dysprosium	156-1	50	43	33	42	19.87	cps
Dysprosium	156-2	17	17	7	13	43.29	cps
Erbium	164-1	83	53	100	79	29.98	cps
Erbium	164-2	67	43	37	49	32.22	cps
Gadolinium	160-1	57	90	37	61	44.08	cps
Gadolinium	160-2	230	290	303	274	14.23	cps
Holmium	165-1	10550932	10383559	10302217	10412236	1.22	cps
Holmium	165-2	6465998	6419884	6372952	6419611	0.72	cps
Indium	115-1	8162361	8191475	8094906	8149581	0.61	cps
Indium	115-2	2598160	2530471	2540790	2556474	1.43	cps
Iron	56-2	1467148	1462466	1446116	1458577	0.76	cps
Iron	57-2	38215	38432	38305	38318	0.28	cps
Iron	54-2	83811	82883	82876	83190	0.65	cps
Krypton	83-1	183	220	187	197	10.31	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BS Instrumnet Name : P7
Client Sample ID : LCS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:52:15 DataFile Name : 064LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	14324	14688	14304	14439	1.50	cps
Lead	207-1	12462	12666	12823	12650	1.43	cps
Lead	208-1	57027	58045	57935	57669	0.97	cps
Lithium	6-1	1305355	1304800	1291589	1300581	0.60	cps
Magnesium	24-2	385247	384441	381551	383747	0.51	cps
Manganese	55-2	6962	7195	6922	7026	2.10	cps
Molybdenum	94-1	39483	39864	39757	39701	0.49	cps
Molybdenum	95-1	47450	47761	47156	47455	0.64	cps
Molybdenum	96-1	53173	53056	54096	53442	1.07	cps
Molybdenum	97-1	29727	29633	29804	29721	0.29	cps
Molybdenum	98-1	76660	77219	77320	77066	0.46	cps
Neodymium	150-1	37	20	13	23	51.52	cps
Neodymium	150-2	13	13	13	13	0.00	cps
Nickel	60-2	3367	3257	3237	3287	2.13	cps
Phosphorus	31-2	160	183	147	163	11.36	cps
Potassium	39-2	502662	496877	495535	498358	0.76	cps
Rhodium	103-1	7571176	7608494	7550438	7576703	0.39	cps
Rhodium	103-2	4449675	4346510	4380443	4392209	1.20	cps
Scandium	45-1	4386634	4358662	4363988	4369761	0.34	cps
Scandium	45-2	389108	387399	385512	387340	0.46	cps
Selenium	82-1	1879	1843	2007	1910	4.53	cps
Selenium	77-2	150	120	140	137	11.18	cps
Selenium	78-2	710	720	703	711	1.18	cps
Silicon	28-1	285445	289425	291348	288739	1.04	cps
Silver	107-1	22207	22781	22684	22558	1.36	cps
Silver	109-1	21289	21499	22197	21662	2.19	cps
Sodium	23-2	625578	623943	620241	623254	0.44	cps
Strontium	86-1	7132	6848	6945	6975	2.07	cps
Strontium	88-1	57442	56133	57967	57181	1.65	cps
Sulfur	34-1	135783	138279	142588	138883	2.48	cps
Terbium	159-1	10716146	10709838	10704279	10710088	0.06	cps
Terbium	159-2	6441525	6442520	6451345	6445130	0.08	cps
Thallium	203-1	16764	17244	16907	16972	1.45	cps
Thallium	205-1	40015	39902	40203	40040	0.38	cps
Tin	118-1	1720	1633	1833	1729	5.80	cps
Titanium	47-1	12082	12479	12509	12356	1.93	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BS Instrumnet Name : P7
Client Sample ID : LCS572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:52:15 DataFile Name : 064LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	50482	49174	51720	50459	2.52	cps
Vanadium	51-2	34630	34767	34286	34561	0.72	cps
Yttrium	89-1	13125278	13040100	12990591	13051990	0.52	cps
Yttrium	89-2	3220150	3188688	3214412	3207750	0.52	cps
Zinc	66-2	7929	7622	7449	7666	3.17	cps
Zirconium	90-1	33772	33097	33565	33478	1.03	cps
Zirconium	91-1	7469	7595	7529	7531	0.84	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BSD Instrumnet Name : P7
Client Sample ID : LCSD572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:55:31 DataFile Name : 065LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5651	5694	5671	5672	0.38	cps
Antimony	121-1	18686	18626	18853	18721	0.63	cps
Arsenic	75-2	597	540	530	556	6.47	cps
Barium	135-1	24264	23686	23666	23872	1.42	cps
Barium	137-1	40233	40972	41293	40833	1.33	cps
Beryllium	9-1	1395	1256	1361	1337	5.42	cps
Bismuth	209-1	6620838	6582502	6622732	6608691	0.34	cps
Bismuth	209-2	4761890	4647569	4809154	4739538	1.75	cps
Bromine	81-1	10947	10607	10724	10760	1.61	cps
Bromine	81-2	77	67	90	78	15.05	cps
Cadmium	108-1	143	183	177	168	12.77	cps
Cadmium	106-1	2767	2687	2874	2776	3.37	cps
Cadmium	111-1	4043	3929	4221	4065	3.62	cps
Calcium	43-1	12705	13356	12912	12991	2.56	cps
Calcium	44-1	214997	218375	217983	217118	0.85	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	8829	8502	8656	8663	1.89	cps
Cobalt	59-2	6228	6351	6518	6366	2.29	cps
Copper	63-2	9730	10437	10290	10152	3.68	cps
Dysprosium	156-1	33	13	33	27	43.31	cps
Dysprosium	156-2	20	13	3	12	68.66	cps
Erbium	164-1	90	70	73	78	13.78	cps
Erbium	164-2	57	37	43	46	22.36	cps
Gadolinium	160-1	40	50	77	56	34.12	cps
Gadolinium	160-2	270	267	250	262	4.09	cps
Holmium	165-1	10345985	10320594	10569824	10412134	1.32	cps
Holmium	165-2	6496861	6400574	6585618	6494351	1.43	cps
Indium	115-1	8115253	8201853	8199943	8172349	0.61	cps
Indium	115-2	2580612	2540357	2571821	2564263	0.83	cps
Iron	56-2	770865	772152	766229	769749	0.40	cps
Iron	57-2	20351	20531	20477	20453	0.45	cps
Iron	54-2	44064	43349	43526	43646	0.85	cps
Krypton	83-1	223	137	193	184	23.86	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BSD Instrumnet Name : P7
Client Sample ID : LCSD572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:55:31 DataFile Name : 065LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	7272	7315	7709	7432	3.24	cps
Lead	207-1	6552	6362	6532	6482	1.61	cps
Lead	208-1	29756	29469	30310	29845	1.43	cps
Lithium	6-1	1303355	1267131	1294671	1288386	1.47	cps
Magnesium	24-2	190197	190741	191415	190784	0.32	cps
Manganese	55-2	4311	4364	4091	4255	3.41	cps
Molybdenum	94-1	19446	19927	19814	19729	1.27	cps
Molybdenum	95-1	23693	24033	24684	24137	2.09	cps
Molybdenum	96-1	26320	27012	26684	26672	1.30	cps
Molybdenum	97-1	14778	15141	14874	14931	1.26	cps
Molybdenum	98-1	38651	38755	38420	38609	0.44	cps
Neodymium	150-1	30	17	20	22	31.22	cps
Neodymium	150-2	13	17	7	12	41.65	cps
Nickel	60-2	1833	1753	1887	1825	3.68	cps
Phosphorus	31-2	153	193	143	163	16.20	cps
Potassium	39-2	272030	271965	269218	271071	0.59	cps
Rhodium	103-1	7655532	7604687	7630075	7630098	0.33	cps
Rhodium	103-2	4403707	4430235	4411287	4415076	0.31	cps
Scandium	45-1	4407515	4393795	4442144	4414485	0.56	cps
Scandium	45-2	393092	391958	391889	392313	0.17	cps
Selenium	82-1	1134	1191	1171	1165	2.49	cps
Selenium	77-2	93	63	43	67	37.75	cps
Selenium	78-2	530	467	497	498	6.36	cps
Silicon	28-1	247701	249949	248924	248858	0.45	cps
Silver	107-1	11158	11511	11101	11257	1.98	cps
Silver	109-1	10410	10954	10914	10760	2.82	cps
Sodium	23-2	318970	318876	318199	318682	0.13	cps
Strontium	86-1	3677	3747	3837	3754	2.14	cps
Strontium	88-1	30802	30482	30862	30715	0.67	cps
Sulfur	34-1	143567	146181	147554	145767	1.39	cps
Terbium	159-1	10700937	10625961	10676020	10667639	0.36	cps
Terbium	159-2	6496786	6561454	6476225	6511488	0.68	cps
Thallium	203-1	8143	8113	8173	8143	0.37	cps
Thallium	205-1	19684	20035	20075	19931	1.08	cps
Tin	118-1	1577	1660	4732	2656	67.70	cps
Titanium	47-1	6085	6281	6325	6230	2.05	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164572BSD Instrumnet Name : P7
Client Sample ID : LCSD572 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:55:31 DataFile Name : 065LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	25167	25535	25020	25241	1.05	cps
Vanadium	51-2	17540	17484	17454	17492	0.25	cps
Yttrium	89-1	13309686	13349072	13375465	13344741	0.25	cps
Yttrium	89-2	3219498	3226264	3199610	3215124	0.43	cps
Zinc	66-2	3924	4014	4107	4015	2.28	cps
Zirconium	90-1	16856	16506	17073	16812	1.70	cps
Zirconium	91-1	3791	3687	3877	3785	2.51	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-03 Instrumnet Name : P7
Client Sample ID : MYE4L6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:58:46 DataFile Name : 066SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4552061	4602516	4564263	4572947	0.58	cps
Antimony	121-1	2137	2220	2454	2270	7.23	cps
Arsenic	75-2	10074	9847	9870	9930	1.26	cps
Barium	135-1	552251	557909	562271	557477	0.90	cps
Barium	137-1	966755	972437	969758	969650	0.29	cps
Beryllium	9-1	1078	1250	1139	1156	7.56	cps
Bismuth	209-1	6597883	6699008	6260541	6519144	3.52	cps
Bismuth	209-2	4780995	4769424	4773565	4774661	0.12	cps
Bromine	81-1	14564	14844	15408	14939	2.88	cps
Bromine	81-2	977	973	960	970	0.91	cps
Cadmium	108-1	177	197	167	180	8.49	cps
Cadmium	106-1	3147	3074	2727	2983	7.52	cps
Cadmium	111-1	3143	3004	2792	2980	5.92	cps
Calcium	43-1	548564	554518	554677	552587	0.63	cps
Calcium	44-1	8427303	8424642	8412269	8421405	0.10	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	190788	191408	192138	191445	0.35	cps
Cobalt	59-2	168713	168275	169215	168735	0.28	cps
Copper	63-2	321778	321880	322815	322157	0.18	cps
Dysprosium	156-1	47804	49416	48443	48555	1.67	cps
Dysprosium	156-2	37169	37126	37657	37317	0.79	cps
Erbium	164-1	42267	42608	43330	42735	1.27	cps
Erbium	164-2	26742	26625	26221	26530	1.03	cps
Gadolinium	160-1	46049	47152	47651	46951	1.75	cps
Gadolinium	160-2	30857	31546	31773	31392	1.52	cps
Holmium	165-1	10723471	10710222	10162295	10531996	3.04	cps
Holmium	165-2	6658820	6786161	6701064	6715348	0.97	cps
Indium	115-1	8154410	8170811	7778458	8034560	2.76	cps
Indium	115-2	2557045	2599819	2606219	2587694	1.03	cps
Iron	56-2	153659408	152082911	150658318	152133546	0.99	cps
Iron	57-2	3788464	3761323	3796852	3782213	0.49	cps
Iron	54-2	8380112	8348742	8243407	8324087	0.86	cps
Krypton	83-1	217	233	280	243	13.49	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-03 Instrumnet Name : P7
Client Sample ID : MYE4L6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:58:46 DataFile Name : 066SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	168828	170527	171727	170361	0.86	cps
Lead	207-1	141239	142613	142407	142086	0.52	cps
Lead	208-1	647861	660315	658213	655463	1.02	cps
Lithium	6-1	1335555	1305397	1280745	1307232	2.10	cps
Magnesium	24-2	6188436	6233188	6213942	6211855	0.36	cps
Manganese	55-2	3810762	3785148	3697889	3764600	1.57	cps
Molybdenum	94-1	48316	48075	47948	48113	0.39	cps
Molybdenum	95-1	2827	3007	3167	3000	5.67	cps
Molybdenum	96-1	12859	13496	13183	13179	2.42	cps
Molybdenum	97-1	1723	1807	1747	1759	2.44	cps
Molybdenum	98-1	4504	4884	4444	4611	5.18	cps
Neodymium	150-1	83429	84548	85074	84350	1.00	cps
Neodymium	150-2	43597	44711	44938	44415	1.62	cps
Nickel	60-2	77640	77033	79282	77985	1.49	cps
Phosphorus	31-2	25064	25034	23949	24682	2.57	cps
Potassium	39-2	2117058	2116654	2079051	2104254	1.04	cps
Rhodium	103-1	7563965	7553929	7089987	7402627	3.66	cps
Rhodium	103-2	4304104	4439622	4388110	4377279	1.56	cps
Scandium	45-1	4540228	4564948	4305535	4470237	3.20	cps
Scandium	45-2	402344	404442	402194	402993	0.31	cps
Selenium	82-1	572	535	481	529	8.71	cps
Selenium	77-2	140	167	200	169	17.80	cps
Selenium	78-2	343	373	320	346	7.74	cps
Silicon	28-1	21932286	22264334	22045115	22080578	0.76	cps
Silver	107-1	1233	1303	1257	1265	2.82	cps
Silver	109-1	1077	1263	1147	1162	8.11	cps
Sodium	23-2	201684	200369	199864	200639	0.47	cps
Strontium	86-1	204942	208370	207379	206897	0.85	cps
Strontium	88-1	1896603	1872137	1872031	1880257	0.75	cps
Sulfur	34-1	149897	144987	142124	145669	2.70	cps
Terbium	159-1	11007340	10863639	10486936	10785972	2.49	cps
Terbium	159-2	6668013	6763175	6600754	6677314	1.22	cps
Thallium	203-1	1513	1770	1530	1605	8.95	cps
Thallium	205-1	3721	3861	3687	3756	2.45	cps
Tin	118-1	4374	12303	4664	7114	63.21	cps
Titanium	47-1	845892	853630	853990	851171	0.54	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-03 Instrumnet Name : P7
Client Sample ID : MYE4L6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 20:58:46 DataFile Name : 066SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	18580	19070	18750	18800	1.33	cps
Vanadium	51-2	317656	315237	315404	316099	0.43	cps
Yttrium	89-1	13846006	13972558	13193861	13670808	3.06	cps
Yttrium	89-2	3366868	3413753	3315240	3365287	1.46	cps
Zinc	66-2	96895	97788	97717	97466	0.51	cps
Zirconium	90-1	124062	126603	128248	126304	1.67	cps
Zirconium	91-1	27970	29002	28371	28448	1.83	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-04 Instrumnet Name : P7
Client Sample ID : MYE4L7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:01:55 DataFile Name : 067SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	6800666	6795818	6708417	6768300	0.77	cps
Antimony	121-1	2104	2207	2120	2144	2.59	cps
Arsenic	75-2	16009	16890	16673	16524	2.78	cps
Barium	135-1	1024838	1050872	1059318	1045010	1.72	cps
Barium	137-1	1930996	1972495	1965567	1956353	1.14	cps
Beryllium	9-1	1633	1939	1972	1848	10.11	cps
Bismuth	209-1	6905567	6700840	6802673	6803026	1.50	cps
Bismuth	209-2	4862911	4940352	4865554	4889606	0.90	cps
Bromine	81-1	17681	19046	19850	18859	5.82	cps
Bromine	81-2	1930	2013	1980	1975	2.12	cps
Cadmium	108-1	183	220	243	216	14.03	cps
Cadmium	106-1	3177	3230	3224	3210	0.91	cps
Cadmium	111-1	3976	4061	3876	3971	2.32	cps
Calcium	43-1	757965	770234	774861	767687	1.14	cps
Calcium	44-1	11467427	11758592	11773843	11666621	1.48	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	254323	254060	255592	254658	0.32	cps
Cobalt	59-2	241297	241555	239978	240943	0.35	cps
Copper	63-2	650022	653145	649682	650950	0.29	cps
Dysprosium	156-1	73387	74892	75096	74458	1.25	cps
Dysprosium	156-2	55318	56921	56239	56159	1.43	cps
Erbium	164-1	63959	64669	66069	64899	1.65	cps
Erbium	164-2	41254	40027	40953	40745	1.57	cps
Gadolinium	160-1	70342	72483	72004	71609	1.57	cps
Gadolinium	160-2	48657	48316	47962	48312	0.72	cps
Holmium	165-1	11141752	10839870	10963379	10981667	1.38	cps
Holmium	165-2	6928886	6850823	6795921	6858543	0.97	cps
Indium	115-1	8140946	8203746	8247412	8197368	0.65	cps
Indium	115-2	2630852	2604332	2571083	2602089	1.15	cps
Iron	56-2	199913077	201141784	199499190	200184684	0.43	cps
Iron	57-2	4935079	4889200	4957666	4927315	0.71	cps
Iron	54-2	10876078	10838821	10872868	10862589	0.19	cps
Krypton	83-1	243	210	227	227	7.35	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-04 Instrumnet Name : P7
Client Sample ID : MYE4L7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:01:55 DataFile Name : 067SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	287843	291205	295265	291438	1.28	cps
Lead	207-1	240782	249085	247636	245834	1.80	cps
Lead	208-1	1116490	1135463	1143290	1131748	1.22	cps
Lithium	6-1	1389389	1395352	1364065	1382936	1.20	cps
Magnesium	24-2	6771588	6711559	6698025	6727057	0.58	cps
Manganese	55-2	6512087	6507862	6503683	6507877	0.06	cps
Molybdenum	94-1	48219	48948	49483	48883	1.30	cps
Molybdenum	95-1	5701	5678	5781	5720	0.95	cps
Molybdenum	96-1	16349	16770	16363	16494	1.45	cps
Molybdenum	97-1	3487	3474	3550	3504	1.17	cps
Molybdenum	98-1	8446	8626	8689	8587	1.47	cps
Neodymium	150-1	125724	127262	127793	126926	0.85	cps
Neodymium	150-2	67939	67849	67062	67617	0.71	cps
Nickel	60-2	92107	91329	90320	91252	0.98	cps
Phosphorus	31-2	24710	24683	25104	24832	0.95	cps
Potassium	39-2	3823795	3800229	3810824	3811616	0.31	cps
Rhodium	103-1	7355074	7403077	7351934	7370029	0.39	cps
Rhodium	103-2	4418498	4365635	4383232	4389121	0.61	cps
Scandium	45-1	4517357	4494304	4511569	4507743	0.27	cps
Scandium	45-2	409903	413692	412319	411971	0.47	cps
Selenium	82-1	569	618	581	589	4.38	cps
Selenium	77-2	233	280	280	264	10.19	cps
Selenium	78-2	333	447	427	402	15.04	cps
Silicon	28-1	27196239	27432350	27659406	27429332	0.84	cps
Silver	107-1	2937	2940	2800	2893	2.76	cps
Silver	109-1	2807	2790	2727	2775	1.52	cps
Sodium	23-2	223187	223529	221944	222887	0.37	cps
Strontium	86-1	290547	292158	294874	292526	0.75	cps
Strontium	88-1	2642129	2635554	2707854	2661846	1.50	cps
Sulfur	34-1	144466	141202	139930	141866	1.65	cps
Terbium	159-1	11212404	11133071	11156711	11167395	0.36	cps
Terbium	159-2	6786289	6736023	6711857	6744723	0.56	cps
Thallium	203-1	3017	3124	3154	3098	2.32	cps
Thallium	205-1	7242	7659	7602	7501	3.01	cps
Tin	118-1	3897	3907	3911	3905	0.18	cps
Titanium	47-1	1231630	1251408	1255813	1246284	1.03	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-04 Instrumnet Name : P7
Client Sample ID : MYE4L7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:01:55 DataFile Name : 067SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	31257	31674	31875	31602	1.00	cps
Vanadium	51-2	434302	427877	430014	430731	0.76	cps
Yttrium	89-1	14208546	14034669	14063018	14102078	0.66	cps
Yttrium	89-2	3503300	3407884	3430951	3447378	1.44	cps
Zinc	66-2	124742	124748	123872	124454	0.41	cps
Zirconium	90-1	123016	125151	124697	124288	0.90	cps
Zirconium	91-1	27332	27796	27850	27659	1.03	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-05 Instrumnet Name : P7
Client Sample ID : MYE4L8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:05:01 DataFile Name : 068SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4749495	4804880	4704216	4752864	1.06	cps
Antimony	121-1	2074	2194	2260	2176	4.35	cps
Arsenic	75-2	13036	13296	12659	12997	2.47	cps
Barium	135-1	727350	733511	735595	732152	0.59	cps
Barium	137-1	1259098	1282571	1279044	1273571	0.99	cps
Beryllium	9-1	1256	1272	1433	1320	7.44	cps
Bismuth	209-1	7030064	6907565	6794202	6910611	1.71	cps
Bismuth	209-2	4879394	4775001	4810319	4821571	1.10	cps
Bromine	81-1	15605	16960	17631	16732	6.17	cps
Bromine	81-2	1397	1403	1620	1473	8.62	cps
Cadmium	108-1	143	143	157	148	5.21	cps
Cadmium	106-1	3060	2904	2897	2954	3.13	cps
Cadmium	111-1	3419	3466	3302	3395	2.49	cps
Calcium	43-1	526959	533569	532485	531004	0.67	cps
Calcium	44-1	8045783	8123644	8135558	8101662	0.60	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	180010	179609	178438	179353	0.46	cps
Cobalt	59-2	182819	181416	180889	181708	0.55	cps
Copper	63-2	423185	419876	419814	420959	0.46	cps
Dysprosium	156-1	56031	56292	56928	56417	0.82	cps
Dysprosium	156-2	42658	42628	42808	42698	0.23	cps
Erbium	164-1	48005	49223	49343	48857	1.51	cps
Erbium	164-2	30386	30239	30503	30376	0.44	cps
Gadolinium	160-1	54047	53993	54090	54044	0.09	cps
Gadolinium	160-2	35973	36283	35552	35936	1.02	cps
Holmium	165-1	11035456	10848048	10799323	10894276	1.14	cps
Holmium	165-2	6850365	6669419	6709223	6743002	1.41	cps
Indium	115-1	8386326	8222984	8145600	8251637	1.49	cps
Indium	115-2	2598664	2593113	2545836	2579204	1.13	cps
Iron	56-2	158567424	157285461	156262061	157371649	0.73	cps
Iron	57-2	3930948	3890910	3896078	3905979	0.56	cps
Iron	54-2	8599681	8608428	8579158	8595756	0.17	cps
Krypton	83-1	260	270	247	259	4.52	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-05 Instrumnet Name : P7
Client Sample ID : MYE4L8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:05:01 DataFile Name : 068SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	232898	233018	234136	233351	0.29	cps
Lead	207-1	194252	199591	197720	197188	1.37	cps
Lead	208-1	901762	910261	906024	906015	0.47	cps
Lithium	6-1	1400580	1426834	1378297	1401903	1.73	cps
Magnesium	24-2	4960829	4933124	4876999	4923651	0.87	cps
Manganese	55-2	4527719	4634634	4585262	4582538	1.17	cps
Molybdenum	94-1	35062	35917	36017	35666	1.47	cps
Molybdenum	95-1	4587	4881	4654	4707	3.27	cps
Molybdenum	96-1	12559	12572	12569	12567	0.06	cps
Molybdenum	97-1	2757	2640	3034	2810	7.19	cps
Molybdenum	98-1	7135	7552	7072	7253	3.60	cps
Neodymium	150-1	96571	96199	97189	96653	0.52	cps
Neodymium	150-2	50888	51434	51099	51140	0.54	cps
Nickel	60-2	71268	70247	69912	70476	1.00	cps
Phosphorus	31-2	23431	22684	23331	23149	1.75	cps
Potassium	39-2	2707620	2696552	2721171	2708447	0.46	cps
Rhodium	103-1	7520261	7425377	7361338	7435659	1.08	cps
Rhodium	103-2	4412006	4329808	4278278	4340031	1.55	cps
Scandium	45-1	4486078	4461609	4424136	4457274	0.70	cps
Scandium	45-2	397799	399370	393829	397000	0.72	cps
Selenium	82-1	517	507	456	493	6.61	cps
Selenium	77-2	160	210	203	191	14.21	cps
Selenium	78-2	403	347	330	360	10.68	cps
Silicon	28-1	37057293	37225355	37489264	37257304	0.58	cps
Silver	107-1	1783	1917	1880	1860	3.70	cps
Silver	109-1	1653	1813	1803	1757	5.10	cps
Sodium	23-2	178070	177095	177214	177460	0.30	cps
Strontium	86-1	195132	199357	198388	197626	1.12	cps
Strontium	88-1	1741930	1798032	1779795	1773252	1.61	cps
Sulfur	34-1	131850	129285	128280	129805	1.42	cps
Terbium	159-1	11245677	11108079	11112266	11155340	0.70	cps
Terbium	159-2	6668539	6789050	6661760	6706450	1.07	cps
Thallium	203-1	2187	2077	2037	2100	3.70	cps
Thallium	205-1	4831	5134	4954	4973	3.07	cps
Tin	118-1	3947	4074	3831	3951	3.08	cps
Titanium	47-1	1051064	1066330	1068435	1061943	0.89	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-05 Instrumnet Name : P7
Client Sample ID : MYE4L8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:05:01 DataFile Name : 068SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	26426	26109	26573	26370	0.90	cps
Vanadium	51-2	329642	327470	325242	327451	0.67	cps
Yttrium	89-1	13961913	13961016	13750117	13891015	0.88	cps
Yttrium	89-2	3372867	3317750	3349621	3346746	0.83	cps
Zinc	66-2	92283	91602	92917	92267	0.71	cps
Zirconium	90-1	87522	89806	89048	88792	1.31	cps
Zirconium	91-1	19683	20181	20107	19990	1.34	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-06 Instrumnet Name : P7
Client Sample ID : MYE4L9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:08:10 DataFile Name : 069SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	9357636	9296191	9306391	9320073	0.35	cps
Antimony	121-1	1707	1873	1850	1810	4.99	cps
Arsenic	75-2	8162	7842	8112	8039	2.14	cps
Barium	135-1	241580	241996	245027	242868	0.77	cps
Barium	137-1	423614	424689	426384	424896	0.33	cps
Beryllium	9-1	422	472	433	443	5.93	cps
Bismuth	209-1	6778302	6765128	6544416	6695949	1.96	cps
Bismuth	209-2	4728153	4687653	4757165	4724324	0.74	cps
Bromine	81-1	12952	12662	13059	12891	1.59	cps
Bromine	81-2	560	583	613	586	4.57	cps
Cadmium	108-1	97	100	110	102	6.79	cps
Cadmium	106-1	2817	2754	2774	2781	1.16	cps
Cadmium	111-1	2438	2464	2515	2472	1.58	cps
Calcium	43-1	382826	383777	385010	383871	0.29	cps
Calcium	44-1	5830310	5803199	5848126	5827212	0.39	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	547594	539547	547570	544904	0.85	cps
Cobalt	59-2	423353	416780	419609	419914	0.79	cps
Copper	63-2	758000	747657	753302	752986	0.69	cps
Dysprosium	156-1	22742	23189	22799	22910	1.06	cps
Dysprosium	156-2	16463	16480	16683	16542	0.74	cps
Erbium	164-1	27016	27307	26582	26968	1.35	cps
Erbium	164-2	16914	16974	16323	16737	2.15	cps
Gadolinium	160-1	24772	24645	25366	24928	1.54	cps
Gadolinium	160-2	16476	16690	16523	16563	0.68	cps
Holmium	165-1	10843244	10938959	10667655	10816619	1.27	cps
Holmium	165-2	6654131	6633409	6675557	6654366	0.32	cps
Indium	115-1	8088625	8217228	8023976	8109943	1.21	cps
Indium	115-2	2520225	2526389	2557815	2534810	0.80	cps
Iron	56-2	294569916	291246822	294379616	293398785	0.64	cps
Iron	57-2	7449656	7267028	7316373	7344352	1.29	cps
Iron	54-2	16223341	15850090	16092266	16055232	1.18	cps
Krypton	83-1	257	263	220	247	9.46	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-06 Instrumnet Name : P7
Client Sample ID : MYE4L9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:08:10 DataFile Name : 069SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	66675	68096	67614	67462	1.07	cps
Lead	207-1	59476	59841	60909	60075	1.24	cps
Lead	208-1	269617	268981	271948	270182	0.58	cps
Lithium	6-1	1405763	1398393	1386576	1396911	0.69	cps
Magnesium	24-2	13031830	12973753	12997592	13001058	0.22	cps
Manganese	55-2	4708165	4676515	4700922	4695201	0.35	cps
Molybdenum	94-1	14918	15642	15138	15232	2.44	cps
Molybdenum	95-1	1507	1517	1547	1523	1.37	cps
Molybdenum	96-1	8382	8276	8573	8410	1.79	cps
Molybdenum	97-1	880	990	977	949	6.33	cps
Molybdenum	98-1	2160	2117	2117	2131	1.17	cps
Neodymium	150-1	31636	31465	31519	31540	0.28	cps
Neodymium	150-2	16053	16303	16266	16207	0.83	cps
Nickel	60-2	190909	190632	191530	191024	0.24	cps
Phosphorus	31-2	10220	9566	10190	9992	3.69	cps
Potassium	39-2	1187299	1241528	1245854	1224894	2.66	cps
Rhodium	103-1	7368602	7377198	7313721	7353174	0.47	cps
Rhodium	103-2	4302280	4268155	4326000	4298812	0.68	cps
Scandium	45-1	4655633	4678181	4653713	4662509	0.29	cps
Scandium	45-2	417068	414109	414557	415245	0.38	cps
Selenium	82-1	489	414	490	464	9.32	cps
Selenium	77-2	57	90	63	70	25.20	cps
Selenium	78-2	347	267	310	308	13.01	cps
Silicon	28-1	21409461	21616676	21745132	21590423	0.78	cps
Silver	107-1	1003	997	1023	1008	1.38	cps
Silver	109-1	870	947	933	917	4.47	cps
Sodium	23-2	278728	275363	275953	276681	0.65	cps
Strontium	86-1	154191	155696	154545	154811	0.51	cps
Strontium	88-1	1353316	1352436	1356104	1353952	0.14	cps
Sulfur	34-1	122499	124767	127459	124909	1.99	cps
Terbium	159-1	11186435	11159828	10980489	11108917	1.01	cps
Terbium	159-2	6595891	6525721	6695803	6605805	1.29	cps
Thallium	203-1	753	580	717	683	13.37	cps
Thallium	205-1	1607	1433	1500	1513	5.78	cps
Tin	118-1	4164	4367	4384	4305	2.85	cps
Titanium	47-1	863225	874417	872522	870055	0.69	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-06 Instrumnet Name : P7
Client Sample ID : MYE4L9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:08:10 DataFile Name : 069SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	5575	5591	5461	5542	1.28	cps
Vanadium	51-2	815677	813946	814867	814830	0.11	cps
Yttrium	89-1	13638546	13625407	13236572	13500175	1.69	cps
Yttrium	89-2	3295123	3213483	3280337	3262981	1.33	cps
Zinc	66-2	77815	76870	78063	77582	0.81	cps
Zirconium	90-1	39182	39312	39132	39209	0.24	cps
Zirconium	91-1	9023	8826	8779	8876	1.46	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-07 Instrumnet Name : P7
Client Sample ID : MYE4M0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:11:17 DataFile Name : 070SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4128501	4151529	4122305	4134112	0.37	cps
Antimony	121-1	3100	3230	3310	3214	3.30	cps
Arsenic	75-2	21015	20798	20978	20930	0.55	cps
Barium	135-1	841866	855248	862386	853167	1.22	cps
Barium	137-1	1465278	1480053	1493537	1479623	0.96	cps
Beryllium	9-1	989	1017	1011	1006	1.46	cps
Bismuth	209-1	6925547	6774701	6749477	6816575	1.40	cps
Bismuth	209-2	4787634	4729277	4867219	4794710	1.44	cps
Bromine	81-1	25051	27923	29049	27341	7.54	cps
Bromine	81-2	3620	3554	3480	3552	1.97	cps
Cadmium	108-1	200	203	260	221	15.25	cps
Cadmium	106-1	3234	3167	3137	3179	1.56	cps
Cadmium	111-1	3520	3488	3251	3420	4.31	cps
Calcium	43-1	741256	749908	758304	749823	1.14	cps
Calcium	44-1	11319771	11425134	11446287	11397064	0.59	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	203531	200454	200643	201543	0.86	cps
Cobalt	59-2	147132	146971	148631	147578	0.62	cps
Copper	63-2	237641	238623	241869	239378	0.92	cps
Dysprosium	156-1	38165	39566	39616	39115	2.11	cps
Dysprosium	156-2	30149	29969	30303	30140	0.55	cps
Erbium	164-1	26886	27621	27096	27201	1.39	cps
Erbium	164-2	17264	16884	16737	16961	1.60	cps
Gadolinium	160-1	33831	34472	35298	34534	2.13	cps
Gadolinium	160-2	23329	23390	22852	23190	1.27	cps
Holmium	165-1	10868159	10955525	10705857	10843180	1.17	cps
Holmium	165-2	6614689	6633352	6706687	6651576	0.73	cps
Indium	115-1	8207337	8207420	8188033	8200930	0.14	cps
Indium	115-2	2561482	2551686	2566400	2559856	0.29	cps
Iron	56-2	123335942	123194525	122777708	123102725	0.24	cps
Iron	57-2	3104245	3027622	3069238	3067035	1.25	cps
Iron	54-2	6709557	6719780	6721529	6716955	0.10	cps
Krypton	83-1	223	240	217	227	5.30	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-07 Instrumnet Name : P7
Client Sample ID : MYE4M0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:11:17 DataFile Name : 070SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	135581	136251	135525	135786	0.30	cps
Lead	207-1	116191	117556	117382	117043	0.63	cps
Lead	208-1	529580	535441	534400	533140	0.59	cps
Lithium	6-1	1375093	1367723	1380567	1374461	0.47	cps
Magnesium	24-2	5316123	5346196	5373783	5345367	0.54	cps
Manganese	55-2	3915402	3909289	3891645	3905445	0.32	cps
Molybdenum	94-1	41993	41305	42415	41904	1.34	cps
Molybdenum	95-1	3157	3104	2927	3063	3.93	cps
Molybdenum	96-1	11328	12115	11668	11704	3.37	cps
Molybdenum	97-1	1940	1720	1793	1818	6.16	cps
Molybdenum	98-1	4587	4838	4734	4720	2.66	cps
Neodymium	150-1	73621	73279	74093	73664	0.56	cps
Neodymium	150-2	38813	38329	38218	38453	0.82	cps
Nickel	60-2	79060	78997	79590	79216	0.41	cps
Phosphorus	31-2	18828	18748	18872	18816	0.33	cps
Potassium	39-2	1584309	1571694	1580515	1578839	0.41	cps
Rhodium	103-1	7499540	7389568	7394674	7427927	0.84	cps
Rhodium	103-2	4354546	4351932	4344247	4350242	0.12	cps
Scandium	45-1	4493687	4485321	4385653	4454887	1.35	cps
Scandium	45-2	400402	400664	403387	401485	0.41	cps
Selenium	82-1	495	498	562	519	7.31	cps
Selenium	77-2	93	93	83	90	6.42	cps
Selenium	78-2	300	337	343	327	7.14	cps
Silicon	28-1	20234336	20500230	20578954	20437840	0.88	cps
Silver	107-1	2137	2007	2034	2059	3.34	cps
Silver	109-1	1867	1943	1857	1889	2.51	cps
Sodium	23-2	252810	251939	255090	253280	0.64	cps
Strontium	86-1	518652	521548	526944	522381	0.81	cps
Strontium	88-1	4660816	4800133	4747627	4736192	1.49	cps
Sulfur	34-1	176604	180499	182796	179966	1.74	cps
Terbium	159-1	11125823	11010739	11003641	11046734	0.62	cps
Terbium	159-2	6651696	6614461	6657010	6641056	0.35	cps
Thallium	203-1	1520	1503	1397	1473	4.54	cps
Thallium	205-1	3631	3537	3474	3547	2.22	cps
Tin	118-1	4531	4741	4684	4652	2.34	cps
Titanium	47-1	423438	427087	432745	427757	1.10	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-07 Instrumnet Name : P7
Client Sample ID : MYE4M0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:11:17 DataFile Name : 070SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	22890	22796	22282	22656	1.44	cps
Vanadium	51-2	264754	263483	265538	264592	0.39	cps
Yttrium	89-1	13533317	13471924	13356131	13453790	0.67	cps
Yttrium	89-2	3294471	3304334	3315941	3304915	0.33	cps
Zinc	66-2	165023	165077	165914	165338	0.30	cps
Zirconium	90-1	107886	109810	111079	109591	1.47	cps
Zirconium	91-1	24474	24988	25405	24956	1.87	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08 Instrumnet Name : P7
Client Sample ID : MYE4M1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:14:26 DataFile Name : 071SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5166969	5107526	5100077	5124857	0.72	cps
Antimony	121-1	4344	4117	4064	4175	3.56	cps
Arsenic	75-2	26504	26731	26798	26677	0.58	cps
Barium	135-1	979128	990161	993367	987552	0.76	cps
Barium	137-1	1878405	1858650	1893246	1876767	0.92	cps
Beryllium	9-1	1350	1439	1295	1361	5.35	cps
Bismuth	209-1	6876597	6901717	6862362	6880225	0.29	cps
Bismuth	209-2	4805054	4728236	4802723	4778671	0.91	cps
Bromine	81-1	14667	14854	14711	14744	0.66	cps
Bromine	81-2	903	953	907	921	3.03	cps
Cadmium	108-1	457	387	493	446	12.16	cps
Cadmium	106-1	3534	3464	3390	3463	2.07	cps
Cadmium	111-1	6271	6448	6292	6337	1.52	cps
Calcium	43-1	616722	618707	617892	617774	0.16	cps
Calcium	44-1	9437634	9424167	9386250	9416017	0.28	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	195854	195944	197796	196531	0.56	cps
Cobalt	59-2	189902	188575	188700	189059	0.39	cps
Copper	63-2	907320	905145	899731	904065	0.43	cps
Dysprosium	156-1	53990	54783	55272	54681	1.18	cps
Dysprosium	156-2	41070	41625	42163	41619	1.31	cps
Erbium	164-1	39823	40662	40535	40340	1.12	cps
Erbium	164-2	24618	25286	24572	24825	1.61	cps
Gadolinium	160-1	48510	49805	49972	49429	1.62	cps
Gadolinium	160-2	32491	32838	32832	32720	0.61	cps
Holmium	165-1	10959471	10899994	10693676	10851047	1.29	cps
Holmium	165-2	6730802	6726112	6707698	6721537	0.18	cps
Indium	115-1	8261178	8137199	8098221	8165533	1.04	cps
Indium	115-2	2561630	2572005	2530438	2554691	0.85	cps
Iron	56-2	153055991	153274624	152530891	152953835	0.25	cps
Iron	57-2	3812508	3746234	3780015	3779586	0.88	cps
Iron	54-2	8427104	8346202	8275385	8349564	0.91	cps
Krypton	83-1	217	237	230	228	4.47	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08 Instrumnet Name : P7
Client Sample ID : MYE4M1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:14:26 DataFile Name : 071SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	866132	879452	888445	878010	1.28	cps
Lead	207-1	747637	766808	770731	761725	1.62	cps
Lead	208-1	3758243	3764091	3828801	3783711	1.03	cps
Lithium	6-1	1406066	1385234	1387128	1392809	0.83	cps
Magnesium	24-2	4890272	4878391	4885588	4884750	0.12	cps
Manganese	55-2	4838203	4850275	4823732	4837403	0.27	cps
Molybdenum	94-1	72518	71209	72046	71924	0.92	cps
Molybdenum	95-1	2197	2310	2190	2232	3.02	cps
Molybdenum	96-1	16496	16436	16606	16513	0.52	cps
Molybdenum	97-1	1283	1343	1253	1293	3.54	cps
Molybdenum	98-1	3314	3290	3310	3305	0.38	cps
Neodymium	150-1	103580	104893	103651	104041	0.71	cps
Neodymium	150-2	54582	54197	53970	54249	0.57	cps
Nickel	60-2	85482	84744	84654	84960	0.53	cps
Phosphorus	31-2	24269	25077	24182	24510	2.01	cps
Potassium	39-2	3025622	3043493	3063023	3044046	0.61	cps
Rhodium	103-1	7457239	7458309	7328108	7414552	1.01	cps
Rhodium	103-2	4397023	4351563	4335422	4361336	0.73	cps
Scandium	45-1	4500021	4471615	4491599	4487745	0.33	cps
Scandium	45-2	404804	403902	398887	402531	0.79	cps
Selenium	82-1	450	566	474	497	12.32	cps
Selenium	77-2	187	150	133	157	17.41	cps
Selenium	78-2	413	313	343	357	14.39	cps
Silicon	28-1	25094690	25230676	25119740	25148369	0.29	cps
Silver	107-1	5731	6155	6148	6011	4.04	cps
Silver	109-1	5661	5788	5608	5686	1.63	cps
Sodium	23-2	128975	130074	127936	128995	0.83	cps
Strontium	86-1	239071	242601	243948	241873	1.04	cps
Strontium	88-1	2163568	2199997	2196535	2186700	0.92	cps
Sulfur	34-1	158047	154459	150304	154270	2.51	cps
Terbium	159-1	11096418	11228815	10979287	11101507	1.12	cps
Terbium	159-2	6643359	6695014	6578954	6639109	0.88	cps
Thallium	203-1	2194	2280	2297	2257	2.46	cps
Thallium	205-1	5324	5118	5374	5272	2.58	cps
Tin	118-1	5678	5761	5728	5722	0.73	cps
Titanium	47-1	651761	661623	660241	657875	0.81	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08 Instrumnet Name : P7
Client Sample ID : MYE4M1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:14:26 DataFile Name : 071SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	19418	19678	19311	19469	0.97	cps
Vanadium	51-2	285241	284215	285776	285077	0.28	cps
Yttrium	89-1	13881023	13722975	13630966	13744988	0.92	cps
Yttrium	89-2	3304921	3311425	3331819	3316055	0.42	cps
Zinc	66-2	198876	197667	198231	198258	0.31	cps
Zirconium	90-1	189523	191920	191258	190900	0.65	cps
Zirconium	91-1	43474	43558	42906	43313	0.82	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-09 Instrumnet Name : P7
Client Sample ID : MYE4M1D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:17:31 DataFile Name : 072SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5018519	4984106	5021474	5008033	0.41	cps
Antimony	121-1	3934	4054	4214	4067	3.45	cps
Arsenic	75-2	25489	25211	25499	25400	0.64	cps
Barium	135-1	950397	955613	955583	953864	0.31	cps
Barium	137-1	1798418	1767604	1785391	1783804	0.87	cps
Beryllium	9-1	1195	1383	1456	1345	10.03	cps
Bismuth	209-1	6859552	6819655	6707874	6795694	1.16	cps
Bismuth	209-2	4777831	4694105	4788924	4753620	1.09	cps
Bromine	81-1	13526	13770	14254	13850	2.67	cps
Bromine	81-2	883	907	857	882	2.84	cps
Cadmium	108-1	387	453	450	430	8.74	cps
Cadmium	106-1	3347	3437	3297	3360	2.11	cps
Cadmium	111-1	6246	6159	6290	6232	1.07	cps
Calcium	43-1	595698	597596	600794	598029	0.43	cps
Calcium	44-1	9112321	9099948	9156979	9123082	0.33	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	190679	190395	189930	190335	0.20	cps
Cobalt	59-2	183972	182687	180154	182271	1.07	cps
Copper	63-2	877521	869781	872739	873347	0.45	cps
Dysprosium	156-1	51956	52464	52598	52339	0.65	cps
Dysprosium	156-2	40097	40097	39749	39981	0.50	cps
Erbium	164-1	38790	39375	38363	38843	1.31	cps
Erbium	164-2	24001	23770	24238	24003	0.97	cps
Gadolinium	160-1	47647	48641	48781	48356	1.28	cps
Gadolinium	160-2	31856	31823	31068	31582	1.41	cps
Holmium	165-1	10773702	10631759	10662374	10689279	0.70	cps
Holmium	165-2	6690865	6594749	6709223	6664946	0.92	cps
Indium	115-1	8260106	8139382	8018267	8139252	1.49	cps
Indium	115-2	2554144	2517584	2539741	2537156	0.73	cps
Iron	56-2	148321781	148131084	148188098	148213654	0.07	cps
Iron	57-2	3684898	3702088	3663798	3683595	0.52	cps
Iron	54-2	8046384	8141389	8031353	8073042	0.74	cps
Krypton	83-1	237	277	267	260	8.01	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-09 Instrumnet Name : P7
Client Sample ID : MYE4M1D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:17:31 DataFile Name : 072SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	845467	852755	849432	849218	0.43	cps
Lead	207-1	737932	737108	742239	739093	0.37	cps
Lead	208-1	3648440	3649657	3639546	3645881	0.15	cps
Lithium	6-1	1390843	1374866	1386987	1384232	0.60	cps
Magnesium	24-2	4746088	4751046	4729275	4742137	0.24	cps
Manganese	55-2	4645097	4632189	4681374	4652887	0.55	cps
Molybdenum	94-1	71848	72069	72022	71980	0.16	cps
Molybdenum	95-1	1963	2120	2194	2092	5.62	cps
Molybdenum	96-1	16056	15869	15795	15907	0.84	cps
Molybdenum	97-1	1183	1283	1230	1232	4.06	cps
Molybdenum	98-1	3254	3214	3070	3179	3.03	cps
Neodymium	150-1	99008	100260	102170	100479	1.58	cps
Neodymium	150-2	52701	52046	52571	52439	0.66	cps
Nickel	60-2	82753	82442	82797	82664	0.23	cps
Phosphorus	31-2	23758	23748	23715	23741	0.10	cps
Potassium	39-2	2928997	2949175	2994417	2957530	1.13	cps
Rhodium	103-1	7452607	7362408	7243681	7352899	1.43	cps
Rhodium	103-2	4303806	4309034	4289103	4300648	0.24	cps
Scandium	45-1	4503133	4428172	4372564	4434623	1.48	cps
Scandium	45-2	399837	396430	400511	398926	0.55	cps
Selenium	82-1	470	444	500	471	5.92	cps
Selenium	77-2	150	190	180	173	12.01	cps
Selenium	78-2	363	377	327	356	7.28	cps
Silicon	28-1	25024138	25096170	25122954	25081087	0.20	cps
Silver	107-1	5878	5835	5898	5870	0.55	cps
Silver	109-1	5324	5338	5618	5427	3.05	cps
Sodium	23-2	126891	126104	126477	126491	0.31	cps
Strontium	86-1	234445	235010	235074	234843	0.15	cps
Strontium	88-1	2103065	2115948	2116703	2111905	0.36	cps
Sulfur	34-1	146690	144637	144475	145267	0.85	cps
Terbium	159-1	11133315	10932645	10906583	10990848	1.13	cps
Terbium	159-2	6637951	6522848	6609786	6590195	0.91	cps
Thallium	203-1	2234	2280	2190	2235	2.01	cps
Thallium	205-1	5318	5241	5178	5246	1.34	cps
Tin	118-1	5548	5374	5468	5463	1.59	cps
Titanium	47-1	635618	634275	637461	635785	0.25	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-09 Instrumnet Name : P7
Client Sample ID : MYE4M1D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:17:31 DataFile Name : 072SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	18510	19334	19204	19016	2.33	cps
Vanadium	51-2	275877	277139	275923	276313	0.26	cps
Yttrium	89-1	13793220	13465419	13472608	13577082	1.38	cps
Yttrium	89-2	3307496	3328933	3281723	3306051	0.72	cps
Zinc	66-2	191413	191097	192358	191623	0.34	cps
Zirconium	90-1	192683	193081	193990	193251	0.35	cps
Zirconium	91-1	43354	44049	43310	43571	0.95	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08LX5 Instrumnet Name : P7
Client Sample ID : MYE4M1L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 21:20:38 DataFile Name : 073SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	978420	983682	971668	977923	0.62	cps
Antimony	121-1	847	740	753	780	7.45	cps
Arsenic	75-2	5341	5121	5168	5210	2.23	cps
Barium	135-1	185015	184313	183844	184391	0.32	cps
Barium	137-1	319337	322291	322130	321253	0.52	cps
Beryllium	9-1	217	328	278	274	20.31	cps
Bismuth	209-1	6812517	6835785	6711225	6786509	0.98	cps
Bismuth	209-2	4797291	4810714	4822182	4810062	0.26	cps
Bromine	81-1	11341	11274	11124	11247	0.99	cps
Bromine	81-2	307	270	197	258	21.73	cps
Cadmium	108-1	90	73	63	76	17.83	cps
Cadmium	106-1	2527	2677	2527	2577	3.36	cps
Cadmium	111-1	2513	2706	2548	2589	3.98	cps
Calcium	43-1	116009	118172	117141	117107	0.92	cps
Calcium	44-1	1782110	1785177	1805452	1790913	0.71	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	38329	38422	36902	37884	2.25	cps
Cobalt	59-2	35281	36544	35616	35814	1.83	cps
Copper	63-2	173776	173334	171278	172796	0.77	cps
Dysprosium	156-1	9770	9877	10144	9930	1.94	cps
Dysprosium	156-2	7632	7492	7662	7595	1.20	cps
Erbium	164-1	7492	7656	7339	7495	2.11	cps
Erbium	164-2	4621	4457	4597	4559	1.94	cps
Gadolinium	160-1	9116	9153	9343	9204	1.32	cps
Gadolinium	160-2	6342	6445	6038	6275	3.37	cps
Holmium	165-1	10668855	10515970	10678387	10621071	0.86	cps
Holmium	165-2	6580967	6597247	6545872	6574695	0.40	cps
Indium	115-1	8155228	8159989	8135783	8150333	0.16	cps
Indium	115-2	2575443	2597638	2547883	2573655	0.97	cps
Iron	56-2	28942876	29209979	28310368	28821074	1.60	cps
Iron	57-2	742839	740695	735470	739668	0.51	cps
Iron	54-2	1611431	1593697	1587292	1597473	0.78	cps
Krypton	83-1	183	210	217	203	8.67	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08LX5 Instrumnet Name : P7
Client Sample ID : MYE4M1L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 21:20:38 DataFile Name : 073SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	161726	162880	162465	162357	0.36	cps
Lead	207-1	142238	141444	142081	141921	0.30	cps
Lead	208-1	645785	646524	645912	646074	0.06	cps
Lithium	6-1	1350332	1378794	1360775	1363301	1.06	cps
Magnesium	24-2	935661	937881	933177	935573	0.25	cps
Manganese	55-2	919077	913124	909069	913757	0.55	cps
Molybdenum	94-1	13443	13510	13336	13430	0.65	cps
Molybdenum	95-1	613	497	557	556	10.50	cps
Molybdenum	96-1	3034	3354	3210	3199	5.01	cps
Molybdenum	97-1	270	323	330	308	10.68	cps
Molybdenum	98-1	817	800	850	822	3.10	cps
Neodymium	150-1	19116	19407	18993	19172	1.11	cps
Neodymium	150-2	10344	9984	10367	10232	2.10	cps
Nickel	60-2	16279	16382	16646	16436	1.15	cps
Phosphorus	31-2	4434	4644	4487	4522	2.41	cps
Potassium	39-2	640477	639358	637111	638982	0.27	cps
Rhodium	103-1	7579984	7580966	7445361	7535437	1.04	cps
Rhodium	103-2	4412234	4414568	4412427	4413076	0.03	cps
Scandium	45-1	4372368	4350632	4378380	4367127	0.33	cps
Scandium	45-2	392684	393055	391420	392386	0.22	cps
Selenium	82-1	449	412	383	415	7.91	cps
Selenium	77-2	33	17	23	24	34.31	cps
Selenium	78-2	203	220	273	232	15.75	cps
Silicon	28-1	4840329	4867046	5198947	4968774	4.02	cps
Silver	107-1	1237	1287	1163	1229	5.05	cps
Silver	109-1	1100	1080	1150	1110	3.25	cps
Sodium	23-2	36162	36667	37111	36647	1.30	cps
Strontium	86-1	44805	46333	45925	45687	1.73	cps
Strontium	88-1	394928	401663	399211	398601	0.86	cps
Sulfur	34-1	123209	124300	124959	124156	0.71	cps
Terbium	159-1	10813505	10884614	11018681	10905600	0.96	cps
Terbium	159-2	6588304	6578954	6596736	6587998	0.14	cps
Thallium	203-1	553	450	560	521	11.84	cps
Thallium	205-1	1190	1250	1343	1261	6.13	cps
Tin	118-1	2304	2294	2347	2315	1.22	cps
Titanium	47-1	125253	124876	125706	125278	0.33	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-08LX5

Instrumnet Name :

P7

Client Sample ID :

MYE4M1L

Dilution Factor :

5

Date & Time Acquired :

2024-11-02 21:20:38

DataFile Name :

073SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	3654	3611	3434	3566	3.27	cps
Vanadium	51-2	54003	53156	53588	53582	0.79	cps
Yttrium	89-1	13220808	13175276	13126647	13174244	0.36	cps
Yttrium	89-2	3206473	3190774	3213352	3203533	0.36	cps
Zinc	66-2	37941	37738	38286	37988	0.73	cps
Zirconium	90-1	36565	36549	37174	36763	0.97	cps
Zirconium	91-1	7899	8109	8322	8110	2.61	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-10 Instrumnet Name : P7
Client Sample ID : MYE4M1S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:23:49 DataFile Name : 074SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5190310	5181131	5145561	5172334	0.46	cps
Antimony	121-1	583353	586331	596946	588877	1.21	cps
Arsenic	75-2	37287	37100	37547	37311	0.60	cps
Barium	135-1	4202798	4230988	4292177	4241988	1.08	cps
Barium	137-1	7290625	7425938	7492231	7402931	1.39	cps
Beryllium	9-1	46191	46899	46397	46495	0.78	cps
Bismuth	209-1	6819884	6783671	6337025	6646860	4.05	cps
Bismuth	209-2	4761848	4718767	4859810	4780142	1.51	cps
Bromine	81-1	13113	12736	13877	13242	4.39	cps
Bromine	81-2	753	767	783	768	1.96	cps
Cadmium	108-1	5181	5171	5298	5217	1.35	cps
Cadmium	106-1	10020	9813	9837	9890	1.15	cps
Cadmium	111-1	79292	80156	81441	80296	1.35	cps
Calcium	43-1	587504	593424	599118	593349	0.98	cps
Calcium	44-1	8959559	8994126	9093802	9015829	0.77	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	681534	678319	672777	677543	0.65	cps
Cobalt	59-2	2092396	2061268	2066947	2073537	0.80	cps
Copper	63-2	1569269	1539873	1556222	1555122	0.95	cps
Dysprosium	156-1	51086	51972	52518	51859	1.39	cps
Dysprosium	156-2	39552	40003	40622	40059	1.34	cps
Erbium	164-1	38904	39148	39613	39222	0.92	cps
Erbium	164-2	24004	24168	24281	24151	0.58	cps
Gadolinium	160-1	47202	47340	48373	47638	1.34	cps
Gadolinium	160-2	31729	32140	32314	32061	0.94	cps
Holmium	165-1	10818920	10782299	9716544	10439255	6.00	cps
Holmium	165-2	6693706	6634267	6613488	6647154	0.63	cps
Indium	115-1	8135640	8002442	7260625	7799569	6.04	cps
Indium	115-2	2528364	2558864	2588320	2558516	1.17	cps
Iron	56-2	150881351	149835911	149661661	150126308	0.44	cps
Iron	57-2	3768453	3741174	3706918	3738848	0.82	cps
Iron	54-2	8245720	8222839	8086424	8184995	1.05	cps
Krypton	83-1	250	230	247	242	4.42	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-10 Instrumnet Name : P7
Client Sample ID : MYE4M1S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:23:49 DataFile Name : 074SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	927476	929113	949595	935395	1.32	cps
Lead	207-1	810829	809332	825336	815165	1.08	cps
Lead	208-1	3963903	4003361	4085779	4017681	1.55	cps
Lithium	6-1	1370645	1339682	1233692	1314673	5.46	cps
Magnesium	24-2	4686111	4658992	4682819	4675974	0.32	cps
Manganese	55-2	5543273	5522333	5508728	5524778	0.31	cps
Molybdenum	94-1	72558	74046	74187	73597	1.23	cps
Molybdenum	95-1	2217	2187	2157	2187	1.37	cps
Molybdenum	96-1	16743	16319	16580	16547	1.29	cps
Molybdenum	97-1	1247	1307	1263	1272	2.43	cps
Molybdenum	98-1	3317	3374	3097	3263	4.48	cps
Neodymium	150-1	98655	100119	100662	99812	1.04	cps
Neodymium	150-2	53695	52829	53578	53367	0.88	cps
Nickel	60-2	548525	544871	546101	546499	0.34	cps
Phosphorus	31-2	23665	23001	23291	23319	1.43	cps
Potassium	39-2	2955236	2929699	2935071	2940002	0.46	cps
Rhodium	103-1	7406464	7325178	6663907	7131849	5.71	cps
Rhodium	103-2	4343510	4373477	4314438	4343809	0.68	cps
Scandium	45-1	4468476	4424416	4059239	4317377	5.20	cps
Scandium	45-2	398708	396861	396695	397421	0.28	cps
Selenium	82-1	10370	10473	10640	10494	1.30	cps
Selenium	77-2	1013	1143	1117	1091	6.29	cps
Selenium	78-2	3607	3560	3330	3499	4.23	cps
Silicon	28-1	24971022	25158175	25285965	25138387	0.63	cps
Silver	107-1	312571	314031	320098	315566	1.26	cps
Silver	109-1	296015	298774	300866	298552	0.82	cps
Sodium	23-2	124455	123722	125110	124429	0.56	cps
Strontium	86-1	232716	233716	237493	234642	1.07	cps
Strontium	88-1	2114623	2093855	2160486	2122988	1.61	cps
Sulfur	34-1	142365	138639	139607	140204	1.38	cps
Terbium	159-1	11067971	11126555	10007915	10734147	5.87	cps
Terbium	159-2	6565941	6617578	6613766	6599095	0.44	cps
Thallium	203-1	270814	273946	276003	273588	0.96	cps
Thallium	205-1	646917	657374	665310	656534	1.41	cps
Tin	118-1	4948	5238	5268	5151	3.43	cps
Titanium	47-1	617166	627915	633521	626201	1.33	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-10 Instrumnet Name : P7
Client Sample ID : MYE4M1S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:23:49 DataFile Name : 074SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	18416	19004	19254	18891	2.28	cps
Vanadium	51-2	1344775	1327297	1339737	1337270	0.67	cps
Yttrium	89-1	13657864	13634993	12421961	13238273	5.34	cps
Yttrium	89-2	3319201	3303845	3323244	3315430	0.31	cps
Zinc	66-2	396506	398619	396161	397095	0.34	cps
Zirconium	90-1	191646	194621	196837	194368	1.34	cps
Zirconium	91-1	43471	44852	44026	44116	1.58	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08MS Instrumnet Name : P7
Client Sample ID : MYE4M1SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:26:51 DataFile Name : 075SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5926499	5912472	5866389	5901787	0.53	cps
Antimony	121-1	3029193	3069742	3084789	3061241	0.94	cps
Arsenic	75-2	79866	80329	80748	80314	0.55	cps
Barium	135-1	16493378	16906559	16713096	16704344	1.24	cps
Barium	137-1	28647760	28768112	28941770	28785880	0.51	cps
Beryllium	9-1	216660	218176	217295	217377	0.35	cps
Bismuth	209-1	6959553	6861196	6833412	6884720	0.96	cps
Bismuth	209-2	4785012	4809570	4867240	4820607	0.88	cps
Bromine	81-1	12942	13159	13126	13076	0.89	cps
Bromine	81-2	807	860	833	833	3.20	cps
Cadmium	108-1	23669	24407	24264	24113	1.62	cps
Cadmium	106-1	35777	35233	36018	35676	1.13	cps
Cadmium	111-1	366692	365682	365856	366077	0.15	cps
Calcium	43-1	563397	563368	565559	564108	0.22	cps
Calcium	44-1	8502988	8641770	8661464	8602074	1.00	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2514665	2504342	2486739	2501915	0.56	cps
Cobalt	59-2	9197373	9262687	9215590	9225217	0.37	cps
Copper	63-2	3961608	4017905	3969414	3982975	0.77	cps
Dysprosium	156-1	49447	49845	50233	49841	0.79	cps
Dysprosium	156-2	38904	38824	38262	38663	0.90	cps
Erbium	164-1	37179	37460	37400	37346	0.40	cps
Erbium	164-2	23383	23256	23276	23305	0.29	cps
Gadolinium	160-1	46086	46597	46008	46230	0.69	cps
Gadolinium	160-2	30323	31091	30593	30669	1.27	cps
Holmium	165-1	10819167	10832016	10898717	10849967	0.39	cps
Holmium	165-2	6714903	6729716	6689130	6711250	0.31	cps
Indium	115-1	8055926	8066627	8045008	8055854	0.13	cps
Indium	115-2	2530313	2510230	2549412	2529985	0.77	cps
Iron	56-2	151387514	152594054	151457398	151812989	0.45	cps
Iron	57-2	3778021	3746433	3785566	3770007	0.55	cps
Iron	54-2	8392906	8336389	8324538	8351278	0.44	cps
Krypton	83-1	197	217	200	204	5.24	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08MS Instrumnet Name : P7
Client Sample ID : MYE4M1SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:26:51 DataFile Name : 075SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1250811	1266631	1265833	1261092	0.71	cps
Lead	207-1	1086717	1092056	1108285	1095686	1.03	cps
Lead	208-1	5379238	5409672	5468218	5419043	0.83	cps
Lithium	6-1	1352648	1342053	1351804	1348835	0.44	cps
Magnesium	24-2	4482705	4471826	4482237	4478923	0.14	cps
Manganese	55-2	8728471	8636738	8599253	8654821	0.77	cps
Molybdenum	94-1	70652	72136	71922	71570	1.12	cps
Molybdenum	95-1	2144	2327	2247	2239	4.11	cps
Molybdenum	96-1	16519	16643	16259	16474	1.19	cps
Molybdenum	97-1	1307	1240	1273	1273	2.62	cps
Molybdenum	98-1	3577	3641	3547	3588	1.33	cps
Neodymium	150-1	98115	97427	95974	97172	1.13	cps
Neodymium	150-2	51755	51012	51437	51401	0.72	cps
Nickel	60-2	2434746	2446339	2411012	2430699	0.74	cps
Phosphorus	31-2	22246	22156	22410	22271	0.58	cps
Potassium	39-2	2805675	2832669	2810345	2816230	0.51	cps
Rhodium	103-1	7384638	7341144	7332792	7352858	0.38	cps
Rhodium	103-2	4290314	4342984	4306982	4313426	0.62	cps
Scandium	45-1	4417423	4411565	4350156	4393048	0.85	cps
Scandium	45-2	398329	396016	397273	397206	0.29	cps
Selenium	82-1	48293	48018	48205	48172	0.29	cps
Selenium	77-2	4818	4517	4697	4677	3.23	cps
Selenium	78-2	15131	14971	14881	14994	0.85	cps
Silicon	28-1	24250655	24672853	24320928	24414812	0.93	cps
Silver	107-1	1497114	1589143	1489846	1525368	3.63	cps
Silver	109-1	1417145	1431929	1424696	1424590	0.52	cps
Sodium	23-2	122056	120359	120890	121101	0.72	cps
Strontium	86-1	221566	224228	227427	224407	1.31	cps
Strontium	88-1	2011506	2039957	2026567	2026010	0.70	cps
Sulfur	34-1	136581	133736	130831	133716	2.15	cps
Terbium	159-1	11083274	10936363	11004223	11007954	0.67	cps
Terbium	159-2	6650532	6597148	6591027	6612903	0.49	cps
Thallium	203-1	1336198	1349008	1351929	1345712	0.62	cps
Thallium	205-1	3758468	3728649	3765991	3751036	0.53	cps
Tin	118-1	5101	5254	5428	5261	3.11	cps
Titanium	47-1	591948	592769	589861	591526	0.25	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08MS Instrumnet Name : P7
Client Sample ID : MYE4M1SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:26:51 DataFile Name : 075SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	18026	18510	18800	18445	2.12	cps
Vanadium	51-2	5283497	5297885	5245782	5275721	0.51	cps
Yttrium	89-1	13339747	13559025	13413970	13437581	0.83	cps
Yttrium	89-2	3302019	3233811	3288684	3274838	1.10	cps
Zinc	66-2	1183024	1179938	1178602	1180521	0.19	cps
Zirconium	90-1	191501	193166	192051	192239	0.44	cps
Zirconium	91-1	43019	43047	43264	43110	0.31	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-11 Instrumnet Name : P7
Client Sample ID : MYE4M2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:29:46 DataFile Name : 076SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4357025	4303206	4352054	4337428	0.69	cps
Antimony	121-1	4901	4657	4911	4823	2.98	cps
Arsenic	75-2	13470	13423	13480	13457	0.23	cps
Barium	135-1	729387	739902	739180	736156	0.80	cps
Barium	137-1	1256263	1278727	1278511	1271167	1.02	cps
Beryllium	9-1	1111	1089	1172	1124	3.84	cps
Bismuth	209-1	7040970	6942758	7072583	7018771	0.96	cps
Bismuth	209-2	4842036	4895378	4843930	4860448	0.62	cps
Bromine	81-1	30328	32198	33384	31970	4.82	cps
Bromine	81-2	4357	4264	4177	4266	2.11	cps
Cadmium	108-1	247	180	270	232	20.11	cps
Cadmium	106-1	3117	3094	3137	3116	0.70	cps
Cadmium	111-1	4196	4070	4180	4149	1.66	cps
Calcium	43-1	446534	451407	457148	451696	1.18	cps
Calcium	44-1	6824188	6891930	6871682	6862600	0.51	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	186849	189184	187119	187717	0.68	cps
Cobalt	59-2	126104	124380	124360	124948	0.80	cps
Copper	63-2	303971	303491	304181	303881	0.12	cps
Dysprosium	156-1	41327	40759	41297	41128	0.78	cps
Dysprosium	156-2	30927	31459	31208	31198	0.85	cps
Erbium	164-1	32240	32097	31642	31993	0.98	cps
Erbium	164-2	19761	19858	20338	19986	1.55	cps
Gadolinium	160-1	39398	39338	39201	39313	0.26	cps
Gadolinium	160-2	25957	26091	25967	26005	0.29	cps
Holmium	165-1	11041842	11008978	11019119	11023313	0.15	cps
Holmium	165-2	6874137	6894305	6677692	6815378	1.76	cps
Indium	115-1	8344270	8276472	8212960	8277901	0.79	cps
Indium	115-2	2605891	2580704	2573879	2586824	0.65	cps
Iron	56-2	106641108	105094168	105189982	105641753	0.82	cps
Iron	57-2	2627926	2626714	2630118	2628253	0.07	cps
Iron	54-2	5778810	5754439	5793232	5775494	0.34	cps
Krypton	83-1	263	237	213	238	10.52	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-11 Instrumnet Name : P7
Client Sample ID : MYE4M2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:29:46 DataFile Name : 076SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	152821	153028	155835	153894	1.09	cps
Lead	207-1	129840	132051	133128	131673	1.27	cps
Lead	208-1	597374	603722	611733	604276	1.19	cps
Lithium	6-1	1421575	1386240	1368037	1391951	1.96	cps
Magnesium	24-2	4830647	4774860	4790476	4798661	0.60	cps
Manganese	55-2	3605661	3567742	3634170	3602524	0.93	cps
Molybdenum	94-1	45504	45755	45962	45740	0.50	cps
Molybdenum	95-1	3534	3884	3617	3678	4.97	cps
Molybdenum	96-1	12626	12799	12632	12686	0.77	cps
Molybdenum	97-1	2057	2310	2277	2215	6.22	cps
Molybdenum	98-1	5651	5391	5524	5522	2.36	cps
Neodymium	150-1	75434	76526	77213	76391	1.17	cps
Neodymium	150-2	40030	39492	39325	39616	0.93	cps
Nickel	60-2	57821	57135	57409	57455	0.60	cps
Phosphorus	31-2	15558	15091	14797	15149	2.53	cps
Potassium	39-2	2067376	2072140	2075744	2071753	0.20	cps
Rhodium	103-1	7454256	7590248	7558789	7534431	0.94	cps
Rhodium	103-2	4402339	4413936	4351827	4389367	0.75	cps
Scandium	45-1	4496224	4462352	4445339	4467971	0.58	cps
Scandium	45-2	405135	399868	403665	402889	0.67	cps
Selenium	82-1	516	530	545	530	2.74	cps
Selenium	77-2	183	157	113	151	23.38	cps
Selenium	78-2	333	397	350	360	9.12	cps
Silicon	28-1	25442750	25596759	25607130	25548880	0.36	cps
Silver	107-1	2064	1950	2090	2035	3.65	cps
Silver	109-1	1863	1830	1740	1811	3.52	cps
Sodium	23-2	384589	380403	381382	382125	0.57	cps
Strontium	86-1	614315	622829	620484	619209	0.71	cps
Strontium	88-1	5607765	5621403	5663525	5630898	0.52	cps
Sulfur	34-1	162807	162523	162639	162656	0.09	cps
Terbium	159-1	11274124	11338304	11193307	11268578	0.64	cps
Terbium	159-2	6793687	6675317	6595928	6688311	1.49	cps
Thallium	203-1	2210	2227	2234	2224	0.54	cps
Thallium	205-1	5304	5268	5535	5369	2.69	cps
Tin	118-1	3250	3254	3290	3265	0.68	cps
Titanium	47-1	481711	486461	489729	485967	0.83	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-11 Instrumnet Name : P7
Client Sample ID : MYE4M2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:29:46 DataFile Name : 076SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	59117	60771	60319	60069	1.42	cps
Vanadium	51-2	255542	254661	256969	255724	0.46	cps
Yttrium	89-1	13794573	13713112	13623141	13710275	0.63	cps
Yttrium	89-2	3362027	3320538	3318745	3333770	0.73	cps
Zinc	66-2	143969	142515	143400	143295	0.51	cps
Zirconium	90-1	117005	118382	119246	118211	0.96	cps
Zirconium	91-1	26628	26781	27349	26919	1.41	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-12 Instrumnet Name : P7
Client Sample ID : MYE4M4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:32:54 DataFile Name : 077SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5005078	4985060	4957646	4982595	0.48	cps
Antimony	121-1	3300	3310	3404	3338	1.71	cps
Arsenic	75-2	25222	24781	25342	25115	1.18	cps
Barium	135-1	954559	958802	960222	957861	0.31	cps
Barium	137-1	1776145	1787549	1761860	1775185	0.73	cps
Beryllium	9-1	1178	1261	1272	1237	4.17	cps
Bismuth	209-1	6908522	6929210	6870769	6902834	0.43	cps
Bismuth	209-2	4814544	4800059	4845783	4820129	0.48	cps
Bromine	81-1	15949	16403	16966	16439	3.10	cps
Bromine	81-2	1270	1360	1163	1265	7.79	cps
Cadmium	108-1	293	240	237	257	12.39	cps
Cadmium	106-1	3127	3100	3157	3128	0.91	cps
Cadmium	111-1	3487	3372	3521	3460	2.25	cps
Calcium	43-1	267220	266314	266890	266808	0.17	cps
Calcium	44-1	4071716	4098976	4091911	4087534	0.35	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	221476	220360	221847	221228	0.35	cps
Cobalt	59-2	198482	197994	197174	197883	0.33	cps
Copper	63-2	528062	524917	530077	527685	0.49	cps
Dysprosium	156-1	53732	53371	53676	53593	0.36	cps
Dysprosium	156-2	40441	39910	40314	40222	0.69	cps
Erbium	164-1	39710	40275	40990	40325	1.59	cps
Erbium	164-2	25253	25166	25016	25145	0.48	cps
Gadolinium	160-1	50384	51116	51002	50834	0.78	cps
Gadolinium	160-2	33487	33904	33213	33535	1.04	cps
Holmium	165-1	11151664	10951808	10816498	10973323	1.54	cps
Holmium	165-2	6765954	6783626	6736368	6761983	0.35	cps
Indium	115-1	8338251	8214146	8121779	8224725	1.32	cps
Indium	115-2	2595467	2541444	2580657	2572523	1.09	cps
Iron	56-2	165994891	163251524	162884604	164043673	1.04	cps
Iron	57-2	4100312	4065841	4092568	4086240	0.44	cps
Iron	54-2	8972125	9068728	9011223	9017359	0.54	cps
Krypton	83-1	270	283	237	263	9.13	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-12 Instrumnet Name : P7
Client Sample ID : MYE4M4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:32:54 DataFile Name : 077SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	192578	190944	191976	191833	0.43	cps
Lead	207-1	164648	163608	165771	164675	0.66	cps
Lead	208-1	750227	756806	757365	754799	0.53	cps
Lithium	6-1	1373286	1387625	1380607	1380506	0.52	cps
Magnesium	24-2	5006973	4931263	4992216	4976817	0.81	cps
Manganese	55-2	7059199	7047746	7004348	7037098	0.41	cps
Molybdenum	94-1	59986	61395	60357	60579	1.21	cps
Molybdenum	95-1	4878	5028	4904	4936	1.62	cps
Molybdenum	96-1	17394	17300	17010	17235	1.16	cps
Molybdenum	97-1	2920	2964	2867	2917	1.66	cps
Molybdenum	98-1	7625	7449	7329	7468	2.00	cps
Neodymium	150-1	96165	96537	97115	96606	0.50	cps
Neodymium	150-2	50965	50968	50534	50822	0.49	cps
Nickel	60-2	81728	80149	80518	80798	1.02	cps
Phosphorus	31-2	18545	18728	18438	18570	0.79	cps
Potassium	39-2	2612506	2570720	2586228	2589818	0.82	cps
Rhodium	103-1	7480942	7521734	7456836	7486504	0.44	cps
Rhodium	103-2	4386671	4353897	4388829	4376466	0.45	cps
Scandium	45-1	4552294	4532716	4492972	4525994	0.67	cps
Scandium	45-2	405394	402393	405043	404277	0.41	cps
Selenium	82-1	481	434	566	493	13.55	cps
Selenium	77-2	200	180	147	176	15.35	cps
Selenium	78-2	370	323	320	338	8.28	cps
Silicon	28-1	14302049	14457914	14296290	14352084	0.64	cps
Silver	107-1	3357	3711	3520	3529	5.01	cps
Silver	109-1	3280	3257	3314	3284	0.87	cps
Sodium	23-2	80539	79795	79571	79969	0.63	cps
Strontium	86-1	169082	172691	171844	171206	1.10	cps
Strontium	88-1	1553641	1553387	1529376	1545468	0.90	cps
Sulfur	34-1	143677	145040	146895	145204	1.11	cps
Terbium	159-1	11316729	11180145	11040369	11179081	1.24	cps
Terbium	159-2	6689663	6739084	6776619	6735122	0.65	cps
Thallium	203-1	1923	1800	1803	1842	3.81	cps
Thallium	205-1	4047	4124	4334	4168	3.56	cps
Tin	118-1	3804	3804	3821	3809	0.25	cps
Titanium	47-1	515974	521909	519408	519097	0.57	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-12 Instrumnet Name : P7
Client Sample ID : MYE4M4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:32:54 DataFile Name : 077SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	23023	22552	22843	22806	1.04	cps
Vanadium	51-2	283997	282898	284832	283909	0.34	cps
Yttrium	89-1	13899476	13820232	13745324	13821678	0.56	cps
Yttrium	89-2	3374139	3305883	3318647	3332890	1.09	cps
Zinc	66-2	105771	104599	105643	105338	0.61	cps
Zirconium	90-1	156550	155880	156319	156250	0.22	cps
Zirconium	91-1	34611	35052	35309	34991	1.01	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-13 Instrumnet Name : P7
Client Sample ID : MYE4M5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:36:00 DataFile Name : 078SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4745546	4736095	4719863	4733835	0.27	cps
Antimony	121-1	2507	2700	2567	2591	3.82	cps
Arsenic	75-2	17761	17247	16953	17320	2.36	cps
Barium	135-1	806440	816984	823151	815525	1.04	cps
Barium	137-1	1392997	1420705	1415248	1409650	1.04	cps
Beryllium	9-1	1145	1245	1333	1241	7.62	cps
Bismuth	209-1	6933372	6869521	6829416	6877436	0.76	cps
Bismuth	209-2	4783513	4851215	4785948	4806892	0.80	cps
Bromine	81-1	30168	35486	37631	34428	11.16	cps
Bromine	81-2	5198	5111	5508	5272	3.96	cps
Cadmium	108-1	183	187	243	204	16.49	cps
Cadmium	106-1	3020	2944	2787	2917	4.08	cps
Cadmium	111-1	3399	3387	3224	3337	2.93	cps
Calcium	43-1	334926	341281	343001	339736	1.25	cps
Calcium	44-1	5160043	5195285	5199459	5184929	0.42	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	184983	182018	182547	183183	0.86	cps
Cobalt	59-2	138078	139111	136679	137956	0.88	cps
Copper	63-2	260665	258906	258782	259451	0.41	cps
Dysprosium	156-1	44203	44614	44360	44392	0.47	cps
Dysprosium	156-2	33670	34546	33583	33933	1.57	cps
Erbium	164-1	33062	34152	33737	33651	1.63	cps
Erbium	164-2	20502	20719	21023	20748	1.26	cps
Gadolinium	160-1	40646	42143	41933	41574	1.95	cps
Gadolinium	160-2	28232	27601	28035	27956	1.16	cps
Holmium	165-1	10935699	10839374	10784454	10853176	0.71	cps
Holmium	165-2	6738904	6731012	6667113	6712343	0.59	cps
Indium	115-1	8196642	8106124	8067346	8123370	0.82	cps
Indium	115-2	2554215	2543249	2531361	2542942	0.45	cps
Iron	56-2	118200962	118651712	117802512	118218395	0.36	cps
Iron	57-2	2922323	2939359	2936476	2932719	0.31	cps
Iron	54-2	6426777	6364769	6416722	6402756	0.52	cps
Krypton	83-1	153	257	217	209	24.94	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-13 Instrumnet Name : P7
Client Sample ID : MYE4M5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:36:00 DataFile Name : 078SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	143968	146585	146432	145661	1.01	cps
Lead	207-1	122133	124689	124972	123931	1.26	cps
Lead	208-1	562226	571430	572965	568873	1.02	cps
Lithium	6-1	1404551	1326740	1373877	1368389	2.86	cps
Magnesium	24-2	4737786	4764085	4788979	4763617	0.54	cps
Manganese	55-2	4241302	4269934	4302745	4271327	0.72	cps
Molybdenum	94-1	38614	39413	40760	39596	2.74	cps
Molybdenum	95-1	3504	3244	3517	3422	4.51	cps
Molybdenum	96-1	11334	11832	11164	11443	3.03	cps
Molybdenum	97-1	1910	1973	1970	1951	1.83	cps
Molybdenum	98-1	4754	4944	5001	4900	2.64	cps
Neodymium	150-1	81155	82425	82680	82087	1.00	cps
Neodymium	150-2	43514	42153	42471	42712	1.67	cps
Nickel	60-2	61104	60324	60846	60758	0.65	cps
Phosphorus	31-2	16793	16589	16502	16628	0.90	cps
Potassium	39-2	2038992	2030016	2037507	2035505	0.24	cps
Rhodium	103-1	7501365	7363900	7418114	7427793	0.93	cps
Rhodium	103-2	4447412	4356441	4334422	4379425	1.37	cps
Scandium	45-1	4384574	4370658	4411747	4388993	0.48	cps
Scandium	45-2	397650	401649	396936	398745	0.64	cps
Selenium	82-1	629	503	554	562	11.35	cps
Selenium	77-2	143	147	143	144	1.33	cps
Selenium	78-2	387	370	393	383	3.13	cps
Silicon	28-1	22916121	23296456	23166455	23126344	0.84	cps
Silver	107-1	4161	4497	4561	4406	4.88	cps
Silver	109-1	4007	4224	4241	4157	3.13	cps
Sodium	23-2	204860	206377	206877	206038	0.51	cps
Strontium	86-1	463291	473097	477175	471188	1.51	cps
Strontium	88-1	4216832	4230638	4316929	4254800	1.27	cps
Sulfur	34-1	155115	152950	152388	153484	0.94	cps
Terbium	159-1	11105300	11056930	11030662	11064297	0.34	cps
Terbium	159-2	6712045	6672876	6543390	6642770	1.33	cps
Thallium	203-1	7852	8012	8356	8074	3.19	cps
Thallium	205-1	19087	19694	19517	19433	1.61	cps
Tin	118-1	6328	7207	6672	6736	6.58	cps
Titanium	47-1	503362	514621	514213	510732	1.25	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-13 Instrumnet Name : P7
Client Sample ID : MYE4M5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:36:00 DataFile Name : 078SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	21795	22683	22102	22193	2.03	cps
Vanadium	51-2	245771	244527	245193	245164	0.25	cps
Yttrium	89-1	13591597	13535094	13648442	13591711	0.42	cps
Yttrium	89-2	3301791	3274534	3259634	3278653	0.65	cps
Zinc	66-2	137486	139201	136971	137886	0.85	cps
Zirconium	90-1	101712	103391	103901	103001	1.11	cps
Zirconium	91-1	22955	23399	23502	23285	1.25	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-14 Instrumnet Name : P7
Client Sample ID : MYE4M6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:39:09 DataFile Name : 079SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7369272	7286597	7290927	7315599	0.64	cps
Antimony	121-1	1407	1483	1417	1436	2.90	cps
Arsenic	75-2	13853	14427	14444	14241	2.36	cps
Barium	135-1	1196793	1220031	1218603	1211809	1.07	cps
Barium	137-1	2270623	2277280	2287068	2278324	0.36	cps
Beryllium	9-1	2050	1933	1745	1909	8.08	cps
Bismuth	209-1	6875619	6923361	6765856	6854945	1.18	cps
Bismuth	209-2	4847781	4784158	4890570	4840836	1.11	cps
Bromine	81-1	22033	24093	24838	23655	6.14	cps
Bromine	81-2	2884	2867	3007	2919	2.62	cps
Cadmium	108-1	193	150	133	159	19.49	cps
Cadmium	106-1	3307	2847	2787	2980	9.55	cps
Cadmium	111-1	2863	2685	2558	2702	5.68	cps
Calcium	43-1	284418	285581	286600	285533	0.38	cps
Calcium	44-1	4326815	4332988	4301055	4320286	0.39	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	215345	214361	215549	215085	0.30	cps
Cobalt	59-2	194202	191255	190996	192151	0.93	cps
Copper	63-2	279792	281039	277480	279437	0.65	cps
Dysprosium	156-1	70566	71514	70506	70862	0.80	cps
Dysprosium	156-2	53622	53655	52859	53379	0.84	cps
Erbium	164-1	53458	53100	53599	53386	0.48	cps
Erbium	164-2	33370	32889	33153	33137	0.73	cps
Gadolinium	160-1	67321	66942	66982	67082	0.31	cps
Gadolinium	160-2	45025	44363	44190	44526	0.99	cps
Holmium	165-1	10860000	10983681	10761216	10868299	1.03	cps
Holmium	165-2	6789669	6709966	6734119	6744585	0.61	cps
Indium	115-1	8190914	8268438	8237886	8232413	0.47	cps
Indium	115-2	2554049	2554980	2549101	2552710	0.12	cps
Iron	56-2	161969221	162228768	161788278	161995422	0.14	cps
Iron	57-2	4050461	4042456	3982547	4025155	0.92	cps
Iron	54-2	8852353	8779148	8885883	8839128	0.62	cps
Krypton	83-1	277	147	273	232	31.91	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-14 Instrumnet Name : P7
Client Sample ID : MYE4M6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:39:09 DataFile Name : 079SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	169563	172607	171541	171237	0.90	cps
Lead	207-1	142960	145347	145644	144650	1.02	cps
Lead	208-1	660539	669578	670667	666928	0.83	cps
Lithium	6-1	1361446	1383924	1376641	1374004	0.83	cps
Magnesium	24-2	5655346	5730053	5650388	5678596	0.79	cps
Manganese	55-2	5189964	5103881	5113678	5135841	0.92	cps
Molybdenum	94-1	62888	63574	64819	63760	1.54	cps
Molybdenum	95-1	2130	1990	1973	2031	4.24	cps
Molybdenum	96-1	14908	15131	14678	14905	1.52	cps
Molybdenum	97-1	1163	1207	1097	1156	4.80	cps
Molybdenum	98-1	2724	2957	2894	2858	4.22	cps
Neodymium	150-1	130033	129145	132213	130463	1.21	cps
Neodymium	150-2	68482	68927	68006	68472	0.67	cps
Nickel	60-2	87329	87004	87949	87427	0.55	cps
Phosphorus	31-2	14751	15148	14871	14923	1.36	cps
Potassium	39-2	2979827	2983431	2979773	2981010	0.07	cps
Rhodium	103-1	7478924	7404341	7324827	7402697	1.04	cps
Rhodium	103-2	4452377	4342423	4429902	4408234	1.32	cps
Scandium	45-1	4555055	4473829	4490464	4506449	0.95	cps
Scandium	45-2	407772	401291	405124	404729	0.81	cps
Selenium	82-1	572	622	465	553	14.54	cps
Selenium	77-2	187	263	223	224	17.08	cps
Selenium	78-2	420	393	363	392	7.23	cps
Silicon	28-1	22845595	23081984	23121434	23016338	0.65	cps
Silver	107-1	1143	1217	1277	1212	5.51	cps
Silver	109-1	1070	1187	1163	1140	5.42	cps
Sodium	23-2	78036	79035	79420	78830	0.91	cps
Strontium	86-1	242066	246605	246240	244970	1.03	cps
Strontium	88-1	2195244	2228408	2244092	2222581	1.12	cps
Sulfur	34-1	126045	123709	123659	124471	1.10	cps
Terbium	159-1	11176521	11066844	11102577	11115314	0.50	cps
Terbium	159-2	6746294	6581658	6593074	6640342	1.38	cps
Thallium	203-1	2210	2277	2177	2221	2.29	cps
Thallium	205-1	5344	5504	5354	5401	1.66	cps
Tin	118-1	3661	3497	3520	3559	2.48	cps
Titanium	47-1	742270	746782	750021	746358	0.52	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-14 Instrumnet Name : P7
Client Sample ID : MYE4M6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:39:09 DataFile Name : 079SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	21527	21711	22189	21809	1.56	cps
Vanadium	51-2	303009	300757	302630	302132	0.40	cps
Yttrium	89-1	13877827	13910512	13859536	13882625	0.19	cps
Yttrium	89-2	3395234	3361195	3410297	3388909	0.74	cps
Zinc	66-2	87977	87135	88078	87730	0.59	cps
Zirconium	90-1	167347	171390	171714	170151	1.43	cps
Zirconium	91-1	37839	37865	38240	37981	0.59	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-15 Instrumnet Name : P7
Client Sample ID : MYE4M7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:42:15 DataFile Name : 080SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5540375	5652410	5549131	5580638	1.12	cps
Antimony	121-1	3687	3737	3804	3743	1.56	cps
Arsenic	75-2	19069	18819	19086	18991	0.79	cps
Barium	135-1	925275	941732	937281	934763	0.91	cps
Barium	137-1	1727433	1718810	1737552	1727931	0.54	cps
Beryllium	9-1	1483	1517	1417	1472	3.46	cps
Bismuth	209-1	6890062	6964070	6883111	6912414	0.65	cps
Bismuth	209-2	4837833	4876668	4806968	4840490	0.72	cps
Bromine	81-1	15992	16893	16913	16599	3.17	cps
Bromine	81-2	1323	1230	1437	1330	7.78	cps
Cadmium	108-1	307	310	347	321	6.91	cps
Cadmium	106-1	3170	3090	2884	3048	4.85	cps
Cadmium	111-1	4602	4883	4590	4691	3.54	cps
Calcium	43-1	253642	255699	256465	255269	0.57	cps
Calcium	44-1	3872973	3923831	3910336	3902380	0.68	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	207636	208056	209532	208408	0.48	cps
Cobalt	59-2	198562	198549	198339	198483	0.06	cps
Copper	63-2	610032	614013	612501	612182	0.33	cps
Dysprosium	156-1	56778	57705	57852	57445	1.01	cps
Dysprosium	156-2	43423	43119	43256	43266	0.35	cps
Erbium	164-1	43698	44096	44969	44254	1.47	cps
Erbium	164-2	26939	27340	27624	27301	1.26	cps
Gadolinium	160-1	53073	53843	54653	53856	1.47	cps
Gadolinium	160-2	34950	35461	35522	35311	0.89	cps
Holmium	165-1	10714550	10939493	10825744	10826596	1.04	cps
Holmium	165-2	6685737	6684479	6632875	6667697	0.45	cps
Indium	115-1	8173128	8186245	8006869	8122081	1.23	cps
Indium	115-2	2556291	2569440	2585361	2570364	0.57	cps
Iron	56-2	152182458	152503191	152307894	152331181	0.11	cps
Iron	57-2	3785228	3852591	3794705	3810841	0.96	cps
Iron	54-2	8304761	8333757	8242890	8293803	0.56	cps
Krypton	83-1	207	253	233	231	10.13	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-15 Instrumnet Name : P7
Client Sample ID : MYE4M7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:42:15 DataFile Name : 080SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	383073	391997	386793	387288	1.16	cps
Lead	207-1	335447	337154	334297	335633	0.43	cps
Lead	208-1	1510450	1534472	1523514	1522812	0.79	cps
Lithium	6-1	1357814	1342605	1365656	1355358	0.86	cps
Magnesium	24-2	5001755	5037254	4942116	4993708	0.96	cps
Manganese	55-2	5212143	5237123	5229833	5226366	0.25	cps
Molybdenum	94-1	63781	64401	63755	63979	0.57	cps
Molybdenum	95-1	2504	2570	2630	2568	2.47	cps
Molybdenum	96-1	15208	14951	15465	15208	1.69	cps
Molybdenum	97-1	1390	1443	1560	1465	5.94	cps
Molybdenum	98-1	3851	3784	3707	3781	1.90	cps
Neodymium	150-1	106871	107936	107744	107517	0.53	cps
Neodymium	150-2	56439	56600	55807	56282	0.75	cps
Nickel	60-2	87956	87577	88204	87912	0.36	cps
Phosphorus	31-2	20504	21355	20023	20627	3.27	cps
Potassium	39-2	3207643	3201029	3202946	3203873	0.11	cps
Rhodium	103-1	7415043	7368303	7333512	7372286	0.55	cps
Rhodium	103-2	4361090	4357318	4353371	4357260	0.09	cps
Scandium	45-1	4436258	4419875	4436987	4431040	0.22	cps
Scandium	45-2	399331	404044	399355	400910	0.68	cps
Selenium	82-1	600	539	507	549	8.52	cps
Selenium	77-2	180	173	187	180	3.70	cps
Selenium	78-2	360	343	367	357	3.37	cps
Silicon	28-1	28390574	28442900	28691810	28508428	0.56	cps
Silver	107-1	3444	3807	3654	3635	5.02	cps
Silver	109-1	3514	3464	3567	3515	1.47	cps
Sodium	23-2	69362	71508	70159	70343	1.54	cps
Strontium	86-1	176136	178923	180032	178364	1.13	cps
Strontium	88-1	1582337	1639761	1640506	1620868	2.06	cps
Sulfur	34-1	123618	121213	121044	121958	1.18	cps
Terbium	159-1	11180070	11035393	11019264	11078242	0.80	cps
Terbium	159-2	6773164	6739722	6739591	6750826	0.29	cps
Thallium	203-1	3060	3030	3077	3056	0.77	cps
Thallium	205-1	7372	7372	7229	7324	1.13	cps
Tin	118-1	4247	4181	4324	4251	1.69	cps
Titanium	47-1	668463	674596	680349	674469	0.88	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-15 Instrumnet Name : P7
Client Sample ID : MYE4M7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:42:15 DataFile Name : 080SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	22028	21768	22990	22262	2.89	cps
Vanadium	51-2	297620	297838	296698	297385	0.20	cps
Yttrium	89-1	13603807	13639720	13638481	13627336	0.15	cps
Yttrium	89-2	3333498	3338014	3333645	3335052	0.08	cps
Zinc	66-2	130252	131133	130618	130668	0.34	cps
Zirconium	90-1	168296	170062	171147	169835	0.85	cps
Zirconium	91-1	37515	37949	38567	38010	1.39	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV003 Instrumnet Name : P7
Client Sample ID : CCV003 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:45:23 DataFile Name : 081CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	10700793	10591440	10798080	10696771	0.97	cps
Antimony	121-1	4914221	4950635	4979690	4948182	0.66	cps
Arsenic	75-2	244596	247209	251109	247638	1.32	cps
Barium	135-1	6797388	6747953	6794053	6779798	0.41	cps
Barium	137-1	11696079	11610993	11636566	11647880	0.37	cps
Beryllium	9-1	719824	724093	732330	725416	0.88	cps
Bismuth	209-1	6217335	6359856	6319897	6299029	1.17	cps
Bismuth	209-2	4259134	4317657	4388377	4321723	1.50	cps
Bromine	81-1	11451	10607	10290	10783	5.56	cps
Bromine	81-2	160	180	170	170	5.88	cps
Cadmium	108-1	90205	90298	90563	90356	0.21	cps
Cadmium	106-1	130130	129803	129091	129675	0.41	cps
Cadmium	111-1	1216593	1226502	1218032	1220376	0.44	cps
Calcium	43-1	6390081	6358864	6394660	6381202	0.31	cps
Calcium	44-1	102110468	101288598	101640828	101679965	0.41	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1935804	1945486	1940543	1940611	0.25	cps
Cobalt	59-2	3127627	3168142	3186441	3160737	0.95	cps
Copper	63-2	22123340	21977832	22136478	22079217	0.40	cps
Dysprosium	156-1	277	300	307	294	5.35	cps
Dysprosium	156-2	313	390	373	359	11.24	cps
Erbium	164-1	357	390	303	350	12.49	cps
Erbium	164-2	213	200	160	191	14.52	cps
Gadolinium	160-1	237	313	303	284	14.65	cps
Gadolinium	160-2	367	390	433	397	8.53	cps
Holmium	165-1	10601144	10658123	10584197	10614488	0.36	cps
Holmium	165-2	6444533	6428882	6558282	6477232	1.09	cps
Indium	115-1	7676228	7620649	7599986	7632288	0.52	cps
Indium	115-2	2354330	2360703	2376854	2363962	0.49	cps
Iron	56-2	418812154	418279367	421764500	419618674	0.45	cps
Iron	57-2	10494988	10334303	10592267	10473853	1.24	cps
Iron	54-2	23012309	22892415	22991833	22965519	0.28	cps
Krypton	83-1	190	223	210	208	8.08	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV003 Instrumnet Name : P7
Client Sample ID : CCV003 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:45:23 DataFile Name : 081CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	20048155	20681948	20550431	20426845	1.64	cps
Lead	207-1	17638189	18095664	18173110	17968988	1.61	cps
Lead	208-1	79720360	82385970	82368220	81491517	1.88	cps
Lithium	6-1	1344607	1337713	1338070	1340130	0.29	cps
Magnesium	24-2	91933982	91552495	92539122	92008533	0.54	cps
Manganese	55-2	14176839	14294108	14228439	14233129	0.41	cps
Molybdenum	94-1	18762022	18783945	18593395	18713121	0.56	cps
Molybdenum	95-1	27401785	27339554	27283204	27341514	0.22	cps
Molybdenum	96-1	29692145	29870022	29826724	29796297	0.31	cps
Molybdenum	97-1	17107639	17203238	17269444	17193440	0.47	cps
Molybdenum	98-1	44161099	44054179	44202514	44139264	0.17	cps
Neodymium	150-1	447	390	417	418	6.79	cps
Neodymium	150-2	140	147	153	147	4.55	cps
Nickel	60-2	754568	749641	754753	752987	0.39	cps
Phosphorus	31-2	111015	109850	112217	111028	1.07	cps
Potassium	39-2	54203034	53951786	54238088	54130969	0.29	cps
Rhodium	103-1	6832444	6790669	6772002	6798372	0.46	cps
Rhodium	103-2	3899394	3961363	3943081	3934612	0.81	cps
Scandium	45-1	4270276	4201509	4172668	4214818	1.19	cps
Scandium	45-2	376998	382180	382985	380721	0.85	cps
Selenium	82-1	75482	75574	75083	75380	0.35	cps
Selenium	77-2	7088	6862	6815	6922	2.11	cps
Selenium	78-2	22911	23025	22925	22953	0.27	cps
Silicon	28-1	40862380	40672025	40617581	40717329	0.32	cps
Silver	107-1	6309291	6256485	6239976	6268584	0.58	cps
Silver	109-1	5873406	5962590	5793018	5876338	1.44	cps
Sodium	23-2	150978801	148088024	151364148	150143658	1.19	cps
Strontium	86-1	39569839	39397043	39298609	39421830	0.35	cps
Strontium	88-1	345775222	345439555	345098315	345437697	0.10	cps
Sulfur	34-1	524127	515181	513084	517464	1.13	cps
Terbium	159-1	10676020	10971194	10779538	10808917	1.39	cps
Terbium	159-2	6331980	6476506	6469334	6425940	1.27	cps
Thallium	203-1	5033460	5160710	5148341	5114170	1.37	cps
Thallium	205-1	12115217	12431889	12349829	12298978	1.34	cps
Tin	118-1	4177687	4220142	4148179	4182003	0.87	cps
Titanium	47-1	6505407	6470384	6454091	6476627	0.40	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV003 Instrumnet Name : P7
Client Sample ID : CCV003 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:45:23 DataFile Name : 081CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	16288011	16543704	16286099	16372605	0.91	cps
Vanadium	51-2	1754489	1738795	1744190	1745825	0.46	cps
Yttrium	89-1	12854404	12898827	12538911	12764047	1.54	cps
Yttrium	89-2	3023581	3058206	3099613	3060467	1.24	cps
Zinc	66-2	3451484	3456438	3472885	3460269	0.32	cps
Zirconium	90-1	8655969	8613413	8586108	8618497	0.41	cps
Zirconium	91-1	1913222	1929143	1929698	1924021	0.49	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB003 Instrumnet Name : P7
Client Sample ID : CCB003 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:48:04 DataFile Name : 082CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	967	1083	957	1002	7.03	cps
Antimony	121-1	3460	3497	3354	3437	2.17	cps
Arsenic	75-2	17	10	13	13	25.01	cps
Barium	135-1	460	420	417	432	5.58	cps
Barium	137-1	873	753	733	787	9.63	cps
Beryllium	9-1	94	78	78	83	11.55	cps
Bismuth	209-1	6878428	7025736	6910187	6938117	1.12	cps
Bismuth	209-2	4856022	4875398	4907220	4879547	0.53	cps
Bromine	81-1	10237	10180	10207	10208	0.28	cps
Bromine	81-2	110	117	53	93	37.29	cps
Cadmium	108-1	13	20	3	12	68.66	cps
Cadmium	106-1	2664	2670	2674	2669	0.19	cps
Cadmium	111-1	1979	1932	1953	1954	1.20	cps
Calcium	43-1	507	500	523	510	2.36	cps
Calcium	44-1	14407	13904	13433	13915	3.50	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	620	677	597	631	6.52	cps
Cobalt	59-2	213	197	183	198	7.60	cps
Copper	63-2	1410	1520	1303	1411	7.68	cps
Dysprosium	156-1	10	23	20	18	39.03	cps
Dysprosium	156-2	7	13	7	9	43.25	cps
Erbium	164-1	73	90	97	87	13.87	cps
Erbium	164-2	50	47	37	44	15.61	cps
Gadolinium	160-1	30	33	30	31	6.18	cps
Gadolinium	160-2	230	220	253	234	7.30	cps
Holmium	165-1	10671163	10758127	10740056	10723115	0.43	cps
Holmium	165-2	6698662	6699672	6608971	6669102	0.78	cps
Indium	115-1	8183154	8267301	8272845	8241100	0.61	cps
Indium	115-2	2591840	2612336	2622423	2608866	0.60	cps
Iron	56-2	42901	42891	42697	42830	0.27	cps
Iron	57-2	1933	1997	1923	1951	2.04	cps
Iron	54-2	3154	3330	3390	3292	3.74	cps
Krypton	83-1	227	207	187	207	9.68	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB003 Instrumnet Name : P7
Client Sample ID : CCB003 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:48:04 DataFile Name : 082CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1560	1567	1483	1537	3.01	cps
Lead	207-1	1477	1350	1200	1342	10.32	cps
Lead	208-1	6464	6071	5731	6088	6.03	cps
Lithium	6-1	1386619	1379558	1340018	1368732	1.83	cps
Magnesium	24-2	3244	2967	2904	3038	5.95	cps
Manganese	55-2	1107	973	1117	1066	7.51	cps
Molybdenum	94-1	1653	1407	1380	1480	10.18	cps
Molybdenum	95-1	1593	1630	1340	1521	10.38	cps
Molybdenum	96-1	1817	1773	1737	1776	2.26	cps
Molybdenum	97-1	1040	950	783	924	14.09	cps
Molybdenum	98-1	2580	2280	2290	2384	7.15	cps
Neodymium	150-1	20	10	30	20	50.00	cps
Neodymium	150-2	7	7	7	7	0.00	cps
Nickel	60-2	157	103	160	140	22.71	cps
Phosphorus	31-2	163	110	143	139	19.40	cps
Potassium	39-2	50005	49310	49697	49671	0.70	cps
Rhodium	103-1	7625303	7679815	7609933	7638350	0.48	cps
Rhodium	103-2	4484607	4467992	4526698	4493099	0.67	cps
Scandium	45-1	4367996	4437981	4385303	4397093	0.83	cps
Scandium	45-2	394005	390196	393448	392550	0.52	cps
Selenium	82-1	338	377	443	386	13.80	cps
Selenium	77-2	3	0	0	1	173.21	cps
Selenium	78-2	290	237	227	251	13.56	cps
Silicon	28-1	213139	210141	210024	211101	0.84	cps
Silver	107-1	760	623	653	679	10.58	cps
Silver	109-1	617	560	450	542	15.63	cps
Sodium	23-2	26974	25444	26199	26206	2.92	cps
Strontium	86-1	1600	1503	1280	1461	11.23	cps
Strontium	88-1	11311	9530	8803	9881	13.06	cps
Sulfur	34-1	158752	161078	160182	160004	0.73	cps
Terbium	159-1	10969561	10995773	10808172	10924502	0.93	cps
Terbium	159-2	6588305	6579479	6672726	6613503	0.78	cps
Thallium	203-1	773	630	643	682	11.61	cps
Thallium	205-1	1757	1540	1377	1558	12.24	cps
Tin	118-1	2277	2624	2660	2520	8.39	cps
Titanium	47-1	447	420	400	422	5.55	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB003 Instrumnet Name : P7
Client Sample ID : CCB003 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:48:04 DataFile Name : 082CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	427	400	400	409	3.77	cps
Vanadium	51-2	97	97	123	106	14.58	cps
Yttrium	89-1	13209429	13221150	13236229	13222269	0.10	cps
Yttrium	89-2	3253357	3211575	3228269	3231067	0.65	cps
Zinc	66-2	673	670	683	676	1.03	cps
Zirconium	90-1	1680	1600	1490	1590	6.00	cps
Zirconium	91-1	387	443	400	410	7.23	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-16 Instrumnet Name : P7
Client Sample ID : MYE4M8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:51:21 DataFile Name : 083SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8028859	7942222	7876120	7949067	0.96	cps
Antimony	121-1	3697	4244	4027	3990	6.90	cps
Arsenic	75-2	66578	67000	65494	66357	1.17	cps
Barium	135-1	2894641	2953141	2972653	2940145	1.38	cps
Barium	137-1	4991093	5091454	5135692	5072746	1.46	cps
Beryllium	9-1	1533	1811	1695	1680	8.30	cps
Bismuth	209-1	6920323	6967025	6963008	6950119	0.37	cps
Bismuth	209-2	4872089	4951237	4969843	4931056	1.05	cps
Bromine	81-1	13089	13917	14767	13924	6.03	cps
Bromine	81-2	787	763	903	818	9.17	cps
Cadmium	108-1	350	247	317	304	17.32	cps
Cadmium	106-1	3831	3537	3447	3605	5.56	cps
Cadmium	111-1	4460	4237	4218	4305	3.12	cps
Calcium	43-1	488860	497349	503583	496598	1.49	cps
Calcium	44-1	7395968	7581230	7787848	7588349	2.58	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	327020	330107	328822	328650	0.47	cps
Cobalt	59-2	283424	282158	282539	282707	0.23	cps
Copper	63-2	503500	501139	503341	502660	0.26	cps
Dysprosium	156-1	73223	75445	74824	74497	1.54	cps
Dysprosium	156-2	57962	58009	58307	58093	0.32	cps
Erbium	164-1	53345	52187	53652	53061	1.46	cps
Erbium	164-2	32661	32408	32371	32480	0.49	cps
Gadolinium	160-1	67531	67893	68262	67896	0.54	cps
Gadolinium	160-2	45313	45376	45567	45419	0.29	cps
Holmium	165-1	11382441	11133612	11245644	11253899	1.11	cps
Holmium	165-2	6892266	6940705	6924482	6919151	0.36	cps
Indium	115-1	8516161	8611356	8665208	8597575	0.88	cps
Indium	115-2	2686590	2740232	2751800	2726207	1.28	cps
Iron	56-2	236085096	234970237	233122643	234725992	0.64	cps
Iron	57-2	5798613	5830983	5776716	5802104	0.47	cps
Iron	54-2	12747626	12874047	12672429	12764701	0.80	cps
Krypton	83-1	280	280	280	280	0.00	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-16 Instrumnet Name : P7
Client Sample ID : MYE4M8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:51:21 DataFile Name : 083SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	178850	179904	179411	179388	0.29	cps
Lead	207-1	150770	152320	151595	151562	0.51	cps
Lead	208-1	691329	703403	701592	698775	0.93	cps
Lithium	6-1	1382853	1407032	1423642	1404509	1.46	cps
Magnesium	24-2	7568517	7490699	7442447	7500554	0.85	cps
Manganese	55-2	24242446	24157707	24229152	24209769	0.19	cps
Molybdenum	94-1	88539	90196	91849	90194	1.84	cps
Molybdenum	95-1	4537	4657	4531	4575	1.56	cps
Molybdenum	96-1	21199	21810	22100	21703	2.12	cps
Molybdenum	97-1	2317	2274	2307	2299	0.99	cps
Molybdenum	98-1	5695	5701	5895	5763	1.97	cps
Neodymium	150-1	144212	143724	145963	144633	0.81	cps
Neodymium	150-2	75628	76422	76874	76308	0.83	cps
Nickel	60-2	170085	170397	169112	169865	0.39	cps
Phosphorus	31-2	25121	25982	25334	25479	1.76	cps
Potassium	39-2	1892819	1863355	1871980	1876051	0.81	cps
Rhodium	103-1	7782120	7830526	7855054	7822567	0.47	cps
Rhodium	103-2	4653565	4636424	4567437	4619142	0.99	cps
Scandium	45-1	4819569	4949745	4994295	4921203	1.84	cps
Scandium	45-2	445229	442190	439365	442262	0.66	cps
Selenium	82-1	498	510	632	547	13.61	cps
Selenium	77-2	263	240	253	252	4.64	cps
Selenium	78-2	467	500	423	463	8.30	cps
Silicon	28-1	28721300	29052481	29368995	29047592	1.11	cps
Silver	107-1	1430	1503	1507	1480	2.93	cps
Silver	109-1	1317	1467	1423	1402	5.51	cps
Sodium	23-2	113065	112631	111362	112353	0.79	cps
Strontium	86-1	352474	361922	369567	361321	2.37	cps
Strontium	88-1	3208605	3360217	3340993	3303271	2.50	cps
Sulfur	34-1	159024	157027	150304	155452	2.94	cps
Terbium	159-1	11430893	11353250	11542503	11442215	0.83	cps
Terbium	159-2	6938402	6896923	6919268	6918198	0.30	cps
Thallium	203-1	3077	3661	3057	3265	10.50	cps
Thallium	205-1	7469	8123	7429	7674	5.08	cps
Tin	118-1	3841	3804	4131	3925	4.56	cps
Titanium	47-1	727509	741412	758485	742469	2.09	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-16 Instrumnet Name : P7
Client Sample ID : MYE4M8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:51:21 DataFile Name : 083SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	21618	22619	22579	22272	2.55	cps
Vanadium	51-2	500050	500377	497655	499361	0.30	cps
Yttrium	89-1	14495950	14922068	14851839	14756619	1.55	cps
Yttrium	89-2	3594770	3539767	3524818	3553119	1.04	cps
Zinc	66-2	108776	107832	107285	107964	0.70	cps
Zirconium	90-1	231733	240143	241663	237846	2.25	cps
Zirconium	91-1	52095	54019	54474	53529	2.36	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-17 Instrumnet Name : P7
Client Sample ID : MYE4M9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:54:27 DataFile Name : 084SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5693073	5650925	5617289	5653762	0.67	cps
Antimony	121-1	2840	2860	2640	2780	4.38	cps
Arsenic	75-2	17734	18251	17717	17901	1.70	cps
Barium	135-1	695050	697791	701669	698170	0.48	cps
Barium	137-1	1199401	1212545	1214562	1208836	0.68	cps
Beryllium	9-1	1406	1439	1489	1445	2.90	cps
Bismuth	209-1	7174729	7116934	6949938	7080534	1.65	cps
Bismuth	209-2	4998127	5030323	5009719	5012723	0.33	cps
Bromine	81-1	30294	35767	37364	34475	10.75	cps
Bromine	81-2	5104	4971	4968	5014	1.56	cps
Cadmium	108-1	177	280	197	218	25.17	cps
Cadmium	106-1	3374	3000	3204	3193	5.86	cps
Cadmium	111-1	3511	3244	3335	3364	4.04	cps
Calcium	43-1	491812	498090	497614	495838	0.70	cps
Calcium	44-1	7509030	7591064	7640557	7580217	0.88	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	209768	210779	207532	209359	0.79	cps
Cobalt	59-2	141711	141062	142317	141697	0.44	cps
Copper	63-2	290121	291824	289898	290614	0.36	cps
Dysprosium	156-1	45012	45470	45547	45343	0.64	cps
Dysprosium	156-2	34496	34392	34562	34483	0.25	cps
Erbium	164-1	35251	35081	35194	35175	0.25	cps
Erbium	164-2	22532	22078	21547	22052	2.24	cps
Gadolinium	160-1	43320	44129	44380	43943	1.26	cps
Gadolinium	160-2	28726	28750	29331	28935	1.18	cps
Holmium	165-1	11246427	11085954	11237524	11189968	0.81	cps
Holmium	165-2	7011657	7086270	6899967	6999298	1.34	cps
Indium	115-1	8676212	8767757	8616695	8686888	0.88	cps
Indium	115-2	2703983	2713241	2701879	2706368	0.22	cps
Iron	56-2	135765491	134910108	136406381	135693994	0.55	cps
Iron	57-2	3375075	3359051	3373955	3369360	0.27	cps
Iron	54-2	7492554	7414390	7355470	7420805	0.93	cps
Krypton	83-1	263	250	223	246	8.29	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-17 Instrumnet Name : P7
Client Sample ID : MYE4M9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:54:27 DataFile Name : 084SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	145728	146345	147454	146509	0.60	cps
Lead	207-1	124262	125574	125328	125055	0.56	cps
Lead	208-1	572728	574581	578365	575225	0.50	cps
Lithium	6-1	1414713	1404616	1405092	1408141	0.40	cps
Magnesium	24-2	5299128	5266374	5289342	5284948	0.32	cps
Manganese	55-2	3760601	3809493	3763310	3777802	0.73	cps
Molybdenum	94-1	54120	55799	55143	55021	1.54	cps
Molybdenum	95-1	3277	3787	3534	3533	7.22	cps
Molybdenum	96-1	14908	15268	14370	14849	3.04	cps
Molybdenum	97-1	2064	2284	2140	2162	5.16	cps
Molybdenum	98-1	5324	5401	5274	5333	1.20	cps
Neodymium	150-1	81946	82499	83036	82494	0.66	cps
Neodymium	150-2	43938	42577	44012	43509	1.86	cps
Nickel	60-2	71811	72360	72695	72289	0.62	cps
Phosphorus	31-2	16409	16072	16189	16223	1.05	cps
Potassium	39-2	1935794	1935902	1921919	1931205	0.42	cps
Rhodium	103-1	8012608	7943376	7846054	7934013	1.05	cps
Rhodium	103-2	4643828	4642898	4652197	4646308	0.11	cps
Scandium	45-1	4885631	4912916	4882029	4893525	0.35	cps
Scandium	45-2	434475	436050	433049	434525	0.35	cps
Selenium	82-1	521	532	595	549	7.34	cps
Selenium	77-2	143	127	173	148	16.00	cps
Selenium	78-2	360	410	340	370	9.74	cps
Silicon	28-1	31269233	31589400	31767756	31542130	0.80	cps
Silver	107-1	2100	2217	2160	2159	2.70	cps
Silver	109-1	1897	2184	1973	2018	7.36	cps
Sodium	23-2	270179	270868	268806	269951	0.39	cps
Strontium	86-1	397271	399863	402412	399848	0.64	cps
Strontium	88-1	3575893	3599898	3620347	3598713	0.62	cps
Sulfur	34-1	156454	152465	151038	153319	1.83	cps
Terbium	159-1	11591324	11535124	11544587	11557012	0.26	cps
Terbium	159-2	6957085	6973890	6949781	6960252	0.18	cps
Thallium	203-1	1647	1847	1610	1701	7.49	cps
Thallium	205-1	4077	3867	3717	3887	4.65	cps
Tin	118-1	5038	5071	5034	5048	0.40	cps
Titanium	47-1	498405	503488	503153	501682	0.57	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-17 Instrumnet Name : P7
Client Sample ID : MYE4M9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:54:27 DataFile Name : 084SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	26914	27408	27802	27375	1.63	cps
Vanadium	51-2	311121	311291	307013	309808	0.78	cps
Yttrium	89-1	14632952	14689080	14567483	14629838	0.42	cps
Yttrium	89-2	3522308	3506332	3499778	3509473	0.33	cps
Zinc	66-2	120444	119466	118422	119444	0.85	cps
Zirconium	90-1	142752	145199	143651	143867	0.86	cps
Zirconium	91-1	31891	31801	32576	32089	1.32	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-18 Instrumnet Name : P7
Client Sample ID : MYE4N0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:57:36 DataFile Name : 085SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5333394	5381275	5319817	5344829	0.60	cps
Antimony	121-1	6505	6975	6512	6664	4.05	cps
Arsenic	75-2	22467	22661	22354	22494	0.69	cps
Barium	135-1	1016081	1019391	1037062	1024178	1.10	cps
Barium	137-1	1883305	1932058	1955163	1923509	1.91	cps
Beryllium	9-1	1606	1411	1461	1493	6.77	cps
Bismuth	209-1	7099161	7099972	6195004	6798045	7.68	cps
Bismuth	209-2	5068679	4952403	4939499	4986860	1.43	cps
Bromine	81-1	22871	24330	24791	23997	4.18	cps
Bromine	81-2	2530	2444	2467	2480	1.81	cps
Cadmium	108-1	317	377	320	338	9.98	cps
Cadmium	106-1	3670	3754	3244	3556	7.70	cps
Cadmium	111-1	4184	4341	3993	4173	4.17	cps
Calcium	43-1	1162869	1178399	1179876	1173715	0.80	cps
Calcium	44-1	17740016	17944621	18026642	17903759	0.82	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	193011	191767	194409	193062	0.68	cps
Cobalt	59-2	162810	162367	161050	162076	0.56	cps
Copper	63-2	461557	456974	456945	458492	0.58	cps
Dysprosium	156-1	47019	47624	47467	47370	0.66	cps
Dysprosium	156-2	36695	36584	36106	36462	0.86	cps
Erbium	164-1	36732	36340	37009	36694	0.92	cps
Erbium	164-2	22431	22558	22401	22464	0.37	cps
Gadolinium	160-1	43210	43815	44457	43827	1.42	cps
Gadolinium	160-2	29331	28168	28910	28803	2.04	cps
Holmium	165-1	11601208	11442032	9861366	10968202	8.77	cps
Holmium	165-2	7193347	7009999	6951571	7051639	1.79	cps
Indium	115-1	8733155	8662621	7561925	8319234	7.89	cps
Indium	115-2	2746236	2690302	2729080	2721873	1.05	cps
Iron	56-2	139159545	138916198	138236305	138770682	0.34	cps
Iron	57-2	3469455	3437269	3486323	3464349	0.72	cps
Iron	54-2	7696531	7593082	7629304	7639639	0.69	cps
Krypton	83-1	300	227	307	278	15.98	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-18 Instrumnet Name : P7
Client Sample ID : MYE4N0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:57:36 DataFile Name : 085SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	230511	232222	232212	231648	0.43	cps
Lead	207-1	197244	197034	200684	198321	1.03	cps
Lead	208-1	898860	909057	908868	905595	0.64	cps
Lithium	6-1	1398480	1435427	1259902	1364603	6.78	cps
Magnesium	24-2	5642476	5620796	5637622	5633631	0.20	cps
Manganese	55-2	5289822	5263000	5298164	5283662	0.35	cps
Molybdenum	94-1	94103	93939	95874	94639	1.13	cps
Molybdenum	95-1	4037	4597	4361	4332	6.49	cps
Molybdenum	96-1	21726	22017	22351	22031	1.42	cps
Molybdenum	97-1	2760	2827	2444	2677	7.65	cps
Molybdenum	98-1	6355	6805	6688	6616	3.53	cps
Neodymium	150-1	88361	87661	88445	88156	0.49	cps
Neodymium	150-2	46383	46633	45931	46316	0.77	cps
Nickel	60-2	80672	81168	80018	80619	0.72	cps
Phosphorus	31-2	26580	26186	26219	26328	0.83	cps
Potassium	39-2	3030508	3022073	3046664	3033082	0.41	cps
Rhodium	103-1	7889355	8057190	6836620	7594388	8.71	cps
Rhodium	103-2	4587754	4593948	4603545	4595082	0.17	cps
Scandium	45-1	4841403	4924842	4258140	4674795	7.77	cps
Scandium	45-2	431687	434223	429102	431670	0.59	cps
Selenium	82-1	536	598	461	532	12.88	cps
Selenium	77-2	167	140	103	137	23.27	cps
Selenium	78-2	400	393	393	396	0.97	cps
Silicon	28-1	40054712	40440366	40531226	40342101	0.63	cps
Silver	107-1	2204	2364	2437	2335	5.11	cps
Silver	109-1	2164	2277	2234	2225	2.57	cps
Sodium	23-2	164821	164461	164181	164488	0.20	cps
Strontium	86-1	841398	851462	848777	847212	0.62	cps
Strontium	88-1	7552843	7662417	7712977	7642746	1.07	cps
Sulfur	34-1	197266	201203	203766	200745	1.63	cps
Terbium	159-1	11658001	11518994	10145532	11107509	7.53	cps
Terbium	159-2	6920676	6985626	6953048	6953117	0.47	cps
Thallium	203-1	3907	3804	4077	3929	3.51	cps
Thallium	205-1	9116	9340	9380	9279	1.53	cps
Tin	118-1	8846	9376	9687	9303	4.57	cps
Titanium	47-1	655291	665943	664886	662040	0.89	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-18 Instrumnet Name : P7
Client Sample ID : MYE4N0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 21:57:36 DataFile Name : 085SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	18977	19858	19211	19349	2.36	cps
Vanadium	51-2	276892	274319	276750	275987	0.52	cps
Yttrium	89-1	14892757	14845922	12993982	14244220	7.60	cps
Yttrium	89-2	3532220	3542245	3500675	3525047	0.62	cps
Zinc	66-2	145853	146040	146652	146182	0.29	cps
Zirconium	90-1	248234	248179	254733	250382	1.50	cps
Zirconium	91-1	55675	56130	56194	56000	0.51	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-19 Instrumnet Name : P7
Client Sample ID : MYE4N1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:00:42 DataFile Name : 086SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5398093	5316059	5255772	5323308	1.34	cps
Antimony	121-1	16036	15976	16213	16075	0.77	cps
Arsenic	75-2	21028	20688	20337	20684	1.67	cps
Barium	135-1	940304	944080	940948	941777	0.21	cps
Barium	137-1	1768840	1763499	1739451	1757263	0.89	cps
Beryllium	9-1	1745	1628	1650	1674	3.70	cps
Bismuth	209-1	7046381	6965360	6964132	6991958	0.67	cps
Bismuth	209-2	4918562	4973007	4901684	4931084	0.76	cps
Bromine	81-1	17057	17868	18238	17721	3.41	cps
Bromine	81-2	1393	1540	1493	1476	5.08	cps
Cadmium	108-1	453	497	463	471	4.82	cps
Cadmium	106-1	3527	3494	3357	3459	2.60	cps
Cadmium	111-1	7013	7096	7140	7083	0.91	cps
Calcium	43-1	254395	255740	256271	255469	0.38	cps
Calcium	44-1	3934218	3916415	3864004	3904879	0.93	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	220756	217283	218084	218708	0.83	cps
Cobalt	59-2	186936	189153	189386	188491	0.72	cps
Copper	63-2	1249292	1260268	1253293	1254284	0.44	cps
Dysprosium	156-1	57674	59144	59358	58725	1.56	cps
Dysprosium	156-2	44647	44303	44009	44320	0.72	cps
Erbium	164-1	43815	42989	44046	43617	1.27	cps
Erbium	164-2	26879	26716	27323	26973	1.17	cps
Gadolinium	160-1	53354	54509	53666	53843	1.11	cps
Gadolinium	160-2	35545	35913	35228	35562	0.96	cps
Holmium	165-1	11136528	11017613	11117809	11090650	0.58	cps
Holmium	165-2	6854921	6883401	6819445	6852589	0.47	cps
Indium	115-1	8621174	8619156	8483260	8574530	0.92	cps
Indium	115-2	2622907	2639686	2624278	2628957	0.35	cps
Iron	56-2	152966468	155012328	153612761	153863852	0.68	cps
Iron	57-2	3812523	3711242	3784738	3769501	1.39	cps
Iron	54-2	8338078	8399919	8417657	8385218	0.50	cps
Krypton	83-1	240	207	270	239	13.26	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-19 Instrumnet Name : P7
Client Sample ID : MYE4N1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:00:42 DataFile Name : 086SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	727067	741882	729700	732883	1.08	cps
Lead	207-1	642388	641975	650321	644894	0.73	cps
Lead	208-1	2897282	2927705	2912239	2912408	0.52	cps
Lithium	6-1	1405807	1392095	1412170	1403357	0.73	cps
Magnesium	24-2	4853953	4869855	4867278	4863695	0.18	cps
Manganese	55-2	5167337	5240729	5159521	5189196	0.86	cps
Molybdenum	94-1	58674	59092	58991	58919	0.37	cps
Molybdenum	95-1	2754	2877	2827	2819	2.20	cps
Molybdenum	96-1	14577	14554	14551	14561	0.10	cps
Molybdenum	97-1	1690	1557	1777	1675	6.62	cps
Molybdenum	98-1	3897	4074	4084	4018	2.61	cps
Neodymium	150-1	105622	108725	109824	108057	2.02	cps
Neodymium	150-2	57416	56680	56509	56868	0.85	cps
Nickel	60-2	82613	82381	81044	82013	1.03	cps
Phosphorus	31-2	19750	19716	19646	19704	0.27	cps
Potassium	39-2	2990476	2977384	2978167	2982009	0.25	cps
Rhodium	103-1	7890127	7757908	7670306	7772780	1.42	cps
Rhodium	103-2	4512837	4484642	4522785	4506755	0.44	cps
Scandium	45-1	4782852	4830710	4761593	4791718	0.74	cps
Scandium	45-2	420623	424476	422467	422522	0.46	cps
Selenium	82-1	506	617	512	545	11.48	cps
Selenium	77-2	183	163	197	181	9.26	cps
Selenium	78-2	437	380	367	394	9.42	cps
Silicon	28-1	41074941	41566028	41254379	41298449	0.60	cps
Silver	107-1	8806	8826	8996	8876	1.18	cps
Silver	109-1	8116	8409	8453	8326	2.20	cps
Sodium	23-2	115113	113528	112958	113866	0.98	cps
Strontium	86-1	199563	200942	197722	199409	0.81	cps
Strontium	88-1	1757289	1816793	1804951	1793011	1.76	cps
Sulfur	34-1	140220	134216	130626	135021	3.59	cps
Terbium	159-1	11564866	11284019	11238692	11362526	1.56	cps
Terbium	159-2	6830378	6853248	6908302	6863976	0.58	cps
Thallium	203-1	2720	2704	2907	2777	4.07	cps
Thallium	205-1	6655	6552	6979	6728	3.31	cps
Tin	118-1	6361	6161	6411	6311	2.10	cps
Titanium	47-1	765932	771325	765754	767670	0.41	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-19 Instrumnet Name : P7
Client Sample ID : MYE4N1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:00:42 DataFile Name : 086SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	23424	23414	23154	23330	0.66	cps
Vanadium	51-2	292149	292479	292672	292433	0.09	cps
Yttrium	89-1	14549339	14504281	14232641	14428754	1.19	cps
Yttrium	89-2	3419475	3447335	3462838	3443216	0.64	cps
Zinc	66-2	197103	194992	195026	195707	0.62	cps
Zirconium	90-1	154793	156558	154944	155432	0.63	cps
Zirconium	91-1	34320	34644	34571	34512	0.49	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-20 Instrumnet Name : P7
Client Sample ID : MYE4N2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:03:50 DataFile Name : 087SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	11217228	11392901	11433550	11347893	1.01	cps
Antimony	121-1	2417	2497	2390	2435	2.28	cps
Arsenic	75-2	18749	18675	18151	18525	1.76	cps
Barium	135-1	798996	801567	800865	800476	0.17	cps
Barium	137-1	1385745	1392038	1393063	1390282	0.29	cps
Beryllium	9-1	1389	1617	1461	1489	7.82	cps
Bismuth	209-1	6920365	6792745	6827189	6846766	0.96	cps
Bismuth	209-2	4888780	4871361	4877688	4879276	0.18	cps
Bromine	81-1	14921	15535	15385	15280	2.09	cps
Bromine	81-2	850	887	987	908	7.79	cps
Cadmium	108-1	450	487	567	501	11.91	cps
Cadmium	106-1	3497	3227	3340	3355	4.04	cps
Cadmium	111-1	8158	8203	8150	8170	0.35	cps
Calcium	43-1	561135	563179	562922	562412	0.20	cps
Calcium	44-1	8574000	8535381	8657021	8588801	0.72	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	217989	214872	214378	215747	0.91	cps
Cobalt	59-2	522199	527276	524979	524818	0.48	cps
Copper	63-2	673340	678569	670729	674213	0.59	cps
Dysprosium	156-1	44514	45199	44748	44820	0.78	cps
Dysprosium	156-2	32067	31716	32070	31951	0.64	cps
Erbium	164-1	45467	44524	45226	45072	1.09	cps
Erbium	164-2	28058	27871	27841	27923	0.42	cps
Gadolinium	160-1	44494	44400	44888	44594	0.58	cps
Gadolinium	160-2	30072	29909	29448	29809	1.09	cps
Holmium	165-1	11087307	11070284	10976742	11044778	0.54	cps
Holmium	165-2	6943393	6840948	6901397	6895246	0.75	cps
Indium	115-1	8501793	8367659	8415333	8428261	0.81	cps
Indium	115-2	2640615	2623687	2621680	2628661	0.40	cps
Iron	56-2	355780568	355673475	357460981	356305008	0.28	cps
Iron	57-2	8862592	8879153	8872774	8871506	0.09	cps
Iron	54-2	19452239	19513517	19569729	19511828	0.30	cps
Krypton	83-1	250	283	237	257	9.37	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-20 Instrumnet Name : P7
Client Sample ID : MYE4N2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:03:50 DataFile Name : 087SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	238468	241811	241052	240444	0.73	cps
Lead	207-1	204305	208357	207103	206588	1.00	cps
Lead	208-1	935466	945951	946397	942605	0.66	cps
Lithium	6-1	1391099	1412419	1411250	1404923	0.85	cps
Magnesium	24-2	10325776	10365115	10245837	10312243	0.59	cps
Manganese	55-2	11346034	11310344	11326857	11327745	0.16	cps
Molybdenum	94-1	17427	17240	17671	17446	1.24	cps
Molybdenum	95-1	5628	5464	5101	5398	5.00	cps
Molybdenum	96-1	12926	13123	13073	13040	0.78	cps
Molybdenum	97-1	3274	3357	3357	3329	1.45	cps
Molybdenum	98-1	7642	8029	7859	7843	2.47	cps
Neodymium	150-1	65980	66613	65096	65896	1.16	cps
Neodymium	150-2	34095	34085	34579	34253	0.83	cps
Nickel	60-2	78169	77636	77653	77820	0.39	cps
Phosphorus	31-2	15334	15308	15261	15301	0.24	cps
Potassium	39-2	1359246	1375444	1358558	1364416	0.70	cps
Rhodium	103-1	7719765	7685061	7599020	7667949	0.81	cps
Rhodium	103-2	4516451	4439903	4452763	4469706	0.92	cps
Scandium	45-1	4881749	4821895	4852516	4852053	0.62	cps
Scandium	45-2	430311	430553	429365	430076	0.15	cps
Selenium	82-1	529	494	528	517	3.78	cps
Selenium	77-2	117	127	153	132	14.34	cps
Selenium	78-2	313	380	280	324	15.69	cps
Silicon	28-1	43459844	43954163	43456359	43623455	0.66	cps
Silver	107-1	9753	9860	9547	9720	1.64	cps
Silver	109-1	9043	9216	9293	9184	1.40	cps
Sodium	23-2	472267	474571	472285	473041	0.28	cps
Strontium	86-1	385214	390318	390004	388512	0.74	cps
Strontium	88-1	3481685	3512182	3491615	3495161	0.45	cps
Sulfur	34-1	118115	116349	114656	116374	1.49	cps
Terbium	159-1	11483844	11445070	11301669	11410194	0.84	cps
Terbium	159-2	6739892	6768320	6729752	6745988	0.30	cps
Thallium	203-1	2694	2664	2750	2703	1.63	cps
Thallium	205-1	6348	6675	6505	6509	2.51	cps
Tin	118-1	3694	3640	3794	3709	2.10	cps
Titanium	47-1	1529467	1563354	1517987	1536936	1.53	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-20 Instrumnet Name : P7
Client Sample ID : MYE4N2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:03:50 DataFile Name : 087SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	14598	14712	14468	14593	0.84	cps
Vanadium	51-2	653191	650472	647567	650410	0.43	cps
Yttrium	89-1	14571721	14607929	14438551	14539400	0.61	cps
Yttrium	89-2	3516488	3511630	3482531	3503550	0.52	cps
Zinc	66-2	345651	349010	347072	347244	0.49	cps
Zirconium	90-1	39486	38978	39560	39341	0.80	cps
Zirconium	91-1	8529	8673	8763	8655	1.36	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-21 Instrumnet Name : P7
Client Sample ID : MYE4N3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:06:56 DataFile Name : 088SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5720309	5781195	5855590	5785698	1.17	cps
Antimony	121-1	3694	3580	3797	3691	2.94	cps
Arsenic	75-2	25048	25165	25105	25106	0.23	cps
Barium	135-1	1008605	1016032	1021945	1015527	0.66	cps
Barium	137-1	1925785	1902060	1918036	1915294	0.63	cps
Beryllium	9-1	1750	1411	1795	1652	12.69	cps
Bismuth	209-1	6964528	6893663	6973539	6943910	0.63	cps
Bismuth	209-2	4783950	4496058	4286044	4522017	5.53	cps
Bromine	81-1	16826	17057	18101	17328	3.92	cps
Bromine	81-2	1460	1627	1557	1548	5.41	cps
Cadmium	108-1	367	343	363	358	3.53	cps
Cadmium	106-1	3560	3047	3317	3308	7.76	cps
Cadmium	111-1	4863	4443	4651	4652	4.51	cps
Calcium	43-1	503918	506633	505486	505346	0.27	cps
Calcium	44-1	7630210	7783161	7701789	7705053	0.99	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	219152	221442	218653	219749	0.68	cps
Cobalt	59-2	201899	204514	201466	202626	0.81	cps
Copper	63-2	666452	667494	667855	667267	0.11	cps
Dysprosium	156-1	58575	58966	59643	59061	0.91	cps
Dysprosium	156-2	44674	44828	45166	44889	0.56	cps
Erbium	164-1	42986	44139	45049	44058	2.35	cps
Erbium	164-2	27667	27688	27460	27605	0.46	cps
Gadolinium	160-1	53943	54275	53776	53998	0.47	cps
Gadolinium	160-2	36026	36297	36076	36133	0.40	cps
Holmium	165-1	11098688	10961873	11091844	11050802	0.70	cps
Holmium	165-2	6822667	6146597	5932501	6300588	7.37	cps
Indium	115-1	8462658	8459405	8375318	8432460	0.59	cps
Indium	115-2	2586987	2370918	2269202	2409036	6.74	cps
Iron	56-2	165047584	165575171	163503008	164708588	0.65	cps
Iron	57-2	4071484	4076789	4019854	4056042	0.78	cps
Iron	54-2	9004255	9021126	8889184	8971522	0.80	cps
Krypton	83-1	237	270	260	256	6.69	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-21 Instrumnet Name : P7
Client Sample ID : MYE4N3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:06:56 DataFile Name : 088SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	382694	385434	387406	385178	0.61	cps
Lead	207-1	333157	335177	337045	335126	0.58	cps
Lead	208-1	1518296	1526993	1538630	1527973	0.67	cps
Lithium	6-1	1386204	1413220	1379016	1392813	1.29	cps
Magnesium	24-2	5393160	5403505	5399927	5398864	0.10	cps
Manganese	55-2	5779777	5813687	5696681	5763382	1.04	cps
Molybdenum	94-1	69557	70311	71918	70596	1.71	cps
Molybdenum	95-1	3260	3474	3510	3415	3.95	cps
Molybdenum	96-1	17664	17594	17080	17446	1.83	cps
Molybdenum	97-1	1833	2020	2257	2037	10.42	cps
Molybdenum	98-1	4747	4898	5161	4935	4.24	cps
Neodymium	150-1	109367	111406	110798	110523	0.95	cps
Neodymium	150-2	58002	58132	58333	58156	0.29	cps
Nickel	60-2	96612	97528	96068	96736	0.76	cps
Phosphorus	31-2	21826	22710	22450	22329	2.04	cps
Potassium	39-2	3173738	3208628	3169015	3183794	0.68	cps
Rhodium	103-1	7776839	7619653	7545912	7647468	1.54	cps
Rhodium	103-2	4429376	3996681	3910342	4112133	6.76	cps
Scandium	45-1	4679989	4661126	4667293	4669469	0.21	cps
Scandium	45-2	418182	382147	362148	387492	7.33	cps
Selenium	82-1	535	470	445	483	9.63	cps
Selenium	77-2	217	197	230	214	7.82	cps
Selenium	78-2	380	380	343	368	5.76	cps
Silicon	28-1	33608570	33939487	34279222	33942426	0.99	cps
Silver	107-1	3174	3540	3347	3354	5.47	cps
Silver	109-1	3097	3127	3187	3137	1.46	cps
Sodium	23-2	140043	140548	140437	140343	0.19	cps
Strontium	86-1	260459	263038	263835	262444	0.67	cps
Strontium	88-1	2364201	2351838	2385321	2367120	0.72	cps
Sulfur	34-1	126760	125141	126334	126078	0.67	cps
Terbium	159-1	11403929	11326399	11239424	11323251	0.73	cps
Terbium	159-2	6803208	6072275	5917665	6264383	7.55	cps
Thallium	203-1	2724	2674	2670	2689	1.11	cps
Thallium	205-1	6025	6608	6378	6337	4.64	cps
Tin	118-1	5081	5258	5308	5215	2.28	cps
Titanium	47-1	740324	741685	746080	742697	0.41	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-21 Instrumnet Name : P7
Client Sample ID : MYE4N3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:06:56 DataFile Name : 088SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	22372	22753	22699	22608	0.91	cps
Vanadium	51-2	315410	319935	316820	317388	0.73	cps
Yttrium	89-1	14271194	14143370	14184223	14199596	0.46	cps
Yttrium	89-2	3439689	3090240	3039834	3189921	6.83	cps
Zinc	66-2	137069	138667	136749	137495	0.75	cps
Zirconium	90-1	186235	188007	188264	187502	0.59	cps
Zirconium	91-1	42134	42618	41940	42231	0.83	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-22 Instrumnet Name : P7
Client Sample ID : MYE4N4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:10:04 DataFile Name : 089SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	11280988	11346504	11148934	11258808	0.89	cps
Antimony	121-1	3400	3304	3347	3350	1.45	cps
Arsenic	75-2	24146	23939	23819	23968	0.69	cps
Barium	135-1	692019	693987	698534	694846	0.48	cps
Barium	137-1	1199882	1210148	1195210	1201747	0.64	cps
Beryllium	9-1	1417	1189	1250	1285	9.17	cps
Bismuth	209-1	6926566	6741339	6758551	6808819	1.50	cps
Bismuth	209-2	4753877	4761036	4759018	4757977	0.08	cps
Bromine	81-1	14774	14657	15208	14880	1.95	cps
Bromine	81-2	820	767	993	860	13.78	cps
Cadmium	108-1	473	403	430	436	8.11	cps
Cadmium	106-1	3404	3284	3374	3354	1.86	cps
Cadmium	111-1	7336	7297	7385	7339	0.60	cps
Calcium	43-1	476862	481107	479880	479283	0.46	cps
Calcium	44-1	7265264	7256511	7227442	7249739	0.27	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	188786	189518	188405	188903	0.30	cps
Cobalt	59-2	476251	473657	470175	473361	0.64	cps
Copper	63-2	740612	733335	728837	734261	0.81	cps
Dysprosium	156-1	40722	41066	41027	40938	0.46	cps
Dysprosium	156-2	29488	29244	29354	29362	0.42	cps
Erbium	164-1	42618	42999	43735	43117	1.32	cps
Erbium	164-2	26545	26732	25784	26354	1.91	cps
Gadolinium	160-1	41110	41361	41986	41486	1.09	cps
Gadolinium	160-2	27337	27571	27293	27400	0.54	cps
Holmium	165-1	11221301	10956612	11014677	11064197	1.26	cps
Holmium	165-2	6716448	6644103	6736349	6698967	0.72	cps
Indium	115-1	8445894	8311422	8321962	8359759	0.89	cps
Indium	115-2	2569147	2579145	2580257	2576183	0.24	cps
Iron	56-2	363339515	360882381	360890195	361704030	0.39	cps
Iron	57-2	8907858	8879475	8921643	8902992	0.24	cps
Iron	54-2	19742290	19665236	19464775	19624100	0.73	cps
Krypton	83-1	227	237	197	220	9.46	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-22 Instrumnet Name : P7
Client Sample ID : MYE4N4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:10:04 DataFile Name : 089SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	237619	237984	237410	237671	0.12	cps
Lead	207-1	203699	203472	204102	203758	0.16	cps
Lead	208-1	925704	930858	927883	928148	0.28	cps
Lithium	6-1	1380115	1387907	1376381	1381468	0.43	cps
Magnesium	24-2	8848762	8791413	8691044	8777073	0.91	cps
Manganese	55-2	10142264	10023663	10024685	10063537	0.68	cps
Molybdenum	94-1	32135	32994	33014	32714	1.53	cps
Molybdenum	95-1	5104	4938	4864	4969	2.48	cps
Molybdenum	96-1	15719	15722	15612	15684	0.40	cps
Molybdenum	97-1	2937	3150	3037	3041	3.51	cps
Molybdenum	98-1	7379	7345	7579	7434	1.70	cps
Neodymium	150-1	59260	59953	60804	60006	1.29	cps
Neodymium	150-2	31559	31221	30526	31102	1.69	cps
Nickel	60-2	69323	69537	68817	69225	0.53	cps
Phosphorus	31-2	14320	14737	14614	14557	1.47	cps
Potassium	39-2	1318684	1315528	1309102	1314438	0.37	cps
Rhodium	103-1	7694500	7589054	7488767	7590774	1.36	cps
Rhodium	103-2	4401040	4355862	4361862	4372922	0.56	cps
Scandium	45-1	4746584	4728043	4714394	4729674	0.34	cps
Scandium	45-2	417519	420054	416499	418024	0.44	cps
Selenium	82-1	503	541	485	510	5.64	cps
Selenium	77-2	163	150	103	139	22.68	cps
Selenium	78-2	327	300	357	328	8.65	cps
Silicon	28-1	31369009	31908034	31876250	31717764	0.95	cps
Silver	107-1	8996	9667	9527	9396	3.76	cps
Silver	109-1	8703	8846	8896	8815	1.14	cps
Sodium	23-2	364663	363748	362140	363517	0.35	cps
Strontium	86-1	345416	347472	344176	345688	0.48	cps
Strontium	88-1	3124092	3217462	3134845	3158800	1.62	cps
Sulfur	34-1	117252	116100	115667	116340	0.70	cps
Terbium	159-1	11377491	11388664	11019733	11261963	1.86	cps
Terbium	159-2	6713378	6679523	6797142	6730015	0.90	cps
Thallium	203-1	2474	2450	2600	2508	3.22	cps
Thallium	205-1	5838	6008	5635	5827	3.21	cps
Tin	118-1	3524	3744	3814	3694	4.10	cps
Titanium	47-1	2082994	2094356	2146665	2108005	1.61	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-22 Instrumnet Name : P7
Client Sample ID : MYE4N4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:10:04 DataFile Name : 089SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	14742	14722	14905	14789	0.68	cps
Vanadium	51-2	627652	628531	627948	628044	0.07	cps
Yttrium	89-1	14236178	14230440	14178143	14214920	0.22	cps
Yttrium	89-2	3465805	3392006	3424773	3427528	1.08	cps
Zinc	66-2	318142	318099	318616	318286	0.09	cps
Zirconium	90-1	80747	80295	80160	80401	0.38	cps
Zirconium	91-1	18412	18141	17724	18092	1.91	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:13:10 DataFile Name : 090SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1292815	1295049	1303711	1297192	0.44	cps
Antimony	121-1	1167	1183	1170	1173	0.75	cps
Arsenic	75-2	6918	6832	6725	6825	1.42	cps
Barium	135-1	245059	248208	249893	247720	0.99	cps
Barium	137-1	426086	429832	429237	428385	0.47	cps
Beryllium	9-1	356	367	444	389	12.45	cps
Bismuth	209-1	6897284	6723067	6770206	6796852	1.33	cps
Bismuth	209-2	4925514	4871964	4938271	4911916	0.72	cps
Bromine	81-1	11618	11698	11932	11749	1.39	cps
Bromine	81-2	290	290	310	297	3.89	cps
Cadmium	108-1	123	113	90	109	15.71	cps
Cadmium	106-1	2864	2834	2624	2774	4.72	cps
Cadmium	111-1	3022	2978	2814	2938	3.73	cps
Calcium	43-1	156281	158086	157585	157318	0.59	cps
Calcium	44-1	2426388	2427104	2403195	2418896	0.56	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	50591	49651	50434	50225	1.00	cps
Cobalt	59-2	48441	48849	48364	48552	0.54	cps
Copper	63-2	232709	233741	232342	232931	0.31	cps
Dysprosium	156-1	12983	13760	13146	13296	3.08	cps
Dysprosium	156-2	10381	10637	10487	10502	1.23	cps
Erbium	164-1	9680	9940	10120	9914	2.23	cps
Erbium	164-2	6115	6131	6345	6197	2.07	cps
Gadolinium	160-1	12035	12312	12079	12142	1.23	cps
Gadolinium	160-2	8343	8049	8036	8142	2.13	cps
Holmium	165-1	10626364	10788781	10562560	10659235	1.09	cps
Holmium	165-2	6712768	6693572	6705143	6703828	0.14	cps
Indium	115-1	8353186	8330184	8278298	8320556	0.46	cps
Indium	115-2	2664059	2639749	2625207	2643005	0.74	cps
Iron	56-2	39253124	38812727	38613115	38892989	0.84	cps
Iron	57-2	991189	989548	992047	990928	0.13	cps
Iron	54-2	2186717	2126138	2137201	2150019	1.50	cps
Krypton	83-1	193	183	177	184	4.55	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:13:10 DataFile Name : 090SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	217327	219511	217835	218224	0.52	cps
Lead	207-1	188631	189466	189645	189247	0.29	cps
Lead	208-1	860928	861094	862513	861512	0.10	cps
Lithium	6-1	1328114	1377106	1379260	1361493	2.12	cps
Magnesium	24-2	1233367	1242182	1235021	1236857	0.38	cps
Manganese	55-2	1238992	1226097	1232020	1232370	0.52	cps
Molybdenum	94-1	18305	17527	18358	18063	2.58	cps
Molybdenum	95-1	743	757	787	762	2.91	cps
Molybdenum	96-1	4351	4291	4347	4330	0.78	cps
Molybdenum	97-1	443	413	443	433	4.00	cps
Molybdenum	98-1	940	1087	1040	1022	7.33	cps
Neodymium	150-1	25557	26224	25490	25757	1.58	cps
Neodymium	150-2	13393	13817	13690	13633	1.60	cps
Nickel	60-2	22928	22444	22527	22633	1.14	cps
Phosphorus	31-2	6535	6348	6421	6435	1.46	cps
Potassium	39-2	850325	856915	855652	854298	0.41	cps
Rhodium	103-1	7622566	7651971	7549350	7607962	0.69	cps
Rhodium	103-2	4504819	4433446	4527101	4488455	1.09	cps
Scandium	45-1	4501969	4448226	4447707	4465967	0.70	cps
Scandium	45-2	401070	399582	404659	401770	0.65	cps
Selenium	82-1	392	388	480	420	12.39	cps
Selenium	77-2	53	37	33	41	26.06	cps
Selenium	78-2	293	267	290	283	5.13	cps
Silicon	28-1	6455261	6461060	6518748	6478356	0.54	cps
Silver	107-1	1850	1687	1623	1720	6.80	cps
Silver	109-1	1410	1487	1503	1467	3.39	cps
Sodium	23-2	47336	46309	46804	46817	1.10	cps
Strontium	86-1	61729	61571	61401	61567	0.27	cps
Strontium	88-1	534946	536159	541528	537544	0.65	cps
Sulfur	34-1	115089	116174	118593	116619	1.54	cps
Terbium	159-1	11154514	11089114	11041327	11094985	0.51	cps
Terbium	159-2	6756772	6701962	6597430	6685388	1.21	cps
Thallium	203-1	613	690	683	662	6.41	cps
Thallium	205-1	1507	1457	1617	1527	5.36	cps
Tin	118-1	2807	2860	2684	2784	3.26	cps
Titanium	47-1	166083	166966	168499	167183	0.73	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:13:10 DataFile Name : 090SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	4644	4748	4888	4760	2.57	cps
Vanadium	51-2	72513	71873	71826	72071	0.53	cps
Yttrium	89-1	13669259	13552081	13360842	13527394	1.15	cps
Yttrium	89-2	3341665	3319642	3304334	3321880	0.56	cps
Zinc	66-2	52191	51378	51903	51824	0.80	cps
Zirconium	90-1	48068	49269	47868	48402	1.57	cps
Zirconium	91-1	11134	10817	10911	10954	1.49	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-09DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1D Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:16:21 DataFile Name : 091SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1326093	1340701	1325320	1330705	0.65	cps
Antimony	121-1	1020	1167	1190	1126	8.19	cps
Arsenic	75-2	6568	6208	6518	6431	3.03	cps
Barium	135-1	242536	244373	246496	244469	0.81	cps
Barium	137-1	417954	421665	427935	422518	1.19	cps
Beryllium	9-1	311	389	406	369	13.68	cps
Bismuth	209-1	6883402	6845795	6745606	6824934	1.04	cps
Bismuth	209-2	4843785	4872172	4808384	4841447	0.66	cps
Bromine	81-1	11334	10958	11261	11184	1.79	cps
Bromine	81-2	243	307	310	287	13.10	cps
Cadmium	108-1	163	103	110	126	26.19	cps
Cadmium	106-1	2910	2804	2774	2829	2.54	cps
Cadmium	111-1	2983	2884	3041	2969	2.68	cps
Calcium	43-1	154655	156127	157750	156177	0.99	cps
Calcium	44-1	2378354	2400831	2431953	2403713	1.12	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	50099	49685	49682	49822	0.48	cps
Cobalt	59-2	47652	47358	47589	47533	0.33	cps
Copper	63-2	229538	229459	229821	229606	0.08	cps
Dysprosium	156-1	13026	13276	12929	13077	1.37	cps
Dysprosium	156-2	10114	10314	10034	10154	1.42	cps
Erbium	164-1	10040	10104	9850	9998	1.32	cps
Erbium	164-2	6165	5851	6105	6040	2.75	cps
Gadolinium	160-1	12536	11668	12309	12171	3.70	cps
Gadolinium	160-2	8733	8296	8559	8529	2.58	cps
Holmium	165-1	10766992	10858609	10681418	10769006	0.82	cps
Holmium	165-2	6680247	6644866	6651004	6658706	0.28	cps
Indium	115-1	8268141	8282022	8293441	8281201	0.15	cps
Indium	115-2	2659924	2626967	2625016	2637302	0.74	cps
Iron	56-2	38708795	38844684	38559811	38704430	0.37	cps
Iron	57-2	982425	989083	985022	985510	0.34	cps
Iron	54-2	2112857	2172492	2147017	2144122	1.40	cps
Krypton	83-1	200	217	163	193	14.11	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-09DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1D Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:16:21 DataFile Name : 091SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	212218	214774	216486	214493	1.00	cps
Lead	207-1	184620	186947	186431	185999	0.66	cps
Lead	208-1	841737	849526	853338	848200	0.70	cps
Lithium	6-1	1334185	1361017	1361858	1352353	1.16	cps
Magnesium	24-2	1218228	1249659	1238233	1235373	1.29	cps
Manganese	55-2	1213247	1220357	1215500	1216368	0.30	cps
Molybdenum	94-1	19083	18852	19510	19148	1.74	cps
Molybdenum	95-1	660	660	630	650	2.66	cps
Molybdenum	96-1	4357	4371	4434	4387	0.93	cps
Molybdenum	97-1	377	373	400	383	3.79	cps
Molybdenum	98-1	960	973	970	968	0.72	cps
Neodymium	150-1	24885	25784	25717	25462	1.97	cps
Neodymium	150-2	12906	13453	13130	13163	2.09	cps
Nickel	60-2	21315	22076	21365	21586	1.97	cps
Phosphorus	31-2	6115	6255	6051	6140	1.70	cps
Potassium	39-2	852678	852376	853845	852966	0.09	cps
Rhodium	103-1	7716537	7623636	7557087	7632420	1.05	cps
Rhodium	103-2	4465676	4439762	4392198	4432545	0.84	cps
Scandium	45-1	4487465	4446670	4478356	4470830	0.48	cps
Scandium	45-2	403304	401696	403998	402999	0.29	cps
Selenium	82-1	441	425	460	442	3.95	cps
Selenium	77-2	53	43	27	41	32.76	cps
Selenium	78-2	267	340	330	312	12.74	cps
Silicon	28-1	6650415	6599327	6637896	6629212	0.40	cps
Silver	107-1	1600	1617	1487	1568	4.51	cps
Silver	109-1	1417	1543	1597	1519	6.09	cps
Sodium	23-2	44745	45571	44484	44933	1.26	cps
Strontium	86-1	59918	60902	60216	60345	0.84	cps
Strontium	88-1	532446	533052	533548	533015	0.10	cps
Sulfur	34-1	118475	119584	121045	119701	1.08	cps
Terbium	159-1	11034961	10848562	10889026	10924183	0.90	cps
Terbium	159-2	6656165	6606650	6491997	6584937	1.28	cps
Thallium	203-1	670	713	680	688	3.30	cps
Thallium	205-1	1497	1587	1510	1531	3.17	cps
Tin	118-1	2594	2587	2550	2577	0.91	cps
Titanium	47-1	163822	165908	166824	165518	0.93	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-09DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1D Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:16:21 DataFile Name : 091SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	4608	4574	4888	4690	3.67	cps
Vanadium	51-2	72114	72285	72198	72199	0.12	cps
Yttrium	89-1	13539088	13299595	13551673	13463452	1.06	cps
Yttrium	89-2	3319332	3294846	3239533	3284570	1.24	cps
Zinc	66-2	50819	51248	50532	50866	0.71	cps
Zirconium	90-1	50547	50516	50694	50586	0.19	cps
Zirconium	91-1	11661	11368	11361	11463	1.50	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4M1L Dilution Factor : 20
Date & Time Acquired : 2024-11-02 22:19:30 DataFile Name : 092SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	260345	261774	261503	261207	0.29	cps
Antimony	121-1	307	327	303	312	4.04	cps
Arsenic	75-2	1257	1237	1327	1273	3.71	cps
Barium	135-1	49432	49104	49716	49418	0.62	cps
Barium	137-1	85072	83500	86122	84898	1.55	cps
Beryllium	9-1	67	78	111	85	27.15	cps
Bismuth	209-1	6881758	6919053	6931499	6910770	0.37	cps
Bismuth	209-2	4885305	4976919	4503967	4788730	5.24	cps
Bromine	81-1	11134	10767	10981	10961	1.68	cps
Bromine	81-2	137	117	97	117	17.14	cps
Cadmium	108-1	27	33	23	28	18.33	cps
Cadmium	106-1	2600	2690	2627	2639	1.75	cps
Cadmium	111-1	2012	2030	2063	2035	1.27	cps
Calcium	43-1	31306	32000	32408	31905	1.75	cps
Calcium	44-1	523365	530212	531679	528418	0.84	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	11188	11358	11114	11220	1.11	cps
Cobalt	59-2	9840	9880	9930	9883	0.46	cps
Copper	63-2	47462	47201	46987	47216	0.50	cps
Dysprosium	156-1	2547	2527	2520	2531	0.55	cps
Dysprosium	156-2	2044	1943	2144	2044	4.89	cps
Erbium	164-1	2017	2054	2070	2047	1.33	cps
Erbium	164-2	1137	1207	1237	1193	4.30	cps
Gadolinium	160-1	2470	2567	2500	2512	1.97	cps
Gadolinium	160-2	1863	1877	1690	1810	5.75	cps
Holmium	165-1	10892541	10897802	10847304	10879216	0.26	cps
Holmium	165-2	6685451	6819654	6126962	6544023	5.61	cps
Indium	115-1	8571188	8408339	8330990	8436839	1.45	cps
Indium	115-2	2651872	2628106	2427951	2569310	4.79	cps
Iron	56-2	7833267	7881093	7828421	7847593	0.37	cps
Iron	57-2	200002	200157	199364	199841	0.21	cps
Iron	54-2	443315	437726	443081	441374	0.72	cps
Krypton	83-1	180	183	193	186	3.74	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4M1L Dilution Factor : 20
Date & Time Acquired : 2024-11-02 22:19:30 DataFile Name : 092SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	43312	44075	44108	43832	1.03	cps
Lead	207-1	37749	37264	37726	37580	0.73	cps
Lead	208-1	171848	173992	173245	173029	0.63	cps
Lithium	6-1	1338438	1343308	1356587	1346111	0.70	cps
Magnesium	24-2	250301	253923	251867	252030	0.72	cps
Manganese	55-2	245650	246047	248743	246813	0.68	cps
Molybdenum	94-1	3614	3761	3844	3739	3.11	cps
Molybdenum	95-1	243	290	290	274	9.82	cps
Molybdenum	96-1	1057	947	1003	1002	5.49	cps
Molybdenum	97-1	160	227	203	197	17.20	cps
Molybdenum	98-1	480	397	397	424	11.33	cps
Neodymium	150-1	5118	5224	5141	5161	1.09	cps
Neodymium	150-2	2594	2807	2750	2717	4.07	cps
Nickel	60-2	4561	4757	4547	4622	2.54	cps
Phosphorus	31-2	1273	1323	1360	1319	3.30	cps
Potassium	39-2	205629	204818	206596	205681	0.43	cps
Rhodium	103-1	7905672	7789822	7701658	7799051	1.31	cps
Rhodium	103-2	4499029	4560788	4163692	4407836	4.85	cps
Scandium	45-1	4517847	4477907	4496420	4497391	0.44	cps
Scandium	45-2	405224	399554	367683	390820	5.18	cps
Selenium	82-1	476	502	475	484	3.20	cps
Selenium	77-2	10	3	10	8	49.52	cps
Selenium	78-2	213	233	287	244	15.51	cps
Silicon	28-1	1497844	1486070	1484780	1489565	0.48	cps
Silver	107-1	470	410	410	430	8.06	cps
Silver	109-1	393	357	423	391	8.54	cps
Sodium	23-2	21675	22460	22206	22114	1.81	cps
Strontium	86-1	12529	13096	12652	12759	2.34	cps
Strontium	88-1	108157	108318	108785	108420	0.30	cps
Sulfur	34-1	114750	117531	119133	117138	1.89	cps
Terbium	159-1	11186811	11026267	10990309	11067796	0.95	cps
Terbium	159-2	6779962	6675655	6132324	6529314	5.33	cps
Thallium	203-1	217	220	250	229	8.02	cps
Thallium	205-1	640	510	553	568	11.66	cps
Tin	118-1	1773	1877	1697	1782	5.07	cps
Titanium	47-1	33213	33594	34005	33604	1.18	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4520-08LDLX20

Instrumnet Name :

P7

Client Sample ID :

MYE4M1L

Dilution Factor :

20

Date & Time Acquired :

2024-11-02 22:19:30

DataFile Name :

092SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	1047	1003	1023	1024	2.12	cps
Vanadium	51-2	14340	14634	14607	14527	1.12	cps
Yttrium	89-1	13763762	13626124	13510021	13633303	0.93	cps
Yttrium	89-2	3278006	3302329	3055500	3211945	4.24	cps
Zinc	66-2	11191	10767	11027	10995	1.94	cps
Zirconium	90-1	10104	10190	10090	10128	0.54	cps
Zirconium	91-1	2190	2160	2200	2184	0.95	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-10DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1S Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:22:45 DataFile Name : 093SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1351932	1369709	1343855	1355165	0.98	cps
Antimony	121-1	148811	150774	151633	150406	0.96	cps
Arsenic	75-2	9460	9360	9183	9334	1.50	cps
Barium	135-1	999549	1007463	1011502	1006171	0.60	cps
Barium	137-1	1875471	1891450	1899879	1888934	0.66	cps
Beryllium	9-1	11126	11338	11960	11475	3.78	cps
Bismuth	209-1	6900572	6974309	6494613	6789832	3.80	cps
Bismuth	209-2	4875690	4877271	4909135	4887365	0.39	cps
Bromine	81-1	11041	11755	11621	11472	3.31	cps
Bromine	81-2	240	260	303	268	12.09	cps
Cadmium	108-1	1460	1367	1230	1352	8.55	cps
Cadmium	106-1	4631	4544	4207	4461	5.02	cps
Cadmium	111-1	22173	22221	22308	22234	0.31	cps
Calcium	43-1	152463	155881	155332	154559	1.19	cps
Calcium	44-1	2358456	2391483	2380907	2376949	0.71	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	176189	176823	175997	176337	0.25	cps
Cobalt	59-2	526310	523025	525518	524951	0.33	cps
Copper	63-2	390266	390087	391490	390614	0.20	cps
Dysprosium	156-1	12739	12809	13303	12950	2.37	cps
Dysprosium	156-2	10204	10527	10461	10397	1.64	cps
Erbium	164-1	9783	9770	10034	9862	1.51	cps
Erbium	164-2	5991	6051	5768	5937	2.52	cps
Gadolinium	160-1	11985	11969	12249	12068	1.30	cps
Gadolinium	160-2	8473	8176	8216	8288	1.94	cps
Holmium	165-1	10973825	10930152	10251738	10718572	3.78	cps
Holmium	165-2	6799696	6721213	6789135	6770015	0.63	cps
Indium	115-1	8469216	8339042	7902744	8237001	3.60	cps
Indium	115-2	2670228	2622371	2631183	2641261	0.96	cps
Iron	56-2	38895784	38856399	38823176	38858453	0.09	cps
Iron	57-2	993019	984238	991338	989532	0.47	cps
Iron	54-2	2118638	2161448	2155317	2145134	1.08	cps
Krypton	83-1	203	233	247	228	9.74	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-10DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1S Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:22:45 DataFile Name : 093SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	233827	233968	238273	235356	1.07	cps
Lead	207-1	204591	204992	206952	205512	0.61	cps
Lead	208-1	928954	937435	942921	936437	0.75	cps
Lithium	6-1	1363819	1370818	1315171	1349936	2.25	cps
Magnesium	24-2	1214219	1205202	1208135	1209185	0.38	cps
Manganese	55-2	1430379	1420096	1430029	1426835	0.41	cps
Molybdenum	94-1	18839	19263	19703	19268	2.24	cps
Molybdenum	95-1	703	650	783	712	9.42	cps
Molybdenum	96-1	4301	4404	4487	4397	2.13	cps
Molybdenum	97-1	443	430	413	429	3.50	cps
Molybdenum	98-1	967	1043	997	1002	3.85	cps
Neodymium	150-1	25096	25299	25697	25364	1.21	cps
Neodymium	150-2	13160	13380	13323	13287	0.86	cps
Nickel	60-2	143159	143230	143152	143181	0.03	cps
Phosphorus	31-2	6145	6171	6065	6127	0.91	cps
Potassium	39-2	834801	831547	835352	833900	0.25	cps
Rhodium	103-1	7757101	7811385	7255892	7608126	4.03	cps
Rhodium	103-2	4491538	4500310	4514609	4502152	0.26	cps
Scandium	45-1	4495761	4523215	4233657	4417544	3.62	cps
Scandium	45-2	406507	402072	404295	404291	0.55	cps
Selenium	82-1	2961	2952	3052	2989	1.86	cps
Selenium	77-2	277	307	347	310	11.33	cps
Selenium	78-2	1050	990	1197	1079	9.86	cps
Silicon	28-1	6686491	6699837	6709849	6698726	0.17	cps
Silver	107-1	79858	81453	83257	81523	2.09	cps
Silver	109-1	75100	77516	77925	76847	1.99	cps
Sodium	23-2	44464	44808	43809	44360	1.14	cps
Strontium	86-1	60738	61421	60400	60853	0.85	cps
Strontium	88-1	527525	536140	530564	531410	0.82	cps
Sulfur	34-1	125386	125611	123279	124759	1.03	cps
Terbium	159-1	11256830	11235331	10672415	11054858	3.00	cps
Terbium	159-2	6710768	6671881	6669966	6684205	0.34	cps
Thallium	203-1	68581	68759	70073	69138	1.18	cps
Thallium	205-1	162492	164663	166152	164436	1.12	cps
Tin	118-1	2624	2514	2540	2559	2.24	cps
Titanium	47-1	161723	163263	162949	162645	0.50	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-10DLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1S Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:22:45 DataFile Name : 093SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	4938	5078	4851	4955	2.31	cps
Vanadium	51-2	346251	345819	346592	346221	0.11	cps
Yttrium	89-1	13725746	13699972	12749191	13391636	4.16	cps
Yttrium	89-2	3349034	3305622	3290135	3314930	0.92	cps
Zinc	66-2	104187	103817	103562	103855	0.30	cps
Zirconium	90-1	49634	49797	51413	50281	1.96	cps
Zirconium	91-1	11415	11621	11461	11499	0.94	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:25:53 DataFile Name : 094SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1530115	1511766	1535159	1525680	0.81	cps
Antimony	121-1	722880	727928	732739	727849	0.68	cps
Arsenic	75-2	20905	20007	20397	20436	2.20	cps
Barium	135-1	4259418	4253853	4218514	4243929	0.52	cps
Barium	137-1	7291859	7232292	7266175	7263442	0.41	cps
Beryllium	9-1	54440	54802	55398	54880	0.88	cps
Bismuth	209-1	6882071	6958409	6278689	6706389	5.55	cps
Bismuth	209-2	4863535	4928032	4973922	4921830	1.13	cps
Bromine	81-1	11855	11651	11438	11648	1.79	cps
Bromine	81-2	260	250	277	262	5.14	cps
Cadmium	108-1	6318	6241	6128	6229	1.54	cps
Cadmium	106-1	11074	11408	11031	11171	1.85	cps
Cadmium	111-1	94155	96153	96177	95495	1.22	cps
Calcium	43-1	147546	147882	146862	147430	0.35	cps
Calcium	44-1	2260591	2275152	2249906	2261883	0.56	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	650156	648813	642669	647212	0.62	cps
Cobalt	59-2	2391123	2392192	2340948	2374754	1.23	cps
Copper	63-2	990031	984453	982139	985541	0.41	cps
Dysprosium	156-1	12669	12245	12596	12503	1.81	cps
Dysprosium	156-2	9824	9874	10134	9944	1.67	cps
Erbium	164-1	9350	9560	9597	9502	1.40	cps
Erbium	164-2	5711	5788	5778	5759	0.72	cps
Gadolinium	160-1	11565	11778	12125	11823	2.39	cps
Gadolinium	160-2	7952	7952	7806	7903	1.07	cps
Holmium	165-1	10979735	10779249	9723388	10494124	6.43	cps
Holmium	165-2	6704972	6726951	6786695	6739539	0.63	cps
Indium	115-1	8458620	8277643	7446685	8060983	6.69	cps
Indium	115-2	2663445	2634149	2589589	2629061	1.41	cps
Iron	56-2	39354699	38782231	38988489	39041806	0.74	cps
Iron	57-2	995451	988508	995358	993105	0.40	cps
Iron	54-2	2159318	2156884	2142006	2152736	0.44	cps
Krypton	83-1	230	203	217	217	6.15	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:25:53 DataFile Name : 094SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	310709	313127	320198	314678	1.57	cps
Lead	207-1	268380	273653	276040	272691	1.44	cps
Lead	208-1	1234131	1242269	1264156	1246852	1.25	cps
Lithium	6-1	1336694	1373210	1271817	1327240	3.87	cps
Magnesium	24-2	1142609	1130033	1138925	1137189	0.57	cps
Manganese	55-2	2234089	2201758	2216353	2217400	0.73	cps
Molybdenum	94-1	18402	18999	19149	18850	2.10	cps
Molybdenum	95-1	747	760	740	749	1.36	cps
Molybdenum	96-1	4261	4541	4524	4442	3.54	cps
Molybdenum	97-1	393	427	377	399	6.38	cps
Molybdenum	98-1	1053	1063	1107	1074	2.64	cps
Neodymium	150-1	23934	24548	24511	24331	1.42	cps
Neodymium	150-2	12916	12499	12966	12794	2.00	cps
Nickel	60-2	601772	600790	595617	599393	0.55	cps
Phosphorus	31-2	5728	5915	5691	5778	2.07	cps
Potassium	39-2	797116	790774	787264	791718	0.63	cps
Rhodium	103-1	7597529	7655094	6853481	7368701	6.07	cps
Rhodium	103-2	4509346	4493274	4483993	4495538	0.29	cps
Scandium	45-1	4567064	4474152	4018739	4353318	6.74	cps
Scandium	45-2	401561	397535	400969	400022	0.54	cps
Selenium	82-1	12772	13106	12689	12856	1.72	cps
Selenium	77-2	1127	1187	1177	1163	2.76	cps
Selenium	78-2	4031	4284	3991	4102	3.88	cps
Silicon	28-1	6430037	6535253	6496723	6487338	0.82	cps
Silver	107-1	383078	389335	391841	388085	1.16	cps
Silver	109-1	365163	367244	369336	367248	0.57	cps
Sodium	23-2	43638	43602	42900	43380	0.96	cps
Strontium	86-1	56852	56903	58292	57349	1.42	cps
Strontium	88-1	505443	510452	507278	507724	0.50	cps
Sulfur	34-1	121229	123383	120007	121540	1.41	cps
Terbium	159-1	11200462	11172822	10033434	10802239	6.16	cps
Terbium	159-2	6788017	6760377	6766686	6771693	0.21	cps
Thallium	203-1	332785	332232	337704	334240	0.90	cps
Thallium	205-1	786853	790506	808055	795138	1.43	cps
Tin	118-1	2600	2637	2720	2652	2.32	cps
Titanium	47-1	152701	154718	155563	154327	0.95	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4520-08MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4M1SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-02 22:25:53 DataFile Name : 094SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	4494	4684	4624	4601	2.11	cps
Vanadium	51-2	1390687	1370460	1335820	1365655	2.03	cps
Yttrium	89-1	13605192	13630949	12263717	13166620	5.94	cps
Yttrium	89-2	3314474	3298025	3296346	3302948	0.30	cps
Zinc	66-2	304306	300962	304487	303252	0.65	cps
Zirconium	90-1	49068	49844	49239	49384	0.83	cps
Zirconium	91-1	10727	11224	11298	11083	2.80	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BL Instrumnet Name : P7
Client Sample ID : PBS456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:31:57 DataFile Name : 095CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1107	1237	1223	1189	6.02	cps
Antimony	121-1	223	303	297	274	16.17	cps
Arsenic	75-2	27	0	7	11	124.89	cps
Barium	135-1	340	407	377	374	8.92	cps
Barium	137-1	630	710	590	643	9.50	cps
Beryllium	9-1	50	67	39	52	26.97	cps
Bismuth	209-1	5931295	6632014	6688956	6417422	6.58	cps
Bismuth	209-2	4837396	4802723	4819976	4820031	0.36	cps
Bromine	81-1	11181	11515	10844	11180	3.00	cps
Bromine	81-2	63	50	77	63	21.06	cps
Cadmium	108-1	13	7	23	14	58.06	cps
Cadmium	106-1	2187	2347	2164	2232	4.47	cps
Cadmium	111-1	1531	1661	1521	1571	4.97	cps
Calcium	43-1	307	323	297	309	4.36	cps
Calcium	44-1	11465	12089	11962	11839	2.78	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	613	660	610	628	4.45	cps
Cobalt	59-2	197	183	237	206	13.50	cps
Copper	63-2	893	940	1077	970	9.82	cps
Dysprosium	156-1	10	20	23	18	39.03	cps
Dysprosium	156-2	13	23	10	16	44.60	cps
Erbium	164-1	87	63	70	73	16.39	cps
Erbium	164-2	53	47	50	50	6.66	cps
Gadolinium	160-1	17	70	27	38	75.03	cps
Gadolinium	160-2	290	227	257	258	12.29	cps
Holmium	165-1	9457993	10520164	10549026	10175728	6.11	cps
Holmium	165-2	6716772	6592653	6578241	6629222	1.15	cps
Indium	115-1	7303412	8330139	8226589	7953380	7.11	cps
Indium	115-2	2638269	2602885	2638087	2626414	0.78	cps
Iron	56-2	45479	44439	46492	45470	2.26	cps
Iron	57-2	1810	1850	2157	1939	9.78	cps
Iron	54-2	3270	3514	3480	3422	3.86	cps
Krypton	83-1	197	147	237	193	23.32	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BL Instrumnet Name : P7
Client Sample ID : PBS456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:31:57 DataFile Name : 095CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	527	583	527	546	6.00	cps
Lead	207-1	453	550	513	506	9.65	cps
Lead	208-1	2183	2360	2167	2237	4.79	cps
Lithium	6-1	1239710	1340646	1345181	1308512	4.56	cps
Magnesium	24-2	1877	1777	1827	1827	2.74	cps
Manganese	55-2	1343	1310	1400	1351	3.37	cps
Molybdenum	94-1	233	273	330	279	17.42	cps
Molybdenum	95-1	183	147	193	174	14.09	cps
Molybdenum	96-1	253	213	213	227	10.19	cps
Molybdenum	97-1	80	107	117	101	18.75	cps
Molybdenum	98-1	287	207	237	243	16.61	cps
Neodymium	150-1	27	10	13	17	52.93	cps
Neodymium	150-2	7	7	7	7	0.00	cps
Nickel	60-2	167	160	220	182	18.05	cps
Phosphorus	31-2	117	157	137	137	14.63	cps
Potassium	39-2	43950	45588	44983	44840	1.85	cps
Rhodium	103-1	6857972	7619846	7653550	7377123	6.10	cps
Rhodium	103-2	4528873	4484327	4454956	4489385	0.83	cps
Scandium	45-1	3927129	4462268	4440616	4276671	7.08	cps
Scandium	45-2	396574	393922	394986	395161	0.34	cps
Selenium	82-1	334	481	334	383	22.10	cps
Selenium	77-2	0	3	0	1	173.21	cps
Selenium	78-2	223	240	253	239	6.29	cps
Silicon	28-1	199816	198466	199570	199284	0.36	cps
Silver	107-1	207	167	163	179	13.48	cps
Silver	109-1	127	170	147	148	14.67	cps
Sodium	23-2	15908	15291	15555	15585	1.99	cps
Strontium	86-1	547	627	507	560	10.91	cps
Strontium	88-1	2354	2267	2317	2312	1.88	cps
Sulfur	34-1	118384	119382	119512	119093	0.52	cps
Terbium	159-1	9744474	10843267	10949920	10512554	6.35	cps
Terbium	159-2	6597130	6594783	6591196	6594370	0.05	cps
Thallium	203-1	153	127	147	142	9.76	cps
Thallium	205-1	377	357	293	342	12.71	cps
Tin	118-1	1630	1503	1560	1565	4.06	cps
Titanium	47-1	230	227	527	328	52.58	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BL Instrumnet Name : P7
Client Sample ID : PBS456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:31:57 DataFile Name : 095CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	107	87	63	86	25.35	cps
Vanadium	51-2	113	127	153	131	15.53	cps
Yttrium	89-1	11835399	13296221	13191138	12774252	6.38	cps
Yttrium	89-2	3323260	3252608	3231219	3269029	1.47	cps
Zinc	66-2	547	597	580	574	4.43	cps
Zirconium	90-1	450	517	523	497	8.17	cps
Zirconium	91-1	100	97	130	109	16.86	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BS Instrumnet Name : P7
Client Sample ID : LCS456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:35:14 DataFile Name : 096LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	9543	9556	9543	9547	0.08	cps
Antimony	121-1	37064	37616	37619	37433	0.85	cps
Arsenic	75-2	1063	950	1093	1036	7.30	cps
Barium	135-1	47305	48629	47723	47886	1.41	cps
Barium	137-1	83527	82842	83198	83189	0.41	cps
Beryllium	9-1	2967	3217	2906	3030	5.44	cps
Bismuth	209-1	6845504	6770518	6730455	6782159	0.86	cps
Bismuth	209-2	4766843	4888198	4820371	4825137	1.26	cps
Bromine	81-1	10651	11395	11234	11093	3.53	cps
Bromine	81-2	53	67	93	71	28.64	cps
Cadmium	108-1	417	357	360	378	8.93	cps
Cadmium	106-1	3134	3227	2887	3083	5.70	cps
Cadmium	111-1	6749	6970	6863	6861	1.61	cps
Calcium	43-1	25478	26039	26263	25927	1.56	cps
Calcium	44-1	427288	431024	428599	428970	0.44	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	17240	16813	16686	16913	1.72	cps
Cobalt	59-2	12492	12455	12762	12570	1.33	cps
Copper	63-2	19026	18902	19122	19017	0.58	cps
Dysprosium	156-1	43	27	30	33	26.45	cps
Dysprosium	156-2	17	0	20	12	87.67	cps
Erbium	164-1	43	93	80	72	35.85	cps
Erbium	164-2	63	50	57	57	11.76	cps
Gadolinium	160-1	57	40	30	42	31.91	cps
Gadolinium	160-2	287	280	307	291	4.77	cps
Holmium	165-1	10849668	10645980	10673393	10723014	1.03	cps
Holmium	165-2	6490056	6662690	6609523	6587423	1.34	cps
Indium	115-1	8347915	8211355	8225980	8261750	0.91	cps
Indium	115-2	2596170	2670280	2643703	2636717	1.42	cps
Iron	56-2	1466468	1453343	1458789	1459533	0.45	cps
Iron	57-2	37884	37610	38242	37912	0.84	cps
Iron	54-2	82615	81750	82719	82361	0.65	cps
Krypton	83-1	203	217	163	194	14.27	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BS Instrumnet Name : P7
Client Sample ID : LCS456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:35:14 DataFile Name : 096LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	14254	14408	14344	14335	0.54	cps
Lead	207-1	12870	12709	12910	12829	0.83	cps
Lead	208-1	58293	58644	57761	58233	0.76	cps
Lithium	6-1	1370862	1325852	1344553	1347089	1.68	cps
Magnesium	24-2	394896	389037	391920	391951	0.75	cps
Manganese	55-2	6868	7068	7165	7034	2.15	cps
Molybdenum	94-1	39490	39460	39998	39649	0.76	cps
Molybdenum	95-1	48908	48804	48386	48700	0.57	cps
Molybdenum	96-1	53447	54046	54330	53941	0.84	cps
Molybdenum	97-1	29860	30251	30686	30266	1.36	cps
Molybdenum	98-1	76841	77672	77216	77243	0.54	cps
Neodymium	150-1	23	20	20	21	9.11	cps
Neodymium	150-2	10	7	3	7	50.03	cps
Nickel	60-2	3240	3404	3617	3420	5.52	cps
Phosphorus	31-2	150	127	113	130	14.28	cps
Potassium	39-2	506130	498703	503019	502617	0.74	cps
Rhodium	103-1	7689763	7694237	7670709	7684903	0.16	cps
Rhodium	103-2	4476150	4604931	4584210	4555097	1.52	cps
Scandium	45-1	4480416	4400998	4438626	4440014	0.89	cps
Scandium	45-2	390758	407551	398250	398853	2.11	cps
Selenium	82-1	1871	2013	1918	1934	3.74	cps
Selenium	77-2	137	130	140	136	3.76	cps
Selenium	78-2	747	720	757	741	2.56	cps
Silicon	28-1	291939	293780	295285	293668	0.57	cps
Silver	107-1	22998	22534	23532	23022	2.17	cps
Silver	109-1	21052	21960	22060	21691	2.56	cps
Sodium	23-2	643861	633868	643396	640375	0.88	cps
Strontium	86-1	6565	6772	6868	6735	2.30	cps
Strontium	88-1	55725	57415	56555	56565	1.49	cps
Sulfur	34-1	132024	131796	135306	133042	1.48	cps
Terbium	159-1	11057963	10752405	10660566	10823645	1.92	cps
Terbium	159-2	6498325	6661535	6602350	6587403	1.25	cps
Thallium	203-1	17281	17188	16760	17076	1.63	cps
Thallium	205-1	40216	41092	40066	40458	1.37	cps
Tin	118-1	1733	1683	1780	1732	2.79	cps
Titanium	47-1	12105	12705	12695	12502	2.75	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BS Instrumnet Name : P7
Client Sample ID : LCS456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:35:14 DataFile Name : 096LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	50756	53025	52757	52180	2.38	cps
Vanadium	51-2	35448	34860	35194	35168	0.84	cps
Yttrium	89-1	13345681	13206821	13195703	13249402	0.63	cps
Yttrium	89-2	3214901	3307154	3256324	3259460	1.42	cps
Zinc	66-2	7665	7425	7976	7689	3.59	cps
Zirconium	90-1	33993	34136	34474	34201	0.72	cps
Zirconium	91-1	7475	7579	7575	7543	0.78	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BSD Instrumnet Name : P7
Client Sample ID : LCSD456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:38:27 DataFile Name : 097LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5118	4998	4904	5006	2.14	cps
Antimony	121-1	18078	18796	18285	18386	2.01	cps
Arsenic	75-2	557	473	483	504	9.02	cps
Barium	135-1	23159	24307	23089	23518	2.91	cps
Barium	137-1	39876	40314	40638	40276	0.95	cps
Beryllium	9-1	1472	1495	1333	1433	6.09	cps
Bismuth	209-1	6634075	6696656	6759113	6696615	0.93	cps
Bismuth	209-2	4758893	4742139	4748487	4749839	0.18	cps
Bromine	81-1	11121	11134	11364	11207	1.22	cps
Bromine	81-2	57	63	83	68	20.47	cps
Cadmium	108-1	200	150	147	166	18.05	cps
Cadmium	106-1	2960	2817	2787	2855	3.25	cps
Cadmium	111-1	4122	4040	4155	4106	1.44	cps
Calcium	43-1	12428	13072	13189	12897	3.18	cps
Calcium	44-1	214616	215022	216563	215400	0.48	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	8412	8706	8456	8525	1.86	cps
Cobalt	59-2	6248	5908	6135	6097	2.84	cps
Copper	63-2	9607	9710	9887	9734	1.46	cps
Dysprosium	156-1	17	13	17	16	12.40	cps
Dysprosium	156-2	10	7	0	6	91.64	cps
Erbium	164-1	73	60	90	74	20.19	cps
Erbium	164-2	53	53	33	47	24.75	cps
Gadolinium	160-1	57	40	47	48	17.56	cps
Gadolinium	160-2	263	283	293	280	5.46	cps
Holmium	165-1	10498242	10563437	10583225	10548301	0.42	cps
Holmium	165-2	6518440	6442017	6555632	6505363	0.89	cps
Indium	115-1	8278918	8332520	8300766	8304068	0.32	cps
Indium	115-2	2574065	2619958	2597594	2597206	0.88	cps
Iron	56-2	743749	748231	739384	743788	0.59	cps
Iron	57-2	19229	19356	19933	19506	1.92	cps
Iron	54-2	42279	42309	41901	42163	0.54	cps
Krypton	83-1	210	187	150	182	16.60	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BSD Instrumnet Name : P7
Client Sample ID : LCSD456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:38:27 DataFile Name : 097LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	7262	7792	6992	7349	5.54	cps
Lead	207-1	6452	6355	6532	6446	1.37	cps
Lead	208-1	29022	29539	29273	29278	0.88	cps
Lithium	6-1	1362226	1359335	1363113	1361558	0.15	cps
Magnesium	24-2	192383	191533	192687	192201	0.31	cps
Manganese	55-2	3774	3867	3934	3858	2.08	cps
Molybdenum	94-1	19383	19610	19516	19503	0.58	cps
Molybdenum	95-1	23449	23829	23900	23726	1.02	cps
Molybdenum	96-1	26387	26541	26714	26547	0.62	cps
Molybdenum	97-1	14941	14784	14457	14728	1.68	cps
Molybdenum	98-1	37932	38116	38066	38038	0.25	cps
Neodymium	150-1	17	20	17	18	10.81	cps
Neodymium	150-2	3	7	3	4	43.40	cps
Nickel	60-2	1790	1717	1753	1753	2.09	cps
Phosphorus	31-2	117	143	123	128	10.86	cps
Potassium	39-2	271172	270459	271264	270965	0.16	cps
Rhodium	103-1	7698886	7730976	7697816	7709226	0.24	cps
Rhodium	103-2	4498029	4539014	4465852	4500965	0.81	cps
Scandium	45-1	4465631	4486063	4432670	4461455	0.60	cps
Scandium	45-2	401767	396446	402290	400167	0.81	cps
Selenium	82-1	1151	1143	1174	1156	1.37	cps
Selenium	77-2	100	47	53	67	43.59	cps
Selenium	78-2	530	450	417	466	12.51	cps
Silicon	28-1	244182	245507	248013	245901	0.79	cps
Silver	107-1	11028	11241	11318	11196	1.34	cps
Silver	109-1	10804	10767	10794	10789	0.18	cps
Sodium	23-2	321561	319110	320549	320407	0.38	cps
Strontium	86-1	3484	3520	3424	3476	1.40	cps
Strontium	88-1	28471	28728	28418	28539	0.58	cps
Sulfur	34-1	139325	140566	141576	140489	0.80	cps
Terbium	159-1	10908949	10834911	10881854	10875238	0.34	cps
Terbium	159-2	6498250	6525759	6521740	6515250	0.23	cps
Thallium	203-1	8106	8363	8449	8306	2.15	cps
Thallium	205-1	19644	20055	20279	19993	1.61	cps
Tin	118-1	1607	1750	1703	1687	4.33	cps
Titanium	47-1	6218	6141	6031	6130	1.53	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164456BSD Instrumnet Name : P7
Client Sample ID : LCSD456 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:38:27 DataFile Name : 097LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	25164	25541	25491	25399	0.81	cps
Vanadium	51-2	17460	16659	17467	17195	2.70	cps
Yttrium	89-1	13384480	13394196	13321831	13366836	0.29	cps
Yttrium	89-2	3207353	3229834	3278055	3238414	1.12	cps
Zinc	66-2	4037	4034	4147	4073	1.58	cps
Zirconium	90-1	16116	16773	17257	16715	3.43	cps
Zirconium	91-1	3670	3917	3877	3822	3.47	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-01 Instrumnet Name : P7
Client Sample ID : MYE4K0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:41:45 DataFile Name : 098SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8922759	8919844	8956654	8933086	0.23	cps
Antimony	121-1	8783	8583	8499	8622	1.69	cps
Arsenic	75-2	36652	37086	36418	36719	0.92	cps
Barium	135-1	1073623	1078434	1077423	1076493	0.24	cps
Barium	137-1	2036250	2026007	2057298	2039851	0.78	cps
Beryllium	9-1	2261	2011	2056	2109	6.32	cps
Bismuth	209-1	6412656	6513573	6563626	6496618	1.18	cps
Bismuth	209-2	4615248	4607839	4634707	4619265	0.30	cps
Bromine	81-1	17918	18672	18338	18309	2.06	cps
Bromine	81-2	1427	1633	1533	1531	6.75	cps
Cadmium	108-1	340	307	357	334	7.61	cps
Cadmium	106-1	3190	3287	3127	3201	2.52	cps
Cadmium	111-1	4886	4970	4976	4944	1.02	cps
Calcium	43-1	475327	474712	476132	475390	0.15	cps
Calcium	44-1	7240437	7201304	7183460	7208400	0.40	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	364769	362043	359214	362009	0.77	cps
Cobalt	59-2	282432	278978	276917	279442	1.00	cps
Copper	63-2	993942	989233	981529	988235	0.63	cps
Dysprosium	156-1	80328	81173	80137	80546	0.68	cps
Dysprosium	156-2	61192	59669	59509	60123	1.55	cps
Erbium	164-1	67304	67582	68071	67653	0.57	cps
Erbium	164-2	41759	41391	41063	41405	0.84	cps
Gadolinium	160-1	77060	76819	77593	77157	0.51	cps
Gadolinium	160-2	52200	50400	50681	51094	1.90	cps
Holmium	165-1	10756736	10665215	10678349	10700100	0.46	cps
Holmium	165-2	6434982	6534758	6425775	6465172	0.93	cps
Indium	115-1	8085209	8073133	8069734	8076026	0.10	cps
Indium	115-2	2479721	2495218	2449094	2474677	0.95	cps
Iron	56-2	290065289	291760602	288748849	290191580	0.52	cps
Iron	57-2	7196476	7190021	7106052	7164183	0.70	cps
Iron	54-2	15925835	15751327	15736418	15804527	0.67	cps
Krypton	83-1	243	277	257	259	6.48	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-01 Instrumnet Name : P7
Client Sample ID : MYE4K0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:41:45 DataFile Name : 098SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	374178	376086	378496	376254	0.58	cps
Lead	207-1	321291	324230	327270	324263	0.92	cps
Lead	208-1	1460859	1478222	1486379	1475153	0.88	cps
Lithium	6-1	1339819	1337551	1313102	1330157	1.11	cps
Magnesium	24-2	7646582	7656862	7631214	7644886	0.17	cps
Manganese	55-2	6436358	6325620	6324320	6362099	1.01	cps
Molybdenum	94-1	51845	52791	52303	52313	0.91	cps
Molybdenum	95-1	10220	10300	10010	10177	1.47	cps
Molybdenum	96-1	22855	23012	22888	22918	0.36	cps
Molybdenum	97-1	6248	6428	6248	6308	1.65	cps
Molybdenum	98-1	15792	16206	15932	15977	1.32	cps
Neodymium	150-1	141132	139109	141260	140500	0.86	cps
Neodymium	150-2	73098	73004	73112	73071	0.08	cps
Nickel	60-2	115457	114342	112965	114255	1.09	cps
Phosphorus	31-2	37867	37456	37432	37585	0.65	cps
Potassium	39-2	6333008	6338987	6281841	6317946	0.50	cps
Rhodium	103-1	7370724	7333213	7297597	7333845	0.50	cps
Rhodium	103-2	4316351	4254382	4259750	4276828	0.80	cps
Scandium	45-1	4679120	4654344	4651050	4661505	0.33	cps
Scandium	45-2	415901	411677	406059	411212	1.20	cps
Selenium	82-1	607	627	667	634	4.78	cps
Selenium	77-2	290	287	280	286	1.78	cps
Selenium	78-2	443	533	523	500	9.87	cps
Silicon	28-1	30914975	30779935	30493095	30729335	0.70	cps
Silver	107-1	6555	5978	5941	6158	5.59	cps
Silver	109-1	6221	5855	5641	5906	4.97	cps
Sodium	23-2	254483	257873	257617	256658	0.74	cps
Strontium	86-1	408395	412870	410231	410498	0.55	cps
Strontium	88-1	3735943	3705782	3691624	3711116	0.61	cps
Sulfur	34-1	136271	133971	134274	134838	0.93	cps
Terbium	159-1	10877441	10772102	10823513	10824352	0.49	cps
Terbium	159-2	6515506	6513366	6467024	6498632	0.42	cps
Thallium	203-1	3627	3634	3747	3669	1.84	cps
Thallium	205-1	8553	8713	8696	8654	1.02	cps
Tin	118-1	22568	22591	22251	22470	0.85	cps
Titanium	47-1	1298445	1278249	1283653	1286783	0.81	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-01 Instrumnet Name : P7
Client Sample ID : MYE4K0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:41:45 DataFile Name : 098SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	32767	33599	33168	33178	1.25	cps
Vanadium	51-2	570858	570753	567401	569671	0.35	cps
Yttrium	89-1	14227929	14075586	14122471	14141996	0.55	cps
Yttrium	89-2	3452568	3401983	3402488	3419013	0.85	cps
Zinc	66-2	187046	183634	185690	185457	0.93	cps
Zirconium	90-1	128547	129128	128634	128769	0.24	cps
Zirconium	91-1	28157	28124	28501	28261	0.74	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-02 Instrumnet Name : P7
Client Sample ID : MYE4K1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:44:49 DataFile Name : 099SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	6795191	6849255	6780402	6808283	0.53	cps
Antimony	121-1	10541	10614	10818	10657	1.35	cps
Arsenic	75-2	24784	24564	23949	24432	1.77	cps
Barium	135-1	845647	856137	854157	851980	0.65	cps
Barium	137-1	1605357	1650226	1629279	1628287	1.38	cps
Beryllium	9-1	1850	1806	1722	1793	3.62	cps
Bismuth	209-1	6655282	5857787	6780300	6431123	7.78	cps
Bismuth	209-2	4740807	4580617	4681139	4667521	1.73	cps
Bromine	81-1	12872	12802	12922	12866	0.47	cps
Bromine	81-2	680	580	567	609	10.17	cps
Cadmium	108-1	213	223	227	221	3.14	cps
Cadmium	106-1	3210	2607	3147	2988	11.10	cps
Cadmium	111-1	3111	2685	3081	2959	8.03	cps
Calcium	43-1	582249	591609	593846	589235	1.04	cps
Calcium	44-1	8888372	9071041	9003919	8987777	1.03	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	349055	350716	344975	348249	0.85	cps
Cobalt	59-2	254052	251779	248017	251282	1.21	cps
Copper	63-2	439884	436846	435427	437386	0.52	cps
Dysprosium	156-1	69454	72285	70436	70725	2.03	cps
Dysprosium	156-2	53043	53538	53033	53205	0.54	cps
Erbium	164-1	57852	60126	60286	59421	2.29	cps
Erbium	164-2	37126	37026	37133	37095	0.16	cps
Gadolinium	160-1	66574	67582	67964	67373	1.07	cps
Gadolinium	160-2	44888	44256	44009	44385	1.02	cps
Holmium	165-1	10634256	9298874	10793127	10242086	8.01	cps
Holmium	165-2	6706821	6585161	6547988	6613323	1.26	cps
Indium	115-1	8116912	7137234	8191128	7815091	7.53	cps
Indium	115-2	2547706	2558939	2551900	2552848	0.22	cps
Iron	56-2	293863949	293284889	292354682	293167840	0.26	cps
Iron	57-2	7294706	7090442	7205600	7196916	1.42	cps
Iron	54-2	15993990	16026682	15793878	15938183	0.79	cps
Krypton	83-1	280	287	197	254	19.71	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-02 Instrumnet Name : P7
Client Sample ID : MYE4K1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:44:49 DataFile Name : 099SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	223344	230022	230146	227838	1.71	cps
Lead	207-1	188215	192848	190061	190375	1.23	cps
Lead	208-1	867326	891039	882974	880446	1.37	cps
Lithium	6-1	1316738	1181743	1302000	1266827	5.85	cps
Magnesium	24-2	6140340	6199627	6137126	6159031	0.57	cps
Manganese	55-2	4931622	4909614	4901287	4914174	0.32	cps
Molybdenum	94-1	70036	72649	73466	72051	2.49	cps
Molybdenum	95-1	5808	6361	6025	6065	4.60	cps
Molybdenum	96-1	21870	22431	22187	22162	1.27	cps
Molybdenum	97-1	3600	3807	3794	3734	3.10	cps
Molybdenum	98-1	9176	9210	9276	9221	0.55	cps
Neodymium	150-1	124531	127100	125109	125580	1.07	cps
Neodymium	150-2	65618	65558	65247	65474	0.30	cps
Nickel	60-2	108888	108623	107529	108347	0.67	cps
Phosphorus	31-2	30023	30601	29826	30150	1.34	cps
Potassium	39-2	3332355	3291797	3278529	3300894	0.85	cps
Rhodium	103-1	7495592	6473193	7533683	7167489	8.39	cps
Rhodium	103-2	4381250	4403216	4389513	4391326	0.25	cps
Scandium	45-1	4637793	3977930	4681517	4432413	8.89	cps
Scandium	45-2	420528	414313	415167	416670	0.81	cps
Selenium	82-1	561	449	613	541	15.49	cps
Selenium	77-2	237	247	273	252	7.52	cps
Selenium	78-2	450	487	447	461	4.81	cps
Silicon	28-1	25764344	26181500	26195273	26047039	0.94	cps
Silver	107-1	1590	1873	1803	1756	8.41	cps
Silver	109-1	1407	1543	1753	1568	11.14	cps
Sodium	23-2	204178	202736	202388	203101	0.47	cps
Strontium	86-1	251229	256138	258251	255206	1.41	cps
Strontium	88-1	2264122	2327984	2337261	2309789	1.72	cps
Sulfur	34-1	129258	129950	131144	130117	0.73	cps
Terbium	159-1	11057756	9641407	11173948	10624371	8.03	cps
Terbium	159-2	6659244	6530997	6551445	6580562	1.05	cps
Thallium	203-1	2560	2660	2700	2640	2.73	cps
Thallium	205-1	6522	6428	6518	6489	0.82	cps
Tin	118-1	19280	20768	20972	20340	4.54	cps
Titanium	47-1	1808715	1822628	1863064	1831469	1.54	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-02 Instrumnet Name : P7
Client Sample ID : MYE4K1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:44:49 DataFile Name : 099SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	30899	32022	31233	31385	1.84	cps
Vanadium	51-2	728309	725343	723505	725719	0.33	cps
Yttrium	89-1	14028849	12276971	14164465	13490095	7.80	cps
Yttrium	89-2	3444335	3457573	3398005	3433304	0.91	cps
Zinc	66-2	113061	110763	110539	111454	1.25	cps
Zirconium	90-1	182171	188695	187477	186115	1.86	cps
Zirconium	91-1	40623	41605	41719	41316	1.46	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03 Instrumnet Name : P7
Client Sample ID : MYE4K2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:47:57 DataFile Name : 100SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7137766	7022162	7102727	7087551	0.84	cps
Antimony	121-1	4707	4888	4781	4792	1.89	cps
Arsenic	75-2	20297	20070	20898	20422	2.09	cps
Barium	135-1	849279	856459	871177	858972	1.30	cps
Barium	137-1	1606121	1623495	1599984	1609867	0.76	cps
Beryllium	9-1	1556	1472	1422	1483	4.54	cps
Bismuth	209-1	6925859	6876930	6827314	6876701	0.72	cps
Bismuth	209-2	4647965	4395723	4638516	4560735	3.14	cps
Bromine	81-1	15865	16776	16816	16486	3.26	cps
Bromine	81-2	1130	1243	1257	1210	5.75	cps
Cadmium	108-1	167	180	183	177	4.99	cps
Cadmium	106-1	2917	2954	3047	2973	2.26	cps
Cadmium	111-1	3425	3508	3498	3477	1.30	cps
Calcium	43-1	701864	717018	715947	711610	1.19	cps
Calcium	44-1	10671712	10861620	10861607	10798313	1.02	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	283123	281211	279153	281163	0.71	cps
Cobalt	59-2	249899	245883	247312	247698	0.82	cps
Copper	63-2	551579	550148	552527	551418	0.22	cps
Dysprosium	156-1	66419	67103	68439	67320	1.53	cps
Dysprosium	156-2	50343	48875	50283	49834	1.67	cps
Erbium	164-1	57515	59107	59750	58791	1.96	cps
Erbium	164-2	35375	35749	36026	35717	0.92	cps
Gadolinium	160-1	64893	65251	65502	65215	0.47	cps
Gadolinium	160-2	43541	43186	42885	43204	0.76	cps
Holmium	165-1	10911127	10977409	10738245	10875594	1.14	cps
Holmium	165-2	6601727	6273976	6601688	6492464	2.91	cps
Indium	115-1	8353042	8479088	8331063	8387731	0.95	cps
Indium	115-2	2586663	2423418	2551168	2520416	3.41	cps
Iron	56-2	244333723	244698190	243270283	244100732	0.30	cps
Iron	57-2	6021370	6008505	6041841	6023906	0.28	cps
Iron	54-2	13451578	13182627	13385934	13340046	1.05	cps
Krypton	83-1	253	277	267	266	4.41	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03 Instrumnet Name : P7
Client Sample ID : MYE4K2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:47:57 DataFile Name : 100SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	261983	262429	265438	263283	0.71	cps
Lead	207-1	218706	225829	218903	221146	1.83	cps
Lead	208-1	1019058	1027230	1023751	1023346	0.40	cps
Lithium	6-1	1276102	1300763	1331718	1302861	2.14	cps
Magnesium	24-2	6487007	6372297	6477104	6445469	0.99	cps
Manganese	55-2	6381492	6292732	6390422	6354882	0.85	cps
Molybdenum	94-1	35576	36559	36896	36344	1.89	cps
Molybdenum	95-1	6278	6255	5921	6151	3.24	cps
Molybdenum	96-1	15578	15395	15211	15395	1.19	cps
Molybdenum	97-1	3827	3764	3774	3788	0.90	cps
Molybdenum	98-1	9196	9240	9376	9271	1.01	cps
Neodymium	150-1	116459	115542	117531	116511	0.85	cps
Neodymium	150-2	60104	58989	59843	59645	0.98	cps
Nickel	60-2	90628	90528	90622	90593	0.06	cps
Phosphorus	31-2	29215	28764	28674	28884	1.00	cps
Potassium	39-2	4991271	4951415	4978719	4973802	0.41	cps
Rhodium	103-1	7626741	7695956	7596002	7639566	0.67	cps
Rhodium	103-2	4438990	4232977	4471957	4381308	2.96	cps
Scandium	45-1	4645599	4690500	4678490	4671529	0.50	cps
Scandium	45-2	425879	399350	416531	413920	3.25	cps
Selenium	82-1	594	580	538	571	5.05	cps
Selenium	77-2	227	207	217	217	4.62	cps
Selenium	78-2	483	510	467	487	4.49	cps
Silicon	28-1	35617078	36095573	36393000	36035217	1.09	cps
Silver	107-1	2644	3060	3067	2924	8.30	cps
Silver	109-1	2577	2787	2920	2761	6.27	cps
Sodium	23-2	303361	301346	303391	302699	0.39	cps
Strontium	86-1	338602	345588	345692	343294	1.18	cps
Strontium	88-1	3068550	3103039	3134576	3102055	1.06	cps
Sulfur	34-1	139241	140334	140028	139867	0.40	cps
Terbium	159-1	11160429	11460880	11096494	11239267	1.73	cps
Terbium	159-2	6638646	6216332	6644823	6499934	3.78	cps
Thallium	203-1	2760	2840	2777	2793	1.51	cps
Thallium	205-1	6428	6398	6745	6524	2.95	cps
Tin	118-1	18312	18592	18319	18407	0.87	cps
Titanium	47-1	1294541	1320412	1334775	1316576	1.55	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03 Instrumnet Name : P7
Client Sample ID : MYE4K2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:47:57 DataFile Name : 100SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	32764	32262	33261	32762	1.52	cps
Vanadium	51-2	493040	493304	492869	493071	0.04	cps
Yttrium	89-1	14476730	14472378	14670756	14539955	0.78	cps
Yttrium	89-2	3489410	3259406	3454524	3401113	3.64	cps
Zinc	66-2	123243	123035	120531	122270	1.23	cps
Zirconium	90-1	89041	89400	90795	89745	1.03	cps
Zirconium	91-1	20020	20201	20755	20325	1.88	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-04 Instrumnet Name : P7
Client Sample ID : MYE4K2D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:51:02 DataFile Name : 101SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7103653	7141633	7149940	7131742	0.35	cps
Antimony	121-1	4898	4854	4811	4854	0.89	cps
Arsenic	75-2	20498	20788	20608	20631	0.71	cps
Barium	135-1	854606	871635	873961	866734	1.22	cps
Barium	137-1	1622929	1644341	1659966	1642412	1.13	cps
Beryllium	9-1	1650	1595	1583	1609	2.22	cps
Bismuth	209-1	6821050	6809894	6746668	6792537	0.59	cps
Bismuth	209-2	4574582	4841371	4652543	4689499	2.93	cps
Bromine	81-1	16409	17160	16720	16763	2.25	cps
Bromine	81-2	1113	1230	1207	1183	5.22	cps
Cadmium	108-1	237	233	180	217	14.68	cps
Cadmium	106-1	3124	2887	3110	3040	4.37	cps
Cadmium	111-1	3554	3451	3558	3521	1.73	cps
Calcium	43-1	704330	718440	724901	715890	1.47	cps
Calcium	44-1	10811722	10947667	11030862	10930084	1.01	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	286036	285942	283900	285292	0.42	cps
Cobalt	59-2	248370	249484	248228	248694	0.28	cps
Copper	63-2	558009	557098	554723	556610	0.30	cps
Dysprosium	156-1	67136	68539	68435	68037	1.15	cps
Dysprosium	156-2	49604	50761	50624	50330	1.26	cps
Erbium	164-1	59563	59764	59596	59641	0.18	cps
Erbium	164-2	35886	35819	36307	36004	0.73	cps
Gadolinium	160-1	66905	66272	67645	66941	1.03	cps
Gadolinium	160-2	42785	43387	43390	43187	0.81	cps
Holmium	165-1	10832797	10995977	10807501	10878758	0.94	cps
Holmium	165-2	6465674	6717935	6539238	6574282	1.97	cps
Indium	115-1	8302970	8333149	8339120	8325080	0.23	cps
Indium	115-2	2481621	2627079	2609939	2572880	3.09	cps
Iron	56-2	246886003	245136296	243190416	245070905	0.75	cps
Iron	57-2	6040185	6141850	6054507	6078847	0.91	cps
Iron	54-2	13429382	13438069	13312941	13393464	0.52	cps
Krypton	83-1	270	300	277	282	5.58	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-04 Instrumnet Name : P7
Client Sample ID : MYE4K2D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:51:02 DataFile Name : 101SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	260228	265177	264262	263222	1.00	cps
Lead	207-1	220378	226531	226035	224315	1.52	cps
Lead	208-1	1021262	1037670	1041047	1033326	1.02	cps
Lithium	6-1	1276838	1295982	1256669	1276496	1.54	cps
Magnesium	24-2	6513722	6464975	6515179	6497959	0.44	cps
Manganese	55-2	6366789	6366464	6306460	6346571	0.55	cps
Molybdenum	94-1	36408	37257	37114	36926	1.23	cps
Molybdenum	95-1	6335	6208	5981	6175	2.90	cps
Molybdenum	96-1	15932	15852	16166	15983	1.02	cps
Molybdenum	97-1	3457	3711	3891	3686	5.91	cps
Molybdenum	98-1	9026	9386	9283	9232	2.01	cps
Neodymium	150-1	117511	117891	119407	118270	0.85	cps
Neodymium	150-2	61895	60436	60807	61046	1.24	cps
Nickel	60-2	90665	91926	91725	91439	0.74	cps
Phosphorus	31-2	29468	29355	29108	29310	0.63	cps
Potassium	39-2	4994578	4976924	4937783	4969762	0.58	cps
Rhodium	103-1	7647128	7677025	7681692	7668615	0.24	cps
Rhodium	103-2	4322684	4557138	4442113	4440645	2.64	cps
Scandium	45-1	4718191	4729291	4674145	4707209	0.62	cps
Scandium	45-2	408291	427570	418593	418151	2.31	cps
Selenium	82-1	631	588	565	595	5.64	cps
Selenium	77-2	187	187	220	198	9.73	cps
Selenium	78-2	440	370	463	424	11.44	cps
Silicon	28-1	36169192	36458837	36396124	36341384	0.42	cps
Silver	107-1	2687	2997	2957	2880	5.86	cps
Silver	109-1	2567	2770	3037	2791	8.45	cps
Sodium	23-2	301834	304407	307187	304476	0.88	cps
Strontium	86-1	343115	345092	348839	345682	0.84	cps
Strontium	88-1	3097571	3151804	3134123	3127833	0.88	cps
Sulfur	34-1	138205	138070	139520	138598	0.58	cps
Terbium	159-1	11000280	11401507	11425617	11275801	2.12	cps
Terbium	159-2	6453317	6759419	6617372	6610036	2.32	cps
Thallium	203-1	2777	2817	2874	2823	1.72	cps
Thallium	205-1	6465	6635	6575	6558	1.32	cps
Tin	118-1	17878	18656	19003	18512	3.11	cps
Titanium	47-1	1332889	1335559	1354806	1341085	0.89	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-04 Instrumnet Name : P7
Client Sample ID : MYE4K2D Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:51:02 DataFile Name : 101SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	33556	33823	34187	33855	0.94	cps
Vanadium	51-2	494182	497220	493871	495091	0.37	cps
Yttrium	89-1	14365665	14600414	14356535	14440871	0.96	cps
Yttrium	89-2	3403303	3514923	3475244	3464490	1.63	cps
Zinc	66-2	123677	123247	123374	123433	0.18	cps
Zirconium	90-1	90490	92321	90295	91035	1.23	cps
Zirconium	91-1	20411	20378	20575	20454	0.52	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03LX5 Instrumnet Name : P7
Client Sample ID : MYE4K2L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 22:54:10 DataFile Name : 102SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1351316	1348371	1379988	1359892	1.28	cps
Antimony	121-1	947	953	993	964	2.62	cps
Arsenic	75-2	3751	4111	4031	3964	4.77	cps
Barium	135-1	164185	164826	165925	164978	0.53	cps
Barium	137-1	282770	287276	284810	284952	0.79	cps
Beryllium	9-1	400	433	328	387	13.94	cps
Bismuth	209-1	6895182	6300417	6753327	6649642	4.67	cps
Bismuth	209-2	4863494	4743762	4865096	4824117	1.44	cps
Bromine	81-1	11898	12118	12489	12168	2.45	cps
Bromine	81-2	397	257	303	319	22.36	cps
Cadmium	108-1	57	30	67	51	37.09	cps
Cadmium	106-1	2537	2290	2470	2432	5.25	cps
Cadmium	111-1	2060	1849	1963	1957	5.39	cps
Calcium	43-1	134963	134432	135663	135019	0.46	cps
Calcium	44-1	2054891	2059929	2058916	2057912	0.13	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	53357	52936	53291	53195	0.43	cps
Cobalt	59-2	47110	46562	46539	46737	0.69	cps
Copper	63-2	106344	106203	106613	106387	0.20	cps
Dysprosium	156-1	13003	13173	12953	13043	0.89	cps
Dysprosium	156-2	9957	9880	9617	9818	1.82	cps
Erbium	164-1	11431	10928	10978	11112	2.50	cps
Erbium	164-2	6898	6898	7155	6984	2.12	cps
Gadolinium	160-1	12145	12809	12629	12528	2.74	cps
Gadolinium	160-2	8383	8076	8539	8333	2.83	cps
Holmium	165-1	10882914	9790452	10618968	10430778	5.46	cps
Holmium	165-2	6728610	6593701	6663891	6662067	1.01	cps
Indium	115-1	8393867	7492293	8153738	8013299	5.83	cps
Indium	115-2	2621600	2601657	2626392	2616550	0.50	cps
Iron	56-2	46172593	45667364	45378733	45739563	0.88	cps
Iron	57-2	1170799	1164286	1172829	1169305	0.38	cps
Iron	54-2	2531915	2549653	2499831	2527133	1.00	cps
Krypton	83-1	163	220	187	190	14.99	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03LX5 Instrumnet Name : P7
Client Sample ID : MYE4K2L Dilution Factor : 5
Date & Time Acquired : 2024-11-02 22:54:10 DataFile Name : 102SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	51092	51885	52314	51764	1.20	cps
Lead	207-1	42864	43436	43038	43113	0.68	cps
Lead	208-1	198723	199439	200744	199636	0.51	cps
Lithium	6-1	1272646	1196159	1262163	1243656	3.33	cps
Magnesium	24-2	1219276	1217678	1221511	1219488	0.16	cps
Manganese	55-2	1191265	1184796	1191228	1189096	0.31	cps
Molybdenum	94-1	7012	6948	6905	6955	0.77	cps
Molybdenum	95-1	1207	1260	1257	1241	2.41	cps
Molybdenum	96-1	3027	3107	3220	3118	3.12	cps
Molybdenum	97-1	783	897	710	797	11.81	cps
Molybdenum	98-1	1907	1947	1847	1900	2.65	cps
Neodymium	150-1	21343	22521	22698	22188	3.32	cps
Neodymium	150-2	11558	11481	11652	11564	0.74	cps
Nickel	60-2	17353	16866	17407	17209	1.73	cps
Phosphorus	31-2	5788	5568	5935	5763	3.20	cps
Potassium	39-2	1022608	1021831	1022689	1022376	0.05	cps
Rhodium	103-1	7709203	6898483	7540876	7382854	5.80	cps
Rhodium	103-2	4546717	4437552	4538470	4507579	1.35	cps
Scandium	45-1	4439803	3981195	4344032	4255010	5.69	cps
Scandium	45-2	396624	397355	397367	397115	0.11	cps
Selenium	82-1	513	428	399	447	13.33	cps
Selenium	77-2	63	40	37	47	31.13	cps
Selenium	78-2	267	250	263	260	3.39	cps
Silicon	28-1	7009112	7080638	7083772	7057841	0.60	cps
Silver	107-1	613	590	643	616	4.34	cps
Silver	109-1	613	530	590	578	7.44	cps
Sodium	23-2	72034	72650	72932	72539	0.63	cps
Strontium	86-1	65424	65183	65100	65236	0.26	cps
Strontium	88-1	566698	572394	570226	569773	0.50	cps
Sulfur	34-1	118237	117736	120658	118877	1.31	cps
Terbium	159-1	11075219	9982773	10882811	10646934	5.48	cps
Terbium	159-2	6550413	6622704	6670247	6614455	0.91	cps
Thallium	203-1	657	677	613	649	4.99	cps
Thallium	205-1	1547	1470	1550	1522	2.97	cps
Tin	118-1	5852	4651	4808	5103	12.80	cps
Titanium	47-1	249460	253054	251651	251388	0.72	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-03LX5

Instrumnet Name :

P7

Client Sample ID :

MYE4K2L

Dilution Factor :

5

Date & Time Acquired :

2024-11-02 22:54:10

DataFile Name :

102SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	6475	6675	6408	6519	2.13	cps
Vanadium	51-2	91394	91616	92736	91915	0.78	cps
Yttrium	89-1	13632221	12293534	13366922	13097559	5.41	cps
Yttrium	89-2	3285195	3221650	3299150	3268665	1.26	cps
Zinc	66-2	24023	23672	24126	23940	0.99	cps
Zirconium	90-1	16863	16960	17240	17021	1.15	cps
Zirconium	91-1	3791	4074	3831	3898	3.93	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-05 Instrumnet Name : P7
Client Sample ID : MYE4K2S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:57:20 DataFile Name : 103SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7151029	7119272	7103816	7124706	0.34	cps
Antimony	121-1	583077	591614	593887	589526	0.97	cps
Arsenic	75-2	31801	31220	31390	31470	0.95	cps
Barium	135-1	4189421	4170646	4135749	4165272	0.65	cps
Barium	137-1	7218196	7180093	7147149	7181813	0.50	cps
Beryllium	9-1	43617	44135	45026	44259	1.61	cps
Bismuth	209-1	6793078	6829645	6740632	6787785	0.66	cps
Bismuth	209-2	4774439	4739912	4731025	4748459	0.48	cps
Bromine	81-1	15885	16299	16623	16269	2.27	cps
Bromine	81-2	1200	1113	1143	1152	3.82	cps
Cadmium	108-1	4991	5058	4934	4994	1.24	cps
Cadmium	106-1	9680	9657	9803	9713	0.81	cps
Cadmium	111-1	77887	77375	77841	77701	0.36	cps
Calcium	43-1	679296	688572	694521	687463	1.12	cps
Calcium	44-1	10259772	10452044	10475049	10395622	1.14	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	766486	765494	761427	764469	0.35	cps
Cobalt	59-2	2161442	2147738	2139625	2149602	0.51	cps
Copper	63-2	1166979	1161324	1156537	1161613	0.45	cps
Dysprosium	156-1	65693	66841	66704	66413	0.94	cps
Dysprosium	156-2	49015	49032	49112	49053	0.11	cps
Erbium	164-1	57749	57230	59108	58029	1.67	cps
Erbium	164-2	35893	35558	34706	35386	1.73	cps
Gadolinium	160-1	64521	65566	64464	64850	0.96	cps
Gadolinium	160-2	41271	42003	42153	41809	1.13	cps
Holmium	165-1	10867835	10815202	10792346	10825128	0.36	cps
Holmium	165-2	6657486	6652701	6609295	6639827	0.40	cps
Indium	115-1	8244543	8280684	8116143	8213790	1.05	cps
Indium	115-2	2579414	2532365	2508622	2540134	1.42	cps
Iron	56-2	237491783	237709390	237134070	237445081	0.12	cps
Iron	57-2	5956093	5863400	5888486	5902660	0.81	cps
Iron	54-2	13087226	13093479	12979077	13053261	0.49	cps
Krypton	83-1	323	293	267	294	9.63	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-05 Instrumnet Name : P7
Client Sample ID : MYE4K2S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:57:20 DataFile Name : 103SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	342276	352731	345340	346782	1.55	cps
Lead	207-1	292541	294941	299710	295731	1.23	cps
Lead	208-1	1335479	1364855	1354655	1351663	1.10	cps
Lithium	6-1	1316101	1283481	1275734	1291772	1.66	cps
Magnesium	24-2	6169737	6188788	6208346	6188957	0.31	cps
Manganese	55-2	6985962	7015213	6917445	6972873	0.72	cps
Molybdenum	94-1	35493	35399	35981	35624	0.88	cps
Molybdenum	95-1	5928	5985	6048	5987	1.00	cps
Molybdenum	96-1	15165	15321	15502	15329	1.10	cps
Molybdenum	97-1	3547	3817	3547	3637	4.29	cps
Molybdenum	98-1	9186	8706	8876	8923	2.73	cps
Neodymium	150-1	113452	116668	113341	114487	1.65	cps
Neodymium	150-2	59049	59314	58631	58998	0.58	cps
Nickel	60-2	558767	554847	554445	556019	0.43	cps
Phosphorus	31-2	27996	28590	28450	28345	1.10	cps
Potassium	39-2	4720468	4822235	4678883	4740529	1.56	cps
Rhodium	103-1	7556140	7522629	7416008	7498259	0.98	cps
Rhodium	103-2	4424305	4416340	4359564	4400070	0.80	cps
Scandium	45-1	4546071	4551551	4534300	4543974	0.19	cps
Scandium	45-2	409680	406862	405712	407418	0.50	cps
Selenium	82-1	9412	9305	9656	9458	1.91	cps
Selenium	77-2	997	1020	1057	1024	2.95	cps
Selenium	78-2	3000	3060	3194	3085	3.21	cps
Silicon	28-1	34615924	34924278	35281310	34940504	0.95	cps
Silver	107-1	310334	316612	318752	315233	1.39	cps
Silver	109-1	293434	301220	301947	298867	1.58	cps
Sodium	23-2	292003	291883	291169	291685	0.15	cps
Strontium	86-1	329092	332335	333443	331623	0.68	cps
Strontium	88-1	2932119	2990043	3029464	2983875	1.64	cps
Sulfur	34-1	135861	134277	133224	134454	0.99	cps
Terbium	159-1	11044575	11298046	11120152	11154258	1.17	cps
Terbium	159-2	6619193	6668388	6540179	6609254	0.98	cps
Thallium	203-1	269260	273471	275400	272710	1.15	cps
Thallium	205-1	640189	652542	651204	647978	1.05	cps
Tin	118-1	18021	18335	18298	18218	0.94	cps
Titanium	47-1	1272657	1275643	1273934	1274078	0.12	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-05 Instrumnet Name : P7
Client Sample ID : MYE4K2S Dilution Factor : 1
Date & Time Acquired : 2024-11-02 22:57:20 DataFile Name : 103SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	32052	33215	32994	32754	1.89	cps
Vanadium	51-2	1537384	1535651	1545288	1539441	0.33	cps
Yttrium	89-1	14141854	14128878	14120499	14130410	0.08	cps
Yttrium	89-2	3417323	3397565	3339334	3384741	1.20	cps
Zinc	66-2	325702	323822	321542	323689	0.64	cps
Zirconium	90-1	86171	89128	87975	87758	1.70	cps
Zirconium	91-1	19513	19463	19590	19522	0.33	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03MS Instrumnet Name : P7
Client Sample ID : MYE4K2SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:00:23 DataFile Name : 104SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7692481	7778057	7696049	7722196	0.63	cps
Antimony	121-1	2949006	2954307	2960847	2954720	0.20	cps
Arsenic	75-2	73832	75651	73467	74317	1.57	cps
Barium	135-1	16420102	16582739	16470200	16491014	0.51	cps
Barium	137-1	28337391	28995183	28554863	28629146	1.17	cps
Beryllium	9-1	202996	205517	207576	205363	1.12	cps
Bismuth	209-1	6856909	6880093	6784004	6840335	0.73	cps
Bismuth	209-2	4725448	4727508	4674937	4709298	0.63	cps
Bromine	81-1	15762	16046	16136	15981	1.22	cps
Bromine	81-2	1170	1160	1080	1137	4.34	cps
Cadmium	108-1	23092	23232	23733	23352	1.44	cps
Cadmium	106-1	34270	34544	34785	34533	0.75	cps
Cadmium	111-1	352879	359988	363499	358788	1.51	cps
Calcium	43-1	649638	656149	661574	655787	0.91	cps
Calcium	44-1	9864933	10003120	9881777	9916610	0.76	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2598034	2616848	2581990	2598957	0.67	cps
Cobalt	59-2	9319724	9246741	9210324	9258930	0.60	cps
Copper	63-2	3656862	3654858	3666341	3659354	0.17	cps
Dysprosium	156-1	63908	64604	65572	64695	1.29	cps
Dysprosium	156-2	47232	47490	48203	47642	1.05	cps
Erbium	164-1	54757	56242	56838	55946	1.92	cps
Erbium	164-2	34700	34426	34713	34613	0.47	cps
Gadolinium	160-1	61407	62030	64025	62487	2.19	cps
Gadolinium	160-2	41254	40759	40642	40885	0.79	cps
Holmium	165-1	10917799	10870981	10703493	10830758	1.04	cps
Holmium	165-2	6562514	6604948	6617320	6594927	0.44	cps
Indium	115-1	8129420	8048863	8080998	8086427	0.50	cps
Indium	115-2	2498031	2494519	2498589	2497046	0.09	cps
Iron	56-2	235818836	236530710	235448563	235932703	0.23	cps
Iron	57-2	5831643	5798031	5816217	5815297	0.29	cps
Iron	54-2	12869529	12985375	12774995	12876633	0.82	cps
Krypton	83-1	280	273	310	288	6.79	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03MS Instrumnet Name : P7
Client Sample ID : MYE4K2SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:00:23 DataFile Name : 104SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	661406	680618	676486	672837	1.50	cps
Lead	207-1	574110	584215	579698	579341	0.87	cps
Lead	208-1	2627793	2684634	2672606	2661678	1.13	cps
Lithium	6-1	1301708	1295377	1295158	1297414	0.29	cps
Magnesium	24-2	5944338	5939224	5914929	5932831	0.26	cps
Manganese	55-2	9940196	9975700	9995170	9970356	0.28	cps
Molybdenum	94-1	34113	34427	34434	34325	0.53	cps
Molybdenum	95-1	6041	5791	6111	5981	2.81	cps
Molybdenum	96-1	15045	14824	15034	14968	0.83	cps
Molybdenum	97-1	3454	3534	3534	3507	1.32	cps
Molybdenum	98-1	8883	9183	9230	9098	2.07	cps
Neodymium	150-1	111043	112521	113448	112338	1.08	cps
Neodymium	150-2	57306	56553	56366	56741	0.88	cps
Nickel	60-2	2444984	2459973	2419792	2441583	0.83	cps
Phosphorus	31-2	26299	27001	26359	26553	1.46	cps
Potassium	39-2	4522858	4533939	4480289	4512362	0.63	cps
Rhodium	103-1	7403797	7408323	7355338	7389153	0.40	cps
Rhodium	103-2	4343159	4339984	4295665	4326269	0.61	cps
Scandium	45-1	4515773	4448436	4453481	4472563	0.84	cps
Scandium	45-2	398326	399525	396988	398280	0.32	cps
Selenium	82-1	48422	48873	48966	48754	0.60	cps
Selenium	77-2	4574	4787	4657	4673	2.30	cps
Selenium	78-2	14747	14811	15054	14871	1.09	cps
Silicon	28-1	33236996	33769210	33672820	33559675	0.84	cps
Silver	107-1	1475288	1580270	1613696	1556418	4.64	cps
Silver	109-1	1397383	1422238	1420331	1413317	0.98	cps
Sodium	23-2	276254	278313	278242	277603	0.42	cps
Strontium	86-1	312660	319908	320325	317631	1.36	cps
Strontium	88-1	2886407	2884294	2867686	2879462	0.36	cps
Sulfur	34-1	131490	129449	128820	129920	1.07	cps
Terbium	159-1	11086316	10998159	11172878	11085784	0.79	cps
Terbium	159-2	6607720	6539898	6569133	6572250	0.52	cps
Thallium	203-1	1285494	1311216	1312549	1303086	1.17	cps
Thallium	205-1	3564341	3640386	3613648	3606125	1.07	cps
Tin	118-1	17270	17574	17671	17505	1.19	cps
Titanium	47-1	1214957	1229227	1235719	1226634	0.87	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03MS Instrumnet Name : P7
Client Sample ID : MYE4K2SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:00:23 DataFile Name : 104SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	31464	33051	32774	32429	2.61	cps
Vanadium	51-2	5490544	5471936	5432874	5465118	0.54	cps
Yttrium	89-1	13977139	13849543	13877354	13901345	0.48	cps
Yttrium	89-2	3338226	3297699	3317506	3317810	0.61	cps
Zinc	66-2	1091831	1098025	1085499	1091785	0.57	cps
Zirconium	90-1	82986	84055	84732	83924	1.05	cps
Zirconium	91-1	19176	18805	19810	19264	2.64	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-06 Instrumnet Name : P7
Client Sample ID : MYE4K3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:03:19 DataFile Name : 105SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	6080260	6140860	6089438	6103519	0.54	cps
Antimony	121-1	9847	9820	9633	9767	1.19	cps
Arsenic	75-2	26437	26053	26046	26179	0.85	cps
Barium	135-1	804630	814327	821335	813431	1.03	cps
Barium	137-1	1392527	1413478	1416916	1407640	0.94	cps
Beryllium	9-1	1489	1506	1656	1550	5.92	cps
Bismuth	209-1	6829416	6925255	6722651	6825774	1.48	cps
Bismuth	209-2	4752878	4704365	4746072	4734439	0.55	cps
Bromine	81-1	16119	16459	16830	16469	2.16	cps
Bromine	81-2	1260	1243	1210	1238	2.06	cps
Cadmium	108-1	283	250	263	266	6.32	cps
Cadmium	106-1	3107	3020	3144	3090	2.05	cps
Cadmium	111-1	4160	4236	4286	4228	1.50	cps
Calcium	43-1	553213	563884	570899	562665	1.58	cps
Calcium	44-1	8454319	8550861	8591304	8532161	0.82	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	249809	249237	247058	248702	0.58	cps
Cobalt	59-2	216858	215677	215444	215993	0.35	cps
Copper	63-2	717275	715318	715247	715946	0.16	cps
Dysprosium	156-1	60406	60362	61584	60784	1.14	cps
Dysprosium	156-2	46647	45794	45761	46068	1.09	cps
Erbium	164-1	50645	50959	51558	51054	0.91	cps
Erbium	164-2	31933	30804	31659	31465	1.87	cps
Gadolinium	160-1	58143	58481	58789	58471	0.55	cps
Gadolinium	160-2	37911	38115	38112	38046	0.31	cps
Holmium	165-1	10708774	10892579	10836477	10812610	0.87	cps
Holmium	165-2	6642883	6579289	6601402	6607858	0.49	cps
Indium	115-1	8111894	8074930	8123446	8103423	0.31	cps
Indium	115-2	2545638	2529345	2566116	2547033	0.72	cps
Iron	56-2	219001803	220725603	217403183	219043530	0.76	cps
Iron	57-2	5477657	5407290	5424402	5436450	0.68	cps
Iron	54-2	12052393	12010374	11903959	11988909	0.64	cps
Krypton	83-1	257	280	270	269	4.35	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-06 Instrumnet Name : P7
Client Sample ID : MYE4K3 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:03:19 DataFile Name : 105SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	319272	320229	322041	320514	0.44	cps
Lead	207-1	273309	276658	282145	277371	1.61	cps
Lead	208-1	1252863	1260613	1269380	1260952	0.66	cps
Lithium	6-1	1290312	1280911	1289357	1286860	0.40	cps
Magnesium	24-2	5994789	5957924	5981178	5977964	0.31	cps
Manganese	55-2	4764130	4843202	4758295	4788542	0.99	cps
Molybdenum	94-1	39747	40576	40623	40315	1.22	cps
Molybdenum	95-1	7682	7555	7612	7616	0.83	cps
Molybdenum	96-1	16893	17634	17414	17314	2.20	cps
Molybdenum	97-1	4707	4707	4711	4709	0.04	cps
Molybdenum	98-1	11651	11431	11571	11551	0.96	cps
Neodymium	150-1	107284	110240	110281	109268	1.57	cps
Neodymium	150-2	56660	56098	56539	56432	0.52	cps
Nickel	60-2	95713	96313	95246	95757	0.56	cps
Phosphorus	31-2	34188	34389	34268	34282	0.29	cps
Potassium	39-2	4624531	4670380	4570516	4621809	1.08	cps
Rhodium	103-1	7316528	7381603	7258121	7318751	0.84	cps
Rhodium	103-2	4409848	4414901	4452079	4425609	0.52	cps
Scandium	45-1	4418852	4477165	4384224	4426747	1.06	cps
Scandium	45-2	407168	404143	405359	405557	0.38	cps
Selenium	82-1	558	599	600	586	4.10	cps
Selenium	77-2	230	193	200	208	9.40	cps
Selenium	78-2	427	423	387	412	5.38	cps
Silicon	28-1	33126166	33090235	33300723	33172375	0.34	cps
Silver	107-1	6131	5548	5181	5620	8.53	cps
Silver	109-1	5471	4821	4661	4984	8.61	cps
Sodium	23-2	439593	442506	442906	441668	0.41	cps
Strontium	86-1	362738	367197	370565	366833	1.07	cps
Strontium	88-1	3294509	3329385	3322507	3315467	0.56	cps
Sulfur	34-1	144061	144428	145694	144728	0.59	cps
Terbium	159-1	10925529	11119495	10992413	11012479	0.89	cps
Terbium	159-2	6493800	6634102	6751777	6626560	1.95	cps
Thallium	203-1	2654	2684	2747	2695	1.77	cps
Thallium	205-1	6515	6592	6562	6556	0.59	cps
Tin	118-1	24187	24484	24534	24402	0.77	cps
Titanium	47-1	1061037	1075304	1088783	1075041	1.29	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-06

Instrumnet Name :

P7

Client Sample ID :

MYE4K3

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:03:19

DataFile Name :

105SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	25959	25858	26723	26180	1.81	cps
Vanadium	51-2	444291	446180	441330	443934	0.55	cps
Yttrium	89-1	13746449	13814788	13737450	13766229	0.31	cps
Yttrium	89-2	3356842	3377856	3380448	3371715	0.38	cps
Zinc	66-2	134681	135734	135676	135364	0.44	cps
Zirconium	90-1	96705	98893	98816	98138	1.27	cps
Zirconium	91-1	22127	21760	22367	22085	1.39	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-07 Instrumnet Name : P7
Client Sample ID : MYE4K4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:06:28 DataFile Name : 106SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7502374	7460512	7431465	7464784	0.48	cps
Antimony	121-1	5611	5444	5741	5599	2.66	cps
Arsenic	75-2	22981	22607	23195	22928	1.30	cps
Barium	135-1	850722	867507	867416	861882	1.12	cps
Barium	137-1	1625916	1490601	1619058	1578525	4.83	cps
Beryllium	9-1	1750	1817	1589	1719	6.81	cps
Bismuth	209-1	6874807	6710226	6884589	6823207	1.44	cps
Bismuth	209-2	4680348	4627361	4683845	4663851	0.68	cps
Bromine	81-1	17527	17968	18278	17924	2.11	cps
Bromine	81-2	1353	1507	1453	1438	5.41	cps
Cadmium	108-1	213	217	200	210	4.20	cps
Cadmium	106-1	3094	3134	2960	3063	2.96	cps
Cadmium	111-1	4003	4125	3962	4030	2.11	cps
Calcium	43-1	542556	549063	548444	546688	0.66	cps
Calcium	44-1	8176933	8339744	8368287	8294988	1.24	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	288054	287915	285889	287286	0.42	cps
Cobalt	59-2	235421	234629	235059	235037	0.17	cps
Copper	63-2	689855	698277	689131	692421	0.73	cps
Dysprosium	156-1	67712	68322	69062	68365	0.99	cps
Dysprosium	156-2	51404	51015	50530	50983	0.86	cps
Erbium	164-1	59677	59355	59439	59490	0.28	cps
Erbium	164-2	36441	36274	36304	36339	0.25	cps
Gadolinium	160-1	67448	68319	67746	67837	0.65	cps
Gadolinium	160-2	44089	43905	43150	43715	1.14	cps
Holmium	165-1	10663003	10640471	10641024	10648166	0.12	cps
Holmium	165-2	6548274	6534186	6621152	6567871	0.71	cps
Indium	115-1	8093966	7875504	8055437	8008302	1.46	cps
Indium	115-2	2513345	2496142	2522337	2510608	0.53	cps
Iron	56-2	238424390	234512510	235793736	236243545	0.84	cps
Iron	57-2	5864964	5865148	5825663	5851925	0.39	cps
Iron	54-2	12950979	12838585	12898068	12895877	0.44	cps
Krypton	83-1	223	293	260	259	13.52	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-07 Instrumnet Name : P7
Client Sample ID : MYE4K4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:06:28 DataFile Name : 106SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	285057	288557	290983	288199	1.03	cps
Lead	207-1	241206	248968	247113	245762	1.65	cps
Lead	208-1	1115914	1135650	1138338	1129967	1.08	cps
Lithium	6-1	1276362	1282709	1263693	1274255	0.76	cps
Magnesium	24-2	6284537	6272799	6302235	6286524	0.24	cps
Manganese	55-2	5685074	5648068	5636940	5656694	0.45	cps
Molybdenum	94-1	38671	39861	39226	39253	1.52	cps
Molybdenum	95-1	6775	6738	6815	6776	0.57	cps
Molybdenum	96-1	16536	16543	17247	16775	2.43	cps
Molybdenum	97-1	4331	4091	4091	4171	3.32	cps
Molybdenum	98-1	10204	10410	10487	10367	1.41	cps
Neodymium	150-1	119061	121108	121814	120661	1.19	cps
Neodymium	150-2	61895	61694	62079	61889	0.31	cps
Nickel	60-2	90011	90078	90397	90162	0.23	cps
Phosphorus	31-2	28433	27605	27161	27733	2.33	cps
Potassium	39-2	5042263	5036068	4939456	5005929	1.15	cps
Rhodium	103-1	7414727	7225241	7485487	7375152	1.82	cps
Rhodium	103-2	4362757	4330457	4352616	4348610	0.38	cps
Scandium	45-1	4479043	4413373	4519557	4470657	1.20	cps
Scandium	45-2	404780	405059	401215	403684	0.53	cps
Selenium	82-1	587	562	555	568	2.92	cps
Selenium	77-2	220	187	223	210	9.66	cps
Selenium	78-2	427	363	447	412	10.55	cps
Silicon	28-1	33575923	34116279	34148651	33946951	0.95	cps
Silver	107-1	3941	4331	3924	4065	5.66	cps
Silver	109-1	3944	3781	3811	3845	2.26	cps
Sodium	23-2	237026	236160	235624	236270	0.30	cps
Strontium	86-1	303442	307850	308901	306731	0.94	cps
Strontium	88-1	2765325	2817227	2777403	2786652	0.97	cps
Sulfur	34-1	133789	132478	131399	132556	0.90	cps
Terbium	159-1	11035975	10809618	11090579	10978724	1.36	cps
Terbium	159-2	6627323	6575499	6601749	6601524	0.39	cps
Thallium	203-1	2977	2897	2884	2919	1.73	cps
Thallium	205-1	6922	7112	7032	7022	1.36	cps
Tin	118-1	19300	19380	20017	19566	2.01	cps
Titanium	47-1	1229523	1238099	1253580	1240401	0.98	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-07 Instrumnet Name : P7
Client Sample ID : MYE4K4 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:06:28 DataFile Name : 106SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	29656	30291	31003	30317	2.22	cps
Vanadium	51-2	481277	479271	473813	478121	0.81	cps
Yttrium	89-1	13890445	13792242	13999636	13894108	0.75	cps
Yttrium	89-2	3381736	3366575	3358147	3368819	0.35	cps
Zinc	66-2	136760	136769	135851	136460	0.39	cps
Zirconium	90-1	94626	95756	96676	95686	1.07	cps
Zirconium	91-1	21205	21489	22210	21635	2.39	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-08 Instrumnet Name : P7
Client Sample ID : MYE4K5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:09:35 DataFile Name : 107SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	6041394	6145545	6063551	6083497	0.90	cps
Antimony	121-1	6418	6718	6665	6600	2.43	cps
Arsenic	75-2	25449	25432	25562	25481	0.28	cps
Barium	135-1	751756	759516	759906	757060	0.61	cps
Barium	137-1	1303123	1322959	1320919	1315667	0.83	cps
Beryllium	9-1	1628	1589	1211	1476	15.59	cps
Bismuth	209-1	6843776	6831830	6768083	6814563	0.60	cps
Bismuth	209-2	4637184	4697913	4678184	4671094	0.66	cps
Bromine	81-1	17918	18672	18685	18425	2.38	cps
Bromine	81-2	1447	1473	1500	1473	1.81	cps
Cadmium	108-1	200	233	210	214	7.98	cps
Cadmium	106-1	3147	3147	2887	3060	4.91	cps
Cadmium	111-1	4190	4083	4014	4096	2.16	cps
Calcium	43-1	495886	501713	504495	500698	0.88	cps
Calcium	44-1	7560130	7627562	7633438	7607044	0.54	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	261541	260531	261022	261031	0.19	cps
Cobalt	59-2	203771	202940	202500	203071	0.32	cps
Copper	63-2	606071	604237	601388	603899	0.39	cps
Dysprosium	156-1	56295	57474	56031	56600	1.36	cps
Dysprosium	156-2	41401	41240	40749	41130	0.83	cps
Erbium	164-1	48049	48744	48494	48429	0.73	cps
Erbium	164-2	29398	29344	28776	29173	1.18	cps
Gadolinium	160-1	53140	54994	54536	54223	1.78	cps
Gadolinium	160-2	35729	34860	35739	35442	1.42	cps
Holmium	165-1	10874869	10901996	10883334	10886733	0.13	cps
Holmium	165-2	6468438	6521605	6565107	6518383	0.74	cps
Indium	115-1	8106493	8110555	7984194	8067081	0.89	cps
Indium	115-2	2502762	2474818	2501984	2493188	0.64	cps
Iron	56-2	199247884	198492824	199401884	199047530	0.24	cps
Iron	57-2	4986586	4915467	4935325	4945792	0.74	cps
Iron	54-2	10831138	10777132	10813719	10807330	0.26	cps
Krypton	83-1	190	250	273	238	18.08	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-08 Instrumnet Name : P7
Client Sample ID : MYE4K5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:09:35 DataFile Name : 107SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	295434	299906	303434	299591	1.34	cps
Lead	207-1	254784	257028	255967	255926	0.44	cps
Lead	208-1	1161018	1170183	1180832	1170678	0.85	cps
Lithium	6-1	1289652	1277910	1294924	1287495	0.68	cps
Magnesium	24-2	5822732	5866742	5846324	5845266	0.38	cps
Manganese	55-2	4289435	4272287	4267891	4276537	0.27	cps
Molybdenum	94-1	33164	33993	33659	33605	1.24	cps
Molybdenum	95-1	8949	9093	8843	8962	1.40	cps
Molybdenum	96-1	17424	17414	17350	17396	0.23	cps
Molybdenum	97-1	5284	5464	5618	5456	3.06	cps
Molybdenum	98-1	14007	14447	13877	14110	2.12	cps
Neodymium	150-1	98857	97860	99148	98622	0.69	cps
Neodymium	150-2	50393	50269	51012	50558	0.79	cps
Nickel	60-2	89052	89803	90179	89678	0.64	cps
Phosphorus	31-2	29358	29746	29074	29393	1.15	cps
Potassium	39-2	4968853	4976357	4922194	4955801	0.59	cps
Rhodium	103-1	7391252	7434097	7410762	7412037	0.29	cps
Rhodium	103-2	4318877	4317210	4296735	4310941	0.29	cps
Scandium	45-1	4471461	4514848	4416876	4467729	1.10	cps
Scandium	45-2	398866	398614	396724	398068	0.29	cps
Selenium	82-1	673	581	562	605	9.78	cps
Selenium	77-2	203	187	190	193	4.56	cps
Selenium	78-2	390	383	367	380	3.16	cps
Silicon	28-1	34918029	35052904	34818279	34929737	0.34	cps
Silver	107-1	3847	3941	3931	3906	1.31	cps
Silver	109-1	3384	3664	3727	3592	5.09	cps
Sodium	23-2	441854	443586	442205	442549	0.21	cps
Strontium	86-1	298075	300089	301734	299966	0.61	cps
Strontium	88-1	2730232	2733570	2766449	2743417	0.73	cps
Sulfur	34-1	147375	146080	144206	145887	1.09	cps
Terbium	159-1	11045401	11318362	10967871	11110545	1.66	cps
Terbium	159-2	6507056	6551126	6556478	6538220	0.41	cps
Thallium	203-1	2337	2387	2530	2418	4.15	cps
Thallium	205-1	5591	5848	5901	5780	2.87	cps
Tin	118-1	18859	19036	20504	19467	4.64	cps
Titanium	47-1	997525	1016081	1017683	1010429	1.11	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-08 Instrumnet Name : P7
Client Sample ID : MYE4K5 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:09:35 DataFile Name : 107SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	23972	23725	24863	24186	2.48	cps
Vanadium	51-2	402530	398053	399827	400137	0.56	cps
Yttrium	89-1	13769109	13871730	13836437	13825759	0.38	cps
Yttrium	89-2	3341747	3349327	3320929	3337334	0.44	cps
Zinc	66-2	144323	144871	142875	144023	0.72	cps
Zirconium	90-1	76391	77640	78344	77458	1.28	cps
Zirconium	91-1	17667	17581	17754	17667	0.49	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-09 Instrumnet Name : P7
Client Sample ID : MYE4K6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:12:45 DataFile Name : 108SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	6187896	6023609	6123211	6111572	1.35	cps
Antimony	121-1	5001	5301	5358	5220	3.67	cps
Arsenic	75-2	19797	18899	19453	19383	2.34	cps
Barium	135-1	753717	764371	777730	765272	1.57	cps
Barium	137-1	1309256	1334670	1335448	1326458	1.12	cps
Beryllium	9-1	1295	1361	1439	1365	5.30	cps
Bismuth	209-1	6735783	6806273	6695803	6745953	0.83	cps
Bismuth	209-2	4609525	4609150	4648693	4622456	0.49	cps
Bromine	81-1	16589	17203	16756	16850	1.88	cps
Bromine	81-2	1327	1297	1327	1317	1.32	cps
Cadmium	108-1	207	177	170	184	10.59	cps
Cadmium	106-1	2887	2774	2734	2798	2.84	cps
Cadmium	111-1	3366	3321	3298	3329	1.04	cps
Calcium	43-1	564677	573183	573541	570467	0.88	cps
Calcium	44-1	8595964	8703799	8736826	8678863	0.85	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	228881	228841	229650	229124	0.20	cps
Cobalt	59-2	197326	197830	198579	197912	0.32	cps
Copper	63-2	515602	513334	514325	514420	0.22	cps
Dysprosium	156-1	58521	59016	60212	59250	1.47	cps
Dysprosium	156-2	44126	44323	44119	44189	0.26	cps
Erbium	164-1	50969	50501	51113	50861	0.63	cps
Erbium	164-2	30363	31258	30944	30855	1.47	cps
Gadolinium	160-1	56386	57179	57685	57083	1.15	cps
Gadolinium	160-2	37266	36701	36765	36911	0.84	cps
Holmium	165-1	10762093	10891645	10818138	10823958	0.60	cps
Holmium	165-2	6538780	6386353	6500102	6475078	1.22	cps
Indium	115-1	8022446	8060934	8013034	8032138	0.32	cps
Indium	115-2	2540602	2482825	2498873	2507433	1.19	cps
Iron	56-2	194509577	194540444	194712317	194587446	0.06	cps
Iron	57-2	4847967	4843980	4821807	4837918	0.29	cps
Iron	54-2	10612011	10692807	10646224	10650347	0.38	cps
Krypton	83-1	217	293	213	241	18.77	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-09 Instrumnet Name : P7
Client Sample ID : MYE4K6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:12:45 DataFile Name : 108SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	224109	229105	228876	227364	1.24	cps
Lead	207-1	189516	192866	192220	191534	0.93	cps
Lead	208-1	872066	886094	886373	881511	0.93	cps
Lithium	6-1	1259567	1261969	1255432	1258989	0.26	cps
Magnesium	24-2	5321758	5332220	5375514	5343164	0.53	cps
Manganese	55-2	4799990	4802637	4756005	4786211	0.55	cps
Molybdenum	94-1	33375	33869	34424	33889	1.55	cps
Molybdenum	95-1	5304	5701	5554	5520	3.63	cps
Molybdenum	96-1	14097	14451	14247	14265	1.24	cps
Molybdenum	97-1	3417	3400	3507	3442	1.67	cps
Molybdenum	98-1	8596	8806	9150	8851	3.16	cps
Neodymium	150-1	102700	104222	105215	104046	1.22	cps
Neodymium	150-2	54639	53899	52618	53718	1.90	cps
Nickel	60-2	78836	78722	78755	78771	0.07	cps
Phosphorus	31-2	29539	28784	29054	29126	1.31	cps
Potassium	39-2	4611210	4588993	4577224	4592476	0.38	cps
Rhodium	103-1	7308545	7435220	7315195	7352987	0.97	cps
Rhodium	103-2	4306104	4273646	4313421	4297724	0.49	cps
Scandium	45-1	4374274	4434646	4400018	4402979	0.69	cps
Scandium	45-2	400766	396346	395772	397628	0.69	cps
Selenium	82-1	555	533	537	542	2.18	cps
Selenium	77-2	187	160	203	183	11.92	cps
Selenium	78-2	333	420	433	396	13.73	cps
Silicon	28-1	27665177	27949605	28140545	27918443	0.86	cps
Silver	107-1	2640	2767	2670	2692	2.46	cps
Silver	109-1	2560	2680	2730	2657	3.29	cps
Sodium	23-2	204884	206215	205276	205459	0.33	cps
Strontium	86-1	255768	260306	261696	259257	1.20	cps
Strontium	88-1	2291767	2368412	2374048	2344742	1.96	cps
Sulfur	34-1	137959	140034	140186	139393	0.89	cps
Terbium	159-1	10844112	11017687	10859584	10907127	0.88	cps
Terbium	159-2	6482778	6531054	6443252	6485695	0.68	cps
Thallium	203-1	2577	2584	2604	2588	0.54	cps
Thallium	205-1	6218	6475	6448	6380	2.21	cps
Tin	118-1	18278	18195	18272	18248	0.25	cps
Titanium	47-1	990304	1011644	1010495	1004148	1.20	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-09 Instrumnet Name : P7
Client Sample ID : MYE4K6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:12:45 DataFile Name : 108SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	25638	24953	25371	25321	1.36	cps
Vanadium	51-2	385228	384820	384789	384946	0.06	cps
Yttrium	89-1	13592819	13807729	13741119	13713889	0.80	cps
Yttrium	89-2	3312566	3310056	3309632	3310751	0.05	cps
Zinc	66-2	116960	116500	116154	116538	0.35	cps
Zirconium	90-1	82936	84474	84260	83890	0.99	cps
Zirconium	91-1	18695	19042	19169	18969	1.29	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV004 Instrumnet Name : P7
Client Sample ID : CCV004 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:15:52 DataFile Name : 109CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	10474788	10488053	10585789	10516210	0.58	cps
Antimony	121-1	4941090	5035485	5055832	5010803	1.22	cps
Arsenic	75-2	245620	245874	244898	245464	0.21	cps
Barium	135-1	6742186	6839507	6796080	6792591	0.72	cps
Barium	137-1	11483013	11685650	11735822	11634828	1.15	cps
Beryllium	9-1	664463	678506	684445	675805	1.52	cps
Bismuth	209-1	6423707	6373176	6442958	6413280	0.56	cps
Bismuth	209-2	4393871	4356326	4372122	4374107	0.43	cps
Bromine	81-1	10417	10134	9777	10109	3.17	cps
Bromine	81-2	80	97	107	94	14.27	cps
Cadmium	108-1	88521	88904	89400	88941	0.50	cps
Cadmium	106-1	127483	129851	127557	128297	1.05	cps
Cadmium	111-1	1202555	1220040	1221608	1214734	0.87	cps
Calcium	43-1	6195647	6233226	6237927	6222266	0.37	cps
Calcium	44-1	98777812	100045119	99317929	99380286	0.64	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1953208	1953422	1932776	1946469	0.61	cps
Cobalt	59-2	3140463	3132235	3145433	3139377	0.21	cps
Copper	63-2	21725655	21774058	21876460	21792057	0.35	cps
Dysprosium	156-1	363	310	287	320	12.28	cps
Dysprosium	156-2	447	413	390	417	6.84	cps
Erbium	164-1	370	333	353	352	5.21	cps
Erbium	164-2	230	247	203	227	9.64	cps
Gadolinium	160-1	267	277	273	272	1.87	cps
Gadolinium	160-2	417	353	427	399	9.97	cps
Holmium	165-1	10664872	10782909	10902835	10783539	1.10	cps
Holmium	165-2	6488645	6533786	6503076	6508502	0.35	cps
Indium	115-1	7631398	7718935	7661631	7670655	0.58	cps
Indium	115-2	2393751	2342727	2369283	2368587	1.08	cps
Iron	56-2	416975140	419602247	419636434	418737940	0.36	cps
Iron	57-2	10492673	10509985	10378664	10460441	0.68	cps
Iron	54-2	22734657	23075093	23040423	22950058	0.82	cps
Krypton	83-1	187	210	223	207	8.98	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV004 Instrumnet Name : P7
Client Sample ID : CCV004 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:15:52 DataFile Name : 109CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	20577565	20477695	20386703	20480654	0.47	cps
Lead	207-1	17825192	17972821	18016225	17938079	0.56	cps
Lead	208-1	81577896	81762334	81905693	81748641	0.20	cps
Lithium	6-1	1226770	1253938	1272358	1251022	1.83	cps
Magnesium	24-2	87964949	88397499	88465005	88275818	0.31	cps
Manganese	55-2	14006220	14291261	14218874	14172118	1.05	cps
Molybdenum	94-1	18429012	18766560	18737209	18644260	1.00	cps
Molybdenum	95-1	27125318	27416830	27705802	27415983	1.06	cps
Molybdenum	96-1	29560800	29711235	30155638	29809224	1.04	cps
Molybdenum	97-1	16981480	16956074	17266540	17068031	1.01	cps
Molybdenum	98-1	43496528	43740831	44014968	43750775	0.59	cps
Neodymium	150-1	440	410	453	434	5.11	cps
Neodymium	150-2	167	127	177	157	16.89	cps
Nickel	60-2	750788	748918	745706	748471	0.34	cps
Phosphorus	31-2	107464	106702	107443	107203	0.41	cps
Potassium	39-2	53199584	53352736	53341519	53297946	0.16	cps
Rhodium	103-1	6802758	6806618	6818022	6809133	0.12	cps
Rhodium	103-2	3981961	3982048	3989119	3984376	0.10	cps
Scandium	45-1	4129169	4131397	4157085	4139217	0.37	cps
Scandium	45-2	377748	377303	383262	379438	0.87	cps
Selenium	82-1	73452	74051	74028	73844	0.46	cps
Selenium	77-2	6585	6635	6918	6713	2.68	cps
Selenium	78-2	22213	22788	22644	22548	1.33	cps
Silicon	28-1	42821209	43181041	43561898	43188049	0.86	cps
Silver	107-1	6202334	6228147	6218437	6216306	0.21	cps
Silver	109-1	5933577	5917758	6000918	5950751	0.74	cps
Sodium	23-2	149049311	150438858	148056774	149181648	0.80	cps
Strontium	86-1	39051444	39356396	39519703	39309181	0.60	cps
Strontium	88-1	343238582	346725835	345662515	345208977	0.52	cps
Sulfur	34-1	497670	498777	498394	498280	0.11	cps
Terbium	159-1	10902001	10803535	11041289	10915608	1.09	cps
Terbium	159-2	6563932	6476563	6531992	6524162	0.68	cps
Thallium	203-1	5090953	5149736	5142302	5127664	0.62	cps
Thallium	205-1	12291281	12245285	12403599	12313388	0.66	cps
Tin	118-1	4162568	4185521	4224524	4190871	0.75	cps
Titanium	47-1	6325233	6378278	6374438	6359316	0.47	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV004 Instrumnet Name : P7
Client Sample ID : CCV004 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:15:52 DataFile Name : 109CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	16489880	16753631	16884634	16709382	1.20	cps
Vanadium	51-2	1744865	1762734	1729782	1745793	0.94	cps
Yttrium	89-1	12783898	12907630	12979163	12890231	0.77	cps
Yttrium	89-2	3087387	3105922	3086229	3093180	0.36	cps
Zinc	66-2	3420723	3440662	3419157	3426848	0.35	cps
Zirconium	90-1	8448573	8599448	8755931	8601317	1.79	cps
Zirconium	91-1	1937813	1930015	1940180	1936002	0.27	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB004 Instrumnet Name : P7
Client Sample ID : CCB004 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:18:36 DataFile Name : 110CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1500	1540	1557	1532	1.90	cps
Antimony	121-1	3400	3360	3247	3336	2.38	cps
Arsenic	75-2	23	7	30	20	60.08	cps
Barium	135-1	560	683	670	638	10.61	cps
Barium	137-1	1123	1093	1100	1106	1.43	cps
Beryllium	9-1	128	83	83	98	26.14	cps
Bismuth	209-1	7213564	7202409	7120951	7178974	0.70	cps
Bismuth	209-2	4985098	5007846	5048512	5013819	0.64	cps
Bromine	81-1	9850	10571	10504	10308	3.86	cps
Bromine	81-2	80	67	93	80	16.66	cps
Cadmium	108-1	7	17	30	18	65.83	cps
Cadmium	106-1	2714	2730	2664	2703	1.28	cps
Cadmium	111-1	2002	1996	1927	1975	2.11	cps
Calcium	43-1	517	550	573	547	5.21	cps
Calcium	44-1	15375	15612	15575	15521	0.82	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	657	753	680	697	7.24	cps
Cobalt	59-2	293	250	380	308	21.51	cps
Copper	63-2	1927	1970	1693	1863	7.99	cps
Dysprosium	156-1	30	27	17	24	28.38	cps
Dysprosium	156-2	10	10	23	14	53.28	cps
Erbium	164-1	53	93	80	76	26.96	cps
Erbium	164-2	47	37	63	49	27.55	cps
Gadolinium	160-1	33	53	40	42	24.12	cps
Gadolinium	160-2	207	267	287	253	16.43	cps
Holmium	165-1	11137919	11109249	10902149	11049772	1.16	cps
Holmium	165-2	6735053	6769023	6812582	6772220	0.57	cps
Indium	115-1	8414728	8440899	8410819	8422149	0.19	cps
Indium	115-2	2696845	2648019	2609418	2651427	1.65	cps
Iron	56-2	62265	61127	64707	62700	2.92	cps
Iron	57-2	2370	2294	2327	2330	1.65	cps
Iron	54-2	4344	4107	4494	4315	4.52	cps
Krypton	83-1	217	153	167	179	18.66	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB004 Instrumnet Name : P7
Client Sample ID : CCB004 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:18:36 DataFile Name : 110CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1920	1783	1660	1788	7.28	cps
Lead	207-1	1767	1603	1560	1643	6.63	cps
Lead	208-1	7404	7127	6721	7084	4.85	cps
Lithium	6-1	1333438	1356773	1336608	1342273	0.94	cps
Magnesium	24-2	4047	3730	4001	3926	4.36	cps
Manganese	55-2	1570	1690	1763	1675	5.83	cps
Molybdenum	94-1	1543	1507	1613	1555	3.49	cps
Molybdenum	95-1	1673	1530	1533	1579	5.18	cps
Molybdenum	96-1	1930	1800	1767	1832	4.71	cps
Molybdenum	97-1	1120	890	1003	1004	11.45	cps
Molybdenum	98-1	2650	2614	2407	2557	5.13	cps
Neodymium	150-1	30	23	40	31	26.97	cps
Neodymium	150-2	7	20	10	12	56.76	cps
Nickel	60-2	200	193	220	204	6.79	cps
Phosphorus	31-2	183	163	117	154	22.15	cps
Potassium	39-2	49520	49146	49644	49437	0.52	cps
Rhodium	103-1	7753978	7802525	7661445	7739316	0.93	cps
Rhodium	103-2	4543278	4511328	4521996	4525534	0.36	cps
Scandium	45-1	4430106	4385990	4352300	4389465	0.89	cps
Scandium	45-2	392145	387784	391927	390619	0.63	cps
Selenium	82-1	364	408	465	412	12.38	cps
Selenium	77-2	3	0	0	1	173.21	cps
Selenium	78-2	243	283	217	248	13.54	cps
Silicon	28-1	212933	210030	210560	211174	0.73	cps
Silver	107-1	810	760	657	742	10.54	cps
Silver	109-1	750	650	553	651	15.10	cps
Sodium	23-2	31482	30844	30717	31015	1.32	cps
Strontium	86-1	1730	1663	1547	1647	5.64	cps
Strontium	88-1	12362	12058	11868	12096	2.06	cps
Sulfur	34-1	155424	156613	158024	156687	0.83	cps
Terbium	159-1	11374243	11383932	11358583	11372252	0.11	cps
Terbium	159-2	6855858	6771606	6722147	6783204	1.00	cps
Thallium	203-1	947	883	620	817	21.21	cps
Thallium	205-1	1960	2017	1723	1900	8.19	cps
Tin	118-1	2244	2367	2437	2349	4.17	cps
Titanium	47-1	587	647	620	618	4.87	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB004 Instrumnet Name : P7
Client Sample ID : CCB004 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:18:36 DataFile Name : 110CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	557	610	510	559	8.95	cps
Vanadium	51-2	217	183	220	207	9.81	cps
Yttrium	89-1	13420849	13463121	13405934	13429968	0.22	cps
Yttrium	89-2	3278137	3228285	3242973	3249798	0.79	cps
Zinc	66-2	557	597	617	590	5.18	cps
Zirconium	90-1	1597	1627	1487	1570	4.70	cps
Zirconium	91-1	420	367	337	374	11.27	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-10 Instrumnet Name : P7
Client Sample ID : MYE4K7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:21:51 DataFile Name : 111SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8378270	8299258	8247415	8308314	0.79	cps
Antimony	121-1	11254	12089	11972	11772	3.84	cps
Arsenic	75-2	41354	41514	41224	41364	0.35	cps
Barium	135-1	994499	1004505	1014254	1004419	0.98	cps
Barium	137-1	1829346	1877222	1887555	1864707	1.67	cps
Beryllium	9-1	1795	1878	1822	1832	2.32	cps
Bismuth	209-1	7171295	6983092	7000741	7051709	1.47	cps
Bismuth	209-2	4887115	4972715	4925951	4928594	0.87	cps
Bromine	81-1	16636	17080	17651	17122	2.97	cps
Bromine	81-2	1300	1213	1427	1313	8.17	cps
Cadmium	108-1	407	493	440	447	9.79	cps
Cadmium	106-1	3801	3821	3647	3756	2.53	cps
Cadmium	111-1	6940	7230	6772	6981	3.32	cps
Calcium	43-1	468759	483993	491380	481377	2.40	cps
Calcium	44-1	7135682	7412366	7482027	7343359	2.49	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	356541	353832	355213	355196	0.38	cps
Cobalt	59-2	266616	266847	264526	265996	0.48	cps
Copper	63-2	1333185	1327046	1324472	1328234	0.34	cps
Dysprosium	156-1	71364	71936	72224	71841	0.61	cps
Dysprosium	156-2	53599	53612	53675	53629	0.08	cps
Erbium	164-1	59506	59118	58940	59188	0.49	cps
Erbium	164-2	36808	36016	36147	36324	1.17	cps
Gadolinium	160-1	67344	67334	68540	67739	1.02	cps
Gadolinium	160-2	44671	44474	45072	44739	0.68	cps
Holmium	165-1	11555285	11366542	11370737	11430855	0.94	cps
Holmium	165-2	7042272	6835305	6872345	6916641	1.60	cps
Indium	115-1	8567944	8675337	8625056	8622779	0.62	cps
Indium	115-2	2670686	2677460	2660773	2669639	0.31	cps
Iron	56-2	271355049	269002596	268384049	269580565	0.58	cps
Iron	57-2	6706264	6636647	6609629	6650846	0.75	cps
Iron	54-2	14783947	14668892	14758177	14737005	0.41	cps
Krypton	83-1	247	257	317	273	13.85	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-10 Instrumnet Name : P7
Client Sample ID : MYE4K7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:21:51 DataFile Name : 111SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	533181	530582	538379	534047	0.74	cps
Lead	207-1	465603	462729	463446	463926	0.32	cps
Lead	208-1	2116069	2100016	2112994	2109693	0.40	cps
Lithium	6-1	1366306	1356950	1329900	1351052	1.40	cps
Magnesium	24-2	7078096	7015842	7044119	7046019	0.44	cps
Manganese	55-2	6450752	6401612	6334194	6395519	0.91	cps
Molybdenum	94-1	61907	63349	63430	62895	1.36	cps
Molybdenum	95-1	8939	9136	9226	9101	1.61	cps
Molybdenum	96-1	22921	23229	23322	23157	0.91	cps
Molybdenum	97-1	5344	5298	5531	5391	2.29	cps
Molybdenum	98-1	13566	14244	14337	14049	2.99	cps
Neodymium	150-1	126650	128909	129431	128330	1.15	cps
Neodymium	150-2	67092	66178	65943	66404	0.91	cps
Nickel	60-2	123349	122055	122028	122477	0.62	cps
Phosphorus	31-2	32191	31857	32134	32060	0.56	cps
Potassium	39-2	6134697	6044132	6043889	6074239	0.86	cps
Rhodium	103-1	7853616	7899004	7955991	7902870	0.65	cps
Rhodium	103-2	4599545	4635582	4636915	4624014	0.46	cps
Scandium	45-1	4863601	5008071	5052888	4974853	1.99	cps
Scandium	45-2	446441	445659	443508	445202	0.34	cps
Selenium	82-1	760	819	753	777	4.68	cps
Selenium	77-2	310	273	270	284	7.80	cps
Selenium	78-2	497	500	520	506	2.50	cps
Silicon	28-1	32769002	33582615	33973707	33441775	1.84	cps
Silver	107-1	9403	9466	9133	9334	1.90	cps
Silver	109-1	9030	8446	8453	8643	3.88	cps
Sodium	23-2	205250	205381	204080	204904	0.35	cps
Strontium	86-1	270296	278002	279691	275996	1.81	cps
Strontium	88-1	2461798	2496405	2546361	2501521	1.70	cps
Sulfur	34-1	170466	168073	161459	166666	2.80	cps
Terbium	159-1	11642172	11753388	11654659	11683406	0.52	cps
Terbium	159-2	6908058	6940956	6897355	6915456	0.33	cps
Thallium	203-1	4354	4271	4121	4248	2.78	cps
Thallium	205-1	10237	9714	9180	9710	5.44	cps
Tin	118-1	23690	24851	25222	24588	3.25	cps
Titanium	47-1	1193028	1239458	1257209	1229898	2.69	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-10 Instrumnet Name : P7
Client Sample ID : MYE4K7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:21:51 DataFile Name : 111SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	31531	32045	32169	31915	1.06	cps
Vanadium	51-2	526802	524439	525921	525720	0.23	cps
Yttrium	89-1	14842286	15053397	15069519	14988401	0.85	cps
Yttrium	89-2	3630146	3506430	3570105	3568894	1.73	cps
Zinc	66-2	205267	203446	201426	203380	0.94	cps
Zirconium	90-1	152595	157703	160044	156781	2.43	cps
Zirconium	91-1	34574	35894	35740	35403	2.04	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-11 Instrumnet Name : P7
Client Sample ID : MYE4K8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:24:58 DataFile Name : 112SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8038705	8087335	8049777	8058606	0.32	cps
Antimony	121-1	10027	10531	10347	10302	2.47	cps
Arsenic	75-2	35546	35586	34587	35240	1.60	cps
Barium	135-1	1059498	1074889	1068155	1067514	0.72	cps
Barium	137-1	1985934	2000334	1979886	1988718	0.53	cps
Beryllium	9-1	1922	1972	1817	1904	4.17	cps
Bismuth	209-1	7031126	5961618	7021074	6671273	9.21	cps
Bismuth	209-2	4806677	4863411	4836147	4835411	0.59	cps
Bromine	81-1	20367	20715	21486	20856	2.74	cps
Bromine	81-2	1833	1823	1727	1795	3.28	cps
Cadmium	108-1	357	370	363	363	1.84	cps
Cadmium	106-1	3497	3084	3374	3318	6.40	cps
Cadmium	111-1	6006	5607	5921	5844	3.60	cps
Calcium	43-1	485903	496326	495463	492564	1.17	cps
Calcium	44-1	7398818	7611177	7418972	7476322	1.57	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	339983	339048	338846	339292	0.18	cps
Cobalt	59-2	267014	264996	263229	265079	0.71	cps
Copper	63-2	1097512	1098712	1097519	1097914	0.06	cps
Dysprosium	156-1	69045	69568	71233	69949	1.63	cps
Dysprosium	156-2	51922	51705	52176	51934	0.45	cps
Erbium	164-1	57150	58743	58351	58081	1.43	cps
Erbium	164-2	35101	34867	34436	34801	0.97	cps
Gadolinium	160-1	65804	67247	67465	66838	1.35	cps
Gadolinium	160-2	44056	43554	42996	43535	1.22	cps
Holmium	165-1	11290786	9538000	11243052	10690613	9.34	cps
Holmium	165-2	6840052	6758100	6712178	6770110	0.96	cps
Indium	115-1	8601183	7215940	8510941	8109355	9.56	cps
Indium	115-2	2653175	2593079	2579071	2608442	1.51	cps
Iron	56-2	246054396	246683043	247508583	246748674	0.30	cps
Iron	57-2	6192973	6123495	6100571	6139013	0.78	cps
Iron	54-2	13438966	13535478	13562161	13512202	0.48	cps
Krypton	83-1	257	270	273	267	3.31	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-11 Instrumnet Name : P7
Client Sample ID : MYE4K8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:24:58 DataFile Name : 112SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	427245	437064	439607	434639	1.50	cps
Lead	207-1	374830	380364	379430	378208	0.78	cps
Lead	208-1	1693780	1734251	1727433	1718488	1.26	cps
Lithium	6-1	1328276	1129992	1294868	1251045	8.49	cps
Magnesium	24-2	6994852	7034151	6968474	6999159	0.47	cps
Manganese	55-2	6628149	6678759	6665000	6657302	0.39	cps
Molybdenum	94-1	49584	50487	50192	50088	0.92	cps
Molybdenum	95-1	10557	11091	10794	10814	2.47	cps
Molybdenum	96-1	22147	22841	22991	22660	1.99	cps
Molybdenum	97-1	6525	6848	6752	6708	2.48	cps
Molybdenum	98-1	16720	17220	17244	17061	1.73	cps
Neodymium	150-1	121582	125818	124779	124060	1.78	cps
Neodymium	150-2	65133	63793	63827	64251	1.19	cps
Nickel	60-2	115836	115491	115568	115632	0.16	cps
Phosphorus	31-2	36587	36153	36002	36247	0.84	cps
Potassium	39-2	5683951	5700768	5694047	5692922	0.15	cps
Rhodium	103-1	7957324	6619290	7793787	7456800	9.79	cps
Rhodium	103-2	4561840	4495924	4538874	4532213	0.74	cps
Scandium	45-1	4963324	4221816	4848564	4677901	8.53	cps
Scandium	45-2	434546	437017	432622	434728	0.51	cps
Selenium	82-1	622	628	670	640	4.07	cps
Selenium	77-2	273	237	223	244	10.59	cps
Selenium	78-2	470	403	383	419	10.83	cps
Silicon	28-1	35720906	36857768	36292597	36290424	1.57	cps
Silver	107-1	8262	8476	8523	8420	1.65	cps
Silver	109-1	7655	7919	7889	7821	1.84	cps
Sodium	23-2	270946	271792	268945	270561	0.54	cps
Strontium	86-1	346129	353213	347875	349072	1.06	cps
Strontium	88-1	3127699	3188592	3164084	3160125	0.97	cps
Sulfur	34-1	160873	157054	156115	158014	1.59	cps
Terbium	159-1	11503954	9645914	11497852	10882573	9.84	cps
Terbium	159-2	6795058	6783510	6800485	6793018	0.13	cps
Thallium	203-1	3544	3771	3574	3629	3.39	cps
Thallium	205-1	8680	8586	8249	8505	2.66	cps
Tin	118-1	23523	23960	23249	23577	1.52	cps
Titanium	47-1	1145165	1177484	1170723	1164457	1.46	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-11 Instrumnet Name : P7
Client Sample ID : MYE4K8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:24:58 DataFile Name : 112SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	30264	30913	31036	30738	1.35	cps
Vanadium	51-2	485213	480516	480973	482234	0.54	cps
Yttrium	89-1	15261443	12761890	14660763	14228032	9.17	cps
Yttrium	89-2	3530964	3522862	3477771	3510532	0.82	cps
Zinc	66-2	200906	200606	199568	200360	0.35	cps
Zirconium	90-1	117210	122193	119131	119511	2.10	cps
Zirconium	91-1	26467	26861	26487	26605	0.83	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-12 Instrumnet Name : P7
Client Sample ID : MYE4K9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:28:04 DataFile Name : 113SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8676368	8579857	8522702	8592976	0.90	cps
Antimony	121-1	10431	10754	11028	10737	2.78	cps
Arsenic	75-2	38430	37584	37561	37858	1.31	cps
Barium	135-1	1100834	1108507	1100265	1103202	0.42	cps
Barium	137-1	2087224	2056248	2087497	2076989	0.86	cps
Beryllium	9-1	1478	1450	1661	1530	7.50	cps
Bismuth	209-1	6916285	7087651	7047172	7017036	1.28	cps
Bismuth	209-2	4825886	4815231	4714792	4785303	1.28	cps
Bromine	81-1	18575	19620	19323	19173	2.81	cps
Bromine	81-2	1530	1490	1380	1467	5.30	cps
Cadmium	108-1	350	327	367	348	5.78	cps
Cadmium	106-1	3404	3610	3124	3379	7.23	cps
Cadmium	111-1	4678	4979	4567	4741	4.50	cps
Calcium	43-1	637651	648198	652881	646243	1.21	cps
Calcium	44-1	9818181	9969363	9883223	9890256	0.77	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	456131	451097	447829	451686	0.93	cps
Cobalt	59-2	320428	317891	318474	318931	0.42	cps
Copper	63-2	974314	966308	959008	966543	0.79	cps
Dysprosium	156-1	61942	62384	62542	62290	0.50	cps
Dysprosium	156-2	45146	45216	45343	45235	0.22	cps
Erbium	164-1	53857	53569	54328	53918	0.71	cps
Erbium	164-2	32879	31642	32174	32232	1.92	cps
Gadolinium	160-1	59747	59030	61156	59978	1.80	cps
Gadolinium	160-2	38636	38881	38523	38680	0.47	cps
Holmium	165-1	11304874	11143009	11081741	11176541	1.03	cps
Holmium	165-2	6798800	6848478	6586724	6744667	2.06	cps
Indium	115-1	8538023	8477362	8438980	8484788	0.59	cps
Indium	115-2	2591663	2606324	2566977	2588322	0.77	cps
Iron	56-2	288328516	285448422	284575849	286117596	0.69	cps
Iron	57-2	7112998	7128347	7045267	7095537	0.62	cps
Iron	54-2	15723335	15812422	15662361	15732706	0.48	cps
Krypton	83-1	263	330	280	291	11.92	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-12 Instrumnet Name : P7
Client Sample ID : MYE4K9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:28:04 DataFile Name : 113SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	315059	321248	322692	319666	1.27	cps
Lead	207-1	274099	275095	280558	276584	1.26	cps
Lead	208-1	1244522	1260157	1262150	1255610	0.77	cps
Lithium	6-1	1312627	1330581	1289857	1311022	1.56	cps
Magnesium	24-2	9276903	9236601	9151509	9221671	0.69	cps
Manganese	55-2	8087893	8093620	8074691	8085401	0.12	cps
Molybdenum	94-1	48256	49741	48761	48919	1.54	cps
Molybdenum	95-1	8473	8779	8643	8632	1.78	cps
Molybdenum	96-1	20237	20678	20751	20556	1.35	cps
Molybdenum	97-1	5338	5271	5038	5215	3.02	cps
Molybdenum	98-1	13199	13126	12899	13075	1.20	cps
Neodymium	150-1	107163	107630	109092	107962	0.93	cps
Neodymium	150-2	54445	54408	55321	54725	0.95	cps
Nickel	60-2	153823	154103	154214	154047	0.13	cps
Phosphorus	31-2	31840	32137	31412	31796	1.15	cps
Potassium	39-2	4562850	4506243	4464929	4511341	1.09	cps
Rhodium	103-1	7911725	7812069	7821508	7848434	0.70	cps
Rhodium	103-2	4491169	4542313	4466764	4500082	0.86	cps
Scandium	45-1	4927673	4846504	4895496	4889891	0.84	cps
Scandium	45-2	427305	431393	428813	429170	0.48	cps
Selenium	82-1	637	521	619	592	10.55	cps
Selenium	77-2	187	200	173	187	7.14	cps
Selenium	78-2	427	430	417	424	1.63	cps
Silicon	28-1	29077120	29301239	29422977	29267112	0.60	cps
Silver	107-1	5054	4944	5188	5062	2.41	cps
Silver	109-1	4701	4754	5088	4848	4.32	cps
Sodium	23-2	278599	277974	275756	277443	0.54	cps
Strontium	86-1	393527	400703	397441	397224	0.90	cps
Strontium	88-1	3580708	3611423	3575021	3589050	0.55	cps
Sulfur	34-1	153798	153771	154286	153952	0.19	cps
Terbium	159-1	11452712	11301557	11290009	11348093	0.80	cps
Terbium	159-2	6749449	6726090	6723330	6732956	0.21	cps
Thallium	203-1	3010	2850	2880	2914	2.92	cps
Thallium	205-1	7195	7239	6838	7091	3.10	cps
Tin	118-1	22851	22705	23386	22981	1.56	cps
Titanium	47-1	1129576	1151959	1153244	1144926	1.16	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-12 Instrumnet Name : P7
Client Sample ID : MYE4K9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:28:04 DataFile Name : 113SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	25064	24830	25668	25187	1.72	cps
Vanadium	51-2	628027	621758	616528	622105	0.93	cps
Yttrium	89-1	14883986	14832489	14717135	14811203	0.58	cps
Yttrium	89-2	3498882	3490291	3442412	3477195	0.88	cps
Zinc	66-2	149079	147493	147140	147904	0.70	cps
Zirconium	90-1	119225	119696	119653	119525	0.22	cps
Zirconium	91-1	26544	27382	27366	27097	1.77	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-13 Instrumnet Name : P7
Client Sample ID : MYE4L0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:31:10 DataFile Name : 114SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7441447	7420339	7363579	7408455	0.54	cps
Antimony	121-1	6895	7055	7219	7056	2.29	cps
Arsenic	75-2	29005	28477	28320	28601	1.25	cps
Barium	135-1	1107817	1120603	1127689	1118703	0.90	cps
Barium	137-1	2063742	2096375	2064103	2074740	0.90	cps
Beryllium	9-1	1967	1817	1995	1926	4.97	cps
Bismuth	209-1	6879011	6953039	6834016	6888689	0.87	cps
Bismuth	209-2	4673439	4702763	4731920	4702707	0.62	cps
Bromine	81-1	15598	16319	16526	16148	3.02	cps
Bromine	81-2	1187	1133	1150	1157	2.36	cps
Cadmium	108-1	273	273	267	271	1.42	cps
Cadmium	106-1	3164	3344	3384	3297	3.56	cps
Cadmium	111-1	3370	3373	3515	3419	2.43	cps
Calcium	43-1	815307	828655	836280	826747	1.28	cps
Calcium	44-1	12481270	12635694	12620065	12579010	0.68	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	301960	303874	301273	302369	0.45	cps
Cobalt	59-2	241647	239208	240422	240426	0.51	cps
Copper	63-2	470321	469430	468208	469320	0.23	cps
Dysprosium	156-1	75495	76513	77981	76663	1.63	cps
Dysprosium	156-2	56891	57025	57541	57152	0.60	cps
Erbium	164-1	62077	62529	62596	62401	0.45	cps
Erbium	164-2	37938	38088	37674	37900	0.55	cps
Gadolinium	160-1	71880	72252	72318	72150	0.33	cps
Gadolinium	160-2	46480	46808	46985	46758	0.55	cps
Holmium	165-1	11046132	10976361	10824848	10949114	1.03	cps
Holmium	165-2	6616405	6609409	6631617	6619144	0.17	cps
Indium	115-1	8343029	8301215	8345739	8329994	0.30	cps
Indium	115-2	2565916	2551740	2593854	2570503	0.83	cps
Iron	56-2	225712330	225966330	226422457	226033706	0.16	cps
Iron	57-2	5617611	5640842	5601127	5619860	0.36	cps
Iron	54-2	12479619	12463874	12478068	12473854	0.07	cps
Krypton	83-1	307	300	270	292	6.68	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-13 Instrumnet Name : P7
Client Sample ID : MYE4L0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:31:10 DataFile Name : 114SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	217751	219887	221378	219672	0.83	cps
Lead	207-1	181923	181161	184136	182407	0.85	cps
Lead	208-1	838796	842702	853045	844848	0.87	cps
Lithium	6-1	1320831	1328327	1299632	1316263	1.13	cps
Magnesium	24-2	6539657	6530899	6539891	6536816	0.08	cps
Manganese	55-2	5808348	5837924	5819398	5821890	0.26	cps
Molybdenum	94-1	73731	74994	75731	74819	1.35	cps
Molybdenum	95-1	5778	5868	5778	5808	0.90	cps
Molybdenum	96-1	21653	21349	22297	21766	2.22	cps
Molybdenum	97-1	3490	3540	3630	3554	2.00	cps
Molybdenum	98-1	8769	8843	8916	8843	0.83	cps
Neodymium	150-1	134813	136365	138725	136634	1.44	cps
Neodymium	150-2	71135	70719	70435	70763	0.50	cps
Nickel	60-2	119951	118934	118635	119173	0.58	cps
Phosphorus	31-2	26276	26500	26997	26591	1.39	cps
Potassium	39-2	3711284	3738062	3722891	3724079	0.36	cps
Rhodium	103-1	7616548	7565667	7461046	7547754	1.05	cps
Rhodium	103-2	4403146	4456483	4484537	4448055	0.93	cps
Scandium	45-1	4652087	4670095	4641815	4654666	0.31	cps
Scandium	45-2	414177	414043	413261	413827	0.12	cps
Selenium	82-1	549	507	595	551	8.00	cps
Selenium	77-2	280	240	193	238	18.24	cps
Selenium	78-2	393	417	453	421	7.18	cps
Silicon	28-1	33228224	33650662	33797687	33558858	0.88	cps
Silver	107-1	2037	1930	2284	2084	8.70	cps
Silver	109-1	1893	1957	1963	1938	1.99	cps
Sodium	23-2	215057	215702	214151	214970	0.36	cps
Strontium	86-1	334783	333041	339740	335855	1.03	cps
Strontium	88-1	3001802	3080225	3039898	3040642	1.29	cps
Sulfur	34-1	140707	142531	142204	141814	0.69	cps
Terbium	159-1	11156730	11232270	11234298	11207766	0.39	cps
Terbium	159-2	6655526	6631004	6658380	6648303	0.23	cps
Thallium	203-1	2934	3044	3160	3046	3.72	cps
Thallium	205-1	7235	7185	7399	7273	1.53	cps
Tin	118-1	19917	20138	21202	20419	3.37	cps
Titanium	47-1	1340363	1358174	1334836	1344458	0.91	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-13 Instrumnet Name : P7
Client Sample ID : MYE4L0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:31:10 DataFile Name : 114SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	29499	29683	29813	29665	0.53	cps
Vanadium	51-2	455303	448995	445435	449911	1.11	cps
Yttrium	89-1	14402507	14513915	14403632	14440018	0.44	cps
Yttrium	89-2	3417812	3442232	3481406	3447150	0.93	cps
Zinc	66-2	114407	114162	115176	114582	0.46	cps
Zirconium	90-1	192757	194861	197610	195076	1.25	cps
Zirconium	91-1	43083	44246	44514	43948	1.73	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-14 Instrumnet Name : P7
Client Sample ID : MYE4L1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:34:15 DataFile Name : 115SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8254700	8259303	8257941	8257315	0.03	cps
Antimony	121-1	5985	5971	6175	6044	1.88	cps
Arsenic	75-2	33331	33077	32516	32974	1.26	cps
Barium	135-1	1005797	1018394	1027851	1017347	1.09	cps
Barium	137-1	1882014	1918296	1935012	1911774	1.42	cps
Beryllium	9-1	1939	2033	1845	1939	4.87	cps
Bismuth	209-1	6788853	6867793	5964761	6540469	7.65	cps
Bismuth	209-2	4733252	4731525	4708923	4724567	0.29	cps
Bromine	81-1	17594	18251	18618	18155	2.86	cps
Bromine	81-2	1430	1563	1503	1499	4.46	cps
Cadmium	108-1	313	240	313	289	14.66	cps
Cadmium	106-1	3044	3074	2607	2908	8.98	cps
Cadmium	111-1	4445	4681	4199	4441	5.42	cps
Calcium	43-1	421818	426605	425489	424637	0.59	cps
Calcium	44-1	6404116	6458728	6430767	6431204	0.42	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	329017	328211	329993	329074	0.27	cps
Cobalt	59-2	257639	257948	254878	256822	0.66	cps
Copper	63-2	864250	857226	858612	860029	0.43	cps
Dysprosium	156-1	76926	78494	77676	77699	1.01	cps
Dysprosium	156-2	57336	57574	56316	57075	1.17	cps
Erbium	164-1	66099	66627	66809	66512	0.55	cps
Erbium	164-2	39750	40783	39740	40091	1.49	cps
Gadolinium	160-1	74564	74785	76071	75140	1.08	cps
Gadolinium	160-2	49099	49203	49343	49215	0.25	cps
Holmium	165-1	10930743	10802716	9575555	10436338	7.17	cps
Holmium	165-2	6599687	6807473	6631388	6679516	1.68	cps
Indium	115-1	8139277	8226922	7302237	7889479	6.47	cps
Indium	115-2	2511057	2517643	2550539	2526413	0.84	cps
Iron	56-2	277076256	276581756	276738683	276798898	0.09	cps
Iron	57-2	6840390	6852519	6881010	6857973	0.30	cps
Iron	54-2	15135794	14960921	14986326	15027680	0.63	cps
Krypton	83-1	297	267	270	278	5.92	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-14 Instrumnet Name : P7
Client Sample ID : MYE4L1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:34:15 DataFile Name : 115SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	345864	354670	356543	352359	1.62	cps
Lead	207-1	298323	304569	305934	302942	1.34	cps
Lead	208-1	1359214	1386990	1389487	1378564	1.22	cps
Lithium	6-1	1283635	1323931	1171213	1259593	6.28	cps
Magnesium	24-2	6808935	6636591	6616343	6687289	1.58	cps
Manganese	55-2	5174472	5184393	5130811	5163225	0.55	cps
Molybdenum	94-1	46490	46958	46547	46665	0.55	cps
Molybdenum	95-1	7289	7762	7789	7613	3.70	cps
Molybdenum	96-1	19500	19346	19597	19481	0.65	cps
Molybdenum	97-1	4914	4607	4737	4753	3.24	cps
Molybdenum	98-1	11755	11835	11585	11725	1.09	cps
Neodymium	150-1	135255	136372	138408	136678	1.17	cps
Neodymium	150-2	70033	69701	69313	69682	0.52	cps
Nickel	60-2	103662	101987	102037	102562	0.93	cps
Phosphorus	31-2	28981	28804	28296	28694	1.24	cps
Potassium	39-2	5395898	5379054	5377853	5384268	0.19	cps
Rhodium	103-1	7558281	7552526	6565321	7225376	7.91	cps
Rhodium	103-2	4399198	4401900	4378530	4393209	0.29	cps
Scandium	45-1	4583951	4608041	4036999	4409664	7.32	cps
Scandium	45-2	409813	408602	406609	408341	0.40	cps
Selenium	82-1	570	657	568	598	8.46	cps
Selenium	77-2	193	267	247	236	16.09	cps
Selenium	78-2	423	463	443	443	4.51	cps
Silicon	28-1	32152439	32531640	32549848	32411309	0.69	cps
Silver	107-1	4961	5011	5021	4998	0.64	cps
Silver	109-1	4851	5028	4904	4928	1.84	cps
Sodium	23-2	237901	239503	237854	238419	0.39	cps
Strontium	86-1	257373	261195	261991	260186	0.95	cps
Strontium	88-1	2349859	2354573	2372606	2359012	0.51	cps
Sulfur	34-1	127510	123907	120372	123930	2.88	cps
Terbium	159-1	11227763	11218488	10058182	10834811	6.21	cps
Terbium	159-2	6721715	6725752	6632468	6693312	0.79	cps
Thallium	203-1	3210	3120	3400	3244	4.41	cps
Thallium	205-1	8062	7912	7659	7878	2.59	cps
Tin	118-1	21356	21807	21706	21623	1.10	cps
Titanium	47-1	1278512	1290675	1304072	1291086	0.99	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-14 Instrumnet Name : P7
Client Sample ID : MYE4L1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:34:15 DataFile Name : 115SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	35614	35678	36446	35913	1.29	cps
Vanadium	51-2	521986	525763	520208	522652	0.54	cps
Yttrium	89-1	14314199	14182561	12546736	13681165	7.20	cps
Yttrium	89-2	3401445	3366754	3412693	3393631	0.71	cps
Zinc	66-2	154203	153780	151775	153253	0.85	cps
Zirconium	90-1	112184	114820	116260	114422	1.81	cps
Zirconium	91-1	25412	26140	26334	25962	1.87	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-15 Instrumnet Name : P7
Client Sample ID : MYE4L2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:37:22 DataFile Name : 116SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	9370546	9180792	9206367	9252568	1.11	cps
Antimony	121-1	5618	6141	5848	5869	4.47	cps
Arsenic	75-2	37588	36993	37564	37381	0.90	cps
Barium	135-1	1137686	1146294	1151955	1145312	0.63	cps
Barium	137-1	2141640	2133879	2150269	2141929	0.38	cps
Beryllium	9-1	2195	1917	2145	2085	7.10	cps
Bismuth	209-1	6900572	6720944	6885816	6835778	1.46	cps
Bismuth	209-2	4738830	4778997	4770901	4762909	0.45	cps
Bromine	81-1	18622	19016	19950	19196	3.55	cps
Bromine	81-2	1800	1647	1700	1716	4.54	cps
Cadmium	108-1	310	290	390	330	16.03	cps
Cadmium	106-1	3117	3240	3240	3199	2.23	cps
Cadmium	111-1	5029	4842	4989	4953	1.98	cps
Calcium	43-1	467152	472008	473591	470917	0.71	cps
Calcium	44-1	7103924	7226091	7234304	7188106	1.02	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	368099	370515	369537	369384	0.33	cps
Cobalt	59-2	317067	314215	313633	314972	0.58	cps
Copper	63-2	986309	988148	978284	984247	0.53	cps
Dysprosium	156-1	86746	87427	87030	87068	0.39	cps
Dysprosium	156-2	64313	65140	63523	64325	1.26	cps
Erbium	164-1	74346	74256	75502	74701	0.93	cps
Erbium	164-2	45835	46226	45400	45820	0.90	cps
Gadolinium	160-1	83513	84402	85039	84318	0.91	cps
Gadolinium	160-2	55001	55717	55178	55299	0.67	cps
Holmium	165-1	10838116	10667960	10946547	10817541	1.30	cps
Holmium	165-2	6719879	6783035	6628720	6710545	1.16	cps
Indium	115-1	8182771	8054873	8209619	8149088	1.01	cps
Indium	115-2	2544609	2484338	2548028	2525658	1.42	cps
Iron	56-2	313113142	313825662	314760835	313899880	0.26	cps
Iron	57-2	7713447	7770949	7774277	7752891	0.44	cps
Iron	54-2	17185259	17123539	16915045	17074614	0.83	cps
Krypton	83-1	257	270	270	266	2.90	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-15 Instrumnet Name : P7
Client Sample ID : MYE4L2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:37:22 DataFile Name : 116SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	398947	398737	403006	400230	0.60	cps
Lead	207-1	336756	342615	344216	341196	1.15	cps
Lead	208-1	1547134	1569838	1571857	1562943	0.88	cps
Lithium	6-1	1290388	1284912	1302298	1292532	0.69	cps
Magnesium	24-2	7726950	7697072	7779874	7734632	0.54	cps
Manganese	55-2	6097969	6021915	6013944	6044609	0.77	cps
Molybdenum	94-1	38574	38554	38701	38610	0.21	cps
Molybdenum	95-1	7835	7702	7902	7813	1.30	cps
Molybdenum	96-1	18762	18842	19346	18984	1.67	cps
Molybdenum	97-1	4794	4888	4904	4862	1.22	cps
Molybdenum	98-1	11525	11735	12212	11824	2.98	cps
Neodymium	150-1	148435	153286	151786	151169	1.64	cps
Neodymium	150-2	78236	77502	76999	77579	0.80	cps
Nickel	60-2	118020	115709	118360	117363	1.23	cps
Phosphorus	31-2	32675	32361	32682	32572	0.56	cps
Potassium	39-2	6239906	6221996	6249948	6237283	0.23	cps
Rhodium	103-1	7387919	7286351	7440343	7371537	1.06	cps
Rhodium	103-2	4385548	4363775	4344616	4364646	0.47	cps
Scandium	45-1	4547893	4468196	4587763	4534617	1.34	cps
Scandium	45-2	412068	410076	409029	410391	0.38	cps
Selenium	82-1	698	621	599	639	8.08	cps
Selenium	77-2	307	287	270	288	6.38	cps
Selenium	78-2	437	390	433	420	6.20	cps
Silicon	28-1	29948232	30109709	30409389	30155776	0.78	cps
Silver	107-1	8453	8963	9046	8821	3.64	cps
Silver	109-1	8326	8289	8076	8230	1.64	cps
Sodium	23-2	261170	259940	259800	260303	0.29	cps
Strontium	86-1	287739	291137	291256	290044	0.69	cps
Strontium	88-1	2604201	2684805	2642683	2643897	1.52	cps
Sulfur	34-1	122392	120067	122267	121575	1.08	cps
Terbium	159-1	11173554	11148393	11311753	11211233	0.78	cps
Terbium	159-2	6763832	6655846	6507000	6642226	1.94	cps
Thallium	203-1	3654	3447	3717	3606	3.92	cps
Thallium	205-1	8376	8429	8616	8474	1.49	cps
Tin	118-1	22120	22658	22715	22498	1.46	cps
Titanium	47-1	1276156	1278914	1288394	1281155	0.50	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-15 Instrumnet Name : P7
Client Sample ID : MYE4L2 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:37:22 DataFile Name : 116SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	40297	40210	40705	40404	0.65	cps
Vanadium	51-2	583126	579524	582089	581580	0.32	cps
Yttrium	89-1	14253719	14067631	14217936	14179762	0.70	cps
Yttrium	89-2	3408047	3421969	3439934	3423317	0.47	cps
Zinc	66-2	173778	173803	174008	173863	0.07	cps
Zirconium	90-1	91895	92945	93035	92625	0.68	cps
Zirconium	91-1	20631	20428	21279	20779	2.14	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-16 Instrumnet Name : P7
Client Sample ID : MYE4P6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:40:26 DataFile Name : 117SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7236278	7103993	7129213	7156495	0.98	cps
Antimony	121-1	13169	13763	13763	13565	2.53	cps
Arsenic	75-2	39566	40321	39422	39770	1.21	cps
Barium	135-1	1085529	1102526	1105493	1097849	0.98	cps
Barium	137-1	2032021	2101451	2105893	2079788	1.99	cps
Beryllium	9-1	1817	1783	1839	1813	1.54	cps
Bismuth	209-1	6853641	6878158	6866316	6866038	0.18	cps
Bismuth	209-2	4769756	4773086	4677101	4739981	1.15	cps
Bromine	81-1	17404	18385	18178	17989	2.88	cps
Bromine	81-2	1433	1490	1453	1459	1.97	cps
Cadmium	108-1	233	217	257	236	8.53	cps
Cadmium	106-1	3134	3137	2990	3087	2.71	cps
Cadmium	111-1	3640	3681	3297	3539	5.96	cps
Calcium	43-1	196004	197673	198163	197280	0.57	cps
Calcium	44-1	2995210	3016255	3041961	3017809	0.78	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	264234	262325	261101	262553	0.60	cps
Cobalt	59-2	238811	238182	241365	239453	0.70	cps
Copper	63-2	555167	551650	549366	552061	0.53	cps
Dysprosium	156-1	68821	70194	69504	69506	0.99	cps
Dysprosium	156-2	52565	51959	53066	52530	1.06	cps
Erbium	164-1	50879	51645	52475	51666	1.54	cps
Erbium	164-2	31566	31071	31455	31364	0.83	cps
Gadolinium	160-1	63798	65167	65546	64837	1.42	cps
Gadolinium	160-2	42361	41287	41337	41662	1.45	cps
Holmium	165-1	10753629	10872639	10929562	10851943	0.83	cps
Holmium	165-2	6571741	6518993	6594158	6561631	0.59	cps
Indium	115-1	8044463	8125702	8199213	8123126	0.95	cps
Indium	115-2	2487950	2525031	2499838	2504273	0.76	cps
Iron	56-2	259229316	258108803	255933983	257757367	0.65	cps
Iron	57-2	6371308	6417110	6382118	6390179	0.37	cps
Iron	54-2	14232826	14094646	13990011	14105828	0.86	cps
Krypton	83-1	243	263	223	243	8.22	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-16 Instrumnet Name : P7
Client Sample ID : MYE4P6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:40:26 DataFile Name : 117SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	234019	237989	237873	236627	0.95	cps
Lead	207-1	200585	203611	204749	202981	1.06	cps
Lead	208-1	920077	931649	938579	930102	1.01	cps
Lithium	6-1	1277851	1311881	1295193	1294975	1.31	cps
Magnesium	24-2	5513153	5458719	5458888	5476920	0.57	cps
Manganese	55-2	5623970	5641924	5582368	5616087	0.54	cps
Molybdenum	94-1	63293	62238	63142	62891	0.91	cps
Molybdenum	95-1	4571	4491	4471	4511	1.17	cps
Molybdenum	96-1	19366	19089	18662	19039	1.86	cps
Molybdenum	97-1	2710	2607	2707	2675	2.20	cps
Molybdenum	98-1	6668	6575	6812	6685	1.78	cps
Neodymium	150-1	135571	136470	135595	135878	0.38	cps
Neodymium	150-2	68978	69735	69460	69391	0.55	cps
Nickel	60-2	112908	110534	111188	111543	1.10	cps
Phosphorus	31-2	18171	18361	18141	18224	0.66	cps
Potassium	39-2	3219547	3195064	3162495	3192369	0.90	cps
Rhodium	103-1	7457450	7420816	7382199	7420155	0.51	cps
Rhodium	103-2	4360038	4380179	4343773	4361330	0.42	cps
Scandium	45-1	4439930	4464384	4414564	4439626	0.56	cps
Scandium	45-2	396922	397051	394802	396258	0.32	cps
Selenium	82-1	479	456	483	473	3.01	cps
Selenium	77-2	273	290	227	263	12.47	cps
Selenium	78-2	353	393	433	393	10.17	cps
Silicon	28-1	29192175	29192026	29074295	29152832	0.23	cps
Silver	107-1	2600	2484	2560	2548	2.33	cps
Silver	109-1	2310	2134	2220	2221	3.98	cps
Sodium	23-2	85067	85388	83645	84700	1.09	cps
Strontium	86-1	208029	207883	211639	209183	1.02	cps
Strontium	88-1	1879856	1881537	1893268	1884887	0.39	cps
Sulfur	34-1	112054	111382	110533	111323	0.68	cps
Terbium	159-1	11040407	11157049	11044087	11080514	0.60	cps
Terbium	159-2	6688912	6589488	6484900	6587766	1.55	cps
Thallium	203-1	3144	3204	3074	3140	2.07	cps
Thallium	205-1	7399	7776	7395	7523	2.91	cps
Tin	118-1	17688	18258	18145	18030	1.68	cps
Titanium	47-1	973407	981068	982163	978879	0.49	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-16 Instrumnet Name : P7
Client Sample ID : MYE4P6 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:40:26 DataFile Name : 117SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	31384	32239	32556	32060	1.89	cps
Vanadium	51-2	480889	479891	476440	479073	0.49	cps
Yttrium	89-1	13739684	13923929	13644464	13769359	1.03	cps
Yttrium	89-2	3316218	3322853	3354316	3331129	0.61	cps
Zinc	66-2	110747	111090	110196	110677	0.41	cps
Zirconium	90-1	162157	165133	164914	164068	1.01	cps
Zirconium	91-1	36669	36759	37301	36910	0.93	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-17 Instrumnet Name : P7
Client Sample ID : MYE4P7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:43:34 DataFile Name : 118SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	13079421	12974781	12969592	13007931	0.48	cps
Antimony	121-1	14140	14040	14170	14117	0.48	cps
Arsenic	75-2	97581	97843	98114	97846	0.27	cps
Barium	135-1	1471328	1612736	1602359	1562141	5.05	cps
Barium	137-1	2757909	2751903	2808136	2772649	1.11	cps
Beryllium	9-1	1595	1695	1717	1669	3.90	cps
Bismuth	209-1	6460461	6458442	6515321	6478075	0.50	cps
Bismuth	209-2	4434975	4384235	4444527	4421246	0.73	cps
Bromine	81-1	11134	11331	10921	11129	1.84	cps
Bromine	81-2	433	423	363	407	9.31	cps
Cadmium	108-1	270	217	233	240	11.37	cps
Cadmium	106-1	3407	3154	3157	3239	4.49	cps
Cadmium	111-1	3286	3013	3080	3126	4.56	cps
Calcium	43-1	4148753	4096720	4162963	4136145	0.84	cps
Calcium	44-1	67252657	67051959	67600137	67301585	0.41	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3699299	3635334	3647609	3660748	0.93	cps
Cobalt	59-2	873533	868025	864818	868792	0.51	cps
Copper	63-2	1049986	1047489	1044517	1047330	0.26	cps
Dysprosium	156-1	80636	82031	81763	81477	0.91	cps
Dysprosium	156-2	54251	54753	54452	54485	0.46	cps
Erbium	164-1	83779	85513	85640	84977	1.22	cps
Erbium	164-2	51070	51253	50919	51081	0.33	cps
Gadolinium	160-1	85730	87172	85747	86216	0.96	cps
Gadolinium	160-2	55499	54877	54680	55019	0.78	cps
Holmium	165-1	10529181	10698518	10580003	10602567	0.82	cps
Holmium	165-2	6352249	6301179	6361609	6338346	0.51	cps
Indium	115-1	7656664	7819656	7694220	7723513	1.11	cps
Indium	115-2	2353894	2352469	2357630	2354664	0.11	cps
Iron	56-2	592566578	579160351	578121698	583282876	1.38	cps
Iron	57-2	14521758	14569309	14354203	14481756	0.78	cps
Iron	54-2	32177108	31830571	31923429	31977036	0.56	cps
Krypton	83-1	323	237	283	281	15.43	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-17 Instrumnet Name : P7
Client Sample ID : MYE4P7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:43:34 DataFile Name : 118SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	70846	72860	71550	71752	1.42	cps
Lead	207-1	61190	62165	62174	61843	0.91	cps
Lead	208-1	279747	284207	284109	282688	0.90	cps
Lithium	6-1	1243241	1224604	1184730	1217525	2.46	cps
Magnesium	24-2	23145167	22829904	22933253	22969441	0.70	cps
Manganese	55-2	21021372	20688478	20809894	20839915	0.81	cps
Molybdenum	94-1	44056	44852	44648	44518	0.93	cps
Molybdenum	95-1	7042	6845	7158	7015	2.26	cps
Molybdenum	96-1	22841	22768	22387	22665	1.08	cps
Molybdenum	97-1	4044	4031	4094	4056	0.82	cps
Molybdenum	98-1	10060	9887	9900	9949	0.97	cps
Neodymium	150-1	103372	104305	102754	103477	0.75	cps
Neodymium	150-2	52782	51992	52390	52388	0.75	cps
Nickel	60-2	896079	887456	887180	890238	0.57	cps
Phosphorus	31-2	15141	14794	14981	14972	1.16	cps
Potassium	39-2	546092	543349	541967	543803	0.39	cps
Rhodium	103-1	6915449	6891342	6867148	6891313	0.35	cps
Rhodium	103-2	4099547	4048298	4001278	4049708	1.21	cps
Scandium	45-1	4709419	4703140	4713763	4708774	0.11	cps
Scandium	45-2	419534	419585	417223	418781	0.32	cps
Selenium	82-1	566	623	523	570	8.78	cps
Selenium	77-2	233	213	263	237	10.63	cps
Selenium	78-2	373	343	347	354	4.64	cps
Silicon	28-1	29103024	29204196	29147392	29151537	0.17	cps
Silver	107-1	923	790	927	880	8.86	cps
Silver	109-1	657	783	843	761	12.52	cps
Sodium	23-2	142701	141941	143296	142646	0.48	cps
Strontium	86-1	489634	497648	496276	494519	0.87	cps
Strontium	88-1	4476677	4491372	4541949	4503333	0.76	cps
Sulfur	34-1	153237	150988	147991	150739	1.75	cps
Terbium	159-1	10789077	10932251	10893458	10871595	0.68	cps
Terbium	159-2	6287835	6233532	6441187	6320851	1.70	cps
Thallium	203-1	2324	2530	2594	2482	5.69	cps
Thallium	205-1	6161	5978	6121	6087	1.58	cps
Tin	118-1	16106	16633	16633	16457	1.85	cps
Titanium	47-1	1776262	1770759	1782689	1776570	0.34	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-17 Instrumnet Name : P7
Client Sample ID : MYE4P7 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:43:34 DataFile Name : 118SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	24486	24710	25601	24932	2.37	cps
Vanadium	51-2	1509732	1529248	1528045	1522342	0.72	cps
Yttrium	89-1	13666749	13680573	13559449	13635590	0.49	cps
Yttrium	89-2	3289727	3267002	3218406	3258378	1.12	cps
Zinc	66-2	148156	147106	148638	147967	0.53	cps
Zirconium	90-1	107675	109517	109168	108787	0.90	cps
Zirconium	91-1	24728	25035	24945	24902	0.63	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-18 Instrumnet Name : P7
Client Sample ID : MYE4P8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:46:35 DataFile Name : 119SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	10722732	10572580	10671487	10655599	0.72	cps
Antimony	121-1	12402	12569	12909	12627	2.05	cps
Arsenic	75-2	56755	56514	57210	56826	0.62	cps
Barium	135-1	1695271	1688652	1761775	1715233	2.36	cps
Barium	137-1	2920178	3004620	2998337	2974378	1.58	cps
Beryllium	9-1	2550	2372	2461	2461	3.61	cps
Bismuth	209-1	6229427	6827668	6826752	6627949	5.21	cps
Bismuth	209-2	4680640	4590752	4612314	4627902	1.01	cps
Bromine	81-1	14888	16012	16212	15704	4.55	cps
Bromine	81-2	1187	1310	1137	1211	7.37	cps
Cadmium	108-1	673	583	577	611	8.84	cps
Cadmium	106-1	3507	3754	4001	3754	6.57	cps
Cadmium	111-1	7001	7497	7871	7456	5.85	cps
Calcium	43-1	322630	334887	339245	332254	2.59	cps
Calcium	44-1	4913778	5132594	5123449	5056607	2.45	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	429817	426342	424057	426739	0.68	cps
Cobalt	59-2	356138	357688	357476	357101	0.24	cps
Copper	63-2	1338326	1336523	1339578	1338142	0.11	cps
Dysprosium	156-1	94441	95858	97059	95786	1.37	cps
Dysprosium	156-2	71578	71889	72285	71917	0.49	cps
Erbium	164-1	74195	74346	75073	74538	0.63	cps
Erbium	164-2	45290	45397	44999	45229	0.46	cps
Gadolinium	160-1	87689	88893	90969	89184	1.86	cps
Gadolinium	160-2	57775	58311	57966	58017	0.47	cps
Holmium	165-1	9790833	11067539	10771644	10543339	6.34	cps
Holmium	165-2	6707526	6536131	6647821	6630493	1.31	cps
Indium	115-1	7182733	8166320	8190019	7846357	7.33	cps
Indium	115-2	2508285	2505607	2493289	2502394	0.32	cps
Iron	56-2	349640368	344158608	344131915	345976964	0.92	cps
Iron	57-2	8665688	8558718	8490083	8571496	1.03	cps
Iron	54-2	18917000	18861731	18685185	18821305	0.64	cps
Krypton	83-1	310	273	290	291	6.31	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-18 Instrumnet Name : P7
Client Sample ID : MYE4P8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:46:35 DataFile Name : 119SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	747477	736791	749680	744649	0.93	cps
Lead	207-1	639318	651429	655004	648584	1.27	cps
Lead	208-1	2922948	3193058	2973492	3029833	4.74	cps
Lithium	6-1	1099235	1254837	1228456	1194176	6.97	cps
Magnesium	24-2	8894906	8771100	8914438	8860148	0.88	cps
Manganese	55-2	9851049	9774855	9816643	9814182	0.39	cps
Molybdenum	94-1	100307	104774	105519	103534	2.72	cps
Molybdenum	95-1	6665	6438	6832	6645	2.97	cps
Molybdenum	96-1	27836	28678	28942	28486	2.03	cps
Molybdenum	97-1	3757	4034	4117	3969	4.75	cps
Molybdenum	98-1	9813	10007	10180	10000	1.84	cps
Neodymium	150-1	179107	186141	183707	182985	1.95	cps
Neodymium	150-2	95347	94148	95725	95073	0.87	cps
Nickel	60-2	175180	173231	176230	174880	0.87	cps
Phosphorus	31-2	31523	31218	31138	31293	0.65	cps
Potassium	39-2	5449684	5443704	5430113	5441167	0.18	cps
Rhodium	103-1	6578007	7255296	7403007	7078770	6.21	cps
Rhodium	103-2	4293911	4386496	4381232	4353879	1.19	cps
Scandium	45-1	4090364	4634374	4589263	4438000	6.80	cps
Scandium	45-2	419451	415268	419074	417931	0.55	cps
Selenium	82-1	576	614	582	591	3.48	cps
Selenium	77-2	387	270	347	334	17.73	cps
Selenium	78-2	477	487	420	461	7.80	cps
Silicon	28-1	33459974	34938493	35536310	34644926	3.09	cps
Silver	107-1	8403	8919	9070	8797	3.98	cps
Silver	109-1	8236	8202	8499	8312	1.96	cps
Sodium	23-2	133606	132100	132662	132790	0.57	cps
Strontium	86-1	302258	312570	318552	311127	2.65	cps
Strontium	88-1	2728588	2859366	2874111	2820688	2.84	cps
Sulfur	34-1	134304	135417	134183	134634	0.51	cps
Terbium	159-1	10156986	11162100	11146159	10821748	5.32	cps
Terbium	159-2	6453223	6626872	6625408	6568501	1.52	cps
Thallium	203-1	5458	5728	5785	5657	3.09	cps
Thallium	205-1	13580	13760	13650	13664	0.66	cps
Tin	118-1	29113	30131	32112	30452	5.01	cps
Titanium	47-1	1251589	1303569	1330410	1295189	3.09	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-18 Instrumnet Name : P7
Client Sample ID : MYE4P8 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:46:35 DataFile Name : 119SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	44875	44778	45083	44912	0.35	cps
Vanadium	51-2	664051	665010	659644	662901	0.43	cps
Yttrium	89-1	12615139	14256115	14167579	13679611	6.75	cps
Yttrium	89-2	3480705	3454475	3433821	3456334	0.68	cps
Zinc	66-2	250305	249587	248624	249505	0.34	cps
Zirconium	90-1	262053	272308	275183	269848	2.56	cps
Zirconium	91-1	59048	61418	60876	60447	2.05	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-19 Instrumnet Name : P7
Client Sample ID : MYE4P9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:49:40 DataFile Name : 120SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	6321557	6292714	6256967	6290413	0.51	cps
Antimony	121-1	17387	17955	17814	17719	1.67	cps
Arsenic	75-2	42691	42985	42725	42800	0.38	cps
Barium	135-1	1004219	1019276	1017657	1013717	0.82	cps
Barium	137-1	1887076	1918018	1900189	1901761	0.82	cps
Beryllium	9-1	1445	1595	1572	1537	5.27	cps
Bismuth	209-1	6858012	6978014	6936681	6924235	0.88	cps
Bismuth	209-2	4740349	4024416	4832234	4532333	9.76	cps
Bromine	81-1	13099	13503	13760	13454	2.48	cps
Bromine	81-2	597	703	683	661	8.58	cps
Cadmium	108-1	440	363	493	432	15.12	cps
Cadmium	106-1	3494	3247	3297	3346	3.90	cps
Cadmium	111-1	6395	6366	6583	6448	1.83	cps
Calcium	43-1	178335	179160	179460	178985	0.33	cps
Calcium	44-1	2694549	2741746	2742340	2726212	1.01	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	269293	272000	267340	269544	0.87	cps
Cobalt	59-2	237230	235973	233321	235508	0.85	cps
Copper	63-2	909467	913042	912319	911609	0.21	cps
Dysprosium	156-1	59693	61390	61424	60835	1.63	cps
Dysprosium	156-2	45607	45952	45838	45799	0.38	cps
Erbium	164-1	45614	45795	46052	45820	0.48	cps
Erbium	164-2	27844	28025	27544	27804	0.87	cps
Gadolinium	160-1	55365	56480	55596	55814	1.05	cps
Gadolinium	160-2	36217	36471	36110	36266	0.51	cps
Holmium	165-1	10902168	10942143	10801267	10881859	0.67	cps
Holmium	165-2	6616634	5589557	6672031	6292741	9.69	cps
Indium	115-1	8221136	8222610	8282813	8242186	0.43	cps
Indium	115-2	2575052	2141330	2531905	2416096	9.89	cps
Iron	56-2	243331036	243736490	243283096	243450207	0.10	cps
Iron	57-2	6052038	6019162	6017338	6029513	0.32	cps
Iron	54-2	13315893	13313474	13302004	13310457	0.06	cps
Krypton	83-1	253	173	220	216	18.64	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-19 Instrumnet Name : P7
Client Sample ID : MYE4P9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:49:40 DataFile Name : 120SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	573665	580036	584007	579236	0.90	cps
Lead	207-1	504598	512563	514193	510451	1.01	cps
Lead	208-1	2282465	2315840	2327822	2308709	1.02	cps
Lithium	6-1	1214914	1232203	1230140	1225752	0.77	cps
Magnesium	24-2	5490146	5510797	5505098	5502013	0.19	cps
Manganese	55-2	5516452	5568702	5602426	5562527	0.78	cps
Molybdenum	94-1	58356	58382	59370	58703	0.98	cps
Molybdenum	95-1	6178	5758	5811	5916	3.87	cps
Molybdenum	96-1	19476	19767	19360	19534	1.07	cps
Molybdenum	97-1	3754	3650	3594	3666	2.21	cps
Molybdenum	98-1	8796	8593	9346	8912	4.38	cps
Neodymium	150-1	116456	117582	117756	117265	0.60	cps
Neodymium	150-2	60456	60322	60536	60438	0.18	cps
Nickel	60-2	110950	111716	110671	111112	0.49	cps
Phosphorus	31-2	18802	19112	18755	18890	1.03	cps
Potassium	39-2	3241939	3269027	3259903	3256956	0.42	cps
Rhodium	103-1	7527577	7510278	7513383	7517079	0.12	cps
Rhodium	103-2	4411041	3699644	4416077	4175587	9.87	cps
Scandium	45-1	4489987	4503875	4427681	4473848	0.91	cps
Scandium	45-2	397704	340342	402876	380307	9.13	cps
Selenium	82-1	629	630	599	619	2.92	cps
Selenium	77-2	160	150	197	169	14.55	cps
Selenium	78-2	397	443	360	400	10.44	cps
Silicon	28-1	32736031	33080570	33020552	32945718	0.56	cps
Silver	107-1	5541	5471	5664	5559	1.76	cps
Silver	109-1	5251	5231	5341	5274	1.11	cps
Sodium	23-2	94175	94377	93605	94053	0.43	cps
Strontium	86-1	171772	175017	176373	174387	1.36	cps
Strontium	88-1	1556942	1575674	1586136	1572917	0.94	cps
Sulfur	34-1	117310	114868	114290	115489	1.39	cps
Terbium	159-1	11062882	11092175	11237377	11130812	0.84	cps
Terbium	159-2	6600303	5609985	6588943	6266411	9.07	cps
Thallium	203-1	3287	3324	3267	3293	0.87	cps
Thallium	205-1	7676	7535	7912	7708	2.47	cps
Tin	118-1	25686	26314	26538	26179	1.69	cps
Titanium	47-1	924595	940095	936319	933670	0.87	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-19 Instrumnet Name : P7
Client Sample ID : MYE4P9 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:49:40 DataFile Name : 120SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	26172	26807	26964	26648	1.57	cps
Vanadium	51-2	458693	463192	455106	458997	0.88	cps
Yttrium	89-1	13951431	13880126	13860254	13897270	0.35	cps
Yttrium	89-2	3286304	2855605	3312061	3151323	8.14	cps
Zinc	66-2	198068	199170	197923	198387	0.34	cps
Zirconium	90-1	148589	151529	152478	150866	1.34	cps
Zirconium	91-1	33508	34357	34491	34119	1.56	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-20 Instrumnet Name : P7
Client Sample ID : MYE4Q0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:52:46 DataFile Name : 121SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	13831966	13910337	13642839	13795047	1.00	cps
Antimony	121-1	21917	22264	22738	22306	1.85	cps
Arsenic	75-2	76861	77625	75805	76764	1.19	cps
Barium	135-1	666122	680920	682166	676403	1.32	cps
Barium	137-1	1163306	1179011	1190176	1177498	1.15	cps
Beryllium	9-1	1500	1589	1572	1554	3.04	cps
Bismuth	209-1	6875993	6774348	6824047	6824796	0.74	cps
Bismuth	209-2	4627444	4658870	4565154	4617156	1.03	cps
Bromine	81-1	12956	13783	13783	13507	3.54	cps
Bromine	81-2	720	647	727	698	6.36	cps
Cadmium	108-1	570	607	533	570	6.43	cps
Cadmium	106-1	3570	3394	3507	3490	2.56	cps
Cadmium	111-1	8935	8743	9240	8973	2.80	cps
Calcium	43-1	549527	558581	561900	556669	1.15	cps
Calcium	44-1	8397518	8434098	8527789	8453135	0.79	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	238338	237125	234557	236673	0.82	cps
Cobalt	59-2	611689	608815	615568	612024	0.55	cps
Copper	63-2	1051563	1053077	1049540	1051393	0.17	cps
Dysprosium	156-1	41906	43049	43296	42750	1.74	cps
Dysprosium	156-2	29548	29975	30259	29928	1.20	cps
Erbium	164-1	49353	48397	50016	49255	1.65	cps
Erbium	164-2	30697	29909	29758	30121	1.67	cps
Gadolinium	160-1	44387	45310	45828	45175	1.62	cps
Gadolinium	160-2	29130	29815	29528	29491	1.17	cps
Holmium	165-1	11133364	11057665	10849706	11013578	1.33	cps
Holmium	165-2	6724397	6685508	6479971	6629959	1.98	cps
Indium	115-1	8326470	8260524	8291298	8292764	0.40	cps
Indium	115-2	2522285	2555154	2493342	2523594	1.23	cps
Iron	56-2	522773592	521768432	515285059	519942361	0.78	cps
Iron	57-2	12966891	12844433	12811956	12874426	0.63	cps
Iron	54-2	28201768	28137950	27757302	28032340	0.86	cps
Krypton	83-1	293	270	293	286	4.72	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-20 Instrumnet Name : P7
Client Sample ID : MYE4Q0 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:52:46 DataFile Name : 121SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	279403	288147	288401	285317	1.80	cps
Lead	207-1	241948	247621	247328	245632	1.30	cps
Lead	208-1	1091420	1127280	1134721	1117807	2.07	cps
Lithium	6-1	1225400	1219048	1204078	1216175	0.90	cps
Magnesium	24-2	9845089	9884428	9738604	9822707	0.77	cps
Manganese	55-2	11657461	11801259	11742973	11733898	0.62	cps
Molybdenum	94-1	23482	24711	24147	24113	2.55	cps
Molybdenum	95-1	8196	9000	9093	8763	5.63	cps
Molybdenum	96-1	20391	21149	21039	20859	1.96	cps
Molybdenum	97-1	5281	5454	5604	5447	2.97	cps
Molybdenum	98-1	12909	13259	13333	13167	1.72	cps
Neodymium	150-1	58353	59113	58280	58582	0.79	cps
Neodymium	150-2	29621	30319	30263	30068	1.29	cps
Nickel	60-2	76071	76738	76892	76567	0.57	cps
Phosphorus	31-2	16939	17971	17203	17371	3.08	cps
Potassium	39-2	1209925	1205269	1206916	1207370	0.20	cps
Rhodium	103-1	7612319	7383989	7447361	7481223	1.58	cps
Rhodium	103-2	4365284	4412006	4301806	4359698	1.27	cps
Scandium	45-1	4726334	4626750	4616240	4656441	1.30	cps
Scandium	45-2	413794	411018	404321	409711	1.19	cps
Selenium	82-1	517	570	505	531	6.45	cps
Selenium	77-2	110	127	117	118	7.12	cps
Selenium	78-2	363	380	357	367	3.28	cps
Silicon	28-1	32021890	32637697	32748887	32469491	1.21	cps
Silver	107-1	14547	14481	15111	14713	2.35	cps
Silver	109-1	13386	13990	13710	13695	2.21	cps
Sodium	23-2	448432	445975	442458	445622	0.67	cps
Strontium	86-1	415706	425166	426342	422405	1.38	cps
Strontium	88-1	3776354	3816665	3893209	3828743	1.55	cps
Sulfur	34-1	110690	113779	114626	113032	1.83	cps
Terbium	159-1	11532964	11264867	11230223	11342685	1.46	cps
Terbium	159-2	6532931	6614743	6498795	6548823	0.91	cps
Thallium	203-1	2210	2344	2494	2349	6.04	cps
Thallium	205-1	5631	5868	5581	5693	2.69	cps
Tin	118-1	17157	17941	17828	17642	2.40	cps
Titanium	47-1	2754911	2811172	2840075	2802053	1.55	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-20

Instrumnet Name :

P7

Client Sample ID :

MYE4Q0

Dilution Factor :

1

Date & Time Acquired :

2024-11-02 23:52:46

DataFile Name :

121SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	16754	16864	16854	16824	0.36	cps
Vanadium	51-2	894441	894035	880782	889753	0.87	cps
Yttrium	89-1	14488664	14141691	14090291	14240215	1.52	cps
Yttrium	89-2	3375688	3365483	3346719	3362630	0.44	cps
Zinc	66-2	424943	423326	422157	423475	0.33	cps
Zirconium	90-1	50403	53876	51674	51984	3.38	cps
Zirconium	91-1	11752	11324	12022	11699	3.00	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-21 Instrumnet Name : P7
Client Sample ID : MYE4Q1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:55:54 DataFile Name : 122SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7906502	7956384	7850314	7904400	0.67	cps
Antimony	121-1	14571	14424	14851	14615	1.48	cps
Arsenic	75-2	56483	56413	56403	56433	0.08	cps
Barium	135-1	1306246	1332751	1329790	1322929	1.10	cps
Barium	137-1	2454855	2455076	2483414	2464448	0.67	cps
Beryllium	9-1	1906	1967	2006	1959	2.57	cps
Bismuth	209-1	6892726	6965360	6904984	6921024	0.56	cps
Bismuth	209-2	4860247	4812379	4792733	4821786	0.72	cps
Bromine	81-1	14587	15345	14994	14975	2.53	cps
Bromine	81-2	807	980	923	903	9.78	cps
Cadmium	108-1	533	543	430	502	12.49	cps
Cadmium	106-1	3300	3354	3490	3382	2.90	cps
Cadmium	111-1	6128	6138	6276	6181	1.34	cps
Calcium	43-1	202763	208763	209063	206863	1.72	cps
Calcium	44-1	3108015	3139502	3156468	3134662	0.78	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	343590	340194	337561	340448	0.89	cps
Cobalt	59-2	317004	314516	314143	315221	0.49	cps
Copper	63-2	1148285	1139272	1141658	1143072	0.41	cps
Dysprosium	156-1	75904	77482	77385	76924	1.15	cps
Dysprosium	156-2	57929	59468	58756	58718	1.31	cps
Erbium	164-1	57337	59041	59636	58671	2.03	cps
Erbium	164-2	35689	36016	35686	35797	0.53	cps
Gadolinium	160-1	70419	71947	72342	71569	1.42	cps
Gadolinium	160-2	45567	45968	45871	45802	0.46	cps
Holmium	165-1	10907086	11082275	10994280	10994547	0.80	cps
Holmium	165-2	6775295	6712444	6733814	6740518	0.47	cps
Indium	115-1	8187988	8286972	8070826	8181929	1.32	cps
Indium	115-2	2556237	2538280	2579909	2558142	0.82	cps
Iron	56-2	299758302	297669669	297405696	298277889	0.43	cps
Iron	57-2	7458810	7401599	7447509	7435973	0.41	cps
Iron	54-2	16469974	16248199	16180638	16299604	0.93	cps
Krypton	83-1	297	240	280	272	10.70	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-21 Instrumnet Name : P7
Client Sample ID : MYE4Q1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:55:54 DataFile Name : 122SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	588595	600850	604762	598069	1.41	cps
Lead	207-1	515717	526888	532074	524893	1.59	cps
Lead	208-1	2342416	2394792	2413224	2383477	1.54	cps
Lithium	6-1	1251398	1227759	1199144	1226101	2.13	cps
Magnesium	24-2	7082950	7070158	6975775	7042961	0.83	cps
Manganese	55-2	7544728	7525134	7499102	7522988	0.30	cps
Molybdenum	94-1	81009	81786	83077	81957	1.27	cps
Molybdenum	95-1	6088	5788	6381	6086	4.88	cps
Molybdenum	96-1	24377	24511	24330	24406	0.38	cps
Molybdenum	97-1	3761	3781	3751	3764	0.41	cps
Molybdenum	98-1	9163	9400	9213	9259	1.35	cps
Neodymium	150-1	146697	147387	149432	147839	0.96	cps
Neodymium	150-2	76634	74197	76439	75757	1.79	cps
Nickel	60-2	147860	148039	147319	147740	0.25	cps
Phosphorus	31-2	19336	19683	19589	19536	0.92	cps
Potassium	39-2	3335419	3317333	3264654	3305802	1.11	cps
Rhodium	103-1	7479574	7390515	7364321	7411470	0.82	cps
Rhodium	103-2	4388654	4364196	4360424	4371091	0.35	cps
Scandium	45-1	4497050	4503034	4530208	4510097	0.39	cps
Scandium	45-2	404451	403345	399448	402415	0.65	cps
Selenium	82-1	527	551	518	532	3.29	cps
Selenium	77-2	217	270	267	251	11.90	cps
Selenium	78-2	380	400	400	393	2.94	cps
Silicon	28-1	33799034	34346854	34272747	34139545	0.87	cps
Silver	107-1	6138	6311	6301	6250	1.56	cps
Silver	109-1	5898	5801	5671	5790	1.97	cps
Sodium	23-2	95735	94819	94353	94969	0.74	cps
Strontium	86-1	210747	212572	213293	212204	0.62	cps
Strontium	88-1	1892386	1949899	1925749	1922678	1.50	cps
Sulfur	34-1	111180	107391	107917	108830	1.89	cps
Terbium	159-1	11157293	11158833	11228965	11181697	0.37	cps
Terbium	159-2	6646476	6678171	6666417	6663688	0.24	cps
Thallium	203-1	4324	4477	4454	4419	1.87	cps
Thallium	205-1	10321	10578	10481	10460	1.24	cps
Tin	118-1	23202	23843	23406	23484	1.40	cps
Titanium	47-1	1125125	1138342	1143596	1135687	0.84	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-21 Instrumnet Name : P7
Client Sample ID : MYE4Q1 Dilution Factor : 1
Date & Time Acquired : 2024-11-02 23:55:54 DataFile Name : 122SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	36316	36834	36533	36561	0.71	cps
Vanadium	51-2	612077	605841	606429	608116	0.57	cps
Yttrium	89-1	13921728	14042624	13914376	13959576	0.52	cps
Yttrium	89-2	3387979	3347502	3368270	3367917	0.60	cps
Zinc	66-2	167696	165202	166962	166620	0.77	cps
Zirconium	90-1	211341	215751	215041	214045	1.11	cps
Zirconium	91-1	47075	47898	48426	47800	1.42	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 23:58:59 DataFile Name : 123SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1697441	1692137	1697752	1695777	0.19	cps
Antimony	121-1	1183	1270	1260	1238	3.83	cps
Arsenic	75-2	4921	5098	4874	4964	2.37	cps
Barium	135-1	214745	218443	218602	217263	1.00	cps
Barium	137-1	369606	375425	379171	374734	1.29	cps
Beryllium	9-1	450	400	367	406	10.34	cps
Bismuth	209-1	7007858	7092896	7031397	7044050	0.62	cps
Bismuth	209-2	4358137	4866199	4901392	4708576	6.46	cps
Bromine	81-1	11845	12248	12315	12136	2.10	cps
Bromine	81-2	357	343	377	359	4.68	cps
Cadmium	108-1	53	57	50	53	6.25	cps
Cadmium	106-1	2644	2530	2564	2579	2.26	cps
Cadmium	111-1	2177	2138	2180	2165	1.06	cps
Calcium	43-1	164919	167479	168790	167062	1.18	cps
Calcium	44-1	2516500	2551026	2556862	2541463	0.86	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	67383	67959	67798	67713	0.44	cps
Cobalt	59-2	59584	59390	59742	59572	0.30	cps
Copper	63-2	135594	135463	135198	135418	0.15	cps
Dysprosium	156-1	16857	17064	17070	16997	0.71	cps
Dysprosium	156-2	12709	12322	12592	12541	1.58	cps
Erbium	164-1	15472	15272	14985	15243	1.61	cps
Erbium	164-2	9186	9287	9273	9249	0.59	cps
Gadolinium	160-1	16410	16747	17054	16737	1.92	cps
Gadolinium	160-2	10954	10651	10821	10809	1.41	cps
Holmium	165-1	10893170	11030804	10929027	10951001	0.65	cps
Holmium	165-2	6071241	6864567	6665912	6533907	6.32	cps
Indium	115-1	8293068	8247761	8325156	8288662	0.47	cps
Indium	115-2	2309222	2576374	2557561	2481052	6.01	cps
Iron	56-2	59424276	58463116	59270434	59052609	0.87	cps
Iron	57-2	1497988	1464220	1474523	1478910	1.17	cps
Iron	54-2	3229127	3160577	3165232	3184978	1.20	cps
Krypton	83-1	150	240	207	199	22.88	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2 Dilution Factor : 4
Date & Time Acquired : 2024-11-02 23:58:59 DataFile Name : 123SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	68622	68216	70287	69042	1.59	cps
Lead	207-1	57695	58254	57695	57881	0.56	cps
Lead	208-1	265429	268464	269504	267799	0.79	cps
Lithium	6-1	1231172	1234234	1229610	1231672	0.19	cps
Magnesium	24-2	1527237	1510501	1519594	1519110	0.55	cps
Manganese	55-2	1542093	1517492	1526046	1528543	0.82	cps
Molybdenum	94-1	9203	9186	9176	9189	0.15	cps
Molybdenum	95-1	1703	1663	1630	1666	2.20	cps
Molybdenum	96-1	3961	4041	4097	4033	1.70	cps
Molybdenum	97-1	883	967	1107	986	11.45	cps
Molybdenum	98-1	2324	2484	2554	2454	4.81	cps
Neodymium	150-1	29377	29614	29628	29540	0.48	cps
Neodymium	150-2	15385	15278	15202	15288	0.60	cps
Nickel	60-2	22190	22340	21889	22140	1.04	cps
Phosphorus	31-2	6938	6732	6655	6775	2.16	cps
Potassium	39-2	1258210	1197100	1185406	1213572	3.22	cps
Rhodium	103-1	7533068	7424816	7440571	7466152	0.78	cps
Rhodium	103-2	4053456	4500784	4416673	4323638	5.50	cps
Scandium	45-1	4237876	4181413	4214823	4211370	0.67	cps
Scandium	45-2	349857	387351	383003	373404	5.49	cps
Selenium	82-1	493	412	475	460	9.28	cps
Selenium	77-2	57	77	90	74	22.53	cps
Selenium	78-2	283	327	260	290	11.66	cps
Silicon	28-1	8535640	8651828	8701570	8629679	0.99	cps
Silver	107-1	887	773	833	831	6.82	cps
Silver	109-1	637	727	740	701	8.02	cps
Sodium	23-2	81374	82559	82271	82068	0.75	cps
Strontium	86-1	82553	84129	84581	83754	1.27	cps
Strontium	88-1	725026	731210	737219	731152	0.83	cps
Sulfur	34-1	107300	107233	109153	107895	1.01	cps
Terbium	159-1	11318269	11279325	11244137	11280577	0.33	cps
Terbium	159-2	6027755	6703032	6561097	6430628	5.54	cps
Thallium	203-1	723	740	633	699	8.21	cps
Thallium	205-1	1707	1783	1813	1768	3.11	cps
Tin	118-1	5494	5991	5911	5799	4.60	cps
Titanium	47-1	308660	313010	315513	312394	1.11	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-03DLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4K2

Dilution Factor :

4

Date & Time Acquired :

2024-11-02 23:58:59

DataFile Name :

123SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8363	8416	8369	8383	0.35	cps
Vanadium	51-2	116065	115088	117322	116158	0.96	cps
Yttrium	89-1	13333569	13259150	13334009	13308909	0.32	cps
Yttrium	89-2	2989722	3288439	3193888	3157350	4.84	cps
Zinc	66-2	30267	30428	30651	30449	0.63	cps
Zirconium	90-1	21740	22320	22340	22133	1.54	cps
Zirconium	91-1	4951	4774	4914	4880	1.91	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-04DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2D Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:02:10 DataFile Name : 124SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1695032	1644846	1671039	1670306	1.50	cps
Antimony	121-1	1047	1280	1283	1203	11.28	cps
Arsenic	75-2	5061	4848	4834	4914	2.59	cps
Barium	135-1	209289	212333	212661	211427	0.88	cps
Barium	137-1	362032	366963	371712	366902	1.32	cps
Beryllium	9-1	439	361	400	400	9.72	cps
Bismuth	209-1	7006818	7046444	6810602	6954621	1.82	cps
Bismuth	209-2	4807239	5028866	4819476	4885194	2.55	cps
Bromine	81-1	12499	11872	12315	12228	2.64	cps
Bromine	81-2	383	327	373	361	8.38	cps
Cadmium	108-1	53	50	47	50	6.66	cps
Cadmium	106-1	2524	2440	2400	2455	2.56	cps
Cadmium	111-1	2163	2063	2004	2077	3.85	cps
Calcium	43-1	162796	165566	165413	164592	0.95	cps
Calcium	44-1	2511732	2548473	2541030	2533745	0.77	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	66743	66824	65923	66497	0.75	cps
Cobalt	59-2	59099	58838	58758	58898	0.30	cps
Copper	63-2	132483	132621	132258	132454	0.14	cps
Dysprosium	156-1	16356	16563	17157	16692	2.49	cps
Dysprosium	156-2	12169	12396	12039	12201	1.48	cps
Erbium	164-1	14701	14127	15082	14637	3.28	cps
Erbium	164-2	9080	8820	9013	8971	1.51	cps
Gadolinium	160-1	16069	15859	16173	16034	1.00	cps
Gadolinium	160-2	10738	10751	10821	10770	0.42	cps
Holmium	165-1	10943058	10954838	10790421	10896106	0.84	cps
Holmium	165-2	6673137	6764448	6631027	6689537	1.02	cps
Indium	115-1	8256701	8252333	8221473	8243502	0.23	cps
Indium	115-2	2570946	2617647	2539803	2576132	1.52	cps
Iron	56-2	57999077	57807231	57385609	57730639	0.54	cps
Iron	57-2	1424941	1438650	1446348	1436646	0.75	cps
Iron	54-2	3154020	3173995	3136190	3154735	0.60	cps
Krypton	83-1	227	180	213	207	11.63	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-04DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2D Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:02:10 DataFile Name : 124SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	64971	66404	66904	66093	1.52	cps
Lead	207-1	54586	55208	56765	55519	2.02	cps
Lead	208-1	254638	257747	259874	257420	1.02	cps
Lithium	6-1	1236455	1227244	1211805	1225168	1.02	cps
Magnesium	24-2	1502377	1491151	1474518	1489348	0.94	cps
Manganese	55-2	1503321	1492230	1497415	1497656	0.37	cps
Molybdenum	94-1	8743	8833	9183	8919	2.61	cps
Molybdenum	95-1	1407	1560	1637	1535	7.63	cps
Molybdenum	96-1	3854	3897	4001	3917	1.92	cps
Molybdenum	97-1	863	1057	977	966	10.06	cps
Molybdenum	98-1	2484	2314	2540	2446	4.82	cps
Neodymium	150-1	27998	28696	29387	28694	2.42	cps
Neodymium	150-2	14911	14671	15021	14868	1.21	cps
Nickel	60-2	22280	22056	21833	22056	1.01	cps
Phosphorus	31-2	6565	6668	6491	6575	1.35	cps
Potassium	39-2	1224941	1256698	1231571	1237737	1.35	cps
Rhodium	103-1	7620969	7467257	7436062	7508096	1.32	cps
Rhodium	103-2	4472887	4480256	4436692	4463278	0.52	cps
Scandium	45-1	4294814	4239880	4261769	4265488	0.65	cps
Scandium	45-2	381053	379071	378804	379643	0.32	cps
Selenium	82-1	393	532	443	456	15.45	cps
Selenium	77-2	60	57	50	56	9.17	cps
Selenium	78-2	290	320	257	289	10.97	cps
Silicon	28-1	8365070	8521668	8620884	8502541	1.52	cps
Silver	107-1	777	753	803	778	3.22	cps
Silver	109-1	710	683	843	746	11.50	cps
Sodium	23-2	79879	80167	79624	79890	0.34	cps
Strontium	86-1	81507	83508	82389	82468	1.22	cps
Strontium	88-1	710816	719794	720844	717151	0.77	cps
Sulfur	34-1	114008	113085	116406	114499	1.50	cps
Terbium	159-1	11340613	11134667	11253807	11243029	0.92	cps
Terbium	159-2	6686865	6834790	6539785	6687147	2.21	cps
Thallium	203-1	707	797	750	751	5.99	cps
Thallium	205-1	1670	1743	1630	1681	3.42	cps
Tin	118-1	5921	5654	5568	5715	3.22	cps
Titanium	47-1	305885	309701	310810	308799	0.84	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-04DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2D Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:02:10 DataFile Name : 124SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8126	8559	8359	8348	2.60	cps
Vanadium	51-2	115612	114950	115618	115393	0.33	cps
Yttrium	89-1	13457301	13317854	13280766	13351974	0.70	cps
Yttrium	89-2	3281739	3203261	3241702	3242234	1.21	cps
Zinc	66-2	30337	30274	30000	30204	0.59	cps
Zirconium	90-1	21265	21626	21913	21602	1.50	cps
Zirconium	91-1	4537	4878	4891	4769	4.20	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4K2L Dilution Factor : 20
Date & Time Acquired : 2024-11-03 00:05:18 DataFile Name : 125SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	302501	303487	301984	302658	0.25	cps
Antimony	121-1	290	323	277	297	8.10	cps
Arsenic	75-2	893	903	803	867	6.36	cps
Barium	135-1	37536	38395	38886	38272	1.79	cps
Barium	137-1	66112	65862	66163	66046	0.24	cps
Beryllium	9-1	89	111	83	94	15.56	cps
Bismuth	209-1	6908044	6930417	6827668	6888709	0.78	cps
Bismuth	209-2	4873650	4793794	4771734	4813059	1.11	cps
Bromine	81-1	11468	11424	11221	11371	1.16	cps
Bromine	81-2	117	133	120	123	7.15	cps
Cadmium	108-1	20	17	13	17	20.01	cps
Cadmium	106-1	2417	2410	2394	2407	0.50	cps
Cadmium	111-1	1775	1729	1776	1760	1.53	cps
Calcium	43-1	30237	30110	31082	30476	1.73	cps
Calcium	44-1	495313	495599	495491	495468	0.03	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	12699	12398	12695	12598	1.37	cps
Cobalt	59-2	10944	10824	10737	10835	0.96	cps
Copper	63-2	24413	25261	24554	24743	1.84	cps
Dysprosium	156-1	2987	3040	3040	3023	1.02	cps
Dysprosium	156-2	2157	2274	2134	2188	3.43	cps
Erbium	164-1	2634	2594	2614	2614	0.77	cps
Erbium	164-2	1613	1613	1610	1612	0.12	cps
Gadolinium	160-1	2947	3004	2964	2971	0.98	cps
Gadolinium	160-2	2157	2180	2110	2149	1.66	cps
Holmium	165-1	10867454	10767678	10637383	10757505	1.07	cps
Holmium	165-2	6631713	6540534	6539219	6570489	0.81	cps
Indium	115-1	8205738	8236508	8161247	8201164	0.46	cps
Indium	115-2	2585366	2549871	2575208	2570148	0.71	cps
Iron	56-2	10416218	10437792	10390410	10414807	0.23	cps
Iron	57-2	261840	264536	265718	264031	0.75	cps
Iron	54-2	586465	584365	582590	584473	0.33	cps
Krypton	83-1	187	217	190	198	8.31	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4K2L Dilution Factor : 20
Date & Time Acquired : 2024-11-03 00:05:18 DataFile Name : 125SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	12486	12322	12299	12369	0.82	cps
Lead	207-1	10444	10814	10611	10623	1.75	cps
Lead	208-1	48006	48811	48467	48428	0.83	cps
Lithium	6-1	1245396	1243104	1242823	1243774	0.11	cps
Magnesium	24-2	277321	273908	274077	275102	0.70	cps
Manganese	55-2	269147	266435	266023	267202	0.64	cps
Molybdenum	94-1	1897	1823	1790	1837	2.97	cps
Molybdenum	95-1	387	407	357	383	6.57	cps
Molybdenum	96-1	807	1013	907	909	11.37	cps
Molybdenum	97-1	233	210	283	242	15.47	cps
Molybdenum	98-1	617	553	573	581	5.57	cps
Neodymium	150-1	5254	5328	4954	5179	3.82	cps
Neodymium	150-2	2770	2647	2710	2709	2.28	cps
Nickel	60-2	4404	3987	4004	4132	5.71	cps
Phosphorus	31-2	1410	1477	1367	1418	3.91	cps
Potassium	39-2	258954	258707	257144	258268	0.38	cps
Rhodium	103-1	7657726	7720291	7537297	7638438	1.22	cps
Rhodium	103-2	4419901	4433586	4438060	4430516	0.21	cps
Scandium	45-1	4214528	4264404	4120452	4199795	1.74	cps
Scandium	45-2	372986	368623	374163	371924	0.78	cps
Selenium	82-1	403	392	417	404	3.15	cps
Selenium	77-2	3	10	13	9	57.30	cps
Selenium	78-2	197	223	273	231	16.84	cps
Silicon	28-1	1705734	1718348	1730229	1718104	0.71	cps
Silver	107-1	183	210	233	209	11.98	cps
Silver	109-1	200	140	160	167	18.33	cps
Sodium	23-2	27798	26947	27815	27520	1.80	cps
Strontium	86-1	14904	15655	15308	15289	2.46	cps
Strontium	88-1	130886	130277	131371	130845	0.42	cps
Sulfur	34-1	109637	109620	110899	110052	0.67	cps
Terbium	159-1	11268922	11217699	10858739	11115120	2.01	cps
Terbium	159-2	6433150	6600040	6518830	6517340	1.28	cps
Thallium	203-1	193	210	243	216	11.81	cps
Thallium	205-1	530	517	540	529	2.21	cps
Tin	118-1	2194	2220	2307	2240	2.65	cps
Titanium	47-1	55943	55853	56762	56186	0.89	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03LDLX20 Instrumnet Name : P7
Client Sample ID : MYE4K2L Dilution Factor : 20
Date & Time Acquired : 2024-11-03 00:05:18 DataFile Name : 125SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	1547	1520	1570	1546	1.62	cps
Vanadium	51-2	20734	20708	20597	20680	0.35	cps
Yttrium	89-1	13212656	13127087	12978625	13106123	0.90	cps
Yttrium	89-2	3190578	3131125	3154828	3158844	0.95	cps
Zinc	66-2	6025	6075	5618	5906	4.24	cps
Zirconium	90-1	4234	4094	4357	4228	3.12	cps
Zirconium	91-1	987	953	967	969	1.73	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-05DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2S Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:08:32 DataFile Name : 126SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1683878	1696558	1709505	1696647	0.76	cps
Antimony	121-1	142110	143572	146588	144090	1.58	cps
Arsenic	75-2	7692	7689	7445	7609	1.86	cps
Barium	135-1	933293	944244	942529	940022	0.63	cps
Barium	137-1	1770141	1743477	1789768	1767795	1.31	cps
Beryllium	9-1	10014	10281	10832	10376	4.02	cps
Bismuth	209-1	7001719	6451054	7047755	6833509	4.86	cps
Bismuth	209-2	4940665	4779309	4781744	4833906	1.91	cps
Bromine	81-1	11828	12172	11895	11965	1.52	cps
Bromine	81-2	267	350	290	302	14.23	cps
Cadmium	108-1	1393	1407	1217	1339	7.92	cps
Cadmium	106-1	4514	3881	4374	4256	7.82	cps
Cadmium	111-1	20706	19973	20365	20348	1.80	cps
Calcium	43-1	160916	163170	164868	162985	1.22	cps
Calcium	44-1	2469479	2537437	2501668	2502861	1.36	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	182173	184093	182769	183011	0.54	cps
Cobalt	59-2	505434	506425	503304	505055	0.32	cps
Copper	63-2	282624	283170	287223	284339	0.88	cps
Dysprosium	156-1	16283	16386	16313	16327	0.33	cps
Dysprosium	156-2	12219	11955	12029	12067	1.13	cps
Erbium	164-1	14154	14621	14397	14391	1.62	cps
Erbium	164-2	8940	8709	8396	8682	3.14	cps
Gadolinium	160-1	15262	15856	15922	15680	2.32	cps
Gadolinium	160-2	10567	10311	10698	10525	1.87	cps
Holmium	165-1	10693962	10158329	10899785	10584025	3.62	cps
Holmium	165-2	6761760	6546673	6434182	6580872	2.53	cps
Indium	115-1	8255460	7659703	8317594	8077586	4.50	cps
Indium	115-2	2597396	2550738	2499114	2549082	1.93	cps
Iron	56-2	56279779	56858866	57603947	56914197	1.17	cps
Iron	57-2	1409754	1428058	1462767	1433526	1.88	cps
Iron	54-2	3143340	3100059	3154157	3132519	0.91	cps
Krypton	83-1	180	170	233	194	17.51	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-05DLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2S Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:08:32 DataFile Name : 126SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	86275	88425	87036	87245	1.25	cps
Lead	207-1	72874	74959	74017	73950	1.41	cps
Lead	208-1	335823	344593	341062	340493	1.30	cps
Lithium	6-1	1229962	1168019	1221110	1206364	2.78	cps
Magnesium	24-2	1466544	1471129	1501361	1479678	1.28	cps
Manganese	55-2	1659692	1688205	1691624	1679841	1.04	cps
Molybdenum	94-1	8483	8779	8726	8663	1.83	cps
Molybdenum	95-1	1483	1613	1547	1548	4.20	cps
Molybdenum	96-1	3784	3731	4124	3879	5.50	cps
Molybdenum	97-1	1027	923	927	959	6.12	cps
Molybdenum	98-1	2307	2354	2340	2334	1.03	cps
Neodymium	150-1	28579	29097	28205	28627	1.56	cps
Neodymium	150-2	14855	14588	14491	14644	1.29	cps
Nickel	60-2	136196	136851	134386	135811	0.94	cps
Phosphorus	31-2	6675	6288	6715	6559	3.59	cps
Potassium	39-2	1231588	1222842	1217663	1224031	0.58	cps
Rhodium	103-1	7430044	6992244	7565649	7329312	4.09	cps
Rhodium	103-2	4390794	4321790	4399128	4370571	0.97	cps
Scandium	45-1	4224296	3993429	4271803	4163176	3.58	cps
Scandium	45-2	375935	372106	378692	375578	0.88	cps
Selenium	82-1	2546	2629	2643	2606	2.01	cps
Selenium	77-2	217	230	230	226	3.41	cps
Selenium	78-2	857	890	907	884	2.88	cps
Silicon	28-1	8321048	8456394	8453328	8410257	0.92	cps
Silver	107-1	74577	78052	77184	76604	2.36	cps
Silver	109-1	72524	73391	73730	73215	0.85	cps
Sodium	23-2	76680	77082	78204	77322	1.02	cps
Strontium	86-1	79583	80703	80448	80245	0.73	cps
Strontium	88-1	694871	707071	704366	702102	0.91	cps
Sulfur	34-1	117686	115469	115398	116184	1.12	cps
Terbium	159-1	10990779	10376264	11158983	10842009	3.80	cps
Terbium	159-2	6592961	6728400	6500654	6607338	1.73	cps
Thallium	203-1	67476	67499	68109	67695	0.53	cps
Thallium	205-1	160612	163341	161146	161700	0.89	cps
Tin	118-1	5404	5594	5675	5558	2.50	cps
Titanium	47-1	299131	308198	306497	304609	1.58	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4519-05DLX4

Instrumnet Name :

P7

Client Sample ID :

MYE4K2S

Dilution Factor :

4

Date & Time Acquired :

2024-11-03 00:08:32

DataFile Name :

126SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	8266	8333	8249	8283	0.53	cps
Vanadium	51-2	368673	367223	365981	367292	0.37	cps
Yttrium	89-1	13251521	12419320	13410661	13027167	4.09	cps
Yttrium	89-2	3183031	3160469	3183797	3175765	0.42	cps
Zinc	66-2	79406	79694	80298	79799	0.57	cps
Zirconium	90-1	20868	21072	20982	20974	0.49	cps
Zirconium	91-1	4818	5158	4391	4789	8.02	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:11:39 DataFile Name : 127SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1835625	1831566	1833405	1833532	0.11	cps
Antimony	121-1	678619	683535	687723	683292	0.67	cps
Arsenic	75-2	18088	17991	17794	17958	0.83	cps
Barium	135-1	4008875	4060372	4068148	4045798	0.80	cps
Barium	137-1	6924133	7006843	6971504	6967494	0.60	cps
Beryllium	9-1	47573	48960	48258	48264	1.44	cps
Bismuth	209-1	6855764	6873350	6824671	6851262	0.36	cps
Bismuth	209-2	4719849	4627444	4766052	4704448	1.50	cps
Bromine	81-1	11975	11818	11681	11825	1.24	cps
Bromine	81-2	357	303	370	343	10.27	cps
Cadmium	108-1	5741	5631	5801	5725	1.51	cps
Cadmium	106-1	10541	10200	10694	10478	2.41	cps
Cadmium	111-1	88920	89655	90257	89611	0.75	cps
Calcium	43-1	154308	155275	157319	155634	0.99	cps
Calcium	44-1	2386539	2382244	2398035	2388939	0.34	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	625144	622905	623565	623871	0.18	cps
Cobalt	59-2	2277905	2231664	2230249	2246606	1.21	cps
Copper	63-2	850567	852034	845938	849513	0.37	cps
Dysprosium	156-1	15485	15679	15779	15648	0.95	cps
Dysprosium	156-2	11808	11918	11632	11786	1.23	cps
Erbium	164-1	13747	14074	13543	13788	1.94	cps
Erbium	164-2	8469	8609	8376	8485	1.38	cps
Gadolinium	160-1	15008	15482	15215	15235	1.56	cps
Gadolinium	160-2	10210	10217	10080	10169	0.76	cps
Holmium	165-1	10775589	10647772	10704809	10709390	0.60	cps
Holmium	165-2	6480086	6504562	6540058	6508235	0.46	cps
Indium	115-1	8137485	8037389	8084240	8086371	0.62	cps
Indium	115-2	2464853	2496353	2489352	2483520	0.67	cps
Iron	56-2	56801641	56486841	55950606	56413029	0.76	cps
Iron	57-2	1416439	1389742	1381176	1395786	1.32	cps
Iron	54-2	3148178	3083584	3106692	3112818	1.05	cps
Krypton	83-1	157	190	167	171	10.00	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:11:39 DataFile Name : 127SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	166263	168011	170592	168289	1.29	cps
Lead	207-1	142724	144385	145139	144082	0.86	cps
Lead	208-1	655906	665468	665389	662254	0.83	cps
Lithium	6-1	1223606	1234484	1215688	1224593	0.77	cps
Magnesium	24-2	1391956	1396872	1420838	1403222	1.10	cps
Manganese	55-2	2439067	2410218	2372376	2407220	1.39	cps
Molybdenum	94-1	8509	8209	8579	8433	2.33	cps
Molybdenum	95-1	1517	1527	1580	1541	2.21	cps
Molybdenum	96-1	3821	3674	3967	3821	3.84	cps
Molybdenum	97-1	880	890	860	877	1.74	cps
Molybdenum	98-1	2297	2364	2364	2341	1.64	cps
Neodymium	150-1	26986	27273	27507	27255	0.96	cps
Neodymium	150-2	14614	14121	14090	14275	2.06	cps
Nickel	60-2	564826	562130	566953	564636	0.43	cps
Phosphorus	31-2	6311	6548	6488	6449	1.91	cps
Potassium	39-2	1181859	1165593	1168409	1171954	0.74	cps
Rhodium	103-1	7487855	7322107	7353583	7387848	1.19	cps
Rhodium	103-2	4352827	4370021	4310385	4344411	0.71	cps
Scandium	45-1	4190999	4138558	4114566	4148041	0.94	cps
Scandium	45-2	372866	371431	372829	372375	0.22	cps
Selenium	82-1	11958	12203	12327	12163	1.54	cps
Selenium	77-2	1140	1083	1067	1097	3.50	cps
Selenium	78-2	4031	3911	3891	3944	1.92	cps
Silicon	28-1	8014265	8134506	8110602	8086458	0.79	cps
Silver	107-1	365222	370723	370220	368721	0.82	cps
Silver	109-1	345029	347500	351927	348152	1.00	cps
Sodium	23-2	75169	76020	75882	75690	0.60	cps
Strontium	86-1	76149	77332	78635	77372	1.61	cps
Strontium	88-1	668566	675094	677302	673654	0.67	cps
Sulfur	34-1	115321	114673	112776	114257	1.16	cps
Terbium	159-1	11002515	10908179	10924027	10944907	0.46	cps
Terbium	159-2	6509516	6450050	6428156	6462574	0.65	cps
Thallium	203-1	318873	323498	321619	321330	0.72	cps
Thallium	205-1	762342	771497	771233	768357	0.68	cps
Tin	118-1	5394	5805	5474	5558	3.91	cps
Titanium	47-1	291650	292941	293576	292722	0.34	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4519-03MSDLX4 Instrumnet Name : P7
Client Sample ID : MYE4K2SRE Dilution Factor : 4
Date & Time Acquired : 2024-11-03 00:11:39 DataFile Name : 127SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	7899	8216	7919	8011	2.21	cps
Vanadium	51-2	1316865	1290118	1301065	1302683	1.03	cps
Yttrium	89-1	13166049	13087082	12946771	13066634	0.85	cps
Yttrium	89-2	3133358	3143303	3136635	3137765	0.16	cps
Zinc	66-2	267511	265393	263802	265569	0.70	cps
Zirconium	90-1	19900	20361	20481	20247	1.51	cps
Zirconium	91-1	4721	4681	4561	4654	1.79	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BL Instrumnet Name : P7
Client Sample ID : PBS452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:14:44 DataFile Name : 128CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1117	1320	1403	1280	11.52	cps
Antimony	121-1	507	537	487	510	4.93	cps
Arsenic	75-2	17	7	13	12	41.65	cps
Barium	135-1	547	420	513	493	13.31	cps
Barium	137-1	777	763	863	801	6.78	cps
Beryllium	9-1	44	39	50	44	12.50	cps
Bismuth	209-1	6838199	5640322	6669622	6382714	10.16	cps
Bismuth	209-2	4824617	4734959	4747092	4768889	1.02	cps
Bromine	81-1	11291	10731	10864	10962	2.67	cps
Bromine	81-2	70	97	83	83	16.00	cps
Cadmium	108-1	17	7	23	16	53.90	cps
Cadmium	106-1	2474	2027	2257	2252	9.92	cps
Cadmium	111-1	1727	1441	1593	1587	9.01	cps
Calcium	43-1	360	373	447	393	11.86	cps
Calcium	44-1	12032	11084	11685	11600	4.13	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	680	690	713	694	2.46	cps
Cobalt	59-2	223	207	177	202	11.69	cps
Copper	63-2	1040	1137	1160	1112	5.72	cps
Dysprosium	156-1	30	30	30	30	0.00	cps
Dysprosium	156-2	7	13	7	9	43.25	cps
Erbium	164-1	73	70	97	80	18.17	cps
Erbium	164-2	33	40	47	40	16.68	cps
Gadolinium	160-1	57	37	47	47	21.43	cps
Gadolinium	160-2	243	270	220	244	10.24	cps
Holmium	165-1	10687938	8625969	10545766	9953224	11.57	cps
Holmium	165-2	6530221	6421619	6541869	6497903	1.02	cps
Indium	115-1	8190855	6562452	8068842	7607383	11.92	cps
Indium	115-2	2554494	2502291	2548273	2535019	1.12	cps
Iron	56-2	74729	79594	81669	78664	4.53	cps
Iron	57-2	2617	2974	2844	2811	6.42	cps
Iron	54-2	5071	5238	5368	5225	2.85	cps
Krypton	83-1	173	200	183	186	7.26	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BL Instrumnet Name : P7
Client Sample ID : PBS452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:14:44 DataFile Name : 128CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	720	627	750	699	9.20	cps
Lead	207-1	580	593	600	591	1.72	cps
Lead	208-1	2630	2570	2687	2629	2.22	cps
Lithium	6-1	1264079	1027636	1229345	1173687	10.88	cps
Magnesium	24-2	1670	1653	1703	1676	1.52	cps
Manganese	55-2	1183	1493	1520	1399	13.38	cps
Molybdenum	94-1	247	220	327	264	20.99	cps
Molybdenum	95-1	153	133	163	150	10.18	cps
Molybdenum	96-1	200	210	220	210	4.76	cps
Molybdenum	97-1	77	120	87	94	24.02	cps
Molybdenum	98-1	210	247	243	233	8.69	cps
Neodymium	150-1	27	23	27	26	7.55	cps
Neodymium	150-2	13	7	13	11	34.61	cps
Nickel	60-2	183	213	183	193	8.96	cps
Phosphorus	31-2	97	120	127	114	13.76	cps
Potassium	39-2	43121	43184	42165	42823	1.33	cps
Rhodium	103-1	7492732	6059008	7427255	6992998	11.58	cps
Rhodium	103-2	4397830	4394022	4341808	4377887	0.72	cps
Scandium	45-1	4156062	3360950	4146939	3887983	11.74	cps
Scandium	45-2	370197	369777	368394	369456	0.26	cps
Selenium	82-1	386	312	364	354	10.66	cps
Selenium	77-2	0	0	0	0	0.00	cps
Selenium	78-2	197	167	200	188	9.78	cps
Silicon	28-1	186195	189384	191617	189065	1.44	cps
Silver	107-1	380	330	270	327	16.86	cps
Silver	109-1	470	273	283	342	32.37	cps
Sodium	23-2	14921	15168	14997	15029	0.84	cps
Strontium	86-1	510	583	520	538	7.39	cps
Strontium	88-1	2107	2144	2360	2204	6.21	cps
Sulfur	34-1	108407	108549	109207	108721	0.39	cps
Terbium	159-1	10948381	8932294	10831324	10237333	11.05	cps
Terbium	159-2	6508709	6492166	6475342	6492072	0.26	cps
Thallium	203-1	167	190	173	177	6.80	cps
Thallium	205-1	360	413	420	398	8.27	cps
Tin	118-1	1627	1677	1810	1705	5.57	cps
Titanium	47-1	330	290	297	306	7.01	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BL Instrumnet Name : P7
Client Sample ID : PBS452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:14:44 DataFile Name : 128CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	67	63	70	67	5.00	cps
Vanadium	51-2	200	140	160	167	18.33	cps
Yttrium	89-1	12937658	10488218	12797445	12074440	11.39	cps
Yttrium	89-2	3154355	3131125	3071166	3118882	1.38	cps
Zinc	66-2	560	643	657	620	8.45	cps
Zirconium	90-1	523	463	473	487	6.61	cps
Zirconium	91-1	93	93	80	89	8.66	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BS Instrumnet Name : P7
Client Sample ID : LCS452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:17:59 DataFile Name : 129LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8689	8662	8846	8733	1.13	cps
Antimony	121-1	36941	37118	36901	36986	0.31	cps
Arsenic	75-2	877	1047	1033	986	9.59	cps
Barium	135-1	46894	46977	46713	46861	0.29	cps
Barium	137-1	81652	82048	81864	81855	0.24	cps
Beryllium	9-1	2411	2550	2600	2521	3.88	cps
Bismuth	209-1	6839385	6721048	6990813	6850415	1.97	cps
Bismuth	209-2	4750318	4709860	4731400	4730526	0.43	cps
Bromine	81-1	10627	10367	11338	10777	4.66	cps
Bromine	81-2	60	67	77	68	12.38	cps
Cadmium	108-1	310	393	327	343	12.84	cps
Cadmium	106-1	2967	2914	2800	2894	2.94	cps
Cadmium	111-1	6107	5925	6326	6119	3.28	cps
Calcium	43-1	24126	23929	24233	24096	0.64	cps
Calcium	44-1	395521	398944	398774	397746	0.49	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	16132	15705	16372	16070	2.10	cps
Cobalt	59-2	12242	12078	11705	12008	2.29	cps
Copper	63-2	17831	18255	17751	17945	1.51	cps
Dysprosium	156-1	13	23	20	19	26.96	cps
Dysprosium	156-2	10	13	10	11	17.30	cps
Erbium	164-1	87	63	70	73	16.39	cps
Erbium	164-2	33	50	53	46	23.52	cps
Gadolinium	160-1	47	33	63	48	31.46	cps
Gadolinium	160-2	280	267	223	257	11.54	cps
Holmium	165-1	10560768	10551370	10960595	10690911	2.19	cps
Holmium	165-2	6495946	6364754	6311225	6390642	1.49	cps
Indium	115-1	8042909	7940813	8295930	8093217	2.26	cps
Indium	115-2	2485144	2484233	2503748	2491042	0.44	cps
Iron	56-2	1409532	1390282	1377679	1392498	1.15	cps
Iron	57-2	36588	36314	36207	36369	0.54	cps
Iron	54-2	81311	78245	79961	79839	1.92	cps
Krypton	83-1	173	177	213	188	11.82	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BS Instrumnet Name : P7
Client Sample ID : LCS452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:17:59 DataFile Name : 129LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	14364	15085	14491	14647	2.63	cps
Lead	207-1	13043	13056	12439	12846	2.74	cps
Lead	208-1	58643	59354	57588	58529	1.52	cps
Lithium	6-1	1256246	1233937	1262517	1250900	1.20	cps
Magnesium	24-2	363882	360590	360305	361592	0.55	cps
Manganese	55-2	6962	6441	6261	6555	5.55	cps
Molybdenum	94-1	38557	39099	38277	38644	1.08	cps
Molybdenum	95-1	47065	46413	46209	46563	0.96	cps
Molybdenum	96-1	51055	51644	51517	51405	0.60	cps
Molybdenum	97-1	29156	29463	29510	29376	0.65	cps
Molybdenum	98-1	74250	74532	75155	74646	0.62	cps
Neodymium	150-1	30	13	27	23	37.81	cps
Neodymium	150-2	13	7	10	10	33.30	cps
Nickel	60-2	3244	3077	3214	3178	2.80	cps
Phosphorus	31-2	127	137	103	122	14.00	cps
Potassium	39-2	473807	466723	466293	468941	0.90	cps
Rhodium	103-1	7387831	7381918	7671481	7480410	2.21	cps
Rhodium	103-2	4372845	4327509	4356704	4352353	0.53	cps
Scandium	45-1	4115155	4102346	4292993	4170165	2.56	cps
Scandium	45-2	373358	367028	366676	369020	1.02	cps
Selenium	82-1	1861	1810	1781	1817	2.25	cps
Selenium	77-2	147	133	183	154	16.77	cps
Selenium	78-2	660	593	583	612	6.81	cps
Silicon	28-1	271346	275155	276714	274405	1.01	cps
Silver	107-1	22010	22638	22501	22383	1.47	cps
Silver	109-1	20491	21019	21416	20975	2.21	cps
Sodium	23-2	592812	583778	584986	587192	0.84	cps
Strontium	86-1	6211	6768	6515	6498	4.29	cps
Strontium	88-1	53714	54755	54487	54319	0.99	cps
Sulfur	34-1	117094	120669	124687	120817	3.14	cps
Terbium	159-1	10796249	10673485	11236307	10902014	2.71	cps
Terbium	159-2	6472864	6371374	6393043	6412427	0.83	cps
Thallium	203-1	16847	17331	16797	16992	1.74	cps
Thallium	205-1	41095	41817	41707	41540	0.94	cps
Tin	118-1	1687	1913	1780	1793	6.35	cps
Titanium	47-1	11564	11371	11578	11504	1.01	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

PB164452BS

Instrumnet Name :

P7

Client Sample ID :

LCS452

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:17:59

DataFile Name :

129LCSE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	53001	52761	52690	52817	0.31	cps
Vanadium	51-2	32979	33293	33156	33143	0.48	cps
Yttrium	89-1	12782349	12567619	13279250	12876406	2.83	cps
Yttrium	89-2	3122615	3061793	3081176	3088528	1.01	cps
Zinc	66-2	7202	7025	7375	7201	2.43	cps
Zirconium	90-1	32229	32559	32583	32457	0.61	cps
Zirconium	91-1	7375	7482	7135	7331	2.42	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BSD Instrumnet Name : P7
Client Sample ID : LCSD452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:21:14 DataFile Name : 130LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	5671	5781	6025	5826	3.11	cps
Antimony	121-1	18068	18235	18295	18199	0.65	cps
Arsenic	75-2	510	427	573	503	14.61	cps
Barium	135-1	23075	23630	23206	23304	1.24	cps
Barium	137-1	40116	41123	40798	40679	1.26	cps
Beryllium	9-1	1322	1261	1356	1313	3.65	cps
Bismuth	209-1	6686625	6743837	6658758	6696406	0.65	cps
Bismuth	209-2	4652398	4386254	4590170	4542940	3.06	cps
Bromine	81-1	10727	10694	10594	10672	0.65	cps
Bromine	81-2	43	73	57	58	26.01	cps
Cadmium	108-1	197	187	140	174	17.34	cps
Cadmium	106-1	2694	2687	2600	2660	1.96	cps
Cadmium	111-1	3828	3954	4032	3938	2.62	cps
Calcium	43-1	11951	12215	12338	12168	1.62	cps
Calcium	44-1	200946	205313	205715	203991	1.30	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	8316	8449	8746	8504	2.59	cps
Cobalt	59-2	6021	6135	5981	6046	1.32	cps
Copper	63-2	9580	9320	9166	9355	2.24	cps
Dysprosium	156-1	23	33	43	33	30.00	cps
Dysprosium	156-2	10	13	30	18	60.28	cps
Erbium	164-1	60	60	77	66	14.68	cps
Erbium	164-2	27	40	47	38	26.95	cps
Gadolinium	160-1	43	57	37	46	22.36	cps
Gadolinium	160-2	197	323	240	253	25.41	cps
Holmium	165-1	10518849	10355498	10356165	10410171	0.90	cps
Holmium	165-2	6256362	5980978	6363191	6200177	3.18	cps
Indium	115-1	8066972	7976627	8125590	8056396	0.93	cps
Indium	115-2	2532078	2360710	2531509	2474765	3.99	cps
Iron	56-2	754627	755813	768012	759484	0.98	cps
Iron	57-2	20150	19987	19723	19953	1.08	cps
Iron	54-2	42954	42055	44148	43052	2.44	cps
Krypton	83-1	157	233	183	191	20.37	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BSD Instrumnet Name : P7
Client Sample ID : LCSD452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:21:14 DataFile Name : 130LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	7119	7209	7659	7329	3.95	cps
Lead	207-1	6668	6568	6425	6554	1.87	cps
Lead	208-1	29199	29239	29813	29417	1.17	cps
Lithium	6-1	1239556	1215959	1202045	1219186	1.56	cps
Magnesium	24-2	182734	180640	180579	181318	0.68	cps
Manganese	55-2	3977	3844	4501	4107	8.45	cps
Molybdenum	94-1	18839	19293	19096	19076	1.19	cps
Molybdenum	95-1	22143	23425	23896	23155	3.92	cps
Molybdenum	96-1	25713	26117	26344	26058	1.23	cps
Molybdenum	97-1	14120	14531	14848	14499	2.52	cps
Molybdenum	98-1	37124	37612	37712	37482	0.84	cps
Neodymium	150-1	23	27	20	23	14.29	cps
Neodymium	150-2	13	10	13	12	15.73	cps
Nickel	60-2	1610	1693	1777	1693	4.92	cps
Phosphorus	31-2	103	147	147	132	18.92	cps
Potassium	39-2	257645	257171	256025	256947	0.32	cps
Rhodium	103-1	7476591	7374883	7462537	7438004	0.74	cps
Rhodium	103-2	4281331	4156305	4289208	4242282	1.76	cps
Scandium	45-1	4166530	4194334	4194797	4185220	0.39	cps
Scandium	45-2	374315	353305	371733	366451	3.13	cps
Selenium	82-1	1111	1088	1043	1081	3.23	cps
Selenium	77-2	77	77	83	79	4.87	cps
Selenium	78-2	507	467	487	487	4.11	cps
Silicon	28-1	231598	233503	236444	233849	1.04	cps
Silver	107-1	11281	11668	11261	11403	2.01	cps
Silver	109-1	10417	10431	10928	10592	2.75	cps
Sodium	23-2	302631	301575	298939	301049	0.63	cps
Strontium	86-1	3387	3354	3384	3375	0.54	cps
Strontium	88-1	27436	27850	27553	27613	0.77	cps
Sulfur	34-1	126182	126703	129194	127360	1.26	cps
Terbium	159-1	10730962	10788401	10639480	10719614	0.70	cps
Terbium	159-2	6250732	6063093	6301279	6205035	2.02	cps
Thallium	203-1	8393	8353	8473	8406	0.73	cps
Thallium	205-1	20105	19888	20566	20186	1.71	cps
Tin	118-1	1557	1650	1717	1641	4.90	cps
Titanium	47-1	5761	6228	6138	6042	4.10	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : PB164452BSD Instrumnet Name : P7
Client Sample ID : LCSD452 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:21:14 DataFile Name : 130LLCC.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	25922	26019	25658	25866	0.72	cps
Vanadium	51-2	16479	16472	16796	16583	1.12	cps
Yttrium	89-1	12829788	12759608	12784696	12791364	0.28	cps
Yttrium	89-2	3088805	2942182	3081306	3037431	2.72	cps
Zinc	66-2	3797	4017	3887	3901	2.84	cps
Zirconium	90-1	16272	15992	16449	16238	1.42	cps
Zirconium	91-1	3620	3684	3644	3649	0.88	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01 Instrumnet Name : P7
Client Sample ID : MYE4H7 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:24:29 DataFile Name : 131SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	12200366	12192004	12211642	12201337	0.08	cps
Antimony	121-1	25773	26014	25787	25858	0.52	cps
Arsenic	75-2	84033	83245	82957	83411	0.67	cps
Barium	135-1	2146884	2186788	2152473	2162048	1.00	cps
Barium	137-1	3678662	3768152	3725148	3723987	1.20	cps
Beryllium	9-1	2884	3017	3356	3086	7.89	cps
Bismuth	209-1	6708436	6673576	6532096	6638036	1.41	cps
Bismuth	209-2	4608172	4631919	4612459	4617517	0.27	cps
Bromine	81-1	13990	14447	13827	14088	2.28	cps
Bromine	81-2	1000	1053	880	978	9.08	cps
Cadmium	108-1	750	717	830	766	7.61	cps
Cadmium	106-1	3814	3624	3924	3787	4.01	cps
Cadmium	111-1	9510	9724	9884	9706	1.94	cps
Calcium	43-1	287992	286980	286919	287297	0.21	cps
Calcium	44-1	4405756	4449454	4330503	4395237	1.37	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	537277	531426	535600	534768	0.56	cps
Cobalt	59-2	486844	485728	482997	485190	0.41	cps
Copper	63-2	1803586	1808301	1810481	1807456	0.19	cps
Dysprosium	156-1	116599	119086	119298	118328	1.27	cps
Dysprosium	156-2	89704	90019	90459	90061	0.42	cps
Erbium	164-1	90303	88907	90282	89831	0.89	cps
Erbium	164-2	55329	54847	55443	55206	0.57	cps
Gadolinium	160-1	109490	110696	111035	110407	0.74	cps
Gadolinium	160-2	71682	70798	71608	71363	0.69	cps
Holmium	165-1	10454263	10704351	10601163	10586592	1.19	cps
Holmium	165-2	6505325	6414718	6495622	6471888	0.77	cps
Indium	115-1	7866535	7872903	7753514	7830984	0.86	cps
Indium	115-2	2413874	2410654	2384402	2402976	0.67	cps
Iron	56-2	486118686	481961233	481017620	483032513	0.56	cps
Iron	57-2	11955260	11982309	11979886	11972485	0.13	cps
Iron	54-2	26458725	26251202	26378720	26362882	0.40	cps
Krypton	83-1	277	247	300	274	9.74	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01 Instrumnet Name : P7
Client Sample ID : MYE4H7 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:24:29 DataFile Name : 131SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	886132	894161	898929	893074	0.72	cps
Lead	207-1	783673	784212	800011	789299	1.18	cps
Lead	208-1	3873839	3853625	3940886	3889450	1.17	cps
Lithium	6-1	1236682	1233809	1224305	1231599	0.53	cps
Magnesium	24-2	10843889	10952209	10673289	10823129	1.30	cps
Manganese	55-2	11348309	11336779	11253589	11312892	0.46	cps
Molybdenum	94-1	111601	113741	112931	112758	0.96	cps
Molybdenum	95-1	9530	9737	9383	9550	1.86	cps
Molybdenum	96-1	34738	35136	34925	34933	0.57	cps
Molybdenum	97-1	5728	5941	5658	5776	2.56	cps
Molybdenum	98-1	14053	13970	13947	13990	0.40	cps
Neodymium	150-1	223270	227037	226535	225614	0.91	cps
Neodymium	150-2	117504	117760	117444	117570	0.14	cps
Nickel	60-2	219868	220514	218921	219768	0.36	cps
Phosphorus	31-2	29288	29712	29582	29527	0.74	cps
Potassium	39-2	5294414	5217333	5177368	5229705	1.14	cps
Rhodium	103-1	7082127	7099672	6950873	7044224	1.15	cps
Rhodium	103-2	4172745	4164148	4143585	4160159	0.36	cps
Scandium	45-1	4509691	4477501	4390880	4459358	1.38	cps
Scandium	45-2	400549	396728	396647	397975	0.56	cps
Selenium	82-1	618	695	695	670	6.62	cps
Selenium	77-2	350	390	373	371	5.41	cps
Selenium	78-2	573	610	500	561	9.98	cps
Silicon	28-1	39657260	39529006	38943426	39376564	0.97	cps
Silver	107-1	9650	9096	9063	9270	3.56	cps
Silver	109-1	9523	8679	8626	8943	5.63	cps
Sodium	23-2	125362	125496	126488	125782	0.49	cps
Strontium	86-1	319412	322108	319653	320391	0.47	cps
Strontium	88-1	2902998	2893235	2907879	2901370	0.26	cps
Sulfur	34-1	119570	117908	116889	118122	1.15	cps
Terbium	159-1	10773191	10833897	10716409	10774499	0.55	cps
Terbium	159-2	6360821	6391466	6415669	6389319	0.43	cps
Thallium	203-1	62402	64351	64475	63743	1.82	cps
Thallium	205-1	152302	154800	154227	153776	0.85	cps
Tin	118-1	32630	32580	32851	32687	0.44	cps
Titanium	47-1	1652605	1808885	1606748	1689413	6.27	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01 Instrumnet Name : P7
Client Sample ID : MYE4H7 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:24:29 DataFile Name : 131SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	52714	53611	54310	53545	1.49	cps
Vanadium	51-2	940937	939575	939079	939864	0.10	cps
Yttrium	89-1	13945415	13830714	13791068	13855732	0.58	cps
Yttrium	89-2	3410460	3301807	3341698	3351322	1.64	cps
Zinc	66-2	272925	269243	271001	271057	0.68	cps
Zirconium	90-1	289660	291219	290478	290453	0.27	cps
Zirconium	91-1	64846	66275	65070	65397	1.18	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-02 Instrumnet Name : P7
Client Sample ID : MYE4H7D Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:27:33 DataFile Name : 132SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	11870511	11901982	11855368	11875953	0.20	cps
Antimony	121-1	25239	25513	26491	25748	2.56	cps
Arsenic	75-2	81157	80651	81284	81030	0.41	cps
Barium	135-1	2067597	2089635	2145828	2101020	1.92	cps
Barium	137-1	3641738	3605938	3715493	3654390	1.53	cps
Beryllium	9-1	2784	2834	2867	2828	1.48	cps
Bismuth	209-1	6848355	6722547	5958351	6509751	7.40	cps
Bismuth	209-2	4620118	4683054	4733273	4678815	1.21	cps
Bromine	81-1	13229	13863	13963	13685	2.91	cps
Bromine	81-2	933	997	1003	978	3.95	cps
Cadmium	108-1	627	737	867	743	16.16	cps
Cadmium	106-1	3741	3794	3460	3665	4.89	cps
Cadmium	111-1	9301	9215	9398	9305	0.98	cps
Calcium	43-1	271244	275199	276276	274239	0.97	cps
Calcium	44-1	4129558	4158113	4242714	4176795	1.41	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	517782	519305	519411	518833	0.18	cps
Cobalt	59-2	474728	471902	473360	473330	0.30	cps
Copper	63-2	1753961	1770541	1754612	1759704	0.53	cps
Dysprosium	156-1	115217	118125	118245	117196	1.46	cps
Dysprosium	156-2	86212	87426	87916	87185	1.01	cps
Erbium	164-1	87213	89732	89256	88734	1.51	cps
Erbium	164-2	53435	53114	53830	53460	0.67	cps
Gadolinium	160-1	106061	108139	110233	108144	1.93	cps
Gadolinium	160-2	71156	70215	70067	70479	0.84	cps
Holmium	165-1	10724768	10855006	9413061	10330945	7.72	cps
Holmium	165-2	6558625	6689149	6609237	6619004	0.99	cps
Indium	115-1	7860043	7923847	6873484	7552458	7.80	cps
Indium	115-2	2475246	2430323	2470626	2458731	1.01	cps
Iron	56-2	467103206	474070313	472530780	471234766	0.78	cps
Iron	57-2	11697496	11728486	11634274	11686752	0.41	cps
Iron	54-2	25693449	25667036	25823932	25728139	0.33	cps
Krypton	83-1	260	300	220	260	15.39	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-02 Instrumnet Name : P7
Client Sample ID : MYE4H7D Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:27:33 DataFile Name : 132SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	881463	884363	897140	887655	0.94	cps
Lead	207-1	774328	782383	790910	782540	1.06	cps
Lead	208-1	3839728	3822360	3896437	3852842	1.01	cps
Lithium	6-1	1241861	1233327	1106637	1193942	6.34	cps
Magnesium	24-2	10552554	10624958	10535455	10570989	0.45	cps
Manganese	55-2	10971583	11152989	11093728	11072767	0.84	cps
Molybdenum	94-1	107182	108746	110361	108763	1.46	cps
Molybdenum	95-1	8202	8956	9096	8752	5.49	cps
Molybdenum	96-1	32803	34477	34020	33767	2.56	cps
Molybdenum	97-1	5498	5424	5464	5462	0.67	cps
Molybdenum	98-1	13243	13757	13760	13586	2.19	cps
Neodymium	150-1	218358	223138	224855	222117	1.52	cps
Neodymium	150-2	113405	113782	114725	113971	0.60	cps
Nickel	60-2	213954	215345	212784	214028	0.60	cps
Phosphorus	31-2	28881	29158	29034	29024	0.48	cps
Potassium	39-2	4994187	5081566	5059850	5045201	0.90	cps
Rhodium	103-1	7034825	7101479	6216211	6784172	7.27	cps
Rhodium	103-2	4236538	4222994	4234556	4231363	0.17	cps
Scandium	45-1	4374064	4372144	3911447	4219218	6.32	cps
Scandium	45-2	402748	401511	403427	402562	0.24	cps
Selenium	82-1	619	583	685	629	8.24	cps
Selenium	77-2	390	410	370	390	5.13	cps
Selenium	78-2	513	537	557	536	4.05	cps
Silicon	28-1	38249975	38875352	38948149	38691159	0.99	cps
Silver	107-1	8419	8903	8969	8764	3.43	cps
Silver	109-1	7762	7989	8309	8020	3.43	cps
Sodium	23-2	123776	126421	124794	124997	1.07	cps
Strontium	86-1	302854	308070	312748	307891	1.61	cps
Strontium	88-1	2759756	2808370	2837542	2801889	1.40	cps
Sulfur	34-1	115569	116685	118008	116754	1.05	cps
Terbium	159-1	10960961	11058677	9610350	10543329	7.68	cps
Terbium	159-2	6583761	6577827	6574860	6578816	0.07	cps
Thallium	203-1	56286	58157	57758	57400	1.72	cps
Thallium	205-1	135046	136510	136387	135981	0.60	cps
Tin	118-1	31003	32039	32607	31883	2.55	cps
Titanium	47-1	1546628	1570887	1575600	1564372	0.99	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4525-02

Instrumnet Name :

P7

Client Sample ID :

MYE4H7D

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:27:33

DataFile Name :

132SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	53155	53781	54160	53699	0.94	cps
Vanadium	51-2	920538	915764	920122	918808	0.29	cps
Yttrium	89-1	13632319	13786731	12236265	13218438	6.46	cps
Yttrium	89-2	3309013	3356256	3318239	3327836	0.75	cps
Zinc	66-2	261662	265148	263222	263344	0.66	cps
Zirconium	90-1	275000	282967	294450	284139	3.44	cps
Zirconium	91-1	61464	62717	64695	62959	2.59	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01LX5 Instrumnet Name : P7
Client Sample ID : MYE4H7L Dilution Factor : 5
Date & Time Acquired : 2024-11-03 00:30:35 DataFile Name : 133SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	2414608	2396523	2393105	2401412	0.48	cps
Antimony	121-1	5228	5164	5068	5153	1.56	cps
Arsenic	75-2	16463	16439	16753	16552	1.06	cps
Barium	135-1	391156	392760	394922	392946	0.48	cps
Barium	137-1	681312	681334	684812	682486	0.30	cps
Beryllium	9-1	606	650	656	637	4.30	cps
Bismuth	209-1	6998784	6996162	7093250	7029399	0.79	cps
Bismuth	209-2	4834232	4793274	4869904	4832470	0.79	cps
Bromine	81-1	10787	10394	10260	10480	2.61	cps
Bromine	81-2	213	207	280	233	17.38	cps
Cadmium	108-1	130	137	163	143	12.31	cps
Cadmium	106-1	2794	2654	2774	2740	2.76	cps
Cadmium	111-1	3428	3475	3357	3420	1.74	cps
Calcium	43-1	55765	55554	56474	55931	0.86	cps
Calcium	44-1	907971	914807	914485	912421	0.42	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	105257	104881	104731	104956	0.26	cps
Cobalt	59-2	95820	97005	97081	96635	0.73	cps
Copper	63-2	346236	347113	344142	345830	0.44	cps
Dysprosium	156-1	23880	23603	23657	23713	0.62	cps
Dysprosium	156-2	17374	18002	17828	17735	1.83	cps
Erbium	164-1	17404	17972	17761	17712	1.62	cps
Erbium	164-2	10864	11108	10834	10935	1.37	cps
Gadolinium	160-1	21306	21707	21734	21582	1.11	cps
Gadolinium	160-2	14394	14261	14337	14331	0.47	cps
Holmium	165-1	10872029	10885011	10803802	10853614	0.40	cps
Holmium	165-2	6571416	6610591	6707755	6629921	1.06	cps
Indium	115-1	8105568	8043645	8027396	8058869	0.51	cps
Indium	115-2	2487080	2460761	2555349	2501063	1.95	cps
Iron	56-2	95558462	95949825	94369509	95292599	0.86	cps
Iron	57-2	2363813	2400798	2346577	2370396	1.17	cps
Iron	54-2	5231492	5195057	5154895	5193814	0.74	cps
Krypton	83-1	187	207	203	199	5.39	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01LX5 Instrumnet Name : P7
Client Sample ID : MYE4H7L Dilution Factor : 5
Date & Time Acquired : 2024-11-03 00:30:35 DataFile Name : 133SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	184395	183264	183929	183863	0.31	cps
Lead	207-1	158918	161420	165315	161884	1.99	cps
Lead	208-1	728901	729061	736415	731459	0.59	cps
Lithium	6-1	1234091	1255562	1294001	1261218	2.41	cps
Magnesium	24-2	2130514	2145440	2113480	2129812	0.75	cps
Manganese	55-2	2257553	2227264	2186451	2223756	1.60	cps
Molybdenum	94-1	22330	22170	22424	22308	0.58	cps
Molybdenum	95-1	1893	1960	1923	1926	1.73	cps
Molybdenum	96-1	6945	6862	7032	6946	1.22	cps
Molybdenum	97-1	1233	1190	1163	1196	2.96	cps
Molybdenum	98-1	2900	2937	3004	2947	1.78	cps
Neodymium	150-1	44664	45202	45339	45069	0.79	cps
Neodymium	150-2	23570	22915	22538	23008	2.27	cps
Nickel	60-2	45094	43891	43961	44315	1.52	cps
Phosphorus	31-2	5648	5568	5908	5708	3.12	cps
Potassium	39-2	1105245	1096350	1102527	1101374	0.41	cps
Rhodium	103-1	7343881	7315774	7357811	7339155	0.29	cps
Rhodium	103-2	4313105	4250610	4430358	4331358	2.11	cps
Scandium	45-1	4102739	4151577	4093672	4115996	0.76	cps
Scandium	45-2	371608	375182	366974	371254	1.11	cps
Selenium	82-1	468	483	425	459	6.58	cps
Selenium	77-2	57	70	60	62	11.15	cps
Selenium	78-2	280	283	267	277	3.19	cps
Silicon	28-1	7838149	7771849	7751918	7787306	0.58	cps
Silver	107-1	1957	1933	1793	1895	4.66	cps
Silver	109-1	1733	1817	1860	1803	3.57	cps
Sodium	23-2	37579	37613	37228	37473	0.57	cps
Strontium	86-1	62843	63435	62378	62885	0.84	cps
Strontium	88-1	545258	552144	551108	549503	0.68	cps
Sulfur	34-1	107834	107162	106243	107080	0.75	cps
Terbium	159-1	10912310	10982817	11115101	11003409	0.94	cps
Terbium	159-2	6624263	6531767	6554994	6570341	0.73	cps
Thallium	203-1	12943	12950	13123	13005	0.79	cps
Thallium	205-1	31319	31466	31253	31346	0.35	cps
Tin	118-1	7472	7419	7872	7588	3.27	cps
Titanium	47-1	313033	314582	315957	314524	0.47	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4525-01LX5

Instrumnet Name :

P7

Client Sample ID :

MYE4H7L

Dilution Factor :

5

Date & Time Acquired :

2024-11-03 00:30:35

DataFile Name :

133SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	11128	11175	10674	10992	2.51	cps
Vanadium	51-2	184267	182128	181673	182689	0.76	cps
Yttrium	89-1	12995677	12957971	12829250	12927632	0.68	cps
Yttrium	89-2	3117888	3120186	3163191	3133755	0.81	cps
Zinc	66-2	56004	56219	55914	56046	0.28	cps
Zirconium	90-1	58012	57348	56940	57433	0.94	cps
Zirconium	91-1	12746	12892	12656	12765	0.94	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-03 Instrumnet Name : P7
Client Sample ID : MYE4H7S Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:33:45 DataFile Name : 134SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8928355	8865931	8888659	8894315	0.36	cps
Antimony	121-1	578828	584380	586274	583161	0.66	cps
Arsenic	75-2	71588	71009	71159	71252	0.42	cps
Barium	135-1	4710756	4772075	4742061	4741631	0.65	cps
Barium	137-1	8186264	8271941	8225749	8227985	0.52	cps
Beryllium	9-1	42447	42241	42642	42443	0.47	cps
Bismuth	209-1	6960324	6979242	6932789	6957452	0.34	cps
Bismuth	209-2	4776437	4807822	4773752	4786004	0.40	cps
Bromine	81-1	12455	12919	13349	12908	3.46	cps
Bromine	81-2	713	783	710	736	5.63	cps
Cadmium	108-1	5171	5454	4958	5194	4.80	cps
Cadmium	106-1	9690	9873	9760	9774	0.95	cps
Cadmium	111-1	77479	78250	79785	78505	1.50	cps
Calcium	43-1	202161	204836	204278	203758	0.69	cps
Calcium	44-1	3080999	3083971	3078932	3081301	0.08	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	844908	840983	840636	842175	0.28	cps
Cobalt	59-2	2188549	2152785	2177145	2172826	0.84	cps
Copper	63-2	1946305	1947441	1922025	1938590	0.74	cps
Dysprosium	156-1	87131	86568	86917	86872	0.33	cps
Dysprosium	156-2	63456	65726	65180	64787	1.83	cps
Erbium	164-1	64920	66387	66363	65890	1.28	cps
Erbium	164-2	40308	39753	40127	40063	0.71	cps
Gadolinium	160-1	80234	79993	81147	80458	0.76	cps
Gadolinium	160-2	52297	52548	51946	52264	0.58	cps
Holmium	165-1	11046646	10942047	10810570	10933088	1.08	cps
Holmium	165-2	6667094	6615623	6624030	6635582	0.42	cps
Indium	115-1	7981039	8044211	7885072	7970107	1.01	cps
Indium	115-2	2469645	2442908	2449536	2454030	0.57	cps
Iron	56-2	348391808	347177035	348457595	348008813	0.21	cps
Iron	57-2	8658650	8624992	8611666	8631769	0.28	cps
Iron	54-2	19081620	18934328	18822862	18946270	0.69	cps
Krypton	83-1	260	303	223	262	15.27	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-03 Instrumnet Name : P7
Client Sample ID : MYE4H7S Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:33:45 DataFile Name : 134SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	743316	748484	758166	749989	1.01	cps
Lead	207-1	655299	665699	676653	665883	1.60	cps
Lead	208-1	2965173	3284200	3311180	3186851	6.04	cps
Lithium	6-1	1250058	1265059	1253960	1256359	0.62	cps
Magnesium	24-2	7670370	7648378	7766067	7694938	0.81	cps
Manganese	55-2	8912338	8866510	8850569	8876472	0.36	cps
Molybdenum	94-1	79440	82092	80332	80621	1.67	cps
Molybdenum	95-1	6502	6615	6818	6645	2.42	cps
Molybdenum	96-1	24644	25292	24898	24945	1.31	cps
Molybdenum	97-1	4127	4031	4237	4132	2.50	cps
Molybdenum	98-1	10040	10274	10424	10246	1.89	cps
Neodymium	150-1	163870	164606	166891	165122	0.95	cps
Neodymium	150-2	85625	85088	83938	84884	1.02	cps
Nickel	60-2	599294	597469	596831	597865	0.21	cps
Phosphorus	31-2	21031	20884	20811	20909	0.54	cps
Potassium	39-2	3707558	3663221	3714415	3695065	0.75	cps
Rhodium	103-1	7177203	7211574	7184502	7191093	0.25	cps
Rhodium	103-2	4293770	4205800	4260330	4253300	1.04	cps
Scandium	45-1	4301850	4341958	4278712	4307506	0.74	cps
Scandium	45-2	390658	388001	386669	388443	0.52	cps
Selenium	82-1	8940	8940	9133	9004	1.23	cps
Selenium	77-2	1220	1023	970	1071	12.29	cps
Selenium	78-2	3107	2894	2970	2990	3.61	cps
Silicon	28-1	28115655	28246925	28289159	28217246	0.32	cps
Silver	107-1	300226	304881	304394	303167	0.84	cps
Silver	109-1	287146	288174	289570	288297	0.42	cps
Sodium	23-2	94793	94531	95430	94918	0.49	cps
Strontium	86-1	224919	229694	229583	228065	1.20	cps
Strontium	88-1	2072999	2074531	2075211	2074247	0.05	cps
Sulfur	34-1	115277	116487	116016	115927	0.53	cps
Terbium	159-1	11299848	11146478	11109487	11185271	0.90	cps
Terbium	159-2	6620845	6536706	6536987	6564846	0.74	cps
Thallium	203-1	314781	315630	318814	316408	0.67	cps
Thallium	205-1	747769	762049	774152	761323	1.73	cps
Tin	118-1	24240	24040	24982	24421	2.03	cps
Titanium	47-1	1146150	1151378	1156181	1151236	0.44	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

P4525-03

Instrumnet Name :

P7

Client Sample ID :

MYE4H7S

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:33:45

DataFile Name :

134SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	39876	39646	40424	39982	1.00	cps
Vanadium	51-2	1648546	1643810	1666792	1653049	0.73	cps
Yttrium	89-1	13656772	13635286	13526780	13606279	0.51	cps
Yttrium	89-2	3303323	3268632	3251401	3274452	0.81	cps
Zinc	66-2	389505	389232	389681	389472	0.06	cps
Zirconium	90-1	205825	208860	208871	207852	0.84	cps
Zirconium	91-1	46353	46339	47062	46585	0.89	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01MS Instrumnet Name : P7
Client Sample ID : MYE4H7SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:36:46 DataFile Name : 135SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	9308679	9415785	9267239	9330568	0.82	cps
Antimony	121-1	2991741	3020595	3025750	3012696	0.61	cps
Arsenic	75-2	110386	111148	110427	110654	0.39	cps
Barium	135-1	16596411	16788621	16841906	16742312	0.77	cps
Barium	137-1	29159202	29061549	28862026	29027592	0.52	cps
Beryllium	9-1	194611	195715	194308	194878	0.38	cps
Bismuth	209-1	6887981	7018264	6967254	6957833	0.94	cps
Bismuth	209-2	4651586	4748820	4636997	4679134	1.30	cps
Bromine	81-1	12579	13063	12559	12733	2.24	cps
Bromine	81-2	737	670	743	717	5.66	cps
Cadmium	108-1	23526	24180	23890	23865	1.37	cps
Cadmium	106-1	35192	35386	34992	35190	0.56	cps
Cadmium	111-1	358585	363642	362296	361507	0.72	cps
Calcium	43-1	190047	193456	192842	192115	0.95	cps
Calcium	44-1	2913392	2943609	2914999	2924000	0.58	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2544528	2513348	2519977	2525951	0.65	cps
Cobalt	59-2	8890930	8844409	8866592	8867310	0.26	cps
Copper	63-2	4181986	4204638	4148149	4178258	0.68	cps
Dysprosium	156-1	83285	84452	85746	84494	1.46	cps
Dysprosium	156-2	64303	63473	63138	63638	0.94	cps
Erbium	164-1	62968	64920	64193	64027	1.54	cps
Erbium	164-2	38199	38700	39402	38767	1.56	cps
Gadolinium	160-1	76765	78173	78146	77695	1.04	cps
Gadolinium	160-2	50333	50377	50735	50482	0.44	cps
Holmium	165-1	10809750	10873268	10603508	10762175	1.31	cps
Holmium	165-2	6544557	6568709	6436374	6516547	1.08	cps
Indium	115-1	7837971	7827956	7794573	7820167	0.29	cps
Indium	115-2	2429796	2430007	2381748	2413850	1.15	cps
Iron	56-2	338192042	339035582	336294468	337840697	0.42	cps
Iron	57-2	8386809	8373560	8355527	8371965	0.19	cps
Iron	54-2	18484723	18525555	18417725	18476001	0.29	cps
Krypton	83-1	240	230	280	250	10.59	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01MS Instrumnet Name : P7
Client Sample ID : MYE4H7SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:36:46 DataFile Name : 135SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1082710	1104101	1097704	1094838	1.00	cps
Lead	207-1	944673	963928	967117	958573	1.27	cps
Lead	208-1	4635584	4779693	4798021	4737766	1.88	cps
Lithium	6-1	1275051	1250892	1238132	1254692	1.49	cps
Magnesium	24-2	7263180	7305043	7225156	7264459	0.55	cps
Manganese	55-2	11547916	11665108	11599795	11604273	0.51	cps
Molybdenum	94-1	75838	78063	77346	77082	1.47	cps
Molybdenum	95-1	6515	6435	6602	6517	1.28	cps
Molybdenum	96-1	23903	24357	24678	24313	1.60	cps
Molybdenum	97-1	3981	3984	3961	3975	0.32	cps
Molybdenum	98-1	9990	10347	9863	10067	2.49	cps
Neodymium	150-1	158302	159394	160916	159537	0.82	cps
Neodymium	150-2	81594	82178	81802	81858	0.36	cps
Nickel	60-2	2355947	2362103	2345229	2354426	0.36	cps
Phosphorus	31-2	20016	19990	19813	19940	0.55	cps
Potassium	39-2	3494657	3511879	3460887	3489141	0.74	cps
Rhodium	103-1	6988542	7075302	6966119	7009988	0.82	cps
Rhodium	103-2	4216818	4164165	4138111	4173031	0.96	cps
Scandium	45-1	4206722	4171828	4141726	4173425	0.78	cps
Scandium	45-2	377213	379266	372047	376175	0.99	cps
Selenium	82-1	45113	46252	46059	45808	1.33	cps
Selenium	77-2	4537	4331	4474	4447	2.38	cps
Selenium	78-2	13927	14270	13820	14006	1.68	cps
Silicon	28-1	26565900	26879418	26835970	26760429	0.63	cps
Silver	107-1	1411659	1422218	1431923	1421933	0.71	cps
Silver	109-1	1339396	1349109	1352922	1347142	0.52	cps
Sodium	23-2	91171	90792	90085	90683	0.61	cps
Strontium	86-1	216282	218264	218406	217651	0.55	cps
Strontium	88-1	1974494	1985067	1978102	1979221	0.27	cps
Sulfur	34-1	111298	112181	112027	111835	0.42	cps
Terbium	159-1	11116040	10983869	11010833	11036914	0.63	cps
Terbium	159-2	6561867	6541625	6539447	6547646	0.19	cps
Thallium	203-1	1402987	1425271	1432306	1420188	1.08	cps
Thallium	205-1	3985327	3959573	3942641	3962514	0.54	cps
Tin	118-1	22988	23566	23142	23232	1.29	cps
Titanium	47-1	1082145	1092924	1088272	1087780	0.50	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4525-01MS Instrumnet Name : P7
Client Sample ID : MYE4H7SRE Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:36:46 DataFile Name : 135SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	38753	39151	39515	39140	0.97	cps
Vanadium	51-2	5338324	5294119	5280537	5304327	0.57	cps
Yttrium	89-1	13165087	13232024	13143829	13180313	0.35	cps
Yttrium	89-2	3137858	3214673	3184905	3179145	1.22	cps
Zinc	66-2	1105697	1107896	1102251	1105281	0.26	cps
Zirconium	90-1	194950	199255	199334	197846	1.27	cps
Zirconium	91-1	44333	44798	45179	44770	0.95	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4526-01 Instrumnet Name : P7
Client Sample ID : MYE4M3 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:39:44 DataFile Name : 136SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8997521	8972587	8939550	8969886	0.32	cps
Antimony	121-1	27453	27503	27346	27434	0.29	cps
Arsenic	75-2	48011	49071	48415	48499	1.10	cps
Barium	135-1	1235159	1238591	1248000	1240584	0.54	cps
Barium	137-1	2327630	2339496	2325327	2330818	0.33	cps
Beryllium	9-1	2311	2078	2217	2202	5.33	cps
Bismuth	209-1	6895036	7040346	6980157	6971846	1.05	cps
Bismuth	209-2	4716728	4720828	4711191	4716249	0.10	cps
Bromine	81-1	13193	13179	13593	13322	1.76	cps
Bromine	81-2	753	763	653	723	8.41	cps
Cadmium	108-1	547	450	520	506	9.88	cps
Cadmium	106-1	3400	3417	3434	3417	0.49	cps
Cadmium	111-1	5428	5321	5367	5372	1.00	cps
Calcium	43-1	250844	253062	251493	251800	0.45	cps
Calcium	44-1	3822251	3862545	3868218	3851005	0.65	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	384963	383233	383453	383883	0.25	cps
Cobalt	59-2	279891	281573	278813	280092	0.50	cps
Copper	63-2	877931	885521	879426	880959	0.46	cps
Dysprosium	156-1	85139	86692	87004	86278	1.16	cps
Dysprosium	156-2	65009	65391	64504	64968	0.69	cps
Erbium	164-1	67482	67904	67656	67681	0.31	cps
Erbium	164-2	41923	40956	41545	41475	1.17	cps
Gadolinium	160-1	80600	79671	82440	80904	1.74	cps
Gadolinium	160-2	52250	52454	52321	52342	0.20	cps
Holmium	165-1	10686241	10804489	10824810	10771846	0.69	cps
Holmium	165-2	6648107	6497090	6601402	6582200	1.17	cps
Indium	115-1	7693915	7920275	7896235	7836808	1.59	cps
Indium	115-2	2462340	2421516	2420374	2434743	0.98	cps
Iron	56-2	309688029	309745842	308953649	309462507	0.14	cps
Iron	57-2	7659532	7650914	7620661	7643702	0.27	cps
Iron	54-2	16821759	16868129	16832834	16840907	0.14	cps
Krypton	83-1	200	247	317	254	23.08	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4526-01 Instrumnet Name : P7
Client Sample ID : MYE4M3 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:39:44 DataFile Name : 136SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	465927	472356	469825	469370	0.69	cps
Lead	207-1	405999	409442	409556	408332	0.49	cps
Lead	208-1	1838803	1855801	1863862	1852822	0.69	cps
Lithium	6-1	1215755	1209461	1214335	1213184	0.27	cps
Magnesium	24-2	6603369	6637059	6642238	6627555	0.32	cps
Manganese	55-2	6493189	6472187	6432551	6465976	0.48	cps
Molybdenum	94-1	113848	114728	115262	114613	0.62	cps
Molybdenum	95-1	6982	7212	7088	7094	1.62	cps
Molybdenum	96-1	30822	30645	30993	30820	0.56	cps
Molybdenum	97-1	4254	4597	4361	4404	3.99	cps
Molybdenum	98-1	10921	11144	11121	11062	1.11	cps
Neodymium	150-1	163533	166059	166912	165501	1.06	cps
Neodymium	150-2	84401	85189	83399	84329	1.06	cps
Nickel	60-2	142846	142631	142664	142714	0.08	cps
Phosphorus	31-2	23645	24039	24036	23907	0.95	cps
Potassium	39-2	4496607	4471421	4462594	4476874	0.39	cps
Rhodium	103-1	6930170	7075811	7083478	7029819	1.23	cps
Rhodium	103-2	4220923	4206379	4243539	4223614	0.44	cps
Scandium	45-1	4114861	4218438	4233966	4189088	1.55	cps
Scandium	45-2	382700	383571	383000	383091	0.12	cps
Selenium	82-1	546	567	488	534	7.62	cps
Selenium	77-2	303	337	283	308	8.75	cps
Selenium	78-2	450	457	370	426	11.33	cps
Silicon	28-1	18173662	18113281	18332606	18206517	0.62	cps
Silver	107-1	6578	5878	5648	6035	8.03	cps
Silver	109-1	6045	5775	5171	5663	7.90	cps
Sodium	23-2	121874	123000	121582	122152	0.61	cps
Strontium	86-1	273035	275697	280190	276307	1.31	cps
Strontium	88-1	2498301	2548726	2524386	2523804	1.00	cps
Sulfur	34-1	122651	123568	125133	123784	1.01	cps
Terbium	159-1	11015246	11056423	10941565	11004411	0.53	cps
Terbium	159-2	6529251	6502118	6469916	6500428	0.46	cps
Thallium	203-1	4888	4597	4417	4634	5.12	cps
Thallium	205-1	10955	10944	11081	10993	0.69	cps
Tin	118-1	27757	28749	28732	28412	2.00	cps
Titanium	47-1	1421712	1430013	1445775	1432500	0.85	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4526-01 Instrumnet Name : P7
Client Sample ID : MYE4M3 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:39:44 DataFile Name : 136SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	44885	45915	45006	45269	1.24	cps
Vanadium	51-2	575705	577870	577005	576860	0.19	cps
Yttrium	89-1	13238430	13366662	13340709	13315267	0.51	cps
Yttrium	89-2	3262275	3226394	3224324	3237664	0.66	cps
Zinc	66-2	183630	185095	183489	184071	0.48	cps
Zirconium	90-1	294011	299665	299961	297879	1.13	cps
Zirconium	91-1	66744	67699	67209	67217	0.71	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4526-02 Instrumnet Name : P7
Client Sample ID : MYE4M3D Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:42:48 DataFile Name : 137SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	9052579	9037599	9031512	9040563	0.12	cps
Antimony	121-1	26682	27279	26932	26964	1.11	cps
Arsenic	75-2	49803	49582	49950	49779	0.37	cps
Barium	135-1	1266911	1282057	1289841	1279603	0.91	cps
Barium	137-1	2361049	2410967	2430755	2400924	1.50	cps
Beryllium	9-1	2117	2117	2489	2241	9.59	cps
Bismuth	209-1	7084800	7010605	6915411	7003606	1.21	cps
Bismuth	209-2	4777977	4710172	4682846	4723665	1.04	cps
Bromine	81-1	12752	13526	13660	13313	3.68	cps
Bromine	81-2	797	777	770	781	1.78	cps
Cadmium	108-1	507	583	603	564	9.04	cps
Cadmium	106-1	3774	3681	3444	3633	4.68	cps
Cadmium	111-1	5794	5598	5739	5710	1.77	cps
Calcium	43-1	255575	256861	258190	256875	0.51	cps
Calcium	44-1	3882726	3947929	3953265	3927973	1.00	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	390394	392641	392309	391781	0.31	cps
Cobalt	59-2	282362	283963	283415	283247	0.29	cps
Copper	63-2	897921	891452	894318	894564	0.36	cps
Dysprosium	156-1	86383	87842	86957	87061	0.84	cps
Dysprosium	156-2	66162	66312	66125	66199	0.15	cps
Erbium	164-1	68343	67843	70125	68770	1.74	cps
Erbium	164-2	41802	41635	42020	41819	0.46	cps
Gadolinium	160-1	81756	82468	82876	82367	0.69	cps
Gadolinium	160-2	53070	53047	53923	53347	0.94	cps
Holmium	165-1	11017480	11000952	10972644	10997025	0.21	cps
Holmium	165-2	6658191	6642960	6567146	6622766	0.74	cps
Indium	115-1	8065513	7907564	7920543	7964540	1.10	cps
Indium	115-2	2410902	2452508	2442093	2435168	0.89	cps
Iron	56-2	315646502	315049802	315565869	315420724	0.10	cps
Iron	57-2	7820417	7828406	7851085	7833302	0.20	cps
Iron	54-2	17114822	17137733	17150984	17134513	0.11	cps
Krypton	83-1	220	267	207	231	13.63	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4526-02 Instrumnet Name : P7
Client Sample ID : MYE4M3D Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:42:48 DataFile Name : 137SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	465291	477328	476912	473177	1.44	cps
Lead	207-1	404284	410089	414118	409497	1.21	cps
Lead	208-1	1857879	1884276	1881813	1874656	0.78	cps
Lithium	6-1	1242816	1221039	1250320	1238058	1.23	cps
Magnesium	24-2	6699600	6746238	6769805	6738548	0.53	cps
Manganese	55-2	6641181	6704435	6677010	6674208	0.48	cps
Molybdenum	94-1	114476	116542	117983	116334	1.52	cps
Molybdenum	95-1	7162	7302	7339	7267	1.28	cps
Molybdenum	96-1	31323	30996	31434	31251	0.73	cps
Molybdenum	97-1	4614	4607	4751	4657	1.74	cps
Molybdenum	98-1	11324	11391	11381	11366	0.32	cps
Neodymium	150-1	166653	165515	168025	166731	0.75	cps
Neodymium	150-2	85024	86352	84270	85215	1.24	cps
Nickel	60-2	143993	145185	144292	144490	0.43	cps
Phosphorus	31-2	23812	24693	24139	24215	1.84	cps
Potassium	39-2	4456412	4541916	4536666	4511665	1.06	cps
Rhodium	103-1	7228557	7070407	7157571	7152178	1.11	cps
Rhodium	103-2	4263137	4230819	4246013	4246656	0.38	cps
Scandium	45-1	4283968	4276470	4255393	4271944	0.35	cps
Scandium	45-2	383643	385618	390004	386422	0.84	cps
Selenium	82-1	575	538	632	582	8.08	cps
Selenium	77-2	253	307	330	297	13.25	cps
Selenium	78-2	407	363	450	407	10.66	cps
Silicon	28-1	18060500	18372282	18559903	18330895	1.38	cps
Silver	107-1	5561	5588	5544	5564	0.39	cps
Silver	109-1	5261	5398	5104	5254	2.79	cps
Sodium	23-2	122234	122240	123399	122624	0.55	cps
Strontium	86-1	277833	283706	283425	281655	1.18	cps
Strontium	88-1	2559328	2586688	2560469	2568828	0.60	cps
Sulfur	34-1	127177	127453	124848	126492	1.13	cps
Terbium	159-1	11221811	11120190	11188726	11176909	0.46	cps
Terbium	159-2	6600979	6514042	6533157	6549392	0.70	cps
Thallium	203-1	4587	4631	4327	4515	3.63	cps
Thallium	205-1	11291	10991	11215	11166	1.40	cps
Tin	118-1	27857	28682	28902	28480	1.93	cps
Titanium	47-1	1447757	1462874	1463694	1458108	0.62	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : P4526-02 Instrumnet Name : P7
Client Sample ID : MYE4M3D Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:42:48 DataFile Name : 137SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	44906	45681	45875	45487	1.13	cps
Vanadium	51-2	586170	584920	584212	585101	0.17	cps
Yttrium	89-1	13649420	13597498	13513135	13586684	0.51	cps
Yttrium	89-2	3290330	3217183	3269415	3258976	1.16	cps
Zinc	66-2	185453	187464	185794	186237	0.58	cps
Zirconium	90-1	301807	305067	306772	304549	0.83	cps
Zirconium	91-1	67323	68181	69855	68453	1.88	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV005 Instrumnet Name : P7
Client Sample ID : CCV005 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:45:54 DataFile Name : 138CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	10091006	10094519	10040007	10075177	0.30	cps
Antimony	121-1	4901124	4964753	5029620	4965165	1.29	cps
Arsenic	75-2	241097	238814	241500	240470	0.60	cps
Barium	135-1	6690246	6767521	6760280	6739349	0.63	cps
Barium	137-1	11563437	11678207	11696301	11645982	0.62	cps
Beryllium	9-1	654373	663419	673514	663769	1.44	cps
Bismuth	209-1	6459004	6459712	6551534	6490083	0.82	cps
Bismuth	209-2	3941814	4515392	4461489	4306232	7.36	cps
Bromine	81-1	10304	9813	10110	10076	2.45	cps
Bromine	81-2	67	73	57	66	12.79	cps
Cadmium	108-1	87566	89085	89450	88700	1.13	cps
Cadmium	106-1	126112	128125	128741	127659	1.08	cps
Cadmium	111-1	1196918	1200888	1210349	1202718	0.57	cps
Calcium	43-1	5965173	5991244	6044087	6000168	0.67	cps
Calcium	44-1	94978689	95066459	95192032	95079060	0.11	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1877908	1880345	1902565	1886939	0.72	cps
Cobalt	59-2	3087162	3064617	3050827	3067535	0.60	cps
Copper	63-2	21436974	21356336	21456436	21416582	0.25	cps
Dysprosium	156-1	333	337	300	323	6.27	cps
Dysprosium	156-2	403	363	460	409	11.88	cps
Erbium	164-1	367	363	340	357	4.07	cps
Erbium	164-2	277	230	240	249	9.87	cps
Gadolinium	160-1	267	353	257	292	18.19	cps
Gadolinium	160-2	420	353	397	390	8.67	cps
Holmium	165-1	10685116	10741886	10740361	10722454	0.30	cps
Holmium	165-2	5922817	6647077	6518612	6362835	6.07	cps
Indium	115-1	7493264	7569121	7637988	7566791	0.96	cps
Indium	115-2	2080016	2378764	2320471	2259750	7.01	cps
Iron	56-2	409866274	406141874	406153647	407387265	0.53	cps
Iron	57-2	10321653	10190133	10252588	10254791	0.64	cps
Iron	54-2	22508364	22225387	22456137	22396629	0.67	cps
Krypton	83-1	197	247	203	216	12.59	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCV005 Instrumnet Name : P7
Client Sample ID : CCV005 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:45:54 DataFile Name : 138CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	20760323	20713469	21273985	20915926	1.49	cps
Lead	207-1	18313813	18467173	18611696	18464228	0.81	cps
Lead	208-1	82924596	83384168	84682075	83663613	1.09	cps
Lithium	6-1	1214575	1218113	1247998	1226896	1.50	cps
Magnesium	24-2	85225874	85357044	84664495	85082471	0.43	cps
Manganese	55-2	13873443	13616896	13819460	13769933	0.98	cps
Molybdenum	94-1	18272954	18441745	18436978	18383892	0.52	cps
Molybdenum	95-1	26757365	26775080	26695720	26742722	0.16	cps
Molybdenum	96-1	29133711	29025819	29337842	29165791	0.54	cps
Molybdenum	97-1	16837350	16866431	16985583	16896455	0.46	cps
Molybdenum	98-1	43232264	43476016	43621046	43443109	0.45	cps
Neodymium	150-1	467	420	510	466	9.67	cps
Neodymium	150-2	153	157	137	149	7.20	cps
Nickel	60-2	730083	728494	729410	729329	0.11	cps
Phosphorus	31-2	104409	104701	105396	104836	0.48	cps
Potassium	39-2	51140374	50931416	51061176	51044322	0.21	cps
Rhodium	103-1	6687312	6615307	6651748	6651456	0.54	cps
Rhodium	103-2	3507438	4016121	3968153	3830571	7.33	cps
Scandium	45-1	4005355	3918497	3907874	3943909	1.36	cps
Scandium	45-2	324208	366838	368427	353158	7.10	cps
Selenium	82-1	71860	72020	72908	72262	0.78	cps
Selenium	77-2	6632	6575	6558	6588	0.58	cps
Selenium	78-2	22657	21920	22317	22298	1.66	cps
Silicon	28-1	41512926	41700524	41702865	41638772	0.26	cps
Silver	107-1	6100849	6134714	6239037	6158200	1.17	cps
Silver	109-1	5866383	5771840	5835430	5824551	0.83	cps
Sodium	23-2	142721748	143401461	142846808	142990006	0.25	cps
Strontium	86-1	37761595	38707306	38221686	38230196	1.24	cps
Strontium	88-1	335770535	337566642	337597068	336978082	0.31	cps
Sulfur	34-1	493954	492065	490143	492054	0.39	cps
Terbium	159-1	10834047	10859246	11086072	10926455	1.27	cps
Terbium	159-2	5850200	6585864	6520051	6318705	6.44	cps
Thallium	203-1	5283878	5189800	5327710	5267129	1.34	cps
Thallium	205-1	12457454	12533856	12678194	12556501	0.89	cps
Tin	118-1	4110618	4159518	4264331	4178156	1.88	cps
Titanium	47-1	6075961	6095838	6102483	6091428	0.23	cps

LB Number :

LB133265

Operator :

Jaswal

Lab Sample ID :

CCV005

Instrumnet Name :

P7

Client Sample ID :

CCV005

Dilution Factor :

1

Date & Time Acquired :

2024-11-03 00:45:54

DataFile Name :

138CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	17116231	17167422	17384443	17222699	0.83	cps
Vanadium	51-2	1682510	1670259	1697552	1683440	0.81	cps
Yttrium	89-1	12417331	12441572	12380733	12413212	0.25	cps
Yttrium	89-2	2711446	3085186	3043600	2946744	6.95	cps
Zinc	66-2	3357154	3400457	3405601	3387737	0.79	cps
Zirconium	90-1	8371689	8419870	8375177	8388912	0.32	cps
Zirconium	91-1	1906694	1869335	1924431	1900153	1.48	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB005 Instrumnet Name : P7
Client Sample ID : CCB005 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:48:35 DataFile Name : 139CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1367	1533	1650	1517	9.39	cps
Antimony	121-1	3300	3360	3280	3314	1.26	cps
Arsenic	75-2	37	17	27	27	37.50	cps
Barium	135-1	727	703	617	682	8.50	cps
Barium	137-1	1390	1177	1093	1220	12.54	cps
Beryllium	9-1	111	106	94	104	8.18	cps
Bismuth	209-1	7289340	7201014	7100700	7197018	1.31	cps
Bismuth	209-2	4907574	4994110	4927470	4943051	0.92	cps
Bromine	81-1	10544	10120	10214	10293	2.16	cps
Bromine	81-2	63	87	77	76	15.50	cps
Cadmium	108-1	10	13	13	12	15.73	cps
Cadmium	106-1	2524	2540	2570	2545	0.93	cps
Cadmium	111-1	1866	1876	1903	1882	1.02	cps
Calcium	43-1	520	490	420	477	10.77	cps
Calcium	44-1	14808	13567	13210	13861	6.05	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	733	833	677	748	10.61	cps
Cobalt	59-2	293	293	297	294	0.65	cps
Copper	63-2	1670	1747	1720	1712	2.27	cps
Dysprosium	156-1	17	7	27	17	59.99	cps
Dysprosium	156-2	13	7	13	11	34.61	cps
Erbium	164-1	60	57	90	69	26.65	cps
Erbium	164-2	87	63	60	70	20.76	cps
Gadolinium	160-1	47	50	37	44	15.61	cps
Gadolinium	160-2	243	277	217	246	12.24	cps
Holmium	165-1	11019577	10980497	10847991	10949355	0.82	cps
Holmium	165-2	6758024	6775543	6655027	6729531	0.97	cps
Indium	115-1	8326973	8334054	8245018	8302015	0.60	cps
Indium	115-2	2574818	2589972	2571202	2578664	0.39	cps
Iron	56-2	62539	70528	73623	68897	8.30	cps
Iron	57-2	2264	2484	2504	2417	5.51	cps
Iron	54-2	4397	4407	5051	4619	8.11	cps
Krypton	83-1	167	130	183	160	17.05	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB005 Instrumnet Name : P7
Client Sample ID : CCB005 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:48:35 DataFile Name : 139CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1910	1777	1557	1748	10.21	cps
Lead	207-1	1660	1463	1330	1485	11.18	cps
Lead	208-1	7554	6804	6341	6900	8.88	cps
Lithium	6-1	1284371	1353828	1304251	1314150	2.72	cps
Magnesium	24-2	3664	4054	4511	4076	10.40	cps
Manganese	55-2	1450	1703	1670	1608	8.56	cps
Molybdenum	94-1	1590	1463	1610	1555	5.12	cps
Molybdenum	95-1	1737	1547	1520	1601	7.38	cps
Molybdenum	96-1	1817	1817	1767	1800	1.60	cps
Molybdenum	97-1	1043	1073	937	1018	7.06	cps
Molybdenum	98-1	2717	2674	2507	2632	4.21	cps
Neodymium	150-1	17	23	20	20	16.65	cps
Neodymium	150-2	10	13	23	16	44.60	cps
Nickel	60-2	200	177	173	183	7.92	cps
Phosphorus	31-2	130	100	127	119	13.83	cps
Potassium	39-2	46186	46604	46925	46572	0.80	cps
Rhodium	103-1	7478188	7499680	7470240	7482703	0.20	cps
Rhodium	103-2	4431744	4443236	4461097	4445359	0.33	cps
Scandium	45-1	4187565	4183207	4159523	4176765	0.36	cps
Scandium	45-2	376099	371743	372544	373462	0.62	cps
Selenium	82-1	414	440	388	414	6.29	cps
Selenium	77-2	3	0	0	1	173.21	cps
Selenium	78-2	207	260	230	232	11.51	cps
Silicon	28-1	204469	204679	201979	203709	0.74	cps
Silver	107-1	883	750	657	763	14.92	cps
Silver	109-1	787	597	627	670	15.24	cps
Sodium	23-2	27248	26960	26486	26898	1.43	cps
Strontium	86-1	1830	1643	1670	1715	5.89	cps
Strontium	88-1	14084	12232	11034	12450	12.34	cps
Sulfur	34-1	154967	153606	154812	154462	0.48	cps
Terbium	159-1	11270744	11387217	11184577	11280846	0.90	cps
Terbium	159-2	6610217	6632562	6648147	6630309	0.29	cps
Thallium	203-1	977	887	687	850	17.46	cps
Thallium	205-1	2277	1990	1733	2000	13.59	cps
Tin	118-1	2360	2244	2440	2348	4.21	cps
Titanium	47-1	587	557	580	574	2.74	cps

LB Number : LB133265 Operator : Jaswal
Lab Sample ID : CCB005 Instrumnet Name : P7
Client Sample ID : CCB005 Dilution Factor : 1
Date & Time Acquired : 2024-11-03 00:48:35 DataFile Name : 139CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	553	613	453	540	14.97	cps
Vanadium	51-2	183	153	157	164	10.00	cps
Yttrium	89-1	13011001	12968061	12880487	12953183	0.51	cps
Yttrium	89-2	3144183	3149090	3123316	3138863	0.44	cps
Zinc	66-2	650	640	760	683	9.74	cps
Zirconium	90-1	1883	1667	1460	1670	12.68	cps
Zirconium	91-1	423	373	297	364	17.51	cps