

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133264

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:23:58 PM
Supervise By	Mohan Bera	Supervise On	11/7/2024 12:56:17 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP83050,MP83049		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE002	TUNE	11/03/24 14:07		Jaswal	OK
2	S00	S0	CAL1	11/03/24 14:49		Jaswal	OK
3	S02	S02	CAL3	11/03/24 14:55		Jaswal	OK
4	S03	S03	CAL4	11/03/24 14:59		Jaswal	OK
5	S04	S04	CAL5	11/03/24 15:02		Jaswal	OK
6	S05	S05	CAL6	11/03/24 15:05		Jaswal	OK
7	S06	S06	CAL7	11/03/24 15:07		Jaswal	OK
8	S07	S07	CAL8	11/03/24 15:10		Jaswal	OK
9	S08	S08	CAL9	11/03/24 15:13		Jaswal	OK
10	ICV002	ICV002	ICV	11/03/24 15:23		Jaswal	OK
11	ICB002	ICB002	ICB	11/03/24 15:31		Jaswal	OK
12	ICSA002	ICSA002	ICSA	11/03/24 15:40		Jaswal	OK
13	ICSAB002	ICSAB002	ICSAB	11/03/24 15:43		Jaswal	OK
14	CCV006	CCV006	CCV	11/03/24 15:46		Jaswal	OK
15	CCB006	CCB006	CCB	11/03/24 15:48		Jaswal	OK
16	PB164585BL	PBS585	MB	11/03/24 15:52		Jaswal	OK
17	PB164585BS	LCS585	LCS	11/03/24 15:55		Jaswal	OK
18	PB164585BSD	LCSD585	LCSD	11/03/24 15:58		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133264

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:23:58 PM
Supervise By	Mohan Bera	Supervise On	11/7/2024 12:56:17 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP83050,MP83049

19	PB164596BL	PBS596	MB	11/03/24 16:01		Jaswal	OK
20	PB164596BS	LCS596	LCS	11/03/24 16:05		Jaswal	OK
21	PB164596BSD	LCSD596	LCSD	11/03/24 16:08		Jaswal	OK
22	P4522-01RE	MYE4R1	SAM	11/03/24 16:11	Not use	Jaswal	Not Ok
23	P4522-02	MYE4R1D	DUP	11/03/24 16:14	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
24	P4522-01L	MYE4R1L	SD	11/03/24 16:17	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
25	P4522-03	MYE4R1S	MS	11/03/24 16:21	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
26	P4522-01MS	MYE4R1SRE	MS	11/03/24 16:24	MS Fail For Ag (Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
27	P4522-01	MYE4R1	SAM	11/03/24 16:27	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
28	P4522-04	MYE4R2	SAM	11/03/24 16:30		Jaswal	OK
29	P4522-05	MYE4R3	SAM	11/03/24 16:34		Jaswal	OK
30	P4522-06	MYE4R4	SAM	11/03/24 16:37		Jaswal	OK
31	P4522-07	MYE4R5	SAM	11/03/24 16:40		Jaswal	OK
32	P4522-08	MYE4R6	SAM	11/03/24 16:43		Jaswal	OK
33	P4522-09	MYE4R7	SAM	11/03/24 16:46		Jaswal	OK
34	P4522-10	MYE4R8	SAM	11/03/24 16:49		Jaswal	OK
35	P4522-11	MYE4S0	SAM	11/03/24 16:52		Jaswal	OK
36	P4522-12	MYE4S1	SAM	11/03/24 16:55		Jaswal	OK
37	P4522-13	MYE4S2	SAM	11/03/24 16:58		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133264

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:23:58 PM
Supervise By	Mohan Bera	Supervise On	11/7/2024 12:56:17 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP83050,MP83049		

38	P4522-14	MYE4S3	SAM	11/03/24 17:01		Jaswal	OK
39	P4522-15	MYE4S4	SAM	11/03/24 17:04		Jaswal	OK
40	P4522-16	MYE4S5	SAM	11/03/24 17:08		Jaswal	OK
41	P4522-17	MYE4S6	SAM	11/03/24 17:11		Jaswal	OK
42	P4522-18	MYE4S7	SAM	11/03/24 17:14		Jaswal	OK
43	P4522-19	MYE4S8	SAM	11/03/24 17:17		Jaswal	OK
44	P4522-20	MYE4S9	SAM	11/03/24 17:20		Jaswal	OK
45	P4522-21	MYE4T0	SAM	11/03/24 17:23		Jaswal	OK
46	CCV007	CCV007	CCV	11/03/24 17:26		Jaswal	OK
47	CCB007	CCB007	CCB	11/03/24 17:29		Jaswal	OK
48	P4522-22	MYE4T1	SAM	11/03/24 17:32		Jaswal	OK
49	P4522-01DL	MYE4R1	SAM	11/03/24 17:35	4X for Ba,Co,Ni,V	Jaswal	Confirms
50	P4522-02DL	MYE4R1D	DUP	11/03/24 17:38	4X for Ba,Co,Ni,V	Jaswal	Confirms
51	P4522-01LDL	MYE4R1L	SD	11/03/24 17:41	20X for Ba,Co,Ni,V	Jaswal	Confirms
52	P4522-03DL	MYE4R1S	MS	11/03/24 17:45	4X for Ba,Co,Ni,V	Jaswal	Confirms
53	P4522-01MSDL	MYE4R1SRE	MS	11/03/24 17:48	4X for Ba,Co,Ni,V	Jaswal	Confirms
54	P4523-01	MYE4T2	SAM	11/03/24 17:51		Jaswal	OK
55	P4523-02	MYE4T3	SAM	11/03/24 17:54		Jaswal	OK
56	P4523-03	MYE4T4	SAM	11/03/24 17:57		Jaswal	OK
57	P4523-04	MYE4T5	SAM	11/03/24 18:00		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133264

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:23:58 PM
Supervise By	Mohan Bera	Supervise On	11/7/2024 12:56:17 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP83050,MP83049		

58	P4523-05	MYE4T6	SAM	11/03/24 18:03		Jaswal	OK
59	P4523-06	MYE4T7	SAM	11/03/24 18:06		Jaswal	OK
60	P4523-07	MYE4T8	SAM	11/03/24 18:09	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
61	P4523-08	MYE4T8D	DUP	11/03/24 18:12	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
62	P4523-07L	MYE4T8L	SD	11/03/24 18:15	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
63	P4523-09	MYE4T8S	MS	11/03/24 18:19	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
64	P4523-07MS	MYE4T8SRE	MS	11/03/24 18:22	MS Fail For Ag(Below RL),Pb,Ba,Co,Ni,V High	Jaswal	Dilution
65	P4523-10	MYE4T9	SAM	11/03/24 18:25		Jaswal	OK
66	P4523-11	MYE4W8	SAM	11/03/24 18:28		Jaswal	OK
67	P4523-12	MYE4W9	SAM	11/03/24 18:31		Jaswal	OK
68	P4523-13	MYE4X0	SAM	11/03/24 18:34		Jaswal	OK
69	P4523-14	MYE4X1	SAM	11/03/24 18:37		Jaswal	OK
70	P4523-15	MYE4X2	SAM	11/03/24 18:40		Jaswal	OK
71	P4523-16	MYE4X3	SAM	11/03/24 18:43		Jaswal	OK
72	P4523-17	MYE4X4	SAM	11/03/24 18:46		Jaswal	OK
73	P4523-18	MYE4X5	SAM	11/03/24 18:49		Jaswal	OK
74	CCV008	CCV008	CCV	11/03/24 18:52		Jaswal	OK
75	CCB008	CCB008	CCB	11/03/24 18:55		Jaswal	OK
76	P4523-19	MYE4X6	SAM	11/03/24 18:58		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133264

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:23:58 PM
Supervise By	Mohan Bera	Supervise On	11/7/2024 12:56:17 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP83050,MP83049		

77	P4523-20	MYE4X7	SAM	11/03/24 19:01		Jaswal	OK
78	P4523-21	MYE4X8	SAM	11/03/24 19:05		Jaswal	OK
79	P4523-22	MYE4X9	SAM	11/03/24 19:08		Jaswal	OK
80	PB164452BL	PBS452	MB	11/03/24 19:11		Jaswal	OK
81	PB164452BS	LCS452	LCS	11/03/24 19:14		Jaswal	OK
82	PB164452BSD	LCSD452	LCSD	11/03/24 19:17		Jaswal	OK
83	P4525-01	MYE4H7	SAM	11/03/24 19:21	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
84	P4525-02	MYE4H7D	DUP	11/03/24 19:24	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
85	P4525-01L	MYE4H7L	SD	11/03/24 19:27	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
86	P4525-03	MYE4H7S	MS	11/03/24 19:30	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
87	P4525-01MS	MYE4H7SRE	MS	11/03/24 19:33	MS Fail For Ag (Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
88	P4526-01	MYE4M3	SAM	11/03/24 19:36	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
89	P4526-02	MYE4M3D	DUP	11/03/24 19:40	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
90	P4526-01L	MYE4M3L	SD	11/03/24 19:44	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
91	P4526-03	MYE4M3S	MS	11/03/24 19:47	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
92	P4526-01MS	MYE4M3SRE	MS	11/03/24 19:50	MS Fail For Ag (Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
93	P4527-01	MYE4R9	SAM	11/03/24 19:53	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
94	P4527-02	MYE4R9D	DUP	11/03/24 19:56	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
95	P4527-01L	MYE4R9L	SD	11/03/24 19:59	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133264

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:23:58 PM
Supervise By	Mohan Bera	Supervise On	11/7/2024 12:56:17 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP83050,MP83049		

96	P4527-03	MYE4R9S	MS	11/03/24 20:03	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
97	P4527-01MS	MYE4R9SRE	MS	11/03/24 20:06	MS Fail For As,Pb,Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
98	P4544-01	MYE5G1	SAM	11/03/24 20:09	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
99	P4544-02	MYE5G1D	DUP	11/03/24 20:12	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
100	P4544-01L	MYE5G1L	SD	11/03/24 20:15	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
101	P4544-03	MYE5G1S	MS	11/03/24 20:18	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
102	CCV009	CCV009	CCV	11/03/24 20:25		Jaswal	OK
103	CCB009	CCB009	CCB	11/03/24 20:28		Jaswal	OK
104	P4544-01MS	MYE5G1SRE	MS	11/03/24 20:32	MS Fail For Ag (Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
105	P4525-01DL	MYE4H7	SAM	11/03/24 20:34	4X For Ba,Co,Ni,V	Jaswal	Confirms
106	P4525-02DL	MYE4H7D	DUP	11/03/24 20:38	4X For Ba,Co,Ni,V	Jaswal	Confirms
107	P4525-01LDL	MYE4H7L	SD	11/03/24 20:41	4X For Ba,Co,Ni,V	Jaswal	Confirms
108	P4525-03DL	MYE4H7S	MS	11/03/24 20:44	4X For Ba,Co,Ni,V	Jaswal	Confirms
109	P4525-01MSDL	MYE4H7SRE	MS	11/03/24 20:47	4X For Ba,Co,Ni,V	Jaswal	Confirms
110	P4526-01DL	MYE4M3	SAM	11/03/24 20:50	4X For Ba,Co,Ni,V	Jaswal	Confirms
111	P4526-02DL	MYE4M3D	DUP	11/03/24 20:53	4X For Ba,Co,Ni,V	Jaswal	Confirms
112	P4526-01LDL	MYE4M3L	SD	11/03/24 20:57	4X For Ba,Co,Ni,V	Jaswal	Confirms
113	P4526-03DL	MYE4M3S	MS	11/03/24 21:00	4X For Ba,Co,Ni,V	Jaswal	Confirms
114	P4526-01MSDL	MYE4M3SRE	MS	11/03/24 21:03	4X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133264

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:23:58 PM
Supervise By	Mohan Bera	Supervise On	11/7/2024 12:56:17 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP83050,MP83049

115	P4527-01DL	MYE4R9	SAM	11/03/24 21:06	4X For Ba,Co,Ni,V	Jaswal	Confirms
116	P4527-02DL	MYE4R9D	DUP	11/03/24 21:09	4X For Ba,Co,Ni,V	Jaswal	Confirms
117	P4527-01LDL	MYE4R9L	SD	11/03/24 21:12	4X For Ba,Co,Ni,V	Jaswal	Confirms
118	P4527-03DL	MYE4R9S	MS	11/03/24 21:16	4X For Ba,Co,Ni,V	Jaswal	Confirms
119	P4527-01MSDL	MYE4R9SRE	MS	11/03/24 21:19	4X For Ba,Co,Ni,V	Jaswal	Confirms
120	P4544-01DL	MYE5G1	SAM	11/03/24 21:22	4X For Ba,Co,Ni,V	Jaswal	Confirms
121	P4544-02DL	MYE5G1D	DUP	11/03/24 21:25	4X For Ba,Co,Ni,V	Jaswal	Confirms
122	P4544-01LDL	MYE5G1L	SD	11/03/24 21:28	4X For Ba,Co,Ni,V	Jaswal	Confirms
123	P4544-03DL	MYE5G1S	MS	11/03/24 21:31	4X For Ba,Co,Ni,V	Jaswal	Confirms
124	P4544-01MSDL	MYE5G1SRE	MS	11/03/24 21:35	4X For Ba,Co,Ni,V	Jaswal	Confirms
125	P4523-07DL	MYE4T8	SAM	11/03/24 21:38	4X For Ba,Co,Ni,V	Jaswal	Confirms
126	P4523-08DL	MYE4T8D	DUP	11/03/24 21:41	4X For Ba,Co,Ni,V	Jaswal	Confirms
127	P4523-07LDL	MYE4T8L	SD	11/03/24 21:44	4X For Ba,Co,Ni,V	Jaswal	Confirms
128	P4523-09DL	MYE4T8S	MS	11/03/24 21:47	4X For Ba,Co,Ni,V	Jaswal	Confirms
129	P4523-07MSDL	MYE4T8SRE	MS	11/03/24 21:51	4X For Ba,Co,Ni,V	Jaswal	Confirms
130	CCV010	CCV010	CCV	11/03/24 21:54		Jaswal	OK
131	CCB010	CCB010	CCB	11/03/24 21:58		Jaswal	OK
132	P4522-01A	MYE4R1ARE	PS	11/03/24 23:04	PS For Ag(0.02ml mp83071-10ml sample)	Jaswal	OK
133	P4523-07ARE	MYE4T8ARE	PS	11/03/24 23:07	Not use	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133264

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:23:58 PM
Supervise By	Mohan Bera	Supervise On	11/7/2024 12:56:17 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP83050,MP83049

134	P4525-01A	MYE4H7ARE	PS	11/03/24 23:10	PS For Ag(0.02ml mp83071-10ml sample)	Jaswal	OK
135	P4526-01A	MYE4M3ARE	PS	11/03/24 23:13	PS For Ag(0.02ml mp83071-10ml sample)	Jaswal	OK
136	P4523-07A	MYE4T8ARE	PS	11/03/24 23:16	PS For Ag,Pb(0.02ml mp83071,0.025ml mp83074-10ml sample)	Jaswal	OK
137	P4527-01A	MYE4R9ARE	PS	11/03/24 23:19	PS For As,Pb,Ag(0.0036ml m5801,0.02ml mp83071,0.018ml mp83074ml -10ml sample)	Jaswal	OK
138	P4544-01A	MYE5G1ARE	PS	11/03/24 23:22	PS For Ag(0.02ml mp83071-10ml sample)	Jaswal	OK
139	P4520-08A	MYE4M1ARE	PS	11/03/24 23:30	PS For Pb(0.022ml mp83074-10ml sample)	Jaswal	OK
140	CCV011	CCV011	CCV	11/03/24 23:34		Jaswal	OK
141	CCB011	CCB011	CCB	11/03/24 23:38		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133265

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:12:34 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83035,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	
Chk Standard	MP83049,MP83050

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	11/02/24 17:07		Jaswal	OK
2	S00	S0	CAL1	11/02/24 17:41		Jaswal	OK
3	S02	S02	CAL3	11/02/24 17:48		Jaswal	OK
4	S03	S03	CAL4	11/02/24 17:51		Jaswal	OK
5	S04	S04	CAL5	11/02/24 17:54		Jaswal	OK
6	S05	S05	CAL6	11/02/24 17:57		Jaswal	OK
7	S06	S06	CAL7	11/02/24 18:00		Jaswal	OK
8	S07	S07	CAL8	11/02/24 18:02		Jaswal	OK
9	S08	S08	CAL9	11/02/24 18:05		Jaswal	OK
10	ICV001	ICV001	ICV	11/02/24 18:09		Jaswal	OK
11	ICB001	ICB001	ICB	11/02/24 18:12		Jaswal	OK
12	ICSA001	ICSA001	ICSA	11/02/24 18:16		Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	11/02/24 18:19		Jaswal	OK
14	CCV001	CCV001	CCV	11/02/24 18:22		Jaswal	OK
15	CCB001	CCB001	CCB	11/02/24 18:28		Jaswal	OK
16	PB164454BL	PBS454	MB	11/02/24 18:31		Jaswal	OK
17	PB164454BS	LCS454	LCS	11/02/24 18:34		Jaswal	OK
18	PB164454BSD	LCSD454	LCSD	11/02/24 18:38		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133265

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:12:34 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83035,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard			
Chk Standard	MP83049,MP83050		

19	P4484-01	MYE4H0	SAM	11/02/24 18:41		Jaswal	OK
20	P4484-02	MYE4H1	SAM	11/02/24 18:44		Jaswal	OK
21	P4484-03	MYE4H2	SAM	11/02/24 18:47	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
22	P4484-04	MYE4H2D	DUP	11/02/24 18:50	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
23	P4484-03L	MYE4H2L	SD	11/02/24 18:53	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
24	P4484-03DL	MYE4H2	SAM	11/02/24 19:02	4X For Ba,Co,Ni,V	Jaswal	Confirms
25	P4484-04DL	MYE4H2D	DUP	11/02/24 19:06	4X For Ba,Co,Ni,V	Jaswal	Confirms
26	P4484-03LDL	MYE4H2L	SD	11/02/24 19:09	20X For Ba,Co,Ni,V	Jaswal	Confirms
27	P4484-05	MYE4H2S	MS	11/02/24 19:15	Ba,Co,Ni,V High in MS RE	Jaswal	Dilution
28	P4484-03MS	MYE4H2SRE	MS	11/02/24 19:18	MS Fail For As(412.39ppb) (IN Original) & Ni(373.54ppb),V(355.84pp b) in (4X Dilution),Ba,Co,Ni,V High	Jaswal	Dilution
29	P4484-05DL	MYE4H2S	MS	11/02/24 19:21	4X For Ba,Co,Ni,V	Jaswal	Confirms
30	P4484-03MSDL	MYE4H2SRE	MS	11/02/24 19:24	4X For Ba,Co,Ni,V	Jaswal	Confirms
31	P4484-06	MYE4H3	SAM	11/02/24 19:27		Jaswal	OK
32	P4484-07	MYE4H4	SAM	11/02/24 19:30		Jaswal	OK
33	P4484-08	MYE4H5	SAM	11/02/24 19:34		Jaswal	OK
34	P4484-09	MYE4H6	SAM	11/02/24 19:37		Jaswal	OK
35	P4484-10	MYE4H8	SAM	11/02/24 19:40		Jaswal	OK
36	P4484-11	MYE4H9	SAM	11/02/24 19:43		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133265

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:12:34 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83035,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard			
Chk Standard	MP83049,MP83050		

37	P4484-12	MYE4J0	SAM	11/02/24 19:46		Jaswal	OK
38	P4484-13	MYE4J1	SAM	11/02/24 19:49		Jaswal	OK
39	P4484-14	MYE4J2	SAM	11/02/24 19:52		Jaswal	OK
40	P4484-15	MYE4J3	SAM	11/02/24 19:55		Jaswal	OK
41	P4484-16	MYE4J4	SAM	11/02/24 19:58		Jaswal	OK
42	CCV002	CCV002	CCV	11/02/24 20:01		Jaswal	OK
43	CCB002	CCB002	CCB	11/02/24 20:04		Jaswal	OK
44	P4484-17	MYE4J5	SAM	11/02/24 20:07		Jaswal	OK
45	P4484-18	MYE4J6	SAM	11/02/24 20:10		Jaswal	OK
46	P4484-19	MYE4J7	SAM	11/02/24 20:13		Jaswal	OK
47	P4484-20	MYE4J8	SAM	11/02/24 20:16		Jaswal	OK
48	P4484-21	MYE4J9	SAM	11/02/24 20:20		Jaswal	OK
49	P4484-22	MYE4L3	SAM	11/02/24 20:23		Jaswal	OK
50	P4520-01	MYE4L4	SAM	11/02/24 20:36		Jaswal	OK
51	P4520-02	MYE4L5	SAM	11/02/24 20:39		Jaswal	OK
52	PB164572BL	PBS572	MB	11/02/24 20:48		Jaswal	OK
53	PB164572BS	LCS572	LCS	11/02/24 20:52		Jaswal	OK
54	PB164572BSD	LCSD572	LCSD	11/02/24 20:55		Jaswal	OK
55	P4520-03	MYE4L6	SAM	11/02/24 20:58		Jaswal	OK
56	P4520-04	MYE4L7	SAM	11/02/24 21:01		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133265

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:12:34 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83035,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard			
Chk Standard	MP83049,MP83050		

57	P4520-05	MYE4L8	SAM	11/02/24 21:05		Jaswal	OK
58	P4520-06	MYE4L9	SAM	11/02/24 21:08		Jaswal	OK
59	P4520-07	MYE4M0	SAM	11/02/24 21:11		Jaswal	OK
60	P4520-08	MYE4M1	SAM	11/02/24 21:14	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
61	P4520-09	MYE4M1D	DUP	11/02/24 21:17	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
62	P4520-08L	MYE4M1L	SD	11/02/24 21:20	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
63	P4520-10	MYE4M1S	MS	11/02/24 21:23	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
64	P4520-08MS	MYE4M1SRE	MS	11/02/24 21:26	MS Fail For Pb(107.25ppb),Ba,Co,Ni,V High	Jaswal	Dilution
65	P4520-11	MYE4M2	SAM	11/02/24 21:29		Jaswal	OK
66	P4520-12	MYE4M4	SAM	11/02/24 21:32		Jaswal	OK
67	P4520-13	MYE4M5	SAM	11/02/24 21:36		Jaswal	OK
68	P4520-14	MYE4M6	SAM	11/02/24 21:39		Jaswal	OK
69	P4520-15	MYE4M7	SAM	11/02/24 21:42		Jaswal	OK
70	CCV003	CCV003	CCV	11/02/24 21:45		Jaswal	OK
71	CCB003	CCB003	CCB	11/02/24 21:48		Jaswal	OK
72	P4520-16	MYE4M8	SAM	11/02/24 21:51		Jaswal	OK
73	P4520-17	MYE4M9	SAM	11/02/24 21:54		Jaswal	OK
74	P4520-18	MYE4N0	SAM	11/02/24 21:57		Jaswal	OK
75	P4520-19	MYE4N1	SAM	11/02/24 22:00		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133265

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:12:34 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83035,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard			
Chk Standard	MP83049,MP83050		

76	P4520-20	MYE4N2	SAM	11/02/24 22:03		Jaswal	OK
77	P4520-21	MYE4N3	SAM	11/02/24 22:06		Jaswal	OK
78	P4520-22	MYE4N4	SAM	11/02/24 22:10		Jaswal	OK
79	P4520-08DL	MYE4M1	SAM	11/02/24 22:13	4X For Ba,Co,Ni,V	Jaswal	Confirms
80	P4520-09DL	MYE4M1D	DUP	11/02/24 22:16	4X For Ba,Co,Ni,V	Jaswal	Confirms
81	P4520-08LDL	MYE4M1L	SD	11/02/24 22:19	20X For Ba,Co,Ni,V	Jaswal	Confirms
82	P4520-10DL	MYE4M1S	MS	11/02/24 22:22	4X For Ba,Co,Ni,V	Jaswal	Confirms
83	P4520-08MSDL	MYE4M1SRE	MS	11/02/24 22:25	4X For Ba,Co,Ni,V	Jaswal	Confirms
84	PB164456BL	PBS456	MB	11/02/24 22:31		Jaswal	OK
85	PB164456BS	LCS456	LCS	11/02/24 22:35		Jaswal	OK
86	PB164456BSD	LCSD456	LCSD	11/02/24 22:38		Jaswal	OK
87	P4519-01	MYE4K0	SAM	11/02/24 22:41		Jaswal	OK
88	P4519-02	MYE4K1	SAM	11/02/24 22:44		Jaswal	OK
89	P4519-03	MYE4K2	SAM	11/02/24 22:47	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
90	P4519-04	MYE4K2D	DUP	11/02/24 22:51	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
91	P4519-03L	MYE4K2L	SD	11/02/24 22:54	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
92	P4519-05	MYE4K2S	MS	11/02/24 22:57	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
93	P4519-03MS	MYE4K2SRE	MS	11/02/24 23:00	Ba,Co,Ni,V High	Jaswal	Dilution
94	P4519-06	MYE4K3	SAM	11/02/24 23:03		Jaswal	OK
95	P4519-07	MYE4K4	SAM	11/02/24 23:06		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133265

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:12:34 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83035,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard			
Chk Standard	MP83049,MP83050		

96	P4519-08	MYE4K5	SAM	11/02/24 23:09		Jaswal	OK
97	P4519-09	MYE4K6	SAM	11/02/24 23:12		Jaswal	OK
98	CCV004	CCV004	CCV	11/02/24 23:15		Jaswal	OK
99	CCB004	CCB004	CCB	11/02/24 23:18		Jaswal	OK
100	P4519-10	MYE4K7	SAM	11/02/24 23:21		Jaswal	OK
101	P4519-11	MYE4K8	SAM	11/02/24 23:24		Jaswal	OK
102	P4519-12	MYE4K9	SAM	11/02/24 23:28		Jaswal	OK
103	P4519-13	MYE4L0	SAM	11/02/24 23:31		Jaswal	OK
104	P4519-14	MYE4L1	SAM	11/02/24 23:34		Jaswal	OK
105	P4519-15	MYE4L2	SAM	11/02/24 23:37		Jaswal	OK
106	P4519-16	MYE4P6	SAM	11/02/24 23:40		Jaswal	OK
107	P4519-17	MYE4P7	SAM	11/02/24 23:43		Jaswal	OK
108	P4519-18	MYE4P8	SAM	11/02/24 23:46		Jaswal	OK
109	P4519-19	MYE4P9	SAM	11/02/24 23:49		Jaswal	OK
110	P4519-20	MYE4Q0	SAM	11/02/24 23:52		Jaswal	OK
111	P4519-21	MYE4Q1	SAM	11/02/24 23:55		Jaswal	OK
112	P4519-03DL	MYE4K2	SAM	11/02/24 23:58	4X For Ba,Co,Ni,V	Jaswal	Confirms
113	P4519-04DL	MYE4K2D	DUP	11/03/24 00:02	4X For Ba,Co,Ni,V	Jaswal	Confirms
114	P4519-03LDL	MYE4K2L	SD	11/03/24 00:05	20X For Ba,Co,Ni,V	Jaswal	Confirms
115	P4519-05DL	MYE4K2S	MS	11/03/24 00:08	4X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133265

Review By	Sarabjit Jaswal	Review On	11/5/2024 7:12:34 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83035,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard			
Chk Standard	MP83049,MP83050		

116	P4519-03MSDL	MYE4K2SRE	MS	11/03/24 00:11	4X For Ba,Co,Ni,V	Jaswal	Confirms
117	PB164452BL	PBS452	MB	11/03/24 00:14	NOT USE -	Jaswal	Not Ok
118	PB164452BS	LCS452	LCS	11/03/24 00:17	NOT USE -	Jaswal	Not Ok
119	PB164452BSD	LCSD452	LCSD	11/03/24 00:21	NOT USE -	Jaswal	Not Ok
120	P4525-01	MYE4H7	SAM	11/03/24 00:24	NOT USE - Ba,Co,Ni,V High in MSRE	Jaswal	Not Ok
121	P4525-02	MYE4H7D	DUP	11/03/24 00:27	NOT USE -Ba,Co,Ni,V High in MSRE	Jaswal	Not Ok
122	P4525-01L	MYE4H7L	SD	11/03/24 00:30	NOT USE -Ba,Co,Ni,V High in MSRE	Jaswal	Not Ok
123	P4525-03	MYE4H7S	MS	11/03/24 00:33	NOT USE -Ba,Co,Ni,V High in MSRE	Jaswal	Not Ok
124	P4525-01MS	MYE4H7SRE	MS	11/03/24 00:36	NOT USE -MS Fail For As(158.31ppb),Pb(114.30p pb),Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Not Ok
125	P4526-01	MYE4M3	SAM	11/03/24 00:39	Not Use	Jaswal	Not Ok
126	P4526-02	MYE4M3D	DUP	11/03/24 00:42	Not Use	Jaswal	Not Ok
127	CCV005	CCV005	CCV	11/03/24 00:45		Jaswal	OK
128	CCB005	CCB005	CCB	11/03/24 00:48		Jaswal	OK