

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: _____ SDG No.: MYE4R1
 Instrument ID: P5 Start Date: 11/12/2024
 Analytical Method: ICP-AES Run Batch: LB133417
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.020	0	200	190	-5	38750	36000	-7
Antimony	0	0.020	0	60	54	-10	1250	1000	-20
Arsenic	0	-0	0	10	13	30	1250	1100	-12
Barium	0	-0.010	0	200	210	5	2500	2400	-4
Beryllium	0	-0	0	5	5.4	8	125	120	-4
Cadmium	0	-0	0	5	5.3	6	625	610	-2
Calcium	0	-0.61	0	5000	5500	10	62500	62000	-1
Chromium	0	-0	0	10	11	10	3875	4000	3
Cobalt	0	0.0	0	50	51	2	625	590	-6
Copper	0	0.010	0	25	20	-20	3750	3600	-4
Iron	0	-0.030	0	100	120	20	37500	39000	4
Lead	0	0.0	0	10	10	0	6250	6000	-4
Magnesium	0	0.15	0	5000	5200	4	62500	58000	-7
Manganese	0	-0	0	15	17	13	3750	3700	-1
Nickel	0	-0	0	40	43	8	625	600	-4
Potassium	0	0.77	0	5000	4800	-4	37500	34000	-9
Selenium	0	0.010	0	35	35	0	1250	1100	-12
Silver	0	0.0	0	10	11	10	312.5	300	-4
Sodium	0	-0.37	0	5000	5500	10	37500	36000	-4
Thallium	0	-0	0	25	26	4	1250	1200	-4
Vanadium	0	-0.010	0	50	61	22	625	610	-2
Zinc	0	0.010	0	60	56	-7	3750	3400	-9

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No.: _____ SDG No.: MYE4R1
 Instrument ID: P5 Start Date: 11/12/2024
 Analytical Method: ICP-AES Run Batch: LB133417
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	77500	75000	-3	155000	160000	3	310000	320000	3
Antimony	2500	2200	-12	5000	4700	-6	10000	11000	10
Arsenic	2500	2300	-8	5000	4900	-2	10000	10000	0
Barium	5000	4900	-2	10000	10000	0	20000	20000	0
Beryllium	250	250	0	500	510	2	1000	1000	0
Cadmium	1250	1200	-4	2500	2500	0	5000	5000	0
Calcium	125000	130000	4	250000	260000	4	500000	510000	2
Chromium	7750	7700	-1	15500	16000	3	31000	31000	0
Cobalt	1250	1200	-4	2500	2500	0	5000	5000	0
Copper	7500	7300	-3	15000	15000	0	30000	30000	0
Iron	75000	79000	5	150000	160000	7	300000	310000	3
Lead	12500	12000	-4	25000	25000	0	50000	50000	0
Magnesium	125000	120000	-4	250000	250000	0	500000	510000	2
Manganese	7500	7600	1	15000	15000	0	30000	30000	0
Nickel	1250	1200	-4	2500	2500	0	5000	5000	0
Potassium	75000	71000	-5	150000	150000	0	300000	310000	3
Selenium	2500	2200	-12	5000	4800	-4	10000	11000	10
Silver	625	590	-6	1250	1200	-4	2500	2600	4
Sodium	75000	74000	-1	150000	150000	0	300000	310000	3
Thallium	2500	2500	0	5000	5100	2	10000	9900	-1
Vanadium	1250	1200	-4	2500	2500	0	5000	5000	0
Zinc	7500	6800	-9	15000	14000	-7	30000	32000	7

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51817 MA No.: SDG No.: MYE4R1
Instrument ID: P5 Start Date: 11/12/2024
Analytical Method: ICP-AES Run Batch: LB133417
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	800000	800000	0						
Calcium	800000	780000	-3						
Iron	800000	780000	-3						
Magnesium	800000	790000	-1						
Sodium	800000	790000	-1						