

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE4T2
 Analytical Method: ICP-MS ICSA Source: EPA-0803
 Instrument ID: P7 ICSB Source: EPA-0803
 Run Batch: LB133264
 Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	100000	100000	89000	89	87000	87
Antimony	1.5	22	1.0	67	19	85
Arsenic	0.1	19	0.38	380	19	102
Barium	1.2	22	1.3	107	18	84
Beryllium	0	19	0.34		20	106
Cadmium	0.7	20	0.15	21	18	89
Calcium	100000	100000	98000	98	96000	96
Chromium	21	40	20	94	39	97
Cobalt	1	20	1.2	121	20	99
Copper	8	25	7.0	88	25	98
Iron	100000	100000	100000	100	98000	98
Lead	4	25	3.6	91	22	86
Magnesium	100000	100000	97000	97	95000	95
Manganese	7	27	7.8	111	27	99
Nickel	6	24	5.5	91	24	101
Potassium	100000	100000	96000	96	94000	94
Selenium	0.3	19	0.06	20	18	95
Silver	0	18	0.06		17	94
Sodium	100000	100000	98000	98	96000	96
Thallium	0	21	0.07		18	85
Vanadium	0.5	19	0.19	38	19	102
Zinc	11	29	11	102	29	100