

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51817 MA No. : SDG No.: MYE4H7
 Analytical Method: ICP-AES ICSA Source: EPA-1211
 Instrument ID: P5 ICSB Source: EPA-0710
 Run Batch: LB133243
 Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	255000	247000	240000	94	250000	99
Antimony	0	618	1.1		560	91
Arsenic	0	104	3.2		94	90
Barium	6	537	3.2	54	540	100
Beryllium	0	495	1.2		550	111
Cadmium	1	972	1.7	174	1000	107
Calcium	245000	235000	250000	101	250000	108
Chromium	52	542	50	96	600	111
Cobalt	0	476	1.4		520	109
Copper	2	511	-6.8	-341	490	97
Iron	101000	99300	100000	102	110000	106
Lead	0	49	-4.3		48	98
Magnesium	255000	248000	250000	99	260000	104
Manganese	7	507	11	159	540	106
Nickel	2	954	4.5	224	1100	110
Potassium	0	0	11		-280	
Selenium	0	46	5.6		53	115
Silver	0	201	-0.043		190	95
Sodium	0	0	220		410	
Thallium	0	108	-13		87	81
Vanadium	0	491	12		520	106
Zinc	0	952	11		950	100