

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GARD04

Lab Code: CHEM Case No.: P4528 SAS No.: P4528 SDG No.: P4528

Instrument ID: MSVOA_N Calibration Date/Time: 10/25/2024 10:04

Lab File ID: VN084504.D Init. Calib. Date(s): 10/08/2024 10/08/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:43 11:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Acrolein	0.434	0.431		-0.69	
Acrylonitrile	1.051	0.953		-9.32	
1,2-Dichloroethane-d4	2.415	2.412	0.01	-0.12	
Toluene-d8	1.390	1.390	0.01	0	
4-Bromofluorobenzene	0.495	0.511	0.2	3.23	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.