

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GARD04
 Lab Code: CHEM Case No.: P4528 SAS No.: P4528 SDG NO.: P4528
 Lab File ID: VN084504.D Date Analyzed: 10/25/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:04
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
241023062-02-VOA	0	0.00				
241023062-05-TRIP-BLANK	0	0.00				
VN1025WBL01	0	0.00				
VN1025WBS01	0	0.00				
VN1025WBSD01	0	0.00				

IS4 =

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.