

Surrogate Summary

SDG No.: **P4528**

Client: **Garden State Laboratories, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4528-01	241023068-02-VOA	1,2-Dichloroethane-d4	50	47.3	95	74	125
		Dibromofluoromethane	50	51.8	104	75	124
		Toluene-d8	50	49.5	99	86	113
		4-Bromofluorobenzene	50	44.7	89	77	121
P4528-02	241023062-05-TRIP-BLANK	1,2-Dichloroethane-d4	50	60.3	121	74	125
		Dibromofluoromethane	50	57.8	116	75	124
		Toluene-d8	50	50.1	100	86	113
		4-Bromofluorobenzene	50	45.5	91	77	121
VN1025WBL02	VN1025WBL02	1,2-Dichloroethane-d4	50	52.2	104	74	125
		Dibromofluoromethane	50	51.4	103	75	124
		Toluene-d8	50	46.5	93	86	113
		4-Bromofluorobenzene	50	43.0	86	77	121
VN1025WBS02	VN1025WBS02	1,2-Dichloroethane-d4	50	47.7	95	74	125
		Dibromofluoromethane	50	50.8	102	75	124
		Toluene-d8	50	49.7	99	86	113
		4-Bromofluorobenzene	50	51.4	103	77	121
VN1025WBSD0	VN1025WBSD02	1,2-Dichloroethane-d4	50	51.2	102	74	125
		Dibromofluoromethane	50	52.8	106	75	124
		Toluene-d8	50	50.4	101	86	113
		4-Bromofluorobenzene	50	50.8	102	77	121
VN1028WBL01	VN1028WBL01	1,2-Dichloroethane-d4	50	49.7	99	74	125
		Dibromofluoromethane	50	51.4	103	75	124
		Toluene-d8	50	49.5	99	86	113
		4-Bromofluorobenzene	50	44.9	90	77	121
VN1028WBS01	VN1028WBS01	1,2-Dichloroethane-d4	50	52.4	105	74	125
		Dibromofluoromethane	50	56.5	113	75	124
		Toluene-d8	50	53.9	108	86	113
		4-Bromofluorobenzene	50	56.4	113	77	121