

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GARD04
 Lab Code: CHEM Case No.: P4528 SAS No.: P4528 SDG NO.: P4528
 Lab File ID: VN084503.D Date Analyzed: 10/25/2024
 Instrument ID: MSVOA_N Time Analyzed: 08:59
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	182036	8.22	306519	9.10	274612	11.87
UPPER LIMIT	364072	8.724	613038	9.6	549224	12.365
LOWER LIMIT	91018	7.724	153260	8.6	137306	11.365
EPA SAMPLE NO.						
241023062-05-TRIP-BLANK	125033	8.23	236367	9.10	203498	11.87
VN1025WBL02	159284	8.22	289382	9.10	249149	11.87
VN1025WBS02	171552	8.22	293994	9.10	259694	11.87
VN1025WBSD02	162151	8.22	284141	9.10	253498	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.