

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GARD04  
 Lab Code: CHEM Case No.: P4528 SAS No.: P4528 SDG NO.: P4528  
 Lab File ID: VN084529.D Date Analyzed: 10/28/2024  
 Instrument ID: MSVOA\_N Time Analyzed: 10:05  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                  | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT # | IS3<br>AREA # | RT #   |
|------------------|---------------|-------|---------------|------|---------------|--------|
| 12 HOUR STD      | 166557        | 8.22  | 279237        | 9.10 | 255910        | 11.87  |
| UPPER LIMIT      | 333114        | 8.724 | 558474        | 9.6  | 511820        | 12.365 |
| LOWER LIMIT      | 83278.5       | 7.724 | 139619        | 8.6  | 127955        | 11.365 |
| EPA SAMPLE NO.   |               |       |               |      |               |        |
| 241023068-02-VOA | 156261        | 8.22  | 256636        | 9.10 | 218316        | 11.87  |
| VN1028WBL01      | 165769        | 8.22  | 293503        | 9.10 | 253666        | 11.87  |
| VN1028WBS01      | 157919        | 8.22  | 266847        | 9.10 | 239542        | 11.87  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.