

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: THE LEARNING CIRCLE YMCA **SDG No.:** P4539

Contract: THEL05 **Lab Code:** CHEM **Case No.:** P4539 **SAS No.:** P4539

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	109	100	109	90 - 110	P	11/06/2024	22:37	LB133345
	Lead	220	200	110	90 - 110	P	11/06/2024	22:37	LB133345

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: THE LEARNING CIRCLE YMCA **SDG No.:** P4539

Contract: THEL05 **Lab Code:** CHEM **Case No.:** P4539 **SAS No.:** P4539

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	4790	5000	96	90 - 110	P	11/06/2024	22:56	LB133345
	Lead	2430	2500	97	90 - 110	P	11/06/2024	22:56	LB133345
CCV02	Copper	4920	5000	98	90 - 110	P	11/06/2024	23:41	LB133345
	Lead	2390	2500	96	90 - 110	P	11/06/2024	23:41	LB133345
CCV03	Copper	4910	5000	98	90 - 110	P	11/07/2024	00:23	LB133345
	Lead	2400	2500	96	90 - 110	P	11/07/2024	00:23	LB133345
CCV04	Copper	4880	5000	98	90 - 110	P	11/07/2024	01:04	LB133345
	Lead	2410	2500	97	90 - 110	P	11/07/2024	01:04	LB133345
CCV05	Copper	4840	5000	97	90 - 110	P	11/07/2024	01:46	LB133345
	Lead	2400	2500	96	90 - 110	P	11/07/2024	01:46	LB133345
CCV06	Copper	4900	5000	98	90 - 110	P	11/07/2024	02:27	LB133345
	Lead	2420	2500	97	90 - 110	P	11/07/2024	02:27	LB133345
CCV07	Copper	4990	5000	100	90 - 110	P	11/07/2024	03:09	LB133345
	Lead	2430	2500	97	90 - 110	P	11/07/2024	03:09	LB133345
CCV08	Copper	4880	5000	98	90 - 110	P	11/07/2024	03:39	LB133345
	Lead	2420	2500	97	90 - 110	P	11/07/2024	03:39	LB133345