



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: THE LEARNING CIRCLE YMCA

SDG No.: P4539

Contract: THEL05

Lab Code: CHEM

Case No.: P4539

SAS No.: P4539

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Copper	1.85	2.0	92	70 - 130	P	11/06/2024	23:07	LB133345
	Lead	0.97	1.0	97	70 - 130	P	11/06/2024	23:07	LB133345