

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: THE LEARNING CIRCLE YMCA

SDG No.: P4539

Contract: THEL05

Lab Code: CHEM

Case No.: P4539

SAS No.: P4539

ICS Source: EPA

Instrument ID: P7

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	7.91	8.0	99	4	12	11/06/2024	22:44	LB133345
	Lead	3.86	4.0	96	2	6	11/06/2024	22:44	LB133345
ICSAB01	Copper	23.5	25.0	94	21	29	11/06/2024	22:52	LB133345
	Lead	22.1	25.0	88	21.3	28.8	11/06/2024	22:52	LB133345