

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>THE LEARNING CIRCLE YMCA</u>	level: <u>low</u>	sdg no.: <u>P4539</u>
contract: <u>THEL05</u>	lab code: <u>CHEM</u>	case no.: <u>P4539</u> sas no.: <u>P4539</u>
matrix: <u>Water</u>	sample id: <u>P4539-20</u>	client id: <u>INFANT-2-SINK-BMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>P4539-20MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	5580		220		5000	107		P
Lead	ug/L	70 - 130	2940		0.43	J	2500	118		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: THE LEARNING CIRCLE YMCA **level:** low **sdg no.:** P4539
contract: THEL05 **lab code:** CHEM **case no.:** P4539 **sas no.:** P4539
matrix: Water **sample id:** P4539-20 **client id:** INFANT-2-SINK-BMSD
Percent Solids for Sample: NA **Spiked ID:** P4539-20MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	5460		220		5000	105		P
Lead	ug/L	70 - 130	2890		0.43	J	2500	116		P