

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>THE LEARNING CIRCLE YMCA</u>	level: <u>low</u>	sdg no.: <u>P4540</u>
contract: <u>THEL05</u>	lab code: <u>CHEM</u>	case no.: <u>P4540</u> sas no.: <u>P4540</u>
matrix: <u>Water</u>	sample id: <u>P4540-15</u>	client id: <u>PREKINDERGARTEN-1-SINK-AMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>P4540-15MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	5470		91.2		5000	108		P
Lead	ug/L	70 - 130	2930		0.43	J	2500	117		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: THE LEARNING CIRCLE YMCA **level:** low **sdg no.:** P4540
contract: THEL05 **lab code:** CHEM **case no.:** P4540 **sas no.:** P4540
matrix: Water **sample id:** P4540-15 **client id:** PREKINDERGARTEN-1-SINK-AMSD
Percent Solids for Sample: NA **Spiked ID:** P4540-15MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	5620		91.2		5000	111		P
Lead	ug/L	70 - 130	2950		0.43	J	2500	118		P

metals
- 5a -
MATRIX SPIKE SUMMARY

client: THE LEARNING CIRCLE YMCA **level:** low **sdg no.:** P4540
contract: THEL05 **lab code:** CHEM **case no.:** P4540 **sas no.:** P4540
matrix: Water **sample id:** P4540-25 **client id:** FB-10-24-24MS
Percent Solids for Sample: NA **Spiked ID:** P4540-25MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	6110		0.89	J	5000	122		P
Lead	ug/L	70 - 130	3180		1.00	U	2500	127		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: THE LEARNING CIRCLE YMCA **level:** low **sdg no.:** P4540
contract: THEL05 **lab code:** CHEM **case no.:** P4540 **sas no.:** P4540
matrix: Water **sample id:** P4540-25 **client id:** FB-10-24-24MSD
Percent Solids for Sample: NA **Spiked ID:** P4540-25MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	70 - 130	5720		0.89	J	5000	114		P
Lead	ug/L	70 - 130	3020		1.00	U	2500	121		P