

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: THE LEARNING CIRCLE YMCA

SDG No.: P4540

Contract: THEL05

Lab Code: CHEM

Case No.: P4540

SAS No.: P4540

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164604BS							
Copper	ug/L	5000	4950		99	85 - 115	P
Lead	ug/L	2500	2470		99	85 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: THE LEARNING CIRCLE YMCA

SDG No.: P4540

Contract: THEL05

Lab Code: CHEM

Case No.: P4540

SAS No.: P4540

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164605BS							
Copper	ug/L	5000	5090		102	85 - 115	P
Lead	ug/L	2500	2480		99	85 - 115	P