

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133123

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/25/2024	09:42
2	CAL2	1	Water	NA	NA	20.2	7.00	10/25/2024	09:43
3	CAL3	1	Water	NA	NA	20.2	10.02	10/25/2024	09:45
4	ICV	1	Water	NA	NA	20.2	7.00	10/25/2024	09:50
5	CCV1	1	Water	NA	NA	20.3	2.01	10/25/2024	10:00
6	P4548-01	1	Water	NA	NA	20.3	5.81	10/25/2024	10:10
7	P4548-03	1	Water	NA	NA	20.1	6.72	10/25/2024	10:15
8	P4548-05	1	Water	NA	NA	20.1	6.73	10/25/2024	10:17
9	P4548-07	1	Water	NA	NA	20.3	6.75	10/25/2024	10:19
10	P4549-02	1	Water	NA	NA	20.7	6.97	10/25/2024	10:20
11	P4549-03	1	Water	NA	NA	20.6	7.32	10/25/2024	10:25
12	P4549-04	1	Water	NA	NA	20.2	7.28	10/25/2024	10:30
13	P4549-04DUP	1	Water	NA	NA	20.3	7.29	10/25/2024	10:31
14	CCV2	1	Water	NA	NA	20.3	12.02	10/25/2024	10:35