

Preparation Blank Summary

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: P4548

Project: Ansonia Landfill 2024

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID:	LB133119BLW						
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	10/25/2024
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	10/25/2024
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	10/25/2024
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	10/25/2024
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	10/25/2024
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	10/25/2024
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	10/25/2024
Sample ID:	LB133120BL						
Turbidity	NTU	< 0.5000	0.5000	U	0.075	1.0	10/25/2024
Sample ID:	LB133122BL						
BOD5	mg/L	< 0.2000	0.2000	U	0.17	2.0	10/25/2024
Sample ID:	LB133165BL						
TSS	mg/L	< 2.0000	2.0000	U	1	4	10/28/2024
Sample ID:	LB133166BL						
TDS	mg/L	< 5.0000	5.0000	U	1.0	10	10/25/2024
Sample ID:	LB133279BLW						
Alkalinity	mg/L	< 1.0000	1.0000	U	1	2	11/04/2024
Sample ID:	PB164383BL						
Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	10/28/2024
Sample ID:	PB164408BL						
TKN	mg/L	< 0.2500	0.2500	U	0.18	0.5	10/28/2024