

Duplicate Sample Summary

Client:	Lockwood, Kessler & Bartlett, Inc.	SDG No.:	P4548
Project:	Ansonia Landfill 2024	Sample ID:	P4548-01
Client ID:	MW-1DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Turbidity	NTU	+/-20	0.38	J	0.38	J	1	1.59		10/25/2024
BOD5	mg/L	+/-20	0.17	U	0.17	U	1	0		10/25/2024
Ammonia as N	mg/L	+/-20	0.045	U	0.045	U	1	0		10/28/2024
TKN	mg/L	+/-20	0.18	U	0.18	U	1	0		10/28/2024
Alkalinity	mg/L	+/-20	49.5		48.7		1	2		11/04/2024

Duplicate Sample Summary

Client:	Lockwood, Kessler & Bartlett, Inc.	SDG No.:	P4548
Project:	Ansonia Landfill 2024	Sample ID:	P4548-01
Client ID:	MW-1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Chloride	mg/L	+/-20	143	OR	143	OR	1	0		10/25/2024
Sulfate	mg/L	+/-20	45.2	OR	44.7	OR	1	1		10/25/2024
Bromide	mg/L	+/-20	9.90		9.40		1	5		10/25/2024
Fluoride	mg/L	+/-20	2.10		2.00		1	5		10/25/2024
Nitrate	mg/L	+/-20	3.80		3.60		1	5		10/25/2024
Orthophosphate as P	mg/L	+/-20	4.80		4.50		1	6		10/25/2024
Nitrite	mg/L	+/-20	3.00		2.80		1	7		10/25/2024
Ammonia as N	mg/L	+/-20	0.98		0.99		1	1		10/28/2024
TKN	mg/L	+/-20	4.80		4.80		1	0		10/28/2024

Duplicate Sample Summary

Client:	Lockwood, Kessler & Bartlett, Inc.	SDG No.:	P4548
Project:	Ansonia Landfill 2024	Sample ID:	P4548-07
Client ID:	MW-4DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TSS	mg/L	+/-5	5.60		5.70		1	1.77		10/28/2024
TDS	mg/L	+/-5	822		840		1	2.17		10/25/2024

Duplicate Sample Summary

Client:	Lockwood, Kessler & Bartlett, Inc.	SDG No.:	P4548
Project:	Ansonia Landfill 2024	Sample ID:	P4549-04
Client ID:	TT-069-IDWGW-20241024DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
pH	pH	+/-20	7.28		7.29		1	0.14		10/25/2024