



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Lockwood, Kessler & Bartlett, Inc.

SDG No.: P4548

Contract: LOCK01

Lab Code: CHEM

Case No.: P4548

SAS No.: P4548

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Antimony	47.7	50.0	95	40 - 160	P	10/29/2024	12:40	LB133198
	Arsenic	18.4	20.0	92	40 - 160	P	10/29/2024	12:40	LB133198
	Barium	90.0	100	90	40 - 160	P	10/29/2024	12:40	LB133198
	Beryllium	5.49	6.0	92	40 - 160	P	10/29/2024	12:40	LB133198
	Cadmium	5.32	6.0	89	40 - 160	P	10/29/2024	12:40	LB133198
	Chromium	10.2	10.0	102	40 - 160	P	10/29/2024	12:40	LB133198
	Cobalt	28.2	30.0	94	40 - 160	P	10/29/2024	12:40	LB133198
	Copper	20.6	20.0	103	40 - 160	P	10/29/2024	12:40	LB133198
	Iron	102	100	102	40 - 160	P	10/29/2024	12:40	LB133198
	Lead	9.96	12.0	83	40 - 160	P	10/29/2024	12:40	LB133198
	Manganese	18.9	20.0	94	40 - 160	P	10/29/2024	12:40	LB133198
	Nickel	37.3	40.0	93	40 - 160	P	10/29/2024	12:40	LB133198
	Potassium	1920	2000	96	40 - 160	P	10/29/2024	12:40	LB133198
	Selenium	19.2	20.0	96	40 - 160	P	10/29/2024	12:40	LB133198
	Silver	10.6	10.0	106	40 - 160	P	10/29/2024	12:40	LB133198
	Sodium	1780	2000	89	40 - 160	P	10/29/2024	12:40	LB133198
	Thallium	37.7	40.0	94	40 - 160	P	10/29/2024	12:40	LB133198
	Vanadium	36.0	40.0	90	40 - 160	P	10/29/2024	12:40	LB133198
	Zinc	38.5	40.0	96	40 - 160	P	10/29/2024	12:40	LB133198